



**HAL**  
open science

## 3D/4D ultrasound registration of bone

Jonathan Schers, Jocelyne Troccaz, Vincent Daanen, Céline Fouard,  
Christopher Plaskos, Pascal Kilian

► **To cite this version:**

Jonathan Schers, Jocelyne Troccaz, Vincent Daanen, Céline Fouard, Christopher Plaskos, et al..  
3D/4D ultrasound registration of bone. IEEE International Ultrasonic Symposium, 2007, Oct 2007,  
New-York, United States. pp.2519 - 2522, 10.1109/ULTSYM.2007.634 . hal-00207587

**HAL Id: hal-00207587**

**<https://hal.science/hal-00207587v1>**

Submitted on 17 Jan 2008

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# 3D/4D Ultrasound Registration of Bone

Jonathan Schers<sup>\*†</sup>, Jocelyne Troccaz<sup>\*</sup>, Vincent Daanen<sup>†</sup>, Céline Fouard<sup>\*</sup>, Christopher Plaskos<sup>‡</sup> and Pascal Kilian<sup>‡</sup>

<sup>\*</sup>TIMC-IMAG Laboratory, GMCAO Department

La Tronche, FRANCE

Email: Jonathan.Schers@imag.fr

<sup>†</sup>Formerly TIMC-IMAG Laboratory, GMCAO Department

La Tronche, FRANCE

<sup>‡</sup>Praxim-Medivision

La Tronche, FRANCE

**Abstract**—This paper presents a method to reduce the invasiveness of Computer Assisted Orthopaedic Surgery (CAOS) using ultrasound. In this goal, we need to develop a method for 3D/4D ultrasound registration. The preliminary results of this study suggest that the development of a robust and “realtime” 3D/4D ultrasound registration is feasible.

## I. INTRODUCTION

The Development of 3D ultrasound offers interesting prospects for surgical navigation. Indeed, this non-invasive data and “realtime” imaging technology could reduce the invasiveness of Computer Assisted Orthopaedic Surgery (CAOS) by replacing the optical trackers which, when pinned into the bones (Fig. 1), could increase the risk of infection and cause extra pain owing to extra incisions and exposed bone surfaces. This implies the development of a method for rigid registration in order to track a bony structure. This method has to be at the same time robust and “realtime”.

However, due to the speckles, the shadowing effects and the poor quality of the images, the registration of this imaging modality is a challenging process. Consequently, the research on intra-modality registration of ultrasound images is scarcely reported in the literature [1], [2], but there appears to be a greater interest in work related to the application of ultrasound images [3].

To our Knowledge there is no report on the evaluation of a method for tracking bone structures without invasive trackers.

This article is organized as follows. Section 2 presents the material and the method. Evaluation and results are presented in section 3. Section 4 concludes the article and deals with the future work.

## II. MATERIAL AND METHOD

### A. Material

Ultrasound data have been obtained with a General Electric Voluson 730 Pro, equipped with a 3D RSP 6-16 MHz probe. The size of the ultrasound volumes is  $199 \times 199 \times 199$  voxels, and the voxel size is 0.28 mm. Each volume acquisition takes almost 5s.

### B. Method

The essential idea of the method that we propose consists in producing a coarse segmentation of the interface between the




Fig. 1. Optical trackers which are rigidly pinned in to the bone during an anterior cruciate ligament surgery.

bone and the soft tissues to select a region of interest around the structures containing information.

1) *Automatic coarse segmentation*: Several types of information, based on the physics of the ultrasound imaging – (1) the bone appears to be hyper-echoic because the difference between the acoustic impedance of the bone and the soft tissues is important, (2) due to the bones have a high absorption rate and there is no imaging possible beyond them, an acoustic shadow is found behind the interface and (3) only the more or less horizontal interfaces can be seen in images because the reflexion of ultrasound is almost specular –, have been used in four different steps. The different steps have been summarized in the Fig. 3.

In a first step, the images have been binarized according to Otsu's thresholding method which allows a good approximation of the separation between the echogenic zone and the shadow zone [4].

In a second step, a horizontal sobel filter has been applied




Fig. 2. 4D ultrasound image of an anterior tibial tuberosity, i.e. 3 orthogonal planes obtained in “realtime”.

to detect interfaces. We have decided to use this filter because the bone interfaces which can be found are nearly horizontal.

In a third step, the images have been averaged to remove the noise. And, eventually, fusioning has been used relying on two conditions:

- the pixel value is greater than Otsu’s threshold;
- the pixel belongs to an interface (i.e. it is a local maximum on the ultrasound beam).

A dilatation has also been performed to increase the number of selected pixels. The results of this processing is a region of interest containing information.

2) *Registration*: The performance of using a voxel-based registration for image registration has been recognized in the literature [5]. Consequently, and because to obtain an accurate segmentation of the bone in ultrasound imaging for a surface-based registration is a challenging process, a voxel-based registration approach has been used.

The images have then been compared according to the normalised cross-correlation (NCC) similarity measure. We have chosen this similarity measure because in the literature [6], this measure seems to be the best for monomodality registration.  $\bar{A}$  represents the mean value of the image  $A$  and  $\bar{B}$  represents the mean value of the image  $B$ . The NCC is defined by as follow:




Fig. 3. Image processing to obtain the automatic coarse segmentation.

$$NCC = \frac{\sum_i (A(i) - \bar{A})(B'(i) - \bar{B}')}{\sqrt{\sum_i (A(i) - \bar{A})^2 \sum_i (B'(i) - \bar{B}')^2}}$$

Among the optimization algorithms, we have selected the simplex method of Nelder and Mead [7] for its robustness and convergence time that’s why this method is often used for image registration [8], [2], [9], [10], [11] and [6]. This method optimizes all the parameters in the same time. The simplex method of Nelder and Mead consists in updating a  $n$ -dimensional simplex such as the value of the similarity measure decrease. At each step, the simplex is modified by geometric operations. A new vertex is accepted according to the value of the similarity measure.

The size of the initial simplex in a parameter space is an important step using simplex optimization. For rigid transformation the space is 6-dimensional. Before determining the size of the initial simplex, we need to process a normalization so that the displacement in the parameters space is approximately the same as in the spatial space. This relationship is checked for translation. However, it is not true for rotation, where the displacement of a voxel is dependent on its distance from the axis rotation. In the data sets we have used in this study, a unit parameter corresponded to 1 mm of translation and  $1^\circ$  of rotation.

There are two conditions for the size of the initial simplex:

- In principle, it should be greater than the unit dimension along each axis so that it does not get stuck in a local minimum;




Fig. 4. Correlation between  $T_{error}$  and the similarity measure found during registration (60 images).

- It should not be greater than the capture range, otherwise, the convergence may not occur.

The size in each direction has been initialized between 3 and 5 units.

The initial attitude corresponds to the result of the previous optimization (for each sequence of images) if registration is successful.

### III. EVALUATION AND RESULTS

#### A. Evaluation method

In order to evaluate our method, we simulated 4D data by reslicing orthogonal 2D slices from the 3D ultrasound volumes. This simulated data were translated and rotated in the maximum range of  $\pm 10\text{mm}$  –except according to the probe’s axis  $\pm 5\text{mm}$ – /  $\pm 12^\circ$  –except according to the perpendicular axis of the probe’s axis  $\pm 6^\circ$ – which represents a known reference transformation  $T_{reference}$ . The registration of the 4D data to the reference volume provides a transformation  $T_{registration}$ . The composition of  $T_{reference}$  and  $T_{registration}$  gives a transformation error  $T_{error}$  of the  $T_{registration}$ :

$$T_{error} = T_{reference}^{-1} T_{registration}$$

From this transformation error  $T_{error}$  we extract a translation vector and 3 Euler angles which correspond to the error in each direction. It should be noted that  $T_{error}$  is correlated to the similarity measure  $\alpha$  found during registration (Fig. 4): when the error is less than 1mm and  $1^\circ$  then  $\alpha$  is closed to the absolute maximum, i.e.  $0.95 \leq \alpha \leq 1$ .

So as to assess our results, we have defined 2 criteria allowing to validate a registration:

- the similarity measure is over 0.95;
- the errors in translation and rotation are inferior to 1 mm and  $1^\circ$ .

TABLE I  
PERCENTAGE OF SUCCESS FOR EACH DATA SETS (60 IMAGES PER DATA SETS)

Data Sets	Success (%)	Total Time (s)
1	68.3	696
2	63.3	703
3	61.7	729

In the following sections, we will maintain that a registration is proved successful if these 2 criteria are established.

#### B. Results

The proposed method has been tested on ultrasound knee images. Three data sets have been used. Each set contains 60 images (i.e. 2 orthogonal slices). The table I shows the results which we have obtained. In our data, the registration that we propose is successful in almost 65% of the cases, what is encouraging for the future work. Furthermore, the average time per registration is almost 10s per images, what lets imagine that “realtime” is possible. Indeed, our code is not optimized.

### IV. CONCLUSION AND FUTURE WORK

This first evaluation of the method shows that rigid registration of 3D/4D ultrasound data works in approximately 65% of the cases in a reasonable time. The average time of a registration is approximately 10s. These initial results suggest that the development of a robust and “realtime” 3D/4D ultrasound registration method is feasible.

Current work deals with improvement of accuracy and the robustness.

We also need to evaluate and compare our method with a “gold standard”. The “gold standard” will be defined by the optical trackers pinned in to the bones.

### REFERENCES

- [1] M. M. J. Letteboer, M. A. Viergever, and W. J. Niessen, “Rigid Registration of 3D Ultrasound Data of Brain Tumours,” in *CARS*, ser. International Congress Series, H. U. Lemke, M. W. Vannier, K. Inamura, A. G. Farman, K. Doi, and J. H. C. Reiber, Eds., vol. 1256. Elsevier, 2003, pp. 433–439.
- [2] R. Shekhar and V. Zagrodsky, “Mutual Information-Based Rigid and Nonrigid Registration of Ultrasound Volumes,” *IEEE Transactions on Medical Imaging*, vol. 21, no. 1, pp. 9–22, January 2002.
- [3] J. P. W. Pluim and J. M. Fitzpatrick, “Image registration,” *IEEE Trans. Med. Imaging*, vol. 22, no. 11, pp. 1341–1343, 2003.
- [4] V. Daanen, J. Tonetti, and J. Troccaz, “An information fusion method for the automatic delineation of the bone-soft tissues interface in ultrasound images,” in *ECCV Workshops CVAMIA and MMBIA*, ser. Lecture Notes in Computer Science, M. Sonka, I. A. Kakadiaris, and J. Kybic, Eds., vol. 3117. Springer, 2004, pp. 218–229.
- [5] J. Maintz and M. Viergever, “A survey of medical image registration,” *Medical Image Analysis*, vol. 2, no. 1, pp. 1–36, 1998. [Online]. Available: [citeseer.ist.psu.edu/maintz98survey.html](http://citeseer.ist.psu.edu/maintz98survey.html)
- [6] J. Kim, J. Fessler, K.L.Lam, J. Balter, and R. T. Haken, “A Feasibility Study on Mutual Information based set-up Error Estimation for Radiotherapy,” *Medical Physics*, vol. 28, no. 12, pp. 2507–2517, 2001.
- [7] J. A. Nelder and R. Mead, “A simplex method for function minimization,” *Computer Journal*, no. 7, pp. 308–313, 1965.

- [8] P. Slomka, J. Mandel, D. Downey, and A. Fenster, "Evaluation of Voxel-based Registration of 3D Power Doppler Ultrasound and 3D Magnetic Resonance Angiographic Images of Carotid Arteries," *Ultrasound in Medicine and Biology*, vol. 27, no. 7, pp. 945–955, 2001.
- [9] P. Radau, P. Slomka, P. Julin, L. Svensson, and L. Wahlund, "Evaluation of Linear Registration Algorithms for Brain SPECT and the Errors due to Hypoperfusion Lesions," *Medical Physics*, vol. 28, no. 8, pp. 1660–1668, 2001.
- [10] F. Maes, D. Vandermeulen, and P. Suetens, "Comparative Evaluation of Multiresolution Optimization Strategies for Multimodality Image Registration by Maximization of Mutual Information," *Medical Image Analysis*, vol. 3, no. 4, pp. 373–386, December 1999.
- [11] C. Meyer, J. Boes, B. Kim, P. Bland, G. LeCarpentier, J. Fowlkes, M. Roubidoux, and P. Carson, "Semiautomatic Registration of Volumetric Ultrasound Scans," *Ultrasound in Medicine and Biology*, vol. 25, no. 3, pp. 339–347, 1999.