



HAL
open science

A novel paediatric game-based visual fields assessor

Tariq M Aslam, Waheeda Rahman, David Henson, Peng T Khaw

► **To cite this version:**

Tariq M Aslam, Waheeda Rahman, David Henson, Peng T Khaw. A novel paediatric game-based visual fields assessor. *British Journal of Ophthalmology*, 2011, 95 (7), pp.921. 10.1136/bjo.2010.198135 . hal-00628363

HAL Id: hal-00628363

<https://hal.science/hal-00628363>

Submitted on 3 Oct 2011

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

A novel paediatric game-based visual fields assessor

Tariq M Aslam^{1,2}, FRCSEd (Ophth), DM (OXON)
Waheeda Rahman¹, FRCOphth
David Henson², PhD, FBCO
Peng T Khaw¹, PhD, FRCOphth.

1. NIHR Biomedical Research Centre Moorfields Eye Hospital and UCL Institute of Ophthalmology
2. NIHR Biomedical Research Centre Manchester Royal Eye Hospital and University of Manchester

Financial support: This paper was funded by a research grant from the NIHR Biomedical Research Centre Moorfields Eye Hospital and UCL Institute of Ophthalmology and Moorfields Trustees. The funding organisation had no role in the design or conduct of this research. No conflicting relationship exists for any author.

Licence for Publication: The Corresponding Author has the right to grant on behalf of all authors and does grant on behalf of all authors, an exclusive licence (or non exclusive for government employees) on a worldwide basis to the BMJ Publishing Group Ltd to permit this article (if accepted) to be published in BJO and any other BMJPGJ products and sublicences such use and exploit all subsidiary rights, as set out in our licence (<http://group.bmj.com/products/journals/instructions-for-authors/licence-forms>).

Competing Interest: None declared.

Running Head : Game-based perimetry

Address for reprints : Mr Tariq M Aslam, Manchester Royal Eye Hospital, Oxford Road, Manchester, U.K. M13 9WL

Measurement of visual fields can be prohibitively difficult in children, particularly below the age of 10-11 years.^{1, 2} Problems include lack of central fixation,³ short attention span³⁻⁵ and intolerance. Recent systems have demonstrated user interaction^{2, 6} may improve tolerance and reliability in children and have demonstrated the potential for personal computer testing.⁷ Other recent innovations include eye tracking technology, proposed in 1989⁸ and developed at the University of Tokyo.⁹ However, this requires expensive apparatus and has not yet been validated.¹⁰

Our aim was to determine whether a computer game system could be devised to measure visual fields in children whilst they play and enjoy an entertaining game.

Methods

The study had full ethics Committee approval and informed consent was obtained from all participants and parents. The development of the game-based visual fields assessor relied on multi-disciplinary input and combined international perimetric society standards¹¹ with computer game science. The final system uses a supra-threshold examination strategy with fixed intensity. Background luminance and other psychophysical constants were set for this feasibility study based on literature review and estimations from adult trials of the test.

The software was developed using BBC basic /windows v5.80a and was conceived and written by the first author. Game design acknowledged Bateman and Boon's¹² game play types and Salen and Zimmerman¹³ and Mulligan & Patrovsky¹⁴ theories on player psychologies. We adopted the iterative game design process advocated by Salen and Zimmerman¹³ and Fullerton et al.,¹⁵ which relies on gaining feedback from players early on and repeated game redesign - our game resulted from 17 cycles of appraisal and reprogramming. Development included designing instructions for explaining how to play the 'Crazy Castle' game and a cohesive storyline that ties in all that is required of the player.

The complete game structure consists of a large wooden castle (1m³) with a set of occludable spectacle frames attached to a large central window on an exterior castle wall (fig. 1). A closed wooden drawbridge barrier obstructs view through the window into the castle. When a child's forehead is correctly aligned in the glasses (fig. 2,3) a switch activates opening of the drawbridge which then reveals a computer game on a screen (fig.3) inside the castle, at a set distance from the spectacle frames. If the correct alignment is lost, the drawbridge instantly closes, hiding the game screen from view

The child uses an attached keypad to play the game, which is separated into four stages, allowing for the child to learn to play whilst maintaining position and for the software to set individualised variables. The game premise involves a small central character of a wizard with a plunger. When a tomato hovers under this plunger, the child presses the appropriately textured console button to squash the tomato (central fixation game). This central game is moved to different corners of the screen and peripheral field targets appear intermittently as ghosts in remaining corners. The child must continue the central fixation game but is encouraged to press a separate button when he sees these peripheral ghosts (peripheral target game). These ghosts (fig 4) appear at 10, 20 and 30 degrees from the central fixation game at each of 4 axes (see fig. 4 and 5). They are presented for 200 milliseconds and increase in size if no response is provoked. Each tested peripheral target is graded 0 (not seen), 1 (5mm size seen), 2 (4mm), 3 (3mm) or 4 (2mm).¹⁶ Missed stimuli are responded to with repeat testing. In accordance with perimetric standards and computer game theory, there are multiple programming measures, animations, sound effects and graphics incorporated to optimise central fixation importance, reduce false negatives and false positives and optimise utility of the game. The examiner is able to monitor progress on a separate laptop adjacent to the castle. Once tested, the children were invited to respond to a questionnaire about their experience with the visual field exam.

Results

19 children were recruited from an outpatient department at the Richard Desmond Children's Eye Centre at Moorfields Eye Hospital. Tests were carried out in the Fight for Sight NIHR Children's Research Facility. 2 children failed to complete the exam due to hardware and software failures that were subsequently corrected.

23 fields were completed, in 23 eyes of 17 children (6 bilateral). All completed fields had physiological blind spots found. Normal fields were defined through earlier pilot studies on normal adults. Median time for completion of the tests including training stages was 4.5 minutes (range 3.11 -8.0 minutes). There were 8 girls and 9 boys, aged 4 - 14yrs (5 children aged 4-6, 8 aged 7-10 and 4 aged 10-14).

Two four year olds with expected normal fields showed general delay in response which, according to our preset definitions, was classed as abnormal. The problems for these two children included, firstly, requiring a greater time to learn the game than was currently programmed into the system. Furthermore the requirement of two separate buttons to respond in the course of the game was perhaps too complex and a single button option might be more

apt in this age group. However, it was very promising that they both still fully completed the tests, and we feel that reducing game difficulty for this age would counter the problems that arose.

For all 15 other eyes with clinically expected normal fields (all normal eyes of children aged over 4), the children were, at their first attempt, able to complete the assessment giving normal visual fields. This included 3 children aged 6 and under. For the 5 eyes with expected glaucomatous loss, all 5 visual fields measured showed abnormal fields. Discrimination potential and acceptability of test are demonstrated in Fig. 4 and Fig. 5. In questionnaires, all children enjoyed playing the game and gave positive feedback.

Discussion

The gaming system has shown great promise, in this feasibility study, to provide clinically useful information in a way that is accessible and enjoyable for the child with no discomfort or potential for harm. Developments need to be made in terms of shortfalls in hardware and software as well as in optimising the psychophysical consistency and validity of peripheral targets. Ultimately we anticipate this system being particularly useful for perimetry in targeted groups of children in terms of age, aptitude, personality and physical ability. Formal validation and further reliability studies will be required before the system can be presented as a fully validated and reliable clinical tool. The final technology would potentially have a greater diagnostic precision and sensitivity to identify visual fields loss than existing tests. The test would be quicker and more reliable and therefore be a lesser drain on resources. It would be more affordable than alternative proposals in development and more enjoyable to experience for the child than any current system. It has potential for worldwide adoption due to its basis upon widely available and affordable technology. In modified form, the game-based system could also be utilised for other visual function measurements such as acuity and contrast sensitivity.

Fig. 1 Paediatric game-based vision assessor

Fig. 2 Child waiting to play game, holding control pad then correctly positioned to see and play game

Fig. 3 Examples of images seen in game; a) Tomato correctly positioned to mash, b) Tomato being mashed, c) peripheral ghost being lasered

Fig. 4 Discrimination potential. Fields measured of 14 yr old boy with no ocular pathology in RE, but with glaucoma and suspected field defects found in his LE that corresponded with those found on Humphrey testing. Note in RE almost all targets were seen normally (score 3 or 4). In LE there are several areas of lost field (7 out of the 24 tested areas score less than 3). LE field was done after the RE

Fig. 5 Acceptability of test. This 11 yr old child had documented poor compliance with standard fields tests and none were found acceptable. Goldman fields were not successful and Humphreys fields showed multiple areas of field loss with poor fixation. On his first attempt with our game-based fields analyser, however, he returned good peripheral fields whilst demonstrating good central target fixation, in accordance with his clinically expected visual function.


Acknowledgements

This paper was funded by a research grant from the NIHR Biomedical Research Centre Moorfields Eye Hospital and UCL Institute of Ophthalmology and Moorfields Trustees. Support of the NIHR Manchester BRC is acknowledged.

We would like to acknowledge input and comments from professionals including optometrists, ophthalmologists, vision scientists and paediatric nurses over the course of this study. Young members of the public also gave key input during the development phase of the game, with particularly useful and lengthy input from Humzah Jessa and Zoha Jessa.

References

1. Wabbels BK, Wilscher S. Feasibility and outcome of automated static perimetry in children using continuous light increment perimetry (CLIP) and fast threshold strategy. *Acta Ophthalmol Scand* 2005; 83:664-9.
2. Marraffa M, Pucci V, Marchini G, Morselli S, Bellucci R, Bonomi L. HPR perimetry and Humphrey perimetry in glaucomatous children. *Doc Ophthalmol* 1995; 89:383-6.
3. Tschopp C, Safran AB, Viviani P, Bullinger A, Reicherts M, Mermoud C. Automated visual field examination in children aged 5-8 years. Part I: Experimental validation of a testing procedure. *Vision Res* 1998; 38:2203-10.
4. Blumenthal EZ, Haddad A, Horani A, Anteby I. The reliability of frequency-doubling perimetry in young children. *Ophthalmology* 2004; 111:435-9.
5. Morales J, Brown SM. The feasibility of short automated static perimetry in children. *Ophthalmology* 2001; 108:157-62.
6. Martin L. Rarebit and frequency-doubling technology perimetry in children and young adults. *Acta Ophthalmol Scand* 2005; 83:670-7.
7. Damato B, Groenewald C. Multifixation campimetry on line: a perimeter for the detection of visual field loss using the internet. *Br J Ophthalmol* 2003; 87:1296-8.
8. Trope GE, Eizenman M, Coyle E. Eye movement perimetry in glaucoma. *Can J Ophthalmol* 1989; 24:197-9.
9. Nunokawa K, Ideguchi N, Nakano Y. The influence of fixation on new visual field measurement using saccadic eye movement. *International Congress Series Vision 2005 - Proceedings of the International Congress held between 4 and 7 April 2005 in London, UK 2005*; 1282:674-678.
10. Murray IC, Fleck BW, Brash HM, Macrae ME, Tan LL, Minns RA. Feasibility of saccadic vector optokinetic perimetry: a method of automated static perimetry for children using eye tracking. *Ophthalmology* 2009; 116:2017-26.
11. Imaging.and.Perimetry.Society. *IPS Standards and guidelines 2010*, 2010.
12. Bateman CaB, R. *21st Century Game Design*. Hingham, Mass.: Charles River Media, 2006.
13. Salen K, Zimmerman, E. *Rules of Play: Game Design Fundamentals*. Cambridge, MA.: MIT Press, 2003.
14. Mulligan J, Patrovsky, B. *Developing Online Games: An Insider's Guide*: New Riders Publishing, 2003.
15. Fullerton T, Swain, C., and Hoffman, S. *Games Design Workshop: Designing , Prototypingand Playtesting games*. San Francisco: CMP Books, 2004.
16. Henson D. *Visual Fields*. Oxford: Oxford Medical Publications, 1993.









fig 3(a)



fig 3(b)




fig 3(c)

3			4			2
	3		3		3	
		3	1	3		
3	2	3	x	4	2	4
		3	3	3		
	4		2		3	
3			1			1

Left eye (glaucoma)

3			2			4
	4		4		4	
		4	4	3		
4	3	3	x	4	4	4
		3	3	3		
	4		4		4	
3			3			4

Right eye (normal)

4			4			4
	3		4		4	
		4	4	4		
4	4	4	X	04	4	4
		4	4	4		
	4		3		4	
2			4			4