

Evaluation of CHROMagarTM KPC for the detection of carbapenemase-producing Enterobacteriaceae in rectal surveillance cultures

Theofano Panagea, Irene Galani, Maria Souli, Panagiota Adamou, Anastasia Antoniadou, Helen Giamarellou

▶ To cite this version:

Theofano Panagea, Irene Galani, Maria Souli, Panagiota Adamou, Anastasia Antoniadou, et al.. Evaluation of CHROMagarTM KPC for the detection of carbapenemase-producing Enterobacteriaceae in rectal surveillance cultures. International Journal of Antimicrobial Agents, 2011, 37 (2), pp.124. 10.1016/j.ijantimicag.2010.10.010. hal-00659904

HAL Id: hal-00659904 https://hal.science/hal-00659904

Submitted on 14 Jan2012

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Accepted Manuscript

Title: Evaluation of CHROMagarTM KPC for the detection of carbapenemase-producing Enterobacteriaceae in rectal surveillance cultures

Authors: Theofano Panagea, Irene Galani, Maria Souli, Panagiota Adamou, Anastasia Antoniadou, Helen Giamarellou



PII: DOI: Reference:	S0924-8579(10)(doi:10.1016/j.ijan ANTAGE 3450	00457-7 ntimicag.201	0.10.0	10	
To appear in:	International	Journal	of	Antimicrobial	Agents
Received date: Accepted date:	23-9-2010 1-10-2010				

Please cite this article as: Panagea T, Galani I, Souli M, Adamou P, Antoniadou A, Giamarellou H, Evaluation of CHROMagarTM KPC for the detection of carbapenemase-producing Enterobacteriaceae in rectal surveillance cultures, *International Journal of Antimicrobial Agents* (2010), doi:10.1016/j.ijantimicag.2010.10.010

This is a PDF file of an unedited manuscript that has been accepted for publication. As a service to our customers we are providing this early version of the manuscript. The manuscript will undergo copyediting, typesetting, and review of the resulting proof before it is published in its final form. Please note that during the production process errors may be discovered which could affect the content, and all legal disclaimers that apply to the journal pertain.

Evaluation of CHROMagar[™] KPC for the detection of carbapenemase-producing Enterobacteriaceae in rectal surveillance cultures

Theofano Panagea *, Irene Galani, Maria Souli, Panagiota Adamou, Anastasia Antoniadou, Helen Giamarellou¹

4th Department of Internal Medicine, Athens University School of Medicine, University General Hospital ATTIKON, Chaidari, Greece

ARTICLE INFO

Article history:

Received 23 September 2010

Accepted 1 October 2010

Keywords:

Imipenem-supplemented MacConkey agar

KPC detection

VIM detection

Klebsiella pneumoniae

Enterobacter spp.

* Corresponding author. Present address: 'Attikon' Hospital, Rimini 1, 12 462 Chaidari, Athens, Greece. Tel.: +30 210 583 1986; fax: +30 210 532 6446. *E-mail address*: fpanagea@yahoo.gr (T. Panagea).

¹ Present address: 6th Department of Internal Medicine, Diagnostic and Therapeutic Center of Athens 'Hygeia', 4 Erythrou Stavrou Str. and Kifissias Ave., 151 23 Maroussi, Greece.

> بر ر

ABSTRACT

In this study, the performance of the chromogenic medium CHROMagar[™] KPC was evaluated and was compared with in-house daily-prepared MacConkey agar plates supplemented with imipenem (1 mg/L) for the detection of carbapenemase-producing Enterobacteriaceae. In this surveillance study, rectal swabs were cultured on both media and polymerase chain reaction (PCR) for *bla*_{KPC} and *bla*_{VIM} was used to confirm the genotype of growing colonies of Enterobacteriaceae. CHROMagar KPC was also tested with 17 genotypically characterised carbapenemase-producing and non-producing Gram-negative bacteria. It was shown that CHROMagar allows rapid detection of carbapenemase-producing Enterobacteriaceae, although bla_{KPC}- and *bla*_{VIM}-harbouring isolates could not be differentiated by colour or colony morphology. The positive and negative predictive values of the tested methods for the detection of carbapenemase-producing Enterobacteriaceae were, respectively, 100% and 98.8% for CHROMagar KPC and 94.7% and 88.6% for imipenem-supplemented MacConkey agar. CHROMagar KPC medium is a useful screening medium for carbapenemase-producing Enterobacteriaceae in stools in settings with a high proportion of patients colonised with a variety of carbapenemase-producers.

1. Introduction

Infections caused by carbapenem-resistant Enterobacteriaceae are an emerging problem associated with high rates of morbidity and mortality, particularly among critically ill patients [1,2]. Carbapenem-resistant Enterobacteriaceae are usually resistant not only to β -lactam antimicrobials but also to most other classes of antimicrobial agents [2].

The most important mechanism of resistance to carbapenems is carbapenemase production. KPC and VIM are the most prevalent carbapenemases produced by Enterobacteriaceae not only in Greece but also in several other countries worldwide [3–8].

Patients colonised with carbapenem-resistant Enterobacteriaceae are thought to be the source of transmission in healthcare settings [7,9]. Surveillance cultures are useful in identifying such patients in order to implement infection control measures rapidly. Accurate detection of colonisation with carbapenem-resistant Enterobacteriaceae within a short time from sampling contributes to the effectiveness of the above measures as well as to the adequacy of the prescribed empirical antimicrobial treatment in severely ill patients [10,11].

Commercially prepared media for isolation of vancomycin-resistant enterococci and meticillin-resistant *Staphylococcus aureus* (MRSA) have facilitated cultivation techniques for rapid evaluation of gastrointestinal colonisation. Until recently, detection of carbapenem-resistant Enterobacteriaceae has been troublesome and was performed using in-house prepared selective media such as agar or tryptic soy

broth containing a 10 μ g disk of imipenem, meropenem or ertapenem [9,11,12] or with polymerase chain reaction (PCR)-based techniques [13], each with its own advantages and disadvantages. CHROMagarTM KPC (Hy-Labs, Rehovot, Israel) is a commercially prepared chromogenic solid medium supplemented with agents that inhibit the growth of carbapenem-sensitive bacteria. Following 24 h of incubation, carbapenem-resistant Enterobacteriaceae colonies appear with different colours according to their specific enzymatic properties: *Escherichia coli* appear as red colonies, *Klebsiella* spp., *Enterobacter* spp. and *Citrobacter* spp. as metallic blue and *Pseudomonas* spp. as translucent cream colonies [14].

In the Intensive Care Unit (ICU) of University General Hospital 'Attikon' (Athens, Greece), in order to achieve early detection of carbapenem-resistant Gram-negative bacteria, surveillance cultures (rectal swabs and bronchial secretions) are performed twice weekly in all ICU patients. Samples are plated on in-house prepared MacConkey agar supplemented with antibiotics. In the present study, the performance of CHROMagar KPC was compared with that of the in-house daily-prepared MacConkey agar plates supplemented with imipenem for the detection of carbapenem-resistant Enterobacteriaceae.

2. Material and methods

2.1. Preliminary studies

To evaluate the reliability of the new CHROMagar KPC and its applicability for detecting carbapenem-resistant Gram-negative pathogens harbouring different resistance mechanisms, 17 genotypically characterised clinical isolates were studied,

including 12 carbapenemase-producing Enterobacteriaceae, 3 carbapenemaseproducing non-fermenters and 2 meropenem-resistant non-carbapenemase producing *Pseudomonas aeruginosa* isolates (Table 1). The colour and morphological characteristics of the colonies grown on CHROMagar KPC were recorded after 24 h and 48 h of incubation in ambient air at 35 °C.

2.2. Comparative evaluation of screening methods

Rectal swabs from patients hospitalised in four different tertiary hospitals in Athens, Greece, during February 2009 were promptly placed in Stuart transport medium (Cultiplast; LPT Italiana SpA, Milan, Italy) and were transported within 12 h to the Clinical Microbiology and Infectious Diseases Laboratory of the 4th Department of Internal Medicine, University General Hospital 'Attikon'. Swabs were then plated sequentially both on commercially prepared CHROMagar KPC and on in-house daily-prepared MacConkey agar plates (Becton-Dickinson, Cockeysville, MD) supplemented with 1 mg/L imipenem. To ensure equal inoculation, each swab was plated alternately on both media. Plates were incubated at 35 °C in ambient air and were examined for growth at 24 h and 48 h. All morphologically different colonies from imipenem-supplemented MacConkey agar and all differently coloured colonies (blue, red or with a brown halo according to the manufacturer) from CHROMagar KPC were subcultured on MacConkey agar and were submitted to identification and susceptibility testing using a BD Phoenix automated microbiology system (Becton Dickinson Diagnostic Systems, Sparks, MD). Metallo-β-lactamase and KPC production were screened using the ethylene diamine tetra-acetic acid (EDTA)imipenem approximation disk synergy test and the boronic acid disk test, respectively

[18,19]. The presence of bla_{KPC} and bla_{VIM} was confirmed by PCR with specific primers [15,16].

2.3. Detection limit

Four genotypically characterised clinical strains of carbapenem-resistant *Klebsiella pneumoniae* [imipenem minimum inhibitory concentration (MIC) \geq 16 mg/L] known to possess *bla*_{KPC-2} or *bla*_{VIM-1} were included in the experiments to assess the detection limit of the CHROMagar KPC screening method. Starting with an initial inoculum of ca. 5 × 10⁷ colony-forming units (CFU)/mL, serial 10-fold dilutions of the four *K*. *pneumoniae* isolates were prepared in normal saline. Viable cells were counted by the colony count technique according to Clinical and Laboratory Standards Institute (CLSI) methodology [20].

To assess possible interference from other organisms that can inhabit the gastrointestinal tract, the following organisms were added to each dilution (ca. 5×10^7 CFU/mL): a clinical isolate of imipenem-resistant *P. aeruginosa* harbouring *bla*_{VIM-}₁; a clinical isolate of extended-spectrum β-lactamase-producing *K. pneumoniae*; and *E. coli* ATCC 35218. Then, 100 µL of each of the culture mixtures was plated on CHROMagar KPC and incubated for 24–48 h for colony counts; recovery of steel blue colonies was recorded. The detection limit of CHROMagar KPC was the lowest concentration of the isolate that resulted in recovery of steel blue colonies.

2.4. Analysis of the results

Swabs positive for any carbapenemase-producing Enterobacteriaceae by both methods (red or metallic blue colonies on CHROMagar KPC and growth on imipenem-supplemented MacConkey agar and a positive PCR for bla_{VIM} and/or bla_{KPC}) were characterised as 'true positive' and those negative by both methods (no red or metallic blue colonies on CHROMagar KPC and no growth of Enterobacteriaceae on imipenem-supplemented MacConkey agar) were characterised as 'true negative'. All swabs yielding discrepant results between the two testing methods were finally characterised as 'true positive' or 'true negative' with regard to the presence of carbapenemase-producing Enterobacteriaceae by PCR for bla_{VIM} and bla_{KPC} , as described above (Fig. 1). The positive predictive value (PPV) and negative predictive value (NPV) of both methods was calculated.

3. Results

3.1. Preliminary studies

All tested strains grew on CHROMagar KPC. *Klebsiella pneumoniae*, *Enterobacter cloacae* and *Enterobacter aerogenes* harbouring *bla*_{VIM} and/or *bla*_{KPC} yielded large steel blue colonies at 24 h. *bla*_{VIM}- or *bla*_{KPC}-harbouring *E. coli* and *Citrobacter freundii* grew as pink colonies or pink with darker centre colonies, respectively, whereas *Proteus mirabilis* grew as colonies with a brown halo. The two types of carbapenemases (VIM or KPC) could not be differentiated by the colony morphology of producing strains (Table 1). The single strain of *E. cloacae* harbouring both carbapenemases was recently isolated from rectal surveillance specimens of two ICU patients (unpublished data).

3.2. Comparative evaluation of screening methods

In total, 126 rectal swabs from 121 patients hospitalised in four different tertiary hospitals in Athens during February 2009 were examined. A second sample collected during a 15-day period from five patients was included in the study because of changes in the recovered microbial flora. Of the 121 patients, 79 (65.3%) were in the ICU and 42 (34.7%) were in internal medicine and surgical wards.

3.2.1. Imipenem-supplemented MacConkey agar

Among 126 samples cultured, 38 (30.2%) yielded at least one strain of the family Enterobacteriaceae, 66 samples (52.4%) yielded no growth (Fig. 1) and the remaining 22 samples (17.5%) grew only non-Enterobacteriaceae. In total, 93 isolates were recovered from imipenem-supplemented MacConkey agar. Of 44 Enterobacteriaceae strains, 4 were imipenem- and meropenem-susceptible [21] and were identified as *E. coli* negative for *bla*_{VIM} or *bla*_{KPC} by PCR (2 of them were present in the sample with carbapenemase-producing Enterobacteriaceae). The other 40 isolates were non-susceptible to imipenem and/or meropenem and were identified as KPC-producing *K. pneumoniae* (n = 30), VIM-producing *K. pneumoniae* (n = 7), KPC-producing *Enterobacter* spp. (n = 2; both coexisted in the sample with KPC-producing *K. pneumoniae*) and VIM-producing *P. mirabilis* (n = 1). Among 49 non-fermenting Gram-negative bacilli, 9 were found to be imipenem-susceptible and were identified as *P. aeruginosa*; the remaining 40 carbapenem-non-susceptible isolates were identified as *P. aeruginosa* or *Acinetobacter baumannii*.

3.2.2. CHROMagar KPC

Metallic blue colonies were obtained from 45 samples (Fig. 1). In 10 rectal samples, CHROMagar KPC detected carbapenem-resistant *K. pneumoniae* harbouring *bla*_{KPC} (*n* = 6) or *bla*_{VIM} (*n* = 4) that the in-house prepared imipenem-supplemented MacConkey agar plate failed to detect (Fig. 1). In total, 47 isolates were recovered, all identified as *K. pneumoniae*. Forty-six were imipenem- and/or meropenemresistant and one was imipenem-intermediate (MIC = 8 mg/L) and meropenemsusceptible (MIC = 2 mg/L). Thirty-five isolates (74.5%) harboured *bla*_{KPC} and 12 isolates (25.5%) harboured *bla*_{VIM}. White and light green colonies were also recovered from 35 samples but were not tested further as they were thought to be non-Enterobacteriaceae.

Concordant results between the two methods were documented for 113 specimens (89.7%), whereas discordant results were recorded for 13 specimens (10.3%). The PPV and NPV of the tested methods for the detection of carbapenemase-producing Enterobacteriaceae were, respectively, 100% and 98.8% for CHROMagar KPC and 94.7% and 88.6% for imipenem-supplemented MacConkey agar (Fig. 1).

Moreover, 42 (91.3%) of the 46 true positive samples were recorded in ICU patients.

All carbapenem-non-susceptible *K. pneumoniae* strains were isolated after 24 h of incubation on either agar plate, except for two isolates that grew the first day on CHROMagar KPC and the second day on imipenem-supplemented MacConkey agar.

Klebsiella pneumoniae isolates harbouring bla_{KPC} were 100% non-susceptible to imipenem and 94.6% non-susceptible to meropenem, whilst those harbouring bla_{VIM} were 92.9% non-susceptible to imipenem and 50% non-susceptible to meropenem according to CLSI breakpoints [21] (Table 2).

3.3. Detection limit experiment

The two KPC-2-producing *K. pneumoniae* isolates were detected as metallic blue colonies on CHROMagar KPC with a detection limit of 4×10^{1} – 9×10^{2} CFU/mL, whilst for the two VIM-1-producing *K. pneumoniae* strains the detection limit was 7×10^{2} – 5×10^{3} CFU/mL.

4. Discussion

Rapid detection of antibiotic-resistant nosocomial pathogens in the gastrointestinal tract of patients is considered a necessary step in successful infection control protocols [2,22]. It provides useful information on the prevalence of colonisation by multiresistant pathogens and dictates strict implementation of additional infection control measures. Obtaining results in a short time interval and taking immediate action in terms of infection control prevents further dissemination of resistant strains as well as predicting the real pathogen in the case of ventilator-associated pneumonia or bacteraemia [10].

In the present study, CHROMagar KPC medium was evaluated for its performance in detecting 17 genotypically characterised clinical isolates. It was shown that all tested Gram-negative bacteria grew on CHROMagar KPC and it was possible to detect

KPC and/or VIM carbapenemase-producing Enterobacteriaceae in just 24 h. Nevertheless, differentiation between bla_{KPC} - and bla_{VIM} -harbouring Enterobacteriaceae was not possible by colour or colony morphology. Likewise, differentiation of carbapenemase-producing *P. aeruginosa* from meropenem-nonsusceptible strains not producing a carbapenemase was not possible. *Citrobacter freundii* and *E. coli* strains with low MICs to carbapenems (intermediate or susceptible) showed poor growth even after 48 h of incubation and were less efficiently detected (Table 1).

In this surveillance study conducted in four institutions with a high percentage of patients colonised with bla_{KPC} - and/or bla_{VIM} -positive Enterobacteriaceae (true positive specimens, 36.5%), CHROMagar KPC showed an excellent PPV and NPV of 100% and 98.8%, respectively, in detecting carbapenemase-producing Enterobacteriaceae compared with in-house prepared medium that showed a PPV of 94.7% and a NPV of 88.6%. The lower NPV of MacConkey agar necessitates confirmatory PCR testing and final results are available \geq 48 h after sampling.

It should be noted that with the use of imipenem-supplemented MacConkey agar it was possible to distinguish a variety of carbapenemase-producing species of Enterobacteriaceae growing simultaneously on the plate owing to the different colony morphologies, whereas using CHROMagar KPC carbapenemase-producing *K. pneumoniae* and carbapenemase-producing *Enterobacter* spp. present in the same sample were indistinguishable. *Klebsiella pneumoniae* bla_{KPC}- or bla_{VIM}-harbouring isolates could not be differentiated by colour or colony morphology in either of the tested media. For these reasons, CHROMagar KPC did not detect 2 of the 37 KPC-

producing and 2 of the 14 VIM-producing *K. pneumoniae* strains that were isolated from imipenem-supplemented MacConkey agar because they coexisted with other carbapenemase-producers. CHROMagar failed to detect one *bla*_{VIM}-positive *P. mirabilis* isolate that was recovered on imipenem-supplemented MacConkey agar with imipenem and meropenem MICs of \geq 16 mg/L since *P. mirabilis* colonies producing a brown halo could be missed in a heavy growth of blue and white colonies.

In this study, the detection limit was determined only for CHROMagar KPC, which was able to detect low concentrations of KPC-2- and VIM-1-producing *K*. *pneumoniae* in stools (4×10^{1} and 7×10^{2} CFU/mL, respectively).

Until now, detection of carbapenem-resistant Enterobacteriaceae in stools has been mainly performed with in-house prepared media and the use of carbapenems either in disks on the surface of the agar or diluted in the agar or with PCR-based techniques. The instability of β -lactam antimicrobials added to the medium is a problem that needs to be overcome. For this purpose, daily fresh prepared imipenem-supplemented MacConkey agar plates were used in University General Hospital 'Attikon'. By this method, confirmed results were available in 48 h. The use of disks on the surface of the agar in institutions where carbapenem-resistant microorganisms are endemic causes further delay, since crowding of colonies close to the carbapenem disk necessitates subcultures in order to isolate and test growing colonies.

The US Centers for Disease Control and Prevention (CDC) recently proposed a protocol for the detection of carbapenem-resistant *Klebsiella* spp. and *E. coli* in stool, according to which samples are first inoculated in trypticase soy broth with a meropenem or ertapenem disk, incubated overnight and subsequently cultured on MacConkey agar. Lactose-fermenting colonies are then screened for carbapenemase production using a phenotypic test such as the modified Hodge test [11]. This method has been shown to detect carbapenem-resistant *K. pneumoniae* KPC-2 in concentrations ranging between 1.7×10^6 to 2.7×10^0 , depending on the imipenem MIC of the strain [12]. Results by this method are available after not less than 3–4 days, therefore delaying the implementation of appropriate infection control measures.

Samra et al. [23] compared CHROMagar KPC to MacConkey agar with carbapenem disks and direct PCR for *bla*_{KPC} for rapid detection of carbapenem-resistant KPCproducing Enterobacteriaceae from rectal swabs. The sensitivity and specificity relative to PCR were 100% and 98.4%, respectively, for CHROMagar KPC and 92.7% and 95.9%, respectively, for MacConkey agar. Imipenem-supplemented MacConkey agar was evaluated by Schechner et al. [13] for the detection of KPCproducing Enterobacteriaceae from rectal swabs in comparison with PCR for *bla*_{KPC}. The sensitivity and specificity of that medium were 87.5% and 99.4%, respectively. PCR showed higher sensitivity and specificity and provided confirmed results sooner [13]. Nevertheless, PCR is an expensive method to use routinely and requires qualified personnel. In Greece, commercially available PCR kits detecting KPC and VIM carbapenemases cost ca. €40 per sample, whilst CHROMagar and imipenemsupplemented MacConkey agar plates cost €8 and €0.40, respectively.

To our knowledge, this is the first study evaluating the performance of CHROMagar KPC medium in the detection of VIM-producing Enterobacteriaceae in rectal samples. A potential limitation of this study is the absence of any reference criterion by which samples may be characterised as truly negative. Therefore, the true NPV may be lower than reported. The reliability and applicability for the detection of other carbapenem-resistant Gram-negative bacilli, such as *Acinetobacter* spp. and *P. aeruginosa*, need to be evaluated for both of the studied media.

In conclusion, CHROMagar KPC medium is a very useful screening medium both for KPC and VIM carbapenemase-producing Enterobacteriaceae in stools. Taking into account that the time to detection of resistance is crucial in infection control policies, CHROMagar KPC, being much cheaper than PCR, efficiently identifies colonised patients in a much shorter time compared with imipenem-supplemented MacConkey agar, thus permitting immediate implementation of infection control measures to prevent further dissemination, and in the case of infection directs therapy away from β -lactam antibiotics.

Funding

None.

Competing interests

None declared.

Ethical approval

Not required.

References

- [1] Schwaber MJ, Carmeli Y. Carbapenem-resistant Enterobacteriaceae: a potential threat. JAMA 2008;300:2911–3.
- [2] Centers for Disease Control and Prevention (CDC). Guidance for control of infections with carbapenem-resistant or carbapenemase-producing Enterobacteriaceae in acute care facilities. MMWR Morb Mort Wkly Rep 2009;58:256–60.
- [3] Giakoupi P, Maltezou H, Polemis M, Pappa O, Saroglou G, Vatopoulos A; Greek System for the Surveillance of Antimicrobial Resistance. KPC-2-producing *Klebsiella pneumoniae* infections in Greek hospitals are mainly due to a hyperepidemic clone. Euro Surveill 2009;14:pii: 19218.
- [4] Vatopoulos A. High rates of metallo-β-lactamase-producing *Klebsiella* pneumoniae in Greece—a review of the current evidence. Euro Surveill 2008;13:pii: 8023.
- [5] Walsh TR. Clinically significant carbapenemases: an update. Curr Opin Infect Dis 2008;21:367–71.
- [6] Queenan AM, Bush K. Carbapenemases: the versatile β-lactamases. Clin Microbiol Rev 2007;20:440–58.
- [7] Samra Z, Ofir O, Lishtzinsky Y, Madar-Shapiro L, Bishara J. Outbreak of carbapenem-resistant *Klebsiella pneumoniae* producing KPC-3 in a tertiary medical centre in Israel. Int J Antimicrob Agents 2007;30:525–9.
- [8] Souli M, Galani I, Antoniadou A, Papadomichelakis E, Poulakou G, Panagea T, et al. An outbreak of infection due to β-lactamase *Klebsiella pneumoniae* carbapenemase 2-producing *K. pneumoniae* in a Greek university hospital:

molecular characterization, epidemiology, and outcomes. Clin Infect Dis 2010;50:364–73.

- [9] Calfee D, Jenkins SG. Use of active surveillance cultures to detect asymptomatic colonization with carbapenem-resistant *Klebsiella pneumoniae* in intensive care unit patients. Infect Control Hosp Epidemiol 2008;29:966–8.
- [10] Papadomichelakis E, Kontopidou F, Antoniadou A, Poulakou G, Koratzanis E, Kopterides P, et al. Screening for resistant Gram-negative microorganisms to guide empiric therapy of subsequent infection. Intensive Care Med 2008;34:2169– 75.
- [11] US Centers for Disease Control and Prevention (CDC), Department of Health and Human Services. Laboratory protocol for detection of carbapenem-resistant or carbapenemase-producing, *Klebsiella* spp. and *E. coli* from rectal swabs. http://www.cdc.gov/ncidod/dhqp/pdf/ar/Klebsiella_or_Ecoli.pdf [accessed 27 June 2010].
- [12] Landman D, Salvani JK, Bratu S, Quale J. Evaluation of techniques for detection of carbapenem-resistant *Klebsiella pneumoniae* in stool surveillance cultures. J Clin Microbiol 2005;43:5639–41.
- [13] Schechner V, Straus-Robinson K, Schwartz D, Pfeffer I, Tarabeia J, Moskovich R, et al. Evaluation of PCR-based testing for surveillance of KPCproducing carbapenem-resistant members of the Enterobacteriaceae family. J Clin Microbiol 2009;47:3261–5.
- [14] CROMagar[™] KPC for detection of Gram-negative bacteria with a reduced susceptibility to most of the carbapenem agents. http://www.chromagar.com/ [accessed 27 June 2010].

- [15] Bradford PA, Bratu S, Urban C, Visalli M, Mariano N, Landman D, et al. Emergence of carbapenem-resistant *Klebsiella* species possessing the class A carbapenem-hydrolyzing KPC-2 and inhibitor-resistant TEM-30 β-lactamases in New York City. Clin Infect Dis 2004;39:55–60.
- [16] Poirel L, Naas T, Nicolas D, Collet L, Bellais S, Cavallo JD, et al. Characterization of VIM-2, a carbapenem-hydrolyzing metallo-β-lactamase and its plasmid- and integron-borne gene from a *Pseudomonas aeruginosa* clinical isolate in France. Antimicrob Agents Chemother 2000;44:891–7.
- [17] Woodford N, Ellington MJ, Coelho JM, Turton JF, Ward ME, Brown S, et al. Multiplex PCR for genes encoding prevalent OXA carbapenemases in *Acinetobacter* spp. Int J Antimicrob Agents 2006;27:351–3.
- [18] Galani I, Rekatsina PD, Hatzaki D, Plachouras D, Souli M, Giamarellou H. Evaluation of different laboratory tests for the detection of metallo-β-lactamase production in Enterobacteriaceae. J Antimicrob Chemother 2008;61:548–53.
- [19] Tsakris A, Kristo I, Poulou A, Themeli-Digalaki K, Ikonomidis A, Petropoulou
 D, et al. Evaluation of boronic acid disk tests for differentiating KPC-possessing
 Klebsiella pneumoniae in the clinical laboratory. J Clin Microbiol 2009;47:362–7.
- [20] Clinical and Laboratory Standards Institute. Methods for dilution antimicrobial susceptibility tests for bacteria that grow aerobically; approved standard—eighth edition. Document M07-A8. Wayne, PA: CLSI; 2009.
- [21] Clinical and Laboratory Standards Institute. Performance standards for antimicrobial susceptibility testing; twentieth informational supplement. Document M100-S20. Wayne, PA: CLSI; 2010.

- [22] Lucet JC, Decre D, Fichelle A, Joly-Guillou ML, Pernet M, Deblangy C, et al. Control of a prolonged outbreak of extended-spectrum β-lactamase-producing Enterobacteriaceae in a university hospital. Clin Infect Dis 1999;29:1411–8.
- [23] Samra Z, Bahar J, Madar-Shapiro L, Aziz N, Israel S, Bishara J. Evaluation of CHROMagar KPC for rapid detection of carbapenem-resistant Enterobacteriaceae. J Clin Microbiol 2008;46:3110–1.

Fig. 1. Processing of 126 rectal swabs by culture on imipenem-containing MacConkey agar (IMP-McConkey agar) and CHROMagarTM KPC for the detection of carbapenemase-producing Enterobacteriaceae, and final interpretation of the results by polymerase chain reaction (PCR) using bla_{KPC} - and bla_{VIM} -specific primers.

Table 1

Colony morphology of genotypically characterised carbapenem-resistant strains on

CHROMagar[™] KPC

Microorganism	Genotype ^a	MIC (mg/L)			Colony morphology
		IPM	MEM	ERT	
Escherichia coli	bla _{∨ıм}	8	1	1	Poor growth, small pink colonies
E. coli	<i>bla</i> _{KPC}	>8	>8	>4	Pink colonies
Klebsiella pneumoniae	<i>Ыа</i> крс	>8	>8	>4	Steel blue colonies
K. pneumoniae	<i>bla</i> _{KPC}	>8	>8	>4	Steel blue colonies
K. pneumoniae	bla _{∨IM}	>8	8	>4	Steel blue colonies
Enterobacter cloacae	bla _{∨IM}	2	>8	>4	Steel blue colonies
E. cloacae	<i>bla_{vim}</i> and <i>bla_{кPC}</i>	>8	8	4	Steel blue colonies
Enterobacter aerogenes	bla _{KPC}	>8	>8	>4	Steel blue colonies
E. aerogenes	bla _{KPC}	>8	>8	>4	Steel blue colonies
Citrobacter freundii	bla _{VIM}	4	≤1	2	Poor growth, small pink colonies with dark centre
Proteus mirabilis	bla _{VIM}	>8	>8	>4	Colonies with brown halo
P. mirabilis	<i>bla</i> _{VIM}	4	4	0.5	Colonies with brown halo
Pseudomonas aeruginosa	bla _{VIM}	>8	>8	N/D	White colonies
P. aeruginosa	None	4	>8	N/D	White colonies
P. aeruginosa	None	4	>8	N/D	White colonies
Acinetobacter	bla _{∨IM}	>8	>8	>8	White colonies
baumannii					
A. baumannii	bla _{OXA-58}	>8	>8	>8	White colonies

MIC, minimum inhibitory concentration; IPM, imipenem; MEM, meropenem; ERT,

ertapenem; N/D, not determined.

^a Genotype was confirmed by polymerase chain reaction (PCR) using specific primers and sequencing [15–17].

، میں ب

Table 2

Recovery and minimum inhibitory concentration (MIC) distribution of Klebsiella pneumoniae strains isolated using imipenem-

supplemented MacConkey agar and CHROMagarTM KPC

Isolates	Recovered on:	No. of	MIC (mg/L)									
		strains	Imipenem (no. of			Meropenem (no. of						
			strains)			strains)						
			≤1	2	4	8	≥16	≤1	2	4	8	≥16
KPC-producing K.	Both plates	28	_	_	_	4	24	_	_	_	_	28
pneumoniae (N = 37)	Imipenem-supplemented	2	_	_	_	_	2	_	_	-	_	2
	MacConkey agar only											
	CHROMagar KPC only	7	_	_	_	2	5	_	2	_	_	5
VIM-producing K.	Both plates	5	_	_	1	1	3	_	_	1	1	3
pneumoniae (N = 14)	Imipenem-supplemented	2	_	_	_	_	2	_	_	_	_	2
	MacConkey agar only											
	CHROMagar KPC only	7	-	—	_	—	7	4	1	1	1	_

