



**HAL**  
open science

## **Large differences between test strategies for the detection of anti- antibodies are revealed by comparing eight ELISAs and five immunoblots**

C. W. Ang, D. W. Notermans, M. Hommes, A. M. Simoons-Smit, T. Herremans

### **► To cite this version:**

C. W. Ang, D. W. Notermans, M. Hommes, A. M. Simoons-Smit, T. Herremans. Large differences between test strategies for the detection of anti- antibodies are revealed by comparing eight ELISAs and five immunoblots. *European Journal of Clinical Microbiology and Infectious Diseases*, 2011, 30 (8), pp.1027-1032. <10.1007/s10096-011-1157-6>. <hal-00663398>

**HAL Id: hal-00663398**

**<https://hal.science/hal-00663398v1>**

Submitted on 27 Jan 2012

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

**Large differences between test strategies for the detection of anti-*Borrelia* antibodies are revealed by comparing eight ELISA's and five immunoblots**

C.W. Ang, VUMC, Amsterdam, The Netherlands

D.W. Notermans, Centre for Infectious Disease Control Netherlands, National Institute for Public Health and the Environment (RIVM), Bilthoven, The Netherlands

M. Hommes, VUMC, Amsterdam, The Netherlands

A.M. Simoons-Smit, VUMC, Amsterdam, The Netherlands

T. Herremans, Centre for Infectious Disease Control Netherlands, National Institute for Public Health and the Environment (RIVM), Bilthoven, The Netherlands

Running title: *Borrelia* ELISA's and immunoblots

Key words: Lyme; immunoblot; neuroborreliosis; recombinant antigens; diagnostic test

**ABSTRACT**

We investigated the influence of assay choice on the results in a two-tier testing algorithm for detection of anti-*Borrelia* antibodies. 89 serum samples from clinically well-defined patients were tested in eight different ELISA systems based on whole cell antigens, whole cell antigens supplemented with VlsE and assays using exclusively recombinant proteins. A subset of samples was tested in five immunoblots: one whole cell blot, one whole cell blot supplemented with VlsE and three recombinant blots. The number of IgM and/or IgG positive ELISA results in the group of patients suspected of *Borrelia* infection ranged from 34% to 59%. The percentage of positives in cross-reactivity controls ranged from 0-38%. Comparison of immunoblots yielded large differences in inter-test agreement and showed at best a moderate agreement between tests. Remarkably, some immunoblots gave positive results in samples that had been tested negative by all eight ELISA's. The percentage of positive blots following a positive ELISA result depended heavily on the choice of ELISA-immunoblot combination. We conclude that the assays used to detect anti-*Borrelia* antibodies have widely divergent sensitivity and specificity. The choice of ELISA-immunoblot combination severely influences the number of positive results, making exchange of test-results between laboratories with different methodologies hazardous.

## Introduction

Lyme disease is caused by *Borrelia* spp. In Europe, infection is mostly caused by *B. afzelii* and *B. garinii*, while in the United States, *Borrelia burgdorferi* sensu stricto is the causative agent [1]. Lyme disease manifests in a myriad of clinical ways, including erythema migrans, arthritis, carditis and neuroborreliosis [1]. Extracutaneous Lyme disease requires laboratory confirmation by culture, PCR or antibody determination [2,3]. Culture is only available in a limited number of laboratories, and the value of PCR in the diagnosis of various forms of Lyme disease is of limited use [2,3]. Therefore, serological assays are the main method used to diagnose extracutaneous forms of Lyme disease.

Current guidelines for the diagnosis of Lyme disease include a two-tier testing algorithm [2,3]. First an ELISA is performed, followed by confirmation of positive ELISA results with an immunoblot. This two-step procedure was initiated because first generation ELISA's for the detection of anti-*Borrelia* antibodies lacked specificity. The inclusion of a second, more specific, serological method made it possible to exclude false positive ELISA samples [2,4]. Many diagnostic assays are currently commercially available, and manufacturers have developed them to increase their sensitivity and specificity. During the last decade, assays using a peptide from the sixth invariant region (C6) of the variable major protein-like sequence-expressed (VlsE) of *B. burgdorferi* have been shown to be promising [5,6]. Laboratories can choose between ELISA's and immunoblots using sonicated whole cell antigens, whole cell antigens combined with recombinant antigens (VlsE C6 peptide) and exclusively recombinant antigens. Due to this array of serological tests, there are an almost indefinite number of possible combinations between ELISA and immunoblot in a two-tier testing scheme. Comparing anti-*Borrelia* test results between laboratories and studies may be impossible if tests with widely diverging sensitivities and specificities are used [7].

The aim of the present study was to compare a wide range of ELISA assays and immunoblots, based on either whole cell or recombinant antigens for detecting anti-*Borrelia*

antibodies. We also aimed to investigate the influence of assay choice on results in a two-tier testing algorithm (ELISA followed by immunoblot). Therefore, we tested serum samples in eight ELISA systems and five immunoblots, covering the entire spectrum of native and recombinant antigens.

## Patients and methods

### Patients

Serum samples were selected from 89 clinically well-defined individuals. Fifty-nine samples were from patients suspected of *Borrelia* infection (skin manifestations, n = 8; neurological symptoms, n = 26; arthritic symptoms, n = 11; ocular symptoms, n = 4; other, n = 10).

Fourteen samples were from healthy controls, and 16 came from patients with a high possibility for cross-reacting antibodies (syphilis patients, n = 10; *Mycoplasma pneumoniae*-infected patients based on symptoms consistent with *M. pneumoniae* infection and a positive result for anti-*M. pneumoniae* IgM and IgG with a Virion/Serion ELISA, n = 6).

### Methods

Serum samples were tested in eight different ELISA systems. Three assays were based on sonicated whole cell antigens (Diacheck/Moran anti-*Borrelia*, VIDAS, and Virion/Serion ELISA Classic *Borrelia burgdorferi*), three assays with sonicate whole cell antigens supplemented with VlsE for IgG anti-*Borrelia* antibodies (Dade Behring Enzygnost Lyme link VlsE, Euroimmun Anti-*Borrelia* plus VlsE ELISA, and Genzyme Virotech *Borrelia afzelii* + VlsE ELISA) and two assays using recombinant proteins (Immunetics C6 Lyme ELISA Kit, and Mikrogen recomWell *Borrelia*). A subset of samples from 31 patients suspected of *Borrelia* infection were also tested in five different immunoblots. This group consisted of the following patients: skin manifestations, n = 3; neurological symptoms, n = 15; arthritic symptoms, n = 6; ocular symptoms, n = 2; other, n = 5. One whole cell blot (homemade using *B. afzelii* strain A39 cell sonicate, RIVM), one whole cell blot supplemented with VlsE (Viramed *Borrelia* "MiQ" + VlsE ViraBlot) and three recombinant blots (Euroimmun Euroline-

RN-AT, Mikrogen recom Line *Borrelia*, and Genzyme Virotech *Borrelia* Europe Line). A total of 31 samples were tested in all immunoblots.

Manufacturer-suggested cut-off levels and interpretation criteria were used for ELISA's and immunoblots. Statistical analysis was performed using SPSS version 16.0 (SPSS Inc, Chicago, Ill).

## Results

As expected, there was considerable discordance between the eight ELISA's. We tested 89 samples from patients and controls on all eight ELISA's. Of the complete set of serum samples 35/89 (39%) were negative in all assays, while 16/89 (18%) were positive in all assays. The remaining 38/89 (43%) samples were positive in 1 to 7 ELISA's.

In the 59 patients that were suspected of *Borrelia* infection we observed a wide range of positive results, with percentages of positive ELISA's varying between 34-61% (Table 1). We did not observe a relation between the fraction of positive results and the nature of antigen used for the ELISA. The specificity of the ELISA's also varied widely. Although we had only small numbers of positive tests in healthy controls, some ELISA's produced up to 38% of positive tests in the cross-reactivity group (syphilis and *Mycoplasma pneumonia*-infected patients).

We aggregated results from the IgM and IgG tests and assessed them using a kappa statistic to determine agreement between the ELISA's. The kappa-values ranged from 0.41 (moderate agreement) to 0.79 (substantial to good agreement), emphasizing the differences between the ELISA's (Table 2). The choice of antigen does not seem to influence the level of agreement. Even the lowest kappa-values were observed between two 'whole cell + VlsE' ELISA's (0.43).

We tested a subset of 31 serum samples from patients suspected of *Borrelia* infection in all five immunoblots. Samples were from patients with positive and negative ELISA results, allowing us to investigate the specificity of the immunoblots. In general, we observed much lower agreement for the immunoblots than for the ELISA's. Kappa values ranged from 0 (poor agreement) to 0.84 (good agreement), indicating that for many samples the outcome of the immunoblot is highly dependent on the choice of manufacturer (Table 3). Inter-blot agreement was disappointingly low for IgM, and much higher for IgG (Table 3). Interestingly, recombinant blots did not have a higher agreement than whole cell blots, and there was limited agreement even between recombinant blots. The highest agreement was for the homemade whole cell blot with the Mikrogen recombinant blot. Additional analysis on the individual band level revealed similarly poor agreement, even in immunoblots containing recombinant antigens.

When performing eight different ELISA's and five different blots, there are 40 possible ELISA x blot combinations. Thirty-one samples were tested in all 40 combinations. A score of 0 indicates a negative result in all ELISA's and all blots, while a score of 40 indicates a positive result in all ELISA's and all blots. A score between 0-40 indicates that not all possible combinations yielded a positive result (i.e. disagreement between various ELISA x blot combinations). Of this small sample cohort, 20/31 (65%) either had a score of 0 or 40, indicating perfect agreement irrespective of the ELISA x blot combination used. Discordant interpretations were generated in the other 35%.

The influence of assay choice is further illustrated by investigation of the relationship between each ELISA and the fraction of positive blots. Surprisingly, we found anti-*Borrelia* immunoblot reactivity in samples that were negative in all eight ELISA's. These are samples that normally would not have been tested in immunoblot. Again, this was not dependent on the nature of the antigen used for the immunoblot. For the Euroimmun immunoblot, 4/11 (36%) of the ELISA negative samples were blot-positive. Some immunoblots also seem to

lack sensitivity, since samples that were positive in six to all eight of the tested ELISA's remained negative in all immunoblots. Some of these samples were from Lyme disease patients with a short duration of symptoms, confirming that ELISA's may have a higher sensitivity than immunoblots during the early phase of a *Borrelia* infection.

For some ELISA x blot combinations, only about half of the ELISA-positive samples could be confirmed by immunoblot (e.g. VIDAS ELISA x Virotech immunoblot, Table 4). The quality of the other ELISA's was so high that the majority of ELISA-positive samples were confirmed with immunoblot (e.g. Diacheck/Moran and Enzygnost ELISA's). When taking into account the lack of specificity of a number of the immunoblots, it is clear that the combination of a non-specific ELISA with a non-specific blot will lead to a high fraction of presumably false-positive test results.

The ELISA test-value is the final factor influencing the fraction of positive confirmatory blots. Figure 1 depicts an example – values for the VIDAS and Immunetics C6 Lyme ELISA according to the immunoblot results of a whole cell blot (homemade) and a recombinant blot (Mikrogen). For the VIDAS x homemade blot combination, it is difficult to indicate a cut-off value for the VIDAS ELISA with a good separation between blot positives and blot negatives. When using the Immunetics ELISA as a screening tool, it becomes clear that irrespective of the blot method used, samples with an index >4 are almost always blot-positive. These characteristics make it possible to define groups of ELISA-positive serum samples that do not need immunoblot confirmation.

## Discussion

We studied the influence of the choice of detection method on the results of *Borrelia* serology. We found that *Borrelia* ELISA's and immunoblots for detecting anti-*Borrelia* antibodies have widely divergent sensitivity and specificity, and that immunoblots generally show limited agreement. Analysis of a large number of ELISA x immunoblot combinations revealed large differences between various test strategies in a two-tier testing algorithm.

Although we only studied a limited number of serum samples, our extensive approach allowed us to draw several conclusion based on our observations.

Theoretically, the use of recombinant antigens should lead to increased specificity and possibly increased sensitivity as well. This doesn't seem to be true for the currently available ELISA's and immunoblots for the detection of anti-*Borrelia* antibodies. We could not find a clear relationship between the fraction of positive tests, the specificity and the nature of the antigen used for the serological tests. ELISA's using sonicated whole cell antigens can be sensitive and specific, while recombinant ELISA's may lack specificity. Therefore, manufacturer claims for the superior performance of assays using recombinant antigens for the detection of *Borrelia* antibodies must be interpreted with caution.

A two-tier testing algorithm for the detection of anti-*Borrelia* antibodies is recommended world-wide [2,3,6]. However, there are several reasons to reappraise the additional value of an immunoblot confirmatory test in a two-tier testing scheme.

First, the lack of specificity of some immunoblots is counterintuitive. The immunoblot is used as a confirmatory test, although it can be argued that it is merely a supplemental test due to the interdependence of ELISA's and immunoblots [8]. Theoretically, the use of recombinant antigens should allow discrimination between aspecific antibody reactivity, cross-reactive antibodies and true anti-*Borrelia* antibodies [4]. The presence of commercially available immunoblots with low specificity diminishes the value of the immunoblot as a confirmatory test [8]. Furthermore, the two-tier testing scheme was originally proposed to overcome the lack of specificity of *Borrelia* ELISA's. This study has shown that not all of the newer generation ELISA's using recombinant *Borrelia* antigens have improved specificity than older serological assays [9,10].

Second, the low level of agreement between the different immunoblots is very disappointing, especially for IgM. This low level of agreement, even at the individual band level, makes it hard to compare immunoblot results from different manufacturers.

Third, a mismatch between immunoblot and ELISA may occur during the early phase of infection. There are numerous examples – from this and other studies – in which patients with early Lyme disease were initially ELISA-positive and blot-negative [11]. In such cases, immunoblot seroconversion can only be documented in a follow-up sample, and sometimes even this option is blocked because antibiotic treatment may interfere with the development of the anti-*Borrelia* antibody response [12]. This is an example of better sensitivity in the ELISA compared to the immunoblots. Without detailed knowledge about clinical manifestations and illness duration, reporting these cases as 'negative' could lead to erroneous conclusions.

Finally, several groups can be discriminated based on the ELISA value [10] – a 'high positive' group exhibiting clinical symptoms consistent with a diagnosis of Lyme disease and which can be reported as 'positive' without confirmatory testing, a 'low positive' group in which confirmatory testing may be helpful, and lastly a negative group that does not require any further investigation. We do not advocate abandoning the use of immunoblots to confirm anti-*Borrelia* antibodies, but we do think that only a selection of samples needs confirmatory blotting. Furthermore, knowledge about the lower sensitivity of immunoblots compared to some of the ELISA's is indispensable in interpreting results.

In conclusion, ELISA's and immunoblots for detecting anti-*Borrelia* antibodies have widely divergent sensitivity and specificity, and immunoblots for detecting anti-*Borrelia* antibodies have only limited agreement. Therefore, the choice of ELISA-immunoblot combination severely influences the number of positive results, making exchange of test-results between laboratories with different methodologies hazardous. The widespread availability of more

specific and sensitive assays for the detection of anti-*Borrelia* antibodies will open the way for a reappraisal of the two-tier testing system.

## Acknowledgements

This work has been presented at the 20<sup>th</sup> ECCMID, Vienna, Austria April 2010. The authors would like to acknowledge Stephen Johnston for editing the final manuscript.

## References

1. Stanek G, Fingerle V, Hunfeld KP, Jaulhac B, Kaiser R, Krause A, Kristoferitsch W, O'Connell S, Ornstein K, Strle F, Gray J (2010) Lyme borreliosis: Clinical case definitions for diagnosis and management in Europe. *Clin Microbiol Infect*:in press
2. Wilske B, Zöller L, Brade V, Eiffert H, Göbel UB, Stanek G, Pfister HW (2000) MiQ-12 Lyme-Borreliosis (English internet version). In: Mauch H, Lütticken R, Gatermann S (eds) *Qualitätsstandards in der mikrobiologisch-infektiologischen Diagnostik*. Urban & Fisher Verlag, München Jena,
3. Brouqui P, Bacellar F, Baranton G, Birtles RJ, Bjoersdorff A, Blanco JR, Caruso G, Cinco M, Fournier PE, Francavilla E, Jensenius M, Kazar J, Laferl H, Lakos A, Lotric Furlan S, Maurin M, Oteo JA, Parola P, Perez-Eid C, Peter O, Postic D, Raoult D, Tellez A, Tselentis Y, Wilske B (2004) Guidelines for the diagnosis of tick-borne bacterial diseases in Europe. *Clin Microbiol Infect* 10 (12):1108-1132
4. Wilske B, Fingerle V, Schulte-Spechtel U (2007) Microbiological and serological diagnosis of Lyme borreliosis. *FEMS immunology and medical microbiology* 49 (1):13-21
5. Liang FT, Steere AC, Marques AR, Johnson BJ, Miller JN, Philipp MT (1999) Sensitive and specific serodiagnosis of Lyme disease by enzyme-linked immunosorbent assay with a peptide based on an immunodominant conserved region of *Borrelia burgdorferi* vlsE. *Journal of clinical microbiology* 37 (12):3990-3996
6. Steere AC, McHugh G, Damle N, Sikand VK (2008) Prospective study of serologic tests for Lyme disease. *Clin Infect Dis* 47 (2):188-195
7. Bakken LL, Callister SM, Wand PJ, Schell RF (1997) Interlaboratory comparison of test results for detection of Lyme disease by 516 participants in the Wisconsin State Laboratory

of Hygiene/College of American Pathologists Proficiency Testing Program. Journal of clinical microbiology 35 (3):537-543

8. Wormser GP, Carbonaro C, Miller S, Nowakowski J, Nadelman RB, Sivak S, Agüero-Rosenfeld ME (2000) A limitation of 2-stage serological testing for Lyme disease: enzyme immunoassay and immunoblot assay are not independent tests. Clin Infect Dis 30 (3):545-548
9. Jansson C, Carlsson SA, Granlund H, Wahlberg P, Nyman D (2005) Analysis of *Borrelia burgdorferi* IgG antibodies with a combination of IgG ELISA and VlsE C6 peptide ELISA. Clin Microbiol Infect 11 (2):147-150
10. Smismans A, Goossens VJ, Nulens E, Bruggeman CA (2006) Comparison of five different immunoassays for the detection of *Borrelia burgdorferi* IgM and IgG antibodies. Clin Microbiol Infect 12 (7):648-655
11. Wormser GP, Nowakowski J, Nadelman RB, Visintainer P, Levin A, Agüero-Rosenfeld ME (2008) Impact of clinical variables on *Borrelia burgdorferi*-specific antibody seropositivity in acute-phase sera from patients in North America with culture-confirmed early Lyme disease. Clin Vaccine Immunol 15 (10):1519-1522
12. Agüero-Rosenfeld ME, Nowakowski J, Bittker S, Cooper D, Nadelman RB, Wormser GP (1996) Evolution of the serologic response to *Borrelia burgdorferi* in treated patients with culture-confirmed erythema migrans. Journal of clinical microbiology 34 (1):1-9

**Table 1 -  
Performance of  
eight ELISA's in  
three patient  
groups**

<b>ELISA manufacturer</b>	<b>Antigen used for ELISA</b>	<b>Number of positive samples (%)</b>	<b>Total number of tested samples</b>
-------------------------------	-----------------------------------	---------------------------------------	---

		Patients suspected for <i>Borrelia</i> infection	Cross- reactivity controls	Healthy controls	
<b>Diacheck/Moran</b>	Whole cell	20/59 (34%)	2/16 (13%)	1/14 (7%)	89
<b>VIDAS</b>	Whole cell	31/59 (53%)	4/16 (25%)	1/14 (7%)	89
<b>Virion/Serion</b>	Whole cell	24/59 (41%)	1/16 (6%)	0/14	89
<b>Enzygnost</b>	Whole cell + VlsE	23/59 (39%)	0/16	0/14	89
<b>Euroimmun</b>	Whole cell + VlsE	29/59 (49%)	3/16 (19%)	0/14	89
<b>Virotech</b>	Whole cell + VlsE	35/59 (59%)	6/16 (38%)	0/14	89
<b>Immunetics</b>	Recombinant	22/59 (37%)	0/16	0/14	89
<b>Mikrogen</b>	Recombinant	24/59 (41%)	3/16 (19%)	0/14	89

**Table 2 - Agreement between ELISA's for detecting IgM and/or IgG anti-*Borrelia* antibodies (Kappa-values)**

ELISA manufacturer	Antigen used for ELISA	Diacheck/ Moran	VIDAS	Virion/ Serion	Enzygnost	Euroimmun	Virotech	Immunetics
<b>Diacheck/ Moran</b>	Whole cell	-	-	-	-	-	-	-
<b>VIDAS</b>	Whole cell	0.53	-	-	-	-	-	-
<b>Virion/Serion</b>	Whole cell	0.67	0.69	-	-	-	-	-
<b>Enzygnost</b>	Whole cell + VlsE	0.71	0.62	0.78	-	-	-	-
<b>Euroimmun</b>	Whole cell + VlsE	0.71	0.45	0.56	0.56	-	-	-
<b>Virotech</b>	Whole cell + VlsE	0.44	0.65	0.57	0.43	0.47	-	-

<b>Immunetics</b>	Recombinant	0.74	0.60	0.64	0.86	0.53	0.41	-
<b>Mikrogen</b>	Recombinant	0.79	0.53	0.63	0.68	0.67	0.44	0.65

**Table 3 -  
Agreement  
between  
immunoblots for  
detecting anti-  
*Borrelia*  
antibodies  
(Kappa-values)**

IgM and IgG  
combined

<b>Blot</b>	<b>Blot type</b>	<b>Home made</b>	<b>Virablot</b>	<b>Euroimmun</b>	<b>Mikrogen</b>	<b>Virotech</b>
Home made	Whole cell	-	-	-	-	-
Virablot	Whole cell + VlsE	0.55	-	-	-	-
Euroimmun	Recombinant	0.45	0.24	-	-	-
Mikrogen	Recombinant	0.74	0.42	0.29	-	-
Virotech	Recombinant	0.66	0.60	0.25	0.55	-

IgM

<b>Blot</b>	<b>Blot type</b>	<b>Home made</b>	<b>Virablot</b>	<b>Euroimmun</b>	<b>Mikrogen</b>	<b>Virotech</b>
Home made	Whole cell	-	-	-	-	-
Virablot	Whole cell + VlsE	-1.57	-	-	-	-
Euroimmun	Recombinant	0.04	0.20	-	-	-
Mikrogen	Recombinant	0.42	0	0.26	-	-
Virotech	Recombinant	0.20	0.46	0.39	0.34	-

IgG

<b>Blot</b>	<b>Blot type</b>	<b>Home made</b>	<b>Virablot</b>	<b>Euroimmun</b>	<b>Mikrogen</b>	<b>Virotech</b>
Home made	Whole cell	-	-	-	-	-
Virablot	Whole cell + VlsE	0.43	-	-	-	-
Euroimmun	Recombinant	0.43	0.24	-	-	-

Mikrogen	Recombinant	0.84	0.27	0.43	-	-
Virotech	Recombinant	0.71	0.63	0.30	0.56	-

**Table 4 -  
Fractions of blot  
confirmed  
samples for 40  
ELISA x  
immunoblot  
combinations**

ELISA manufacturer	Antigen used for ELISA	Number of positive samples in ELISA/total number of samples	BLOT				
			Whole cell	Whole cell + VisE	Recombinant		
					home made	Virablot	Euroimmun
<b>Diacheck/Moran</b>	Whole cell	12/31	11/12 (92%)	9/12 (75%)	11/12 (92%)	12/12 (100%)	9/12 (75%)
<b>VIDAS</b>	Whole cell	19/31	11/19 (58%)	12/19 (63%)	13/19 (68%)	14/19 (74%)	10/19 (53%)
<b>Virion/Serion</b>	Whole cell	15/31	11/15 (73%)	11/15 (73%)	13/15 (87%)	12/15 (80%)	9/15 (60%)
<b>Enzygnost</b>	Whole cell + VisE	12/31	11/12 (92%)	10/12 (83%)	10/12 (83%)	12/12 (100%)	10/12 (83%)
<b>Euroimmun</b>	Whole cell + VisE	14/31	11/14 (79%)	11/14 (79%)	12/14 (86%)	12/14 (86%)	9/14 (64%)
<b>Virotech</b>	Whole cell + VisE	17/31	11/17 (65%)	11/17 (65%)	13/17 (77%)	13/17 (77%)	9/17 (53%)
<b>Immunetics</b>	Recombinant	13/31	11/13 (85%)	10/13 (77%)	10/13 (77%)	13/13 (100%)	10/13 (77%)
<b>Mikrogen</b>	Recombinant	13/31	11/13 (85%)	9/13 (69%)	11/13 (85%)	12/13 (92%)	9/13 (69%)

**Figure 1 – ELISA test value in relation to immunoblot results for detection of anti-*Borrelia* antibodies**