



**HAL**  
open science

## Self control of cell electroporation by dielectrophoresis transition

Fériel Hamdi, Olivier Français, Marion Woytasik, Elisabeth Dufour-Gergam,  
Bruno Le Pioufle

► **To cite this version:**

Fériel Hamdi, Olivier Français, Marion Woytasik, Elisabeth Dufour-Gergam, Bruno Le Pioufle. Self control of cell electroporation by dielectrophoresis transition. Nanobiotech, 2011, Montreux, Switzerland. hal-00739183

**HAL Id: hal-00739183**

**<https://hal.science/hal-00739183>**

Submitted on 18 Feb 2013

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## SELF CONTROL OF CELL ELECTROPORATION BY DIELECTROPHORESIS TRANSITION

Feriel S. Hamdi<sup>1,2</sup>, Olivier Français<sup>1</sup>, Marion Woytasik<sup>2</sup>, Elisabeth Dufour-Gergam<sup>2</sup>, Bruno Le Pioufle<sup>1</sup>

<sup>1</sup> Ecole Normale Supérieure de Cachan, CNRS, SATIE, UMR 8029, Cachan, FRANCE

<sup>2</sup> Univ Paris-Sud, CNRS, Institut d'Electronique Fondamentale, UMR 8622, Orsay, France

This paper deals with the transition of the dielectric properties of cells as a marker of their permeabilization. A demonstration of the transition between positive and negative DiElectroPhoresis (DEP) as a reaction to electroporation pulses is achieved in a microfluidic device.

Electropermeabilization of cell membrane is broadly used in cell biology for electrotransfert of genes and other molecules in vitro and in vivo [1] or for electrofusion [2]. The success rate of this technique is nevertheless highly variable, depending on the type of cells, its status in the cell cycle and on the method used. In the case of the use of conventional electroporation cuvet, the electroporation rate remains unsatisfying (if we apply an electric field for which 90% of the cells survive, the efficiency remains below 20% [3]). Several attempts to propose biochips dedicated to the electroporation have been proposed to address this issue [4]. The advantages of using a dedicated microfluidic device are i) the capability to control precisely the electrical field at the scale of the single cell [4] ii) the possibility to parallelize ([5], [6]) or serialize [7] the single cell electroporation function in order to achieve the high number of cells with a good ratio of permeabilization.

We present in this paper a novel method, implemented on a microfluidic device, that permits to control automatically and independently the level of electropermeabilization of cells. The cells are disposed in successive traps, using DEP force. Once the membrane permeabilization occurs, the dielectrical properties of cells change. If the electric field is maintained, the cells are automatically and separately expelled from the high electrical field zone. In that way, the cells are self-controlling their permeabilization level, prior to further utilization.

### **Fabrication of the device**

The microfabrication of the device is performed on a quartz substrate. Electrodes (10 nm Ti/ 240 nm Pt) are patterned on this substrate thanks to the lift-off of a Titanium/Platinum layer sputtered on a negative resist. Channels are defined in 25µm thick SU8 thanks to photolithography. The device is packaged using a membrane of PDMS bounded irreversibly after silanization. The several steps of the fabrication process are shown on figure 1.

The figure 2 shows the simulation within the device that includes insulating SU8 structures that generate the field gradients necessary for the cell handling.

### **Experiment and results**

The mouse fibroblast cells NIH3T3 were injected within the fluidic channel (buffer conductivity= 272µS/cm). A sinusoidal field excitation applied on the electrodes ( $f=1\text{MHz}$ ,  $V=30\text{V}$ ) permitted the trapping of cells by positive DEP (figure 3.a). Buffer was injected to eject untrapped cells. Electroporation pulses (10 pulses, duration 100µs,  $V=75\text{V}$ ) were superposed to the DEP excitation. Maintaining the same level and frequency of excitation, the permeabilized cells moved progressively to the low electric field areas thanks to negative DEP (as shown on figure 3.b).

### **Acknowledgements**

We thank Frederic Subra for cell culture. This work was performed thanks to the financial support of PRES Universud, CNANO'Ile de France, and the ANR PNANO Nanopulsebiochip.




Figure 1: Process of fabrication: (a,b,c).photolithography of negative resist AZ 5214, d. Ti and Pt sputtering, e. lift-off,( f,g,h). SU8 photolithography, i. silanization and packaging with PDMS




Figure 2: Simulation of the electric field




Figure 3: **a.** Before electroporation: Cells are trapped by positive DEP, **b.** Once electroporated, cells are ejected from the electroporation area by negative DEP (contour detection with Image J ©)

## REFERENCES:

- [1] "In vivo gene electroinjection and expression in rat liver", Richard Heller et al, FEBS Letters, 389(3):225-8, (1996)
- [2] "High yield parallel electrofusion device based on field constriction at an orifice array", Washizu et al, Miniaturized Systems for Chemistry and Life Sciences, October 12-16 2008, San Diego.
- [3] "Energy dissipation as a key factor for the electroporation of protoplasts", Chen G.Y, A.J. Conner, Jian Wang, A.G. Fautrier and R.J. Field, Molecular Biotechnology, **10**, 209-216 (1998).
- [4] "Microfabricated electroporation chip for single cell membrane permeabilization", Y. Huang and B. Rubinsky. Sensors and Actuators A, **89**, 242-249 (2001).
- [5] "Living cells captured on a bio-microsystem devoted to DNA injection", B. Le Pioufle, P. Surbled, H. Nagai, Y. Murakami, H. S. Chun, E. Tamiya and H. Fujita. Materials Science and Engineering: C, **12**, (1-2), 77-81 (2000).
- [6] "Toward single cell controlled electroporation in a microfluidic device", U. Seger, S. Gawad, A. Scheer and P. Renaud in, MicroTAS, Nara, pp 796-798 (2002)
- [7] "Flow-through micro-electroporation chip for high efficiency single-cell genetic manipulation", Y. Huang and B. Rubinsky. Sensors and Actuators A: Physical, **104**, (3), 205-212 (2003).