

Biomechanics for computer-assisted surgery

Yohan Payan

▶ To cite this version:

Yohan Payan. Biomechanics for computer-assisted surgery. Computer Methods in Biomechanics and Biomedical Engineering, 2012, 15 (S1), pp.8-9. hal-00739712

HAL Id: hal-00739712

https://hal.science/hal-00739712

Submitted on 8 Oct 2012

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Biomechanics for Computer-Assisted Surgery

Yohan PAYAN*†

† UJF-Grenoble 1 / CNRS / TIMC-IMAG UMR 5525, Grenoble, F-38041, France

Keywords: soft tissues; patient-specific; constitutive law; finite element model

1 Context

TIMC-IMAG Laboratory (www-timc.imag.fr) is a 250 people laboratory devoted to translational and fundamental research at the intersection between Medicine, Information Science and Technology (from Applied Mathematics to Computer Science and Robotics). In that lab, since the 80's, the Computer-Assisted Medical Intervention group is developing devices to assist the physician or the surgeon in the successful execution of diagnostic or therapeutic gestures by minimizing invasiveness whilst improving accuracy.

Computer-Assisted Surgery (CAS) is now a mature domain. Researchers, clinicians and industrial partners have developed CAS applications by building links with classical domains such as Computer Science, Robotics, Image Processing and Mathematics. Orthopaedics was the first clinical domain mainly addressed by the pioneer CAS applications [1]. The reason for this was probably that bones are the human body structures which were considered as the most easily includable into a CAS application: they were assumed to be rigid, i.e. with a fixed 3D geometry, they are strongly identifiable onto Computed Tomography exams, and their relative position during surgery is easily tractable by fixing rigid bodies onto their external surfaces (these rigid bodies being for example tracked with the use of an optical device, thus providing "surgical navigation").

The connection to Biomechanics (i.e. the Mechanics of living tissues) is more recent. Biomechanicians were first asked to work onto CAS applications when orthopaedic surgeons were looking for tools able to predict risks of fractures in the case of prosthetic implants. In that case, bonny structures could no more be considered as rigid but on the contrary had to be modeled as a deformable continuum with a nonhomogeneous distribution of the internal stresses. For example, a patient-specific Finite Element model of the femur could be designed to estimate the internal stresses generated by a hip prosthesis and therefore to help limit In fracture risks [2]. these continuous biomechanical models, bones were usually considered as linear elastic material that underwent deformations, which permitted easy calculation of numerical solutions.

More recently, CAS has addressed a larger spectrum of clinical domains such as cardiology, neurosurgery, urology or abdominal surgery. For these applications, biomechanics faces a new challenge since the involved tissues are required to move and be deformed by stress generated by clinical actions. Moreover, soft tissues are difficult to model accurately since they typically exhibit complex, time dependent, non-linear, inhomogeneous and anisotropic behaviors. Most of the corresponding biomechanical models need to include large deformation effects and viscohyperelastic constitutive laws. Such models are very computationally demanding and are therefore limited to pre-operative use, since the simulations often require many minutes or hours to compute. Our group did contribute to such pre-operative use of biomechanical models, for example in the

domain of orthognatic surgery [3], tongue cancer treatment [4] or orbital surgery [5].

More recently, we have addressed the new frontier that biomechanics is now facing with the development of CAS devices that can provide intra-operative assistance [6]. The underlying idea is to use patient-specific biomechanical models during surgery, i.e. in the operating theater. In that case, three main challenges need to be solved to be compatible with the clinical constraints:

- (1) patient-specific models should be easily generated (no more than some minutes to elaborate such a model);
- (2) patient-specific constitutive equations of the soft tissues have to be estimated through in vivo experiments, some of them only being possible during surgery if the organs are not accessible pre-operatively (e.g. the brain tissues);
- (3) the implementation of the models should provide real-time (or at least interactive) numerical simulations.

2 Generation of patient-specific models

In order to face the time constraints for the design of patient-specific Finite Element (FE) models, our group has proposed the Mesh Matching algorithm [7] followed by the Mesh-Match-and-Repair (MMRep) approach [8]. The idea consists in maintaining the advantages of a manual design of

the FE mesh, while introducing an automatic mesh conformation process. The global algorithm is based on the following strategy:

- 1. A Finite Element "generic" (or "atlas") biomechanical model is manually built. This step is long, tedious, but is done once.
- 2. Patient data are collected (US, CT, or MRI).
- 3. The generic FE mesh is automatically conformed to patient morphology (extracted from the data), by the mean of a local elastic registration.
- 4. The new patient mesh is regularized so that it can be used for FE analysis. This patient mesh has a topology similar to the generic mesh (same number of elements and same element types).

3 Estimation of the constitutive law

In order to estimate the *in vivo* constitutive behavior of human soft tissues, our group has developed the Light Aspiration device for in vivo Soft TIssue Characterization (LASTIC), based on the pipette aspiration principle and consisting in measuring the tissues deformations induced by a negative pressure. It is built in a very compact metallic cylinder of 33 mm in height and 34 mm in diameter [9]. The surgeon holds the instrument and establishes contact with the tissues surface while the device measures the negative pressures and displacement responses. The LASTIC device can undergo a full sterilization and the data processing is sufficiently fast to provide an interactive estimation of the soft tissues constitutive equation. LASTIC has already been used to evaluate the constitutive behavior of forearm skin, tongue [10] and brain tissues [11].

4 Interactive-time numerical simulations

Last, but not the least, is the constraint of a quasireal-time computation of the simulations provided by the models, in order to be used during surgery. This can be ensured when tissues deformations are small (e.g. brain deformations during large skull opening [12]) but it is still challenging for models with large visco-hyper-elastic frameworks (see [13] for example). This bottleneck will definitively need to be studied in the future.

References

- [1] Taylor R., Lavallée S., Burdea G. & Mosges R. Computer-Integrated Surgery. Technology and clinical applications, Cambridge, MA: MIT Press, 1996.
- [2] Weinans, H.; Huiskes, R.; Grootenboer, H. J.: Effects of fit and bonding characteristics of femoral stems on adaptive bone remodeling. *J. Biomech. Eng.*, 116, (1994), 393–400.
- [3] Chabanas M., Luboz V. & Payan Y. (2003). Patient specific Finite Element model of the face soft tissue for computer-assisted

- maxillofacial surgery, *Medical Image Analysis*, Vol. 7, Issue 2, pp. 131-151.
- [4] Buchaillard S., Brix M., Perrier P. & Payan Y. (2007). Simulations of the consequences of tongue surgery on tongue mobility: Implications for speech production in post-surgery conditions. *International Journal of Medical Robotics and Computer Assisted Surgery*, Vol. 3(3), pp. 252-261.
- [5] Luboz V. Swider P., Ambard D., Boutault F. & Payan Y. (2007). Intra-operative quantification of the surgical gesture in orbital surgery: application to the proptosis reduction. *Clinical Biomechanics*, Vol. 22, pp. 298-303.
- [6] Payan Y., ed. (2012). Soft Tissue Biomechanical Modeling for Computer Assisted Surgery. Springer-Verlag, ISBN 978-3-642-29013-8.
- [7] Couteau B., Payan Y. & Lavallée S. (2000). The Mesh-Matching algorithm: an automatic 3D mesh generator for finite element structures, *Journal of Biomechanics*, vol. 33/8, 1005-1009.
- [8] Bucki M., Lobos C. & Payan Y. (2010). A Fast and Robust Patient Specific Finite Element Mesh Registration Technique: Application to 60 Clinical Cases. *Medical Image Analysis*, Vol. 14, pp. 303–317.
- [9] Schiavone P., Promayon E. & Payan Y. (2010). LASTIC: a Light Aspiration device for in vivo Soft TIssue Characterization. *Lecture Notes in Computer Science*, Vol. 5958, pp. 1-10.
- [10] Schiavone P., Boudou T., Promayon E., Perrier P. & Payan Y. (2008). A light sterilizable pipette device for the in vivo estimation of human soft tissues constitutive laws. Proceedings of the 30th Annual International Conference of the IEEE Engineering in Medicine and Biology Society, IEEE EMBS 2008, pp. 4298-4301.
- [11] Schiavone P., Chassat F., Boudou T., Promayon E., Valdivia F. & Payan Y. (2009). In vivo measurement of human brain elasticity using a light aspiration device. *Medical Image Analysis*, Vol. 13, pp. 673–678.
- [12] Bucki M., Palombi O., Bailet M., Payan Y. (2012). Doppler Ultrasound Driven Biomechanical Model of the Brain for Intraoperative Brain-Shift Compensation: A Proof of Concept in Clinical Conditions. In Soft Tissue Biomechanical Modeling for Computer Assisted Surgery, Y. Payan Editor, Springer, pp. 135-165, ISBN 978-3-642-29013-8.
- [13] Stavness I., Lloyd J., Payan Y. & Fels S. (2011). Coupled Hard-Soft Tissue Simulation with Contact and Constraints Applied to Jaw-Tongue-Hyoid Dynamics. *International Journal for Numerical Methods in Biomedical Engineering*, Vol. 27, pp. 367–390.