



A new method for the in vivo identification of mechanical properties in arteries from cine MRI images: theoretical framework and validation.

Alexandre Franquet, Stéphane Avril, Rodolphe Le Riche, Pierre Badel, Fabien Schneider, Zhi-Yong Li, Chrisitan Boissier, Jean-Pierre Favre

► To cite this version:

Alexandre Franquet, Stéphane Avril, Rodolphe Le Riche, Pierre Badel, Fabien Schneider, et al.. A new method for the in vivo identification of mechanical properties in arteries from cine MRI images: theoretical framework and validation.. IEEE Transactions on Medical Imaging, 2013, 32 (8), pp.1448 - 1461. 10.1109/TMI.2013.2257828 . hal-00805124

HAL Id: hal-00805124

<https://hal.science/hal-00805124>

Submitted on 27 Mar 2013

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

1 A new method for the *in vivo* identification
2 of mechanical properties in arteries from cine
3 MRI images: theoretical framework and
4 validation.

5
6 Alexandre Franquet, Stéphane Avril, Rodolphe Le Riche, Pierre Badel, Fa-
7 bien C Schneider, Zhi Yong Li, Christian Boissier and Jean Pierre Favre

- 8 1. A. Franquet, S. Avril and P. Badel are with the CIS-EMSE, CNRS UMR 5146, Ecole Nationale Supérieure
9 des Mines, Saint-Etienne F-42023, France. e-mail: avril@emse.fr
- 10 2. R. Le Riche is with CNRS UMR 6158 and the Ecole nationale Supérieure des Mines, Saint-Etienne F-42023
- 11 3. Z.Y. Li is with the School of Biological Science and Medical Engineering, Southeast University, Nanjing
12 210096, China
- 13 4. F.C. Schneider, C. Boissier and J.P. Favre are with the Radiology Department, Vascular Department, and
14 Cardiovascular surgery Department, respectively, University Hospital of Saint Etienne France and GRT-
15 PAR UR4672 Jean Monnet University, Saint Etienne F-42000, France.

16 **This work has been submitted to the IEEE for possible publication. Copy-**
17 **right may be transferred without notice, after which this version may no**
18 **longer be accessible**

19 **Abstract**

20 Quantifying the stiffness properties of soft tissues is essential for the di-
21 agnosis of many cardiovascular diseases such as atherosclerosis. In these
22 pathologies it is widely agreed that the arterial wall stiffness is an indicator
23 of vulnerability. The present paper focuses on the carotid artery and pro-
24 poses a new inversion methodology for deriving the stiffness properties of
25 the wall from cine-MRI (Magnetic Resonance Imaging) data. We address
26 this problem by setting-up a cost function defined as the distance between
27 the modeled pixel signals and the measured ones. Minimizing this cost

function yields the unknown stiffness properties of both the arterial wall and the surrounding tissues. The sensitivity of the identified properties to various sources of uncertainty is studied. Validation of the method is performed on a rubber phantom. The elastic modulus identified using the developed methodology lies within a mean error of 9.6 %. It is then applied to two young healthy subjects as a proof of practical feasibility, with identified values of 625 kPa and 587 kPa for one of the carotid of each subject.

Keywords Finite elements, mechanical properties, elasticity, inverse method, identification, artery, *in vivo* analysis.

1 Introduction

Quantifying the stiffness properties of soft tissues is essential for the diagnosis of many diseases. This is especially important for cardiovascular pathologies where it has been shown that the stiffness of the arterial wall is an indicator of the stroke risk for diseases such as atherosclerosis [1]. This could be used to improve the diagnosis in clinical practice.

The mechanical properties of arteries have been a research topic of major attention for thirty years ([2, 3, 4, 5, 6, 7, 8, 9, 10]). Many authors have characterized the mechanical properties of excised arteries using either tensile tests, indentation tests or inflation tests. However the results reported in the literature are not very consistent. The large scattering observed can be attributed to the differences in experimental conditions and conservation processes, to the anisotropic and behavioral non linearities ([5, 11, 12]), or to the natural inter-individual variability.

Therefore, the identification of the mechanical properties *in vivo* provides another essential point of view. *In vivo* identification techniques are more constraining because they are usually non-intrusive. An overview of the existing methods and results of identification is presented in Table 1 which will be thoroughly discussed in the “Discussion”. In clinical applications, for assessing arterial stiffness, it is still commonplace to use the Moens-Korteweg theoretical equation relating the Pulse Wave Velocity and the incremental modulus in a round cylinder [13]. Recent developments based on the propagation of shear waves in soft tissues have enabled clinicians to improve the evaluation of the arterial stiffness [14]. Although these types of methods are well suited for rapid

61 medical examinations, they suffer, among other things, from a poor precision
62 because of the strong assumptions made about the mechanical behavior of ar-
63 teries.

64 In order to improve precision, medical imaging techniques can be employed.
65 The basic method consists in tracking the diameter changes in the images and
66 relating these to blood pressure variations ([15, 16, 17, 18, 19, 20, 21, 22, 23, 24,
67 25, 14]). The advantage is that a non-linear behavior can possibly be identified
68 [26] however the precision remains questionable due to the assumptions about
69 the geometry of the artery.

70 Imaging techniques providing access to kinematic fields (displacements,
71 strains) offer an attractive flexibility and enable the identification of more com-
72 plex mechanical behaviors with a limited number of mechanical assumptions.
73 However, they require the development of appropriate inverse methods [27].
74 The most natural one is the Finite Element Model Updating (FEMU) method
75 which has been applied several times to data acquired *in vivo*.

76 The FEMU method consists in finding the parameters of a Finite Element
77 (FE) model that minimize the gap between numerical and experimental kine-
78 matic fields (displacements, strains). This method has been successfully ap-
79 plied to identify the heterogeneous elastic properties of a pig artery with an
80 atherosclerotic plaque ([28, 29]). The elastic properties and the contours of an
81 atherosclerotic plaque in the human coronary artery were also determined in a
82 similar way using a Parametric FEMU method (P-FEMU) [30]. Using the same
83 idea, other authors [31] have designed a specific algorithm where the contours
84 are recovered by a watershed method and the heterogeneous elastic properties
85 are found at the same time.

86 All these applications of the FEMU method are based on experimental strain
87 fields obtained by IntraVascular UltraSounds (IVUS). IVUS is an ultrasonic
88 imaging technique where the ultrasound probe is inserted inside the arterial
89 conduit using a catheter. Specific algorithms have been developed for deriv-
90 ing radial strains from the Radio Frequency (RF) signals with a suitable spatial
91 resolution (pixel size ~ 0.1 mm). But the IVUS technique is very invasive due to
92 the use of catheters thereby limiting its applications.

93 Other non-invasive imaging techniques are available for mapping the strain
94 fields but their spatial resolution is inferior. For instance, strain maps can be
95 measured *in vivo* in the vicinity of the carotid artery using a specific Mag-

netic Resonance Imaging (MRI) sequence: the DENSE (Displacement ENcoding with Stimulated Echoes) sequence [32]. The resulting pixel size is over 0.5 mm and moreover requires highly skilled operators.

Despite the great potential of MRI for these types of problems, the pixel size is usually not fine enough to track the strain fields of the arterial wall, limiting the use of MRI for the purpose of accurately identifying the mechanical behavior of arterial walls.

In this paper we address this issue by setting-up a model relating the temporal variations of the pixel MR signal to the strains and deformations of the tissue. Using this approach, we are able to combine the advantages of a full field based identification method with a non invasive imaging device. A novel inverse method is developed called Magnitude-Based FEMU (MB-FEMU). The cost function is defined as the distance between the modeled pixel signals and the measured ones. Minimizing this cost function yields the unknown stiffness properties of the arterial wall and possibly of the surrounding tissues. This type of method has been developed in the context of elastography ([33, 34]) for assessing hard nodules in breast cancers. However the issue of the low spatial resolution of MRI compared with the size of the specimen may prevent from using this method on arteries. In this study we used the partial volume effect that occurs in MRI, which corresponds to the partial filling of the pixels by different materials.

After presenting the theoretical and numerical foundations of the method, the method is validated on a rubber phantom and then applied to two subjects as a proof of practical feasibility.

2 Methodology

The approach is based on the minimization of a cost function which measures the distance between a template image and a registered image. The template image is the experimental image of a deformed artery provided by the magnitude of a cine Phase-Contrast MR sequence. The registered image is obtained from the experimental image of the undeformed artery and a displacement field provided by a FE analysis. After introducing the theoretical framework, the implementation of the method is described and then the experimental set up used to validate the method is presented.

Table 1: Literature summary of methods, imaging systems and Young’s moduli of arteries identified *in vivo*.

Team	Year	Artery	Imaging system	Number of subjects	Methodology / index	Value (kPa)
Riley et al.	1992	Common carotid	B-mode echo	419	Variation of diameter	701 ± 324
Laurent et al.	1994	Radial artery	RF US signal	22	Variation of diameter / E_{inc}	2680 ± 1810
Brands et al.	1999	Common carotid	RF US signal	1	Variation of diameter / E_{inc}	480
Aggoun et al.	2000	Common carotid	B-mode echo	21	Variation of diameter / E_{inc}	187 ± 67
Bussy et al.	2000	Common carotid	RF US signal	40	Variation of diameter / E_{inc}	710 ± 290
Boutouyrie et al.	2001	Common carotid	RF US signal	15	Variation of diameter / E_{inc}	431 ± 243
Selzer et al.	2001	Common carotid	B-mode echo	24	Variation of diameter	630 ± 230
Kanai et al.	2003	Common carotid	RF US signal	2	Variation of diameter.	960
Boutouyrie et al.	2004	Common carotid	RF US signal	16	Variation of diameter / E_{inc}	230 ± 110
Hasegawa and Kanai	2004	Common carotid	RF US signal	1	Variation of diameter	1400
Baldewsing et al.	2005	Coronary artery	IVUS	1	Strain based FEMU	188
Taviani et al.	2008	Common carotid	MRI	3	Lumen’s area based FEMU	250
Avril et al.	2009	Common carotid	MRI	1	Pulse wave Analysis	99
Couade et al.	2010	Common carotid	RF US signal	1	Shear waves propagation	402

2.1 Theoretical framework

2.1.1 Kinematics principles

In continuum mechanics a body can be viewed as a set of elements called material points that occupy regions in space. The reference and current configurations of a body refer respectively to the undeformed or initial configuration of the body, denoted $\mathcal{C}_0$, and to the deformed or current configuration, denoted $\mathcal{C}_t$, after the application of forces and moments. The position of a material point in the initial configuration $\mathcal{C}_0$ is written $\underline{X}$ while the position of a material point in the current configuration $\mathcal{C}_t$ is written $\underline{x}$. $\underline{x}$ is a function of the coordinates of

the material points and of time such as $\underline{x} = \Phi_\theta(\underline{X}) = \underline{X} + \underline{U}(\underline{X})$. Similarly it is possible to write $\underline{X} = \Phi_\theta^{-1}(\underline{x}) = \underline{x} - \underline{U}(\underline{X})$ where the transformation Φ_θ is a function of the material properties θ and of $\underline{U}$ the displacement. It is possible to introduce the function $\underline{u}$ which will be used in this paper so that $\underline{X}$ is a function of $\underline{x}$ only:

$$\begin{aligned}\underline{X} &= \underline{x} - \underline{U}(\underline{X}) \\ &= \underline{x} - \underline{U}(\Phi_\theta^{-1}(\underline{x})) \\ &= \underline{x} - \underline{u}(\underline{x})\end{aligned}\tag{1}$$

2.1.2 Signal magnitude in MRI

In medical imaging, contrast results from materials which are different in nature. A cine-MRI sequence allows to measure the structure's response at different times of the cardiac cycle. We assumed that the magnitude at position $\underline{x}$ of the signal in the deformed configuration $\mathcal{C}_t$ is identical to the magnitude of the signal at position $\Phi_\theta^{-1}(\underline{x})$ in the initial configuration $\mathcal{C}_0$. This is justified by the fact that $\underline{x}$ and $\Phi_\theta^{-1}(\underline{x})$ are the positions of the same material particle in two different configurations. This means that:

$$I(\underline{x}) = I_0(\Phi_\theta^{-1}(\underline{x}))\tag{2}$$

where I is the magnitude in the current configuration $\mathcal{C}_t$ and I_0 is the magnitude in the initial configuration $\mathcal{C}_0$. An implementation of Φ_θ will be proposed later in the present article based on FE displacement fields.

2.1.3 Identification of material properties

Equation (2) is verified if the transformation Φ_θ corresponds to the real experimental transformation. Part of the inverse method is based on the evaluation of a cost function which calculates the distance between a measured quantity and an estimation of this quantity which is a function of unknown parameters. In our case the transformation Φ_θ of the structure is unknown. It is affected by the material parameters which can therefore be identified by minimizing the following distance or cost function J_2 :

$$J_2(\theta) = \frac{1}{2} \iint_{\Omega} \left(I(\underline{x}) - I_0(\Phi_\theta^{-1}(\underline{x})) \right)^2 d\underline{x}\tag{3}$$

162 The domain Ω corresponds to the imaged body. I_0 and I are experimental im-
 163 ages measured by the imaging system in the configurations $\mathcal{C}_0$ and $\mathcal{C}_t$, respec-
 164 tively. An important feature of our method is that the distance J_2 is directly
 165 expressed in terms of image magnitudes, as opposed to displacements, strains,
 166 stresses, or energies.

167 Minimizing Equation (3) on θ gives the material properties for which the
 168 image associated to the transformation Φ_θ is the closest to the measured image.

169 2.2 Implementation

170 The proposed MB-FEMU methodology is an instance of the Finite Element
 171 Model Updating method. This method can be described by the association
 172 of: (i) experimental data, (ii) a numerical model based on finite elements and
 173 (iii) a cost function which measures the distance between (i) and (ii). MRI was
 174 chosen for providing the experimental data. The numerical model consists
 175 in computing the image of the deformed geometry from displacement fields
 176 which were estimated by a FE computation. The cost function is an elabora-
 177 tion of Equation (3). These three aspects are now further detailed. Later in the
 178 article, the method is applied *in vitro* to a rubber phantom and *in vivo* to two
 179 healthy subjects.

180 2.2.1 Experimental data

181 A Phase Contrast sequence was used to provide cine-MRI data [35]. A bipolar
 182 magnetic gradient is applied to the imaged body during the excitation phase
 183 which means that a first gradient is applied along one direction followed by
 184 the application of a second gradient in the opposite direction during the same
 185 time frame. The moving protons and the stationary protons are respectively
 186 dephased and rephased by the second gradient. The phase information of this
 187 sequence is often used to obtain the flow velocity in the out-of-plane direction.
 188 In our case however, only the magnitude of the signal is considered since it is
 189 mainly correlated to the T_1 intrinsic property of the materials.

190 2.2.2 Numerical model

191 Finite element modeling was chosen to model the structure (phantom and ar-
 192 teries). 2D plane strain was assumed inasmuch as out-of-plane displacements

193 were small compared to in-plane displacements and MR slice thickness. Quasi
 194 static conditions were used as heart beats have a low frequency (approximately
 195 1 Hz). In the loading interval, only the tangent behavior of both the phan-
 196 tom and the arteries was considered: linear elasticity in the finite deformations
 197 framework was assumed for all materials.

198 2.2.3 Cost function description

199 The FEMU process iteratively adjusts the mechanical parameters of a model
 200 based on its mechanical response as obtained by solving FE models until con-
 201 vergence is reached. The iterations are created by the minimization of a dis-
 202 tance with respect to the parameters:

$$\min_{\theta} J_2(\theta) \quad (4)$$

203 In practice the spatial resolution of MRI depends on the acquisition time,
 204 which should be small enough to satisfy the constraints of a clinical imple-
 205 mentation, and on the number of $\underline{k}$ sets which can be measured. Equation (3)
 206 is then a discrete sum with N terms (= number of pixels). In order to save
 207 computation time we assumed that the pixels were undeformable and that the
 208 transformation of a pixel is a translation only. As a consequence, the inverse
 209 transformation Φ_{θ}^{-1} can be simplified for a pixel centered at coordinates $\underline{\zeta}_k$ as:

$$\Phi_{\theta}^{-1}(\underline{\zeta}_k) \approx \underline{\zeta}_k - \underline{u}_{\theta}(\underline{\zeta}_k) \quad (5)$$

210 where $\underline{u}_{\theta}$ is the displacement in the current configuration $\mathcal{C}_t$ (see notation in-
 211 troduced in Equation (1)) which depends on the mechanical properties θ .

212 Fig. 1 describes the calculation of the signal magnitude provided by the
 213 pixel k which, following the above assumptions, is a discrete sum:

$$I_0(\Phi_{\theta}^{-1}(\underline{\zeta}_k)) \approx I_0(\underline{\zeta}_k - \underline{u}_{\theta}(\underline{\zeta}_k)) = \sum_{j=1}^4 \alpha_{kj}(\theta) \cdot I_0(\underline{\zeta}_{kj}) \quad (6)$$

214 As illustrated in Fig. 1, $k1, \dots, k4$ are the pixels which, in the current con-
 215 figuration, intersect the initial image $I_0(\underline{\zeta}_{kj})$. $\alpha_{k1}, \dots, \alpha_{k4}$ are the normalized
 216 intersection areas.

217 The whole image can be reconstructed in this way. However, in order to
 218 provide more sensitivity to the method, a restrained set of pixels called “control





Figure 1: Illustration of the principle applied to generate the current image from the reference image and the computed displacements $u_\theta(\xi_k)$ in order to recover the coefficients α_{kj} for each pixel. In this example, the magnitude of the pixel in the middle of the deformed image is calculated using the coefficients $\{\alpha_{k1}, \alpha_{k2}, \alpha_{k3}, \alpha_{k4}\} = \{50\%, 20\%, 10\%, 20\%\}$ (see Equation 6).

pixels" was selected. Control pixels were located in areas where a large change in signal magnitude is likely to be observed i.e. at the lumen contours. A binary mask is defined to select the control pixels (see Fig. 2a) so that N is less than the total number of pixels. The cost function is normalized by the number of pixels considered.


Finally, the cost function is written as:

$$\begin{aligned}
 J_2(\theta) &= \frac{1}{2 \cdot N} \sum_{k=1}^N \left(I(\xi_k) - I_0 \left(\Phi_\theta^{-1}(\xi_k) \right) \right)^2 \\
 &= \frac{1}{2 \cdot N} \sum_{k=1}^N \left(I(\xi_k) - \sum_{j=1}^4 \alpha_{kj}(\theta) \cdot I_0(\xi_{kj}) \right)^2
 \end{aligned} \tag{7}$$


The minimization of Equation (7) is an iterative process where different combinations of material parameters θ were tested in a FE model. The updated vector of parameters ($\theta \leftarrow \theta + \Delta\theta$) was determined by a Levenberg-Marquardt algorithm with bounds handling [36]. It requires gradients of the computed intensity $I_0(\Phi_\theta^{-1}(\xi_k))$ with respect to the parameters θ which were estimated by backward finite differences (or forward differences at lower bounds). The identification process stopped when one of the following termination criteria was satisfied: $J_2(\theta) \leq \varepsilon_a = 10^{-7}$ (accuracy on J_2 was reached) or $\|\Delta\theta\| \leq \varepsilon_b = 10^{-3}$ (the step size was too small, no more improvement was expected.). We



(a) Binary mask which highlights the control pixels.



(b) Initial filtered image $I_0(\xi_k)$.



(c) Deformed filtered image $I(\xi_k)$.

Figure 2: Example of images processing during the phantom study. A mask is defined to select the control pixels.

previously showed the importance of setting these two criteria ([37]).

The method is summarized in the flow chart of Fig. 3.

2.3 Validation of the MB-FEMU method

2.3.1 Experimental subjects

Phantom description An arterial rubber phantom was used to perform *in vitro* experimentation. The phantom was a 50-cm-long tube with a 7 mm inner diameter and 10 mm outer diameter made of PVA gel (1 freeze-thaw-cycle), surrounded by gadolinium enriched water mimicking the surrounding tissue (see Fig. 4). A peristaltic pump generated a pulsating flow mimicking physiological pressure while the wall motion was imaged.

Patients description Two healthy human subjects were selected for the experiment. They were 25-year-old and 24-year-old healthy subjects (185 cm/85 kg and 183 cm/80 kg, respectively). Their common carotid arteries were imaged using MRI.




Figure 3: Flow chart of the MB-FEMU methodology to identify mechanical and other parameters. From the cine Phase Contrast MRI images, initial (diastole) and deformed (systole) images were chosen. A FE model is created from the initial image. Then for a given set of initial parameters θ_0 a FE model is computed to provide a displacement field $U_\theta(\underline{X})$. This field is applied to the initial image $I_0(\underline{\xi}_k)$ according to the method described in the text in order to generate a registered image, $I_0(\Phi_\theta^{-1}(\underline{\xi}_k))$. The cost function J_2 is the distance between this registered image and the experimental deformed image, calculated for a given region of interest (binary mask: control pixels). A Levenberg-Marquardt with bounds handling is used to minimize the cost function J_2 with respect to the parameters θ . The procedure stops if the desired accuracy on J_2 is reached (ε_a) or if the step size is too small (ε_b) to expect significant further improvements of the cost function J_2 .

2.3.2 Reference data

Measurement technique A 3T Siemens Verio system MRI scanner was used for the measurements. A 2D spin-echo FLASH sequence was used to record a




Figure 4: Description of the rubber phantom experiment.

single 3 mm thick slice with a matrix size of 256×256 giving in-plane dimensions $0.27 \text{ mm} \times 0.27 \text{ mm}$ for the *in vitro* experiments and $0.586 \text{ mm} \times 0.586 \text{ mm}$ for the *in vivo* study. The image acquisition was synchronized either with the signal command for the phantom or with the physiological signal for the subjects. Regarding the physiological signal, infrared spectroscopy was used to measure the blood flow in one of the subject's finger. We obtained $N_t = 50$ and $N_t = 43$ snapshots for the phantom and for the subjects respectively, evenly distributed throughout the cardiac cycle. It corresponds to a mean sampling frequency of 50 Hz and 40 Hz, respectively. The magnitude of the signal was digitized with a 12 bits resolution (integer between 0 and 4 095).

Choice of experimental images The raw magnitude image was a 3D array of size $256 \times 256 \times N_t$, denoted $I_n^{raw}(\xi_l)$ for l varying from 0 to 65 535, and n varying from 0 to $N_t - 1$. The images were filtered along the time axis using a Gaussian filter with a kernel size $\sigma = 1.5$ to produce $I_n(\xi_k)$.

A region of interest of 63×63 pixels was selected around the vessel. The first image $I_0(\xi_k)$ with k varying from 0 to 3 968 was chosen as the initial image. The deformed image was selected through the images $I_n(\xi_k)$ for $1 \leq n \leq N_t - 1$ according to the following definition:

$$\max \left(\left| I_n(\xi_k) - I_0(\xi_k) \right| \right) \quad (8)$$

An example of experimental images is shown in Fig. 5.




Figure 5: Example of a magnitude image obtained by PC MRI. The raw image is filtered after having chosen a region of interest.

Choice of control pixels A binary mask was defined before the identification procedure to discriminate the control pixels (see Fig. 2): (i) pixels which contain the moving contours were selected automatically. (ii) Additional control pixels were added by dilating the previous binary mask according to:

$$M_{(ii)} = M_{(i)} \oplus S \quad (9)$$

where $M_{(ii)}$ is the final binary mask used; $M_{(i)}$ is the binary mask at step (i); and S is a square structuring element of size 3×3 with $S_{ij} = 1$. The choice of a limited set of pixels allows to improve the sensitivity of the cost function.

2.3.3 Description of the FE models

The 2D FE geometry was either derived from the known geometry of the phantom or segmented from PC images. The contours of the artery were extracted in order to separate the arterial wall from the surrounding tissue and then separately estimate the elastic moduli of each. Quadratic triangular elements were used for the mesh generation (we previously showed that this type of elements provides accurate displacement fields for this type of problems [37]). A uniform pressure was applied to the phantom or artery inner walls. The implicit solver of Abaqus© [38] was used.

Phantom A 2D cross section of the phantom was modeled. It was made up of 1 100 finite elements surrounded by approximately 5 000 similar elements mimicking the fluid.

The boundary conditions were enforced in the experiment. A maximal pressure of 9.33 kPa (70 mmHg) was applied on the phantom's inner wall. The free edges of the surrounding box were fixed in the FE model (see Fig. 4).

Linear elastic material properties were used for the phantom. The Poisson's ratio was set to 0.49 while the elastic modulus has to be recovered. The external water had no mechanical influence because the external pressure applied on the phantom was constant during the test. It was nonetheless incorporated into the model in order to provide the necessary external displacement fields. Linear elasticity was considered with a Poisson's ratio of 0 and an elastic modulus of 10^{-6} kPa.

Subjects The FE geometry of each subject (see Fig. 6) was derived from the PC MR images. The contours of the neck, trachea, vertebral column and jugular vein were segmented manually from the global undeformed image (see Fig. 5a). The inner wall of the artery was recovered automatically using an algorithm based on watershed and Fourier polynomial descriptors: (i) gradients of the mean magnitude along the cardiac cycle were calculated using finite differences; (ii) watershed sources were chosen in the artery's lumen, in the vein's lumen and in the surrounding tissue; (iii) "contour pixels" were then deduced; (iv) the final contour was the approximation of these pixels' centers with a first degree Fourier descriptor. The thickness of the artery was measured by echography: the IMT (Intima-Media Thickness) was $e_A = 0.45$ mm




Figure 6: FE meshes were made of triangular quadratic plane strain elements. The arteries were composed of approximately 1 000 elements while the surrounding tissues were composed of approximately 7 000 elements. A uniform pressure was applied to the inner lumen of the carotid artery. Vertical symmetry was assumed. Vertebral column was fixed (see bold black lines).

310 and $e_B = 0.43$ mm for each subject respectively. The outer contour of the artery
 311 was deduced from the inner contour and from the IMT.

312 The artery was represented using approximately 1 000 elements and the
 313 surrounding tissue with about 7 000 elements. Both the artery and the sur-
 314 rounding tissue were assumed linear elastic in the range of loading. Their
 315 Poisson's ratio was set to 0.49 as arteries were quasi incompressible. The elastic

modulus of the artery was unknown and had to be identified while the elastic modulus of the surrounding tissue was set to 30 kPa. The influence of this modulus on the identified elastic modulus of the artery is studied herein (see Section “Results”).

Regarding the boundary conditions, sagittal symmetry conditions were imposed for each model (horizontal symmetry). The vertebral column was assumed undeformable and the corresponding contour was fixed. Other contours were left free. The mean measured pulse pressure (differential pressure between diastolic and systolic states), measured before and after the exam with a digital sphygmomanometer, was applied to the artery’s lumen ($\Delta P_A = 11.33 \pm 0.13$ kPa with $P_A^{diastole} = 72$ mmHg and $P_A^{systole} = 157$ mmHg, and $\Delta P_B = 6.67 \pm 0.13$ kPa with $P_B^{diastole} = 84$ mmHg and $P_B^{systole} = 134$ mmHg). As a first approximation, we assumed that the blood pressure in the carotid was equal to the blood pressure in the brachial artery. The values of these pulse pressures were discussed below. The reference state was set to the diastolic state.

2.3.4 Definition of the parametric study

A parametric study was carried out to evaluate the sensitivity of the described methodology on its internal parameters. The uniqueness of the solution was verified by changing the initial elastic modulus. The influences of (i) the choice of control pixels, (ii) potential errors on the estimation of physical parameters (mechanical properties of the surrounded tissues or measured pressure) and (iii) potential errors on the geometry of the artery were studied. The reference configuration is described in Table 2.

3 Results

In this section we focus on the final identified elastic moduli and on the values of the cost function at the optima. Results are given as mean $\pm$ standard deviation. A full identification procedure requires approximately 5 min (8 Levenberg-Marquardt iterations).

Table 2: Reference test configuration. Tests were conducted by modifying one parameter at a time.

	Phantom	Subject A	Subject B
Initial elastic modulus of the vessel	300 kPa	500 kPa	500 kPa
Choice of control pixels	automatic ($N = 704$)	automatic ($N^A = 130$)	automatic ($N^B = 120$)
Smoothing filter	temporal Gaussian with $\sigma = 1.5$	temporal Gaussian with $\sigma = 1.5$	temporal Gaussian with $\sigma = 1.5$
Pressure applied	9.33 kPa	11.33 kPa	6.666 kPa
Elastic modulus of the surrounding tissue	0 kPa	30 kPa	30 kPa
Thickness of the vessel considered	1.5 mm	0.45 mm	0.43 mm

3.1 Validation of the MB-FEMU method with the phantom

3.1.1 Validation of the value of the identified properties

A rubber phantom was used to validate the method. Its elastic properties in the range of loading were determined by an uni-axial tensile test (elastic modulus $E = 250$ kPa and Poisson's ratio $\nu = 0.49$). 32 tests combining different choices of control pixels, different initial elastic moduli and different smoothing parameters were conducted. The details of these results are presented herein. The mean identified elastic modulus is $E = 274$ kPa ± 19.8 which corresponds to an average error $\varepsilon = 9.6\%$.

3.1.2 Uniqueness of solution

Our identification is based on solving the minimization problem (see Equation 4) with the Levenberg-Marquardt algorithm which is efficient for least squares formulations but which may get trapped at local minima of J_2 . In order to study the uniqueness of the solution to the minimization problem, five different initial values of the elastic modulus between 150 and 350 kPa have been input to the identification algorithm. Results are $E_{mean} = 270$ kPa ± 1.09 and $J_2^{mean} = 3.34 \cdot 10^4 \pm 2.64$. Note that the standard deviations are very low which means that the identification is marginally affected by a change of the initial value of the elastic modulus.

3.1.3 Effect of the choice of control pixels


The choice of the pixels considered in the cost function may have an influence on the identified elastic modulus. The control pixels can be either selected automatically using the procedure described previously or chosen manually by defining a binary mask which contains the vessel. In this section, 15 different binary masks were tested. For each one of them, a different number of control pixels was assigned (from 704 to 1320 control pixels). Results are $E_{mean} = 280 \text{ kPa} \pm 4.93$ which means that the variations of the binary mask don't affect the identified modulus. Note nevertheless that the best identified elastic modulus occurs when the control pixels are chosen automatically (704 control pixels; $E = 271 \text{ kPa}$). The value of the cost function at optimum tends to slightly decreases with the increase of the number of control pixels (results are $J_2^{mean} = 2.67 \cdot 10^4 \pm 0.23 \cdot 10^4$). However there is no correlation between this decrease of values and a change on the identified moduli.

3.1.4 Effect of the smoothing filter


We also studied the effect of filtering the data before the identification procedure (see Fig. 7a and 7b). Two types of filters were tested: temporal filters (Gaussian filter as described previously with various kernel sizes), and spatial filters (Wiener and median with different kernel sizes). The overall results accounting for both filters are $E_{mean} = 268 \text{ kPa} \pm 31.09$ and $J_2^{mean} = 3.27 \cdot 10^4 \pm 0.88 \cdot 10^4$. The temporal filter tends to increase the estimated elastic modulus (Fig. 7a). The elastic modulus ranges from $E = 265 \text{ kPa}$ to $E = 341 \text{ kPa}$ while the value of the cost function decreases linearly with the increase of kernel size σ (Fig. 7b). Median spatial filtering yields moduli close to the tensile test reference value (kernels 3×3 and 5×5 : $E = 256 \text{ kPa}$ and $E = 254 \text{ kPa}$). Wiener spatial filtering underestimates the elastic modulus (kernels 3×3 and 5×5 : $E = 231 \text{ kPa}$ and $E = 217 \text{ kPa}$).

3.1.5 Effect of measurement errors in the pressure


Errors of plus or minus 15 % were considered for the nominal pressure P in the phantom (see Fig. 8a). Results show that the value of the cost function slightly decreases with the increase of pressure. Nevertheless, the standard deviation is very low which makes the cost function almost unaffected ($J_2^{mean} =$



(a) Identified elastic modulus for the phantom.




(b) Values of the cost function at the optima for the phantom.




(c) Identified elastic modulus for subject A.

Figure 7: Effect of the raw images filtering. The filter was applied to both the initial and the deformed images. Final identified elastic moduli (squares, left axis) and the values of the cost function at the optima (diamonds, right axis) are shown for the phantom and the two subjects.


396 $3.34 \cdot 10^4 \pm 3.74$). The identified elastic modulus ranges from $E_{min} = 231$ kPa
 397 to $E_{max} = 307$ kPa which corresponds to errors of -7.6% to $+22.8\%$. There
 398 is a clear linear relationship between the pressure and the identified modulus



(d) Values of the cost function at the optima for subject A.




(e) Identified elastic modulus for subject B.




(f) Values of the cost function at the optima for subject B.

Figure 7: (continued with last three figures)


399 ($r = 0.99$).



(a) Phantom.



(b) Subject A.



(c) Subject B.

Figure 8: Effect of the applied pressure in the FE model for the phantom and both subjects. Squares and diamonds correspond to the identified elastic moduli (left axis) and to the values of the cost function at the optima (right axis), respectively.

400 3.2 Healthy subjects study

401 For the subjects, the same numerical experiments were conducted for the vari-
402 ations of the identification parameters and for the effect of the estimation of
403 pressure on the identified modulus.

404 3.2.1 Identification parameter variations

405 **Uniqueness of solution** The initial elastic modulus of the artery was changed
406 six times to study the uniqueness of the solution in the range 400 to 900 kPa.
407 Results are $E_{mean}^A = 625 \text{ kPa} \pm 0.27$ and $E_{mean}^B = 586 \text{ kPa} \pm 3.25$. The value of
408 the cost function at the optimum is unaffected: $J_{2A}^{mean} = 166.38 \pm 5.10^{-4}$ and
409 $J_{2B}^{mean} = 187.48 \pm 0.10$. As well as the results on the phantom, the identified
410 elastic modulus is independent of the initial elastic modulus.


411 **Effect of the choice of control pixels** The choice of different sets of control
412 pixels doesn't affect the identified elastic modulus (results are $E_{mean}^A =$
413 $630 \text{ kPa} \pm 13.50$ and $E_{mean}^B = 589 \text{ kPa} \pm 9.66$ for 15 sets of control pixels which
414 contain from 101/89 to 220/167 pixels for A/B, respectively). In each case the
415 identified elastic moduli are similar while the values of the cost function de-
416 crease with the increase of the number of control pixels. This observation is
417 similar to what we found with the phantom.

418 **Effect of the smoothing filter** Results are summarized in Fig. 7c, 7d, 7e and
419 7f. The average results are $E_{mean}^A = 602 \text{ kPa} \pm 31.45$ and $E_{mean}^B = 598 \text{ kPa} \pm$
420 33.38 . The identified moduli range from 555 kPa to 664 kPa for subject A and
421 from 547 kPa to 670 kPa for subject B. The minimum and maximum values
422 were obtained in both cases for the 3×3 median filter and for the 2.5 Gaussian
423 filter, respectively. The value of the cost function at optimum decreases with
424 the augmentation of the "strength" of the filtering although the comparison
425 between the temporal and the spatial filters is not easy. However the filtering
426 of images always reduces the value of the cost function at optimum compared
427 to using the raw images.


3.2.2 Effect of the estimation of physical parameters

Effect of measurement errors in the pressure Pulse blood pressure (difference between maximal and minimal pressures) was measured with a digital sphygmomanometer ($\Delta P_A = 11.33$ kPa and $\Delta P_B = 6.67$ kPa). Similarly to the phantom validation, a variation of plus or minus 15 % were considered for the nominal pressure ΔP applied in the FE model. Results are detailed in Fig. 8b and 8c. Identified elastic moduli range from $E_{min}^A = 517$ kPa to $E_{max}^A = 734$ kPa for subject A and from $E_{min}^B = 487$ kPa to $E_{max}^B = 691$ kPa for subject B. The relationship between the pressure and the identified modulus is linear which confirms the results from the phantom. Nevertheless and contrary to the phantom study, the cost function value at optimum tends to increase linearly with a higher pressure in both subjects.

Effect of the estimation of the mechanical properties of the surrounding tissue The artery was assumed to be surrounded by a soft elastic media. Its elastic properties were supposed to be known. The choice of its elastic properties may have an influence on the identification of the artery's stiffness. For instance, in Fig. 9, we show how the choice of the elastic modulus of the surrounding tissue affects the identification of the elastic modulus of the artery. As the surrounding tissue stiffness is fixed to increasingly higher values, the identified elastic modulus of the artery decreases. The values of the cost function exhibits a second order polynomial shape with a minimum reached for $E_{surrounding}^A \approx 30$ kPa and $E_{surrounding}^B \approx 150$ kPa. This shows that the cost function is sensitive to the stiffness of the surrounding tissue. The modification of the elasticity of the surrounding tissue corresponds somehow to a modification of the pressure applied at the outer boundary of the artery. Previous results on the effect of the pressure shows a linear relationship between the identified moduli or the value of the cost function at optimum and the inner pressure applied. Here the correlation between the elastic modulus of the surrounding tissue and the identified artery's modulus is linear in both cases ($r^A = 0.98$ and $r^B = 0.94$). For the cost function however, the relationship with the elastic properties of the surrounding tissue seems to be a quadratic form.



(a) Subject A.




(b) Subject B.

Figure 9: Influence of the elastic modulus of the surrounding tissue on the identified elastic modulus of the artery (squares, left axis) and on the value of the cost function value at the optima (diamonds, right axis).


3.2.3 Effects of errors in the geometry

Errors in the artery's position The FE model geometry was defined from the undeformed image. Potential errors in the positioning of the artery were considered (see Fig. 10). No tendency appears with the X-axis or Y-axis offsets neither for the identified moduli nor for the values of the cost function at the optima.

Effect of measurement errors in the artery's diameter The inner lumen of the artery is determined automatically (see section "Methodology"). This section aims at studying the effect of a wrong estimation of the inner diameter due to the algorithm described previously. Errors of plus or minus 10 % of the nominal diameter were considered (see Fig. 11). For subject B, the diameter of the artery could not be increased further because it was already in contact



(a) Error of the position of the artery along the X-axis for subject A.




(b) Error of the position of the artery along the Y-axis for subject A.


Figure 10: Study of potential errors in the positioning of the artery in the FE model. Squares represent the final identified elastic moduli (left axis) and diamonds the values of the cost function at the optima (right axis).

with the jugular vein. The resulting elastic moduli identified lie within the following ranges respectively for subjects A and B : $E_{range}^A = [553 \text{ kPa}, 659 \text{ kPa}]$ and $E_{range}^B = [487 \text{ kPa}, 587 \text{ kPa}]$. The identified moduli increase with diameter. The value of the cost function at the optimum decreases when the diameter is reduced.

Effect of measurement errors in the artery's thickness The thickness of the artery was measured with ultrasounds. Errors of plus or minus 10 % in the thickness of the artery were considered (see Fig. 12). The elastic moduli ranges resulting from the identification process are respectively $E_{range}^A = [566 \text{ kPa}, 696 \text{ kPa}]$ and $E_{range}^B = [533 \text{ kPa}, 660 \text{ kPa}]$. The correlation between the identified modulus and the thickness of the artery is linear, with a decrease of the identified



(c) Error of the position of the artery along the X-axis for subject B.



(d) Error of the position of the artery along the Y-axis for subject B.

Figure 10: (continued with last two figures)


modulus when the thickness increases. The same effect can be observed for the value of the cost function at optimum, although the range of variation is very low.

4 Discussion


The discussion is organized in two parts. The first part focuses on the robustness of the MB-FEMU method whereas the second part discusses the elastic properties of arteries obtained.

4.1 Robustness of the MB-FEMU method

In the three cases studied in the present work (the phantom and the two subjects) the robustness of the MB-FEMU method was tested with respect to dif-



(a) Subject A.



(b) Subject B.

Figure 11: Effect of the measurement error of the artery's diameter. Squares represents the elastic moduli (left axis) and diamonds the values of the cost function at the optima (right axis).


ferent sources of uncertainty and artifacts. The normalized input parameters were noted X_i . Assuming that the relationship between the identified elastic modulus and parameter X_i is linear, it is possible to rank the influence of each parameter by calculating the following normalized regression coefficient:

$$a_i = \left| \frac{Cov(X_i, E)}{Var(X_i)} \right| \quad (10)$$


where E is the identified elastic modulus and X_i is the normalized parameter i ($X_i \in [0, 1]$). The higher the coefficient a_i , the higher the influence of the parameter on the identified variable. Results are shown in Table 3.

The sources of uncertainty and artifacts can be sorted into three groups according to their relative influence:

1. The most important parameter is the pressure applied on the vein wall (i) by the surrounding soft tissue on the outer surface, (ii) and by the



(a) Subject A.



(b) Subject B.

Figure 12: Effect of the measurement error of the artery's thickness. External diameter remains constant in all cases and only the internal diameter was modified. The identified elastic moduli (left axis) and the values of the cost function at the optima (right axis) are represented as squares and diamonds, respectively.

blood circulation on the inner surface. It is particularly obvious for subject B (see Fig. 9b). The estimation of the stiffness of the surrounding tissue and of the pressure applied is crucial to accurately identify the elastic modulus of the artery. It can be observed in Fig. 9a and 9b that the value of the cost function evolves when the stiffness of the surrounding tissue is increased in the model. A minimum is reached for the values $E_{surrounding}^A \approx 30$ kPa and $E_{surrounding}^B \approx 150$ kPa. This result suggests that the identification process can be used to determine the elastic moduli of both the artery and the surrounding tissue simultaneously. In this case the uniqueness of the solution would not be guaranteed and the computation time would increase. The asymmetric displacement of the artery between the configurations (see Fig. 13) suggests the presence of a stiffer

515 tissue on the right hand side of the artery (tendons for instance).

516 2. Less important but not negligible are the sources of uncertainty that may
517 arise from the geometry definition. On one hand, we observe that the
518 error made on the artery position does not significantly significantly af-
519 fect the identified elastic modulus. This constitutes one of the strength
520 of the MB-FEMU methodology. On the other hand, the diameter and the
521 thickness have a moderate but significant influence and thus, must be
522 characterized accurately. It is possible to define a parametric model as
523 in [30] to recover simultaneously both the elastic modulus and the ge-
524 ometry of the artery (in this case the uniqueness of the solution would
525 not be guaranteed and the computation time would probably increase).
526 Note that the entire thickness of the artery is not taken into account here
527 (only the Intima-Media Thickness is considered) but that extra informa-
528 tion on the total thickness could help reduce this uncertainty as the total
529 wall thickness can be measured by ultrasounds or MRI with a Turbo Spin
530 Echo flash sequence (Blackblood-FATSAT) [39].

531 3. The least important sources of uncertainty results from the so-called iden-
532 tification parameters: (i) position and number of control pixels, (ii) initial
533 elastic modulus. The induced errors can therefore be neglected as com-
534 pared to those induced by the other sources. Identification results (very
535 close one to the other) obtained from different sets of initial parameters
536 are a good hint that the solution of the problem is very likely to be unique.
537 Only the kernel sizes of the filters must be chosen with care because an
538 increase of the kernel size decreases the apparent deformation of the im-
539 ages between the initial and deformed configurations (see Fig. 7).

540 **4.2 Identified elastic moduli**

541 Arterial compliance has been widely studied over the last decades. Stiffness
542 values found in the literature are widespread for two main reasons: (i) subjects
543 doing the experiment are different (important inter-individual differences), (ii)
544 methods are different. The review of [40] on the calculation of arterial stiffness
545 indexes in clinical practice mentioned nine different stiffness indexes from the
546 “elastic modulus” to the “stiffness index”. Recently [41] listed existing meth-




Figure 13: Difference between experimental deformed and initial images for subject B. A white crown appears around the artery. The asymmetry of displacement of the artery can be due to the high stiffness of tissues at the right hand side (tendons, etc.).

Table 3: Normalized regression coefficients a_i , measuring the influence of the different parameters on the identified elastic modulus of the artery, are calculated with Equation (10).

	Subject A	Subject B
E surrounding tissue	339.87	487.76
Pressure	192.64	181.41
Thickness	111.31	108.76
Diameter	90.64	76.96
Control pixels	43.61	26.40
Translation dX	11.85	44.48
Translation dY	8.40	10.18
Initial elastic modulus	0.39	2.84

ods which are used in clinical practice to assess arterial stiffness. The same author defined in [16] the incremental modulus of an artery such as:

$$E_{inc} = \left(3 \left(1 + \frac{R_d^2}{(R_d + IMT)^2 - R_d^2} \right) \right) \cdot \frac{R_d^2 \cdot \Delta P}{R_s^2 - R_d^2} \quad (11)$$

where R_s and R_d are respectively the systolic (highest pressure) and diastolic (lowest pressure) radii of the artery; IMT is the Intima-Media Thickness; ΔP is the pulse pressure between diastole and systole. This formula was determined for an isotropic, incompressible, semi closed thick cylinder with an inner pressure applied.

554 4.2.1 Comparison with ultrasounds measurements

555 The MB-FEMU method presented in this paper yields elastic moduli of 625 kPa
 556 and 587 kPa when applied to two different young healthy subjects. In ad-
 557 dition an echography was conducted on these subjects in order to acquire
 558 movies of their carotid arteries during three heart beats. Initial and deformed
 559 images were extracted manually for both subjects (see Fig. 14). Pulse blood
 560 pressures were measured during the exam with a digital sphygmomanome-
 561 ter: $\Delta P^A = 8.27$ kPa and $\Delta P^B = 6.40$ kPa. Arteries lumens were segmented
 562 manually on the initial and deformed ultrasounds images and the correspond-
 563 ing incremental modulus was calculated with Equation (11). The implemen-
 564 tation of the MB-FEMU method was based on the assumption of plane strain
 565 in a cross section of the artery which is different from the assumptions used
 566 for deriving Equation (11). In order to make some equivalence between the
 567 methods, initial and deformed areas (and subsequently radii) were segmented
 568 directly from the results of the FE computation at the optima for calculating
 569 E_{inc} using Equation (11). The pressures used in the case of the FE model were
 570 $\{\Delta P_A, \Delta P_B\} = \{11.33 \text{ kPa}, 6.666 \text{ kPa}\}$. Results are $E_{inc}^{echo} = 374$ kPa; $E_{inc}^{FE} =$
 571 496 kPa and $E_{inc}^{echo} = 434$ kPa; $E_{inc}^{FE} = 509$ kPa for the subjects A and B, respec-
 572 tively.

573 It can be noted that the results obtained with both imaging systems are
 574 consistent with the values reported in the literature (see below) though small
 575 differences can be noted. The differences between the two imaging systems
 576 can be justified by three main reasons. First, the boundary conditions enforced
 577 in each model are different because the ultrasound probe applied a pressure
 578 on the neck during the exam. Secondly, the height on the neck at which the
 579 ultrasound images were taken is not managed and is potentially different from
 580 the height of the MRI. Third, the blood pressure of the patients were sensibly
 581 different during the MRI and the echography exams ($\Delta P_A^{echo} = 8.27$ kPa vs
 582 $\Delta P_A^{MRI} = 11.33$ kPa; $\Delta P_B^{echo} = 6.40$ kPa vs $\Delta P_B^{MRI} = 6.666$ kPa). This poten-
 583 tially could impact the behavior of the arterial wall which is non-linear. Dif-
 584 ferences in pulse pressures between MRI and echography as well as higher
 585 pulse pressure in MRI were observed several times for some subjects. It can
 586 be explained by some psychological stress induced by the noisy and confined
 587 environment.

4.2.2 Comparison with the literature




Figure 14: Ultrasounds image (echography) of the carotid arteries of both subjects. Initial (diastole) and deformed (systole) images were extracted manually from movies taken during three heart beats.

Artery's stiffness A survey of different *in vivo* elastic moduli identified by other authors is reported in Table 1. Different imaging systems were used: B-mode echography (see Fig. 14), Radio-Frequency UltraSounds signal, IntraVascular UltraSounds or MRI. The reported elastic moduli are in good agreement with our results in most of the studies except for: (i) [16] ($E = 2680 \pm 1810$ kPa) where the difference may be explained by the location of the artery (radial artery at the wrist instead of carotid artery); (ii) [25] ($E = 1400$ kPa) where the elastic modulus depends on the excitation frequency of the ultrasounds because a viscoelastic behavior is considered; (iii) [42] ($E = 188$ kPa) where a different artery was also studied (pathological coronary artery); (iv) [18] ($E_{inc} =$

187 $\pm$ 67 kPa) where the subjects were significantly younger (9-year-old); (v)
 [13] ($E = 99$ kPa) where the authors suggest that the small elastic modulus
 found may be related to a bad estimation of the thickness.

MB-FEMU method vs other methods The methods used in the literature for
 quantifying the artery's stiffness *in vivo* are generally based on the variation
 of the artery's diameter during heart beats. In these cases authors generally
 used theoretical formula by considering several mechanical and geometrical
 assumptions: [16, 17, 18, 19, 20, 24] calculated the incremental modulus accord-
 ing to Equation (11); [15, 21] applied the Laplace law by considering a 2D plane
 stress artery; [22, 25] derived an analytical formula of the circumferential elas-
 tic modulus using the assumptions that the artery is incompressible, isotropic
 and that the arterial pressure decreases linearly from the artery's inner diam-
 eter to adventitia). Other studies derived the artery's stiffness from velocity
 Phase Contrast MR images and Moens-Korteweg equation which relates the
 stiffness of a round artery to the pulse wave velocity and a round artery's stiff-
 ness [13] or studied the propagation of shear waves induced by an ultrasonic
 device to deduce shear modulus (denoted μ) and the elastic modulus (denoted
 E) according to $E = 3 \cdot \mu$ [14].

Few studies measured the stiffness of the human carotid arteries by taking
 into account the surrounding tissue and the environment of the artery (verte-
 bral column, tendons, etc.). [42] used a strain-based FEMU method to recover
 both the contours and the elastic moduli of the constituents of an atheroscle-
 rotic plaque. This method is very invasive since the IVUS technique was em-
 ployed to obtain strain fields in the wall of the artery. [43] used another FEMU
 method with a cost function defined as a difference of areas but the areas were
 segmented manually from Phase Contrast images which may be critical with
 regard to the pixel size.

Our method and these alternative FEMU methods have in common the ad-
 vantage that they are not based on strong mechanical assumptions. Indeed the
 arteries are known to behave highly non-linearly ([5, 44, 45, 12]). Only the tan-
 gent (or incremental) behavior is considered for the moment by linearizing the
 stress/strain relationship of arteries between diastole and systole. The identi-
 fied elastic modulus is therefore a function of the diastolic and pulse pressures
 due to the intrinsic non-linear behavior of arteries. More complex constitu-

633 tive equations could potentially be identified by including into the FE model
634 anisotropy, viscoelastic effects, heterogeneity, etc. The limit is the question of
635 the identifiability and uniqueness of the solution that should be investigated
636 for each other constitutive equations. Compared to alternative FEMU meth-
637 ods reported in the literature, the MB-FEMU method has also the advantage of
638 being compatible with non-invasive imaging techniques such as MRI. Indeed,
639 clinical MRI was used and the impact on the patient is very limited (standard
640 and routinely-used exam for a patient before carotid endarterectomy).

641 Close to our method, the “hyper elastic warping” is a technique that was
642 developed a few years ago ([46, 47, 48]). The objective is to perform image
643 registration in order to quantify the kinematics of the transformation within
644 a deforming body. Our approach is similar in the sense that we perform im-
645 age registration as well, with the difference that our main objective is not to
646 quantify the kinematics of the deformations but to identify the constitutive
647 properties of the deforming body.

648 5 Conclusions

649 This study shows that identifying elastic moduli in common carotid arteries *in*
650 *vivo* from cine-MRI images is possible. An original method named the Mag-
651 nitude Based Finite Element Model Updating (MB-FEMU) methodology has
652 been thoroughly described, validated on a rubber phantom and applied to two
653 healthy subjects. In our implementation MRI and Phase Contrast magnitude
654 images were used to provide the experimental data. The phase information of
655 the Phase Contrast sequence could be used for improving the boundary con-
656 ditions in the FE model as a perspective. We would like to bring the reader’s
657 attention to the fact that the MB-FEMU can be easily extended to other imaging
658 techniques. The identified elastic moduli on healthy subjects are in good agree-
659 ment with those found in the literature, especially if we consider the differences
660 in hypotheses made that we discussed in the article. The MB-FEMU method
661 offers several advantages compared to existing methods: it makes a maximum
662 use of experimental data since the FE results are used only as a perturbation of
663 the experimental acquisitions; it is possible to extract arterial stiffness from a
664 routine MRI exam on any patient; it can readily be extended to more complex
665 tissue constitutive behaviors. Heterogeneous geometries can be possibly iden-

666 tified if an accurate FE model can be created, although the uniqueness of the
 667 minimization problem will have to be verified. In the future this method will
 668 be employed to study the modification of the mechanical behavior of arteries
 669 with age and cardiovascular disease. Applications for the diagnosis of vul-
 670 nerable atherosclerotic plaques in the carotid is probably the most promising
 671 challenge for this method.

672 Acknowledgment

673 We are grateful to the two volunteers for their time and implication.
 674 We would like to thank IEEE Transactions on Medical Imaging (please visit
 675 <http://ieeexplore.ieee.org/xpl/RecentIssue.jsp?punumber=42>)

676 References

- 677 [1] B. A. Kingwell, T. K. Wadell, T. L. Medley, J. D. Cameron, and A. M. Dart,
 678 "Large artery stiffness predicts ischemic threshold in patients with coro-
 679 nary artery disease," *Journal of the American College of Cardiology*, vol. 40,
 680 no. 4, pp. 773–779, 2002.
- 681 [2] R. T. Lee, A. J. Grodzinsky, E. H. Frank, R. D. Kamm, and F. J. Schoen,
 682 "Structure-dependent dynamic mechanical behavior of fibrous caps from
 683 human atherosclerotic plaques," *Circulation*, vol. 83, no. 5, pp. 1764–1770,
 684 1991.
- 685 [3] A. Duprey, K. Khanafer, M. Schlicht, S. Avril, D. Williams, and R. Berguer,
 686 "In vitro characterisation of physiological and maximum elastic modu-
 687 lus of ascending thoracic aortic aneurysms using uniaxial tensile testing,"
 688 *European Journal of Vascular and Endovascular Surgery*, vol. 39, no. 6, pp.
 689 700–707, 2010.
- 690 [4] D. Beattie, C. Xu, R. Vito, S. Glagov, and M. C. Whang, "Mechanical anal-
 691 ysis of heterogeneous, atherosclerotic human aorta," *Journal of Biomechan-
 692 ical Engineering*, vol. 120, no. 5, pp. 602–607, 1998.
- 693 [5] Y. C. Fung, *Biomechanics : Mechanical properties of living tissues*, 2nd ed.
 694 New-York: Springer-Verlag, 1993.

- 695 [6] G. A. Holzapfel, G. Sommer, and P. Regitnig, "Anisotropic Mechanical
696 Properties of Tissue Components in Human Atherosclerotic Plaques,"
697 *Journal of Biomechanical Engineering*, vol. 126, no. 5, pp. 657–665, 2004.
- 698 [7] C. L. Lendon, M. J. Davies, P. D. Richardson, and G. V. R. Born, "Testing of
699 small connective tissue specimens for the determination of the mechanical
700 behaviour of atherosclerotic plaques," *Journal of Biomechanical Engineering*,
701 vol. 15, pp. 27–33, 1993.
- 702 [8] H. M. Loree, B. J. Tobias, L. J. Gibson, R. D. Kamm, D. M. Small, and R. T.
703 Lee, "Mechanical properties of model atherosclerotic lesion lipid pools,"
704 *Arteriosclerosis Thrombosis and Vascular Biology*, vol. 14, pp. 230–234, 1994.
- 705 [9] S. R. H. Barrett, M. P. F. Sutcliffe, S. Howarth, Z.-Y. Li, and J. H. Gillard,
706 "Experimental measurement of the mechanical properties of carotid
707 atherothrombotic plaque fibrous cap," *Journal of Biomechanics*, vol. 42, pp.
708 1650–1655, 2009.
- 709 [10] J. O. V. Delgadillo, "Mechanical properties of arterial wall," Ph.D. disser-
710 tation, University of British Columbia, 2008.
- 711 [11] G. A. Holzapfel, "Biomechanics of Soft Tissues with Application to Ar-
712 terial Walls," in *Mathematical and Computational Modeling of Biological Sys-*
713 *tems*, J. A. C. Martins and E. A. C. Borges Pires, Eds. Coimbra, Portugal:
714 Centro Internacional de Matemática CIM, 2002, pp. 1–37.
- 715 [12] E. Peña, V. Alastrué, A. Laborda, M. A. Martínez, and M. Doblaré, "A
716 constitutive formulation of vascular tissue mechanics including viscoelas-
717 ticity and softening behaviour," *Journal of Biomechanics*, vol. 43, no. 5, pp.
718 984–989, 2010.
- 719 [13] S. Avril, J. M. Huntley, and R. Cusack, "In-vivo measurements of blood
720 viscosity and wall stiffness in the carotid using PC-MRI," *European Journal*
721 *of Computational Mechanics*, vol. 18, pp. 9–20, 2009.
- 722 [14] M. Couade, M. Pernot, C. Prada, E. Messas, J. Emmerich, P. Bruneval,
723 A. Criton, M. Fink, and M. Tanter, "Quantitative assessment of arterial
724 wall biomechanical properties using shear wave imaging," *Ultrasound in*
725 *Medicine & Biology*, vol. 36, no. 10, pp. 1662–1676, 2010.

- 726 [15] W. A. Riley, R. W. Barnes, G. W. Evans, and G. L. Burke, "Ultrasonic Mea-
727 surement of the Elastic Modulus of the Common Carotid Artery," *Stroke*,
728 vol. 23, no. 7, pp. 952–956, 1992.
- 729 [16] S. Laurent, X. Girerd, J. J. Mourad, P. Lacolley, L. Beck, P. Boutouyrie, J. P.
730 Mignot, and M. Safar, "Elastic modulus of the radial artery wall material
731 is not increased in patients with essential hypertension," *Arteriosclerosis*
732 *and Thrombosis*, vol. 14, no. 7, pp. 1223–1231, 1994.
- 733 [17] P. J. Brands, A. P. G. Hoeks, J. Willigers, C. Willekes, and R. S. Reneman,
734 "An integrated system for the non-invasive assessment of vessel wall and
735 hemodynamic properties of large arteries by means of ultrasound," *Euro-*
736 *pean Journal of Ultrasounds*, vol. 9, pp. 257–266, 1999.
- 737 [18] Y. Aggoun, D. Sidi, B. I. Levy, S. Lyonnet, J. Kachaner, and D. Bonnet, "Me-
738 chanical properties of the common carotid artery in Williams syndrome,"
739 *Heart*, vol. 84, pp. 290–293, 2000.
- 740 [19] C. Bussy, P. Boutouyrie, P. Lacolley, P. Challande, and S. Laurent, "Intrinsic
741 stiffness of the carotid arterial wall material in essential hypertensives,"
742 *Hypertension*, vol. 35, no. 5, pp. 1049–1054, 2000.
- 743 [20] P. Boutouyrie, D. P. Germain, A.-I. Tropeano, B. Laloux, F. Carenzi,
744 M. Zidi, X. Jeunemaitre, and S. Laurent, "Compressibility of the
745 Carotid Artery in Patients With Pseudoxanthoma Elasticum," *Hyperten-*
746 *sion*, vol. 38, no. 5, pp. 1181–1184, 2001.
- 747 [21] R. H. Selzer, W. J. Mack, P. L. Lee, H. Kwong-Fu, and H. N. Hodis,
748 "Improved common carotid elasticity and intima-media thickness mea-
749 surements from computer analysis of sequential ultrasound frames,"
750 *Atherosclerosis*, vol. 154, no. 1, pp. 185–193, 2001.
- 751 [22] H. Kanai, H. Hasegawa, M. Ichiki, F. Tezuka, and Y. Koiwa, "Elastic-
752 ity imaging of atheroma with transcutaneous ultrasound: preliminary
753 study," *Circulation*, vol. 107, no. 24, pp. 3018–3021, 2003.
- 754 [23] C. Stephanis, D. Mourmouras, and D. Tsagadopoulos, "On the elastic
755 properties of arteries," *Journal of Biomechanics*, vol. 36, no. 11, pp. 1727–
756 1731, 2003.

- [24] P. Boutouyrie, D. P. Germain, J.-N. Fiessinger, B. Laloux, J. Perdu, and S. Laurent, "Increased Carotid Wall Stress in Vascular Ehlers-Danlos Syndrome," *Circulation*, vol. 109, no. 12, pp. 1530–1535, 2004.
- [25] H. Hasegawa and H. Kanai, "Measurement of Elastic Moduli of the Arterial Wall at Multiple Frequencies by Remote Actuation for Assessment of Viscoelasticity," *Japanese Journal of Applied Physics*, vol. 43, no. 5B, pp. 3197–3203, 2004.
- [26] I. Masson, "Contribution à la modélisation mécanique du comportement dynamique hyperélastique et anisotrope de la paroi artérielle," Thèse de doctorat, Université Paris 12 Val de Marne, 2008.
- [27] S. Avril, M. Bonnet, A.-S. Bretelle, M. Grédiac, F. Hild, P. Ienny, F. Latourte, D. Lemosse, S. Pagano, E. Pagnacco, and F. Pierron, "Overview of Identification Methods of Mechanical Parameters Based on Full-field Measurements," *Experimental Mechanics*, vol. 48, no. 4, pp. 381–402, 2008.
- [28] K. B. Chandran, J. H. Mun, K. K. Choi, J. S. Chen, A. Hamilton, A. Nagaraj, and D. D. McPherson, "A method for in-vivo analysis for regional arterial wall material property alterations with atherosclerosis: preliminary results," *Medical Engineering & Physics*, vol. 25, no. 4, pp. 289–298, 2003.
- [29] A. J. Hamilton, H. Kim, A. Nagaraj, J.-H. H. Mun, L. L. Yan, S. I. Roth, D. D. McPherson, and K. B. Chandran, "Regional material property alterations in porcine femoral arteries with atheroma development," *Journal of Biomechanics*, vol. 38, no. 12, pp. 2354–2364, 2005.
- [30] R. A. Baldewsing, M. G. Danilouchkine, F. Mastik, J. A. Shaar, P. W. Serruys, and A. F. W. van der Steen, "An Inverse Method for Imaging the Local Elasticity of Atherosclerotic Coronary Plaques," *IEEE Transactions on Information Technology in Biomedicine*, vol. 12, no. 3, pp. 277–289, 2008.
- [31] S. Le Floch, J. Ohayon, P. Tracqui, G. Finet, A. M. Gharib, R. L. Maurice, G. Cloutier, and R. I. Pettigrew, "Vulnerable Atherosclerotic Plaque Elasticity Reconstruction Based on a Segmentation-driven Optimization Procedure using Strain Measurements: Theoretical Framework," *IEEE Transactions on Medical Imaging*, vol. 28, no. 7, pp. 1126–1137, 2009.

- 789 [32] A. P. Lin, E. Bennett, L. E. Wisk, M. Gharib, S. E. Fraser, H. Wen, and
790 E. Benett, "Circumferential Strain in the Wall of the Common Carotid
791 Artery: Comparing Displacement-Encoded and Cine MRI in Volunteers,"
792 *Magnetic resonance in Medicine*, vol. 60, no. 1, pp. 8–13, 2008.
- 793 [33] M. I. Miga, "A new approach to elastography using mutual information
794 and finite elements," *Physics in Medicine and Biology*, vol. 48, no. 4, pp.
795 467–480, 2003.
- 796 [34] C. W. Washington and M. I. Miga, "Modality independent elastography
797 (MIE): a new approach to elasticity imaging," *IEEE Transactions on Medical
798 Imaging*, vol. 23, no. 9, pp. 1117–1128, 2004.
- 799 [35] M. A. Bernstein, K. F. King, and X. J. Zhou, *Handbook of MRI pulse sequences*.
800 Elsevier Inc., 2004.
- 801 [36] F. Guyon and R. Le Riche, "Least Squares Parameter Estimation and the
802 Levenberg-Marquardt Algorithm : Deterministic Analysis, Sensitivities
803 and Numerical Experiments," Institut National des Sciences Appliquées,
804 Tech. Rep., 2000.
- 805 [37] A. Franquet, S. Avril, R. Le Riche, and P. Badel, "Identification of het-
806 erogeneous elastic properties in stenosed arteries: a numerical plane
807 strain study," *Computer Methods in Biomechanics and Biomedical Engineer-
808 ing*, vol. 15, no. 1, pp. 49–58, 2012.
- 809 [38] ABAQUS 6.8, *Abaqus Theory Manual*, Dassault Systèmes Simulia Corp.
810 Providence (RI,USA), Ed., 2008.
- 811 [39] M. R. Skilton, L. Boussel, F. Bonnet, S. Bernard, P. C. Douek, P. Moulin, and
812 A. Serusclat, "Carotid intimamedia and adventitial thickening: Compari-
813 son of new and established ultrasound and magnetic resonance imaging
814 techniques," *Atherosclerosis*, vol. 21, pp. 405–410, 2011.
- 815 [40] I. S. Mackenzie, I. B. Wilkinson, and J. R. Cockcroft, "Assessment of Arte-
816 rial Stiffness in Clinical Practice," *QJM*, vol. 95, no. 2, pp. 67–74, 2002.
- 817 [41] S. Laurent, J. Cockcroft, L. Van Bortel, P. Boutouyrie, C. Giannattasio,
818 D. Hayoz, B. Pannier, C. Vlachopoulos, I. Wilkinson, and H. Struijker-
819 Boudier, "Expert Consensus Document on Arterial Stiffness: Method-

- ological Issues and Clinical Applications," *European Heart Journal*, vol. 27,
no. 21, pp. 2588–2605, 2006.
- [42] R. A. Baldewsing, J. A. Schaar, F. Mastik, C. W. J. Oomens, and A. F. W.
van der Steen, "Assessment of vulnerable plaque composition by match-
ing the deformation of a parametric plaque model to measured plaque
deformation," *IEEE Transactions on Medical Imaging*, vol. 24, no. 4, pp. 514–
528, 2005.
- [43] V. Taviani, M. P. F. Sutcliffe, P. Wong, Z.-Y. Li, V. Young, M. J. Graves, and
J. H. Gillard, "In vivo non-invasive high resolution MR-based method for
the determination of the elastic modulus of arterial vessels," *Annual Inter-
national Conference of the IEEE Engineering in Medicine and Biology Society*,
vol. 2008, pp. 5569–5572, 2008.
- [44] G. A. Holzapfel and H. W. Weizsacker, "Biomechanical behavior of the
arterial wall and its numerical characterization," *Computers in Biology and
Medicine*, vol. 28, no. 4, pp. 377–392, 1998.
- [45] J. D. Humphrey, *Cardiovascular Solid Mechanics : Cells, Tissues and Organs*.
New-York: Springer-Verlag, 2002.
- [46] A. I. Veress, N. Phatak, and J. A. Weiss, "Deformable Image Registra-
tion with Hyperelastic Warping," in *Handbook of Biomedical Image Analysis*,
3rd ed., J. S. Suri, D. L. Wilson, and S. Laxminarayan, Eds. New-York:
Kluwer Academic/Plenum, 2004, pp. 487–534.
- [47] J. A. Weiss and A. I. Veress, "Strain measurement using deformable image
registration," in *Mechanics of Biological Tissue*, G. A. Holzapfel and R. W.
Ogden, Eds., 2006, pp. 489–501.
- [48] N. S. Phatak, S. A. Maas, A. I. Veress, N. A. Pack, E. V. R. Di Bella, and J. A.
Weiss, "Strain measurmeent in the left ventricule during systoo le with
deformable image registration," *Medical Image Analysis*, vol. 13, no. 2, pp.
354–361, 2009.