



**HAL**  
open science

## Gate simulation of a complete proton treatment combined with prompt-gamma monitoring

P. Gueth, L. Grevillot, D. Dauvergne, N. Freud, J.-M. Létang, C. Ray, E.  
Testa, D. Sarrut

► **To cite this version:**

P. Gueth, L. Grevillot, D. Dauvergne, N. Freud, J.-M. Létang, et al.. Gate simulation of a complete proton treatment combined with prompt-gamma monitoring. 2012 Nuclear Science Symposium, Medical Imaging Conference, Oct 2012, Anaheim, United States. hal-00838620

**HAL Id: hal-00838620**

**<https://hal.science/hal-00838620v1>**

Submitted on 26 Jun 2013

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## Gate simulation of a complete proton treatment combined with prompt-gamma monitoring.

*Pierre Gueth<sup>1</sup>, Loic Grevillot<sup>1</sup>, Denis Dauvergne<sup>2</sup>, Nicolas Freud<sup>1</sup>, Jean-Michel Létang<sup>1</sup>, Cédric Ray<sup>2</sup>, Étienne Testa<sup>2</sup>, David Sarrut<sup>1,3</sup>*

1 – Université de Lyon, CREATIS ; CNRS UMR5220 ; Inserm U1044 ; INSA-Lyon ; Université Lyon 1, France

2 – IPNL, CNRS UMR5822, IN2P, Université Lyon 1, France

3 – Léon Bérard cancer center, France

### Purpose.

Several systems are studied to monitor the deposited dose using prompt-gamma (PG) emitted during proton nuclear interactions. It was shown that ion and PG ranges are correlated. However, measuring spot-by-spot Bragg Peak positions in a scanned beam setup is still an open issue. We present a Monte-Carlo (MC) simulation of a treatment fraction combined with an innovative PG camera design. We used patient data imported from a clinical TPS (Xio, Elekta).

### Methods.

We developed a spot scanning beam model based exclusively on calibration data required by the TPS. This model includes nozzle and beam line characteristics. We developed a DICOM plan to Gate macro conversion tool and a “dose to water” option to allow dose distribution comparison between TPS and MC. A collimated multi-slit PG detector, under development at IPNL-Lyon, was simulated. It is composed of tungsten collimators, BGO scintillators coupled with photo-multipliers. Each PG event is filtered (TOF and energy) to create a depth PG signal. Simulations were computed using the new Gate v6.2 release.

### Results.

The beam model reproduced measured SOBP ranges within 0.7 mm [1]. The dose distributions obtained with Xio and Gate were in good overall agreement. Stopping power differences up to 3% were observed, due to the HU conversion procedure [2]. For a 2 Gy fraction, spot-by-spot PG monitoring estimates the Bragg peak position with a 0.6 mm standard deviation [3].

### Conclusion.

A complete MC simulation of a clinical proton treatment plan, including the active scanning delivery system, patient data and PG camera is feasible with Gate. It represents an interesting tool to assess and optimize spot-by-spot PG-based dose monitoring for clinical treatment plans. This work is carried out within the framework of the ENVISION FP7 project.

[1] L. Grevillot, D. Bertrand, F. Dessy, N. Freud, and D. Sarrut, "A Monte Carlo pencil beam scanning model for proton treatment plan simulation using GATE/GEANT4", *Physics in Medicine and Biology*, vol. 56, no. 16, pp. 5203, 2011

[2] L. Grevillot, D. Bertrand, F. Fessy, N. Freud, and D. Sarrut, "GATE as a GEANT4-based Monte Carlo platform for the evaluation of proton pencil beam scanning treatment plans", *Physics in Medicine and Biology*, vol. 57, no. 13.

[3] P. Gueth, E. Testa, JM Letang, N. Freud and D. Sarrut, "Perturbation detection using prompt  $\gamma$  in protontherapy", *PTCOG*, 2012