GEOTRACES intercalibration of neodymium isotopes and rare earth element concentrations in seawater and suspended particles. Part 1: reproducibility of results for the international intercomparison

T. van de Flierdt, K. Pahnke, H. Amakawa, P. Andersson, C. Basak, B. Coles, C. Colin, K. Crocket, M. Frank, N. Frank, et al.

HAL Id: hal-00985401
https://hal.science/hal-00985401

Submitted on 30 Apr 2014

GEOTRACES intercalibration of neodymium isotopes and rare earth element concentrations in seawater and suspended particles. Part 1: reproducibility of results for the international intercomparison

Tina van de Flierdt[1], Katharina Pahnke[2,3], and GEOTRACES intercalibration participants: Hiroshi Amakawa[4,5], Per Andersson[6], Chandranath Basak[7], Barry Coles[1], Christophe Colin[8,9], Kirsty Crocket[1], Martin Frank[10], Norbert Frank[9], Steven L. Goldstein[11,12], Vineet Goswami[13], Brian A. Haley[10,14], Ed C. Hathorne[10], Sidney R. Hemming[11,12], Gideon M. Henderson[15], Catherine Jeandel[16], Kevin Jones[11,12,17], Katharina Kreissig[1], Francois Lacan[16], Myriam Lambelet[1], Ellen E. Martin[7], Derrick R. Newkirk[7], Haijme Obata[4], Leopoldo Pena[11], Alexander M. Piotrowski[18], Catherine Pradoux[16], Howie D. Scher[19], Hans Schöberg[6], Sunil Kumar Singh[13],Torben Stichel[2,10], Hirofumi Tazoe[20], Derek Vance[21], and Jingjing Yang[15]*

[1]Department of Earth Science and Engineering, Imperial College London, South Kensington Campus, London, SW7 2AZ, UK

[2]Department of Geology and Geophysics, University of Hawaii, Honolulu, HI 96822, USA

[3]Max Planck Research Group for Marine Isotope Geochemistry, University of Oldenburg, Carl-von-Ossietzky-Str. 9-11, 26129 Oldenburg, Germany

[4]Atmosphere and Ocean Research Institute, University of Tokyo, 5-1-5 Kashiwanoha, Kashiwa, Chiba 277-8564, Japan

[5]Department of Geosciences, National Taiwan University, No. 1, Sec. 4, Roosevelt Road, Taipei, 10617 Taiwan (R.O.C)

[6]Laboratory for Isotope Geology, Swedish Museum of Natural History, Box 50007, 104 05 Stockholm, Sweden

[7]Department of Geological Sciences, University of Florida, Gainesville, FL 32611 USA

[8]Laboratoire des Interactions et Dynamique des Environnements de Surface (IDES), UMR 8148, CNRS-Université de Paris-Sud, Bâtiment 504, 91405 Orsay Cedex, France

[9]Laboratoire des Sciences du Climat et de L'Environnement (LSCE), Laboratoire mixte CNRS-CEA, Avenue de la Terrasse, 91198, Gif-sur-Yvette Cedex, France

[10]IFM-GEOMAR, Leibniz Institute of Marine Sciences at the University of Kiel, Wischhofstraße 1-3, 24148 Kiel, Germany

[11]Lamont-Doherty Earth Observatory of Columbia University, 61 Rt. 9W, Palisades, NY 10964, USA

[12]Department of Earth and Environmental Sciences, Columbia University, 61 Rt. 9W, Palisades, NY 10964, USA

[13]Geosciences Division, Physical Research Laboratory, Navrangpura, Ahmedabad – 380009, India

[14]COAS, OSU, 104 Ocean Admin. Bldg., Corvallis, OR 97209, USA

[15]Department of Earth Sciences, South Parks Road, Oxford, OX1 3AN, UK

[16]LEGOS (CNES/CNRS/IRD/University of Toulouse), 14 Avenue Edouard Belin, 31400, Toulouse, France

[17]ExxonMobil Production Company, 396 West Greens Road, Houston, TX 77067, USA

[18]Godwin Laboratory for Palaeoclimate Research, Department of Earth Sciences, University of Cambridge, Cambridge, CB2 3EQ, UK

[19]Department of Earth and Ocean Sciences and Marine Science Program, University of South Carolina, Columbia, SC 29205, USA

[20]Department of Radiation Chemistry, Institute of Radiation Emergency Medicine, Hirosaki University, 66-1 Hon-cho, Hirosaki, Aomori, Japan

[21]School of Earth Sciences, University of Bristol, Bristol, BS8 1RJ, UK

*Corresponding author: E-mail: tina.vandeflierdt@imperial.ac.uk

Abstract

One of the key activities during the initial phase of the international GEOTRACES program was an extensive international intercalibration effort, to ensure that results for a range of trace elements and isotopes (TEIs) from different cruises and from different laboratories can be compared in a meaningful way. Here we present the results from the intercalibration efforts on neodymium isotopes and rare earth elements in seawater and marine particles. Fifteen different laboratories reported results for dissolved $^{143}Nd/^{144}Nd$ ratios in seawater at three different locations (BATS 15 m, BATS 2000 m, SAFe 3000 m), with an overall agreement within 47 to 57 ppm (2σ standard deviation of the mean). A similar agreement was found for analyses of an unknown pure Nd standard solution carried out by 13 laboratories (56 ppm), indicating that mass spectrometry is the main variable in achieving accurate and precise Nd isotope ratios. Overall, this result is very satisfactory, as the achieved precision is a factor of 40 better than the range of Nd isotopic compositions observed in the global ocean. Intercalibration for dissolved rare earth element concentrations (REEs) by six laboratories for two water depths at BATS yielded a reproducibility of 15% or better for all REE except Ce, which seems to be the most blank-sensitive REE. Neodymium concentrations from 12 laboratories show an agreement within 9%, reflecting the best currently possible reproducibility. Results for Nd isotopic compositions and REE concentrations on marine particles are inconclusive, and should be revisited in the future.

The neodymium isotopic composition of seawater has been of interest to the scientific community for more than three decades. The feasibility to directly measure Nd isotopes in seawater was first demonstrated by Piepgras et al. (1979). It was these measurements, alongside indirect studies of seawater Nd isotopes based on fish debris (DePaolo and Wasserburg 1977) and ferromanganese nodules (O'Nions et al. 1978) from which the picture emerged that the $^{143}Nd/^{144}Nd$ isotope ratio in the different ocean basins is closely coupled to that of the surrounding continents. The provinciality between different ocean basins implied by the early data (see also Piepgras and Wasserburg 1980) has been verified and refined in many studies since, and points to a residence time of Nd in seawater on the order of, or shorter than, the global turnover time of the ocean (500-1000 y; Tachikawa et al. 2003). The lowest values for dissolved Nd isotopes are observed around the old cratons in the North Atlantic (Stordal and Wasserburg 1986), while the highest values are found next to young volcanic areas in the Pacific (Piepgras and Jacobsen 1988; Vance et al. 2004), yielding an overall range in the global ocean of more than 20 epsilon units ($^{143}Nd/^{144}Nd$ ratios are expressed in epsilon units, which denotes the deviation of a sample $^{143}Nd/^{144}Nd$ ratio from the "CHondritic Uniform Reservoir" value in parts per 10000; CHUR = 0.512638; Jacobsen and Wasserburg 1980).

However, uncertainties remain when trying to exploit dissolved Nd isotopes as a tracer for water mass mixing (Goldstein and Hemming 2003), or as a tracer for continental inputs and exchange with the ocean margins (Lacan and Jeandel 2005). Our current understanding of sources, sinks, and especially of the internal cycling of Nd in the ocean is relatively poor, mainly due to a scarce number of observations. As of today, only ~700 measurements on dissolved Nd isotopes are published from the global ocean, more than a quarter of which are surface water samples. Furthermore, the geographic spread of sampling locations is very poor with ~ 50% of all

depth profiles being located in the North Atlantic and the North Pacific Ocean.

The international GEOTRACES program is bound to improve this situation, as its three research objectives are defined as follows (GEOTRACES Science Plan 2006):

(1) To determine global ocean distributions of selected trace elements and isotopes (TEIs), including their concentration, chemical speciation, and physical form, and to evaluate the sources, sinks, and internal cycling of these species to characterize more completely the physical, chemical, and biological processes regulating their distributions;

(2) To understand the processes involved in oceanic trace-element cycles sufficiently well that the response of these cycles to global change can be predicted, and their impact on the carbon cycle and climate understood; and

(3) To understand the processes that control the concentrations of geochemical species used for proxies of the past environment, both in the water column and in the substrates that reflect the water column.

A set of 'key TEIs' (Trace Elements and Isotopes; among them dissolved Nd isotopes) considered central to these broader goals of GEOTRACES were identified and have to be measured on every GEOTRACES cruise (GEOTRACES Science Plan 2006). Before the start of the major field program, a phase of intercalibration activities was launched for a number of TEIs, including all 'key TEIs'.

No international intercalibration has ever been carried out for Nd isotopes – neither for seawater, nor in hard rocks. The hard rock community could overcome this problem by analyses of readily available USGS reference materials (e.g., BCR-1 and 2, BHVO-1 and 2, AGV-1 and 2, G-2, GSP-2; *see* Weis et al. 2006 for a recent example). No reference material however is available that resembles the matrix of seawater and requires analytical procedures to extract small abundances of REE from a large volume of sample. Here we report the results on the GEOTRACES

intercalibration efforts for dissolved Nd isotopes, which include a total of 15 participating laboratories, some of which have a long-standing history of seawater Nd isotope measurements, and some of which are relatively new to the field.

Although not one of the key parameters of the GEOTRACES program, the REE concentrations in seawater have been historically used by a large number of laboratories to decipher lateral and vertical processes in the water column as well as water mass provenance (e.g., Elderfield and Greaves 1982; Piepgras and Jacobsen 1992; Alibo and Nozaki 1999). We therefore used the opportunity offered by the Nd isotope study to gather intercalibration data on dissolved rare earth element concentrations (REEs), as well as Nd isotopic compositions and REE concentrations from marine particles.

Our goal was to assess the precision and accuracy possible for dissolved and particulate Nd isotopes (and REEs). Following documentation and assessment of the data, we will present some recommendations to achieve precise and accurate results on future GEOTRACES cruises. This article is accompanied by a second article that takes a closer look at the seagoing side of achieving accurate and precise results for dissolved and particulate Nd isotopes and REE, by presenting results on systematic tests on different materials and methods involved (Pahnke et al. 2012). The second article also presents the first GEOTRACES baseline profiles for Nd isotopes (and REEs) for the Bermuda Atlantic Time-Series Station (BATS) in the NW Atlantic Ocean and the SAFe station (Sampling and Analysis of Fe) in the Pacific Ocean.

Materials and procedures

Shipboard sampling of seawater and particles

Samples used during this intercalibration exercise were collected on two GEOTRACES intercalibration cruises, carried out on the R/V *Knorr* in June-July 2008 in the Atlantic Ocean (Bermuda-Norfolk, KN193-6), and in May 2009 in the Pacific Ocean (Honolulu-San Diego, KN195-8). The two cruises were chosen to sample fundamentally different water masses and biogeochemical regimes and to occupy previously well characterized time-series stations in the Atlantic Ocean (BATS) and the Pacific Ocean (SAFe).

For intercalibration of Nd isotopes and REE concentrations, both dissolved and particulate samples were collected. For dissolved samples, large volumes of filtered and homogenized seawater were collected from three different locations: BATS deep water (2000 m), BATS sub-surface water (15 m), and SAFe deep water (3000 m). Deep water samples were collected using multiple casts of the trace metal clean GEOTRACES rosette deployed on a Kevlar hydroline, whereas sub-surface water was collected using the UCSC designed "GeoFish" towed sampling system. Filtration of seawater was carried out using 0.2 μm Osmonics cartridge filters in a portable clean van environment. Water from the same water depth was transferred through Teflon tubing into two interconnected 500 L tanks made of fluorinated low density polyethylene, which were

homogenized using an all PFA Teflon diaphragm pump (UCSC SAFe tanks). Homogenized 500-1000 L of seawater were acidified to pH < 2 using ultrapure Seastar HCl. Individual sample containers ranged from 0.5 L to 20 L, and were provided and precleaned by the individual intercalibration participants. Overall, most laboratories received 2 × 10 L seawater from each of the three locations for duplicate analyses of Nd isotopes, and 0.5 L of seawater for REE concentration measurements. Filling of individual sample containers was performed directly from the SAFe tanks under a tent on deck the R/V *Knorr*.

Particle samples were derived from the three following locations: BATS deep water (2000 m), BATS sub-surface water (30 m), and Virgina slope water (98 m). Two complementary pumping systems were deployed for Nd isotope and REE intercalibration purposes: McLane pumps (deep water location at BATS), and MULVFS (Multiple Unit Large Volume Filtration System; other two location) (Maiti et al. 2012; Bishop et al. 2008). Using the McLane pumps, about 600 L seawater were pumped through each 0.45 μm Supor filter (142 mm diameter). Filters were dried onboard in a dedicated clean environment, and cut into halves using a PVC template fitted with a ceramic knife (pie-cutter). Each laboratory participating in the Nd isotope intercalibration was provided half of a filter and half of a corresponding dipped blank. For REE intercalibration, 2 cm punches, corresponding to ~20 L pumped seawater, were taken out of a few of the filter halves for REE measurements (e.g., some of the isotope laboratories received slightly smaller samples). Larger water volumes (>1000 L) were pumped with the MULVFS system, equipped with large quartz fiber filters (QMA; 506.7 cm^2 area, 1 μm pore size). Four 4.5 cm punches were taken out of the filters for Nd isotope samples, representing ~380 L pumped water volume, accompanied by two 2.5 cm punches from dipped blanks. For REE measurements, four 1.2 cm punches (~27 L pumped water volume) were collected from dried filters together with two 1.2 cm punches from dipped blanks. All filters were stored and shipped in precleaned containers.

Shore-based methods to determine dissolved and particulate Nd isotopic compositions and REE concentrations

As this article documents the first intercalibration effort for Nd isotopes (and REE concentrations) in seawater and marine particles, no particular analytical method was prescribed. This approach allowed a variety of routine procedures to be compared (e.g., different sample preconcentration, ion chromatography, and mass spectrometry procedures). Below we briefly summarize the main differences in the analytical methods applied. Methodological information for each anonymized laboratory can be found in Tables 1a and 1b.

Neodymium isotope measurements in seawater samples

To preconcentrate Nd from the seawater matrix most laboratories performed a Fe coprecipitation step, where Fe is added to seawater, most typically as FeCl$_3$ and equilibrated for 24-48 h. A subsequent increase to a pH of ~8, using ultraclean

Table 1a: Details for processing seawater samples for their Nd isotopic composition in different laboratories.

lab	spike	preconcentration of Nd	ion chromatography		mass spectrometry	blank
1	NA	Fe co-precipitation (100 mg Fe for 20L of seawater)	cation exchange	DEP column	TIMS (Nd$^+$)/ MC-ICP-MS	<20 pg
2	mixed ^{150}Nd-^{149}Sm	Fe co-precipitation (8 mg Fe per litre of seawater)	cation exchange	Ln Spec	MC-ICP-MS	30 pg (275 pg*)
3	NA	Fe co-precipitation (50 mg Fe for 10L of seawater)	RE Spec	α-HIBA	TIMS (NdO$^+$)	3 pg
4[a]	NA	Fe co-precipitation (50 mg Fe for 10L of seawater)	cation exchange	Ln Spec	MC-ICP-MS	<2 pg
4[b]	NA	preconcentrated 10L on C18 cartridge	NA	Ln Spec	MC-ICP-MS	ND
5	spiked samples 511 and 311 with 250μg ^{9}Be	Fe co-precipitation	anion exchange, cation exchange, TRU spec column, Ln Spec column, Eichrom prefilter material		TIMS (Nd$^+$)	70 pg
6	^{150}Nd, ^{147}Sm	Fe co-precipitation (14 mg Fe for 3.6L seawater)	anion exchange, TRU Spec, Ln Spec		TIMS (Nd$^+$)	8-12 pg[a] 4 pg[b]
7[a]	NA	Fe co-precipitation (60 mg Fe for 10L seawater)	cation exchange, Mitsubishi resion, Ln Spec		MC-ICP-MS	2.5 pg
7[b]	NA	preconcentrated 10L on C18 cartridge	NA	Ln Spec	MC-ICP-MS	2.5 pg
8	NA	preconcentrated 10L on C18 cartridge	cation exchange	Ln Spec	TIMS (Nd$^+$)/ MC-ICP-MS	187 pg
9	NA	Fe co-precipitation	anion exchange, cation exchange (x2), Ln Spec column		TIMS (Nd$^+$)	120 pg
10	NA	Fe co-precipitation (1 mg Fe per litre of seawater)	TRU Spec	Ln Spec	MC-ICP-MS	ND
11	NA	Fe co-precipitation (50 mg Fe for 10L of seawater)	RE Spec	α-HIBA	TIMS (NdO$^+$)	3 pg
12	NA	preconcentrated 10L on C18 cartridge	TRU Spec	Ln Spec	TIMS (NdO$^+$)	7 pg[c] 20pg[d]
13	NA	preconcentrated 10L on C18 cartridge	cation exchange	Ln Spec	MC-ICP-MS	8 pg
14	NA	Fe co-precipitation (0.5 g Fe per 20L of seawater)	cation exchange	Ln Spec	MC-ICP-MS	ND
15	NA	preconcentrated 10L on C18 cartridge	cation exchange	Ln Spec	TIMS (Nd$^+$)	< 300 pg

* blank of Fe solution, which was measured for its isotopic composition and would imply a correction of 0.02 to 0.1 epsilon units, which the lab refrained from applying; remaining analytical blank (chemistry and mass spectrometry): 30pg
[a] method used for intercalibration samples from BATS
[b] method used for intercalibration samples from SaFE
[c] column chemistry and loading blank
[d] total procedural blank derived from pumping 10L ultraclean water through two cartridges and subtracting the water blank
ND = not determined; NA = not applicable

ammonium hydroxide leads to the formation of iron hydroxides, which in turn, efficiently scavenge rare earth elements out of the seawater solution (e.g., Piepgras and Wasserburg 1987). Depending on the laboratory, between 1 and 25 mg purified Fe are added per liter of seawater. Purification of Fe is typically carried out by isopropyl ether back-extraction (Dodson et al. 1936), or by ion exchange chromatography. Lack of careful purification has been reported as the most likely can-

didate to introduce a significant procedural blank (e.g., a few hundred picograms; see Table 1a). A few laboratories choose to concentrate Nd from the seawater matrix by liquid-liquid extraction, pumping 10 L seawater aliquots, adjusted to a pH of ~3.5, through two coupled Sep-Pack C18 cartridges, each filled with 300 mg of a mixture of 65% bis(2-ethylhexyl) hydrogen phosphate (HDEHP) and 35% 2-ethylhexyl dihydrogen phosphate (H$_2$MEHP), at a speed of 20 mL/min (Sha-

Table 1b: Details for processing seawater samples for Nd concentrations and other rare earth element concentrations in different laboratories.

lab	concentrations	aliquoting	spike	technique	mass spectrometry
1	[REE]	4mL aliquot	no	seaFAST system with a column containing resin with ethylenediaminetriacetic acid and iminodiacetic acid functional groups to preconcentrate REE	ICP-MS
2	[Nd], [Sm]	spiked large volume sample for isotope work with [150]Nd and [149]Sm and processed as described in Table 1a			MC-ICP-MS
3	[REE]	50 to 150mL aliquot	mixed REE spike ([138]La, [142]Ce, [145]Nd, [149]Sm, [153]Eu, [155]Gd, [161]Gd, [167]Er, and [171]Yb)	co-precipitated REE with Fe; RE spec ion chromatography with added ascorbic acid to reduce Fe	MC-ICP-MS
5	[REE]	500 ml aliquot	[145]Nd and Indium	solvent extraction following Shabani et al (1992) except for sample 513, for which REE were co-precipitated with 8mg of or Fe followed by anion exchange chemistry	ICP-MS
6	[Nd], [Sm]	spiked entire sample with [150]Nd and [147]Sm and processed as described in Table 1a			TIMS
8	[REE]	500mL aliquot	[150]Nd, [172]Yb	co-precipitated REE with 2.5mg Fe; anion exchange column	ICP-MS
9	[Nd]	500mL aliquot	[145]Nd	co-precipitated REE with Fe; anion exchange column	ICP-MS
10	[REE]	600 mL aliquot	~20 µg Tm solution (100.2 ppm)	co-precipitated REE with 1ml 1000ppm Fe solution	ICP-MS
11	[Nd]	250 mL aliquot	[146]Nd	co-precipitated REE with Fe; RE Spec column	ICP-MS
12	[Nd]	spiked large volume for isotope work with [150]Nd and processed as described in Table 1a			TIMS
16	[REE]	180-190g of seawater	[145]Nd, [171]Yb	co-precipitated REE with Fe; anion exchange column	ICP-MS
17[*,a]	[REE]	10L of seawater; used a fraction of the REE cut from column chemistry	no	co-precipitated REE with Fe; cation exchange column, Mitsubishi resin column; Re and Rh as internal standards	ICP-MS
17[b]	[Nd]	500mL aliquot	[150]Nd	preconcentration of REE on C-18 cartridges; Ln Spec column	MC-ICP-MS

[*] seawater concentrations were calculated assuming a total sample volume of 10L for the two intercalibration samples from BATS
[a] method used for intercalibration samples from BATS
[b] method used for analyses of unknown REE standard

bani et al. 1992; see Table 1a). Barium was subsequently eluted from the cartridges using 5 mL 0.01M HCl and the REE were collected in 35 mL 6M HCl.

Following preconcentration of Nd, chemical separation of Nd from the sample matrix is required to minimize possible interferences during subsequent analyses by mass spectrometry. Dependent on the column set up, high Fe concentrations in initial sample solutions can be a problem. Hence, a few laboratories reduced the amount of Fe in an initial step by standard anion exchange chromatography (e.g., Hooker et al. 1975), isopropyl ether back extraction (Dodson et al. 1936), or Fe reduction through the use of ascorbic acid. Most laboratories however applied a two-step ion chromatography, with a first column to separate REE from the sample matrix, and a second column to separate REEs from each other (see Table 1a). For the first column, the standard cation exchange procedure using elution with HCl of increasing molarities (e.g., White and Patchett 1984) is the most frequent choice (note: some laboratories use an initial step of eluting some of the sample matrix with HNO₃ before switching over to HCl). The more recently developed Eichrom ion exchange resins, TRU Spec and RE Spec, and elution using HNO₃ and HCl in succession (e.g., Pin and Zalduegui 1997) were used by a number of laboratories. The second column is either based on cation exchange chromatography with α-hydroxyisobutyric acid (α-HIBA) to sequentially elute the various REEs from heavy to light (e.g., Eugster et al. 1970; Lugmair et al. 1975), or a reversed-phase chromatography using beads coated with HDEHP [di(2-ethylhexyl)orthophosphoric acid] (Richard et al. 1976) to sequentially elute the REE from light to heavy. Over the past years the latter has been increasingly replaced by Eichrom's Ln Spec resin (Pin and Zalduegui 1997).

Neodymium isotopic ratios were measured by Thermal Ionization Mass Spectrometry (TIMS: Triton, Finnigan MAT 261 and 262, VG Sector, VG Sector

54) or Multiple Collector Inductively Coupled Plasma Mass Spectrometry (MC-ICP-MS: Nu Plasma or Neptune). While historically TIMS was the method of choice for all laboratories, the past 15 y have seen a quick rise of MC-ICP-MS as a replacement for many routine analyses previously performed by TIMS. Arguments in favor of MC-ICP-MS measurements are the speediness of data collection achieved by significantly reduced analysis time, while maintaining a similar precision to that achieved by TIMS (for a summary on MC-ICP-MS, *see* Halliday et al. 2000). The advantage of TIMS measurements, especially when considering improved ionization during neodymium oxide measurements (NdO^+) compared with metal runs (Nd^+), is the ability to return high precision data on small abundance samples (e.g., Li et al. 2007; Chu et al. 2009; Harvey and Baxter 2009). Extensive literature exists on appropriate measurement setups, especially on eliminating the relatively large mass discrimination effect observed for measurements by MC-ICP-MS (e.g., Vance and Thirlwall 2002; Wombacher and Rehkämper 2003) and on carrying out the numerous oxide interference corrections and using the appropriate oxygen isotopic composition during TIMS measurements (e.g., Thirlwall 1991a, 1991b). Discussing the details of applied mass spectrometry between the 15 participating laboratories goes beyond the scope of this article and would violate agreed anonymity. It is, however, important to note that all laboratories corrected for instrumental mass bias using a $^{146}Nd/^{144}Nd$ of 0.7219 (O'Nions et al. 1977), making the results directly comparable.

Neodymium isotope measurements in marine particles

The main difference in processing marine particles for their Nd isotopic composition, compared with seawater samples, is the initial sample dissolution step. As different digestion and leaching methods of filters are likely to result in varying amounts of REE being removed from the particles on the filters, two methods were prescribed for the particle samples from deep waters at BATS and shallow water at BATS, respectively. For the Supor filters (deep water particles at BATS), a total digest of the filter was targeted by following the method used by Cullen and Sherrell (1999). Briefly, filters were placed in precleaned Teflon vials together with a mixture of concentrated HNO_3 and concentrated HF (19:1). Complete digestion of any particles was achieved by refluxing at 120°C for 4 h. This procedure also leads to a total digestion of the Supor filter, which however does not constitute a major blank issue (*see* "Assessment and discussion of results" below). One laboratory deviated from this procedure and instead followed a digestion method adapted from Landing and Lewis (1991) that avoids total digestion of the filter material (Table 5). The large blank of quartz fiber filters (QMA) in contrast precludes total digestion, and the approach chosen targets a leaching of the REE fraction not bound to the silicate phases (Collier and Edmond 1984). As described in Jeandel et al. (1995), samples were leached in 0.6M HCl at 60°C for 20 h. For the third set of filters from the Virginia slope station, individual laboratories

were free to choose a different method. The only deviation from above outlined methods for samples from the Virginia slope station was performed however by laboratory number 2, where QMA filters were leached for 3.5 h at 90°C in a mixture of 0.005M hydroxylamine hydrochloride, 1.5% acetic acid, and 2 nM EDTA solution (pH of 3.5), with an ultrasonification step performed every hour.

Rare earth element concentration measurements in seawater samples

In general, the procedure for determining REE concentrations on seawater samples can follow a similar methodology to the one described above for the first steps in determining the Nd isotopic composition. Most laboratories use Fe coprecipitation or liquid-liquid extraction to preconcentrate the REE, followed by ion exchange chromatography to separate the REE fraction from the sample matrix (typically anion exchange or RE Spec; Table 1b).

The major difference between isotope and concentrations measurements is the smaller water volume required for concentration analyses (between 4 mL and 600 mL; Table 1b), and the addition of an isotopically enriched tracer ("spike") before preconcentration. This addition is crucial to correct for elemental fractionation, which may happen during preconcentration and/or ion chromatography. The types of spikes used for this intercalibration exercise include multi-element mixed REE isotope spikes, two-element isotope spikes, and the addition of monoisotopic elements (Table 1b). Traditionally, REE concentrations were analyzed by isotope dilution methods and TIMS (e.g., Elderfield and Greaves 1982), but for this intercalibration, REE patterns were mostly analyzed by ICP-MS (Agilent 7500, Element 2, HP-4500, Perkin Elmer Elan DRCII, Yokogawa PMS-2000). Five laboratories report results for Nd concentrations only, which in three cases, were determined on the same large volume sample used for isotopic analyses through spiking before coprecipitation. In these cases, measurements were performed by TIMS/MC-ICP-MS.

Notably, one laboratory (nr 1) used the seaFAST system, a commercially available system (Elemental Scientific) with a column containing a resin with ethylendiaminetriacetic acid and iminodiacetic acid functional groups to preconcentrate REEs. This system can be directly connected to an ICP-MS, and blanks, standards, and samples are passed through the column in the same manner, and eluted directly into the spray chamber of the ICP-MS.

Rare earth element concentration measurements in marine particles

Rare earth element concentrations of marine particles were attempted by four laboratories. Partial return of data, as well as some deviation from recommended methodologies, yielded a poor statistical basis for assessing the results. We, therefore, refrain from reporting the actual values in the article, but will make some qualitative comments on the results in the next sections to encourage and guide future efforts.

Assessment and discussion of results

Neodymium isotope intercalibration in seawater at BATS and SAFe

To achieve comparability of results, all measured $^{143}Nd/^{144}Nd$ ratios have been normalized relative to a JNd_i ratio of 0.512115 (Tanaka et al. 2000) or a La Jolla ratio of 0.511858 (Lugmair et al. 1983) using standard values reported by each laboratory (see caption of Table 2). For laboratories that routinely use other in-house standards than the ones listed above, their reported cross-calibration for their respective standards relative to JNd_i/La Jolla was used. Figs. 1 and 2 and Tables 2 and 3 show the results for the Nd isotopic composition of seawater at three distinct water depths at BATS and SAFe as measured by 15 different laboratories. Sub-surface water at BATS yielded an average Nd isotopic composition of –9.2 ± 0.6, deep water at BATS a value of –13.1 ± 0.6, and the average Nd isotopic composition of deep water at SAFe is –3.2 ± 0.5 (errors:

two sigma standard deviations of the mean). Interpretation of the results concerning comparability to published literature values can be found in the accompanying paper by Pahnke et al. (2012). Here we focus on assessing the agreement achieved between the different laboratories (i.e., reproducibility; reproducibility is here defined as the two sigma standard deviation of the mean of all individual data points considered for a given location).

The reproducibility for the three sample sets is found to be between 47 and 57 ppm. Considering that the external two sigma standard deviation reported by each laboratory for $^{143}Nd/^{144}Nd$ ratios varies between 10 and 100 ppm (see values reported on La Jolla, JNd_i, or other in house standard runs in the caption of Tables 2 and 3), this is an excellent result. However, two thirds of all laboratories report a more narrow range of external reproducibilities between 20 and 40 ppm for their $^{143}Nd/^{144}Nd$ analyses. Two questions emerge at this point: (1) Is a better agreement of natural seawater measurements possible (e.g., what is the reproducibility on natural samples compared to pure standard solutions and how do various methodologies feature in the statistics?), and (2) what is the reason for the larger external errors reported by some laboratories, and how do their values influence the overall statistics.

Starting with the first point, a simple exercise was undertaken to constrain whether the spread in data observed in Figs. 1 and 2 is an artifact of incomplete matrix removal from natural samples during ion chromatography, and subsequent mass spectrometry, or whether the spread in $^{143}Nd/^{144}Nd$ ratios arises from the analyses themselves (i.e., different methods applied during mass spectrometry yielding different degrees of accuracy and precision). Fig. 3 illustrates results obtained for an 'unknown standard solution'. The standard was produced from Nd_2O_3 powder at Imperial College London in an ultraclean lab-

Fig. 1. Dissolved neodymium isotopic composition for two water depths at BATS (NW Atlantic). Results display good agreement of the international community (15 individual laboratories) on measuring the Nd isotopic composition of seawater (i.e., values agree within 57 ppm; two sigma standard deviation of the mean). Details on samples and laboratories are given in Tables 1a and 2. Errors plotted are external two sigma standard deviations reported by each laboratory (based on repeat analyses of isotopic standards; see Table 2 caption), or internal two sigma standard error, depending on which error was the larger one.

Fig. 2. Dissolved neodymium isotopic composition at 3 km water depth at SAFe (NE Pacific). Results display good agreement of the international community (8 individual laboratories) on measuring the Nd isotopic composition of seawater (i.e., all values agree within 47 ppm; two sigma standard deviation of the mean). Details on samples and laboratories are given in Tables 1a and 3. Errors plotted are external two sigma standard deviations reported by each laboratory (based on repeat analyses of isotopic standards; see Table 3 caption).

Table 2: International intercalibration for Nd isotopes in seawater at BATS.

lab	sample ID	sample volume	$^{143}Nd/^{144}Nd$ measured	internal 2σ SE	$^{143}Nd/^{144}Nd$ normalized*	ε_{Nd}[†]	internal 2σ SE	external 2σ SD[‡]	deviation from average
BATS, 2000m									
1	KN193-6-Nd-301	20L	0.511914 ± 0.000011		0.511979	-12.9 ± 0.2		0.4	0.3
2	KN193-6-Nd-305	5L	0.511957 ± 0.000008		0.511966	-13.1 ± 0.2		0.1	0.0
2	KN193-6-Nd-305	5L	0.511968 ± 0.000009		0.511977	-12.9 ± 0.2		0.1	0.2
3	KN193-6-Nd-324	10L	0.511943 ± 0.000015		0.511956	-13.3 ± 0.3		0.4	-0.2
3	KN193-6-Nd-323	10L	0.511941 ± 0.000008		0.511954	-13.3 ± 0.3		0.4	-0.2
4	KN193-6-Nd-314	10L	0.511942 ± 0.000018		0.511949	-13.4 ± 0.3		0.2	-0.3
4	KN193-6-Nd-314	10L	0.511978 ± 0.000020		0.511985	-12.7 ± 0.4		0.2	0.4
5	KN193-6-Nd-310	10L	0.511927 ± 0.000008		0.511959	-13.2 ± 0.2		0.3	-0.1
5	KN193-6-Nd-311	10L	0.511918 ± 0.000008		0.511950	-13.4 ± 0.2		0.3	-0.3
6	KN193-6-Nd-315	10L	0.511974 ± 0.000019		0.511984	-12.8 ± 0.4		0.4	0.4
6	KN193-6-Nd-318	10L	0.511927 ± 0.000022		0.511937	-13.7 ± 0.4		0.4	-0.5
7	KN193-6-Nd-309	10L	0.511936 ± 0.000020		0.511948	-13.5 ± 0.4		0.3	-0.3
7	KN193-6-Nd-308	10L	0.511957 ± 0.000010		0.511969	-13.1 ± 0.2		0.3	0.1
8[§]	KN193-6-Nd-304A	10L	0.512004 ± 0.000018		0.511961	-13.2 ± 0.4		0.2	-0.1
8[§]	KN193-6-Nd-304B	10L	0.511997 ± 0.000027		0.511954	-13.3 ± 0.5		0.2	-0.2
9	KN193-6-Nd312	10L	0.511950 ± 0.000017		0.511967	-13.1 ± 0.3		0.3	0.1
9	KN193-6-Nd313	10L	0.511940 ± 0.000015		0.511957	-13.3 ± 0.3		0.3	-0.1
10	KN193-6-Nd-302	10L	0.512040 ± 0.000014		0.511978	-12.9 ± 0.3		0.8	0.3
10	KN193-6-Nd-303	10L	0.512087 ± 0.000022		0.512003	-12.4 ± 0.4		1.0	0.8
11	KNR193-6-706	10L	0.511937 ± 0.000012		0.511958	-13.3 ± 0.2		0.3	-0.1
11	KNR193-6-711	10L	0.511941 ± 0.000007		0.511962	-13.2 ± 0.1		0.3	-0.1
12	KN193-6-Hf-315a	10L	0.511955 ± 0.000008		0.511968	-13.1 ± 0.2		0.2	0.1
12	KN193-6-Hf-315b	10L	0.511945 ± 0.000009		0.511958	-13.3 ± 0.2		0.2	-0.1
13	KN193-6-Nd-325	10L	0.511950 ± 0.000022		0.511963	-13.2 ± 0.4		0.3	0.0
average BATS 2000m						**-13.1**			
2σ SD						**0.6**			
BATS, 15m									
1	KN193-6-Nd-501	20L	0.512114 ± 0.000011		0.512179	-9.0 ± 0.2		0.4	0.2
2	KN193-6-Nd-505	5L	0.512148 ± 0.000010		0.512157	-9.4 ± 0.2		0.1	-0.2
2	KN193-6-Nd-505	5L	0.512161 ± 0.000008		0.512170	-9.1 ± 0.2		0.1	0.1
3	KN193-6-Nd-526	10L	0.512149 ± 0.000008		0.512162	-9.3 ± 0.2		0.4	-0.1
3	KN193-6-Nd-527	10L	0.512132 ± 0.000011		0.512145	-9.6 ± 0.2		0.4	-0.4
4	KN193-6-Nd-517	10L	0.512159 ± 0.000018		0.512166	-9.2 ± 0.4		0.2	0.0
4	KN193-6-Nd-517	10L	0.512169 ± 0.000018		0.512176	-9.0 ± 0.4		0.2	0.2
5	KN193-6-Nd-510	10L	0.512117 ± 0.000009		0.512149	-9.5 ± 0.2		0.3	-0.3
5*	KN193-6-Nd-511	10L	0.512060 ± 0.000032		0.512092	-10.7 ± 0.6		0.3	-1.5
5	KN193-6-Nd-512 & 513	20L	0.512127 ± 0.000011		0.512159	-9.3 ± 0.2		0.3	-0.1
6	KN193-6-Nd-516	10L	0.512202 ± 0.000030		0.512212	-8.3 ± 0.6		0.4	0.9
6	KN193-6-Nd-518	10L	0.512166 ± 0.000057		0.512176	-9.0 ± 1.1		0.4	0.2
7	KN193-6-Nd-509	10L	0.512178 ± 0.000010		0.512190	-8.7 ± 0.2		0.3	0.5
7	KN193-6-Nd-508	10L	0.512149 ± 0.000015		0.512161	-9.3 ± 0.3		0.3	-0.1
8[§1]	KN193-6-Nd-504	20L	0.512194 ± 0.000021		0.512151	-9.5 ± 0.4		0.2	-0.3
9	KN193-6-Nd-514 & 515	20L	0.512152 ± 0.000014		0.512170	-9.1 ± 0.3		0.3	0.1
10	KN193-6-Nd-502	10L	0.512246 ± 0.000019		0.512162	-9.3 ± 0.4		1.0	-0.1
10	KN193-6-Nd-503	10L	0.512231 ± 0.000011		0.512169	-9.2 ± 0.2		0.8	0.0
12	KN193-6-Hf-539a	10L	0.512155 ± 0.000006		0.512168	-9.2 ± 0.1		0.2	0.0
12	KN193-6-Hf-539b	10L	0.512157 ± 0.000008		0.512170	-9.1 ± 0.2		0.2	0.1
13	KN193-6-Nd-529	10L	0.512145 ± 0.000022		0.512158	-9.4 ± 0.4		0.3	-0.2
15	KN193-6-Nd-521	10L	0.512152 ± 0.000012		0.512153	-9.5 ± 0.2		0.3	-0.3
15	KN193-6-Nd-520	10L	0.512162 ± 0.000011		0.512163	-9.3 ± 0.2		0.3	-0.1
average BATS 15m						**-9.2**			
2σ SD						**0.6**			

* $^{143}Nd/^{144}Nd$ ratios were normalized using the reported standard data for each lab, relative to a JNdi value of 0.512115 (Tanaka et al., 2000) or a La Jolla value of 0.511858 (Lugmair et al., 1983).

[†] ε_{Nd} values were calculated relative to a CHUR of 0.512638 (Jacobsen and Wasserburg, 1980).

[‡] external errors are derived from repeat standard analyses during the measurement session; if internal errors are larger than external errors, these are plotted in Figure 1.

[n] sample was flagged as contaminated by the lab and is not included in the calculated average and not shown in Figure 1.

[§] lab reported that error bars were larger than usual, and speculated that their new Fe solution was not sufficiently clean.

[§1] lab reported sample loss during preconcentration.

lab 1: JNd, $^{143}Nd/^{144}Nd$ ratios of 0.512050 ± 0.000018 (n=184; 15 ppb loads).
lab 2: La Jolla $^{143}Nd/^{144}Nd$ of 0.511849 ± 0.000007 (n=11; 50 ppb solution, uptake of 0.3 ml per analysis).
lab 3: La Jolla $^{143}Nd/^{144}Nd$ of 0.511845 ± 0.000021 (n=10; 15 ng loads).
lab 4: La Jolla $^{143}Nd/^{144}Nd$ of 0.511851 ± 0.000012 (20 ppb solution).
lab 5: La Jolla $^{143}Nd/^{144}Nd$ of 0.511815 ± 0.000010 (n=1; 50ng load run at low beam intensity to match sample beam); JNd, $^{143}Nd/^{144}Nd$ of 0.512083 ± 0.000013 (n=8; 20 ng loads).
lab 6: La Jolla $^{143}Nd/^{144}Nd$ of 0.511848 ± 0.000004 (n=12, loads of 100-400ng). 20 loads of 4-12 ng of an inhouse standard yielded an error of 0.000022.
lab 7: JNd, $^{143}Nd/^{144}Nd$ of 0.512103 ± 0.000014 (n=65; 60 ppb solution).
lab 8: La Jolla $^{143}Nd/^{144}Nd$ of 0.511902 ± 0.000010 (n=8; 15-30 ng loads).
lab 9: JNd, $^{143}Nd/^{144}Nd$ of 0.512098 ± 0.000017 (n=6; 250 ng loads, often runs one filament repeatedly).
lab 10: Second deep water sample and first shallow water sample have been normalized using a JNd, $^{143}Nd/^{144}Nd$ of 0.512199 ± 0.000051 (n=3; 9 ppb solution); remaining ratios are normalized based on JNd, $^{143}Nd/^{144}Nd$ of 0.512177 ± 0.000038 (n=4; 13 ppb solution).
lab 11: La Jolla $^{143}Nd/^{144}Nd$ of 0.511838 ± 0.000015 (n=4; 10 ng loads).
lab 12: JNd, $^{143}Nd/^{144}Nd$ of 0.512102 ± 0.000008 (n=12; 5ng loads).
lab 13: JMC $^{143}Nd/^{144}Nd$ of 0.511110 ± 0.000018 (n=3; 15 ppb solution), calibrated to correspond to a JNd, value of 0.512102.
lab 15: La Jolla $^{143}Nd/^{144}Nd$ of 0.511857 ± 0.000011 (n=4; 10-20 ng loads).

oratory environment using distilled acids only. Two different stock solutions (15ppb Nd in 0.1M HNO_3 and 15 ppm Nd in 4M HNO_3) were prepared from a single digest of the powder, and aliquots were subsequently sent out to each laboratory (15 ppb and 15 ppm solutions dependent on the use of MC-ICP-MS or TIMS respectively). Concentrations were chosen to mimic those typical for Atlantic seawater. Data returns from 13 laboratories yielded an average Nd isotopic composition of –17.3 ± 0.6 (Table 4). The two sigma standard deviation of 56 ppm from the mean of all individual measurements is very similar to the reproducibility obtained on seawater samples (47 to 57 ppm). The two sigma standard deviation calculated for each laboratory on the unknown standard solution varies between 11 and 86 ppm (Table 4), which is similar to the range reported by individual laboratories on La Jolla, JNd$_i$, and other in house standard runs (Tables 2,3; 6-100 ppm). Detailed comparison shows that some laboratories may underestimate their external errors, whereas others report rather conservative errors. Hence, it can be concluded that the major variable in obtaining good agreement for Nd isotopic measurements between different laboratories is mass spectrometry. Although very different preconcentration methods and ion chromatographic protocols are used by the different laboratories (see Table 1a), these parts of the methodology do not seem to add significantly to the uncertainty of the final results. What has a larger effect, however, on the precision possible during mass spectrometry is the concentration of Nd analyzed (i.e., counting statistics). Analyses of 15 ng loads by TIMS typically produce more precise results than 15 ppb runs by MC-ICP-MS, an outcome that is expected due to the higher transmission efficiency of TIMS analyses, especially when performed as NdO$^+$ (e.g., Li et al. 2007; Chu et al. 2009; Harvey and Baxter 2009). However, Nd$^+$ measurements are not necessarily superior to analyses performed on

Table 3: International intercalibration for Nd isotopes in seawater at SAFe.

lab	sample ID	sample volume	^{143}Nd/^{144}Nd measured	internal 2σ SE	^{143}Nd/^{144}Nd normalized*	ε_{Nd}[†]	internal 2σ SE	external 2σ SD[‡]	deviation from average
SAFe, 3000m									
2	KNR195-8-Nd-2143	2.5L	0.512461 ± 0.000006		0.512473	-3.2 ± 0.1		0.1	-0.1
2	KNR195-8-Nd-2144	2.5L	0.512461 ± 0.000006		0.512473	-3.2 ± 0.1		0.1	0.0
2	KNR195-8-Nd-2162	2.5L	0.512469 ± 0.000006		0.512481	-3.1 ± 0.1		0.1	0.1
4	KNR195-8-Nd-2157	5L	0.512473 ± 0.000008		0.512468	-3.3 ± 0.2		0.2	-0.2
4	KNR195-8-Nd-2157	5L	0.512456 ± 0.000010		0.512465	-3.4 ± 0.2		0.2	-0.2
4	KNR195-8-Nd-2157	5L	0.512460 ± 0.000010		0.512469	-3.3 ± 0.2		0.2	-0.1
4	KNR195-8-Nd-2158	5L	0.512429 ± 0.000011		0.512438	-3.9 ± 0.2		0.2	-0.7
6	KNR195-8-Nd-2159	10L	0.512471 ± 0.000008		0.512481	-3.1 ± 0.2		0.4	0.1
6	KNR195-8-Nd-2160	10L	0.512476 ± 0.000006		0.512486	-3.0 ± 0.1		0.4	0.2
7	KNR195-8-Nd-2156	10L	0.512485 ± 0.000009		0.512497	-2.8 ± 0.2		0.3	0.4
8	KNR195-8-Nd-2145	10L	0.512465 ± 0.000007		0.512466	-3.4 ± 0.1		0.3	-0.2
8	KNR195-8-Nd-2146	10L	0.512475 ± 0.000005		0.512476	-3.2 ± 0.1		0.3	0.0
11	KNR195-8-Nd-2202	10L	0.512457 ± 0.000005		0.512478	-3.1 ± 0.1		0.3	0.0
11	KNR195-8-Nd-2305	10L	0.512456 ± 0.000004		0.512475	-3.2 ± 0.1		0.3	0.0
12	KNR195-8-Nd-2306	5L	0.512467 ± 0.000008		0.512480	-3.1 ± 0.2		0.1	0.1
14	KNR195-8-Nd-2147_run1	5L	0.512533 ± 0.000008		0.512484	-3.0 ± 0.2		0.2	0.2
14	KNR195-8-Nd-2147_run2	5L	0.512534 ± 0.000008		0.512484	-3.0 ± 0.2		0.2	0.2
14	KNR195-8-Nd-2148_run1	5L	0.512533 ± 0.000006		0.512483	-3.0 ± 0.1		0.2	0.1
14	KNR195-8-Nd-2148_run2	5L	0.512529 ± 0.000008		0.512480	-3.1 ± 0.2		0.2	0.1
average SAFe 3000m						**-3.2**			
2σ SD						**0.5**			

* ^{143}Nd/^{144}Nd ratios were normalized using the reported standard data for each lab, relative to a JNd$_i$ value of 0.512115
  Tanaka et al. (2000) or a La Jolla value of 0.511858 (Lugmair et al., 1983).

[†] ε_{Nd} values were calculated relative to a CHUR of 0.512638 (Jacobsen and Wasserburg, 1980).

[‡] external errors are derived from repeat standard analyses during the measurement session

lab 2:   La Jolla ^{143}Nd/^{144}Nd of 0.511846 ± 0.000005 (n=16).
lab 4:   Samples were normalised to two measurement sessions yielding JNd$_i$ ^{143}Nd/^{144}Nd of 0.512120 ± 0.000008 (n=7) and
         JNd$_i$ ^{143}Nd/^{144}Nd of 0.512106 ± 0.000005 (n=19) respectively. For chemical separation: see Table 2.
lab 6:   La Jolla ^{143}Nd/^{144}Nd of 0.511848 ± 0.000004 (12 loads of 100-400ng; 19 loads of 10-20ng of an inhouse standard
         yielded an error of 0.000018).
lab 7:   JNdi ^{143}Nd/^{144}Nd of 0.512103 ± 0.000014 (n=45). For chemical separation: see Table 2.
lab 8:   La Jolla ^{143}Nd/^{144}Nd of 0.511857 ± 0.000015 (n=17). Samples were processed in lab 8, but analyzed in lab 2.
lab 11: La Jolla ^{143}Nd/^{144}Nd of 0.511837 ± 0.000013 (n=5).
lab 12: JNdi ^{143}Nd/^{144}Nd of 0.512102 ± 0.000003 (n=5).
lab 14: JNd$_i$ ^{143}Nd/^{144}Nd of 0.512164 ± 0.000011 (n=11; 40 ppb solution).

a sensitive MC-ICP-MS (e.g., a set-up that yields large ion beams even for small abundances of Nd; see Fig. 3 and Table 4). The two laboratories that stand out in Fig. 3 with the largest data spread (numbers 7 and 10), both report results from MC-ICP-MS runs where machine sensitivity hampered a better reproducibility on 15 ppb solutions. In general, a combination of time-resolved analyses, which allows acquisition of isotopic ratios at a faster speed and optimized sample solution concentration to achieve the largest possible Nd beam can potentially improve results (i.e., individual laboratories were requested to run the original 15 ppb solutions).

Returning to the individual results on the Nd isotopic composition of seawater at BATS (Fig. 1; point (2) above), the agreement between laboratories improves from a two sigma

standard deviation of ~60 ppm to ~40 ppm, when omitting the results for laboratories 6 and 10. There is some justification to look at the results like this, as laboratory 6 reported problems with the Fe solution that they used for the seawater analyses from BATS. As outlined above, Fe solution can be a significant source of blank, and this blank can only be neglected for small abundance samples when isotopically characterized as done by Laboratory 2 (see Table 1a). Laboratory 10 reports replicate values for each water depth that overlap with each other, and also overlap within error with the average from all laboratories. However, error bars are large due to poor external reproducibility on low concentration standard runs. The same laboratory reports an improved external reproducibility of 18 to 37 ppm on 100 ppb standard solution. It is

Fig. 3. Neodymium isotopic compositions for an unknown standard solution. The unknown Nd isotope standard was produced from a pure Nd_2O_3 powder. It was distributed in Nd concentrations that mimic seawater concentrations, without requiring further treatment in individual laboratories (i.e., 15 ppm for TIMS laboratories and 15 ppb for MC-ICP-MS laboratories). Analyses by 13 different laboratories yielded a similar reproducibility as reported for seawater analyses (i.e., 56 ppm two sigma standard deviation when taking into account all individual measurements; for seawater results, see Figs. 1 and 2 and Tables 2 and 3). Results furthermore reveal that some laboratories may underestimate their external errors when dealing with low abundance samples (see text for further explanation). Filled symbols: laboratories that carried out analyses by MC-ICP-MS. Open symbols: laboratories that carried out analyses by TIMS. Errors plotted are internal two sigma standard errors.

Fig. 4. Neodymium isotopic composition of marine particles collected during the first intercalibration cruise. Results for particle analyses at three different stations: sub-surface water at BATS (top), Virginia continental slope (middle), and deep water at BATS (bottom). Whereas good agreement between different laboratories can be observed for the samples from the shallow water depth at BATS (two sigma standard deviation of 53 ppm), different leaching/digestion methods used as well as outliers in the analyses make the data for the other two stations inconclusive (two sigma standard deviations worse than 100 ppm; Table 5).

important to note that not all laboratories chose to run their isotopic standards at the same concentrations as the samples (see notes in the caption of Table 2). In detail, Laboratories 7 and 9 report results on significantly higher concentrated standards. In the case of Laboratory 9, ion beam intensity however was controlled at a level similar to the one obtained during sample runs. Laboratory 13 reports a much larger reproducibility on the unknown standard than for their in-house standard (76 versus 25-35 ppm; Tables 2 and 4) for unresolved reasons. Overall, some laboratories may underestimate their external errors by not matching standard and sample concentrations. Many laboratories however yield consistent results between 20 and 40 ppm throughout the intercalibration exercise (see captions of Tables 2-4). Furthermore, failure to reproduce individual samples within one laboratory could hint at unidentified blank issues affecting the accuracy of results (Laboratory 4; Fig. 2).

Neodymium isotope intercalibration in marine particles at BATS

Figure 4 shows the results obtained from six different laboratories on the Nd isotopic composition of marine particulate matter collected at three different locations in the Atlantic Ocean, using two different pumping systems. Following the discussion above on seawater Nd isotope analyses, the different laboratories can be expected to achieve $^{143}Nd/^{144}Nd$ results

that agree within ~60 ppm if two important preconditions are met: (i) the particle composition on each filter was homogenous and reproducible between filters from the same pumping station, and (ii) all laboratories used the same method to either leach particles off the filter, or to perform a total filter and sample digest.

The first point is discussed in more detail in dedicated papers on analyses and sampling of marine particles (Maiti et al. 2012 and Bishop et al. 2012). The latter point is fulfilled for the particle samples from the shallow water depth at BATS, but violated for the other two intercalibration samples (Fig. 4; Table 5). Considering that any particle sample from the ocean is a mixture of biogenic and detrital particles, each of which could have a distinct Nd concentration and isotopic composition and react differently to different chemical leaching/digestion procedures, we would expect to see some dependence on chosen methodology. The actual results are, however, inconclusive. Particles from the shallow water depth at BATS, all leached using the same prescribed methodology following Jeandel et al. (1995), show an average Nd isotopic composition of –9.3 and agree within 0.5 epsilon units (2σ standard deviation of the mean; Fig. 4, Table 5). This result could be interpreted as an excellent agreement for an intercalibration, although statistics are hampered by the small number of data submissions (*n* = 5).

Table 4: Results from individual laboratories on the 'unknown' Nd standard.

lab	$^{143}Nd/^{144}Nd$ measured ± internal 2σ SE	$^{143}Nd/^{144}Nd$ normalized*	ε_{Nd}† ± internal 2σ SE	deviation from average
1	0.511677 ± 0.000008	0.511742	-17.5 ± 0.2	-0.2
1	0.511693 ± 0.000008	0.511740	-17.5 ± 0.2	-0.2
1	0.511721 ± 0.000007	0.511731	-17.7 ± 0.1	-0.4
1	0.511728 ± 0.000005	0.511738	-17.6 ± 0.1	-0.2
average ε_{Nd} and 2σ SD, lab 1:			**-17.6 ± 0.2**	
2	0.511755 ± 0.000011	0.511742	-17.5 ± 0.2	-0.2
2	0.511782 ± 0.000010	0.511769	-17.0 ± 0.2	0.4
2	0.511764 ± 0.000012	0.511751	-17.3 ± 0.2	0.0
2	0.511767 ± 0.000012	0.511754	-17.2 ± 0.2	0.1
2	0.511766 ± 0.000011	0.511753	-17.3 ± 0.2	0.1
2	0.511769 ± 0.000011	0.511756	-17.2 ± 0.2	0.1
2	0.511778 ± 0.000010	0.511765	-17.0 ± 0.2	0.3
2	0.511766 ± 0.000009	0.511753	-17.3 ± 0.2	0.0
2	0.511777 ± 0.000010	0.511764	-17.1 ± 0.2	0.3
2	0.511762 ± 0.000012	0.511749	-17.3 ± 0.2	0.0
2	0.511762 ± 0.000009	0.511749	-17.3 ± 0.2	0.0
average ε_{Nd} and 2σ SD, lab 2:			**-17.2 ± 0.3**	
3	0.511728 ± 0.000009	0.511746	-17.4 ± 0.2	-0.1
3	0.511720 ± 0.000008	0.511738	-17.6 ± 0.2	-0.2
3	0.511720 ± 0.000009	0.511738	-17.6 ± 0.2	-0.2
3	0.511731 ± 0.000008	0.511749	-17.3 ± 0.2	0.0
3	0.511732 ± 0.000007	0.511750	-17.3 ± 0.1	0.0
average ε_{Nd} and 2σ SD, lab 3:			**-17.4 ± 0.2**	
4	0.511733 ± 0.000007	0.511748	-17.3 ± 0.1	0.0
4	0.511740 ± 0.000006	0.511755	-17.2 ± 0.1	0.1
4	0.511745 ± 0.000005	0.511760	-17.1 ± 0.1	0.2
4	0.511737 ± 0.000006	0.511752	-17.2 ± 0.1	0.1
4	0.511745 ± 0.000006	0.511763	-17.0 ± 0.1	0.3
average ε_{Nd} and 2σ SD, lab 4:			**-17.1 ± 0.2**	
6	0.511746 ± 0.000010	0.511756	-17.2 ± 0.2	0.1
6	0.511746 ± 0.000010	0.511756	-17.2 ± 0.2	0.1
6	0.511731 ± 0.000011	0.511741	-17.5 ± 0.2	-0.2
6	0.511750 ± 0.000012	0.511759	-17.1 ± 0.2	0.2
6	0.511729 ± 0.000009	0.511739	-17.5 ± 0.2	-0.2
6	0.511740 ± 0.000013	0.511750	-17.3 ± 0.2	0.0
6	0.511749 ± 0.000014	0.511759	-17.2 ± 0.3	0.2
6	0.511736 ± 0.000011	0.511746	-17.4 ± 0.2	-0.1
6	0.511742 ± 0.000011	0.511752	-17.3 ± 0.2	0.0
6	0.511738 ± 0.000006	0.511748	-17.4 ± 0.1	0.0
6	0.511744 ± 0.000015	0.511754	-17.3 ± 0.3	0.1
6	0.511746 ± 0.000009	0.511755	-17.2 ± 0.2	0.1
average ε_{Nd} and 2σ SD, lab 6:			**-17.3 ± 0.3**	
7	0.511775 ± 0.000013	0.511794	-16.5 ± 0.3	0.9
7	0.511741 ± 0.000013	0.511760	-17.1 ± 0.3	0.2
7	0.511722 ± 0.000011	0.511741	-17.5 ± 0.2	-0.2
7	0.511765 ± 0.000010	0.511784	-16.7 ± 0.2	0.7
7	0.511706 ± 0.000010	0.511725	-17.8 ± 0.2	-0.5
7	0.511764 ± 0.000010	0.511783	-16.7 ± 0.2	0.6
7	0.511726 ± 0.000010	0.511745	-17.4 ± 0.2	-0.1
7	0.511754 ± 0.000012	0.511773	-16.9 ± 0.2	0.4
7	0.511722 ± 0.000012	0.511741	-17.5 ± 0.2	-0.2
7	0.511726 ± 0.000011	0.511745	-17.4 ± 0.2	-0.1
7	0.511726 ± 0.000009	0.511745	-17.4 ± 0.2	-0.1
7	0.511707 ± 0.000010	0.511732	-17.7 ± 0.2	-0.3
7	0.511725 ± 0.000010	0.511750	-17.3 ± 0.2	0.0
7	0.511732 ± 0.000009	0.511757	-17.2 ± 0.2	0.1
7	0.511726 ± 0.000006	0.511751	-17.3 ± 0.1	0.0
average ε_{Nd} and 2σ SD, lab 7:			**-17.2 ± 0.8**	

lab	$^{143}Nd/^{144}Nd$ measured ± internal 2σ SE	$^{143}Nd/^{144}Nd$ normalized*	ε_{Nd}† ± internal 2σ SE	deviation from average
8	0.511754 ± 0.000027	0.511769	-17.0 ± 0.5	0.4
8	0.511731 ± 0.000011	0.511746	-17.4 ± 0.2	-0.1
8	0.511737 ± 0.000007	0.511752	-17.3 ± 0.1	0.0
8	0.511742 ± 0.000008	0.511757	-17.2 ± 0.2	0.1
8	0.511735 ± 0.000008	0.511750	-17.3 ± 0.2	0.0
8	0.511799 ± 0.000019	0.511769	-17.0 ± 0.4	0.4
8	0.511806 ± 0.000024	0.511776	-16.8 ± 0.5	0.5
8	0.511799 ± 0.000006	0.511769	-17.0 ± 0.1	0.4
8	0.511798 ± 0.000012	0.511768	-17.0 ± 0.2	0.3
average ε_{Nd} and 2σ SD, lab 8:			**-17.1 ± 0.4**	
9	0.511725 ± 0.000016	0.511738	-17.5 ± 0.3	-0.2
9	0.511723 ± 0.000014	0.511736	-17.6 ± 0.3	-0.3
9	0.511730 ± 0.000015	0.511743	-17.5 ± 0.3	-0.1
9	0.511732 ± 0.000016	0.511745	-17.4 ± 0.3	-0.1
9	0.511752 ± 0.000015	0.511765	-17.0 ± 0.3	0.3
9	0.511717 ± 0.000017	0.511730	-17.7 ± 0.3	-0.4
9	0.511721 ± 0.000015	0.511734	-17.6 ± 0.3	-0.3
9	0.511716 ± 0.000013	0.511729	-17.7 ± 0.3	-0.4
9	0.511742 ± 0.000014	0.511755	-17.2 ± 0.3	0.1
9	0.511721 ± 0.000014	0.511734	-17.6 ± 0.3	-0.3
average ε_{Nd} and 2σ SD, lab 9:			**-17.5 ± 0.4**	
10	0.511873 ± 0.000012	0.511784	-16.7 ± 0.2	0.7
10	0.511831 ± 0.000012	0.511742	-17.5 ± 0.2	-0.2
10	0.511873 ± 0.000012	0.511784	-16.7 ± 0.2	0.7
10	0.511828 ± 0.000011	0.511739	-17.5 ± 0.2	-0.2
10	0.511828 ± 0.000010	0.511739	-17.5 ± 0.2	-0.2
10	0.511810 ± 0.000010	0.511721	-17.9 ± 0.2	-0.6
10	0.511828 ± 0.000010	0.511739	-17.5 ± 0.2	-0.2
10	0.511850 ± 0.000011	0.511761	-17.1 ± 0.2	0.2
10	0.511865 ± 0.000012	0.511747	-17.4 ± 0.2	-0.1
10	0.511868 ± 0.000011	0.511750	-17.3 ± 0.2	0.0
10	0.511850 ± 0.000010	0.511732	-17.7 ± 0.2	-0.3
10	0.511846 ± 0.000011	0.511728	-17.7 ± 0.2	-0.4
10	0.511822 ± 0.000011	0.511704	-18.2 ± 0.2	-0.9
10	0.511870 ± 0.000011	0.511752	-17.3 ± 0.2	0.0
10	0.511870 ± 0.000011	0.511752	-17.3 ± 0.2	0.0
10	0.511894 ± 0.000011	0.511776	-16.8 ± 0.2	0.5
average ε_{Nd} and 2σ SD, lab 10:			**-17.4 ± 0.9**	
11	0.511730 ± 0.000008	0.511750	-17.3 ± 0.2	0.0
11	0.511728 ± 0.000008	0.511748	-17.4 ± 0.2	0.0
11	0.511736 ± 0.000008	0.511756	-17.2 ± 0.2	0.1
11	0.511720 ± 0.000008	0.511740	-17.5 ± 0.2	-0.2
11	0.511726 ± 0.000009	0.511746	-17.4 ± 0.2	-0.1
average ε_{Nd} and 2σ SD, lab 11:			**-17.4 ± 0.2**	
12	0.511737 ± 0.000003	0.511748	-17.4 ± 0.1	0.0
12	0.511737 ± 0.000002	0.511748	-17.4 ± 0.1	0.0
12	0.511739 ± 0.000003	0.511750	-17.3 ± 0.1	0.0
12	0.511736 ± 0.000004	0.511747	-17.4 ± 0.1	-0.1
12	0.511733 ± 0.000002	0.511744	-17.4 ± 0.1	-0.1
12	0.511732 ± 0.000003	0.511743	-17.4 ± 0.1	-0.1
12	0.511741 ± 0.000002	0.511752	-17.3 ± 0.1	0.0
12	0.511739 ± 0.000003	0.511750	-17.3 ± 0.1	0.0
12	0.511736 ± 0.000003	0.511747	-17.4 ± 0.1	-0.1
average ε_{Nd} and 2σ SD, lab 12:			**-17.4 ± 0.1**	
13	0.511723 ± 0.000016	0.511741	-17.5 ± 0.3	-0.2
13	0.511757 ± 0.000024	0.511775	-16.8 ± 0.5	0.5
13	0.511740 ± 0.000014	0.511758	-17.2 ± 0.3	0.1
13	0.511713 ± 0.000014	0.511731	-17.7 ± 0.3	-0.4
average ε_{Nd} and 2σ SD, lab 13:			**-17.3 ± 0.8**	
14	0.511709 ± 0.000006	0.511749	-17.3 ± 0.1	0.0
14	0.511706 ± 0.000006	0.511745	-17.4 ± 0.1	-0.1
14	0.511695 ± 0.000007	0.511734	-17.6 ± 0.1	-0.3
average ε_{Nd} and 2σ SD, lab 14:			**-17.5 ± 0.3**	

	$^{143}Nd/^{144}Nd$ normalized	ε_{Nd}
average of all measurements	**0.511750**	**-17.3**
2σ SD	**0.000029**	**0.6**

All results represent 15 ng loads for TIMS measurements or runs of 15 ppb solutions for MC-ICP-MS.

* $^{143}Nd/^{144}Nd$ ratios were normalized using the reported standard data for each lab, relative to a JNd_i value of 0.512115 (Tanaka et al., 2000) or a La Jolla value of 0.5115858 (Lugmair et al., 1983).

† ε_{Nd} values were calculated relative to a CHUR of 0.512638 (Jacobsen and Wasserburg, 1980).

lab 1: First two results are normalized to JNd_i $^{143}Nd/^{144}Nd$ ratios of 0.512050 ± 0.000007 and 0.512068 ± 0.000008 respectively. Latter two results are normalized to a JNd_i $^{143}Nd/^{144}Nd$ of 0.512105 ± 0.000015 (n=6).

lab 2: La Jolla $^{143}Nd/^{144}Nd$ of 0.511871 ± 0.000008 (n=16).

lab 3: La Jolla $^{143}Nd/^{144}Nd$ of 0.51184 ± 0.000014 (n=8); JNd_i $^{143}Nd/^{144}Nd$ of 0.512095 ± 0.000011 (n=8). Values were normalized using the reported La Jolla standards.

lab 4: Results normalized to two different JNd_i values: Results 1-4 are normalized to a JNd_i $^{143}Nd/^{144}Nd$ of 0.512100 ± 0.000015 (n=14) and result 5 is normalized to a JNd_i $^{143}Nd/^{144}Nd$ of 0.512097 ± 0.000010 (n=7).

lab 6: La Jolla $^{143}Nd/^{144}Nd$ of 0.511848 ± 0.000009 (n=7).

lab 7: Results from two different measurement sessions. First 11 ratios are normalized using a JNd_i $^{143}Nd/^{144}Nd$ of 0.512096 ± 0.000005 (n=11); remaining ratios are normalized based on JNd_i $^{143}Nd/^{144}Nd$ of 0.512090 ± 0.000006 (n=11) .

lab 8: First five ratios are the result of dynamic mode analysis normalized to a La Jolla $^{143}Nd/^{144}Nd$ of 0.511843 ± 0.000012 (n=12). The last four ratios are the result of static mode analysis normalized to a La Jolla $^{143}Nd/^{144}Nd$ of 0.511888 ± 0.000019 (n=12).

lab 9: JNd_i $^{143}Nd/^{144}Nd$ of 0.512102 ± 0.000015 (n=19).

lab 10: First 11 ratios are normalized using a JNd_i $^{143}Nd/^{144}Nd$ of 0.512090 (n=11); remaining ratios are normalized based on JNd_i $^{143}Nd/^{144}Nd$ of 0.512096 (n=14).

lab 11: La Jolla $^{143}Nd/^{144}Nd$ of 0.511838 ± 0.000015 (n=4).

lab 12: JNd_i $^{143}Nd/^{144}Nd$ of 0.512104 ± 0.000003 (n=4).

lab 13: JMC-321 $^{143}Nd/^{144}Nd$ of 0.511105 ± 0.000013 (n=6), which was calibrated to correspond to a JNd_i value of 0.512097.

lab 14: Results are relative to a linear drift correction on the day of analyses. 40ppb JNd_i $^{143}Nd/^{144}Nd$ of 0.512164 ± 0.000011 (n=11).

Table 5: International intercalibration for Nd isotopes in marine particulates.

lab	sample ID	volume filtered	filter type	method[$]	^{143}Nd/^{144}Nd measured	internal 2σ SE	^{143}Nd/^{144}Nd normalized*	ε$_{Nd}$[†]	internal 2σ SE	external 2σ SD[‡]	Nd [pg/L][$]
BATS, 30m											
2	KN193-6-Nd-133	180L	quartz fibre (QMA)	0.6N HCl leach	0.512162 ± 0.000016		0.512164	-9.2 ± 0.3		0.1	49.90
6	KN193-6-Nd-101	380L	quartz fibre (QMA)	0.6N HCl leach	0.512159 ± 0.000009		0.512169	-9.1 ± 0.2		0.4	55.42
7	KN193-6-Nd-111	380L	quartz fibre (QMA)	0.6N HCl leach	0.512164 ± 0.000016		0.512176	-9.0 ± 0.3		0.3	40.26
10	KN193-6-Nd-115	380L	quartz fibre (QMA)	0.6N HCl leach	0.512220 ± 0.000013		0.512158	-9.4 ± 0.3		0.4	27.01
11	KN193-6-Nd-113	380L	quartz fibre (QMA)	0.6N HCl leach	0.512121 ± 0.000005		0.512140	-9.7 ± 0.1		0.3	35.54
average BATS 30m								**-9.3**			**41.62**
2σ SD								**0.5**			**22.61**
BATS, 2000m											
2	KN193-6-Nd-033	136.8L	0.45 μm Supor	total digest	0.511915 ± 0.000019		0.511917	-14.1 ± 0.4		0.1	39.45
6	KN193-6-Nd-001	270L	0.45 μm Supor	total digest	0.511938 ± 0.000019		0.511948	-13.5 ± 0.4		0.4	51.94
7	KN193-6-Nd-011	300L	0.45 μm Supor	0.6N HCl leach	0.511923 ± 0.000015		0.511935	-13.7 ± 0.3		0.3	36.07
8	KN193-6-Nd-009	270L	0.45 μm Supor	other method[a]	0.511965 ± 0.000025		0.511921	-14.0 ± 0.5		0.3	39.65
11	KN193-6-Nd-013	246.24L	0.45 μm Supor	total digest	0.511814 ± 0.000007		0.511833	-15.7 ± 0.1		0.3	36.46
average BATS 2000m								**-14.2**			**40.71**
2σ SD								**1.8**			**12.98**
Slope, 98m											
2	KN193-6-Nd-231	216.6L	quartz fibre (QMA)	other method[b]	0.512082 ± 0.000011		0.512084	-10.8 ± 0.2		0.1	57.19
6	KN193-6-Nd-201	445L	quartz fibre (QMA)	0.6N HCl leach	0.512015 ± 0.000013		0.512025	-12.0 ± 0.3		0.4	75.35
7	KN193-6-Nd-211	445L	quartz fibre (QMA)	0.6N HCl leach	0.512079 ± 0.000012		0.512091	-10.7 ± 0.2		0.3	38.02
11	KN193-6-Nd-213	445L	quartz fibre (QMA)	0.6N HCl leach	0.512035 ± 0.000005		0.512054	-11.4 ± 0.1		0.3	64.40
average Slope, 98m								**-11.2**			**58.74**
2σ SD								**1.2**			**31.40**

[$] '0.6N HCl leach' follows the method described in Jeandel et al. (1995); 'total digest' follows the method described in Cullen and Sherrell (1999).

* ^{143}Nd/^{144}Nd ratios were normalized relative to a JNdi value of 0.512115 (Tanaka et al., 2000) or a La Jolla value of 0.511858 (Lugmair et al., 1983).

[†] ε$_{Nd}$ values were calculated relative to a CHUR of 0.512638 (Jacobsen and Wasserburg, 1980).

[‡] external errors are derived from repeat standard analyses during the measurement session; if internal errors are larger than external errors, these are plotted in Figure 4.

[$] Nd concentration are corrected for dipped blanks for labs 2, 6, and 8, but not for labs 7, 10, and 11 (no measurements reported on dipped blanks); additional Nd concentrations were reported by lab 16 as follows: BATS, 30m: 65.16 pg/L (corrected for dipped blank); slope: 104.83 pg/L (dipped blank correction for this sample was >200%); for details an external errors on [Nd] methodology see Tables 1b, 6 and 7

[a] total digest using a mixture of HCl/HNO3 and traces of HF (following Landing and Lewis 1991).

[b] 0.005M hydroxylamine hydrochloride in 1.5% acetic acid plus 2nM EDTA solution at pH of 3.5; leached for 3.5hrs on hotplate (90°C) with ultrasonification every hour.

lab 2: Normalized relative to La Jolla ^{143}Nd/^{144}Nd of 0.511856± 0.000007 (n=14).

lab 6: Normalized relative to La Jolla ^{143}Nd/^{144}Nd of 0.511848 ± 0.000004 (12 loads of 100-400ng; 20 loads of an inhouse standard yielded an error of 0.000022 for 4-12 ng loads).

lab 7: Normalized relative to JNd$_i$ ^{143}Nd/^{144}Nd of 0.512103 ± 0.000014 (n=65).

lab 8: Normalized relative to La Jolla ^{143}Nd/^{144}Nd of 0.511902 ± 0.000010 (n=8).

Lab 10: Normalized relative to JNdi ^{143}Nd/^{144}Nd of 0.512177 ± 0.000019 (n=5). Particulate sample from the slope location did not yield enough Nd for an isotopic analyses.

Lab 11: Normalized relative to La Jolla ^{143}Nd/^{144}Nd of 0.511839 ± 0.000014 (n=5).

Applying however the same leaching method at the slope station (same filter type, same pumping system) yielded results that do not agree within error (Laboratories 6 and 7: Fig. 4, Table 5).

Similarly, results from four laboratories for the deep water samples at BATS show excellent agreement (mean Nd isotopic composition of –13.8 ± 0.6) despite the use of different digestion/leaching methods. Two laboratories (numbers 2 and 6) followed the prescribed method for total digestion of particles and filter material described by Cullen and Sherrell (1999). The method used by Laboratory 8 is not too different (see Landing and Lewis 1991), but Laboratory 7 carried out a 0.6M HCl leach. Although using the same digestion method as Laboratories 2 and 6, Laboratory 11 reports a Nd isotopic composition for deep water particles that deviates significantly from the other results (ε$_{Nd}$ = –15.7 ± 0.3).

Fig. 5. Dissolved REE concentrations at two water depths at BATS. Results display good agreement between six international laboratories on measuring the concentrations of REE in seawater (e.g., agreement with 15%, omitting results from laboratory 17; Table 6). (a) REE concentrations normalized to hypothetical seawater of the following composition: La-0.57, Ce-0.7, Pr-0.11, Nd-0.51, Sm-0.1, Eu-0.03, Gd-0.17, Tb-0.03, Dy-0.26, Ho-0.07, Er-0.24, Tm-0.03, Yb-0.18, Lu-0.03 (in [ppt]). (b) REE concentrations normalized to PAAS (Post Archean Australian shale; Nance and Taylor 1976).

At the moment, we can only speculate on the reasons for the inconsistent results reported above, which go along with variable total particle Nd concentrations as well as variable calculated Nd concentrations per pumped amount of seawater (Table 5). Possible explanations include heterogeneous particle distributions on the filters, differences between the filters from individual pumps, or contamination either during sampling, drying, and cutting on the ship, or during processing in the laboratory.

Due to the difficulties with interpreting Nd concentration and isotope data, we refrain at this point from reporting full REE patterns, which were submitted by three laboratories (Laboratories 10, 16, 17).

Rare earth element concentrations in seawater at BATS

Figures 5a and 5b and Table 6 summarize the results for seawater REE concentrations from the two intercalibration samples collected from BATS, analyzed by seven laboratories with significantly different experience levels.

Average REE patterns from all participating laboratories (1,3,5,8,10,16,17), except Laboratory 17 (see discussion below),

show agreement within 12% for deep water at BATS and 15% for shallow water at BATS (relative 2σ SD of the mean = RSD; Table 6). Overall, slightly better agreement is observed for the heavy rare earth elements (HREEs), than for the light rare earth elements (LREE), a result typical for REE measurements. The exception from this overall agreement are Ce concentrations, which display a relative two sigma standard deviation of 44% for individual results from deep waters and 23% for results from shallow waters. A potential reason for this large spread in Ce concentrations reported by different labs are significantly higher blank levels for Ce than for other REEs (4% to 10%). Those laboratories who measured and reported blanks on their REE concentrations found a shale-like REE pattern, implying larger blank contributions to LREE than HREE.

The best fit to the mean REE pattern is observed for Laboratory 3, which reports all REE except La and Ce within 2% RSD. This is not surprising as this laboratory used a mixed REE spike consisting of nine individual REE isotopes. Hence isotope dilution calculations could be performed all along the REE spectrum and not just for one or two isotopes as done by

Table 6: International intercalibration for REE in seawater at BATS.

lab	sample ID	sample no.ª	La [ppt]ᵇ	±2σ SE (%)	Ce [ppt]	±2σ SE (%)	Pr [ppt]	±2σ SE (%)	Nd [ppt]	±2σ SE (%)	Sm [ppt]	±2σ SE (%)	Eu [ppt]	±2σ SE (%)	Gd [ppt]	±2σ SE (%)	Tb [ppt]	±2σ SE (%)	Dy [ppt]	±2σ SE (%)	Ho [ppt]	±2σ SE (%)	Er [ppt]	±2σ SE (%)	Tm [ppt]	±2σ SE (%)	Yb [ppt]	±2σ SE (%)	Lu [ppt]	±2σ SE (%)
BATS, 2000m																														
1	KN193-6-Nd-329	A	3.267	0.135	0.697	0.097	0.548	0.088	2.413	0.126	0.527	0.121	0.143	0.030	0.766	0.124	0.134	0.024	0.919	0.078	0.257	0.025	0.834	0.075	0.129	0.023	0.809	0.076	0.141	0.015
1	KN193-6-Hf-312	B	3.239	0.086	0.574	0.091	0.546	0.078	2.398	0.249	0.555	0.207	0.139	0.023	0.774	0.155	0.129	0.020	0.942	0.153	0.254	0.026	0.835	0.060	0.130	0.015	0.835	0.109	0.144	0.021
2	KN193-6-Nd-305	A1							2.535		0.516																				
2	KN193-6-Nd-305	A2							2.471		0.500																				
3	KN193-6-Nd-323	A	3.435	0.608	0.699	0.957	0.578	0.436	2.533	0.970	0.520	0.781	0.136	1.125	0.772	0.551	0.124	0.557	0.953	0.531	0.247	0.530	0.849	1.105	0.127	0.599	0.827	0.720	0.139	0.717
3	KN193-6-Nd-324	B	3.391	0.732	0.633	0.797	0.568	0.378	2.518	0.571	0.515	0.887	0.133	1.004	0.768	0.499	0.124	0.591	0.952	0.472	0.246	0.705	0.847	1.120	0.126	0.864	0.822	0.659	0.137	0.603
5	KN193-6-Nd-310	A	3.040		1.119		0.549		2.392		0.508		0.129		0.706		0.121		0.933		0.241		0.837		0.117		0.783		0.141	
5	KN193-6-Nd-311	B	3.013		0.923		0.552		2.483		0.503		0.125		0.742		0.118		0.929		0.246		0.824		0.126		0.820		0.139	
6	KN193-6-Nd-315	A							2.500																						
6	KN193-6-Nd-318	B							2.500																						
8	KN193-6-Nd-304	A	3.633	0.063	0.859	0.015	0.620	0.011	2.712	0.058	0.575	0.015	0.149	0.005	0.828	0.026	0.130	0.004	1.001	0.027	0.265	0.007	0.885	0.024	0.130	0.004	0.863	0.021	0.148	0.003
8	KN193-6-Nd-304	B	3.601	0.046	0.831	0.015	0.600	0.009	2.693	0.050	0.564	0.011	0.152	0.004	0.825	0.033	0.129	0.003	1.008	0.022	0.258	0.005	0.889	0.018	0.130	0.002	0.864	0.024	0.146	0.004
9	KN193-6-Nd-333	A							2.540	0.110																				
10	KN193-6-Nd-302	A1	3.105		0.600		0.549		2.376		0.506		0.137		0.772		0.129		0.935		0.249		0.823				0.815		0.137	
10	KN193-6-Nd-302	A2	3.191		0.612		0.558		2.420		0.514		0.142		0.753		0.124		0.940		0.257		0.842				0.842		0.142	
10	KN193-6-Nd-303	B1	3.063		0.583		0.535		2.431		0.478		0.130		0.762		0.122		0.900		0.251		0.835				0.810		0.138	
10	KN193-6-Nd-303	B2	3.246		0.596		0.563		2.485		0.505		0.133		0.795		0.133		0.961		0.257		0.853				0.828		0.141	
11	KN193-6-Nd-320								2.430																						
12	KN193-6-Hf-315	a							2.600																						
12	KN193-6-Hf-315	b							2.560																						
16	KN193-6-Nd-330	A	3.333		0.655		0.591		2.490		0.510		0.141		0.696		0.116		0.921		0.243		0.828		0.125		0.814		0.141	
16	KN193-6-Nd-330	B	3.351		0.660		0.597		2.495		0.507		0.144		0.694		0.115		0.912		0.241		0.822		0.122		0.808		0.141	
17	KN193-6-Nd-309	A	1.229	0.960	0.302	0.600	0.293	2.400	1.355	3.040	0.270	1.640	0.077	5.800	0.410	1.220	0.068	1.060	0.504	4.960	0.122	1.000	0.378	0.580	0.045	3.180	0.243	1.480	0.032	3.320
17	KN193-6-Nd-308	B	1.742	0.660	0.423	0.680	0.410	0.860	1.904	2.720	0.387	2.340	0.108	1.500	0.410	1.360	0.095	2.360	0.720	1.640	0.180	0.600	0.572	2.980	0.068	3.100	0.378	3.460	0.050	3.780
average value (omitting lab 17)			**3.279**		**0.717**		**0.568**		**2.499**		**0.519**		**0.138**		**0.761**		**0.125**		**0.943**		**0.251**		**0.843**		**0.126**		**0.824**		**0.141**	
2σ SD			0.388		0.318		0.050		0.176		0.051		0.015		0.083		0.012		0.062		0.015		0.042		0.008		0.043		0.007	
relative 2σ SD			11.8		44.4		8.9		7.0		9.8		11.2		10.9		9.5		6.5		5.9		5.0		6.4		5.3		4.6	
BATS, 15m																														
1	KN193-6-Nd-532	A	2.072	0.167	1.646	0.145	0.438	0.067	2.034	0.240	0.527	0.170	0.138	0.027	0.780	0.127	0.135	0.016	0.969	0.090	0.254	0.033	0.811	0.068	0.124	0.023	0.739	0.091	0.121	0.013
1	KN193-6-Hf-518	B	1.977	0.143	1.455	0.121	0.426	0.080	1.941	0.176	0.510	0.114	0.138	0.024	0.762	0.106	0.135	0.036	0.957	0.121	0.248	0.037	0.810	0.095	0.122	0.026	0.724	0.096	0.119	0.020
2	KN193-6-Nd505	A1							2.036		0.472																				
2	KN193-6-Nd505	A2							2.038		0.474																				
3	KN193-6-Nd-526	A	2.079	0.636	1.539	0.693	0.438	0.436	2.031	0.630	0.483	0.674	0.134	1.116	0.766	0.619	0.129	0.427	0.971	0.424	0.245	0.692	0.813	0.917	0.117	0.584	0.726	0.713	0.115	0.772
3	KN193-6-Nd-527	B	2.057	0.765	1.543	0.626	0.435	0.557	2.021	0.748	0.480	0.862	0.132	1.217	0.766	0.572	0.128	0.604	0.972	0.461	0.245	0.560	0.815	0.944	0.117	0.625	0.730	0.756	0.116	0.738
5	KN193-6-Nd-510	A	2.005		2.025		0.447		2.071		0.480		0.127		0.755		0.128		0.962		0.249		0.807		0.115		0.718		0.119	
5	KN193-6-Nd-512	C	1.777		1.901		0.407		1.908		0.454		0.120		0.683		0.119		0.900		0.231		0.742		0.105		0.632		0.109	
5	KN193-6-Nd-513	D	1.932		1.828		0.394		1.938		0.445		0.123		0.650		0.109		0.882		0.226		0.790		0.117		0.816		0.136	
6	KN193-6-Nd-516	A							2.040																						
6	KN193-6-Nd-518	B							2.080																						
8	KN193-6-Nd-504	A	2.345	0.028	1.966	0.026	0.491	0.007	2.277	0.044	0.541	0.020	0.146	0.005	0.815	0.025	0.135	0.004	1.057	0.017	0.272	0.004	0.891	0.025	0.126	0.004	0.782	0.019	0.130	0.003
8	KN193-6-Nd-504	B	2.286	0.020	1.823	0.015	0.464	0.007	2.182	0.039	0.502	0.017	0.144	0.004	0.801	0.018	0.129	0.003	0.995	0.026	0.254	0.003	0.826	0.019	0.120	0.002	0.728	0.021	0.119	0.002
8	KN193-6-Nd-504	C	2.297	0.028	1.800	0.023	0.475	0.007			0.503	0.040	0.144	0.006	0.784	0.029	0.128	0.003	1.008	0.022	0.255	0.003	0.845	0.018	0.117	0.003	0.735	0.024	0.118	0.002
9	KN193-6-Nd-536	A							1.880	0.130																				
10	KN193-6-Nd-502	A1	1.928		1.479		0.426		1.990		0.472		0.132		0.736		0.124		0.914		0.232		0.759				0.658		0.108	
10	KN193-6-Nd-502	A2	1.995		1.518		0.422		2.011		0.446		0.133		0.806		0.125		0.931		0.235		0.759				0.673		0.102	
10	KN193-6-Nd-503	B1	1.873		1.444		0.414		1.974		0.445		0.133		0.733		0.125		0.936		0.242		0.772				0.679		0.109	
10	KN193-6-Nd-503	B2	1.992		1.512		0.431		2.000		0.480		0.130		0.781		0.130		0.984		0.244		0.805				0.716		0.114	
11	KN193-6-Nd-523								2.010																						
12	KN193-6-Hf-539	a							2.050																						
12	KN193-6-Hf-539	b							2.100																						
16	KN193-6-Nd-533	A	2.072		1.739		0.461		2.032		0.481		0.143		0.774		0.122		0.951		0.243		0.802		0.116		0.715		0.119	
16	KN193-6-Nd-533	B	2.081		1.728		0.459		2.043		0.484		0.144		0.770		0.121		0.950		0.244		0.804		0.116		0.724		0.118	
17	KN193-6-Nd-508	A	0.590	2.500	0.558	0.380	0.167	2.980	0.819	1.240	0.185	1.440	0.054	3.120	0.297	3.360	0.050	2.400	0.378	2.280	0.095	2.100	0.293	2.400	0.036	2.520	0.189	1.480	0.027	5.120
17	KN193-6-Nd-509	B	1.080	1.160	0.986	0.840	0.293	0.480	1.359	2.560	0.306	9.560	0.086	3.440	0.477	2.360	0.077	1.840	0.576	2.820	0.135	0.520	0.410	1.260	0.050	2.560	0.257	2.240	0.032	3.880
average value (omitting lab 17)			**2.048**		**1.684**		**0.439**		**2.037**		**0.482**		**0.135**		**0.760**		**0.126**		**0.959**		**0.245**		**0.803**		**0.118**		**0.719**		**0.117**	
2σ SD			0.307		0.384		0.052		0.179		0.054		0.013		0.086		0.013		0.085		0.022		0.071		0.011		0.089		0.016	
relative 2σ SD			15.0		22.8		11.8		8.8		11.2		9.8		11.4		10.7		8.9		9.0		8.9		9.0		12.3		13.8	

ª Samples A and B are separate samples from the same station and water depth, numbers indicate repeat analyses of the same sample.
ᵇ All concentrations have been converted to ppt. Originally reported units are as follows: labs 1, 3, 5, 9, 10: [pg/kg]; lab 2: [pmol/kg]; lab 6: [pM]; lab 8: [ppt]; lab 11,16: ng/L (no correction for seawater density).
Lab 1: Each value reported is the average of a number of individual runs: KN193-6-Nd-329, n=10; KN193-6-Hf-312, n=9; KN193-6-Nd-532, n=18; KN193-6-Hf-518, n=12. Errors are given as two sigma standard deviations of replicate analyses.
Lab 2: Processed a subsample of the spiked large volume sample for Sm and Nd concentrations (see Table 1).
Lab 3: Reported an external two sigma reproducibility of 3% for Ce, and <2% for all other REE (n=5; in-house standard).
Lab 5: Note that sample KN193-6-Nd-513 was processed with a different method compared to the other samples (see Table 1b).
Lab 8: Processed a subsample of the spiked large volume sample for Nd concentrations (see Table 1). Estimates a reproducibility of 0.5% of repeat Nd concentration measurements.
Lab 9: Reported value represents the average of three measurements with its two sigma standard deviation. External reproducibility for Nd concentrations is reported as 0.1 ppt based on repeated mixed standard-spike analyses.
Lab 10: Reports a relative standard deviation of 5% for LREE and 0.7% for HREE (n=6).
Lab 11: Reported value represents the average of seven measurements. The external two sigma standard deviation is reported as 0.142.
Lab 12: Processed the spiked large volume sample for Nd concentrations at the same time as measuring the isotopic composition (see Table 1). External error based on 5 repeat BCR analyses: 5.6%.
Lab 16: Reported an external reproducibility of 8.8% for Ce, 4.8% for Eu and <3.6% for all other REE based on repeat in-house seawater standard measurements (n=29).
Lab 17: Reports a two sigma standard deviation on trace metal data of 10% based on repeated runs of rock standard AGV-1.

Table 7: Results from individual laboratories on an 'unknown' REE standard.

lab	La [ppm]	±2σ SE	Ce [ppm]	±2σ SE	Pr [ppm]	±2σ SE	Nd [ppm]	±2σ SE	Sm [ppm]	±2σ SE	Eu [ppm]	±2σ SE	Gd [ppm]	±2σ SE	Tb [ppm]	±2σ SE	Dy [ppm]	±2σ SE	Ho [ppm]	±2σ SE	Er [ppm]	±2σ SE	Tm [ppm]	±2σ SE	Yb [ppm]	±2σ SE	Lu [ppm]	±2σ SE
1	9.570		12.501		1.951		8.386		1.513		0.529		3.557		0.646		4.070		1.376		4.309		0.614		2.868		0.513	
1	10.019		12.243		1.825		8.465		1.751		0.535		2.723		0.555		4.519		1.254		4.020		0.549		3.346		0.484	
1	9.927		12.356		1.947		8.774		1.784		0.501		2.924		0.553		4.437		1.270		4.069		0.560		3.187		0.518	
8	9.107	0.192	11.234	0.240	1.748	0.039	8.053	0.171	1.571	0.040	0.463	0.011	2.791	0.066	0.488	0.011	3.952	0.094	1.111	0.025	3.781	0.089	0.489	0.012	2.709	0.068	0.449	0.011
8	9.032	0.180	11.116	0.180	1.727	0.180	7.949	0.180	1.577	0.180	0.457	0.180	2.786	0.180	0.498	0.180	3.983	0.180	1.116	0.180	3.804	0.180	0.492	0.180	2.740	0.180	0.451	0.180
8	9.033	0.213	11.162	0.264	1.735	0.038	7.883	0.180	1.554	0.040	0.460	0.012	2.769	0.073	0.490	0.011	3.931	0.106	1.096	0.028	3.715	0.102	0.477	0.012	2.675	0.065	0.437	0.012
8	9.038	0.196	11.220	0.248	1.747	0.039	7.990	0.203	1.554	0.046	0.459	0.015	2.774	0.070	0.490	0.012	3.905	0.109	1.104	0.024	3.774	0.097	0.486	0.013	2.703	0.069	0.450	0.011
8	9.148	0.201	11.283	0.249	1.753	0.043	8.072	0.228	1.562	0.046	0.468	0.014	2.799	0.074	0.490	0.013	3.924	0.096	1.112	0.028	3.800	0.110	0.479	0.013	2.698	0.074	0.445	0.009
8	9.040	0.206	11.130	0.264	1.732	0.038	7.923	0.201	1.561	0.044	0.460	0.016	2.768	0.080	0.488	0.011	3.921	0.107	1.105	0.028	3.773	0.100	0.479	0.013	2.685	0.068	0.446	0.013
8	9.173	0.190	11.224	0.237	1.754	0.040	8.059	0.198	1.581	0.037	0.461	0.013	2.762	0.070	0.487	0.011	3.899	0.087	1.102	0.026	3.765	0.098	0.486	0.011	2.666	0.075	0.449	0.009
8	9.089	0.186	11.206	0.233	1.749	0.036	7.984	0.178	1.578	0.043	0.453	0.010	2.738	0.071	0.485	0.011	3.917	0.097	1.094	0.021	3.758	0.094	0.479	0.011	2.676	0.053	0.444	0.009
8	8.935	0.217	11.080	0.254	1.731	0.042	7.912	0.193	1.562	0.043	0.455	0.012	2.767	0.080	0.489	0.012	3.878	0.097	1.098	0.026	3.745	0.088	0.488	0.013	2.693	0.074	0.446	0.010
8	8.856	0.174	11.012	0.228	1.709	0.035	7.748	0.171	1.559	0.043	0.458	0.013	2.739	0.072	0.483	0.010	3.893	0.091	1.106	0.023	3.755	0.091	0.490	0.013	2.714	0.065	0.449	0.012
17	7.958	0.080	9.402	0.150	1.506	0.055	6.714	0.208	1.367	0.031	0.376	0.031	2.520	0.071	0.465	0.016	3.652	0.150	0.991	0.015	3.388	0.203	0.439	0.015	2.607	0.088	0.424	0.003
17	8.102	0.457	9.808	0.347	1.503	0.054	6.799	0.243	1.354	0.158	0.390	0.037	2.520	0.134	0.454	0.020	3.679	0.300	1.022	0.100	3.461	0.104	0.446	0.025	2.668	0.091	0.420	0.023
17	8.295	0.355	9.911	0.299	1.591	0.004	6.775	0.159	1.355	0.031	0.376	0.014	2.473	0.057	0.440	0.014	3.676	0.097	1.012	0.033	3.505	0.301	0.447	0.010	2.595	0.115	0.417	0.017
average	**9.0**		**11.1**		**1.7**		**7.8**		**1.5**		**0.5**		**2.8**		**0.5**		**4.0**		**1.1**		**3.8**		**0.5**		**2.8**		**0.5**	
2σ SD	1.7		2.7		0.4		1.8		0.3		0.1		0.6		0.1		0.7		0.3		0.7		0.1		0.6		0.1	
relative 2σ SD	19.1		24.0		21.7		23.4		21.4		31.1		20.3		26.7		17.1		26.2		18.0		26.6		19.6		18.9	

For details on methods, see Tables 1b and 6.
Average values are calculated from the mean values for all three laboratories.

all other laboratories (Table 1b). A similarly good result is however obtained by Laboratory 10, for the deep water samples where results agree within 3% RSD with the average values. Laboratories 1, 5, and 16 show deviations from the average within 6% to 8% for both water depths, and Laboratory 8 is just outside this margin with agreements within 10% to 13% (RSDs cited include all REE but Ce).

How good is the observed agreement? Taking into account that most individual laboratories quote external errors for their methodologies well below 10%, it might be expected that an agreement within 10% should be possible. The observed agreement is not far from this number, and data returns on an unknown REE standard solution sent to the individual laboratories did not yield an improved agreement over natural samples (Table 7; Fig. 6). The last statement is based on comparison of results from Laboratories 1 and 8, which show an agreement within 10% for most REE, and 16% for all REE (relative 2σ SD).

REE concentrations reported by Lab 17 (not shown in Fig. 5) are significantly lower than results obtained by other laboratories (~30% to 70%). Looking into the detail of the methodology applied by this laboratory (Table 1b), a number of factors could be responsible for this offset. First, the laboratory did not spike their concentration samples, which means that imperfect yields and fractionation of REE during preconcentration and ion chromatography cannot be corrected for. Second, the laboratory used the REE cut from their Nd isotope chemistry to determine REE concentrations, which again stresses that sample recovery up to this point had to be assumed (i.e., 100% yield). Finally, calculations of REE contents were based on the assumption that the starting sample volume was 10 L, without, however, determining the exact weight before Fe coprecipitation. Shipboard sampling easily introduces an error of 10% on this assumption, as no efforts were undertaken to collect exact sample volumes. The most likely explanation for the deviation of Laboratory 17 from the average result is therefore a combination of unaccounted sample loss, and an erroneous assumption on sample volume. It is notable that the laboratory significantly improved their methodology for the analysis of the unknown standard. However, the deviation from the average of the two other laboratories is still 11% to 22%. Further improvement can probably be achieved through close exchange with one of the other laboratories and improved spike calibration.

A better statistical database for evaluating reproducibility of individual REE concentrations can be achieved when considering Nd concentrations only. Besides the seven laboratories that reported full REE pattern for seawater at BATS, three additional laboratories spiked their large volume isotope samples to determine Nd concentrations (Laboratories 2, 6, 12), and two other laboratory (9 and 11) obtained separate 0.25 to 0.5 L aliquots for Nd concentration measurements (Table 6). Results for Nd concentrations from all 11 laboratories (omitting Laboratory 17) are shown in Fig. 7 for the two intercali-


Fig. 6. REE concentrations for an unknown standard solution. Results reported by three laboratories on a pure REE standard solution that was mixed to roughly match seawater concentrations. The standard was sent around with the information that concentrations for Ce are approximately 10 ppm, and that dilutions for analyses should be carried out accordingly. Laboratories 1 and 8 show good agreement, whereas laboratory 17 (not shown in Fig. 5) deviates significantly from the average value.


Fig. 7. Dissolved neodymium concentrations at BATS. Neodymium concentrations were not only analyzed by the laboratories interested in REE concentration analyses, but also by some laboratories that are primarily interested in Nd isotopic compositions. Eleven laboratories (omitting laboratory 17) show an agreement of Nd concentrations for two different water samples within 9%, reflecting the precision currently possible on REE concentration measurements. Error bars reflect external errors (not reported by every laboratory; see Table 6). The average value and its two sigma standard deviation have been calculated from the mean values for each laboratory in order to not weigh the average toward results from laboratories that report larger amounts of individual analyses. For comparison, Table 6 shows the statistics when calculating the average from all individual data points reported.

bration water depths at BATS. Shallow waters yield a Nd concentration of 2.0 ± 0.2 ng/kg, and deep waters are slightly more concentrated at 2.5 ± 0.2 ng/kg. Errors are absolute two sigma standard deviations from the mean of all data, which translate to 9% and 7% RSD, respectively. This result confirms that a 10% envelope is probably the best agreement currently possible for REE analyses between different laboratories.

Summary and recommendations

Below, we summarize the most important observations from this intercalibration exercise and provide some recommendations for future analyses of Nd isotopes and REE concentrations in seawater and marine particles.

Dissolved Nd isotopes

Return of results from 15 laboratories that participated in the dissolved Nd isotope intercalibration at three locations shows agreement within 0.5 to 0.6 epsilon units for ^{143}Nd/^{144}Nd ratios (2σ standard deviation of the mean).

Return of results from 13 laboratories on an unknown Nd isotope standard solution yields a very similar agreement (0.6 epsilon units), suggesting that mass spectrometry is the single most important factor governing interlaboratory precision on Nd isotope analyses.

Recommendation: Considering the global range of Nd isotopes in seawater (i.e., more than 20 epsilon units), the result of this intercalibration is very satisfying and no further action is required. A slight improvement can probably be achieved through taking extra care about the blank of the Fe solution used for coprecipitation. Otherwise improvement will depend on the availability of sensitive mass spectrometers and detailed comparison of data reduction protocols during mass spectrometry. To allow rigorous evaluation of the accuracy and precision of seawater Nd isotope analyses reported by individual laboratories, it is crucial that all laboratories consider for the future to report their external reproducibility based on standard and/or sample analyses of similar concentration to targeted seawater samples.

Dissolved REE concentrations

Results have been returned from seven laboratories for REE concentrations in seawater. One laboratory has obvious analytical issues, which can be identified and have been discussed with and addressed by the laboratory. All other laboratories show good agreement, with all REE (except Ce) reproducing within 15% of the average value.

Cerium reproducibility is significantly worse with deviations of 22% and 44% from the average value, most likely due to blank contributions. Where analyzed, blanks show a shale-like REE pattern, and hence will have a larger impact on LREE compared with HREE.

Neodymium concentrations determined by 12 laboratories agree within 9% (with one exception) and probably mark the best reproducibility currently possible for individual REE concentrations.

Recommendation: It would be good if all laboratories would analyze and report the results on a pure standard solution to verify the main source of uncertainty in the reproducibility of REE patterns. Spike calibration seems to be the most likely candidate for variability together with the exact methodology used to calculate concentrations for unspiked elements. The shapes of the REE pattern seem quite robust, but absolute values should not be interpreted to a level better than 10%.

Particles (Nd isotopes and REE concentrations)

Partial or complete results on Nd isotopes have been obtained from six laboratories for marine particles. Whereas 0.6 M leaches on particles from the open ocean setting at BATS agree very well, the more particle-rich station on the continental slope shows discrepancies in results, even when the same digestion method is used. Neodymium isotope signatures obtained for particles from deep waters agree fairly well although different leaching/digestion methods have been applied (omitting one outlier). It is unclear at this point whether the described inconsistencies result from sample heterogeneity or analytical artifacts.

Results on Nd concentrations show a large spread and seem to confirm the overall difficulties (at present) with precise analysis of REE concentrations and Nd isotopic composition in marine particles.

Recommendation: The particle intercalibration should be revisited taking into account results presented in the specific papers (Bishop et al. 2012; Maiti et al. 2012). For the time being, one has to be careful to only compare data obtained by the same leaching/digestion method and probably even from the same laboratory (see also Pahnke et al. 2012).

References

Alibo, D. S., and Y. Nozaki. 1999. Rare earth elements in seawater: Particle association shale-normalization, and Ce oxidation. Geochim. Cosmochim. Acta 63:363-372 [doi:10.1016/S0016-7037(98)00279-8].

Bishop, J., P. Lam, T. Wood. In press. Getting good particles: Accurate sampling of particles by large volume in-situ filtration. Limnol Oceanogr. Methods.

Bishop, J. K. B., and T. J. Wood. 2008. Particulate matter chemistry and dynamics in the twilight zone at VERTIGO ALOHA and K2 sites. Deep Sea Res. I 55:1684-1706 [doi:10.1016/j.dsr.2008.07.012].

Chu, Z., F. Chen, Y. Yang, and J. Guo. 2009. Precise determination of Sm, Nd concentrations and Nd isotopic compositions at the nanogram level in geological samples by thermal ionization mass spectrometry. J. Anal. At. Spectrom. 24:1534-1544 [doi:10.1039/b904047a].

Collier, R. W., and J. M. Edmond. 1984. The trace element geochemistry of marine biogenic particulate matter. Progr. Oceanogr. 13:113-199 [doi:10.1016/0079-6611(84)90008-9].

Cullen, J. T., and R. M. Sherrell. 1999. Techniques for determination of trace metals in small samples of size-fractionated particulate matter: phytoplankton metals off central California. Mar. Chem. 67:233-247 [doi:10.1016/S0304-4203(99)00060-2].

DePaolo, D. J., and G. J. Wasserburg. 1977. The sources of island arcs as indicated by Nd and Sr isotopic studies. Geophys. Res. Lett. 4:465-468 [doi:10.1029/GL004i010p00465].

Dodson, R. W., G. J. Forney, and E. H. Swift. 1936. The extraction of ferric chloride from hydrochloric acid solutions by isopropyl ether. J. Amer. Chem. Soc. 58:2573-2577

[doi:10.1021/ja01303a058].

Elderfield, H., and M. J. Greaves. 1982. The rare earth elements in seawater. Nature 296:214-219 [doi:10.1038/296214a0].

Eugster, O., F. Tera, D. S. Burnett, and G. J. Wasserburg. 1970. The isotopic composition of gadolinium and neutron capture effect in some meteorites. J. Geophys. Res. 75:2753-2768 [doi:10.1029/JB075i014p02753].

Goldstein, S. L., and S. R. Hemming. 2003. Long-lived isotopic tracers in oceanography, paleoceanography, and ice-sheet dynamics, p. 453-489. *In* H. D. Holland, K. K. Turekian, and H. Elderfield [eds.], Treatise on geochemistry: The oceans and marine geochemistry. Elsevier.

GEOTRACES Planning Group. 2006. GEOTRACES science plan. Scientific Committee on Oceanic Research.

Halliday, A. N., J. N. Christensen, D. -C. Lee, M. Rehkämper, C. M. Hall, and X. Luo. 2000. Multiple collector inductively coupled plasma mass spectrometry, p. 291-328. *In* C. B. Barshick, D. C. Duckworth, and D. H. Smith [eds.], Inorganic mass spectrometry: fundamentals and applications. Marcel Dekker.

Harvey, J., and E. F. Baxter. 2009. An improved method for TIMS high precision neodymium isotope analysis of very small aliquots (1-10 ng). Chem. Geol. 258:251-257 [doi:10.1016/j.chemgeo.2008.10.024].

Hooker, P. J., R. K. O'Nions, and R. J. Pankhurst. 1975. Determination of rare-earth elements in USGS standard rocks by mixed-solvent ion exchange and mass-spectrometric isotope dilution. Chem. Geol. 16:189-196 [doi:10.1016/0009-2541(75)90027-3].

Jacobsen, S. B., and G. J. Wasserburg. 1980. Sm-Nd isotopic evolution of chondrites. Earth Planet. Sci. Lett. 50:139-155 [doi:10.1016/0012-821X(80)90125-9].

Jeandel, C., J. K. Bishop, and A. Zindler. 1995. Exchange of neodymium and its isotopes between seawater and small and large particles in the Sargasso Sea. Geochim. Cosmochim. Acta 59:535-547 [doi:10.1016/0016-7037(94)00367-U].

Lacan, F., and C. Jeandel. 2005. Neodymium isotopes as a new tool for quantifying exchange fluxes at the continent-ocean interface. Earth Planet. Sci. Lett. 232:245-257 [doi:10.1016/j.epsl.2005.01.004].

Landing, W. M., and B. L. Lewis. 1991. Analysis of marine particulate and colloidal material for transition metals, p. 263-272. *In* D. C. Hurd and D. W. Spencer [eds.], Marine particles: analysis and characterization. Geophysical Monograph 63. AGU.

Li, C. -F., F. Chen, and X. -H. Li. 2007. Precise isotopic measurements of sub-nanogram Nd of standard reference material by thermal ionization mass spectrometry using the NdO^+ technique. Int. J. Mass Spectrom. 266:34-41 [doi:10.1016/j.ijms.2007.06.013].

Lugmair, G. W., N. B. Scheinin, and K. Marti. 1975. Sm-Nd age and history of Apollo 17 basalt 75075: Evidence for early differentiation of the lunar exterior. *In* Proc. 6th Lunar Sci.

Conf., Pergamon Press, Inc., New York. p. 1419-1429.

———, T. Shimamura, R. S. Lewis, and E. Anders. 1983. Samarium-146 in the early solar system: evidence from neodymium in the allende meteorite. Science 222:1015-1018 [doi:10.1126/science.222.4627.1015].

Maiti, K., and K. O. Buesseler. 2012. 234th intercalibration.

Nance, W. B., and S. R. Taylor. 1976. Rare earth element patterns and crustal evolution–I. Australian post-Archean sedimentary rocks. Geochim. Cosmochim. Acta 40:1539-1551 [doi:10.1016/0016-7037(76)90093-4].

O'Nions, R. K., P. J. Hamilton, and N. M. Evenson. 1977. Variations in the $^{143}Nd/^{144}Nd$ and $^{87}Sr/^{86}Sr$ ratios in oceanic basalts. Earth Planet. Sci. Lett. 34:13-22 [doi:10.1016/0012-821X(77)90100-5].

O'Nions, R. K., S. R. Carter, R. S. Cohen, N. M. Evensen, and P. J. Hamilton. 1978. Pb, Nd and Sr isotopes in oceanic ferromanganese deposits and ocean floor basalts. Nature 273:435-438 [doi:10.1038/273435a0].

Pahnke, K., T. van de Flierdt, K. Jones, S. R. Hemming, and S. L. Goldstein. 2012. GEOTRACES intercalibration of neodymium isotopes and rare earth element concentrations in seawater and particles—Part 2: systematic tests and baseline profiles. Limnol Oceanogr. Methods 10:252-269 [doi:10.4319/lom.2012.10.252].

Piepgras, D. J., and S. B. Jacobsen. 1988. The isotopic composition of neodymium in the North Pacific. Geochim. Cosmochim. Acta 52:1373-1381 [doi:10.1016/0016-7037(88)90208-6].

———, and S. B. Jacobsen. 1992. The behavior of rare earth elements in seawater: Precise determination of variation in the North Pacific water column. Geochim. Cosmochim. Acta 56:1851-1862 [doi:10.1016/0016-7037(92)90315-A].

———, and G. J. Wasserburg. 1980. Neodymium isotopic variations in seawater. Earth Planet. Sci. Lett. 50:128-138 [doi:10.1016/0012-821X(80)90124-7].

———, and ———. 1987. Rare earth element transport in the western North-Atlantic inferred from Nd isotopic observations. Geochim. Cosmochim. Acta 51:1257-1271 [doi:10.1016/0016-7037(87)90217-1].

———, G. J. Wasserburg, and E. J. Dasch. 1979. The isotopic composition of Nd in different ocean masses. Earth Planet. Sci. Lett. 45:223-236 [doi:10.1016/0012-821X(79)90125-0].

Pin, C., and J. F. S. Zalduegui. 1997. Sequential separation of light rare-earth elements, thorium and uranium by miniaturized extraction chromatography: Application to isotopic analyses of silicate rocks. Anal. Chim. Acta 33:79-89 [doi:10.1016/S0003-2670(96)00499-0].

Richard, P., N. Shimizu, and C. J. Allègre. 1976. $^{143}Nd/^{146}Nd$, a natural tracer: an application to oceanic basalts. Earth Planet. Sci. Lett. 31:269-278 [doi:10.1016/0012-821X(76)90219-3].

Shabani, M. B., T. Akagi, and A. Masuda. 1992. Preconcentration of trace rare-earth elements in seawater by complexation with bis(2-ethylhexyl) hydrogen phosphate and 2-eth-

ylhexyl dihydrogen phosphate adsorbed on a C18 cartridge and determination by inductively coupled plasma mass-spectrometry. Anal. Chem. 64:737-743 [doi:10.1021/ac00031a008].

Stordal, M. C., and G. J. Wasserburg. 1986. Neodymium isotopic study of Baffin Bay water: sources of REE from very old terranes. Earth Planet. Sci. Lett. 77:259-272 [doi:10.1016/0012-821X(86)90138-X].

Tachikawa, K., V. Athias, and C. Jeandel. 2003. Neodymium budget in the modern ocean and paleo-oceanographic implications. J. Geophys. Res. 108:3254 [doi:10.1029/1999JC000285].

Tanaka, T., and others. 2000. JNdi-1: a neodymium isotopic reference in consistency with LaJolla neodymium. Chem. Geol. 168:279-281 [doi:10.1016/S0009-2541(00)00198-4].

Thirlwall, M. F. 1991a. Long-term reproducibility of multicollector Sr and Nd isotope ratio analysis. Chem. Geol. 94:85-104 [doi:10.1016/S0009-2541(10)80021-X].

———. 1991b. High-precision multicollector isotopic analysis of low levels of Nd as oxide. Chem. Geol. 94:13-22 [doi:10.1016/S0009-2541(10)80013-0].

Vance, D., and M. Thirlwall. 2002. An assessment of mass discrimination in MC-ICPMS using Nd isotopes. Chem. Geol. 185:227-240 [doi:10.1016/S0009-2541(01)00402-8].

———, A. E., Scrivner, P. Beney, M. Staubwasser, G. M. Henderson, and N. C. Slowey. 2004. The use of foraminifera as a record of the past neodymium isotope composition of seawater. Paleoceanography 19:PA2009 [doi:10.1029/2003PA000957].

Weis, D., and others. 2006. High-precision isotopic characterization of USGS reference materials by TIMS and MC-ICP-MS. Geochem. Geophys. Geosys. 7:Q08006 [doi:10.1029/2006GC001283].

White, W. M., and J. Patchett. 1984. Hf-Nd-Sr isotopes and incompatible element abundances in island arcs: implications for magma origins and crust-mantle evolution. Earth Planet. Sci. Lett. 67:167-185 [doi:10.1016/0012-821X(84)90112-2].

Wombacher, F., and M. Rehkämper. 2003. Investigation of the mass discrimination of multiple collector ICP-MS using neodymium isotopes and the generalised power law. J. Anal. At. Spectrom. 18:1371-1375 [doi:10.1039/b308403e].