



HAL
open science

Holographic laser Doppler imaging of microvascular blood flow

Caroline Magnain, Amandine Castel, Tanguy Boucneau, Manuel Simonutti, Isabelle Ferezou, Armelle Rancillac, Tania Vitalis, José-Alain Sahel, Michel Paques, Michael Atlan

► **To cite this version:**

Caroline Magnain, Amandine Castel, Tanguy Boucneau, Manuel Simonutti, Isabelle Ferezou, et al.. Holographic laser Doppler imaging of microvascular blood flow. *Journal of the Optical Society of America. A Optics, Image Science, and Vision*, 2014, 31 (12), pp.2723-2735. hal-01086781

HAL Id: hal-01086781

<https://hal.science/hal-01086781v1>

Submitted on 28 Jan 2015

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

Holographic laser Doppler imaging of microvascular blood flow

C. Magnain,¹ A. Castel,¹ T. Boucneau,¹ M. Simonutti,² I. Ferezou,³
A. Rancillac,³ T. Vitalis,³ J. A. Sahel,² M. Paques,² and M. Atlan¹

¹*Institut Langevin. Centre National de la Recherche Scientifique (CNRS) UMR 7587, Institut National de la Santé et de la Recherche Médicale (INSERM) U 979, Université Pierre et Marie Curie (UPMC), Université Paris Diderot. Ecole Supérieure de Physique et de Chimie Industrielles - 10 rue Vauquelin. 75005 Paris, France*

²*Institut de la Vision, INSERM UMR-S 968. CNRS UMR 7210. UPMC. 17 rue Moreau, 75012 Paris. France*

³*Brain Plasticity Unit, CNRS UMR 8249, ESPCI ParisTech, 10 Rue Vauquelin, 75005 Paris, France*

We report on local superficial blood flow monitoring in biological tissue from laser Doppler holographic imaging. In time-averaging recording conditions, holography acts as a narrowband bandpass filter, which, combined with a frequency-shifted reference beam, permits frequency-selective imaging in the radiofrequency range. These Doppler images are acquired with an off-axis Mach-Zehnder interferometer. Microvascular hemodynamic components mapping is performed in the cerebral cortex of the mouse and the eye fundus of the rat with near-infrared laser light without any exogenous marker. These measures are made from a basic inverse-method analysis of local first-order optical fluctuation spectra at low radiofrequencies, from 0 Hz to 100 kHz. Local quadratic velocity is derived from Doppler broadenings induced by fluid flows, with elementary diffusing wave spectroscopy formalism in backscattering configuration. We demonstrate quadratic mean velocity assessment in the 0.1-10 mm/s range *in vitro* and imaging of superficial blood perfusion with a spatial resolution of about 10 micrometers in rodent models of cortical and retinal blood flow.
OCIS codes : 090.0090, 040.2840, 170.1470, 170.3340.

I. INTRODUCTION

A. Motivations

The role of the microcirculation is increasingly being recognized in the pathophysiology of cardiovascular diseases; eg. hypertension [1–3] and diabetes [4–6]. In particular, assessing retinal blood flow can be potentially useful for understanding diabetic retinopathies [7], and study the relationships between vascularization and glaucoma [8–12]. Moreover, extensive use of optical methods is made for monitoring skin microvascular endothelial (dys)function [13–18]. Prevalent optical techniques to monitor microvascular blood flow in clinical studies are laser Doppler probes and spatial speckle contrast imaging. The former is characterized by its high temporal resolution and the latter enables wide-field imaging of superficial microvascular networks. Superficial blood flow monitoring require sensitivity in low light, high temporal resolution, high spatial resolution, and the ability to perform quantitative flow measurements. Most current techniques are limited in their spatial resolution or temporal resolution or both. Hence wide-field optical imaging techniques using laser light and sensor arrays to probe local dynamics with potentially high spatial and temporal resolution are attracting attention for the measurement of blood flow [19].

B. Relationship between local motion and optical fluctuations

The radiofrequency (RF) spectrum of dynamic light fluctuations is affected by microvascular hemodynamics and hence a subject of great interest for blood flow imaging applications. The laser Doppler technique measures

Doppler-shifts and broadenings of quasi-elastically scattered light. Depending on the detection configuration, single scattering or multiple scattering can be targeted to yield Doppler spectra. The observation and interpretation of Doppler broadening of a scattered laser light beam by a fluid in motion *in vitro* [20] and *in vivo* [21] has led to the physical modeling of laser Doppler velocimetry [22, 23]. Mapping retinal hemodynamics *in vivo* has many biomedical applications such as diagnosis of retinal microvasculature disorders [24–26]. Optical instrumentation is well adapted to non-invasive retinal blood flow imaging, because the eye fundus vascular tree is visible through the cornea and is thus potentially accessible for light imaging. The Doppler shift of a monochromatic optical radiation scattered by a moving target is the scalar product of the optical momentum transfer with the target velocity. Doppler shifts are cumulative. In the case of multiple scattering, Algebraic Doppler shifts add-up throughout each optical path, and the broadening of the backscattered radiation still carries a highly valuable information, for velocity assessment [22, 27]. Under coherent illumination, the local motion of scatterers can be probed either by the analysis of local spatial contrast, the local temporal fluctuations, or the local RF Doppler spectrum of the speckle pattern [28]. Single point laser Doppler detection schemes in heterodyne [29–32] or self-mixing [33–36] configurations are prevalent in practical detection of blood flow from wideband temporal optical fluctuations analysis, for superficial blood flow sensing. The design of depth-resolved Doppler-contrast optical coherence tomography schemes [37–40] is also an active field of research with applications in vascular imaging of the posterior segment of the eye. In this article, we limit the description of state-of-the-art detection schemes based on sensor arrays.

C. Spatial speckle contrast analysis

Spatial intensity fluctuation analysis techniques [41–45] can be referred to as time-averaged speckle contrast imaging. These techniques have been adopted over the past decades for imaging of blood flow dynamics in real-time. Their rapid adoption for physiological studies is due to the ability to quantify blood flow changes with good spatial and temporal resolution. Local motion imaging by the analysis of spatial contrast of laser speckle is a simple and robust method to achieve full-field imaging [46–48]. It is a widely used imaging technique to image blood flow in vivo, primarily due to advantages like ease of instrumentation, reproducibility [49], and low cost [50]. It relies on local speckle contrast in time-averaged recording conditions to assess local motion of light scatterers. Motion of scattering particles, like the red blood cells, causes spatial and temporal blurring of the speckle pattern. In speckle contrast imaging, the interference pattern is recorded with a camera, and the blurring of the pattern is quantified to obtain a measure of relative change in flow. Spatial contrast analysis techniques improved to assessment of quantitative relative changes in local blood perfusion with exposure-control of the recorded frames [51, 52] and consideration of the non-fluctuating light component in contrast analysis models [53]. Speckle contrast analysis schemes were used to generate images of blood flow in the rat retina in traditional configuration [54, 55], and via an endoscope [56]. Speckle contrast imaging is now being routinely applied to cerebral blood flow assessment, providing valuable vascular perfusion information in vivo [57–61]; in particular with the advent of multi-exposure speckle imaging [52, 62]. Recently, the technique was used with success to monitor cerebral blood flow during neurosurgery in the human brain [63, 64].

D. Temporal speckle contrast analysis with sensor arrays

Full-field laser Doppler imaging techniques from time-domain optical intensity fluctuations measurements on sensor arrays can enable monitoring of blood perfusion changes [65–71]. In these approaches, laser Doppler signal processing is performed off-chip, which means that large digital data transfers have to be handled from the sensor array to the processing unit of a computer. Recently, image-plane laser Doppler recordings with a high throughput complementary metal-oxide-semiconductor (CMOS) camera in conjunction with short-time discrete Fourier transform calculations by a field programmable gate array (FPGA) reportedly enabled continuous monitoring of blood perfusion in the mm/s range. Full-field flow maps of 480×480 pixels were rendered at a rate of 14 Hz, obtained from image recordings at a frame rate of 14.9 kHz [72]. However, cameras capable of frame rates of up to several kilohertz and data transfer rates still larger than the commercially available ones are a re-

quirement in this approach. Locally processing the laser Doppler signal within the sensor array can prevent large data throughputs off-chip, which is presently a technological bottleneck. To circumvent data transfer issues, fully-integrated CMOS sensor arrays with Doppler signal processing were developed [73], which have the advantage of enabling not only digital but also analog signal processing, which permits capacitive coupling to cancel DC contributions and low-pass filtering of the analog signal to process. Such sensors were demonstrated to achieve 64×64 pixel blood flow images at 1 frame per second [74–76], with a sampling bandwidth of 40 kHz. On-chip laser Doppler signal processing is one of the most promising approaches for the temporal analysis of optical fluctuations for blood flow imaging. It alleviates the technical issue of unmet data throughput required in time-resolved parallel sensing with sensor arrays.

E. Laser Doppler imaging by holography

Early experimental demonstrations of holographic ophthalmoscopy of the eye fundus in vivo [77–81] have shown the potential of this modality for retinal vessels imaging, before the advent of digital acquisition schemes with sensor arrays. A convenient way to perform optical heterodyne detection with video frame rate sensor arrays is to have recourse to holography. For instance, blood flow contrasts were imaged by holography in the cerebral cortex [82] and in the eye fundus [83] of rodents with a standard camera. Time-averaged holography with a frequency-shifted reference beam was shown to enable narrow band detection and imaging of local Doppler broadenings [84–86]. Blood flow contrasts were also successfully rendered with wideband detection of optical fluctuations in holographic configuration [87]. Laser Doppler holography enables narrowband optical heterodyne detection with the benefit of canceling laser amplitude noise, which can be efficiently filtered spatially in off-axis recording configuration [88] at the price of a reduction of the available spatial bandwidth. Hence holographic schemes have a major advantage in comparison with direct image recording schemes, especially for narrowband Doppler detection with video-rate cameras.

F. Presented results

We report on quantitative fluid flow assessment in vitro and in vivo from holographic interferometry, which enables wide-field imaging of a Doppler-shifted radiation with an array detector. Time-averaged holography acts as a narrowband bandpass filter, described in section II, which permits frequency-selective Doppler imaging in the radiofrequency range, in combination with a frequency-shifted reference beam. These Doppler images are acquired with an off-axis Mach-Zehnder interferometer, in reflection configuration [82, 83]. The signal obtained with narrowband, frequency-tunable holo-

graphic interferometry is the first-order power spectrum of the optical field scattered by the object, recorded in heterodyne configuration [89]. Multiply-scattered light yields Doppler spectra from which the directional information is lost because of wave vector randomization during scattering. Nevertheless, it has the advantage of providing spatially-resolved wide-field hemodynamic contrast. Our strategy is to use an inverse-problem approach to derive velocity maps of blood flow from optically-acquired Doppler maps; we make use of elementary results from the diffusing-wave spectroscopy formalism [90] in backscattering configuration in order to retrieve local root mean square velocities [91]. The measured heterodyne signal described in section II is modeled with the diffusing-wave spectroscopy formalism for flowing and diffusing scatterers in sections IV, V, and VI. In vitro flow measurements in a calibrated sample are reported in section IV. In vivo microvascular blood flow mapping in the cerebral cortex of the mouse and the eye fundus of the rat are reported in section V, and section VI, respectively.

II. NARROWBAND HETERODYNE DETECTION OF LIGHT BY FREQUENCY-TUNABLE HOLOGRAPHY

A. Optical fields

In this section, we use a scalar, deterministic representation of light to highlight the temporal filtering features of time-averaged heterodyne holography. The single-frequency laser light wave can be modeled by a complex number E , which represents the oscillating electric field. This field oscillates at the angular frequency ω_L . Polarization properties are neglected in this representation. The amplitude and phase of the field are the absolute value and angle of E . The optical field illuminating the object under investigation is of the form $E = E_0 \exp(i\omega_L t)$, where E_0 is a complex constant and i is the imaginary unit. The probe optical field, backscattered by the preparation under investigation is noted

$$E(t) = \mathcal{E}(t) \exp(i\omega_L t) \quad (1)$$

where the envelope $\mathcal{E}(t)$ carries the RF temporal fluctuations of the field. It is decomposed in discrete frequency components

$$\mathcal{E}(t) = \sum_n \tilde{\mathcal{E}}(\omega_n) \exp(i\omega_n t) \quad (2)$$

With these notations, the discrete first-order RF power spectrum of the object light is

$$s_1(\omega_n) = |\tilde{\mathcal{E}}(\omega_n)|^2 \quad (3)$$

The optical local oscillator (LO) field is a monochromatic wave, coherent with the illumination beam, frequency-shifted by $\Delta\omega$

$$E_{LO}(t) = \tilde{\mathcal{E}}_{LO} \exp(i\Delta\omega t) \exp(i\omega_L t) \quad (4)$$

The irradiance of the light wave, in terms of its electric field E , is $|E|^2 c \epsilon_0 / 2$, where ϵ_0 is the permittivity and c is the speed of light in vacuum, with a relative permittivity and a magnetic permeability of 1. At time t , the optical power (irradiance integrated over the surface d^2 of a pixel) integrated during the exposure time τ_E by the square-law sensor array yields the recorded interferogram

$$I(t) = \frac{1}{2} \epsilon_0 c d^2 \eta \int_{-\tau_E/2}^{\tau_E/2} |E(t+\tau) + E_{LO}(t+\tau)|^2 d\tau \quad (5)$$

where η is the quantum efficiency of a pixel.

B. Bandpass filtering properties of time-averaged holograms

The object field E beats against the LO field E_{LO} and yields the following cross-terms contribution in an off-axis region of the spatial frequency spectrum of the interferograms [88]

$$H(t) = \frac{1}{2} \epsilon_0 c d^2 \eta \int_{-\tau_E/2}^{\tau_E/2} E(t+\tau) E_{LO}^*(t+\tau) d\tau \quad (6)$$

To calculate an image, the recorded interferograms have to be demodulated spatially. To this end, numerical image rendering in the object plane is performed with a discrete Fresnel transform involving one fast Fourier transform of the recorded interferograms [92–96]. Once this operation is performed, Eq. 6 represents quantities back-propagated to the object plane. The integration of the interference pattern over the exposure time acts as a temporal frequency filter. In the Fourier domain, it corresponds to a low bandpass filter centered at $\Delta\omega$ and whose bandwidth is $1/\tau_E$.

$$H(t) = C \sum_n \tilde{\mathcal{E}}(\omega_n) \tilde{\mathcal{E}}_{LO}^* \frac{\sin(\Delta\omega_n \tau_E / 2)}{\Delta\omega_n \tau_E} \exp(i\Delta\omega_n t) \quad (7)$$

where $C = \epsilon_0 c d^2 \eta \tau_E$ and $\Delta\omega_n = \omega_n - \Delta\omega$ is the angular frequency detuning between the local oscillator and the n -th spectral component. Two-phase temporal signal demodulation is performed by calculating the squared magnitude of the difference of two consecutive holograms in order to filter-off stray LO intensity fluctuations from the interferometric cross-terms. The resulting signal is

$$|(H(t+2\pi/\omega_S) - H(t))|^2 = S^2(t) \quad (8)$$

Optical phase fluctuations of the probe field E are considered a stationary process at the time scale of the acquisition of a few frames, such that the signal temporal dependency can be dropped : $S^2(t) \equiv S^2$. Using Eq. 7, Eq. 8 can be rewritten as

$$S^2 = \left| 2C \tilde{\mathcal{E}}_{LO}^* \sum_n \tilde{\mathcal{E}}(\omega_n) b(\omega_n - \Delta\omega) \right|^2 \quad (9)$$

The right member of equation 9 is proportional to the squared magnitude of the discrete convolution of $\tilde{\mathcal{E}}$ with the response function b , assessed at the frequency $\Delta\omega$. Besides a constant phase factor, this response is [97]

$$b(\omega) = \sin(\pi\omega/\omega_S) \sin(\omega\tau_E/2) / (\omega\tau_E) \quad (10)$$

The response for the probe field's irradiance $B(\omega) = |b(\omega)|^2$ is a bandpass filter of width $\sim 1/\tau_E$. Its line shape is reported in Fig. 4 (dotted line) in typical recording conditions, for $\tau_E = 30$ ms and $\omega_S/(2\pi) = 12$ Hz.

C. First-order spectrum assessment

If $1/\tau_E$ is much smaller than the typical cutoff frequency of the RF Doppler broadening of the probe beam, we can consider that only the component of the discrete spectral decomposition of the object field at $\Delta\omega$ is measured

$$S^2 \propto |\tilde{\mathcal{E}}(\Delta\omega)|^2 |\tilde{\mathcal{E}}_{LO}|^2 \quad (11)$$

The average number of probe photons reaching the pixels of size d during the measurement time τ_E and collected by the filter B tuned on $\Delta\omega$, is noted $n(\Delta\omega)$; it satisfies $n(\Delta\omega)\hbar\omega_L \approx |\tilde{\mathcal{E}}(\Delta\omega)|^2 \epsilon_0 c d^2 \tau_E / 2$. The average number of photons of the local oscillator satisfies $n_{LO}\hbar\omega_L \approx |\tilde{\mathcal{E}}_{LO}|^2 \epsilon_0 c d^2 \tau_E / 2$. In the off-axis region of the hologram, where the object-LO cross-term is separated from other interferometric contributions, the shot-noise is the dominating noise, even in low light [98]. The shot-noise N , in high heterodyne gain regime, scales up linearly with η and the average number of photons in the LO channel [98–103]. It satisfies

$$N^2 \propto |\tilde{\mathcal{E}}_{LO}|^2 \quad (12)$$

The domain where this quantity is averaged in practice is shown in Fig. 2. From Eq. 11, Eq. 12, and Eq. 3, a robust narrowband measurement of the RF power spectrum can be performed by forming the following ratio, which does not depend on the LO power

$$S^2/N^2 \propto s_1(\Delta\omega) \quad (13)$$

Where $\Delta\omega$ is the detuning angular frequency of the optical local oscillator. In the case of two-phase demodulation [88, 97] (Eq. 8), the detection frequency is equal to the detuning frequency of the local oscillator.

III. EXPERIMENTAL SETUP

The experimental laser Doppler imaging scheme used for this study is sketched in Fig. 1. It is based on a heterodyne holographic arrangement previously reported [82, 83, 85, 88], which consists of a modified Mach-Zehnder optical interferometer designed for off-axis and frequency-shifting holographic imaging in time-averaged

recording conditions [104]. In vitro and in vivo flows are investigated. The in vitro preparation, reported in section IV, consists of a transparent tube of 1 mm diameter in which an emulsion is injected with a known average velocity. In vivo flow studies of cerebral and retinal blood flow are reported in section V, and section VI. In all these experiments, a continuous laser diode (Mitsubishi ML120G21) emits a monochromatic radiation of wavelength $\lambda = 785$ nm with an average power of 80 mW. The detection consists of wide-field illumination and collection of the cross-polarized backscattered light from the brain along the same optical axis. It is achieved with a polarizing beam splitter, in order to select photons which have undergone at least a few scattering events [105]. This approach is chosen to increase the relative weight of




FIG. 1. Sketch of the optical arrangement : Mach-Zehnder holographic interferometer. The main laser beam of a near infrared laser diode is split into two channels. In the object channel, the optical field E is backscattered by the preparation. In vitro measurements were performed in a transparent tube wherein a controlled flow is induced. For in vivo measurements, the preparation was a cortical window or the eye fundus of rodents. Local motion induces RF broadenings of the optical radiation. These broadenings are quantitatively analysed by narrowband optical detection, at the detuning frequency of the reference channel. In the reference channel, the optical field E_{LO} is frequency-shifted by two acousto-optic modulators from which alternate diffraction orders (± 1) are selected, yielding an optical LO the form of Eq. (4). The sensor array of the camera records the interference pattern I of both optical fields beating against each other, in time-averaging conditions, at a frame rate ω_S and exposure time τ_E , which define the narrowband filter of the detection (Eq. (10)). Images of the preparation are calculated with a standard holographic rendering algorithm involving a numerical Fresnel transform.

multiply scattered Doppler-shifted photons with respect to photons scattered once. The incident light beam is expanded to form a plane wave and diaphragged by a $\sim 5 \text{ mm} \times 5 \text{ mm}$ pupil. Its polarization angle is tuned with a half-wave plate to set the illumination power to $\sim 1 \text{ mW}$. In the reference arm, an attenuator, a half-wave plate, and a beam expander are used to control the LO beam power, polarization angle, and to ensure a flat illumination of the detector. Two acousto-optic modulators (AA Opto Electronic), driven with phase-locked signals at ω_C and $\omega_C + \Delta\omega$, are used to shift the optical frequency of the laser beam from ω_L to $\omega_L + \Delta\omega$, to form a frequency-shifted optical local oscillator. The carrier frequency $\omega_C/(2\pi)$ is set at the peak response of the acousto-optic modulators, at 80 MHz. The backscattered field E is mixed with the LO field E_{LO} with a non-polarizing beam splitter cube. The interference pattern $I(t)$ is measured at time t by a charge-coupled device array detector (PCO pixelfly QE camera, 1394×1024 square pixels of $d = 6.7 \mu\text{m}$, frame rate $\omega_S/(2\pi) = 12 \text{ Hz}$, exposure time $\tau_E \simeq 30 \text{ ms}$, dynamic range 12 bit), set at a distance $\Delta z \simeq 69 \text{ cm}$ from the object plane. A small angular tilt $\sim 1^\circ$ ensures off-axis recording conditions. The temporal modulation of these fringes is controlled by the optical frequency shift $\Delta\omega$ between the reference and the illumination beam. In all the reported experiments, sets of 32 consecutive interferograms $\{I_1, \dots, I_{32}\}$ are recorded for each detuning frequency $\Delta\omega$, for signal averaging purposes. The measurement time of one frequency component is $32/12 = 2.7 \text{ s}$.

IV. IN VITRO FLUID FLOW IMAGING

From now on, throughout sections IV, V, VI, the optical probe field $E(t)$ is described by a stochastic, random variable, as opposed to the model used to derive the temporal filter of the detection in section II. The power spectrum is related to the optical field autocorrelation degree $g_1(t)$. A quantitative derivation of local decay rates of this function can be made from diffusing-wave spectroscopy formalism in backscattering configuration for a semi-infinite medium [90, 106–108]

$$g_1(t) \sim \exp\left(-\gamma\sqrt{k^2\langle\Delta r^2(t)\rangle}\right) \quad (14)$$

where $\langle\Delta r^2(t)\rangle$ is the mean-square displacement of local light scattering particles, γ is a fitting parameter, and k is the optical wave number. Under the assumption that the optical field undergoes a stationary random phase fluctuations, according to the Wiener–Khinchin theorem [109–112], the first-order Doppler spectrum s_1 and the field autocorrelation function g_1 form a Fourier pair

$$s_1(\omega) = \int_{-\infty}^{+\infty} g_1(\tau) \exp(-i\omega\tau) d\tau \quad (15)$$

Detuning the LO frequency of the reference beam of the Mach-Zehnder interferometer by $\Delta\omega$ permits robust nar-




FIG. 2. (a) Squared amplitude hologram of the tube used for in vitro fluid flow calibration measurements. The image of the tube is in the upper left corner. The noise component N^2 (Eq. 12) is averaged spatially in the upper right corner. The twin-image appears in the lower right corner. Self-beating interferometric contributions are gathered in the center of the hologram. (b) zoom in the region of the object. The signal component S^2 (Eq. 8) is spatially-averaged in the intralipid flow region to form Doppler line shapes reported in Fig. 4.

rowband heterodyne imaging of a component of the first-order RF spectrum of the probe field, by forming the quantity $S^2/N^2 \propto s_1(\Delta\omega)$ (Eq.13). In practice, for a given detuning frequency $\Delta\omega$, 32 consecutive interferograms are recorded $\{I_1, \dots, I_{32}\}$. Off-axis hologram rendering in the image plane is performed with discrete Fresnel transformation [92–96]. The difference of consecutive off-axis holograms $H_{n+1} - H_n$ is then calculated in order to form the temporally-demodulated signal S^2 (Eq. 8). This signal is temporally averaged over 32 raw frames. The noise component N^2 (Eq. 12) is temporally averaged in the same manner. Moreover, the quantity N^2 is spatially averaged in the highlighted region of Fig. 2(a), and for the measurement of the Doppler lineshapes, the quantity S^2 is spatially-averaged in the intralipid flow region highlighted in Fig. 2(b). Representative Doppler maps at several detuning frequencies (0 Hz, 200 Hz, 1 kHz, and 2 kHz) for a flow speed of 0.1 mm/s are reported in Fig. 3.

A. Brownian motion of scatterers

In the case of a suspension of diffusing particles for which the mean-square displacement is $\langle\Delta r^2(t)\rangle = 6Dt$, the optical field autocorrelation function can be written as [113]

$$g_1(t) \sim \exp\left(-\gamma\sqrt{6t/\tau_0}\right) \quad (16)$$

where $\tau_0 = (k^2D)^{-1}$, the parameter γ depends on the configuration and D is the spatial diffusion coefficient of a light-scattering particle (diffusivity). For particles in brownian motion, the theoretical first-order optical power spectral density (reported in Fig. 4, for null flow $v_F = 0$) is calculated numerically from Eq. 16 and Eq. 15.




FIG. 3. Doppler images of a flowing intralipid suspension at velocity 0.1 mm/s in a transparent tube at four detuning frequencies $\Delta\omega/(2\pi)$: 0 Hz (a), 200 Hz (b), 1 kHz (c), 2 kHz (d).

B. Convective motion of scatterers

Particles in deterministic (convective) motion with a quadratic mean velocity v_F undergo a local mean-square displacement

$$\langle \Delta r^2(t) \rangle = v_F^2 t^2 \quad (17)$$

The temporal autocorrelation function of the optical field backscattered from a suspension of flowing particles takes the form of an exponential decay; it scales as [114]

$$g_1(t) \sim \exp(-|t|/\tau) \quad (18)$$

where $\tau = (\gamma k v_F)^{-1}$. For these particles in deterministic motion, the normalized first-order power spectral density of optical fluctuations is

$$s_1(\omega) \sim \frac{1}{1 + \omega^2/\omega_F^2} \quad (19)$$

The half-width at half-maximum ω_F of this Lorentzian line shape is proportional to the local quadratic mean velocity v_F

$$\omega_F = \gamma k v_F \quad (20)$$

This quantity is the decay rate of g_1 (Eq. 18). This line shape is the model against which the experimental Doppler measures are fitted, in vitro and in vivo, reported in Fig. 4 (for $v_F \neq 0$) and Fig. 7.




FIG. 4. Normalized spectral lines (a.u., symbols) versus detuning frequency, in Hz. The average perfusion velocity in the tube ranges from 0 mm/s to 7 mm/s (symbols). The measured lines are fitted against Eq. 22 for brownian (Eq. 16) and convective (Eq. 19) motion of the scatterers, plotted as continuous black lines. The apparatus lineshape Eq. 10 is plotted as a dotted line.

C. Measurement of Doppler line shapes

For the detected light composed of dynamic (subscript D) and static (subscript S) components $\mathcal{E} = \mathcal{E}_D + \mathcal{E}_S$ it was shown [53, 89] that the autocorrelation function of the field will take the following form

$$g(t) = \rho |g_1(t)| + (1 - \rho) \quad (21)$$

where $\rho = |\mathcal{E}_D|^2 / (|\mathcal{E}_D|^2 + |\mathcal{E}_S|^2)$ characterizes the relative weight of the dynamic part of the detected light intensity, referred to as pedestal wave [115, 116]. In that case, the holographic measurement of the local RF spectrum of light, defined in Eq. 13, should exhibit the presence of the apparatus function B (Eq. 10) in the non-shifted light component. At a given reconstructed pixel, at a detuning $\Delta\omega$ of the local oscillator, measurements can be described by the following expression

$$S^2/N^2 \propto \rho s_1(\Delta\omega) + (1 - \rho)B(\Delta\omega) \quad (22)$$

D. Experimental measurements of in vitro flows

The setup for in vitro flow assessment consists of a tube of 1 mm diameter in which a 1 part in 10 dilute solution of 10% Intralipid in water is injected with a known average velocity by a calibrated syringe pump. The range of velocities studied with this set up varies from 0 mm.s⁻¹ to 10 mm.s⁻¹. The experimental Doppler broadened lines measured for known velocities v_F were fitted against expression 22. The results for RF spectra obtained for 3 velocities (symbols) are shown on Fig. 4 (black lines). The




FIG. 5. Doppler broadening ω_F/k (mm.s^{-1}) versus average perfusion quadratic velocity v_F (mm.s^{-1}). In vitro measurements of Doppler linewidth are in agreement with the model of convective motion (Eq. 19) within a velocity range of 0.1 mm/s to 10 mm/s.

variation of the half width of the Doppler line ω_F with the flow velocity v_F is investigated. We plotted ω_F/k against v_F in Fig. 5 in order to verify experimentally whether the relation $\omega_F/k = \gamma v_F$ holds. This plot shows that the Doppler width scales linearly with the flow velocity for flows in the 100 $\mu\text{m/s}$ to 10 mm/s range, which is in agreement with previously reported results, in vitro [117–119]. The coefficient γ evaluated from this procedure is

$$\gamma = 1.42 \quad (23)$$

which is consistent with experimental diffusing wave spectroscopy results in semi-infinite scattering media [90]. This value of γ is used to calculate flow maps in section V and section VI. For flow velocities below 100 microns per second, experimentally measured Doppler widths are systematically larger than $k\gamma v_F$, which can be explained by the fact that the dominating motion of the scatterers is diffusive. This is confirmed by fitting the measured spectrum of light scattered from particles undergoing only brownian motion with the discrete Fourier transform of Eq. 16, which yields the Doppler line for null flow velocity reported in Fig. 4. A spatial diffusivity $D = k_B T / (6\pi\mu r) = 1 / (k^2\tau_0) \sim 8 \times 10^{-13} \text{ m}^2.\text{s}^{-1}$ is found for $\gamma = 1.42$. It corresponds to a homogenous suspension of oil droplets of radius $r = 311 \text{ nm}$, in water of dynamic viscosity $\mu = 0.89 \times 10^{-3} \text{ kg.m}^{-1}.\text{s}^{-1}$. The effect of residual motion on Doppler flow measurements is referred to as the biological zero problem [120, 121].

V. IMAGING OF MICROVASCULAR CEREBRAL BLOOD FLOW

Cerebral blood flow imaging in the mouse brain was conducted in strict compliance with approved institutional protocols and in accordance with the

provisions for animal care and use described in the European Communities Council directive of 24 November 1986 (86-16-09/EEC). Two 26g C57BL6J mice were anesthetized with urethane (1.75 mg/g). Paw withdrawal, whisker movement, and eyeblink reflexes were suppressed. The head of the mouse was fixed by using a stereotaxic frame (Stoelting). The skin overlying the right cerebral hemisphere was removed and the bone gently cleaned. A $3 \times 3 \text{ mm}$ craniotomy was made above the primary somatosensory cortex. Extreme care was taken at all times not to damage the cerebral cortex, especially during the removal of the dura. Physiological Ringers solution containing (in mM): 135 NaCl, 5 KCl, 5 HEPES, 1.8 CaCl₂, and 1 MgCl₂, was used during the surgery to prevent the exposed cortical area from drying out. At the end of this surgical procedure, the cortical surface was covered with agarose (1 % in Ringers), and a coverslip was placed on top.

Doppler maps were acquired for logarithmically-spaced LO detunings $\Delta\omega/(2\pi)$ from 1 Hz to 100 kHz, in order to observe the optical fluctuations due to the Doppler effect induced by blood flow. Fig. 6 shows representative Doppler images of the quantity S^2/N^2 (Eq. 13) in the cerebral cortex of a mouse, at five frequency shifts, 9 Hz (a), 167 Hz (b), 1.7 kHz (c), 5.5 kHz (d), 17.5 kHz (e). The speckle was attenuated by averaging Doppler maps over three consecutive frames. The comprehensive data obtained from the frequency sweep is reported in Media 1. At low frequency (a), the background (parenchyma) exhibits a higher signal than the vessels. At higher frequency, the contrast is reversed and vessels exhibit a higher Doppler signal than the background. As the detuning frequency increases (d, e), only the main vessels are revealed. The inhomogeneity of the illumination can also be observed in those figures. First-order RF spectral line shapes, normalized by their value at low frequency are reported in Fig. 7. They are measured and averaged over four regions of interest, labeled from 1 to 4, ranked by increasing blood flow, according to the image shown in the insert. They show the features described herein, i.e. narrow broadenings for slow flows and broader line shapes for larger flows.

We assume that bioflows under investigation lead to convective scatterers' motions much larger than the ones due to diffusive processes, so that the formalism of section IV B is valid. If the local mean square displacement of an elementary scattering volume in the tissue scales up quadratically with time t , relationship 17 holds and we can derive a quadratic mean flow map from a fitting procedure of the measured signal S^2/N^2 against Eq. 22. The temporal correlation function of the backscattered optical field $g_1(t)$ is of the form of Eq. 18, and the first-order fluctuation spectrum has a Lorentzian lineshape, described by Eq. 19. The first-order power spectral density of the fluctuating field backscattered by the cortex agrees well with a Lorentzian line shape Eq. 19 whose half-width at half maximum is $\omega_F = \gamma k v_F$. The param-




FIG. 6. Doppler images of the mouse cerebral cortex at different frequency shifts 9 Hz (a), 167 Hz (b), 1.7 kHz (c), 5.5 kHz (d), 17.5 kHz (e). The frequency sweep is reported in Media 1. Image of a USAF resolution target (f). Arbitrary logarithmic gray scale. Scale bar : 1 mm.

ter γ was assessed from the calibration procedure in vitro (Eq. 23), $k = 2n\pi/\lambda$ and $n = 1.39$ is the mean refractive index of the blood at our studying wavelength [122], which is close to the refractive index of the brain, evaluated at 1.35 for $\lambda=1.1 \mu\text{m}$ [123]. A holographic image of a resolution target, acquired in the same experimental configuration, is reported in Fig. 6 (f), from which we can observe a lateral resolution limit of about 10 microns, which is compatible with the observation of the smallest vessels. Four spectra from the region of interest are reported in Fig. 7. They are normalized by their respective maximal value at low frequency. As we can see,




FIG. 7. Normalized first-order power spectra averaged in the regions of interest (ROI) 1 to 4 (dots correspond to measurements, continuous lines are the result of fitting measured values against Eq. 19). Inset: Doppler map displaying the 4 regions of interest.

the higher the blood flow velocity, the broader the spectra. Those spectra are fitted by Lorentzian lines from Eq. 19 with a robust nonlinear least squares algorithm (MathWorks Matlab Curve Fitting Toolbox) and plotted as solid lines on the same graph. These theoretical lines are in good agreement with the measurements. A map of the quadratic mean blood flow velocity v_F can then be derived by fitting each pixel's Doppler line shape




FIG. 8. Map of the local quadratic mean blood flow velocity v_F derived from the Doppler maps reported in Fig. 6.




FIG. 9. Retinal images at four frequency shifts, 10 Hz (a), 513 Hz (b), 3.0 kHz (c), 6.1 kHz (d). The frequency sweep is reported in Media 2. Scale bar : 1 mm.

with the Lorentzian line of Eq. 19 to assess local values of v_F . The result is reported in Fig. 8. The inhomogeneity of the lighting seen in Fig. 6 is no longer observed. The resulting velocity map is thus independent on the local illumination level. The blood vessels are well resolved, even the smallest ones.

VI. IMAGING OF MICROVASCULAR RETINAL BLOOD FLOW

Retinal blood flow imaging was conducted in strict compliance with approved institutional protocols. Three adult rats were used for the preparations. Anesthesia was induced by intraperitoneal injection of 100 mg/kg ketamine and 25 mg/kg xylazine (both from Sigma-Aldrich). Topical tropicamide (CibaVision) was administered for pupil dilation. Each rat was placed on its side in a clay bed, their right eyes under the illumination beam. The head was supported so that the iris was perpendicular to the illumination axis. After administration of topical oxybuprocaine (CibaVision), a coverslip was applied on a ring surrounding the globe in order to compensate for the cornea curvature [83]. Methylcellulose (Goniosol) was applied as a contact medium.

Doppler maps were acquired for logarithmically-spaced LO detunings $\Delta\omega/(2\pi)$ from 10 Hz to 63 kHz, in order to

observe the optical fluctuations due to the Doppler effect induced by blood flow in the eye fundus. Fig. 9 shows representative Doppler images of the quantity S^2/N^2 (Eq. 13) in the eye fundus of a mouse, at four frequency shifts : 10 Hz (a), 513 Hz (b), 3.0 kHz (c), 6.1 kHz (d). The speckle was attenuated by averaging Doppler maps over three consecutive frames. The comprehensive data obtained from the frequency sweep is reported in Media 2. The contrast reversal from low to high frequency is similar to the observations in the brain, reported in Fig. 6. First-order RF spectral line shapes, normalized by their value at low frequency are reported in Fig. 10. They are measured and averaged over five regions of interest, labeled from 1 to 5.

A. Velocity composition

The global motion of the eye has to be taken into account while considering the velocity of the blood flow. Let \mathbf{v}_G be the global in-plane velocity of the retina that defines eye movement [124, 125]. The scattering particles flowing in a vessel with an in-plane velocity \mathbf{v}_F have a total instant velocity

$$\mathbf{v} = \mathbf{v}_F + \mathbf{v}_G \quad (24)$$

The local velocity outside a vessel is dominated by the global physiological motion at the quadratic mean velocity $v_G = \sqrt{\langle \mathbf{v}_G^2 \rangle}$, where $\langle \rangle$ denotes averaging over the measurement time. We assume that the global physiological motion of the eye, induces a frequency broadening $\omega_G/(2\pi)$ related to its quadratic mean velocity $v_G = \omega_G/(\gamma k)$, where $k = 2\pi n/\lambda$ with n the average refractive index of the eye fundus [126]. The local retinal blood flow induces a frequency broadening $\omega_F/(2\pi)$ related to the quadratic mean velocity $v_F = \sqrt{\langle \mathbf{v}_F^2 \rangle}$. In the neighborhood of a vessel, the local mean square displacement at time t of an elementary scattering volume in the tissue is supposed to be, on average, over many realizations

$$\langle \Delta r^2(t) \rangle \approx (v_F^2 + v_G^2) t^2 \quad (25)$$

under the hypothesis that the averaged cross-term $\langle 2\mathbf{v}_F \cdot \mathbf{v}_G \rangle$ is zero, because the blood flow direction and the global motion direction are independent, and v_G is supposed to be a zero-mean random variable. Under these assumptions, the temporal correlation function of the optical field backscattered from the retina is the sum of two exponential decays

$$g_1(t) \sim (1 - \rho) \exp(-t/\tau_G) + \rho \exp(-t/\tau) \quad (26)$$

where ρ is the weight of light component frequency-shifted by the blood flow and $1 - \rho$ is the weight of light affected by the global motion of the eye only. The decay times satisfy the relationships $\tau_G = 1/(\gamma k v_G)$, $\tau = 1/\sqrt{\tau_F^{-2} + \tau_G^{-2}}$ and $\tau_F = 1/(\gamma k v_F)$. Eq. 26 yields




FIG. 10. Normalized first-order power spectra averaged in the regions of interest labeled from 1 to 5 (dots correspond to measurements, continuous lines are the result of the fitting procedure). Insert : Doppler map showing the 5 regions where spectra are measured.

a first-order power spectral density of the form

$$s_1(\omega) \sim \frac{1 - \rho}{1 + \omega^2/\omega_G^2} + \frac{\rho}{1 + \omega^2/(\omega_G^2 + \omega_F^2)} \quad (27)$$

where $\omega_G = 1/\tau_G$ and $\omega_F = 1/\tau_F$, from which the quadratic mean blood flow velocity v_F can be derived with Eq. 20.

B. Derivation of flow maps

Experimental spectral lines reported in Fig. 10 were averaged over several regions of interest, labeled from 1 to 5, and were fitted by a least squares fitting algorithm with Eq. 27, with the robust nonlinear technique of MathWorks' Matlab Curve Fitting Toolbox. An estimation of the average Doppler broadening ω_F at each point of the image is derived from the fitting parameters of Eq. 27. Once ω_F is assessed, an estimation of the local quadratic mean blood flow velocity v_F can be derived, besides a factor γ , via Eq. 20. The characteristic angular frequency ω_G of the global physiological movements can be determined from a region showing no perfusion. In this case, only one Lorentzian line is needed to describe the first order optical power spectral density. A region is selected, labeled 1 in the insert of Fig. 10. The Doppler line averaged in region 1 is shown in Fig. 10. It leads to a global background Doppler broadening value of $\omega_G/(2\pi) \sim 500$ Hz, and to a corresponding quadratic mean velocity of $v_G \sim 0.2$ mm/s, for a value of γ calibrated in vitro (Eq. 23). This velocity is the threshold value above which local flow quadratic mean velocities for retinal imaging in the reported experimental conditions




FIG. 11. Local quadratic mean velocity map in the eye fundus. Units : mm.s^{-1} . This map was derived from Eq. 20 and a fitting procedure of the Doppler maps against Eq. 27 for $\rho > 35\%$.

can be assessed, for eye fundus imaging of anesthetized rats. The local quadratic mean blood flow velocity map derived from Eq. 27 is shown in Fig. 11. Some artifacts can be seen in the surrounding medium while no vessels are seen. Therefore, the local relative weight of the flow component can be used to wash-out fitting artifacts and enhance local flow maps. The quantity ρ can be assessed from the fitting procedure. For the rms velocity map shown on Fig. 11, only the velocity corresponding to a weight greater than 35% is taken into account. Below the threshold value of 35%, the fitting procedure has the tendency to derive aberrant velocity values from noise.

VII. CONCLUSIONS

We demonstrated that holographic laser Doppler imaging has the potential to enable quantitative assessment of hemodynamic parameters. Cerebral and retinal blood flow was mapped in the superficial microvasculature of rodents. Frequency-scanned narrowband detection of Doppler components with a spectral resolution of a few Hertz was performed; Doppler spectra at radiofrequencies up to 100 kHz for a radiation wavelength of 785 nm revealed contrasts of microvascular blood flow. We derived quantitative quadratic mean blood flow velocity maps by using the first-order power spectrum of optical fluctuations from a basic inverse-problem model involving the diffusing-wave spectroscopy formalism. In vitro validation of this method allowed steady-state rms assessment of fluid flow from $100 \mu\text{m.s}^{-1}$ to 10mm.s^{-1} .

The low speed limit being due either to the dependence of the signal on dynamic background flows or thermal motion. The lateral spatial resolution of about 10 microns is compatible with the visualization of the smallest superficial vessels and arteries, but because of random scattering of light in tissue, flow direction is lost, and no depth sectioning is demonstrated here, in contrast to optical Doppler tomography [127]. Furthermore, strong hypotheses are made in the inverse problem formulation, which include multiple random light scattering in a semi-infinite medium, homogenous optical index of refraction, steady-state flows, uncorrelated velocities, and the presence of a fitting parameter.

Optical techniques provide suitable non-contact ways of obtaining superficial microvascular blood flow images at high resolution. They are good candidates for the development of robust quantitative, non-invasive, non-ionizing motion screening tools. The key benefits of holography with respect to the state-of-the-art optical schemes for blood flow imaging is its propensity to reveal Doppler contrasts of microvascular blood flow in low-light, which is a strong competitive advantage for retinal monitoring, and quantitative fluctuation spectra, from which local dynamical properties can be assessed. Fur-

thermore, holography is suited to the design of robust microrheological imaging tools without any contrast agent, that could be used for clinical exploration of retinal blood flow. The major drawback of time-averaged holographic detection of blood flow is poor temporal resolution, due to sequential frequency-scanning of the Doppler spectrum. In particular, pulsatile flow could not be assessed. Potential ways of circumventing this issue may be to limit the acquisition to one Doppler component with a faster camera, or to have recourse to high-speed Fourier-transform Doppler imaging [128, 129] and/or logarithmic frequency chirps of the detection frequency. Also, local [72] or even on-chip [76, 130] processing of the optical measurements with sensor arrays can be investigated.

ACKNOWLEDGEMENTS

We gratefully acknowledge support from Agence Nationale de la Recherche (ANR-09-JCJC-0113, ANR-11-EMMA-046), Fondation Pierre-Gilles de Gennes (FPGG014), région Ile-de-France (C’Nano, AIMA), the "investments for the future" program (LabEx WIFI: ANR-10-LABX-24, ANR-10-IDEX-0001-02 PSL*), and European Research Council (ERC Synergy HELMHOLTZ).

-
- [1] KM Gooding, JE Tooke, H von Lany, M Mitra, R Ling, CI Ball, D Mawson, K Skinner, and AC Shore. Capillary pressure may predict preclinical changes in the eye. *Diabetologia*, 53(9):2029–2035, 2010.
 - [2] Harry AJ Struijker-Boudier, Bart FJ Heijnen, Yan-Ping Liu, and Jan A Staessen. Phenotyping the microcirculation. *Hypertension*, 60(2):523–527, 2012.
 - [3] H.A.J. Struijker-Boudier and Bart F. J. Heijnen. *The Microcirculation and Hypertension*. Springer Milan, 2012.
 - [4] Marc Rendell, Tom Bergman, Greg O’Donnell, ED Drobny, John Borgos, and Robert F Bonner. Microvascular blood flow, volume, and velocity measured by laser doppler techniques in iddm. *Diabetes*, 38(7):819–824, 1989.
 - [5] O Schlager, A Hammer, A Willfort-Ehringer, M Fritsch, B Rami-Merhar, E Schober, K Nagl, A Giurgea, C Margeta, S Zehetmayer, et al. Microvascular autoregulation in children and adolescents with type 1 diabetes mellitus. *Diabetologia*, 55(6):1633–1640, 2012.
 - [6] PRJ Vas, AQ Green, and G Rayman. Small fibre dysfunction, microvascular complications and glycaemic control in type 1 diabetes: a case-control study. *Diabetologia*, 55(3):795–800, 2012.
 - [7] JE Grunwald, AJ Brucker, SE Grunwald, and CE Riva. Retinal hemodynamics in proliferative diabetic retinopathy. a laser doppler velocimetry study. *Invest. Ophthalmol. Vis. Sci.*, 34(1):66–71, 1993.
 - [8] Nobuyuki Koseki, Makoto Araie, Atsuo Tomidokoro, Miyuki Nagahara, Tomoyuki Hasegawa, Yasuhiro Tamaki, and Seiichiro Yamamoto. A placebo-controlled 3-year study of a calcium blocker on visual field and ocular circulation in glaucoma with low-normal pressure. *Ophthalmology*, 115(11):2049–2057, 2008.
 - [9] Subha T Venkataraman, John G Flanagan, and Chris Hudson. Vascular reactivity of optic nerve head and retinal blood vessels in glaucoma—a review. *Microcirculation*, 17(7):568–581, 2010.
 - [10] Tetsuya Sugiyama, Maho Shibata, Shota Kojima, and Tsunehiko Ikeda. Optic nerve head blood flow in glaucoma. *The Mystery of Glaucoma, T. Kubena, ed. (InTech, 2011)*, pages 207–218, 2011.
 - [11] Doreen Schmidl, Agnes Boltz, Semira Kaya, Rene Werkmeister, Nikolaus Dragostinoff, Michael Lasta, Elzbieta Polska, Gerhard Garhöfer, and Leopold Schmetterer. Comparison of choroidal and optic nerve head blood flow regulation during changes in ocular perfusion pressure. *Investigative Ophthalmology & Visual Science*, 53(8):4337–4346, 2012.
 - [12] John C Hwang, Ranjith Konduru, Xinbo Zhang, Ou Tan, Brian A Francis, Rohit Varma, Mitra Sehi, David S Greenfield, Srinivas R Sadda, and David Huang. Relationship among visual field, blood flow, and neural structure measurements in glaucoma. *Investigative Ophthalmology & Visual Science*, 53(6):3020–3026, 2012.
 - [13] Sandrine Kubli, Bernard Waeber, Anne Dalle-Ave, and François Feihl. Reproducibility of laser doppler imaging of skin blood flow as a tool to assess endothelial function. *Journal of cardiovascular pharmacology*, 36(5):640–648, 2000.
 - [14] Jean-Luc Cracowski, Christopher T Minson, Muriel Salvat-Melis, and John R Halliwill. Methodological issues in the assessment of skin microvascular endothelial

- function in humans. *Trends in pharmacological sciences*, 27(9):503–508, 2006.
- [15] Lacy A Holowatz, Caitlin S Thompson-Torgerson, and W Larry Kenney. The human cutaneous circulation as a model of generalized microvascular function. *Journal of Applied Physiology*, 105(1):370–372, 2008.
- [16] Faisal Khan, Dean Patterson, Jill Belch, Kumiko Hirata, and Chim Lang. Relationship between peripheral and coronary function using laser doppler imaging and transthoracic echocardiography. *Clinical Science*, 115:295–300, 2008.
- [17] Jamie Turner, Jill JF Belch, and Faisal Khan. Current concepts in assessment of microvascular endothelial function using laser doppler imaging and iontophoresis. *Trends in cardiovascular medicine*, 18(4):109–116, 2008.
- [18] Heleen van Herpt, Matthijs Draijer, Erwin Hondebrink, Marianne Nieuwenhuis, Gerard Beerhuizen, Ton van Leeuwen, and Wiendelt Steenbergen. Burn imaging with a whole field laser doppler perfusion imager based on a cmos imaging array. *Burns*, 36(3):389–396, 2010.
- [19] A Humeau-Heurtier, E Guerreschi, P Abraham, and G Mahe. Relevance of laser doppler and laser speckle techniques for assessing vascular function: State of the art and future trends. *IEEE transactions on bio-medical engineering*, 2013.
- [20] Y. Yeh and H. Z. Cummins. Localized fluid flow measurements with an he-ne laser spectrometer. *Appl. Phys. Lett.*, 4:176–179, 1964.
- [21] M. D. Stern. In vivo evaluation of microcirculation by coherent light scattering. *Nature*, 254(5495):56–58, March 1975.
- [22] R. Bonner and R. Nossal. Model for laser doppler measurements of blood flow in tissue. *Applied Optics*, 20:2097–2107, 1981.
- [23] M. Heckmeier, S. E. Skipetrov, G. Maret, and R. Maynard. Imaging of dynamic heterogeneities in multiple-scattering media. *J. Opt. Soc. Am. A*, 14(1):185–191, 1997.
- [24] Gilbert T Feke and Charles E Riva. Laser doppler measurements of blood velocity in human retinal vessels. *JOSA*, 68(4):526–531, 1978.
- [25] L. Schmetterer and M. Wolzt. Ocular blood flow and associated functional deviations in diabetic retinopathy. *Diabetologia*, 42(4):387–405, 1999.
- [26] E. Sato, G.T. Feke, M.N. Menke, and J.W. McMeel. Retinal haemodynamics in patients with age-related macular degeneration. *Eye*, 20:697–702, 2006.
- [27] C Riva, B Ross, and G B. Benedek. Laser doppler measurements of blood flow in capillary tubes and retinal arteries. *Invest. Ophthalmol. Vis. Sci.*, 11(11):936–944, 1972.
- [28] J. D. Briers. Laser doppler and time-varying speckle: a reconciliation. *JOSA A.*, 13:345–350, 1996.
- [29] C. E. Riva, G. T. Feke, B. Eberli, and V. Benary. Bidirectional ldv system for absolute measurement of blood speed in retinal vessels. *Applied Optics*, 18:2301–2306, July 1979.
- [30] G Michelson, B Schmauss, MJ Langhans, J Harazny, and MJ Groh. Principle, validity, and reliability of scanning laser doppler flowmetry. *J. Glaucoma.*, 5(2):99–105, 1996.
- [31] R. D. Ferguson, D. X. Hammer, A. E. Elsner, R. H. Webb, S. A. Burns, and J. J. Weiter. Wide-field retinal hemodynamic imaging with the tracking scanning laser ophthalmoscope. *Optics Express*, 12:5198–+, August 2004.
- [32] Vinayakrishnan Rajan, Babu Varghese, Ton G van Leeuwen, and Wiendelt Steenbergen. Review of methodological developments in laser doppler flowmetry. *Lasers in medical science*, 24(2):269–283, 2009.
- [33] Gert E Nilsson, Torsten Tenland, and P Ake Oberg. Evaluation of a laser doppler flowmeter for measurement of tissue blood flow. *Biomedical Engineering, IEEE Transactions on*, (10):597–604, 1980.
- [34] FFM De Mul, J Van Spijker, D Van der Plas, J Greve, JG Aarnoudse, and TM Smits. Mini laser-doppler (blood) flow monitor with diode laser source and detection integrated in the probe. *Applied optics*, 23(17):2970–2973, 1984.
- [35] Thierry Bosch, Silvano Donati, et al. Optical feedback interferometry for sensing application. *Optical engineering*, 40(1):20–27, 2001.
- [36] Milan Nikolić, Elaine Hicks, Yah Leng Lim, Karl Bertling, and Aleksandar D Rakić. Self-mixing laser doppler flow sensor: an optofluidic implementation. *Applied Optics*, 52(33):8128–8133, 2013.
- [37] J. A. Izatt, M. D. Kulkarni, S. Yazdanfar, J. K. Barton, and A. J. Welch. In vivo bidirectional color doppler flow imaging of picoliter blood volumes using optical coherence tomography. *Optics Letters*, 22:1439–1441, September 1997.
- [38] Rainer Leitgeb, L Schmetterer, W Drexler, A Fercher, R Zawadzki, and T Bajraszewski. Real-time assessment of retinal blood flow with ultrafast acquisition by color doppler fourier domain optical coherence tomography. *Optics Express*, 11(23):3116–3121, 2003.
- [39] L. An and R.K. Wang. In vivo volumetric imaging of vascular perfusion within human retina and choroids with optical micro-angiography. *Optics Express*, 16(15):11438–11452, 2008.
- [40] Lin An, Peng Li, Gongpu Lan, Doug Malchow, and Ruikang K Wang. High-resolution 1050 nm spectral domain retinal optical coherence tomography at 120 khz a-scan rate with 6.1 mm imaging depth. *Biomedical optics express*, 4(2):245, 2013.
- [41] Matthijs Draijer, Erwin Hondebrink, Ton van Leeuwen, and Wiendelt Steenbergen. Review of laser speckle contrast techniques for visualizing tissue perfusion. *Lasers in medical science*, 24(4):639–651, 2009.
- [42] David A Boas and Andrew K Dunn. Laser speckle contrast imaging in biomedical optics. *Journal of biomedical optics*, 15(1):011109–011109, 2010.
- [43] Andrew K Dunn. Laser speckle contrast imaging of cerebral blood flow. *Annals of biomedical engineering*, 40(2):367–377, 2012.
- [44] Kausik Basak, M Manjunatha, and Pranab Kumar Dutta. Review of laser speckle-based analysis in medical imaging. *Medical & biological engineering & computing*, 50(6):547–558, 2012.
- [45] Claudia P. Valdes, Hari M. Varma, Anna K. Kristoffersen, Tanja Dragojevic, Joseph P. Culver, and Turgut Durduran. Speckle contrast optical spectroscopy, a non-invasive, diffuse optical method for measuring microvascular blood flow in tissue. *Biomed. Opt. Express*, 5(8):2769–2784, Aug 2014.
- [46] A. F. Fercher and J. D. Briers. Flow visualisation by means of single-exposure speckle photography. *Opt. Commun.*, 37:326–330, 1981.
- [47] J. A. Briers and S. Webster. Quasi real-time digital version of single-exposure speckle photography for full-

- field monitoring of velocity or flow fields. *Optics Comm.*, 116:36–42, 1995.
- [48] J. D. Briers. Laser doppler, speckle and related techniques for blood perfusion mapping and imaging. *Physiological Measurement*, 22:R35–R66, 2001.
- [49] M Roustit, C Millet, S Blaise, B Dufournet, and JL Czacowski. Excellent reproducibility of laser speckle contrast imaging to assess skin microvascular reactivity. *Microvascular research*, 80(3):505–511, 2010.
- [50] Lisa M. Richards, S. M. Shams Kazmi, Janel L. Davis, Katherine E. Olin, and Andrew K. Dunn. Low-cost laser speckle contrast imaging of blood flow using a webcam. *Biomed. Opt. Express*, 4(10):2269–2283, Oct 2013.
- [51] R. Bandyopadhyay, A. S. Gittings, S. S. Suh, P. K. Dixon, and D. J. Durian. Speckle-visibility spectroscopy: A tool to study time-varying dynamics. *Review of Scientific Instruments*, 76(9):093110, 2005.
- [52] Ashwin B Parthasarathy, W James Tom, Ashwini Gopal, Xiaojing Zhang, Andrew K Dunn, et al. Robust flow measurement with multi-exposure speckle imaging. *Opt. Express*, 16(3):1975–1989, 2008.
- [53] Pavel Zakharov, Andreas Völker, Alfred Buck, Bruno Weber, and Frank Scheffold. Quantitative modeling of laser speckle imaging. *Opt. Lett.*, 31(23):3465–3467, 2006.
- [54] Haiying Cheng and Timothy Q. Duong. Simplified laser-speckle-imaging analysis method and its application to retinal blood flow imaging. *Opt. Lett.*, 32(15):2188–2190, Aug 2007.
- [55] Anja I Srienc, Zeb L Kurth-Nelson, and Eric A Newman. Imaging retinal blood flow with laser speckle flowmetry. *Frontiers in neuroenergetics*, 2, 2010.
- [56] Adrien Ponticorvo, Damon Cardenas, Andrew K Dunn, Daniel Ts’o, and Timothy Q Duong. Laser speckle contrast imaging of blood flow in rat retinas using an endoscope. *Journal of biomedical optics*, 18(9):090501–090501, 2013.
- [57] A. Dunn, H. Bolay, M. A. Moskowitz, and D. A. Boas. Dynamic imaging of cerebral blood flow using laser speckle. *Journal of Cerebral Blood Flow and Metabolism*, 21(3):195–201, 2001.
- [58] A. Dunn, A. Devor, M. Andermann, H. Bolay, M. Moskowitz, A. Dale, and D. Boas. Simultaneous imaging of total cerebral hemoglobin concentration, oxygenation and blood flow during functional activation. *Optics Letters*, 28:28–30, 2003.
- [59] AK Dunn, A Devor, AM Dale, and Boas DA. Spatial extent of oxygen metabolism and hemodynamic changes during functional activation of the rat somatosensory cortex. *Neuroimage*, 27(2):279–90, 2005.
- [60] D.D. Duncan and S.J. Kirkpatrick. Can laser speckle flowmetry be made a quantitative tool? *JOSA A*, 25(8):2088–2094, 2008.
- [61] Pavel Zakharov, AC Völker, MT Wyss, F Haiss, N Calcinaghi, C Zunzunegui, Alfred Buck, Frank Scheffold, Bruno Weber, et al. Dynamic laser speckle imaging of cerebral blood flow. *Opt. Express*, 17(16):13904–13917, 2009.
- [62] Ashwin B Parthasarathy, SM Shams Kazmi, and Andrew K Dunn. Quantitative imaging of ischemic stroke through thinned skull in mice with multi exposure speckle imaging. *Biomedical optics express*, 1(1):246, 2010.
- [63] Ashwin B Parthasarathy, Erica L Weber, Lisa M Richards, Douglas J Fox, and Andrew K Dunn. Laser speckle contrast imaging of cerebral blood flow in humans during neurosurgery: a pilot clinical study. *Journal of biomedical optics*, 15(6):066030–066030, 2010.
- [64] Eva Klijn, Hester C Hulscher, Rutger K Balvers, Wim PJ Holland, Jan Bakker, Arnaud JPE Vincent, Clemens MF Dirven, and Can Ince. Laser speckle imaging identification of increases in cortical microcirculatory blood flow induced by motor activity during awake craniotomy: Clinical article. *Journal of neurosurgery*, 118(2):280–286, 2013.
- [65] A. Serov, W. Steenbergen, and F. de Mul. Laser doppler perfusion imaging with complementary metal oxide semiconductor image sensor. *Optics Letters*, 27:300–302, 2002.
- [66] A. Serov, B. Steinacher, and T. Lasser. Full-field laser doppler perfusion imaging monitoring with an intelligent cmos camera. *Opt. Ex.*, 13(10):3681–3689, 2005.
- [67] Haiying Cheng, Yumei Yan, and Timothy Q Duong. Temporal statistical analysis of laser speckle images and its application to retinal blood-flow imaging. *Optics express*, 16(14):10214, 2008.
- [68] Matthijs Draijer, Erwin Hondebrink, Ton van Leeuwen, and Wiendelt Steenbergen. Twente optical perfusion camera: system overview and performance for video rate laser doppler perfusion imaging. *Optics express*, 17(5):3211–3225, 2009.
- [69] Dake Wang, Adam Moyer, and Matt Henderson. Intensity fluctuation spectra of dynamic laser speckle patterns acquired by a full-field temporal modulation method. *Applied optics*, 51(31):7576–7580, 2012.
- [70] Yaguang Zeng, Mingyi Wang, Guangping Feng, Xianjun Liang, and Guojian Yang. Laser speckle imaging based on intensity fluctuation modulation. *Opt. Lett.*, 38(8):1313–1315, Apr 2013.
- [71] Mingyi Wang, Yaguang Zeng, Xianjun Liang, Guangping Feng, Xuanlong Lu, Junbo Chen, Dingan Han, and Guojian Yang. In vivo label-free microangiography by laser speckle imaging with intensity fluctuation modulation. *Journal of Biomedical Optics*, 18(12):126001–126001, 2013.
- [72] M. Leutenegger, E. Martin-Williams, P. Harbi, T. Thacher, W. Raffoul, M. André, A. Lopez, P. Lasser, and T. Lasser. Real-time full field laser doppler imaging. *Biomedical Optics Express*, 2(6):1470–1477, 2011.
- [73] Quan Gu, Barrie R Hayes-Gill, and Stephen P Morgan. Laser doppler blood flow complementary metal oxide semiconductor imaging sensor with analog on-chip processing. *Applied optics*, 47(12):2061–2069, 2008.
- [74] D He, HC Nguyen, BR Hayes-Gill, Y Zhu, JA Crowe, GF Clough, CA Gill, and SP Morgan. 64 × 64 pixel smart sensor array for laser doppler blood flow imaging. *Optics Letters*, 37(15):3060–3062, 2012.
- [75] Hoang C Nguyen, Barrie R Hayes-Gill, Yiqun Zhu, John A Crowe, Diwei He, and Stephen P Morgan. Low resource processing algorithms for laser doppler blood flow imaging. *Medical engineering & physics*, 33(6):720–729, 2011.
- [76] Diwei He, Hoang C Nguyen, Barrie R Hayes-Gill, Yiqun Zhu, John A Crowe, Cally Gill, Geraldine F Clough, and Stephen P Morgan. Laser doppler blood flow imaging using a cmos imaging sensor with on-chip signal processing. *Sensors*, 13(9):12632–12647, 2013.
- [77] Joseph L Calkins and Carl D Leonard. Holographic recording of a retina using a continuous wave laser. *Investigative Ophthalmology & Visual Science*, 9(6):458–

- 462, 1970.
- [78] R. L. Wiggins, K. D. Vaughan, and G. B. Friedmann. Holography using a fundus camera. *Appl. Opt.*, 11(1):179–181, Jan 1972.
- [79] AN Rosen. Holographic funduscopy with fibre optic illumination. *Optics & Laser Technology*, 7(3):127–129, 1975.
- [80] H Ohzu and T Kawara. Application of holography in ophthalmology. pages 133–146, 1979.
- [81] Allan R Tokuda, David C Auth, and Adam P Bruckner. Holocamera for 3-d micrography of the alert human eye. *Applied Optics*, 19(13):2219–2225, 1980.
- [82] M. Atlan, M. Gross, T. Vitalis, A. Rancillac, B. C. Forget, and A. K. Dunn. Frequency-domain, wide-field laser doppler in vivo imaging. *Optics Letters*, 31(18):2762–2764, 2006.
- [83] M. Simonutti, M. Paques, J. A. Sahel, M. Gross, B. Samson, C. Magnain, and M. Atlan. Holographic laser doppler ophthalmoscopy. *Opt. Lett.*, 35(12):1941–1943, 2010.
- [84] M. Atlan, M. Gross, and J. Leng. Laser doppler imaging of microflow. *Journal of the European Optical Society - Rapid publications*, 1:06025–1, 2006.
- [85] M. Atlan and M. Gross. Laser doppler imaging, revisited. *Review of Scientific Instruments*, 77(11):1161031–1161034, 2006.
- [86] Nicolas Verrier, Daniel Alexandre, and Michel Gross. Laser doppler holographic microscopy in transmission: application to fish embryo imaging. *Opt. Express*, 22(8):9368–9379, Apr 2014.
- [87] M. Atlan, M. Gross, T. Vitalis, A. Rancillac, J. Rossier, and A. C. Boccara. High-speed wave-mixing laser doppler imaging in vivo. *Optics Letters*, 33(8):842–844, 2008.
- [88] M. Atlan and M. Gross. Spatiotemporal heterodyne detection. *Journal of the Optical Society of America A*, 24(9):2701–2709, 2007.
- [89] D.A. Boas and A.G. Yodh. Spatially varying dynamical properties of turbid media probed with diffusing temporal light correlation. *J. Opt. Soc. Am. A*, 14:192–215, 1997.
- [90] D. J. Pine, D. A. Weitz, P. M. Chaikin, and E. Herzogheimer. Diffusing-wave spectroscopy. *Physical Review Letters*, 60(12):1134–1137, 1988.
- [91] AH Gandjbakhche, R. Nossal, and RF Bonner. Scaling relationships for theories of anisotropic random walks applied to tissue optics. *Appl. Opt.*, 32(4):504–516, 1993.
- [92] U. Schnars and W. Juptner. Direct recording of holograms by a ccd target and numerical reconstruction. *Appl. Opt.*, 33:179–181, 1994.
- [93] U. Schnars and W. P. O. Juptner. Digital recording and numerical reconstruction of holograms. *Meas. Sci. Technol.*, 13:R85–R101, 2002.
- [94] Myung K Kim, Lingfeng Yu, and Christopher J Mann. Interference techniques in digital holography. *Journal of Optics A: Pure and Applied Optics*, 8(7):S518–S523, 2006.
- [95] Pascal Picart and Julien Leval. General theoretical formulation of image formation in digital fresnel holography. *JOSA A*, 25(7):1744–1761, 2008.
- [96] Nicolas Verrier and Michael Atlan. Off-axis digital hologram reconstruction: some practical considerations. *Appl. Opt.*, 50(34):H136–H146, Dec 2011.
- [97] Michael Atlan, Pierre Desbiolles, Michel Gross, and Maité Coppey-Moisan. Parallel heterodyne detection of dynamic light-scattering spectra from gold nanoparticles diffusing in viscous fluids. *Opt. Lett.*, 35(5):787–789, 2010.
- [98] M. Gross and M. Atlan. Digital holography with ultimate sensitivity. *Optics Letters*, 32(8):909–911, 2007.
- [99] RL Whitman and A Korpel. Probing of acoustic surface perturbations by coherent light. *Applied Optics*, 8(8):1567–1576, 1969.
- [100] M. Ueda, S. Miida, and T. Sato. Signal-to-noise ratio and smallest detectable vibration amplitude in frequency-translated holography: an analysis. *Applied Optics*, 15(11):2690–2694, 1976.
- [101] Jean-Pierre Monchalain. Heterodyne interferometric laser probe to measure continuous ultrasonic displacements. *Review of scientific instruments*, 56(4):543–546, 1985.
- [102] James W Wagner and James B Spicer. Theoretical noise-limited sensitivity of classical interferometry. *JOSA B*, 4(8):1316–1326, 1987.
- [103] D Royer and E Dieulesaint. Optical probing of the mechanical impulse response of a transducer. *Applied physics letters*, 49(17):1056–1058, 1986.
- [104] P. Picart, J. Leval, D. Mounier, and S. Gougeon. Time-averaged digital holography. *Opt. Lett.*, 28:1900–1902, 2003.
- [105] JM Schmitt, AH Gandjbakhche, and RF Bonner. Use of polarized light to discriminate short-path photons in a multiply scattering medium. *Appl. Opt.*, 31(30):6535, 1992.
- [106] L Rovati, S Cattini, N Zambelli, F Viola, and G Staurenghi. In-vivo diffusing-wave-spectroscopy measurements of the ocular fundus. *Optics express*, 15(7):4030–4038, 2007.
- [107] Yu Lin, Lian He, Yu Shang, and Guoqiang Yu. Non-contact diffuse correlation spectroscopy for noninvasive deep tissue blood flow measurement. *Journal of Biomedical Optics*, 17(1):0105021–0105023, 2012.
- [108] Stefano Cattini, Giovanni Staurenghi, Antonietta Gatti, and Luigi Rovati. In vivo diffuse correlation spectroscopy investigation of the ocular fundus. *Journal of biomedical optics*, 18(5):057001–057001, 2013.
- [109] Norbert Wiener. Generalized harmonic analysis. *Acta Mathematica*, 55(1):117–258, 1930.
- [110] B.E.A. Saleh and M.C. Teich. *Fundamentals of photonics*. Wiley, 1991.
- [111] Joseph W. Goodman. *Statistical Optics*. John Wiley & Sons, 1985.
- [112] D. Kumar and P.M. Bellan. Heterodyne interferometer with unequal path lengths. *Review of scientific instruments*, 77(8):083503–083503, 2006.
- [113] B. J. Berne and R. Pecora. *Dynamic Light Scattering*. Dover, 2000.
- [114] XL Wu, DJ Pine, PM Chaikin, JS Huang, and DA Weitz. Diffusing-wave spectroscopy in a shear flow. *J. Opt. Soc. Am. B*, 7(1):15–20, 1990.
- [115] Haruo Ogiwara. Laser doppler velocimeter with a differential photodiode array. *Applied optics*, 18(10):1533–1538, 1979.
- [116] Yogesh Chandra Agrawal and James R McCullough. Directional pedestal-free laser doppler velocimetry without frequency biasing. part 1. *Applied Optics*, 20(9):1553–1556, 1981.
- [117] Martine Rousseau. Statistical properties of optical fields scattered by random media. application to rotating ground glass. *JOSA*, 61(10):1307–1316, 1971.

- [118] LE Estes, LM Narducci, and RA Tuft. Scattering of light from a rotating ground glass. *JOSA*, 61:1301, 1971.
- [119] B. P. Singh and S. Chopra. Investigations of the autocorrelation function for inhomogeneous scatterers. *JOSAA*, 9:619–620, 1992.
- [120] Jicun Zhong, Alexander M Seifalian, Goran E Salerud, and Gert E Nilsson. A mathematical analysis on the biological zero problem in laser doppler flowmetry. *Biomedical Engineering, IEEE Transactions on*, 45(3):354–364, 1998.
- [121] DP Kernick, JE Tooke, and AC Shore. The biological zero signal in laser doppler fluximetry—origins and practical implications. *Pflügers Archiv*, 437(4):624–631, 1999.
- [122] Dirk J Faber, Maurice CG Aalders, Egbert G Mik, Brett A Hooper, Martin JC Gemert, and Ton G Leeuwen. Oxygen saturation-dependent absorption and scattering of blood. *Physical review letters*, 93(2):028102, 2004.
- [123] Jonas Binding, Juliette Ben Arous, Jean-François Léger, Sylvain Gigan, Claude Boccara, and Laurent Bourdieu. Brain refractive index measured in vivo with high-na defocus-corrected full-field oct and consequences for two-photon microscopy. *Optics express*, 19(6):4833–4847, 2011.
- [124] Lorrin A Riggs, John C Armington, and Floyd Ratliff. Motions of the retinal image during fixation. *JOSA*, 44(4):315–321, 1954.
- [125] Gerald Westheimer and Suzanne P McKee. Visual acuity in the presence of retinal-image motion. *JOSA*, 65(7):847–850, 1975.
- [126] Kate Grieve, Michel Paques, Arnaud Dubois, Jose Sahel, Claude Boccara, and Jean-Francois Le Gargasson. Ocular tissue imaging using ultrahigh-resolution, full-field optical coherence tomography. *Invest. Ophthalmol. Vis. Sci.*, 45(11):4126–4131, 2004.
- [127] Ruikang K. Wang. Directional blood flow imaging in volumetric optical microangiography achieved by digital frequency modulation. *Opt. Lett.*, 33(16):1878–1880, 2008.
- [128] Benjamin Samson, Frédéric Verpillat, Michel Gross, and Michael Atlan. Video-rate laser doppler vibrometry by heterodyne holography. *Opt. Lett.*, 36(8):1449–1451, Apr 2011.
- [129] B. Samson and M. Atlan. Short-time fourier transform laser doppler holography. *Journal of the European Optical Society - Rapid publications*, 8(0), 2013.
- [130] Timothe Laforest, Antoine Dupret, Arnaud Verdant, François Ramaz, Sylvain Gigan, Gilles Tessier, and Emilie Benoit la Guillaume. A 4000 hz cmos image sensor with in-pixel processing for light measurement and modulation. pages 1–4, 2013.