

Design optimization using Statistical Confidence Boundaries of response surfaces: Application to robust design of a biomedical implant

Laetitia Rossi, Jean-Marc Linares, Julien Chaves-Jacob, Jean Mailhé,

Jean-Michel Sprauel

► To cite this version:

Laetitia Rossi, Jean-Marc Linares, Julien Chaves-Jacob, Jean Mailhé, Jean-Michel Sprauel. Design optimization using Statistical Confidence Boundaries of response surfaces: Application to robust design of a biomedical implant. CIRP Annals - Manufacturing Technology, 2014, 63 (1), pp.141–144. hal-01303708

HAL Id: hal-01303708 https://hal.science/hal-01303708

Submitted on 19 Jan2017

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Design optimization using Statistical Confidence Boundaries of response surfaces: Application to robust design of a biomedical implant

Laetitia Rossi^{a,b}, Jean-Marc Linares (2)^{a,b,*}, Julien Chaves-Jacob^a, Jean Mailhe^a, Jean-Michel Sprauel^a

^a Aix-Marseille Universite, CNRS, ISM UMR 7287, 13288 Marseille, Cedex 09, France ^b Biotech International, 305 Allees de Craponne, 13300 Salon de Provence, France

This paper deals with the use of Statistical Confidence Boundaries (SCB) of response surfaces in robust design optimization. An empirical model is therefore selected to describe a real design constraint function. This constraint is thus approximated by a second order polynomial expansion which is fitted to numerical simulations that use a Finite Element Method (FEM). A technique is also proposed to analyze the effects of the uncertainties of the inputs of the simulations. This approach is employed to optimize the design of a biomedical wrist implant. A real optimized implant is then manufactured and tested to validate the numerical model.

Keywords: Design optimization Uncertainty Biomedical

1. Introduction

In the common life, wounds can appear, due to sports activities or accidents in the home. Often, prostheses or implants are then used to restore the functional capacity of the patient [1]. The design of biomedical products is however constrained by many criteria (biocompatibility of the materials, lifespan of the implant, small available geometrical space...) and needs multi-objective optimization methods. An overview of Design Theories and Methodologies (DTM) was already presented in a keynote paper [2]. The prescriptive models for design were thus discussed (canonical design process, morphological analysis and prescriptive models of the design artefacts, Suh's Axiomatic Design and Taguchi Method). These DTM methods can be used to enrich the functional and attributive information of design solutions. Different approaches are then employed to characterize the Design Solution Surface (DSS). The proposed models can thus be classified with two different points of view: Global/local description and Analytical/ Numerical modelling. As example, the optimization of the deflection of a cantilever is presented in Fig. 1. The first approach allows calculating the design solution for any set of input parameters. The quality of the results also only depends on the accuracy of the model. Generally, the time and cost to develop a global analytical model are however significant. For that reason, empirical models are employed in the second approach to describe the real design constraint function. Usually, a polynomial model is then locally best fitted to the real DSS, using a limited number of numerical simulations or real experiments. A design of experiment (DOE) technique can therefore be employed to define the optimal set of input parameters of this approach. Optimization of the

design parameters is however limited to the local domain used for fitting. In the third approach a single numerical simulation is carried out, just to check that the design constraints are satisfied for the given selected set of inputs. This simulation is usually based on a Finite Element Model (FEM). In the last approach, the numerical simulation is repeated many times to get a global view of the DSS. This method, however, only provides a discrete description of the DSS. In the second approach, the quality of the best fit is a central property for the accuracy of the optimization. In some works [3,4], the major components of Engineering Design Optimization (EDO) were classified in five entities: design variables, constraints, objective functions, problem domain and environment. Their uncertainties or variations are propagated to the optimized design solution to check the robustness of the design [5]. This permits also verifying that the design requirements will be satisfied for all manufactured products. Robust design optimization [6-8] can be used to design biomedical products. Roy et al. conclude their keynote paper with this sentence: "there is a lack of research in multi-objective design optimization that deals with



Fig. 1. Different approaches to characterize design solution surfaces.

uncertainty and constraints together" [3]. In this context, the aim of our paper is to use the Statistical Confidence Boundary (SCB) of response surfaces in design of experiment-based optimization.

2. Presentation of the method

This section presents the methodology used to get the response surface and the related SCBs, of a DSS characterized by FEM simulations. As example, the approach will focus on a problem with two factors (U1 and U2). The response surface will be approximated by the second order polynomial empirical model described by Eq. (1).

$$\begin{array}{c} \mathbf{X}\mathbf{X} \\ y \ \mathbf{\chi} & \mathbf{b}_{ij} \ \mathbf{x}_i \ \mathbf{x}_j \\ & \mathbf{i}_{10}^{ij} \mathbf{i}_{1i}^{ij} \end{array} \tag{1}$$

where $x_0 = 1$. x_1 , x_2 are the normalized values of factors U1, U2.

The different possible shapes of response surfaces are drawn in Fig. 2. A DOE strategy permits defining the optimal set of inputs (U1, U2) to be employed for the FEM simulations. The coefficients of the polynomial model are then derived from the FEM results using the following pseudo-inverse calculation:



Fig. 2. Principle of D.O.E based optimization.

 \hat{B} defines the best estimate of the surface response coefficients.

(2)

A propagation method is then implemented to account for the uncertainties of the input parameters and the inaccuracy of the model. The scatter of the material properties, design parameters and manufacturing conditions is assumed random and normal distributed. It is described by standard deviations (**S**).

The inaccuracy of the fitting model is characterized by the root mean square (Rms) of the differences between the FEM simulation results and the mean local polynomial used to describe the DSS. Variations of the environment are not taken into account in this study. The scheme used to propagate the different deviations to the SCB is presented in Fig. 3.



Fig. 3. Uncertainty and Error Propagation scheme.

2.1. Propagation of the inaccuracies of the fitting model

The inaccuracy of the model is represented by the root mean square RmsðyÞ of the best fit residues R calculated through Eq. (3).

$$R \frac{1}{Y} Y \ddot{\mathbf{y}} X \hat{\mathbf{B}}$$
 (3)

The mean square error matrix MSEðB^b of the response surface coefficients is then calculated using following expression:

$$MSE\delta\hat{B}^{\flat} \ \ MSE\delta\delta X^{T} \ \ X^{\flat}^{\mathbf{y}^{1}} \ \ X^{T} \ \ Y^{\flat}$$

$$MSE\delta\hat{B}^{\flat} \ \ \mathcal{M} SE\delta\hat{B}^{\flat} \ \ \mathcal{M} SE\delta^{\flat} \ \ \mathcal{M} SE\delta^{$$

This permits finally evaluating the mean square error $Mse\delta\hat{y}b$ of the design solution \hat{y} estimated for any set of input parameters. Following classical propagation method is used for that purpose:

$$Mseðŷ Þ ¼ J_{\hat{B}} MSEðB \hat{P} J^{T}_{\hat{B}}$$
(5)

with, $J_{\hat{B}}$; Jacobian of function \hat{y} with respect to coefficients b_{ij} .

2.2. Propagation of the uncertainties of the design parameters

The scatter of the design parameters is usually characterized by Tolerance Intervals (TI) that are either imposed by the designer or derived from the capability of the manufacturing process. If a Gaussian Probability Density Function (PDF) is assumed for a given input parameter x_i , the standard deviation S_{xi} can be derived from the tolerance interval TI_i through following expression:

$$\mathbf{S}_{\mathrm{xi}} \overset{\mathrm{Ti}_{\mathrm{i}}}{\sim} \mathbf{6}$$

Assuming the independence of the two parameters x_1 , x_2 , the standard deviations are then propagated to the response surface, using Eq. (7).

$$Varðŷ 1/4 J_{xi} VARð x_i J_{xi}^{T}$$
(7)

with, J_{xi} , Jacobian of function \hat{y} with respect to x_i

VAR
$$\delta x_i \not\models \frac{11^2 = 36}{0}$$
 II $z_1 = \frac{36}{26}$ is the covariance matrix of $x_{i.}$

2.3. Propagation to load and stress, of the scatter of the material properties, in the case of a pure elastic behaviour

Mechanical tests (tensile tests) are usually carried out to characterize the material properties. They allow defining the mean Young's modulus E of the material, and the related standard deviation $S_{\rm E}$. Common constrains imposed in design optimization are the maximum stress (S) or applied load (L) to which the structure must resist. In the case of imposed displacements, and pure elastic behaviour of the material, the resulting applied Load and stress are proportional to Young's modulus. This leads to following relationships:

$$L \ \ k_L E$$
) Varð Lþ $\ \ k_L^2$ Varð Eþ) $\mathbf{S}_L^2 \ \ \delta L = E b^2 \ \mathbf{S}_E^2$ (8)

$$S \ \ k_S E$$
) $Var\delta S \ \ k_S^2 Var\delta E$ **)** $S^2 \ \ \delta S = E \ \ S^2 E$ (9)

2.4. Response surface and Statistical Confidence Boundary

Previous calculations are used to evaluate the design solution \hat{y} for any set of input parameters and estimate its mean square error (Eq. (10)).

Mseðy ¼ J_B MSEðB J^T
$$\underset{\hat{B}}{\overset{}{\text{b}}} J_{Xi}$$
 VARð $X_i \overset{}{\overset{}{\text{b}}} J^{T}_i \overset{}{\underset{}{\text{b}}} \overset{L \text{ or } S^2}{\underbrace{E}}$
VarðE (10)

With smooth design solution surfaces, the model inaccuracy remains small in comparison to the random perturbations of the inputs that are assumed to be normal distributed. The global distribution is therefore close to a Gaussian. The SCBs of the response surface can thus be approximated in the following way:

$$\mathbf{q}$$
SCB ¼ ŷ k MseðyÞ; k 2 for a confidence level of 95%: (11)

3. Application to the design of a biomedical wrist implant

The function of the studied biomedical implant is to substitute the natural ligament between two bones of the wrist. Experiments of the literature show that the Young's modulus E of the ligament is ranged between 50 and 100 MPa [9]. If assuming a uniform PDF and a risk of 5%, this leads to the confidence interval E = 75 10 MPa. The maximum elongation of the ligament and its torsion angle were estimated by Upal et al. [10] to about 0.9 mm and 158 respectively. The material selected for the implant is PolyEther Ether Ketone (PEEK) which is a biocompatible thermoplastic organic polymer. Its Young's modulus was evaluated by tensile tests to $E_0 = 3582$ 127 MPa (confidence level 95%). In relation to the fatigue limit, the maximum stress that can be applied to the material is restricted to 90 MPa. The maximum overall dimensions of the implant were specified by the surgeon to 18 1 8 1 2 mm.

3.1. Implant design driven by material properties

The functionality of the implant is to reproduce the mechanical behaviour of the ligament. This imposes designing an implant with a very low stiffness, which is not directly compatible with the elastic modulus of PEEK. One way to overcome this impediment is to create an elementary pattern with holes. The equivalent Young's modulus of such volume can then be evaluated through a micro-mechanical approach of porous materials. Fig. 4 shows the predictions of the classical Voigt and Hill models. Both models lead to simple linear expressions (Eqs. (12) and (13)) that permit evaluating the density of holes required to adjust the stiffness of the elementary peek pattern to the Young's modulus of the ligament. Hill's approach, which is the most accurate in the case of random repartitions of voids, predicts that the material loses its consistency when the porosity exceeds 50%.



Fig. 4. Modulus evolution versus porosity and topology.

These two equations bring to the fore how it is difficult to obtain an implant that owns stiffness compatible with the elastic behaviour of the natural ligament. To accept the applied displacements (0.9 mm, 158) and reproduce the behaviour of the natural ligament (toe and linear regions in Fig. 5), the design of the proposed implant is based on two springs (with respective rate K1 and K2 and n_1 , n_2 number of windings) mounted in parallel with two stops (e_1 and e_2). Fig. 5 presents the rheological scheme of the implant. The first spring K1 was dimensioned to accept a torsion angle of 158 and an elongation of 0.9 mm. The second spring K2 was calculated to bring an elongation of 0.3 mm. The stop device e_2 is a titanium cable designed to carry the final load.



Fig. 5. Behaviours of the natural ligament and the designed product.

3.2. Optimization methodology

The spring K2 was studied first. Each of its windings consists of an elementary pattern (central part) that was optimized through a DOE methodology. This pattern is shown in Fig. 6. It is characterized by two main parameters: a width U1 and a radius U2. In Fig. 5, the number of windings (n_2) was fixed to 3.



Fig. 6. Design of the elementary pattern of spring K2.

The mechanical behaviour of the elementary pattern was characterized by FEM simulations. Nine configurations were thus studied. The composite DOE matrix, presented in Table 1, was therefore used to define the optimal set of design parameters U1, U2 to be employed for the simulations.

Table 1 Design factors (U1, U2) and related normalized factors (x_1, x_2) for $n_2 = 3$.

U1 (mm)	\mathbf{x}_1	U2 (mm)	x ₂
1.5	ÿ1	0.45	ÿ1
2	1	0.45	ÿ1
1.5	ÿ1	0.65	1
2	1	0.65	1
1.5	0	0.55	C
$\frac{2}{1.75}$	0	0.55	0
1.75	0	0.45	0

For each configuration, the FEM calculations permitted defining the maximum equivalent stress S of the material and the total load L applied to the spring at its maximum imposed elongation of 0.3 mm. The mean response surfaces of the stress (\hat{y}_s) and the load (\hat{y}_L) versus the two factors U1, U2, and the associated SCBs were finally derived from the nine performed simulations. These two response surfaces are shown in Fig. 7.



Fig. 7. Stress and load vs. width U1 and radius U2 and related SCB's.

The same procedure was applied to springs of different numbers of windings (i.e. $n_2 = 2.5$, 3 and 3.5). However, due to the maximum overall dimension limits specified by the surgeon, the best compromise proved to be 3. The challenge was then to maximize the load bearing capacity of the spring while maintaining the equivalent stress below the material limit of 90 MPa. This optimization under constraint is defined by following expressions:

$$\overset{O}{\underset{k}{\text{Max}}} \overset{Max}{\overset{\delta}{\text{y}}_{L}} \overset{b}{\text{while}} \overset{\hat{y}_{S}}{\overset{b}{\text{p}}} \overset{k}{\overset{S}{\text{s}}} \overset{S}{\overset{S}{\text{max}}} \overset{Max}{\overset{Max}} \overset{\delta}{\overset{With}{\text{max}}} \overset{g}{\overset{Hax}{\text{max}}} \overset{g}{\overset{Hax}{}} \overset{g}{\overset{Hax}{\text{max}}} \overset{g}{$$

3.3. Optimization result with uncertainty

8

The manufacturing tolerance intervals of width U1 and radius U2 were fixed to TI = 0.1 mm. An optimization algorithm, based on Eq. (14) was used to define the best configuration of the spring. The width (U1, x_1) and the radius (U2, x_2) of the optimized winding pattern are shown in Eq. (15). The maximum load L held by the spring and its uncertainty were also evaluated.

$$\begin{array}{cccc} x_1 & 1 & U1 & 2 & mm \\ x_2 & 4 & \ddot{y}0:155 & U2 & 4 & 0:534 & mm \end{array} L & 10:16 & 0:81 & (15) \end{array}$$

4. Numerical validation and experimental testing

To confirm the results of the optimization procedure, two checks were made (Fig. 8). First an additional FEM simulation was realized with Abaqus software to validate the results of the D.O.E. approach. It was performed for the spring configuration corresponding to the best design parameters. As expected, the maximum equivalent stress calculated for the largest imposed elongation was found below the allowable material limit of 90 MPa. The load bearing capacity of the optimized spring was also evaluated to 10.24 N. This is very close to the value derived from the response surface. In second, two real springs were machined and tested. The manufactured prototypes were therefore stretched to the maximum elongation of 0.3 mm. Repeated tests permitted



Fig. 8. Numerical and experimental testing

then to define the mean load applied to the part and to estimate the error bars. The load bearing capacity of the springs was thus evaluated to 10.31 0.55 MPa. Fig. 9 summarizes the different results. All values are in good agreement.



Fig. 9. Estimated, FEM and experimental load values.

5. Conclusion

In this paper, a D.O.E. approach was proposed to build empirical polynomial models for design solution surfaces. An error and uncertainty propagation technique was also suggested to define the Statistical Confidence Boundaries of any design constraint. The method was applied successfully to the robust design of a biomedical wrist implant designed to replace a natural ligament. The main property of the device is to permit a large elongation while resisting to significant loads. Due to the great difference of stiffness of the biological matter and the polymer (PEEK) used for the implant, these constrains proved particularly difficult to satisfy. The designed device was therefore based on two springs mounted in parallel with two stops and a titanium wire to hold the final load. An optimization under constraint, based on results of FEM simulations, permitted then to define the best configuration and dimensions of the springs. The results were confirmed by tests carried out on real manufactured prototypes.

References

- Serenó L, Ferrer I, Soy F, Salvador A, Elías-Zúñiga A, Perez T, Ciurana Q (2013) Designing and Prototyping of New Device for Scapholunate Ligament Repair. Procedia CIRP 5:270–275.
- [2] Tomiyama T, Gu PH, Jin Y, Lutters D, Kind Ch, Kimura F (2009) Design Methodologies: Industrial and Educational Applications. CIRP Annals – Manufacturing Technology 58:543–565.
- [3] Roy R, Hinduja S, Teti R (2008) Recent Advances in Engineering Design Optimisation: Challenges and Future Trends. CIRP Annals – Manufacturing Technology 57:697–715.
- [4] Beyer H-G, Sendhoff B (2007) Robust Optimization—A Comprehensive Survey. Computer Methods in Applied Mechanics and Engineering 196:3190–3218.
- [5] Roy R, Azene YT, Farrugia D, Onisa C, Mehnen J (2009) Evolutionary Multi-Objective Design Optimization with Real Life Uncertainty and Constraints. CIRP Annals – Manufacturing Technology 58:169–172.
- [6] Gu P, Lu B, Spiewak S (2004) A New Approach for Robust Design of Mechanical Systems. CIRP Annals – Manufacturing Technology 53:129–133.
- [7] Zhang J, Li SP, Bao NS, Zhang GJ, Xue DY, Gu PH (2010) A Robust Design Approach to Determination of Tolerances of Mechanical Products. CIRP Annals – Manufacturing Technology 59:195–198.
- [8] Dantan JY, Qureshi AJ, Antoine JF, Eisenbart B, Blessing L (2013) Management of Product Characteristics Uncertainty Based on Formal Logic and Characteristics Properties Model. CIRP Annals – Manufacturing Technology 62:147–150.
- [9] Subit D, Chabrand P, Masson C (2009) Micromechanical Model to Predict Damage and Failure in Biological Tissues. Application to the Ligament to Bone Attachment in the Human Knee Joint. Journal of Biomechanics 42:261–265.
- [10] Upal MA, Crisco JJ, Moore DC, Sonenblum SE, Wolfe SW (2006) In Vivo Elongation of Planar and Dorsal Scapholunate Interosseous Ligament. The Journal of Hand Surgery 31:1326–1332.