



HAL
open science

Optical-based probe for real time assessment of RF electrical field during MRI exam

Isabelle Saniour, Gwenaël Gaborit, Lionel Duvillaret, Anne-Laure Perrier, Olivier Beuf

► **To cite this version:**

Isabelle Saniour, Gwenaël Gaborit, Lionel Duvillaret, Anne-Laure Perrier, Olivier Beuf. Optical-based probe for real time assessment of RF electrical field during MRI exam. 25th annual meeting & exhibition ISMRM, Apr 2017, Honolulu, United States. <hal-01693604>

HAL Id: hal-01693604

<https://hal.science/hal-01693604v1>

Submitted on 19 Jun 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

Optical-based probe for real time assessment of RF electrical field during MRI exam

Isabelle Saniour¹, Gwenaël Gaborit^{2,3}, Lionel Duvillearet³, Anne-Laure Perrier², and Olivier Beuf¹

¹Univ. Lyon, CREATIS ; CNRS UMR 5220 ; INSERM U1206 ; INSA-Lyon ; UJM-Saint-Etienne ; Université Lyon1, Villeurbanne, France, ²Univ. Savoie-Mont-Blanc, IMEP-LAHC, Le Bourget-du-Lac, France, ³Kapteos, Sainte-Hélène-du-Lac, France

Synopsis

During MRI exam, Specific Absorption Rate (SAR) is essential to be controlled and can be evaluated by measuring either indirectly for instance the rise in temperature or directly the radiofrequency induced electrical E -field. In the current study, we proposed an optical probe based on the Pockels effect for subcentimeter resolution measurements of the E -field without altering the surrounding media. Measurements were performed at 4.7 T and 3.0 T. Results show that the probe has an excellent linear response and allow a real time estimate of the three components of the E -field produced during MRI examination.

Introduction

Patient safety is controlled during an MRI exam by estimating the global specific absorption rate (SAR). Without dedicated devices such as radiofrequency (RF) traps, the electrical E -field induces RF currents, resulting in heating near metallic conductors. The SAR-parameter is determined either by the E -field or the temperature rise. E -field probes are fast and more sensitive than thermal sensors but hard to implement. Dipole-based probes were largely used to measure the E -field in MRI¹⁻³. However, their metallic structure may perturb the electromagnetic fields and the spatial resolution is limited by the probe's dimensions. Recently, optical-based methods were used to measure the E -field^{4,5}. But, the value of the E -field was not precisely determined, besides only one component of the E -field was estimated. In the following, the characteristics of the optical probe and real time measurements of the three components of the RF E -field are presented.

Methods

Birefringence characterizes crystals that change their refractive indices with an applied E -field. Based on this principle, a sub-cm optical probe using isotropic crystals (ZnTe, BSO...) was designed and built. A probe with isotropic crystal allows the measurement of both x- and y-orientations of the E -field so-called transverse probe. The z-component is measured by a longitudinal probe. A fiber optic was connected to the crystal to drive a laser diode ($\lambda = 1.55 \mu\text{m}$, relative intensity noise (RIN) < 150 dB/Hz) having an incident polarization that changes with the applied E -field. The modulation of the polarization was analyzed using a processing unit (eosens, Kapteos, France) and then converted into an electrical signal using a photodiode. A spectrum analyzer was used to display the envelope of the detected electro-optical (EO) signal on a fixed and narrow frequency bandwidth (30 kHz). First, the E -field measurements were performed on 4.7 T preclinical MRI (Bruker). A high-pass 60 mm-diameter linear transceiver birdcage with capacitors placed on the two end sides was used. The probe was fixed on a 3-axis translation bench (Fig 1). The measurements were performed in air in order to verify the sensing abilities of the probe. The probes were calibrated outside the MRI to obtain the proportionality factor between the EO signal and the E -field. A FLASH sequence with 1 ms Hermite shape RF excitation pulse was run. In order to determine the relation between the emitted magnetic B_1 -field and the detected E -field, the flip angle of the RF pulse was varied and the corresponding E -field was measured. Besides, the two transverse and longitudinal probes were used separately to measure the three components of the E -field along the z-axis. Finally, this measuring protocol was also applied on a 3.0 T clinical MRI (GEHC MR750). MR images of an orange with and without inserted probe were acquired.

Results

Figure 2 exhibits an excellent linearity between E - and B_1 -fields for both probes measuring radial and longitudinal components. The B_1 -field was derived from the shape (Hermite equation) and amplitude (flip angle). The maximum values of the E -field were measured at 1078 V/m and 488 V/m for transverse and longitudinal probes, respectively. The sensitivity was established at $E_{min}=0.27\text{V/m.Hz}^{-1/2}$ for transverse probe and $E_{min}=0.087\text{V/m.Hz}^{-1/2}$ for longitudinal probe. Figure 3 displays the three components of the E -field along z-axis. Two maximum peaks were found at $Z=5\text{cm}$ and $Z=11\text{cm}$ for all curves. Next, the exact shapes of three different excitation pulses having the same excitation time (1 ms) and power (138W) were detected by the optical probe (Fig 4). The difference between the MR images of the orange with and without the inserted probe is negligible and the measured E -field was equal to 198 V/m (Fig 5).

Discussion

The E -field has a linear dependence with respect to the B_1 -field. The highest value of the E -field is obtained when the flip angle is equal to 90° which corresponds to the highest value of the emitted B_1 -field. The EO probe is highly sensitive to the variation of the E -field value regarding the RF pulse shape. The y-component of the E -field is the dominant component noting that the RF B_1 excitation pulse is linear and emitted along the x-axis. The detected peaks along z-direction are spatially separated by about 60 mm which is the birdcage length. The high-concentrations of the E -field happen near the capacitor areas. Finally, based on MR images comparisons, the probe has not noticeable impact on the B_1 -field.

Conclusion

The described optical probe device is suitable to measure E -field within MRI tunnel without altering the B_1 -field. An excellent linearity and a good sensitivity of this optical-based probe were demonstrated. Besides, the probes can be used at different B_0 .

Acknowledgements

This work was supported by the Rhône-Alpes region and performed within the framework of LABEX-PRIMES (ANR-11-LABX-0063), and the program "Investissements d'Avenir" (ANR-11-IDEX-0007) operated by the French National Research Agency.

References

1. H. C. Taylor, et al. Experimental verification of numerically predicted electric field distributions produced by a radiofrequency coil. Phys. Med. Biol. 1997;42(7):1395-1402

2. A. Kangarlu, et al. Electric field measurements and computational modeling at ultrahigh-field MRI. *Magn. Reson. Imaging* 2007; 25(8):1222–1226.
3. P. Nordbeck, et al. Spatial distribution of RF-induced E-fields and implant heating in MRI. *Magn. Reson. Med.* 2008; 60(2):312–319.
4. B. Loader, et al. Evaluation of an optical electric field sensor for measurement of specific absorption rate (SAR) during magnetic resonance imaging. 2012 International Symposium on Electromagnetic Compatibility (EMC EUROPE); 2012
5. S. Reiss, et al. An optical setup for electric field measurements in MRI with high spatial resolution. *Phys. Med. Biol.* 2015; 60(11): 4355–4369

Figures




Fig 1. Schematic of the experimental setup for E -field measurements. The probe was fixed on a three-axis translation bench to move precisely the EO probe within the birdcage coil. The emitted and reflected lasers were transmitted by the same optical fiber. The signal processing unit contains optoelectronics components to analyze the modulation of the polarization of the laser. A photodiode was used to insure the conversion of the optical signal into an electrical signal. A spectrum analyzer and an oscilloscope were used to visualize the envelope of RF signal in the frequency and time domains.




Fig 2. The detected E -field is plotted as a function of the emitted B_1 -field. The B_1 -field was derived from the shape (Hermite equation) and amplitude (flip angle). The factors of proportionality between the EO signal and the E -fields were equal to 417714.9 m^{-1} and 76682.8 m^{-1} for transverse and longitudinal probes, respectively.




Fig 3. The three components of the E -field were measured using a transverse probe (TP) that measures the x and y -components of the E -field and a longitudinal probe (LP) that measures the z -component of the E -field. The probe was placed at $X=-2.5\text{cm}$ and $Y=0\text{cm}$ and moved with a 1 cm step along the z -axis. For the three curves, two high-concentrations located at $Z=5\text{cm}$ and $Z=11\text{cm}$ were found and the minimum was detected at $Z=8\text{cm}$. The distance between the two peaks is 60 mm which corresponds to the locations of the capacitors at distal and proximal sides of the birdcage.




Fig 4. The detected shape of three different RF pulses (Bloc pulse, Hermite pulse and Sinc3 pulse) with the optical probe. The measurements were performed at a fixed point. For the three pulses, the excitation time was 1 ms and the excitation power was 138 W . The E -field peaks were 139.91 V/m , 844.04 V/m and 753.21 V/m for the bloc pulse, Hermite pulse and Sinc3, respectively.




Fig 5. a) MR images of an orange with and without the insertion of the optical probe and the difference between the two images. b) The detected EO which is the envelope of the detected RF signal. 127 MHz corresponds to the proton's resonant frequency at 3.0 T. The orange was located in the center of the body coil; the electrical field peak was 198 V/m. The exact shape of the emitted excitation pulse was detected by the optical probe. The maximal E -field peak was detected near the location of the gradients and was equal to 15680 V/m.