

Parallelization on a 8-GPU server of iterative 3D reconstruction algorithm for dental X-ray tomography

Long Chen, Nicolas Gac, Thomas Rodet, Colombe Maury

▶ To cite this version:

Long Chen, Nicolas Gac, Thomas Rodet, Colombe Maury. Parallelization on a 8-GPU server of iterative 3D reconstruction algorithm for dental X-ray tomography . PUMPS summer school, Jul 2012, Barcelone, Spain. hal-01851992

HAL Id: hal-01851992 https://hal.science/hal-01851992v1

Submitted on 31 Jul 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.











Parallelization on a 8-GPU server of iterative 3D reconstruction algorithm for dental X-ray tomography

Long Chen^{1,2}, Nicolas Gac¹, Thomas Rodet¹, Colombe Maury²

1,Groupe Problèmes Inverses, Laboratoire des Signaux et Systèmes (L2S, UMR 8506 CNRS - SUPELEC - Univ Paris Sud 11) Supélec, Plateau de Moulon, 91192 Gif-sur-Yvette, FRANCE

2, Trophy, Carestream Dental, 4 rue F. Pelloutier - Croissy-Beaubourg 77534, France.

E-mails: lchen@lss.supelec.fr, nicolas.gac@lss.supelec.fr, thomas.rodet@lss.supelec.fr, colombe.maury@carestream.com

PUMPS summer school, july, 2-6, 2012, Barcelona, Spain

Abstract: In dental numeric 3D imaging, classic filtered backprojection (FBP) methods, for exemple, Feldkamp (FDK), is too sensible to the number of projections and noise, also limited in metal artifacts and beam-hardening reduction, so we use a 3D iterative least squares algorithm. Because of the huge size of 3D volumes (512³ to 1024³ voxels) and sinograms (512*512*360), it was too long to be applied in clinical medical imaging. However, thanks to the high performance of computing acceleration by GPU, its appearance in clinic is possible. In our iterative algorithm,we have two principal operations, projection and backprojection. According to our first resultats, the computation time of these operations on GPU is less in two orders (/100) than that on CPU, in particular for backprojection. With Multi-GPU (8 GPUs, Tesla C1060) in a Carri system's server, it decreases by a significant factor compared to single GPU. My Ph.D job consists of developping an iterative reconstruction method to reduce or remove the artifacts above, and obtain a segmentation of reconstructed volume as well.

Cone beam X-ray tomography

- ▶ Description: X-ray tomography consists of reconstruction of the unknown object from the detector data. We take it as an inverse problem.
- ▶ Mathematical modelling: Let f be an unknown object vector, g observation vector, H system matrix and ϵ noise vector.

$$g = Hf + \epsilon \tag{1}$$

► Method: an iterative least squares algorithm.

Iterative least squares(LS) algorithm

To estimate the unknown object f, we minimize the square errors as below.

$$\hat{f} = \operatorname{argmin}_f J_{LS}(f) = \operatorname{argmin}_f ||g - Hf||^2$$
 (2)

the $n+1^{th}$ estimation update $\hat{f}^{(n+1)}$:


$$\hat{f}^{(n+1)} = \hat{f}^{(n)} - \alpha^{(n)} \cdot \nabla J_{LS}(\hat{f}^{(n)}) = \hat{f}^{(n)} + \delta f_{LS}(n)$$
 (3)


where


$$\nabla J_{LS}(\hat{\mathbf{f}}^{(n)}) = 2\mathbf{H}^t(\mathbf{g} - \mathbf{H}\hat{\mathbf{f}}^{(n)}) \tag{4}$$

$$\alpha^{(n)} = \frac{1}{2} \frac{\|\nabla J_{LS}(\hat{f}^{(n)})\|^2}{\|H\nabla J_{LS}(\hat{f}^{(n)})\|^2}$$
(5)

Simpified Least squares algorithm steps, let g* be original data;







Regular sampling projection (H)

Voxel-driven backprojection (\mathbf{H}^t)

GPU & multi-GPU implementation of H & H^t operators


Compared to single GPU, the computation time of $H \& H^t$ operators with multi-GPU decreases by a significant factor roughly in proportion to the number of GPUs


Operation	Computation time(/s)	
	1GPU	8 GPUs
Backprojection	10.9	2.0(5)
Projection	213	28.5(7)

Reconstruction time of volume 1024³ from 256 projections [1]

Reconstruction by FDK & iterative LS


We used the real data of fake tooth, provided by Trophy, Carestream dental, and reconstructed the volume with Feldkamp and our iterative least squares algorithm.





Reconstructed slice of fake tooth (FDK)

Reconstructed slice (LS,iter60)



Profile of reconstructed slice (yellow lines in the top figures)

Future work

- ▶ Optimization of our iterative algorithm with a more accurate model
- ▶ Development of an iterative algorithm of metal artifacts reduction
- Method of beam-hardening reduction
- ► Method of simultaneous reconstruction and segmentation of volume
- ► Acceleration of all our iterative algorithms on multi-GPU

[1] **N. Gac et al**, "Multi GPU parallelization of 3D bayesian CT algorithm", Forum on recent developments in Volume Reconstruction techniques 2011, Poitiers, France.