



A volume-based aspiration method to estimate in-vivo soft tissues stiffness: evaluation of the device with silicone samples

Seyed Ali Elahi, Nathanaël Connesson, Yohan Payan

► To cite this version:

Seyed Ali Elahi, Nathanaël Connesson, Yohan Payan. A volume-based aspiration method to estimate in-vivo soft tissues stiffness: evaluation of the device with silicone samples. 16th European Mechanics of Materials Conference (EMMC 2016), Mar 2018, Nantes, France. pp.464-465. hal-01930219

HAL Id: hal-01930219

<https://hal.science/hal-01930219>

Submitted on 21 Nov 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

A volume-based aspiration method to estimate in-vivo soft tissues stiffness: evaluation of the device with silicone samples

Seyed Ali Elahi^{*1}, Nathanael Connesson^{†1}, and Yohan Payan¹

¹Techniques de l'Ingénierie Médicale et de la Complexité - Informatique, Mathématiques et Applications [Grenoble] (TIMC-IMAG) – Centre National de la Recherche Scientifique : UMR5525, Université Grenoble Alpes – Domaine de la Merci - 38706 La Tronche, France

Abstract

Computer Assisted Medical Interventions (CAMI) involving soft tissues require devices that model such tissues in order to estimate the way they are going to be deformed and/or resected by the surgical gesture. The corresponding biomechanical patient-specific models require an estimation of the constitutive behavior of the soft tissues. Since the mechanical behavior of living tissues varies between *in-vivo* and *ex-vivo* conditions, it is important for the CAMI devices to offer an *in-vivo* estimation with non-traumatic measurements that should undergo sterile conditions. For this purpose, among all the methods proposed in the literature, aspiration/suction is the most widely used technique due to its simplicity and robustness.

For such a technique, a device with a hole is put in contact with the soft tissue while a negative pressure aspires part of this tissue. Knowing the relationship between the negative pressure and the aspired tissue height, an inverse problem is then solved to identify the material mechanical properties. In the literature, the apex height is usually measured with a camera and a mirror or a prism, which induces design difficulties, in particular in regards on the required sterilization process for *in-vivo* measurements. This paper introduces a new method that replaces the optical apex height measurement with a measurement of the aspired tissue volume. The method is referred to as "rate-based method". The camera, mirror and all electronic parts are not required, which makes our device the simplest, lightest and cheapest one could achieve. In particular, this simplification enables the system to meet the severe sterilizations constraints that can be present for some surgeries. Indeed, the proposed device is only composed of a cylindrical plastic chamber (with a hole at its basis), some connection tubes, a syringe pump, a manometer and an aspiration chamber. The idea is to aspirate the tissue inside the chamber using the syringe pump, while measuring the negative pressure and the corresponding removed volume. Such a volume is due to both the aspired tissue volume inside the chamber and the volume changes in the device (air expansion and elasticity of the connections, tubes, syringe, *etc.*). An off-line calibration process is defined to differentiate the volume changes due to tissue aspiration from the volume changes due to the compressibility of the device (tubes, connections, syringe).

Our device was evaluated with the estimation of the constitutive behavior of two bulk cylindrical silicone samples, one being slightly softer than the other one. For each silicone, thin

^{*}Speaker

[†]Corresponding author: nathanael.connesson@gmail.com

flat samples were extracted to characterize their constitutive behavior using a classical traction experiment (MTS Criterion machine, Model 41) and an equibiaxial extension (bulge test). Such experiments respectively provide Young modulus of 16.82kPa and 24.87kPa for silicones #1 and #2.

The aspiration tests on each silicone sample were repeated nine times to check the reproducibility of the results. The inverse method to estimate the constitutive behavior assumed a hyper-elastic Gent model with material constants computed using an updated FE simulation. The results validate the ability of our method to discriminate the two silicones with equivalent Young moduli estimated to 18.35kPa and 26.52kPa for silicones #1 and #2.

Keywords: Aspiration Method, In, vivo Measurement, Soft Tissues Characterization, Experimental Mechanics, Inverse Characterization