



HAL
open science

Intraluminal Ultrasonic Palpation Imaging Technique Revisited for Anisotropic Characterization of Healthy and Atherosclerotic Coronary Arteries: A Feasibility Study

Armida Gómez, Antoine Tacheau, Gerard Finet, Manuel Lagache, Jean-Louis Martiel, S. Lefloch, Saami K. Yazdani, Alex Elias-Zuñiga, Roderic I Pettigrew, Guy Cloutier, et al.

► To cite this version:

Armida Gómez, Antoine Tacheau, Gerard Finet, Manuel Lagache, Jean-Louis Martiel, et al.. Intraluminal Ultrasonic Palpation Imaging Technique Revisited for Anisotropic Characterization of Healthy and Atherosclerotic Coronary Arteries: A Feasibility Study. *Ultrasound in Medicine & Biology*, 2019, 45 (1), pp.35-49. 10.1016/j.ultrasmedbio.2018.08.020 . hal-01943969

HAL Id: hal-01943969

<https://hal.science/hal-01943969>

Submitted on 4 Dec 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

INTRALUMINAL ULTRASONIC PALPATION IMAGING TECHNIQUE REVISITED FOR ANISOTROPIC CHARACTERIZATION OF HEALTHY AND ATHEROSCLEROTIC CORONARY ARTERIES: A FEASIBILITY STUDY

ARMIDA GÓMEZ,^{*} ANTOINE TACHEAU,^{*} GÉRARD FINET,[†] MANUEL LAGACHE,^{‡,§} JEAN-LOUIS MARTIEL,^{*}
SIMON LE FLOCH,[¶] SAAMI K. YAZDANI,^{||} ALEX ELIAS-ZUÑIGA,[#] RODERIC I. PETTIGREW,^{**}
GUY CLOUTIER,^{††} and JACQUES OHAYON^{*,§}

^{*}Laboratory TIMC-IMAG/DyCTiM, UGA, CNRS UMR 5525, Grenoble, France; [†]Department of Hemodynamics and Interventional Cardiology, Hospices Civils de Lyon and Claude Bernard University Lyon1, INSERM Unit 886, Lyon, France;

[‡]Laboratory SYMME, SYMME, University Savoie Mont-Blanc, France; [§]Polytech Annecy-Chambéry, University Savoie Mont-Blanc, Le Bourget du Lac, France; [¶]Laboratory LMGc, CNRS UMR 5508, University of Montpellier II, Montpellier, France;

^{||}Department of Mechanical Engineering, University of South Alabama, Mobile, Alabama, USA; [#]Department of Mechanical Engineering Instituto Tecnológico y de Estudios Superiores de Monterrey, Campus Monterrey, Monterrey, Mexico; ^{**}Texas A&M University Engineering Health (EnHealth), Houston, Texas, USA; and ^{††}Laboratory of Biorheology and Medical Ultrasonics, University of Montreal Hospital Research Center (CRCHUM), Montréal, Québec, Canada

Abstract—Accurate mechanical characterization of coronary atherosclerotic lesions remains essential for the *in vivo* detection of vulnerable plaques. Using intravascular ultrasound strain measurements and based on the mechanical response of a circular and concentric vascular model, E. I. Céspedes, C. L. de Korte and A. F. van der Steen developed an elasticity-palpography technique in 2000 to estimate the apparent stress-strain modulus palpogram of the thick subendoluminal arterial wall layer. More recently, this approach was improved by our group to consider the real anatomic shape of the vulnerable plaque. Even though these two studies highlighted original and promising approaches for improving the detection of vulnerable plaques, they did not overcome a main limitation related to the anisotropic mechanical behavior of the vascular tissue. The present study was therefore designed to extend these previous approaches by considering the orthotropic mechanical properties of the arterial wall and lesion constituents. Based on the continuum mechanics theory prescribing the strain field, an elastic anisotropy index was defined. This new anisotropic elasticity-palpography technique was successfully applied to characterize ten coronary plaque and one healthy vessel geometries of patients imaged *in vivo* with intravascular ultrasound. The results revealed that the anisotropy index-palpograms were estimated with a good accuracy (with a mean relative error of $26.8 \pm 48.8\%$) compared with ground true solutions. (E-mail: jacques.ohayon@univ-grenoble-alpes.fr)

Key Words: Atherosclerosis, Vulnerable plaques, Biomechanics, Orthotropic material, Coronary disease, Plaque detection, Linear incremental elasticity, Inverse problem.

INTRODUCTION

Morphologic and mechanical characterization of coronary atherosclerotic lesions remains an essential step for the diagnosis of rupture-prone plaques and subsequent thrombosis (Alsheikh-Ali et al. 2010; Arbab-Zadeh et al. 2012; Benjamin et al. 2017; Fleg et al. 2012; Libby et al. 2002; Naghavi et al. 2003).

To detect vulnerable plaques (VPs) in a timely manner and prevent adverse coronary events, several endovascular imaging techniques are available to clinicians or in development including intravascular ultrasound (IVUS) (Cheng et al. 2014; Rioufol et al. 2002), optical coherence tomography (Sinclair et al. 2015), magnetic resonance imaging (Cavalcante and Larose 2016; Larose et al. 2005) and near-infrared spectroscopy (Caplan et al. 2006; Madder et al. 2016), as well as emergent hybrid techniques that combine two or more imaging modalities (Abran et al. 2014; Bourantas et al.

Address correspondence to: Jacques Ohayon, Laboratory TIMC-IMAG/DyCTiM, UGA, CNRS UMR 5525, Grenoble, France. E-mail: jacques.ohayon@univ-grenoble-alpes.fr

2017). Despite these technological advances, identifying which VPs will rupture and induce an acute event remains a challenging task.

Previous biomechanical studies have identified peak cap stress amplitude as a strong indicator of plaque vulnerability to rupture (Finet et al. 2004; Gijssen and Migliavacca 2014; Kok et al. 2016; Ohayon et al. 2008; Tang et al. 2009). Others have established that plaque rupture risk is also attributed to additional factors such as the presence of microcalcifications in the thin cap (Cardoso and Weinbaum 2014; Kelly-Arnold et al. 2013; Thondapu et al. 2017), plaque morphology (Akyildiz et al. 2016; Ohayon et al. 2008) and hemodynamics (Ahmadi et al. 2016; Gijssen et al. 2008; Morbiducci et al. 2016), as well as tissue mechanical properties of the atherosclerotic lesion (Guo et al. 2017; Finet et al. 2004; Ohayon et al. 2014).

Although plaque morphology and cap thickness can be directly extracted from endovascular images, the stress distribution within the plaque is impossible to obtain using current medical imaging technologies. However, the local deformation (strain) of the arterial wall and plaque caused by the cyclic intraluminal pressure can be calculated from a series of ultrasound images (Doyley 2012; Golemati et al. 2016; Maurice et al. 2004b). Based on the postulate that local variations in elastic properties of tissues are generally correlated with pathologic regions, strain elastography—a strain field-based technique—was developed. This approach was first introduced by Ophir's group (Céspedes et al. 1993; Ophir et al. 1991) to map the mechanical strain response to an applied force. Nevertheless, complex non-linear geometric effects do not allow a direct correlation between the measured spatial strain field and the mechanical property of the tissue. By using the IVUS strain field inside atherosclerotic plaques (De Korte et al. 2002; Doyley 2012; Maurice et al. 2004a, 2007), several elasticity reconstruction approaches have been developed to characterize the entire lesion (Baldeusung et al. 2007; Deléaval et al. 2013; Doyley 2012; Le Floc'h et al. 2009, 2010, 2012; Nayak et al. 2017; Poree et al. 2017; Richards and Doyley 2011; Richards et al. 2015; Tacheau et al. 2016) and to delineate a subendoluminal layer of the arterial wall called the palpography domain (Céspedes et al. 2000; Deléaval et al. 2013; Doyley et al. 2001; Schaar et al. 2005; Widynski et al. 2014). Both techniques aim to characterize the mechanical properties of atherosclerotic lesions and ultimately differentiate vulnerable from stable plaques.

In the latter elasticity-palpography technique originally developed by Céspedes et al. (2000) and Doyley et al. (2001), and revisited by our group to account for the anatomic shape of the VP (Deléaval et al. 2013), the apparent local stiffness of the subendoluminal arterial

wall layer was computed assuming that all plaque constituents are elastically isotropic. Even though these studies highlighted promising results, they failed to address the anisotropic mechanical behavior of the arterial wall and atherosclerotic lesion (Holzapfel et al. 2005).

Therefore, the purpose of this study was to extend the theoretical framework of the palpography technique implemented by Deléaval et al. (2013) to account for the anisotropic mechanical behavior of the arterial wall. Based on the continuum mechanics theory prescribing the radial strain field, a new anisotropic elasticity-palpography technique (AE-PT) was successfully applied to five modeled plaque geometries, ten coronary lesions and one healthy vessel of patients imaged *in vivo* with IVUS. An elastic anisotropy index (AI) was extracted.

METHODS

Using images of patients who underwent coronary IVUS, cross-sectional geometries of ten atherosclerotic plaques and one healthy vessel were manually extracted based on the echogenicity aspect. Additionally, five idealized plaque morphologies were designed and simulated to validate the proposed AE-PT and to investigate the effects of cap thickness and palpography domain size on computed palpograms. Then, a 3-D finite-element (FE) analysis was conducted for each plaque geometry to compute the exact displacement and strain field distributions. To validate the robustness of palpograms, we synthesized radiofrequency (RF) echo frames by combining the computed FE displacement field with an ultrasound simulation model based on the Field II software (Jensen and Svendsen 1992). For all simulated IVUS RF sequences, the radial strain elastograms were estimated with the Lagrangian speckle model estimator (Maurice et al. 2004a). These radial strain elastograms were used to obtain the inverse solution of palpograms for the IVUS mechanical characterization of anisotropic vascular tissues.

IVUS study and plaque geometries

Patient population. Vulnerable plaque geometries were extracted from IVUS images acquired on patients referred for percutaneous coronary intervention at the Louis Pradel Cardiology Hospital of Lyon after a first acute coronary syndrome with troponin I elevation. Patients' consent was obtained, and the study was approved by the ethical institutional review board of the Cardiology Hospital.

Intravascular ultrasound imaging. A set of one healthy and ten non-ruptured VP morphologies were collected from IVUS scans of the coronary arteries

according to the protocol described by Rioufol et al. (2002). Cross-sectional IVUS images of the atherosclerotic lesions were recorded at sites exhibiting the thinner fibroatheroma cap or at sites with the largest degree of stenosis. IVUS scans were performed with the iLab platform (Boston Scientific, Watertown, MA, USA) equipped with 40-MHz catheters (Atlantis SR Pro 3.6 F, Boston Scientific) and a radial spatial resolution in B-mode of approximately $90 \mu\text{m}$ (Chopard et al. 2010).

IVUS image analysis and plaque morphologic measurements. IVUS echogenicity aspects were used to characterize VP components according to the following observations: (i) highly hypo-echogenic components (or anechogenic zones), suggestive of lipid or cellular deposition; or (ii) homogeneous reflective components, suggestive of fibrosis (Di Mario et al. 1998). A manual segmentation procedure using ImageJ software (National Institutes of Health [NIH], Bethesda, MD, USA) was performed by a cardiologist to extract the internal (*i.e.*, lumen) and external contours of the coronary cross section, as well as the contours of each plaque component (*i.e.*, necrotic core and dense fibrosis).

The morphology of each IVUS lesion image was quantified for plaque area (Pla_{area} , mm^2), lumen area (Lu_{area} , mm^2), soft necrotic core area, dense fibrosis area and degree of stenosis ($\text{Stenos}_{\text{deg}}$, %) as $100 \times \text{Pla}_{\text{area}} / (\text{Pla}_{\text{area}} + \text{Lu}_{\text{area}})$ and cap thickness ($\text{Cap}_{\text{thick}}$, mm), which was defined as the shortest distance between the lumen and the necrotic core.

Idealized geometries. One concentric and one eccentric idealized circular plaque geometries with homogeneous anisotropic and quasi-incompressible atherosclerotic lesions were simulated to validate the theoretical background of the AE-PT (Fig. 1). Moreover, to investigate the effects of cap thickness ($\text{Cap}_{\text{thick}}$) on the resulting anisotropic elasticity palpogram, plaque geometry acquired on a single patient (plaque #2 with a $\text{Cap}_{\text{thick}}$ of $93 \mu\text{m}$) (Fig. 2) was modified by translating the soft necrotic inclusion radially (*i.e.*, along the axis going from the center of mass of the vessel lumen to the center of mass of the necrotic core). Thus, three idealized plaque geometries with $\text{Cap}_{\text{thick}}$ values of 200, 400 and $600 \mu\text{m}$ were modeled (plaques #2A–2C) (Fig. 2).

Forward problem: Spatial displacement and strain field distributions

3-D finite-element simulations. The 2-D digitized contours obtained were transferred into Comsol software (Structural Mechanics Module, Version 4.3b, Comsol, France). To compute the deformed shape, a 3-D static FE simulation was conducted on each thin-sliced

volumetric plaque (slice thickness = 0.1 mm) obtained by extrusion of the cross-sectional plaque profile in the longitudinal direction of the artery. Each geometry was meshed with approximately 5000 six-node prism elements. The barycenter (*i.e.*, the center of mass) of the luminal area was used as the origin of the cylindrical coordinate system (R, θ, Z).

Instantaneous pressures were not recorded during the clinical IVUS exploration of the coronary arteries. Therefore, a blood pressure differential amplitude ΔP between 0.3 and 1 kPa (*i.e.*, 2.2–7.5 mmHg) was assumed, corresponding to a realistic pressure gradient occurring between two successive images (Im^i and Im^{i+1}) of the IVUS sequence (Le Floc’h et al. 2009, 2012). In our computational approach, the deformation of the atherosclerotic plaque between Im^i and Im^{i+1} was obtained by considering the geometry of the lesion at Im^i as the reference configuration. Therefore, FE simulations were performed in linear elasticity with a small loading blood pressure differential ΔP . A free boundary condition was assumed at the external wall of the artery, and sliding conditions were imposed on the two cross sections of the vessel segment ($R-\theta$ planes, $Z=0$ and 0.1 mm). Such boundary conditions constrain the deformation to take place in the planes where Z is equal to a constant. Additionally, displacement limitations were placed upon nodes located on the four cardinal directions of the external wall of the artery to avoid rigid rotations. These latter boundary conditions allow free-radial expansion of the vessel during pressure loading, resulting in more realistic vessel inflation kinematics.

The fibrotic plaque medium and the healthy coronary arterial wall were considered as orthotropic and quasi-incompressible media (Appendix I), with initial mechanical properties coming from measurements (Chagnon et al. 2017; Holzapfel et al. 2005) of the pathologic coronary thick intima layer and the adventitia layer, respectively. The mechanical properties of the dense fibrosis and soft necrotic cores were considered as isotropic and quasi-incompressible (Lee et al. 1991; Loree et al. 1994). Young’s moduli, Poisson ratios and shear moduli for all media are summarized in Table 1.

Simulation of ultrasound images. The open source software Field II (Jensen and Svendsen 1992) was modified to simulate the IVUS acoustic field (Fromageau et al. 2003; Nayak et al. 2017; Poree et al. 2017). Briefly, biological tissues were characterized by their acoustic scattering amplitude. Acoustic models were built by randomly distributing scatterers over each non-deformed medium. The relative acoustic scatterer amplitudes of blood, necrotic core, fibrotic plaque and dense fibrosis were set to 10%, 15%, 35% and 60%, respectively




Fig. 1. Performance of the anisotropic elasticity-palpography technique on idealized homogeneous concentric and eccentric plaque models (plaques #0 and 1, respectively). (a, d) Plaque geometries (*red contours*), palpography domains Ω_{palpo} (*blue contours*) and estimated radial strain elastograms. (b, e) Intravascular ultrasound images with anisotropy index-palpograms $AI(\theta)$. (c, f) Comparisons between exact and estimated $AI(\theta)$ palpograms. fi = fibrotic plaque region.




Fig. 2. Effect of cap thickness (Cap_{thick}) on the resulting anisotropic elasticity palpogram $AI(\theta)$. Simulations were performed with plaque #2. (a, c, e, g) Plaque geometries (*red contours*), palpography domains (*blue contours*) and estimated radial strain elastograms. (b, d, f, h) Comparisons between exact and estimated $AI(\theta)$ palpograms. (i) Comparison between all estimated AI -palpograms. Areas shaded in *yellow* indicate the location of the necrotic core. fi = fibrotic plaque region; nc = necrotic core.

Table 1. Orthotropic and isotropic mechanical properties and anisotropy index values assigned for the healthy wall, fibrous plaque, dense fibrosis and soft necrotic core inclusions

	Young's modulus (kPa)			Anisotropy index (kPa)	Reference (medium)
	E_R	E_θ	E_Z		
Healthy wall	120	80	170	121.8	Holzzapfel et al. (2005) (adventitia)
Pathologic wall					
Fibrous plaque	360	230	330	384.3	Holzzapfel et al. (2005) (thick intima)
Dense fibrosis*		1500		2000	Lee et al. (1991)
Necrotic core*		5		6.7	Loree et al. (1994)

* Isotropic ($E_R = E_\theta = E_Z$).

(Cardoso et al. 2014; Le Floc'h et al. 2010; Maurice et al. 2004a). First, an IVUS image of the non-deformed state (*i.e.*, Im^i) was created. Then the displacement field resulting from the FE simulation was used to track the location of scatterers and to create the post-deformation IVUS image (*i.e.*, Im^{i+1}). A single rotating piezo-electric element (600- μ m diameter, 40-MHz central emission frequency) was modeled. An incremental circumferential rotation step of 0.7° was set to get 512 lines of RF signal for IVUS image reconstruction. A sampling frequency of 400 MHz was mimicked to represent the RF signal (SVMI HD IVUS system, ACIST Medical Systems, Eden Prairie, MN, USA). The arterial wall was also adjusted to set the centroid of the lumen coincident with the probe location.

Intravascular strain elastography. Sequences of simulated RF images were processed using the Lagrangian speckle model estimator. Radial strain fields were computed using a two-step strain imaging algorithm (Majdoulina et al. 2014; Maurice and Bertrand 1999; Maurice et al. 2004a, 2007). First, a block matching method provides local registration on overlapping sub-windows (measurement windows [MWs]). It allows compensating for rigid translation movements using a 2-D cross correlation coefficient. Once MW locations have been registered, the 2-D-deformation matrix is estimated. Estimation is based on the resolution of a first-order optical flow equation for each MW (Majdoulina et al. 2014). Sizes of MWs in radial and circumferential directions were set to 101 pixels (about 194.5 μ m) and 21 pixels (about 14.8°). Shift between MWs was set to preserve an overlap of 90% in the radial and circumferential directions (Le Floc'h et al. 2010; Maurice et al. 2004a). Moreover, a sensitivity study was conducted to investigate the influence of MW size on the accuracy of the AI-palpograms. In agreement with the IVUS imaging studies of Céspedes et al. (2000), Doyley et al. (2001), Deléaval et al. (2013), Widynsky et al. (2014) and Tacheau et al. (2016), a palpography domain thickness

close to 1 mm was used for our simulations, unless otherwise stated.

Inverse problem: The anisotropic elasticity-palpography technique

To account for the orthotropic mechanical properties of the vascular tissue, we defined an elastic anisotropy index (AI) as the ratio of the averaged difference between the radial and circumferential stresses (σ_{RR} and $\sigma_{\theta\theta}$) over the averaged radial strain (ε_{RR}) along the radial axis:

$$AI(\theta) = \frac{\left| \int_{R_i(\theta)}^{R_o(\theta)} \{ \sigma_{RR}(R, \theta) - \sigma_{\theta\theta}(R, \theta) \} dR \right|}{\varepsilon(\theta)} \quad (1)$$

with

$$\varepsilon(\theta) = \left| \int_{R_i(\theta)}^{R_o(\theta)} \varepsilon_{RR}(R, \theta) dR \right| \quad (2)$$

where $R_i(\theta)$ and $R_o(\theta)$ are the inner and outer radii of the considered palpography domain, respectively. Appendix II provides the details of the computation of this index.

Note that the amplitude of $AI(\theta)$ remains constant and equal to

$$AI(\theta) = E_R^0 \left[1 - \frac{1}{4} \left(1 - \frac{E_R^0}{E_\theta^0} + \frac{E_R^0}{E_Z^0} \right) \left(1 + \frac{E_Z^0}{E_R^0} - \frac{E_Z^0}{E_\theta^0} \right) \right]^{-1} \quad (3)$$

when considering a homogeneous pathologic lesion made of the same incompressible orthotropic medium (*i.e.*, when $E_R(R, \theta) = E_R^0$, $E_\theta(R, \theta) = E_\theta^0$ and $E_Z(R, \theta) = E_Z^0$), whatever the plaque geometry and the palpography domain Ω_{palpo} considered (with $R_i(\theta) \leq R \leq R_o(\theta)$).

For a homogeneous incompressible transversely isotropic lesion (*i.e.*, when $E_R(R, \theta) = E_R^0$ and $E_Z(R, \theta) = E_\theta(R, \theta) = E_\theta^0$), the AI amplitude is reduced to the constant

$$AI = \frac{4E_R^0{}^2}{4E_R^0 - E_\theta^0} \quad (4)$$

and to $AI = 4E/3$ for a homogeneous incompressible isotropic plaque (*i.e.*, when $E_R(R, \theta) = E_\theta(R, \theta) = E_Z(R, \theta) = E$).

Table 1 gives the amplitudes of the AI for the considered mechanical properties of the adventitia (healthy arterial wall), thick intima (fibrotic plaque), dense fibrosis and soft necrotic media.

Method used to verify the accuracy of the anisotropic elasticity-palpography technique. Two anisotropic index palpograms were computed: the first, $AI(\theta)$, with the estimated radial strain elastogram, and the second, $AI^*(\theta)$, considering the exact radial strain map. To investigate the accuracy of each palpogram (*i.e.*, $AI(\theta)$ and $AI^*(\theta)$), we compared them with the exact circumferential distribution of the averaged anisotropic material constant $C(R, \theta)$ along the radial axis:

$$AI^{\text{exact}}(\theta) = \frac{1}{R_o(\theta) - R_i(\theta)} \int_{R_i(\theta)}^{R_o(\theta)} C(R, \theta) dR \quad (5)$$

and we computed two mean relative anisotropy index errors (MR_{error}^{AI} and $MR_{\text{error}}^{AI^*}$):

$$MR_{\text{error}}^{K(\theta)} (\%) = \frac{100}{2\pi} \int_{\theta=0}^{2\pi} \left| \frac{K(\theta) - AI^{\text{exact}}(\theta)}{AI^{\text{exact}}(\theta)} \right| d\theta, \quad (6)$$

with $K(\theta) = AI(\theta)$ or $AI^*(\theta)$

$MR_{\text{error}}^{AI^*}$ indicates the best accuracy that could be reached when using the new anisotropic palpography technique, and MR_{error}^{AI} quantifies the cumulative error induced by both the proposed palpography technique and the estimated radial strain elastogram. Therefore, to quantify the error induced only by the radial strain elastogram estimation on the AI-palpogram, we defined the following mean relative error ($MR_{\text{error}}^{AI/AI^*}$):

$$MR_{\text{error}}^{AI/AI^*} (\%) = \frac{100}{2\pi} \int_{\theta=0}^{2\pi} \left| \frac{AI(\theta) - AI^*(\theta)}{AI^*(\theta)} \right| d\theta \quad (7)$$

RESULTS

IVUS study

Seven non-ruptured VPs with soft necrotic cores, one homogeneous plaque, two atherosclerotic lesions with dense fibrosis inclusions and one non-pathologic cross section were identified after extensive IVUS scanning. The geometric characteristics of these 11 cross sections scanned *in vivo* are summarized in Table 2. All computed AI-palpograms presented in the figures were obtained from radial strain elastograms estimated from sequences of simulated RF images.

Performance of the anisotropic elasticity-palpography technique

Validation cases and AI-palpogram accuracies.

Simulations were performed on the idealized concentric and eccentric plaque geometries with homogeneous orthotropic and quasi-incompressible fibrotic atherosclerotic plaques (Fig. 1). AI-palpograms for the concentric and eccentric plaques remain almost constant with mean values (\pm standard deviation [SD]) of 396 ± 9 and 394 ± 14 kPa, both close to the expected theoretical value ($AI = 384.3$ kPa, Table 1) given by eqn (3)—even in the high-strained free-plaque arc length of the eccentric plaque geometry, which is the thinnest region of the palpography domain (*i.e.*, between the angular positions 180° and 360° , Figs. 1e, f). The small amplitudes of the mean relative anisotropy index errors ($MR_{\text{error}}^{AI} = 3.08\%$ and 3.70% , Table 3), together with the small amplitudes of the standard deviations (SD = 1.62% and 2.08% , Table 3), highlight the robustness of the AE-PT algorithm with respect to the IVUS RF sequences when estimating AI-palpograms for concentric and eccentric plaque morphologies, respectively. More generally, the results revealed that all 18 AI-palpograms were estimated with reasonable accuracy (with a mean relative error $MR_{\text{error}}^{AI} = 26.79 \pm 48.84\%$, Table 3) when compared with the exact ones and with a good robustness with respect to the estimated strain-elastogram, as reflected by the mean relative error $MR_{\text{error}}^{AI/AI^*} = 10.24 \pm 14.43\%$ (Table 3).

Detection and characterization of vulnerable plaques with soft inclusions. AI-palpograms were computed for VPs with one (plaques #2, 2A–2C and 12–15, Figs. 2–4) and two distinct (plaques #16 and 17, Fig. 5) necrotic cores. These AI-palpograms were estimated with a mean relative error of $34.36 \pm 57.03\%$ (Table 3). Locations of the necrotic core sites were accurately detected using the AE-PT. The AI-palpogram amplitudes were found close to the AI value of the fibrotic plaque (*i.e.*, close to 384.3 kPa, Table 1) except at soft inclusion locations, for which the AI amplitudes were found to be significantly lower, as expected (Figs. 2–4).

Detection and characterization of stable plaques with dense fibrosis inclusions. For one stable plaque (plaque #19, Fig. 6) with one isolated dense fibrosis inclusion, the proposed palpography technique detected and distinguished the fibrotic plaque from dense fibrosis. Compared with the exact value, the estimated AI amplitude at the dense fibrosis inclusion location was underestimated by almost a factor of 2.

Table 2. Characteristics of the IVUS-detected and modelled atherosclerotic plaques and healthy artery used to test the performance of the proposed anisotropic elasticity–palpography technique.

Plaque No.	Origin of geometry	Cap thickness* (μm)	Area (mm^2)			Lumen	Stenosis (%)
			Necrotic core	Dense fibrosis	Plaque		
0	Model	—	—	—	14.14	1.77	89
1	Model	—	—	—	25.13	3.14	89
2	IVUS	93	1.36	—	8.74	2.79	76
2A	Model	200	1.36	—	8.74	2.79	76
2B	Model	400	1.36	—	8.74	2.79	76
2C	Model	600	1.36	—	8.74	2.79	76
Healthy 1	IVUS	—	—	—	8.45	11.87	42
11	IVUS	—	—	—	9.88	7.59	57
12	IVUS	<90 (55)	3.48	—	14.24	6.06	70
13	IVUS	<90 (63)	4.41	—	15.48	4.28	78
14	IVUS	<90 (49)	1.97	—	19.38	6.80	74
15	IVUS	<90 (48)	1.89	—	17.29	8.44	67
16	IVUS	100	1.41/0.46	—	19.50	7.67	72
17	IVUS	<90 (65)	2.20/0.53	—	29.53	4.34	87
18	IVUS	<90 (60)	—	1.36	6.34	5.99	51
19	IVUS	<90 (88)	—	0.90	5.31	5.35	50

IVUS = intravascular ultrasound.

* Cap thickness was randomly assigned (value in parentheses) when found to be under the limit of the B-mode spatial resolution obtained with the 40-MHz IVUS catheter (*i.e.*, <90 μm).

Detection and characterization of homogeneous stable plaques and healthy vessel wall. AI-palpograms were also computed for one eccentric homogeneous stable plaque (plaque #11, Fig. 7) and one healthy coronary wall (healthy #1, Fig. 8). The AE-PT successfully characterized the orthotropic mechanical properties of the pathologic lesion (made of fibrotic plaque tissue) and of the non-pathologic coronary artery. The two AI-palpograms were found almost constant with mean values ($\pm\text{SD}$) of 360 ± 15 and 134 ± 9 kPa, close to the

expected theoretical values: AI = 384.3 and 121.8 kPa for the pathologic and healthy arteries, respectively (Table 1).

Effects of Cap_{thick} on estimated AI-palpogram.

For a given palpography domain, the Cap_{thick} significantly affects the AI-palpogram at the soft inclusion location (plaques #2 and 2A–2C, Fig. 2). Simulations performed with the idealized VP models in which the Cap_{thick} increases from 93 to 600 μm revealed that the

Table 3. Accuracy of the proposed anisotropic elasticity-palpography technique and effect of the estimated radial strain elastogram (computed with a measurement window size of 101×21 pixels) on the estimated AI-palpograms*

Plaque No.	NC	DF	H	MR _{error} ^{AI} (%)	MR _{error} ^{AI} (%)	MR _{error} ^{AI/AI*} (%)
0	—	—	✓	0.22 \pm 0.14	3.08 \pm 1.62	2.9 \pm 1.49
1	—	—	✓	3.72 \pm 1.72	3.7 \pm 2.08	4.24 \pm 3.2
2	✓	—	—	4.86 \pm 4.53	13.41 \pm 9.45	8.76 \pm 6.06
2A	✓	—	—	5.97 \pm 5.26	11.13 \pm 7.9	5.53 \pm 4.28
2B	✓	—	—	3.74 \pm 2.29	7.84 \pm 4.35	4.22 \pm 3.06
2C	✓	—	—	1.91 \pm 1.55	5.05 \pm 3.85	4.03 \pm 2.79
Healthy 1	—	—	✓	0.65 \pm 0.5	10.03 \pm 7.67	10.49 \pm 7.66
11	—	—	✓	2.45 \pm 2.24	6.66 \pm 3.55	5.91 \pm 3.43
12	✓	—	—	21.71 \pm 27.64	33.16 \pm 52.64	16.4 \pm 19.1
13 (0.75)	✓	—	—	66.6 \pm 52.63	83.23 \pm 71.73	16.1 \pm 21.47
13	✓	—	—	59.43 \pm 57.53	74.14 \pm 67.96	15.98 \pm 18.61
13 (1.25)	✓	—	—	36.96 \pm 41.17	46.45 \pm 55.68	15.84 \pm 17.44
14	✓	—	—	35.21 \pm 62.45	44.87 \pm 78.06	9.96 \pm 11.05
15	✓	—	—	35.54 \pm 73.87	45.19 \pm 92.07	8.49 \pm 7.47
16	✓	—	—	8.72 \pm 9.57	14.46 \pm 17.38	8.11 \pm 9.05
17	✓	—	—	24.86 \pm 22	33.3 \pm 33.19	8.36 \pm 9.24
18	—	✓	—	11.32 \pm 12.96	35.34 \pm 31.27	29.5 \pm 31.75
19	—	✓	—	3.76 \pm 6.38	11.25 \pm 8.91	9.43 \pm 6.69
Mean \pm SD				18.2 \pm 37.99	26.79 \pm 48.84	10.24 \pm 14.43

AI = anisotropy index; DF = dense fibrosis; H = homogeneous; NC = necrotic core; SD = standard deviation.

* See eqns (6) and (7) for definitions of the mean relative errors MR_{error}^{AI}, MR_{error}^{AI*} and MR_{error}^{AI/AI*}. Palpography thickness is 1 mm unless otherwise indicated in parentheses beside the plaque number.




Fig. 3. Performance of the anisotropic elasticity palpography technique in detection of vulnerable plaques with one necrotic core (plaques #12, 14 and 15). (a, d, g) Plaque geometries (*red contours*), palpography domains (*blue contours*) and estimated radial strain-elastograms. (b, e, h) IVUS images with anisotropy index-palpograms $AI(\theta)$. (c, f, i) Comparisons between exact and estimated $AI(\theta)$ palpograms. Areas shaded in *yellow* indicate the locations of the necrotic cores. fi = fibrotic plaque region; nc = necrotic core.

mean AI amplitude at the location of the necrotic core increased from 139 ± 74 to 300 ± 65 kPa (Fig. 2).

Effects of the palpography domain size on estimated AI amplitude. We computed the AI-palpograms for the same VP considering three palpography domains with maximum thicknesses of 0.75, 1 and 1.25 mm (plaque #13, Fig. 4). Increasing the palpography domain thickness from 0.75 to 1.25 mm improved the accuracy of the AI palpogram from $MR_{\text{error}}^{\text{AI}} = 83.23 \pm 71.73\%$ to $MR_{\text{error}}^{\text{AI}} = 46.45 \pm 55.68\%$, respectively. The location of the necrotic core site was accurately detected using all palpography domains.

Sensitivity of AI-palpograms to the accuracy of the estimated strain elastograms. Because the MW size is one of the main parameters driving the accuracy of the computed strain elastograms, we restricted our sensitivity analysis to this parameter. These

computations were performed on the stable plaque #18 with one dense fibrosis inclusion and by considering three MW sizes: 75×15 pixels (*i.e.*, $144.4 \mu\text{m} \times 10.6^\circ$), 101×21 pixels (*i.e.*, $194.5 \mu\text{m} \times 14.8^\circ$) and 121×31 pixels (*i.e.*, $232.9 \mu\text{m} \times 21.8^\circ$). An important palpogram variation was found at the free-plaque arc length. Only for this particular plaque, the AI-palpogram obtained with (101×21) pixels MW size fail to characterize the atherosclerotic lesion, indicating that the accuracy of the AI-palpogram depends on the MW size used when estimating the radial strain elastogram (Fig. 9).

DISCUSSION

The new intraluminal ultrasonic palpation imaging technique successfully quantified the apparent wall stiffness of the anisotropic vascular wall based on the arterial radial strain and blood pressure data set.




Fig. 4. Effect of palpography domain size on the resulting anisotropic elasticity palpogram. Simulations were performed with plaque #13 considering three palpography domain thicknesses. (a, d, g) Plaque geometries (*red contours*); estimated radial strain-elasticograms and palpography domains with thicknesses of 0.75, 1 and 1.25 mm, respectively (*blue contours*). (b, e, h) Intravascular ultrasound images with anisotropy index-palpograms $AI(\theta)$. (c, f, i) Comparisons between exact and estimated $AI(\theta)$ palpograms. Areas shaded in *yellow* indicate the location of the necrotic core. fi = fibrotic plaque region; nc = necrotic core.

Physical meaning of the elastic anisotropy index AI

One of the advantages of AI, which may be viewed as an anisotropic apparent modulus, is to account for the three Young's moduli of the orthotropic atherosclerotic lesion. Even though the AE-PT does not allow identifying individually the three orthotropic material constants, this index, which combines the three orthotropic Young's moduli, is proportional to the radial Young's modulus and depends on the two stiffness ratios: $k_1 = E_\theta / E_R$ and $k_2 = E_Z / E_R$. Equation (3) indicates that the normalized elastic anisotropy index AI/E_R increases with k_2 and decreases with k_1 when $0.5 \leq k_i \leq 2$ ($i=1, 2$). In other words, an increase in AI may be due to either an increase in E_R and E_Z or a decrease in E_θ . Furthermore, the AI amplitude appears to be very sensitive to the presence of mechanical heterogeneities, dropping significantly when predicting soft inclusions in the composite atherosclerotic lesions (Figs. 2–5).

AE-PT highlights the non-linear mechanical response of the vascular tissue

Because the proposed constitutive model refers to the configuration of the plaque at Im^i , rather than the initial zero-stress configuration, the linear tangent elasticity procedure used in our approach (Bonet and Wood 1997) remains acceptable as long as the absolute amplitudes of the incremental strain components estimated between the two successive IVUS images Im^i and Im^{i+1} remain small ($\leq 2\%$) (Le Floc'h et al. 2012). Therefore, the computed AI must be viewed as an anisotropic linearized tangent modulus (Holzapfel 2000) measured for the physiologic pressure range considered. The other advantage of such an incremental linearized technique applied on a series of IVUS images is that it allows quantification of the non-linear mechanical response of the fibrotic plaque tissue during the physiologic increase in blood pressure.




Fig. 5. Performance of the anisotropic elasticity-palpography technique in detection of vulnerable plaques with two necrotic cores (plaques #16 and 17). (a, d) Plaque geometries (*red contours*), palpography domains (*blue contours*) and estimated radial strain elastograms. (b, e) intravascular ultrasound images with anisotropy index-palpogram $AI(\theta)$. (c, f) Comparisons between exact and estimated $AI(\theta)$ palpograms. Areas shaded in *yellow* indicate the locations of the two necrotic cores. fi = fibrotic plaque region; nc = necrotic core.

How should cardiologists choose the palpography domain?

With the AE-PT, the interventional cardiologist may define the palpography domain by considering any arbitrary endoluminal thick layer size and shape. The results clearly indicated that an increase in the palpography domain thickness does not necessarily decrease the accuracy of the AI-palpogram (see Table 3 and Fig. 4, plaque #13). This could be explained by the fact that the palpography domain must be thin enough to obtain a reasonable signal-to-noise (SNR) ratio (Céspedes et al. 2000; Doyley et al. 2001) and large enough for a good performance of the strain imaging algorithm (Maurice et al. 2004a).

What is the advantage of using the AI instead of the fractional anisotropy index for characterizing a VP?

The principal strain and fractional strain anisotropy indexes are two other strain-based metrics that were defined and applied in non-invasive clinical cardiovascular magnetic resonance imaging and ultrasound studies (Nayak et al. 2017; Soleimanifard et al. 2010). However, extracting the anisotropic mechanical properties of the tissue from these indexes is difficult. Furthermore, although the estimations of these indexes require accurate knowledge of all strain tensor components, estimation of the AI requires only the radial strain component, which is the only component that can be extracted from the coronary RF-IVUS signal with reasonable accuracy.




Fig. 6. Performance of the anisotropic elasticity-palpography technique in detection of a stable plaque with one dense fibrosis inclusion (plaque #19). (a) Plaque geometry (*red contour*), palpography domain (*blue contour*) and estimated radial strain elastogram. (b) Intravascular ultrasound images with anisotropy index-palpogram $AI(\theta)$. (c) Comparison between exact and estimated $AI(\theta)$ palpograms. Area shaded in *yellow* indicates the location of the dense fibrosis inclusion. df = dense fibrosis. fi = fibrotic plaque region.




Fig. 7. Performance of the anisotropic elasticity-palpography technique in detection of a stable homogeneous fibrotic plaque (plaque #11). (a) Plaque geometry (*red contours*), palpography domain (*blue contours*) and estimated radial strain elastogram. (b) Intravascular ultrasound image with the anisotropy index-palpogram $AI(\theta)$. (c) Comparison between the exact and estimated $AI(\theta)$ palpograms. *fi* = fibrotic plaque region.




Fig. 8. Performance of the anisotropic elasticity-palpography technique in characterization of a healthy coronary cross section (healthy 1). (a) Artery geometry (*red contours*), palpography domain (*blue contours*) and estimated radial strain elastogram. (b) Intravascular ultrasound image with the anisotropy index-palpogram $AI(\theta)$. (c) Comparison between the exact and the estimated $AI(\theta)$ palpograms.

Is the AE-PT sufficient to detect VPs and diagnose plaque rupture?

The proposed AE-PT cannot diagnose the degree of plaque vulnerability to rupture because it does not allow

measurements of other key geometric factors (Cilla et al. 2012; Ohayon et al. 2008). However, this technique remains reliable for detecting and highlighting complex VPs with one or several soft inclusions (see Figs. 2–5).




Fig. 9. Sensitivity analysis revealing the influence of measurement window (MW) size on the computed anisotropy index-palpogram $AI(\theta)$. All these computations were performed on the dense fibrotic plaque #18. Three MW sizes were considered for these simulations: 75×15 pixels, 101×21 pixels and 121×31 pixels. (a) Artery geometry (*red contour*), palpography domain (*blue contour*) and estimated radial strain map obtained with the 101×21 -pixels MW size. (b) Comparison between exact and estimated $AI(\theta)$ palpograms. The area shaded in *yellow* indicates the location of the dense fibrosis inclusion. *df*= dense fibrosis; *fi* = fibrotic plaque region.

Study limitations

Several limitations deserve to be pointed out. First, the anisotropic palpation technique was tested and validated only on synthesized RF echo frames. Although further animal and clinical studies are needed to extend and strengthen the present findings, our preliminary results appear to be promising and indicate the potential and feasibility of this novel imaging technique. The second limitation of this study is that the palpography model neglected the residual strain, which has a significant effect on the stress amplitude in the thin fibrous cap (Ohayon et al. 2007). However, because the AI is an anisotropic linearized tangent modulus (Holzapfel 2000) measured during physiologic conditions, it implicitly accounts for all existing internal stresses/strains, including the residual ones. Finally, one clinical case presented in this study (plaque #18, Fig. 9) highlighted the importance of considering optimal MW sizes to compute accurate AI-palpograms. A more exhaustive study clearly deserves to be investigated as future work to define an MW size that would optimize the performance of the strain-imaging algorithm of Majdoulina et al. (2014).

Relevance to other clinical applications

Modifying the mechanical properties of the vulnerable atherosclerotic lesion could be the key issue for the stabilization of VPs. Finet et al. (2004) reported that a very slight increase in soft necrotic core stiffness can tilt a VP from instability to stability. Today, therapies based on statins (Libby et al. 2002; Nozue et al. 2012) and other emergent medications (Nicholls et al. 2016) are used by clinicians to enhance plaque stability. However, the effects of all these drugs on the mechanical properties of coronary atherosclerosis lesions still remain unclear. The proposed improved anisotropic elasticity-palpography imaging technique could be used by clinicians to investigate the evolution of the anisotropic mechanical properties of atherosclerotic plaques during drug therapies.

Acknowledgments—Armida Gómez, Jacques Ohayon and Alex Elias-Zuñiga thank the Mexican Science and Technology Council (CONACYT) and Nuevo-León's Innovation and Technology Transfer Institute (I2 T2). Armida Gómez held a doctoral fellowship from CONACYT (2016-19).

CONFLICT OF INTEREST DISCLOSURE

The authors have no conflicts of interest to disclose.

APPENDIX I. CONSTITUTIVE LAW IN LINEAR ELASTICITY FOR INCOMPRESSIBLE ORTHOTROPIC VASCULAR TISSUES

The constitutive law is given using the cylindrical coordinate system (R, θ, Z) . In linear elasticity, the relationships between the material constants of a compressible orthotropic medium and the stress tensor $(\sigma_{ij}$ with $i, j = R, \theta, Z)$ and strain tensor $(\varepsilon_{ij}$ with $i, j = R, \theta, Z)$ components are given by Berthelot (1999):

$$\begin{pmatrix} \varepsilon_{RR} \\ \varepsilon_{\theta\theta} \\ \varepsilon_{ZZ} \\ 2\varepsilon_{\theta Z} \\ 2\varepsilon_{RZ} \\ 2\varepsilon_{R\theta} \end{pmatrix} = \begin{bmatrix} 1/E_R & -\nu_{\theta R}/E_R & -\nu_{ZR}/E_Z & 0 & 0 & 0 \\ -\nu_{R\theta}/E_R & 1/E_\theta & -\nu_{Z\theta}/E_Z & 0 & 0 & 0 \\ -\nu_{RZ}/E_R & -\nu_{\theta Z}/E_\theta & 1/E_Z & 0 & 0 & 0 \\ 0 & 0 & 0 & 1/G_{\theta Z} & 0 & 0 \\ 0 & 0 & 0 & 0 & 1/G_{RZ} & 0 \\ 0 & 0 & 0 & 0 & 0 & 1/G_{R\theta} \end{bmatrix} \begin{pmatrix} \sigma_{RR} \\ \sigma_{\theta\theta} \\ \sigma_{ZZ} \\ \sigma_{\theta Z} \\ \sigma_{RZ} \\ \sigma_{R\theta} \end{pmatrix} \quad (\text{AI.1})$$

where E_i ($i = R, \theta, Z$) are the three Young's moduli associated with the i directions, $G_{ij} = G_{ji}$ ($i, j = R, \theta, Z$ with $i \neq j$) are the three shear moduli associated with the ij planes, and $\nu_{ij} = -\varepsilon_{ij}/\varepsilon_{ii}$ ($i, j = R, \theta, Z$ with $i \neq j$) are the six Poisson's ratios (i.e. the ratios of the strain in the j direction to the strain in the perpendicular i direction). Such orthotropic material also obeys to the reciprocity conditions: $\nu_{ij}/E_i = \nu_{ji}/E_j$ ($i, j = R, \theta, Z$ with $i \neq j$). Thanks to these relationships, the number of material constants drops to nine. Moreover, assuming that the vascular wall incompressibility imposes the following three relationships among the Poisson's ratios:

$$\nu_{R\theta} + \nu_{RZ} = 0, \quad \nu_{\theta R} + \nu_{\theta Z} = 0 \quad \text{and} \quad \nu_{ZR} + \nu_{Z\theta} = 0. \quad (\text{AI.2a} - 2c)$$

So knowing the three Young's moduli E_R , E_θ and E_Z , we can extract the six Poisson's ratios.

APPENDIX II. THE ELASTIC ANISOTROPY INDEX AI

Assuming the vascular media as incompressible and under plane strain constraint ($\varepsilon_{RZ} = \varepsilon_{\theta Z} = \varepsilon_{ZZ} = 0$), one can derive the following relationships (Berthelot 1999):

$$\sigma_{RR}(R, \theta) - \sigma_{\theta\theta}(R, \theta) = C(R, \theta)\varepsilon_{RR}(R, \theta) \quad (\text{AII.1a})$$

with

$$C(R, \theta) = E_R(R, \theta)[1 - \nu_{RZ}(R, \theta)\nu_{ZR}(R, \theta)]^{-1} \quad (\text{AII.1b})$$

$$\nu_{RZ}(R, \theta) = \frac{1}{2} \left[1 - \frac{1}{k_1} + \frac{1}{k_2} \right] \quad (\text{AII.1c})$$

$$\nu_{ZR}(R, \theta) = \frac{1}{2} \left[1 + k_2 - \frac{k_2}{k_1} \right] \quad (\text{AII.1d})$$

$$k_1 = \frac{E_\theta(R, \theta)}{E_R(R, \theta)} \quad (\text{AII.1e})$$

and

$$k_2 = \frac{E_Z(R, \theta)}{E_R(R, \theta)} \quad (\text{AII.1f})$$

where, $\sigma_{RR}(R, \theta)$, $\sigma_{\theta\theta}(R, \theta)$ and $\varepsilon_{RR}(R, \theta)$ are the radial and circumferential stress and radial strain components, respectively.

In this linearized elasticity approach, we assumed that the incremental radial strain $\varepsilon_{RR}(R, \theta)$ is proportional to the imposed change in pressure ΔP and inversely proportional to the anisotropic material constant $C(R, \theta)$:

$$\varepsilon_{RR}(R, \theta) = \frac{\Delta P}{C(R, \theta)} g(R, \theta) \quad (\text{AII.2})$$

where $g(R, \theta)$ is a correcting shape function, which account for the entire plaque morphology including regions of plaque heterogeneities. By substituting Eq. (AII.2) into Eq. (AII.1a), Eq. (AII.1a) can be rewritten as follow:

$$\sigma_{RR}(R, \theta) - \sigma_{\theta\theta}(R, \theta) = \Delta P g(R, \theta) \quad (\text{AII.3})$$

Taking advantage of this expression (Eq. (AII.3)), the elastic anisotropy index AI (Eq. (1)) becomes

$$AI(\theta) = \frac{\Delta P \left| \int_{R_i(\theta)}^{R_o(\theta)} g(R, \theta) dR \right|}{\varepsilon(\theta)} \quad (\text{AII.4})$$

Since we assumed that the unknown morphological function $g(R, \theta)$ does not depend on the mechanical properties and because this function escapes to direct measurements, we approximated it by a correcting *homogeneous* shape function $g^*(R, \theta)$ accounting only for the anatomical cross-sectional plaque geometry. This approximated correcting function was obtained by performing a FE simulation (with a loading blood pressure differential ΔP) in which the pathologic vessel was assumed homogeneous and made of a quasi-incompressible isotropic material. This approximated shape function $g^*(R, \theta)$ was used to estimate the anisotropic palpography index:

$$AI(\theta) = \frac{\Delta P \left| \int_{R_i(\theta)}^{R_o(\theta)} g^*(R, \theta) dR \right|}{\varepsilon(\theta)} \quad (\text{AII.5})$$

REFERENCES

Abran M, Cloutier G, Cardinal MH, Chayer B, Tardif JC, Lesage F. Development of a photoacoustic, ultrasound and fluorescence imaging catheter for the study of atherosclerotic plaque. *IEEE Trans Biomed Circuits Syst* 2014;8:696–703

Ahmadi A, Stone GW, Leipsic J, Serruys PW, Shaw L, Hecht H, Wong G, Nørgaard BL, O’Gara PT, Chandrashekar Y, Narula J. Association of coronary stenosis and plaque morphology with fractional flow reserve and outcomes. *JAMA Cardiol* 2016;1:350–357.

Akyildiz AC, Speelman L, Nieuwstadt HA, van Brummelen H, Virmani R, van der Lugt A, van der Steen AFW, Wentzel JJ, Gijzen FJH. The effects of plaque morphology and material properties on peak cap stress in human coronary arteries. *Comput Methods Biomech Biomed Eng* 2016;19:771–779.

Alsheikh-Ali AA, Kitsios GD, Balk EM, Lau J, Ip S. The vulnerable atherosclerotic plaque: Scope of the literature. *Ann Intern Med* 2010;153:387–395.

Arbab-Zadeh A, Nakano M, Virmani R, Fuster V. Acute coronary events. *Circulation* 2012;125:1147–1156.

Baldewings RA, Schaar JA, Mastik F, van der Steen AFW. Local elasticity imaging of vulnerable atherosclerotic coronary plaques. *Adv Cardiol* 2007;44:35–61.

Benjamin EJ, Blaha MJ, Chiuve SE, Cushman M, Das SR, Deo R, de Ferranti SD, Floyd J, Fornage M, Gillespie C, Isasi CR, Jiménez MC, Jordan LC, Judd SE, Lackland D, Lichtman JH, Lisabeth L, Liu S, Longenecker CT, Mackey RH, Matsushita K, Mozaffarian D, Mussolino ME, Nasir K, Neumar RW, Palaniappan L, Pandey DK, Thiagarajan RR, Reeves MJ, Ritchey M, Rodriguez CJ, Roth GA, Rosamond WD, Sasson C, Towfighi A, Tsao CW, Turner MB, Virani SS, Voeks JH, Willey JZ, Wilkins JT, Wu JH, Alger HM, Wong SS, Muntner P. American Heart Association Statistics Committee and Stroke Statistics Subcommittee. Heart Disease and Stroke Statistics—2017 Update: A report from the American Heart Association. *Circulation* 2017;135:e146–e603.

Berthelot JM. Elastic behavior of an orthotropic composite. *Composite materials. Mechanical Engineering series*. New York: Springer; 1999. p. 186–194.

Bonet J, Wood RD. *Nonlinear continuum mechanics for finite element analysis*. London/New York: Cambridge University Press, 1997.

Bourantas CV, Jaffer FA, Gijzen FJ, Van Soest G, Madden SP, Courtney BK, Fard AM, Tenekcioglu E, Zeng Y, Van Der Steen AFW, Emelianov S, Muller J, Stone PH, Marcu L, Tearney GJ, Serruys PW. Hybrid intravascular imaging: Recent advances, technical considerations, and current applications in the study of plaque pathophysiology. *Eur Heart J* 2017;38:400–412.

Caplan JD, Waxman S, Nesto RW, Muller JE. Near-infrared spectroscopy for the detection of vulnerable coronary artery plaques. *J Am Coll Cardiol* 2006;47:C92–C96.

Cardoso L, Kelly-Arnold A, Maldonado N, Laudier D, Weinbaum S. Effect of tissue properties, shape and orientation of microcalcifications on vulnerable cap stability using different hyperelastic constitutive models. *J Biomech* 2014;47:870–877.

Cardoso L, Weinbaum S. Changing views of the biomechanics of vulnerable plaque rupture: A review. *Ann Biomed Eng* 2014;42:415–431.

Cavalcante JL, Larose E. Intravascular MRI for plaque characterization: Are we close to reality? *Curr Cardiol Rep* 2016;18:89.

Céspedes I, Ophir J, Ponnekanti H, Maklad N. Elastography: Elasticity imaging using ultrasound with application to muscle and breast *in vivo*. *Ultrason Imaging* 1993;15:73–88.

Céspedes EI, De Korte CL, Van Der Steen AFW. Intraluminal ultrasonic palpation: Assessment of local and cross-sectional tissue stiffness. *Ultrasound Med Biol* 2000;26:385–396.

Chagnon G, Ohayon J, Martiel JL, Favier D. Hyperelasticity modeling for incompressible passive biological tissues. In: Payen Y, Ohayon J, (eds). *Biomechanics of living organs*. London/San Diego: Elsevier (Academic Press); 2017. p. 3–30.

Cheng JM, Garcia-Garcia HM, De Boer SPM, Kardys I, Heo JH, Akkerhuis KM, Oemrawsingh RM, Van Domburg RT, Ligthart J, Witberg KT, Regar E, Serruys PW, Van Geuns RJ, Boersma E. In vivo detection of high-risk coronary plaques by radiofrequency intravascular ultrasound and cardiovascular outcome: Results of the ATHEROREMO-IVUS study. *Eur Heart J* 2014;35:639–947.

Chopard R, Boussel L, Motreff p, Rioufol G, Tabib A, Douek P, Meyronet D, Revel D, Finet G. How reliable are 40 MHz IVUS and 64-slice MDCT in characterizing coronary plaque composition?

- An *ex vivo* study with histopathological comparison. *Int J Cardiovasc Imaging* 2010;26:373–383.
- Cilla M, Peña E, Martínez MA. 3-D computational parametric analysis of eccentric atheroma plaque: Influence of axial and circumferential residual stresses. *Biomech Model Mechanobiol* 2012;11:1001–1013.
- De Korte CL, Carlier SG, Mastik F, Doyley MM, Van Der Steen AFW, Serruys PW, Bom N. Morphologic and mechanical information of coronary arteries obtained with intravascular elastography: Feasibility study *in vivo*. *Eur Heart J* 2002;23:405–413.
- Deleaval F, Bouvier A, Finet G, Cloutier G, Yazdani SK, Le Floch S, Clarysse P, Pettigrew RI, Ohayon J. The intravascular ultrasound elasticity–palpography technique revisited: A reliable tool for the *in vivo* detection of vulnerable coronary atherosclerotic plaques. *Ultrasound Med Biol* 2013;39:1469–1481.
- Di Mario C, Gorge G, Peters R, Kearney P, Pinto F, Hausmann D, von Birgelen C, Colombo A, Mudra H, Roelandt J, Erbel R. Clinical application and image interpretation in intracoronary ultrasound. Study Group on Intracoronary Imaging of the Working Group of Coronary Circulation and of the Subgroup on Intravascular Ultrasound of the Working Group of Echocardiography of the European Society of Cardiology. *Eur Heart J* 1998;19:207–229.
- Doyley MM. Model-based elastography: A survey of approaches to the inverse elasticity problem. *Phys Med Biol* 2012;57:R35–R73.
- Doyley MM, Mastik F, de Korte CL, Carlier SG, Céspedes EI, Serruys PW, Bom N, van der Steen AF. Advancing intravascular ultrasound palpation toward clinical applications. *Ultrasound Med Biol* 2001;27:1471–1480.
- Finet G, Ohayon J, Rioufol G. Biomechanical interaction between cap thickness, lipid core composition and blood pressure in vulnerable coronary plaque: Impact on stability or instability. *Coron Artery Dis* 2004;15:13–20.
- Fleg JL, Stone GW, Fayad ZA, Granada JF, Hatsukami TS, Kolodgie FD, Ohayon J, Pettigrew R, Sabatine MS, Tearney GJ, Waxman S, Domanski MJ, Srinivas PR, Narula J. Detection of high-risk atherosclerotic plaque: Report of the NHLBI Working Group on current status and future directions. *JACC Cardiovasc Imaging* 2012;5:941–955.
- Fromageau J, Brusseau E, Vray D, Gimenez G, Delachartre P. Characterization of PVA cryogel for intravascular ultrasound elasticity imaging. *IEEE Trans Ultrason Ferroelectr Freq Control* 2003;50:1318–1324.
- Gijzen FJ, Migliavacca F. Plaque mechanics. *J Biomech* 2014;47:763–764.
- Gijzen FJ, Migliavacca F, Schievano S, Succi L, Petrini L, Thury A, Wentzel JJ, Van Der Steen AFW, Serruys PW, Dubini G. Simulation of stent deployment in a realistic human coronary artery. *Biomed Eng Online* 2008; 7. Available at: <http://www.biomedical-engineering-online.com/content/7/1/23> [cited Mar 6, 2017].
- Golemati S, Gastounioti A, Nikita KS. Ultrasound-image-based cardiovascular tissue motion estimation. *IEEE Rev Biomed Eng* 2016;9:208–218.
- Guo X, Zhu J, Maehara A, Monoly D, Samady H, Wang L, Billiar KL, Zheng J, Yang C, Mintz GS. Quantify patient-specific coronary material property and its impact on stress/strain calculations using *in vivo* IVUS data and 3-D FSI models: A pilot study. *Biomech Model Mechanobiol* 2017;16:333–344.
- Holzappel GA. *Nonlinear solid mechanics: A continuum approach for engineering*. New York: Wiley; 2000.
- Holzappel GA, Sommer G, Gasser CT, Regitnig P. Determination of layer-specific mechanical properties of human coronary arteries with nonatherosclerotic intimal thickening and related constitutive modeling. *Am J Physiol Circ Physiol* 2005;289:H2048–H2058.
- Jensen JA, Svendsen NB. Calculation of pressure fields from arbitrarily shaped, apodized, and excited ultrasound transducers. *IEEE Trans Ultrason Ferroelectr Freq Control* 1992;39:262–267.
- Kelly-Arnold A, Maldonado N, Laudier D, Aikawa E, Cardoso L, Weinbaum S. Revised microcalcification hypothesis for fibrous cap rupture in human coronary arteries. *Proc Natl Acad Sci USA* 2013;110:10741–10746.
- Kok AM, Speelman L, Virmani R, Steen AFW, Gijzen FJH, Wentzel JJ. Peak cap stress calculations in coronary atherosclerotic plaques with an incomplete necrotic core geometry. *Biomed Eng Online* 2016;15:48.
- Larose E, Yeghiazarians Y, Libby P, Yucel EK, Aikawa M, Kacher DF, Aikawa E, Kinlay S, Schoen FJ, Selwyn AP, Ganz P. Characterization of human atherosclerotic plaques by intravascular magnetic resonance imaging. *Circulation* 2005;112:2324–2331.
- Le Floch S, Ohayon J, Tracqui P, Finet G, Gharib AM, Maurice RL, Cloutier G, Pettigrew RI. Vulnerable atherosclerotic plaque elasticity reconstruction based on a segmentation-driven optimization procedure using strain measurements: Theoretical framework. *IEEE Trans Med Imaging* 2009;28:1126–1137.
- Le Floch S, Cloutier G, Finet G, Tracqui P, Pettigrew RI, Ohayon J. On the potential of a new IVUS elasticity modulus imaging approach for detecting vulnerable atherosclerotic coronary plaques: *In vitro* vessel phantom study. *Phys Med Biol* 2010;55:5701–5721.
- Le Floch S, Cloutier G, Saijo Y, Finet G, Yazdani SK, Deleaval F, Rioufol G, Pettigrew RI, Ohayon J. A four-criterion selection procedure for atherosclerotic plaque elasticity reconstruction based on *in vivo* coronary intravascular ultrasound radial strain sequences. *Ultrasound Med Biol* 2012;38:2084–2097.
- Lee R, Grodzinsky A, Frank E, Kamm R, Schoen F, Lee RT, Grodzinsky AJ, Frank EH, Kamm RD, Schoen FJ. Structure-dependent dynamic mechanical behavior of fibrous caps from human structure-dependent dynamic mechanical behavior of fibrous caps from human atherosclerotic plaques. *Circulation* 1991;83:1764–1770.
- Libby P, Ridker PM, Maseri A. Inflammation and atherosclerosis. *Circulation* 2002;105:1135–1143.
- Loree HM, Tobias BJ, Gibson LJ, Kamm RD, Small DM, Lee RT. Mechanical properties of model atherosclerotic lesion lipid pools. *Arterioscler Thromb Vasc Biol* 1994;14:230–234.
- Madder RD, Husaini M, Davis AT, Vanoosterhout S, Khan M, Wohns D, Mcnamara RF, Wolschleger K, Gribar J, Collins JS, Jacoby M, Decker JM, Hendricks M, Sum ST, Madden S, Ware JH, Muller JE. Large lipid-rich coronary plaques detected by near-infrared spectroscopy at non-stented sites in the target artery identify patients likely to experience future major adverse cardiovascular events. *Eur Heart J Cardiovasc Imaging* 2016;17:393–399.
- Majdoulina Y, Ohayon J, Keshavarz-Motamed Z, Roy Cardinal MH, Garcia D, Allard L, Lerouge S, Arsenaault F, Soulez G, Cloutier G. Endovascular shear strain elastography for the detection and characterization of the severity of atherosclerotic plaques: *In vitro* validation and *in vivo* evaluation. *Ultrasound Med Biol* 2014;40:890–903.
- Maurice RL, Bertrand M. Lagrangian speckle model and tissue-motion estimation-theory [ultrasonography]. *IEEE Trans Med Imaging* 1999;18:593–603.
- Maurice RL, Ohayon J, Finet G, Cloutier G. Adapting the Lagrangian speckle model estimator for endovascular elastography: Theory and validation with simulated radio-frequency data. *J Acoust Soc Am* 2004a;116:1276–1286.
- Maurice RL, Ohayon J, Frégnigny Y, Bertrand M, Soulez G, Cloutier G. Noninvasive vascular elastography: Theoretical framework. *IEEE Trans Med Imaging* 2004b;23:164–180.
- Maurice RL, Fromageau J, Brusseau E, Finet G, Rioufol G, Cloutier G. On the potential of the Lagrangian estimator for endovascular ultrasound elastography: *In vivo* human coronary artery study. *Ultrasound Med Biol* 2007;33:1199–1205.
- Morbiducci U, Kok AM, Kwak BR, Stone PH, Steinman DA, Wentzel JJ. Atherosclerosis at arterial bifurcations: Evidence for the role of haemodynamics and geometry. *Thromb Haemost* 2016;115:484–492.
- Naghavi M, Libby P, Falk E, Casscells SW, Litovsky S, Rumberger J, Badimon JJ, Stefanadis C, Moreno P, Pasterkamp G, Fayad Z, Stone PH, Waxman S, Raggi P, Madjid M, Zarrabi A, Burke A, Yuan C, Fitzgerald PJ, Siscovick DS, de Korte CL, Aikawa M, Juhani Airaksinen KE, Assmann G, Becker CR, Chesebro JH, Farb A, Galis ZS, Jackson C, Jang IK, Koenig W, Lodder RA, March K, Demirovic J, Navab M, Puri SG, Reekter MD, Bahr R, Grundy SM, Mehran R, Colombo A, Boerwinkle E, Ballantyne C, Insull W,

- Schwartz RS, Vogel R, Serruys PW, Hansson GK, Faxon DP, Kaul S, Drexler H, Greenland P, Muller JE, Virmani R, Ridker PM, Zipes DP, Shah PK, Willerson JT. From vulnerable plaque to vulnerable patient: A call for new definitions and risk assessment strategies: Part I. *Circulation* 2003;108:1664–1672.
- Nayak R, Huntzicker S, Ohayon J, Carson N, Dogra V, Schifitto G, Doyley MM. Principal strain vascular elastography: Simulation and preliminary clinical evaluation. *Ultrasound Med Biol* 2017;43:682–699.
- Nicholls SJ, Puri R, Anderson T, Ballantyne CM, Cho L, Kastelein JJP, Koenig W, Somaratne R, Kassahun H, Yang J, Wasserman SM, Scott R, Ungi I, Podolec J, Ophuis AO, Cornel JH, Borgman M, Brennan DM, Nissen SE. Effect of evolocumab on progression of coronary disease in statin-treated patients. *JAMA* 2016;316:2373.
- Nozue T, Yamamoto S, Tohyama S, Umezawa S, Kunishima T, Sato A, Miyake S, Takeyama Y, Morino Y, Yamauchi T, Muramatsu T, Hibi K, Sozu T, Terashima M, Michishita I. Statin treatment for coronary artery plaque composition based on intravascular ultrasound radio-frequency data analysis. *Am Heart J* 2012;163:191–199.
- Ohayon J, Dubreuil O, Tracqui P, Le Floch S, Rioufol G, Chalabreysse L, Thivolet F, Pettigrew RI, Finet G. Influence of residual stress/strain on the biomechanical stability of vulnerable coronary plaques: potential impact for evaluating the risk of plaque rupture. *Am J Physiol Heart Circ Physiol* 2007;293:1987–1996.
- Ohayon J, Finet G, Gharib AM, Herzka DA, Tracqui P, Heroux J, Rioufol G, Kotys MS, Elagha A, Pettigrew RI. Necrotic core thickness and positive arterial remodeling index: Emergent biomechanical factors for evaluating the risk of plaque rupture. *Am J Physiol Heart Circ* 2008;295:H717–H727.
- Ohayon J, Finet G, Le Floch S, Cloutier G, Gharib AM, Heroux J, Pettigrew RI. Biomechanics of atherosclerotic coronary plaque: Site, stability and *in vivo* elasticity modeling. *Ann Biomed Eng* 2014;42:269–279.
- Ophir J, Céspedes I, Ponnekanti H, Yazdi Y, Li X. Elastography: A quantitative method for imaging the elasticity of biological tissues. *Ultrason Imaging* 1991;13:111–134.
- Poree J, Chayer B, Soulez G, Ohayon J, Cloutier G. Noninvasive vascular modulography method for imaging the local elasticity of atherosclerotic plaques: Simulation and *in vitro* vessel phantom study. *IEEE Trans Ultrason Ferroelectr Freq Control* 2017;64:1805–1817.
- Richards MS, Doyley MM. Investigating the impact of spatial priors on the performance of model-based IVUS elastography. *Phys Med Biol* 2011;56:7223–46.
- Richards MS, Perucchio R, Doyley MM. Visualizing the stress distribution within vascular tissues using intravascular ultrasound elastography: A preliminary investigation. *Ultrasound Med Biol* 2015;41:1616–1631.
- Rioufol G, Finet G, Gimon I, André-Fouët X, Rossi R, Vialle E, Desjoux E, Convert G, Huret JF, Tabib A. Multiple atherosclerotic plaque rupture in acute coronary syndrome: A three-vessel intravascular ultrasound study. *Circulation* 2002;106:804–808.
- Schaar JA, De Korte CL, Mastik F, Van Damme LCA, Krams R, Serruys PW, Van Der Steen AFW. Three-dimensional palpography of human coronary arteries: Ex vivo validation and in-patient evaluation. *Herz* 2005;30:125–133.
- Sinclair H, Bourantas C, Bagnall A, Mintz GS, Kunadian V. OCT for the identification of vulnerable plaque in acute coronary syndrome. *JACC Cardiovasc Imaging* 2015;8:198–209.
- Soleimanifard S, Abd-Elmoniem KZ, Agarwal HK, Tomas MS, Sasano T, Vonken E, Youssef A, Abraham MR, Abraham TP, Prince JL. Identification of myocardial infarction using three-dimensional strain tensor fractional anisotropy. *Proc IEEE Int Symp Biomed Imaging* 2010;2010:468–471.
- Tacheau A, Le Floch S, Finet G, Doyley MM, Pettigrew RI, Cloutier G, Ohayon J. The imaging modulography technique revisited for high-definition intravascular ultrasound: Theoretical framework. *Ultrasound Med Biol* 2016;42:727–741.
- Tang D, Teng Z, Canton G, Hatsukami TS, Dong L, Huang X, Yuan C. Local critical stress correlates better than global maximum stress with plaque morphologic features linked to atherosclerotic plaque vulnerability: An *in vivo* multi-patient study. *Biomed Eng Online* 2009;8:15.
- Thondapu V, Bourantas CV, Foin N, Jang IK, Serruys PW, Barlis P. Basic science for the clinician: Biomechanical stress in coronary atherosclerosis: Emerging insights from computational modelling. *Eur Heart J* 2017;38:81–92.
- Widynski N, Porée J, Roy Cardinal MH, Ohayon J, Cloutier G, Garcia D. A sequential Bayesian based method for tracking and strain palpography estimation of arteries in intravascular ultrasound images. *Proc IEEE Int Ultrason Symp* 2014;515–518.