



**HAL**  
open science

## Validation of a method to compensate multicenter effects affecting CT radiomic features

Fanny Orhac, Frédérique Frouin, Christophe Nioche, Nicholas Ayache, Irene Buvat

► **To cite this version:**

Fanny Orhac, Frédérique Frouin, Christophe Nioche, Nicholas Ayache, Irene Buvat. Validation of a method to compensate multicenter effects affecting CT radiomic features. Radiology, In press. hal-01953538

**HAL Id: hal-01953538**

**<https://hal.science/hal-01953538>**

Submitted on 13 Dec 2018

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## Validation of a method to compensate multicenter effects affecting CT radiomic

Fanny Orlhac (PhD)<sup>1</sup>, Frédérique Frouin (PhD)<sup>2</sup>, Christophe Nioche (PhD)<sup>2</sup>,  
Nicholas Ayache (PhD)<sup>1</sup>, Irène Buvat (PhD)<sup>2</sup>

1: UCA, Inria Sophia Antipolis – Méditerranée, Epione, 2004 route des Lucioles – BP 93, 06 902 Sophia Antipolis Cedex, France.

2: Imagerie Moléculaire In Vivo, CEA-SHFJ, Inserm, CNRS, Université Paris-Sud, Université Paris-Saclay, 4 place du Général Leclerc, 91 400 Orsay, France.

Corresponding author:

Fanny Orlhac

+33 4 92 38 71 52

[orlhac@gmail.com](mailto:orlhac@gmail.com)

Inria Sophia Antipolis – Méditerranée

2004 route des Lucioles – BP 93

06 902 Sophia Antipolis Cedex, France

**Article type:** Technical Development

**Word count:** 2510 words

**List of all abbreviations:**

**ABS:** Acrylonitrile Butadiene Styrene

**CCR:** Credence Cartridge Radiomics

**FBP:** Filtered Back Projection

**L:** Lung reconstruction algorithm

**PCA:** Principal Component Analysis

**RIDER:** Reference Image Database to Evaluate therapy Response

**S:** Standard reconstruction algorithm

**S3:** Sinogram affirmed iterative reconstruction with a noise reconstruction strength of level 3

**S5:** Sinogram affirmed iterative reconstruction with a noise reconstruction strength of level 5

**VOI:** Volume of Interest

## **Abstract**

**Purpose:** We investigated whether a compensation method could correct for the variations of radiomic feature values caused by the use of different CT protocols.

**Materials and Methods:** Phantom data involving 10 texture patterns and 74 patients (cohort #1: 42 patients; 19 males; mean age, 60.4 years; range, 31-81 years; September-October 2013; cohort #2: 32 patients; 16 males; mean age, 62.1 years; range, 29-82 years; January-September 2007) scanned using different CT protocols were retrospectively included. For any radiomic feature, the compensation approach identified a protocol-specific transformation to express all data in a common space devoid of protocol effects. The differences in statistical distributions between protocols were assessed using Friedman tests before and after compensation. Principal component analyses (PCA) were performed on the phantom data to evaluate the ability to distinguish between texture patterns after compensation.

**Results:** In the phantom data, the statistical distributions of features were different between protocols for all radiomic features and texture patterns ( $p < .05$ ). After compensation, the protocol effect was no longer detectable ( $p > .05$ ). PCA demonstrated that each texture pattern was no longer displayed as different clusters corresponding to different imaging protocols unlike what was observed before compensation. The correction for scanner effect was confirmed in patient data with 100% (10/10 features for cohort #1) and 98% (87/89 features for cohort #2) of p-values less than .05 before compensation, compared to 30% (3/10) and 15% (13/89) after compensation.

**Conclusion:** The compensation successfully realigns feature distributions computed from different CT imaging protocols and should facilitate multicenter radiomic studies.

## Introduction

Since 2012, the concept of Radiomics is expanding (1) in oncology with the objective to characterize tumor heterogeneity from medical images. Radiomics extracts features from medical images that quantify tumor shape, intensity histogram and texture of the lesions more precisely and more accurately than visual assessment by a radiologist, in order to build models involving such features to assist patient management. In particular, texture analysis from CT images has led to promising results to distinguish between tumor lesions with different histopathological characteristics and to predict treatment response or patients survival (2). However, several studies have highlighted the sensitivity of radiomic features to CT acquisition and reconstruction parameters using phantoms (3–8) or patient data (9–12). Indeed, feature values are affected by slice thickness, pixel size, reconstruction kernel, tube voltage, tube current and contrast-enhancement. They also differ between different scanners with the same settings (8). Moreover, the impact of imaging protocols varies according to the texture pattern and the radiomic feature (4).

One of the most widely cited studies in radiomics (13), which included 1019 patients, used different CT imaging protocols involving different CT scanners, different pixel size and slice thickness, with or without intravenous contrast, without accounting for this variability in the data analysis. To reduce that variability, it has been proposed to resample images with a fixed voxel size, to filter the images (5) or to change the definition of features (6,11). These approaches require a modification of the CT images or are not applicable to all radiomic features.

The same issue is encountered in PET imaging, where radiomic features are sensitive to the acquisition protocol and reconstruction algorithm (14). A compensation method was initially described in genomics (15), where the so-called batch effect is the source of variations in measurements caused by the handling of samples by different laboratories, different technicians and on different days. The batch-effect is conceptually similar to variations induced by the scanner or the protocol effects in radiomics. The

compensation method identifies a batch-specific transformation to express all data in a common space devoid of batch effects. It has been shown to be effective in PET to realign the radiomic feature distributions between 3 different protocols for healthy liver tissue and breast lesions, without altering the biological information (16).

The purpose of this study was therefore to determine whether this compensation method could also correct for the CT protocol effect, using phantom and patient data.

## **Materials and Methods**

All patient data were anonymized and publicly available in supplemental data of (9) and (10). All authors had control of the data and information submitted for publication.

### **Phantom experiments**

The phantom data used in our study have been produced by Mackin et al (4) and are publicly available in supplemental data of (4). The Credence Cartridge Radiomics (CCR) phantom consists of 10 layers with different materials corresponding to different texture patterns. This phantom was scanned using 17 different imaging protocols from four medical institutes involving various reconstruction kernels, scan types, slice thickness, pixel spacing, spiral pitch factor and effective milliamperage. Additional information on phantom and acquisition characteristics are provided in Supplemental Tables 1 and 2. For each layer, 16 non-overlapping volumes of interest (VOI) with a cubic volume, on average, of  $8 \text{ cm}^3$  (range:  $[7.6 - 9 \text{ cm}^3]$  corresponding to  $[2708 - 14332 \text{ voxels}]$  depending on the imaging protocols) are also made available in Dicom-RTstruct format. For each VOI and each imaging protocol, we (FO with 7 years and CN with 20 years of research experience in medical imaging) computed 40 radiomic features using the LIFEx freeware (17) ([www.lifexsoft.org](http://www.lifexsoft.org), Inserm, Orsay, France, Supplemental Table 3), with a fixed bin size (18) set to 10

HU between -1 000 UH and 3 000 UH without any spatial resampling. We performed the radiomic analysis for 16 imaging protocols out of 17 due to a reading issue with acquisition CCR1-GE2.

## **Patients**

Publicly available radiomic features from two patient databases (#1 and #2) were used in our study. First set of features was derived from cohort #1 of 42 patients with a lung cancer between September and October 2013 (9), including 19 males (mean age, 60.4 years; range, 31-81 years, Table 1). All patients underwent a CT scan with the same machine and protocol (Supplemental Table 4), and CT images were reconstructed using three algorithms: filtered back projection (FBP), Sinogram Affirmed Iterative Reconstruction (Siemens Healthcare, Forchheim, Germany) with a noise reduction strength of level 3 (called S3 thereafter) and 5 (called S5). For each patient, the dominant tumor lesion was segmented manually three times, twice by a radiologist and once by a technologist. For each of the 3 VOIs per patient and each reconstruction, 15 radiomic features were calculated. We (FO) excluded 5 geometric features (volume, diameter, surface, sphericity and compactness) from the analysis as they mostly depend on the segmentation.

Second set of features was obtained from 32 patients of cohort #2 between January and September 2007 (16 males; mean age, 62.1 years; range, 29-82 years, Table 1) with a lung cancer who underwent two CT scans (Supplemental Table 4) within 15 minutes (10). This dataset was originally collected in the clinical trial NCT00579852 to evaluate the reproducibility of tumor volume and diameter measurements and is part of the Reference Image Database to Evaluate therapy Response (RIDER) project (19). The CT images were reconstructed using 6 protocols combining two reconstruction algorithms (lung and standard abbreviated as L and S) and three slice thicknesses (1.25 mm, 2.5 mm and 5 mm) (10). For one lesion per patient (29 primary and 3 metastatic lesions), a tumor VOI was obtained from a consensus among the

manual segmentations by 3 radiologists. After resampling the VOI voxels to  $0.5 \times 0.5 \times 0.5 \text{ mm}^3$  using a trilinear interpolation, 89 radiomic features were calculated for the 6 imaging protocols (2 reconstructions  $\times$  3 slice thicknesses) and for each of the 64 scans (32 patients with 2 scans).

### Compensation method

To correct for differences in features caused by the various imaging protocols, we (FO) used the ComBat compensation method (15). This method has been used for cortical-thickness measurements from MR images (20) and for radiomic features from different PET protocols (16). It is a data-driven method that identifies the protocol effect assuming that the value of each feature  $y$  measured in VOI  $j$  with imaging protocol  $i$  can be written as:

$$y_{ij} = \alpha + \gamma_i + \delta_i \varepsilon_{ij} \quad \text{Equation 1}$$

where  $\alpha$  is the average value for feature  $y_{ij}$ ,  $\gamma_i$  is an additive protocol effect, and  $\delta_i$  is a multiplicative protocol effect affected by an error term ( $\varepsilon_{ij}$ ). The compensation consists in estimating the model parameters  $\alpha$ ,  $\gamma_i$  and  $\delta_i$  using a maximum likelihood approach based on the set of available observations  $y$ :

$$y_{ij}^{\text{ComBat}} = \frac{y_{ij} - \hat{\alpha} - \hat{\gamma}_i}{\hat{\delta}_i} + \hat{\alpha} \quad \text{Equation 2}$$

where  $\hat{\alpha}$ ,  $\hat{\gamma}_i$  and  $\hat{\delta}_i$  are estimators of  $\alpha$ ,  $\gamma_i$  and  $\delta_i$ .

We used the non-parametric form of the model in which no assumptions are made regarding the laws followed by the parameters. In this setting, ComBat determines a transformation for each feature separately. For each texture pattern of phantom data and of each patient dataset, we used the R (version

3.4.2, R foundation for Statistical Computing, Vienna, Austria) function called ComBat, available at <https://github.com/Jfortin1/ComBatHarmonization> to identify the transformation parameters.

### **Statistical analysis**

To determine whether the protocol setting (independent variable  $i$  in Equation 1) impacted the distributions of radiomic feature values (dependent variables  $y_{ij}$  in Equation 1), we (FO, FF with 30 years of experience) performed 2-sided Friedman tests before and after ComBat compensation for each feature as summarized in Supplemental Table 5. Null hypothesis is that there is no difference between the distributions. Benjamini-Hochberg procedure was used to control the false discovery rate (21). P-values less than .05 were regarded as statistically significant. As the goal of the compensation is to realign the distributions in terms of mean and standard deviation, a p-value of the Friedman test greater than .05 means that the realignment was successful.

For the phantom data, we also performed a principal component analysis (PCA) of the 2 560 samples (16 VOI x 10 texture patterns x 16 imaging protocols) described by 40 variables (radiomic features). PCA was performed before and after ComBat to visualize the impact of the compensation method on the distinction between patterns. We also studied whether two textural patterns could be distinguished when pooling data from the 3 imaging protocols before and after compensation and for balanced and un-balanced groups.

The statistical analysis was performed using R software.

## Results

Patient characteristics are shown in Table 1.

### Phantom experiments

In the phantom data, 399/400 p-values of the Friedman tests performed for all features based on 16 imaging protocols and 10 texture patterns were lower than .05 before compensation (Table 2, Supplemental Table 6). Only one p-value for Skewness was larger than .05 for pattern 7 (dense cork,  $p=.46$ ). After compensation, all p-values of Friedman tests were higher than .05, demonstrating that the protocol effect was no longer detectable.

These results were confirmed visually using the projection of the data in the space spanned by the first 2 principal components of PCA. Figure 1 shows an overlapped between textural patterns before ComBat, due to the large variability of radiomic feature values computed from 16 different CT protocols. For each textural pattern (each color), several clusters corresponding to different CT protocols could be identified. After ComBat, textural patterns could be clearly distinguished and were no longer composed of different clusters, demonstrating that the compensation method properly corrected for the scanner effect while retaining the specific characteristics of each texture pattern. Interestingly, the variance explained by the first two components was higher after ComBat (65.6% versus 53.2%), with approximately the same features contributing to the first two principal components before and after compensation (data not shown).

Based on three CT acquisitions (GE1, P2 and S2, see Supplemental Table 2), Figure 2 shows that when data were pooled without realignment, the sensitivity for distinguishing cork from dense cork was 67% (32/48 VOI) with a specificity of 98% (47/48 VOI) using the cutoff maximizing the Youden index. After ComBat, both sensitivity and specificity were 100% (48/48 VOI). For unbalanced groups, Supplemental Figure 1 shows that the compensation method also yielded a perfect distinction between these two patterns.

## **Patient data**

For patient datasets #1 and #2, 100% (10/10) and 98% (87/89) of Friedman tests respectively were lower than .05 between imaging protocols before ComBat (Table 2, Supplemental Tables 7-8). After ComBat, 30% (3/10) of p-values for dataset #1 and 15% (13/89) of p-values for dataset #2 were lower than .05. Visual inspection of the radiomic feature value distributions when Friedman tests remained significant after ComBat showed that the residual difference between protocols was always small and that the protocol effect was much reduced (Supplemental Figure 2), demonstrating the effectiveness of the compensation. As illustrated in Supplemental Figure 3 for Homogeneity feature, ComBat corrects the protocol effect with a realignment of feature values among the three protocols for dataset #1 and among the 6 protocols for dataset #2. For instance, before ComBat, the plot shows a shift in distribution with greater Homogeneity values for reconstruction S5 than for S3 and FBP. This was expected since S5 involved higher noise reduction. After compensation, the distributions between the three reconstructions better overlapped. Figure 3 shows three examples of realignment of features between different reconstruction algorithms, reconstruction kernels and slice thicknesses.

## **Discussion**

As widely reported in the literature, radiomic features are sensitive to the acquisition and reconstruction parameters of CT images. Feature values are therefore not directly comparable between different imaging protocols, limiting their use in multicenter studies. Here, we demonstrated that the ComBat method can realign radiomic features computed from different CT imaging protocols. Using phantom data, we showed that ComBat removed the scanner and protocol effect while preserving the differences between texture patterns. The correction for the scanner effect was confirmed using patient images reconstructed with different imaging protocols.

The use of this compensation method should facilitate multicentric radiomic analyses that are absolutely needed to demonstrate the practical usefulness of radiomic features for patient management. Data harmonization is currently a hot topic in the international imaging community with increasing awareness of the need to reduce the variability in image quality between centers and machines (22,23). ComBat offers a solution to realign radiomic features with several advantages. ComBat is easily available to all and fast (a function available for free in R software). The transformations are estimated based on the measured feature values, without the need to go back to images or to perform phantom experiments. No learning set is needed. Unlike other options described in the literature, ComBat does not change the feature definitions (6,11) and can therefore be used with all software/algorithms and with any radiomic features. This is illustrated in three datasets using three different implementations for the radiomic feature calculation (Supplemental Table 5). ComBat does not require spatial resampling of the CT images to a single pixel size and/or image filtering (5). It is applicable when only radiomic features values are available or when images are not available. ComBat can account for covariates of interest if the patients scanned with different imaging protocols do not have the same characteristics (eg, different age distributions). ComBat can model these covariates in the compensation process as illustrated in PET imaging with different proportions of cancer subtypes in different departments (16), as long as enough patients with the same characteristics are available.

As demonstrated using the phantom data (Table 2), before compensation, the values of radiomic features were significantly different between imaging protocols for a given pattern. Ignoring this effect in multicentric studies might bias the findings and lower statistical power. A recent study highlighted that the 4 features selected in (13) to build the radiomic signature were highly correlated with tumor volume (24), which might explain why the model remained robust on data from different centers. Using ComBat might help determine whether radiomic features reflecting the lesion biological heterogeneity but affected by the center effect more than the lesion volume also have some predictive value.

Our study has some limitations: our findings should be confirmed on other cancer types, for other imaging protocols and scanners and the actual impact on diagnostic performance on clinical data needs to be demonstrated. Independent multi-center validation of radiomic models is also essential for them to become mainstream (1,25,26).

In summary, ComBat makes it possible to pool radiomic features from different CT protocols. This method appears promising to deal with the center effect in multicenter radiomic studies and to possibly raise the statistical power of such studies. ComBat is data-driven meaning that the transformations identified by ComBat to set all data in a common space should be estimated for each study involving data from different centers/protocols. Our analysis was based on less than 50 patients for each acquisition protocol demonstrating the efficiency of the method even for small patient cohorts. Using simulations in which we gradually removed patient data (results not shown), we found satisfactory results with as few as 20 patients per imaging protocol. The minimum number of patients required per imaging protocol to successfully apply ComBat remains to be comprehensively investigated.

## References

1. Gillies RJ, Kinahan PE, Hricak H. Radiomics: Images Are More than Pictures, They Are Data. *Radiology*. 2016;278(2):563–577. doi: 10.1148/radiol.2015151169.
2. Lubner MG, Smith AD, Sandrasegaran K, Sahani DV, Pickhardt PJ. CT Texture Analysis: Definitions, Applications, Biologic Correlates, and Challenges. *Radiographics*. 2017;37(5):1483–1503. doi: 10.1148/rg.2017170056.
3. Zhao B, Tan Y, Tsai WY, Schwartz LH, Lu L. Exploring Variability in CT Characterization of Tumors: A Preliminary Phantom Study. *Transl Oncol*. 2014;7(1):88–93. doi: 10.1016/j.tranon.2014.09.001.
4. Mackin D, Fave X, Zhang L, et al. Measuring Computed Tomography Scanner Variability of Radiomics Features. *Invest Radiol*. 2015;50(11):757–765. doi: 10.1097/RLI.000000000000180.
5. Mackin D, Fave X, Zhang L, et al. Harmonizing the pixel size in retrospective computed tomography radiomics studies. *PLoS ONE*. 2017;12(9):e0178524. doi: 10.1371/journal.pone.0178524.
6. Shafiq-Ul-Hassan M, Zhang GG, Latifi K, et al. Intrinsic dependencies of CT radiomic features on voxel size and number of gray levels. *Med Phys*. 2017;44(3):1050–1062. doi: 10.1002/mp.12123.
7. Caramella C, Allorant A, Orhac F, et al. Can we trust the calculation of texture indices of CT images? A phantom study. *Med Phys*. 2018;45(4):1529–1536. doi: 10.1002/mp.12809.
8. Berenguer R, Pastor-Juan MDR, Canales-Vázquez J, et al. Radiomics of CT Features May Be Nonreproducible and Redundant: Influence of CT Acquisition Parameters. *Radiology*. 2018;288(2):407–415. doi: 10.1148/radiol.2018172361.
9. Kim H, Park CM, Lee M, et al. Impact of Reconstruction Algorithms on CT Radiomic Features of Pulmonary Tumors: Analysis of Intra- and Inter-Reader Variability and Inter-Reconstruction Algorithm Variability. *PLoS ONE*. 2016;11(10):e0164924. doi: 10.1371/journal.pone.0164924.
10. Lu L, Ehmke RC, Schwartz LH, Zhao B. Assessing Agreement between Radiomic Features Computed for Multiple CT Imaging Settings. *PLoS ONE*. 2016;11(12):e0166550. doi: 10.1371/journal.pone.0166550.
11. Shafiq-Ul-Hassan M, Latifi K, Zhang G, Ullah G, Gillies R, Moros E. Voxel size and gray level normalization of CT radiomic features in lung cancer. *Sci Rep*. 2018;8(1):10545. doi: 10.1038/s41598-018-28895-9.
12. He L, Huang Y, Ma Z, Liang C, Liang C, Liu Z. Effects of contrast-enhancement, reconstruction slice thickness and convolution kernel on the diagnostic performance of radiomics signature in solitary pulmonary nodule. *Sci Rep*. 2016;6:34921. doi: 10.1038/srep34921.
13. Aerts HJWL, Velazquez ER, Leijenaar RTH, et al. Decoding tumour phenotype by noninvasive imaging using a quantitative radiomics approach. *Nat Commun*. 2014;5:4006. doi: 10.1038/ncomms5006.
14. Yan J, Chu-Shern JL, Loi HY, et al. Impact of Image Reconstruction Settings on Texture Features in 18F-FDG PET. *J Nucl Med*. 2015;56(11):1667–1673. doi: 10.2967/jnumed.115.156927.

15. Johnson WE, Li C, Rabinovic A. Adjusting batch effects in microarray expression data using empirical Bayes methods. *Biostatistics*. 2007;8(1):118–127. doi: 10.1093/biostatistics/kxj037.
16. Orlhac F, Boughdad S, Philippe C, et al. A Postreconstruction Harmonization Method for Multicenter Radiomic Studies in PET. *J Nucl Med*. 2018;59(8):1321–1328. doi: 10.2967/jnumed.117.199935.
17. Nioche C, Orlhac F, Boughdad S, et al. LIFEx: A Freeware for Radiomic Feature Calculation in Multimodality Imaging to Accelerate Advances in the Characterization of Tumor Heterogeneity. *Cancer Res*. 2018;78(16):4786–4789. doi: 10.1158/0008-5472.
18. Orlhac F, Soussan M, Chouahnia K, Martinod E, Buvat I. 18F-FDG PET-Derived Textural Indices Reflect Tissue-Specific Uptake Pattern in Non-Small Cell Lung Cancer. *PLoS ONE*. 2015;10(12):e0145063. doi: 10.1371/journal.pone.0145063.
19. Armato SG, Meyer CR, Mcnitt-Gray MF, et al. The Reference Image Database to Evaluate Response to therapy in lung cancer (RIDER) project: a resource for the development of change-analysis software. *Clin Pharmacol Ther*. 2008;84(4):448–456. doi: 10.1038/clpt.2008.161.
20. Fortin J-P, Cullen N, Sheline YI, et al. Harmonization of cortical thickness measurements across scanners and sites. *Neuroimage*. 2018;167:104–120. doi: 10.1016/j.neuroimage.2017.11.024.
21. Benjamini Y, Hochberg Y. Controlling the False Discovery Rate: A Practical and Powerful Approach to Multiple Testing. *Journal of the Royal Statistical Society Series B (Methodological)*. 1995;57(1):289–300. doi: 10.2307/2346101.
22. Sullivan DC, Obuchowski NA, Kessler LG, et al. Metrology Standards for Quantitative Imaging Biomarkers. *Radiology*. 2015;277(3):813–825. doi: 10.1148/radiol.2015142202.
23. O'Connor JPB, Aboagye EO, Adams JE, et al. Imaging biomarker roadmap for cancer studies. *Nat Rev Clin Oncol*. 2017;14(3):169–186. doi: 10.1038/nrclinonc.2016.162.
24. Vallieres M, Visvikis D, Hatt M. Dependency of a validated radiomics signature on tumor volume and potential corrections. *J Nucl Med*. 2018;59(supplement 1):640–640. [http://jnm.snmjournals.org/content/59/supplement\\_1/640](http://jnm.snmjournals.org/content/59/supplement_1/640)
25. Buvat I, Orlhac F, Soussan M. Tumor Texture Analysis in PET: Where Do We Stand? *J Nucl Med*. 2015;56(11):1642–1644. doi: 10.2967/jnumed.115.163469.
26. Reuzé S, Schernberg A, Orlhac F, et al. Radiomics in Nuclear Medicine Applied to Radiation Therapy: Methods, Pitfalls, and Challenges. *Int J Radiat Oncol Biol Phys*. 2018;102(4):1117–1142. doi: 10.1016/j.ijrobp.2018.05.022.

**Table 1:** Patient characteristics.


	Patient dataset #1	Patient dataset #2
<b>Sex-no. (%)</b>		
Male	19 (45%)	16 (50%)
Female	23 (55%)	16 (50%)
<b>Age-yr.</b>		
Mean	60.4	62.1
Range	31-81	29-82
<b>Lesion-no. (%)</b>		
Primary lung lesion	8 (19%)	29 (91%)
Metastatic lesion	34 (81%)	3 (9%)

**Table 2:** number of significant Friedman tests ( $p < .05$ ) without and with compensation in the phantom and clinical datasets.


	w/o realignment	with realignment
<b>Phantom (40 tests for each pattern)</b>		
Pattern 1: 20% filled ABS	40/40 (100%)	0/40 (0%)
Pattern 2: 30% filled ABS	40/40 (100%)	0/40 (0%)
Pattern 3: 40% filled ABS	40/40 (100%)	0/40 (0%)
Pattern 4: 50% filled ABS	40/40 (100%)	0/40 (0%)
Pattern 5: acrylic	40/40 (100%)	0/40 (0%)
Pattern 6: cork	40/40 (100%)	0/40 (0%)
Pattern 7: dense cork	39/40 (98%)	0/40 (0%)
Pattern 8: plaster resin	40/40 (100%)	0/40 (0%)
Pattern 9: rubber particles	40/40 (100%)	0/40 (0%)
Pattern 10: wood	40/40 (100%)	0/40 (0%)
<b>Patient dataset #1 (10 tests)</b>	10/10 (100%)	3/10 (30%)
<b>Patients dataset #2 (89 tests)</b>	87/89 (98%)	13/89 (15%)

ABS: acrylonitrile butadiene styrene


**Figure 1:** Phantom data: principal component scores for 2560 samples corresponding to 16 VOI x 10 texture patterns (in colors) x 16 imaging protocols described by 40 radiomic features on the first two principal components before (A) and after ComBat (B). After compensation, each texture pattern was no longer composed of different clusters, demonstrating that the scanner effect has been correctly compensated.




**Figure 2:** Example of ComBat application in phantom experiments: two texture patterns (cork and dense cork) were scanned using three different imaging protocols with 16 volumes of interest in each case (GE1, P2, S2, see Supplemental Table 2). When pooling all radiomic feature values, the optimal cutoff could not perfectly distinguish the patterns while after compensation of scanner effects, a perfect distinction was observed.




**Figure 3:** Probability density function(%) of Homogeneity before (left) and after ComBat (right) in patient data using two CT reconstruction algorithms (FBP and S5), two reconstruction kernels (Standard and Lung) and two voxel thicknesses (1.25 mm and 5 mm). CT images are from Kim et al, PLoS ONE 2016;11:e0164924 and Lu et al, PLoS ONE 2016;11:e0166550. Displayed p-values are given for Friedman tests.



**Supplemental Figure 1:** Phantom experiments: distinction between two patterns (Cork vs Dense cork) for 3 acquisition protocols with unbalanced groups (different numbers of cork and dense cork VOI for each scanner) without and with realignment.





**Supplemental Figure 2:** Probability density function of 4 radiomic features before (left) and after (right) ComBat for patient dataset #1 (A, B) and #2 (C, D). For these features, the p-values of Friedman test remained statistically significant after ComBat but these plots demonstrate that the correction was effective.





**Supplemental Figure 3:** Boxplot before (left) and after (right) compensation for the homogeneity feature. A) patient dataset #1 for 3 reconstruction algorithms (FBP, S3 and S5). B) patient dataset #2 for 6 different protocols involving three slice thicknesses (1.25 mm, 2.5 mm and 5 mm) and two reconstruction algorithms (lung-L and standard-S). Displayed p-values are given for Friedman tests.



A)



B)



**Supplemental Table 1: phantom characteristics.**

No.	Description	Illustration	
Pattern 1	Acrylonitrile butadiene styrene (ABS) plastic with 20% of air-filled (holes of diameter = 6.0 mm)		
Pattern 2	ABS plastic with 30% of air-filled (holes of diameter = 1.4 mm)		
Pattern 3	ABS plastic with 40% of air-filled (holes of diameter = 1.0 mm)		
Pattern 4	ABS plastic with 50% of air-filled (holes of diameter = 0.9 mm)		
Pattern 5	Solid polymethyl methacrylate (acrylic)		
Pattern 6	Standard cork		
Pattern 7	Dense cork		
Pattern 8	Plaster resin		
Pattern 9	Rubber particles		
Pattern 10	Sycamore wood		

**Supplemental Table 2:** acquisition characteristics for phantom experiments.

Designation	Manufacturer/Model	Reconstruction kernel	Scan type	Voxel size (mm <sup>3</sup> )	Spiral pitch factor	kVp	mAs
GE1	GE/Discovery CT750 HD	standard	helical	0.49x0.49x2.5	0.98	120	81
GE3	GE/Discovery CT750 HD	standard	helical	0.78x0.78x2.5	0.98	120	122
GE4	GE/Discovery ST	standard	helical	0.98x0.98x2.5	1.35	120	143
GE5	GE/LightSpeed RT	standard	helical	0.98x0.98x2.5	0.75	120	1102
GE6	GE/LightSpeed RT16	standard	helical	0.98x0.98x2.5	0.94	120	367
GE7	GE/LightSpeed VCT	standard	helical	0.74x0.74x2.5	0.98	120	82
P1	Philips/Brilliance Big Bore	B	helical	0.98x0.98x3	0.94	120	320
P2	Philips/Brilliance Big Bore	C	helical	0.98x0.98x3	0.94	120	369
P3	Philips/Brilliance Big Bore	B	helical	1.04x1.04x3	0.81	120	320
P4	Philips/Brilliance Big Bore	B	helical	1.04x1.04x3	0.81	120	369
P5	Philips/Brilliance 64	B	helical	0.98x0.98x3	0.67	120	372
S1	Siemens/Sensation Open	B31s	axial	0.52x0.52x2	1.00	120	26-70
S2	Siemens/Somatom Definition Flash	170f, 2	helical	0.54x0.54x3	0.60	120	17-28
T1	Toshiba/Aquilion	FC18	helical	0.63x0.63x3	1.11	120	135
T2	Toshiba/Aquilion	FC18	helical	0.63x0.63x3	1.11	120	135
T3	Toshiba/Aquilion One	FC18	helical	0.98x0.98x3	0.99	120	151

**Supplemental Table 3:** list of radiomic features extracted from phantom data. Complete description is available on [www.lifexsoft.org](http://www.lifexsoft.org).

Category	Features
Histogram	Mean, Standard-Deviation, Maximum, Skewness, Kurtosis, Entropy ( $\log_2$ and $\log_{10}$ ), Energy
Gray-Level Co-occurrence Matrix (GLCM)	Homogeneity, Energy, Contrast, Correlation, Entropy ( $\log_2$ and $\log_{10}$ ), Dissimilarity
Gray-Level Run Length Matrix (GLRLM)	Short-Run Emphasis (SRE), Long-Run Emphasis (LRE), Low Gray-level Run Emphasis (LGRE), High Gray-level Run Emphasis (HGRE), Short-Run Low Gray-level Emphasis (SRLGE), Short-Run High Gray-level Emphasis (SRHGE), Long-Run Low Gray-level Emphasis (LRLGE), Long-Run High Gray-level Emphasis (LRHGE), Gray-Level Non-Uniformity (GLNU), Run-Length Non-Uniformity (RLNU), Run Percentage (RP)
Neighborhood Gray-Level Different Matrix (NGLDM)	Coarseness, Contrast, Busyness
Gray-Level Zone Length Matrix (GLZLM)	Short-Zone Emphasis (SZE), Long-Zone Emphasis (LZE), Low Gray-level Zone Emphasis (LGZE), High Gray-level Zone Emphasis (HGZE), Short-Zone Low Gray-level Emphasis (SZLGE), Short-Zone High Gray-level Emphasis (SZHGE), Long-Zone Low Gray-level Emphasis (LZLGE), Long-Zone High Gray-level Emphasis (LZHGE), Gray-Level Non-Uniformity (GLNU), Zone-Length Non-Uniformity (ZLNU), Zone Percentage (ZP)

**Supplemental Table 4:** acquisition characteristics for patient datasets.

Characteristics	Patient dataset #1	Patient dataset #2
Manufacturer/Model	Siemens/Somatom Definition	GE/LightSpeed 16 or VCT
Voxel size (mm <sup>3</sup> )	0.68x0.68x1.00	[0.50x0.50x1.25]-[0.90x0.90x5]
Pitch factor	1	0.984-1.375
kVp	120	120
mAS	150	298-441

**Supplemental Table 5:** summary of data used in this study.

Dataset	Number of imaging protocols	Number of Volume of Interest	Number of radiomic features	Algorithm/software for the radiomic calculation	Number of Friedman tests
Phantom	16	16 by pattern (10 patterns)	40	LIFEx	40 features x 10 patterns → 400 tests
Patient dataset #1	3	42 patients x 3 segmentations = 126	10	In-house software by (9)	10 features → 10 tests
Patient dataset #2	6	32 patients x 2 acquisitions = 64	89	In-house software by (10)	89 features → 89 tests





**Supplemental Table 7:** p-values of Friedman tests before and after realignment corrected using Benjamini-Hochberg procedure for patient dataset #1. Values in red demonstrate significant differences at  $p < 0.05$ .

	w/o realignment	with realignment
i.Mean (HU)	<.001	.003
i.SD (HU)	<.001	.003
i.Skewness	<.001	.35
i.Kurtosis	<.001	.50
i.Entropy	<.001	.50
i.Homogeneity	<.001	.90
i.GLCM moments	<.001	.32
i.GLCM IDM	<.001	.99
i.GLCM Contrast	<.001	.001
i.GLCM Entropy	<.001	.90

**Supplemental Table 8:** p-values of Friedman tests before and after realignment corrected using Benjamini-Hochberg procedure for patient dataset #2. Values in red demonstrate significant differences at  $p < 0.05$ .

	w/o realignment	with realignment
Uni	<.001	.99
Bi	<.001	.27
Vol	<.001	.01
Intensity_Mean	<.001	.02
Intensity_SD	<.001	.98
Intensity_Skewness	<.001	>.99
Intensity_Kurtosis	<.001	.86
Intensity_Mean_MaxD	<.001	.02
Intensity_SD_MaxD	<.001	.98
Intensity_Skewness_MaxD	<.001	>.99
Intensity_Kurtosis_MaxD	<.001	.98
Shape_Compact-Factor	<.001	.21
Shape_Eccentricity_MaxD	.70	.38
Shape_Solidity_MaxD	<.001	.98
Shape_Round-Factor_MaxD	<.001	.98
Shape_SI2	<.001	.001
Shape_SI3	<.001	.01
Shape_SI4	<.001	.24
Shape_SI5	<.001	.49
Shape_SI6	<.001	.39
Shape_SI7	.08	.27
Shape_SI8	<.001	.14
Shape_SI9	<.001	.10
Boundary_Sigmoid-Amplitude-Mean-d5	<.001	.001
Boundary_Sigmoid-Slope-Mean-d5	<.001	.98
Boundary_Sigmoid-Offset-Mean-d5	<.001	.87
Wavelet_DWT-D	<.001	.86
Wavelet_DWT-V	<.001	.27
Wavelet_DWT-H	<.001	.91
Wavelet_DWT-LD	<.001	.57
Wavelet_DWT-LV	<.001	.39
Wavelet_DWT-LH	<.001	.49
EdgeFreq_Mean-d1	<.001	.98
EdgeFreq_Coarseness-d1	<.001	.001
EdgeFreq_Contrast-d1	<.001	.39
Fractal_Dimension-Mean	<.001	.85

GTDM_Coarseness-d1	<.001	<.001
GTDM_Contrast-d1	<.001	.39
GTDM_Busyness-d1	<.001	.005
GTDM_Complexity-d1	<.001	>.99
GTDM_Strength-d1	<.001	.39
Gabor_Energy-sum-w5	<.001	.97
Gabor_Energy-dir0-w5	<.001	.39
Gabor_Energy-dir45-w5	<.001	.98
Gabor_Energy-dir90-w5	<.001	.98
Gabor_Energy-dir135-w5	<.001	.86
Laws_Energy-1	<.001	.39
Laws_Energy-2	<.001	.77
Laws_Energy-3	<.001	.72
Laws_Energy-4	<.001	.85
Laws_Energy-5	<.001	.64
Laws_Energy-6	<.001	.98
Laws_Energy-7	<.001	.57
Laws_Energy-8	<.001	.39
Laws_Energy-9	<.001	.42
Laws_Energy-10	<.001	.57
Laws_Energy-11	<.001	.85
Laws_Energy-12	<.001	.57
Laws_Energy-13	<.001	.39
Laws_Energy-14	<.001	.21
LoG_MGI-s1	<.001	.48
LoG_Entropy-s1	<.001	.98
LoG_Uniformity-s1	<.001	.98
LoG_MGI-s4	<.001	.57
LoG_Entropy-s4	<.001	.98
LoG_Uniformity-s4	<.001	.14
Run_SPE	<.001	.98
Run_LPE	<.001	.98
Run_GLU	<.001	<.001
Run_PLU	<.001	.07
Run_PP	<.001	.98
Spatial_Corr-d1	<.001	.02
GLCM_ASM-mean-d1	<.001	.57
GLCM_Contrast-mean-d1	<.001	.98
GLCM_Corr-mean-d1	<.001	.71
GLCM_Sum-Squares-mean-d1	<.001	.86
GLCM_Homogeneity-mean-d1	<.001	.98
GLCM_IDM-mean-d1	<.001	.98
GLCM_Sum-Average-mean-d1	<.001	.007

GLCM_Sum-Variance-mean-d1	<.001	.005
GLCM_Sum-Entropy-mean-d1	<.001	.39
GLCM_Entropy-mean-d1	<.001	.41
GLCM_Diff-Variance-mean-d1	<.001	.98
GLCM_Diff-Entropy-mean-d1	<.001	.57
GLCM_IMC1-mean-d1	<.001	.86
GLCM_IMC2-mean-d1	<.001	.86
GLCM_MCC-mean-d1	<.001	.50
GLCM_Max-Prob-mean-d1	<.001	.64
GLCM_Cluster-Tendency-mean-d1	<.001	.52