



HAL
open science

Photoreceptor Detection in In-Vivo Adaptive Optics Images of the Retina: Statistical Validation

Kevin Loquin, Isabelle Bloch, Kiyoko Nakashima, Florence Rossant,
Pierre-Yves Boëlle, Michel Paques

► **To cite this version:**

Kevin Loquin, Isabelle Bloch, Kiyoko Nakashima, Florence Rossant, Pierre-Yves Boëlle, et al.. Photoreceptor Detection in In-Vivo Adaptive Optics Images of the Retina: Statistical Validation. ICIAR: International Conference Image Analysis and Recognition, Jun 2012, Aveiro, Portugal. hal-01959816

HAL Id: hal-01959816

<https://hal.science/hal-01959816>

Submitted on 19 Dec 2018

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

PHOTORECEPTOR DETECTION IN IN-VIVO ADAPTIVE OPTICS IMAGES OF THE RETINA: STATISTICAL VALIDATION

Kevin Loquin^{1*} Isabelle Bloch¹ Kiyoko Nakashima² Florence Rossant³ Pierre-Yves Boelle⁴ Michel Paques²

¹Telecom ParisTech - Institut Telecom - CNRS LTCI - Paris, France

kevin.loquin@gmail.com Isabelle.Bloch@telecom-paristech.fr

²CIC 503 of the XV-XX Hospital DHOS/INSERM - Paris, France

kn@cicoph.org mp@cicoph.org

³ISEP - Paris, France

Florence.Rossant@isep.fr

⁴Hôpital Saint-Antoine - Paris, France

pierre-yves.boelle@sat.aphp.fr

ABSTRACT

This article presents a photoreceptor detection algorithm applied to in-vivo Adaptive Optics (AO) images of the retina obtained from an advanced ophthalmic diagnosis device. Our algorithm is based on a recursive construction of thresholded connected components when the seeds of the recursions are the regional maxima of the deconvoluted image. This algorithm is validated on a gold standard dataset obtained thanks to manual cones detections made by ophthalmologist physicians.

Index Terms— Adaptive Optics, Photoreceptor detection, in vivo diagnosis, retina imaging

1. INTRODUCTION

Adaptive Optics (AO) is a new ophthalmic imaging system inherited from spectacular technological advances in astronomy [1]. An AO system is made up of a deformable mirror which compensates (it can be viewed as some kind of physical deconvolution) for the image defects due to atmospheric distortions. The imaging of the retina with an AO system [2] is particularly interesting because the light distortions due to the iris, pupil and lens system of the eye are similar to the distortions due to atmosphere.

The AO imaging modality technically outperforms any other up-to-date retinal imaging modality. It allows the imaging of the living human retina at the microscopic scale. For instance, the *rtx1* AO system has a resolution of $0.8\mu\text{m}$ per pixel, while usual fundus cameras reach resolutions from 6 to $10\mu\text{m}$ per pixel. This feature is illustrated in Figure 1.

*This work has been supported by French National Research Agency (ANR) through Tecsan Program (project iPhot n° ANR-09-TECS -009). K. Loquin is currently researcher at LITIS, Rouen and LIRMM, Montpellier, France




Fig. 1. Usual eye fundus image vs. Adaptive Optics image of the retina on a pathological eye

This new imaging modality is of prime importance for a clinical use in routine. At a microscopic level, the retina is a stack of several neuronal layers whose aim is to transform the incident light in an electrical signal that will be transmitted to the brain [3]. Within this process, the photosensitive neurons are the photoreceptor cells (rods and cones). Many inherited and acquired diseases or disorders provoke degeneration of the retina [4, 5], a progressive loss of the cones cells. It leads to severe visual handicap. In some diseases such as retinitis pigmentosa, visual loss can occur very early in life. In [6], the authors illustrate the diagnosis power of retinal AO images on degenerative cones dystrophy by comparing visual diagnosis procedures from these AO images (like cone counting, manual delineation of degenerated area,...) with well managed usual diagnosis procedures performed with other modalities.

Automated retinal AO image analysis methods for assisting in vivo diagnosis and follow up of early degenerative cones dystrophy are essential for physicians. In this paper, we propose an automatic photoreceptor detection algorithm on AO images. From this cornerstone procedure of automatic

cone detection on AO images, we can derive many clinically useful features such as photoreceptor density maps or extract other statistics.

The algorithm that we present in this article is an extension of an algorithm already proposed by the authors [7]. This algorithm is based on a recursive construction of thresholded connected components whose seeds are the regional maxima of the image. The stopping rules of this recursive construction are deduced from geometrical arguments: preference is given to brighter spots for closed merged spots which are thus considered as noisy spots. In this present article we propose to use an additional preliminary deconvolution step. The deconvolution of AO images has been proved to be particularly useful [8]: it considerably facilitates the distinction between background and information (cones) on retinal AO images.

Rare other automatic photoreceptors detection algorithms have already been proposed in the literature [9, 10]. Both are based on the detection of regional maxima; in [9], the image is pre-filtered and post morphologically processed and in [10], the regional maxima are processed in a decreasing intensity order to discriminate the dimmest ones. In this present article, we propose a validation of our algorithm by comparing it with the ones presented in [9, 10].

In Section 2, we present our photoreceptor detection algorithm. Then, Section 3 details the performed experiments to validate our tool, before concluding.

2. PHOTORECEPTOR DETECTION ALGORITHM

2.1. Previous version [7]

In the used AO images, the photoreceptors that we aim at detecting are bright spots, i.e. pixels with high grey levels. Hence a first natural step of our algorithm is a simple regional maxima detection. Afterwards, a maximum area of a regional maximum is defined up to a tolerance parameter, denoted by T . This tolerance parameter finds its roots in the component tree decomposition of an image [11]. In this scope, an image is seen as a decomposition of α -level connected components. The level range corresponding to the greyscale range of the image, an image is decomposed in a stack of connected components. The tolerance parameter T that we are using in our algorithm tunes the height of the top part of this stack that we consider as a maximum area.

Note that the proposed algorithm is a truncated adaptation of the watershed by immersion algorithm [12] when applied to the inverted image. In the inverted image, the origins of the basins used in the watershed by immersion are the regional maxima of the original image. In our case, we are not interested in reaching a complete separation of the basins (the maxima) but only in eliciting these basins until a given height, i.e. the maximum areas up to a tolerance parameter.

2.2. What's new ?

Even for adaptive optics which is physically corrected by means of deformable mirrors, an acquisition necessarily transforms the real information into an altered one: a measurement. The relationship between the physical reality and its measurement is supposed to be linear, which leads to model any acquisition by a convolution operator. Deconvolution of a signal is, in some sense, reversing the acquisition process to reach as close as possible, the physical reality. With AO retinal images, the results of the deconvolution are particularly interesting. In [8], the authors propose to learn the parameters of the deconvolution kernel, i.e. the point spread function (PSF) of the sensor model, from the entire image: this is called blind deconvolution. Their results are impressive but their algorithm is too much time consuming in a clinical context. In our case, we propose a simpler approach, less computationally demanding: we use a simple Gaussian convolution model of the PSF with a unique parameter: σ , its standard deviation.




Fig. 2. Deconvolution illustration

Figure 2 illustrates the interest of this technique. The right column shows (for illustration) the obtained deconvolved images from the left column images. On the left column we can also observe the results of our method for the following parameters: $\sigma = 5$; $T = 10$ for the first row and $\sigma = 6$; $T = 25$ for the second row. The results seem to be as good as with the blind deconvolution method. However, we add a new parameter to our method, which makes the physician task more complex.

2.3. Algorithm implementation: an interactive tool for the physicians

This algorithm is very rapid: after a complete cover of the image for finding the regional maxima, the recursive coverage is only applied to these extracted spots which highly diminishes the computational cost. When java-implemented as an imageJ plugin (which is not the most rapid computational framework) the algorithm result is quite instantly displayable. This remark is very important because it allows us to provide an interactive tool to the final user, the physician. This

non algorithmic part of our work is greatly appreciated by the physicians who can thus easily interact with the algorithm and adapt its parameters to his visual expertise in an intuitive way. This feature highly contributes to the fact that, even if still in development, this tool is already in use in a clinical context.

Another way to interact on the photoreceptors detection is to incorporate in our imageJ plugin the possibility to add non automatically detected cones and to remove wrongly detected cones by the hand.

3. VALIDATION

3.1. Material and Methods

The AO images used along this validation procedure are obtained with two prototypic systems: *INOVEO* and *rtx1* manufactured by Imagine Eyes, Orsay, France, <http://www.imagine-eyes.com/content/view/122/124>. Those systems are currently operational in clinical and research settings at the Clinical Investigation Center 503 of the *XV-XX* hospital. According to the used system (*INOVEO* and *rtx1*), in each image, for an axial length of $24.7mm$, one pixel corresponds respectively to a $1.2\mu m \times 1.2\mu m$ and $0.8\mu m \times 0.8\mu m$ spot on the retina.

The gold standard dataset is constructed from 52 patch images of 13 bigger AO images. Thus, for each of the 13 images, 4 patch images were extracted. Among the 4 images of one bigger 2 were exactly the same patch images. Those 52 images were presented in a random order to two ophthalmologist physicians of *XV-XX* hospital (Kiyoko Nakashima and Michel Paques). They marked the cones on these images thanks to a developed imageJ plugin which enabled us to recover the supposedly true cone positions. This procedure enables us to sort the images by quality: we consider that the less important is the difference (computed in terms of sensibility and false positive rate) between the manual detections obtained with the same patch images presented at two different times, the better is the quality of the global including image. Table 1 presents the images by increasing quality from left to right.

10	25	2	17	42	20	32	28	21	23	24	27	51
3	5	16	1	18	13	52	34	26	37	4	31	8
48	44	14	9	33	46	29	11	15	30	47	43	6
49	7	40	35	22	45	36	41	39	12	19	38	50

Table 1. Increasing quality image ordering

Not only we proposed to compare our algorithm with this gold standard but we also proposed to compare it with two other existing algorithms. In [9], the authors proposed a multi-step algorithm. First, they propose to work with an above-threshold version of the image in order to eliminate the dim spots. Afterwards, the regional maxima are detected from a low-pass filtered version of the thresholded image in

order to eliminate the noise. Finally, a morphological dilation is used to merge close enough detected spots. The structuring element is chosen according to the minimal cone spacing. In [10], the same kind of procedure is presented but the regional maxima detection is performed by decreasing greylevel regions in order to favor the brightest detected spots and avoid the threshold step of the method in [9].

In order to compare our method to [9, 10], we used Matlab implementations of these tools. The personal webpage of Kaccie Yiang Li <http://vision.berkeley.edu/roordalab/Kaccie/KaccieResearch.htm> provides the implementation of their method [9], while Jan Kopec, from the company Imagine Eyes provided the implementation of the method of Xue *et al.* [10].

Our comparison were based on (i) the *sensibility* (Se) index which is the percentage of identified cones by the algorithm which are well inside the gold standard dataset ($Se \in [0, 1]$) and (ii) the *false positive rate* (FPR) index which is the ratio of the number of identified cones by the algorithm which are not inside the gold standard dataset (number of false positives) over the number of cones in the gold standard ($FPR > 0$ and unbounded).

3.2. Results




Fig. 3. Maximum of $Se-FPR$ for each image and each compared method

First, we present a graph showing, for each studied method (Xue *et al.* [10] in blue, Li *et al.* [9], Loquin *et al.* [7] in black and our method in green) the maximum of the difference between Se and FPR for the set of possible parameterization for each method. The number of tested parameters were of 3052 for Xue, 1216 for Li, 7680 (30 standard deviations and 256 tolerances) for the method presented in this article and 256 for Loquin 2011. We took the maximum of $Se-FPR$ since the detection is better for higher value of $Se-FPR$. This remark leads to a simple interpretation of this graph: globally (apart for images 6 and 18) our method is better than the ones proposed in [9, 10]. We can observe a striking improvement of the detection with our algorithm on the bad quality images. Indeed, for patch images 3, 5, 7,

10, 48 and 49 which are all belonging to two worst images (cf. Table 1), Figure 3 shows that our method particularly outperforms the other ones.

Second, we present the ROC (Receiver Operating Characteristic) cloud and its associated logarithmic fit curve. The ROC curve has FPR in abscissa and Se in ordinate. Such curve is supposed to measure the ability of a decision making support to diagnose. The bigger the area under the curve the more discriminative is the proposed tool. It can be noticed that, according to this ROC analysis, our algorithm is more able to detect cones than the compared ones [9, 10].




Fig. 4. ROC cloud and fitted (logrithmic) ROC curve for each compared methods

From Figure 4, but in contradiction with Figure 3, we could conclude that the approach without deconvolution (the black curve) is more accurate than the method presented in this article (the black curve). This is due to the fact that the number of tested parameterizations with deconvolution is 30 times bigger than without deconvolution. Therefore, the number of aberrant parameters, thus leading to worst results with deconvolution. However this comparison bias due to the number of tested parameters can not be advocated for comparing the Li and Xue approach with the present one. Indeed, the number of tested parameters is higher in our approach than any of the other tested approaches.

4. CONCLUSION

In this paper, we proposed a new photoreceptor detection algorithm applied to in-vivo AO images of the retina. This algorithm is an extension of our work in [7] where we add a preliminary deconvolution step. Our method is implemented as an imageJ plugin already in use in a clinical setting at the XV-XX hospital. This use shows its early clinical value before any objective evaluation. However, in this paper, we also present a validation procedure of this algorithm which seems to show that our method is more efficient than the available rare litterature in that field especially for poor quality images.

5. REFERENCES

- [1] F. Roddier, *Adaptive Optics in Astronomy*, Cambridge University Press, 1999.
- [2] A. Roorda, F. Romero-Borja, W. Donnelly, H. Queener, T. Hebert, and M. Campbell, "Adaptive optics scanning laser ophthalmoscopy," *Optics Express*, vol. 10(9), pp. 405–412, 2002.
- [3] C.-A. Curcio, K.-R. Sloan, R.-E. Kalina, and A.-E. Hendrickson, "Human photoreceptor topography," *The Journal of comparative neurology*, vol. 292, no. 4, pp. 497–523, 1990.
- [4] S.J. Ryan, D.R. Hinton, A.P. Schachat, and P. Wilkinson, *Retina (3 Volume Set)*, Mosby, 2005.
- [5] P.T. de Jong, "Age-related macular degeneration," *The New England Journal of Medicine*, vol. 355(14), pp. 1474–1485, 2006.
- [6] J. I. Wolfing, M. Chung, J. Carroll, A. Roorda, and D. R. Williams, "High-resolution retinal imaging of cone-rod dystrophy," *Ophthalmology*, vol. 113, pp. 1014–1019, 2006.
- [7] K. Loquin, I. Bloch, K. Nakashima, F. Rossant, and M. Paques, "Photoreceptor detection in in-vivo adaptive optics images of the retina: towards a simple interactive tool for the physician," in *IEEE International Symposium of Biomedical Imaging (ISBI'11)*, Chicago, USA, 2011.
- [8] L. Blanco and L.M. Mugnier, "Marginal blind deconvolution of adaptive optics retinal images," *Opt Express*, vol. 19, no. 23, pp. 23227–39, November 2011.
- [9] K.Y. Li and A. Roorda, "Automated identification of cone photoreceptors in adaptive optics retinal images," *Journal of the Optical Society of America A*, vol. 24(5), pp. 1358–1363, 2007.
- [10] B. Xue, S.S. Choi, N. Doble, and J.S. Werner, "Photoreceptor counting and montaging of enface retinal images from an adaptive optics fundus camera," *Journal of the Optical Society of America A*, vol. 24(5), pp. 1364–1372, 2007.
- [11] M. Couprie, L. Najman, and G. Bertrand, "Quasi-linear algorithms for the topological watershed," *Journal of Mathematical Imaging and Vision*, vol. 22, no. 2-3, pp. 231–249, 2005.
- [12] L. Vincent and P. Soille, "Watersheds in digital spaces: an efficient algorithm based on immersion simulations," *IEEE Trans. Patt. Anal. Mach. Intell.*, vol. 13, no. 6, pp. 583–598, 1991.