



HAL
open science

Thermophysical properties of gas hydrates with stirred, high pressure calorimetric cells

Henry Delroisse, Jean-Philippe Torre, Didier Hailot, Laurent Marlin, Frédéric Plantier, Christophe Dicharry, Stéphan Moreau, Rémi André

► To cite this version:

Henry Delroisse, Jean-Philippe Torre, Didier Hailot, Laurent Marlin, Frédéric Plantier, et al.. Thermophysical properties of gas hydrates with stirred, high pressure calorimetric cells. 21st European Conference on Thermophysical Properties (ECTP21), Sep 2017, Graz, Austria. pp.296. hal-02006167

HAL Id: hal-02006167

<https://hal.science/hal-02006167>

Submitted on 4 Feb 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Open Archive Toulouse Archive Ouverte

OATAO is an open access repository that collects the work of Toulouse researchers and makes it freely available over the web where possible

This is an author's version published in: <http://oatao.univ-toulouse.fr/21869>

To cite this version:

Delroisse, Henry and Torr , Jean-Philippe  and Haillot, Didier and Marlin, Laurent and Plantier, Fr d ric and Dicharry, Christophe and Moreau, St phan and Andr , R mi *Thermophysical properties of gas hydrates with stirred, high pressure calorimetric cells*. In: European Conference on Thermophysical Properties (ECTP21), 3 September 2017 - 8 September 2017 (Graz, Austria)

Any correspondence concerning this service should be sent to the repository administrator: tech-oatao@listes-diff.inp-toulouse.fr

THERMOPHYSICAL PROPERTIES OF GAS HYDRATES WITH STIRRED, HIGH PRESSURE CALORIMETRIC CELLS

H. DELROISSE¹, J.-P. TORRÉ¹, D. HAILLOT², L. MARLIN³, F. PLANTIER¹, C. DICHARRY¹, Stéphan MOREAU⁴, Rémi ANDRÉ⁴

¹ *Univ. Pau & Pays Adour, CNRS, TOTAL – UMR CNRS 5150 – LFC-R – Laboratoire des Fluides Complexes et leurs Réservoirs, Avenue de l'Université, BP 1155 – PAU, F-64013, France.*

² *Univ. Pau & Pays Adour – EA1932 – LaTEP – Laboratoire de Thermique Energétique et Procédés, Rue Jules Ferry, BP7511 – PAU, F-64075, France*

³ *Institut Pluridisciplinaire de Recherche Appliquée dans le domaine du génie pétrolier (IPRA). Atelier de physique. Avenue de l'Université, BP 1155 – PAU, F-64013, France.*

⁴ *SETARAM Instrumentation, 7 Rue de l'Oratoire, CALUIRE, F-69300, France
e-mail (corresponding author): moreau@setaram.com*

High pressure differential scanning calorimetry (HP-DSC) is of importance in several fields involving gas hydrates, such as oil and gas production, flow assurance, carbon dioxide capture and storage, CO₂ hydrates reversible formation/dissociation for refrigeration loops.

However, the technique suffered for some limitations linked to the fact that the gas hydrate formation in the calorimetric cell occurs at the gas-liquids interface, leading to problems such as inefficient gas dissolution, formation of a hydrate crust covering the gas/liquid interface, low hydrate to water conversion, and difficulties to crystallize these compounds even at low temperature. It is for example rather difficult to determine accurately the heat capacities and the enthalpies of formation/dissociation of several systems involving gas hydrates.

To overcome such limitations, we present two prototypes of calorimetric cells equipped with an in-situ mechanical agitation system, which allow performing experiments under pressure (150 bar max). The first one is called MIXCEL®, and was developed for macro-calorimetry analysis (experiments carried out with a BT 2.15 Calvet Calorimeter from SETARAM Instrumentation). The second one, called MICROMIXCEL®, was developed for micro-calorimetry analysis (experiments carried out using a microDSC7 evo from Setaram Instrumentation). In this study, technical details of the two cells and results obtained both at macro and micro scales will be presented, and compared to the case with no agitation. Thermophysical properties of the cyclopentane hydrate (phase change enthalpy, and specific heat) will be given and commented. The use of such novel calorimetric cells opens a wide range of possibilities for complex systems, such as gas hydrates, which must be analysed in both pressurized and agitated conditions.