

Development and application of a novel electrochemical immunosensor for tetracycline screening in honey using a fully integrated electrochemical Bio-MEMS

Nadia El Alami El Hassani, Abdoullatif Baraket, Selim Boudjaoui, Ernandes Taveira Tenório Neto, Joan Bausells, Nezha El Bari, Benachir Bouchikhi, Abdelhamid Elaissari, Abdelhamid Errachid, Nadia Zine

▶ To cite this version:

Nadia El Alami El Hassani, Abdoullatif Baraket, Selim Boudjaoui, Ernandes Taveira Tenório Neto, Joan Bausells, et al.. Development and application of a novel electrochemical immunosensor for tetracycline screening in honey using a fully integrated electrochemical Bio-MEMS. Biosensors and Bioelectronics, 2019, 130, pp.330-337. 10.1016/j.bios.2018.09.052 . hal-02092892

HAL Id: hal-02092892 https://hal.science/hal-02092892

Submitted on 22 Oct 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

Development and application of a novel electrochemical immunosensor for tetracycline screening in honey using a fully integrated electrochemical Bio MEMS

- 4 Nadia El Alami El Hassani^{a,b,c}, Abdoullatif Baraket^c, Selim Boudjaoui^c, Ernandes Taveira
- Tenório Neto ^d, Joan Bausells ^e, Nezha El Bari ^a, Benachir Bouchikhi ^b, Abdelhamid Elaissari
 ^d, Abdelhamid Errachid ^c, Nadia Zine ^{c*}
- ^a Biotechnology Agroalimentary and Biomedical Analysis Group, Department of Biology,
 Faculty of Sciences, Moulay Ismaïl University, BP 11201, Zitoune, Meknes, Morocco.
- ^b Sensor Electronic and Instrumentation Group, Department of Physics, Faculty of Sciences,
 Moulay Ismail University, BP 11201, Zitoune, Meknes, Morocco.
- ^c Université de Lyon, Institut des Sciences Analytiques, UMR 5280, CNRS, Université Lyon
 1, ENS Lyon 5, rue de la Doua, F-69100 Villeurbanne, France.
- ^d Université de Lyon, LAGEP, UMR-5007, CNRS, Université Lyon 1, 5007, 43 Bd 11
 Novembre 1918, F-69622 Villeurbanne, France.
- ^e Instituto de Microelectronica de Barcelona, IMB-CNM (CSIC) Campus UAB, 08193,
- 16 Bellaterra, Barcelona, Spain.
- 17 **Corresponding author:* E-mail address: nadia.zine@univ-lyon1.fr

18 Abstract:

Tetracycline (TC) is a veterinary drug, wildly prescribed for prophylactic and therapeutic 19 purposes. Consequently, its remaining residues in food products have to be regularized. We 20 report in this paper about the development of a novel immunosensor based on an integrated 21 bio micro-electromechanical system (Bio-MEMS) containing eight gold microelectrodes 22 (µWEs), an integrated silver and platinum reference and counter electrodes, respectively. TC 23 immobilization on the μ WEs surface was conducted using three methods. The first through 24 functionalization with 4-aminophenylacetic acid (CMA), the second by functionalization with 25 CMA followed by preconcentration of a new structure of magnetic nanoparticles (MNPs) 26 27 coated with poly (pyrrole-co-pyrrole-2-carboxylic acid) (Py/Py-COOH/MNPs) cross-linked 28 with Ab-TC, and the last one directly through the functionalization with Py/Py-COOH/MNPs. The analyte was quantified by competitive detection with TC immobilized on the µWEs 29 surface toward specific polyclonal antibody (Ab-TC), using a mixture of a fixed concentration 30 of Ab-TC and decreasing levels of TC one from 0.1 pg.mL⁻¹ to 1000 pg.mL⁻¹. Microcontact 31 printing, followed by fluorescence microscopy characterization were performed during the 32 functionalization of the immunosensor surface to certify that the corresponding immune 33 detection process is taking place. This immunosensor was found to be highly sensitive with a 34 limit of detection of 1.2 pg.mL⁻¹ and specific in the presence of interferents. The standard 35 addition method was exploited to detect TC in honey samples. The present immunosensor 36 platform is up-and-coming for TC detection which can dramatically decrease the time of 37 analysis providing a new pathway for advanced immunoassays development in industrial food 38 control. 39

40 Keywords:

41 Bio-MEMS, Magnetic nanoparticles (MNPs), Immunosensor, Soft lithography, micro contact

42 printing, Food control.

43 **1. Introduction**

Tetracycline (TC) is a broad-spectrum antibiotic used as a veterinary drug to prevent 44 45 and to treat diseases caused by Gram-positive and Gram-negative bacteria, atypical organisms such as Chlamydiae, Mycoplasmas, Rickettsia, and protozoan parasites (Eliopoulos et al., 46 2003). The favorable antimicrobial properties of this drug and the absence of significant 47 48 adverse side effects have led to their extensive use in the therapy of both human and animal 49 infections (Chopra and Roberts, 2001). Furthermore, it has been reported as the most intensely growth-promoting agent widely used in animal husbandry (Economou and Gousia, 50 2015). Consequently, despite their proven efficacy, the intensive use of this drug is no longer 51 encouraged since its remain residues can be accumulated in the animal bones, as well as in 52 fish, meat, eggs, and milk, on account of the complexation of TC with Ca^{2+} (Nelson et al., 53 2010). To ensure human food safety, the European Union (EU) has established the maximum 54 residue limits (MRLs) for TC in edible products and milk at $100 \,\mu g.kg^{-1}$, in eggs at 55 $200 \ \mu g.kg^{-1}$ and in the liver at $300 \ \mu g.kg^{-1}$ (European Commission, 1999). However, the 56 57 MRLs for TCs in honey has not yet been established by Codex, since the treatment of 58 honeybee diseases with antibiotics is unallowable in the EU (Kivrak et al., 2016). Despite banning, some countries, such as Switzerland, UK, and Belgium still use TC treatment for 59 hive diseases and have established, therefore, the MRLs for TC ranging from 10 to 50 μ g.kg⁻¹ 60 61 (Bargańska et al., 2011).

However, TC residues can be successfully determined in various food products 62 requiring highly performing sample preparation steps. Several analytical methods currently 63 64 exist for TC detection and quantification. High performance liquid chromatography coupled to mass spectrometry (Pokrant et al., 2018), capillary electrophoresis coupled with 65 electrochemiluminescence (Deng et al., 2012), enzyme immunoassay (Gaurav et al., 2014), 66 67 liquid chromatography-mass spectroscopy (LC-MS) (Desmarchelier et al., 2018) have been described for confirmatory analysis. Other methodologies based on spectroscopy analysis are 68 described (Qin et al., 2016); unfortunately, they suffer from a lack of sensitivity compared to 69 chromatographic techniques (Mitra, 2004). Subsequently, Most of these approaches often 70 71 require the use of extraction techniques, such as liquid-liquid (Desmarchelier et al., 2018) and solid phase extraction (Shi et al., 2011), in order to accomplish both the analytes pre-72 73 concentration and sample cleanup (O'Connor and Aga, 2007). However, the principal limitations of these steps lie in the high costs, the long time of achievement and the 74 requirement of advanced technical skills (Joshi and Anderson, 2012). For this reason, simple, 75 rapid and accurate methods are recommended for the on-site screening of low TC residues 76 77 without needing extraction or clean-up steps. Due to their advantages of high selectivity, rapid detection, and *in-situ* applications, several optical and electrochemical techniques based on 78 79 biosensors have been investigated (Lan et al., 2017). Therefore, aptamer-based sensing techniques were widely used for the food safety determination. Notably, there is a growing 80 rise in the aptasensors fabrication for the TC detection (Han et al., 2018; Tang et al., 2017) 81 with a few applications in honey samples (Wang et al., 2015, 2014). Nevertheless, the main 82 limitations of these systems are related to their relative low detection signals (Yi et al., 2014). 83 Furthermore, in some recent works, the molecularly imprinted sensors had been successfully 84 applied to the analysis of antibiotic residues in honey samples (Bougrini et al., 2016; Lv et al., 85 86 2014). This type of sensors, however, often show relatively low sensitivity and specificity when the imprinted membrane is fragile. Hence, on the last few decades, the development of 87 immunosensors grew tremendously to the detection of drug residues in food. The use of these 88 analytical systems has attracted considerable attention because of the low detection limits and 89 90 the high selectivity for analyzing complex samples (Majdinasab et al., 2017). In this regard, some immunosensors were performed for TC detection in milk samples (Conzuelo et al., 91

92 2013; Liu et al., 2016) with the limits of detection being 0.032, 3.9, and 0.85 ng.mL⁻¹, 93 respectively.

94 Herein, a novel and highly sensitive approach is proposed for the detection of remaining TC residues in honey samples. This immunosensor was conducted on bio-micro-95 electro-mechanical systems (Bio-MEMS) transducers based on gold micro working electrodes 96 97 (μWEs) with fully integrated, reference and counter electrodes. The novelty of this work is related to the functionalization process of µWEs, which was achieved through two- and three-98 dimensional (2D and 3D) shapes. These functionalizations were conducted, in one hand, by 99 electroaddressing diazonium salt and, in the other hand, by electrodepositing of a new 100 structure of magnetic nanoparticles coated poly(pyrrole-co-carboxylic acid) (Py/Py-101 COOH/MNPs), while combining the two methods with a preconcentration technique. The 102 Pv/Pv-COOH/MNPs as solid supports of biomolecules, is the pioneering aspect of this work 103 dedicated to the combination of submicron, magnetic conducting particles. The quantification 104 of this analyte was performed through competitive detection procedure with TC molecules 105 immobilized on the µWEs surface toward polyclonal TC antibody. The prepared 106 immunosensor was successfully applied to enhance the detection limit of TC in complex 107 matrices such as honey. 108


109 **2. Experimental**

110 2.1. Reagents and solutions

Tetracycline (TC), Doxycycline (DX), Oxytetracycline (OX), Chlortetracycline (CT), 111 112 1-Ethyl-3-(3-dimethylaminopropyl)carbodiimide (EDC), Ethanolamine, Sulfuric acid (H₂SO₄) (30 wt.% in H₂O), 4-aminophenylacetic acid (CMA), sodium nitrite (NaNO₂), 113 Sodium cyanoborohydride (95%), hydrochloric acid (37%, HCl), Carbon tetrachloride, 114 heptane, octadecyltrichlorosilane (OTS), Pure ethanol (99.8%), Acetone (99.9%), Potassium 115 hexacyanoferrate II, Potassium hexacyanoferrate III, Phosphate buffer saline (PBS) with pH 116 7.4 were all purchased from Sigma-Aldrich, France. N-Hydroxysuccinimide (NHS) and 117 Hydrogen peroxide (H₂O₂) (35%, v/v) were purchased from Acros Organics, France. 11-118 (triethoxysilyl)undecanal (TESUD) was purchased from abcr GmbH, Germany. 119 Polydimethylsiloxane (PDMS: Sylgard 184) was purchased from Dow Corning, France. The 120 sheep polyclonal antibody raised against TC (Ab-TC) was supplied by Abcam, France. A 121 fluorescent labeled antibody was generated by conjugation of Ab-TC with rhodamine using 122 the rhodamine fast conjugation kit (ab188286) by Abcam, France. Rosemary honey sample, 123 guaranteed tetracycline-free was supplied by Secrets d'Apiculteur, Lyon, France. Honey 124 sample with an unknown concentration of TC was commercially provided. Milli-Q ultrapure 125 water was used in all experiments. 126

127 **2.2 Procedure for immunosensor elaboration**

The development of this immunosensor was conducted by three methods (Fig. 1). Before any functionalization, the Bio-MEMS device surface was cleaned with ethanol in an ultrasonic bath for 10 min, dried under nitrogen flow, then placed under UV-Ozone Cleaner– ProCleanerTM Plus from BioForce for 30 min to remove all organic contaminants.



132

Fig.1. Schematic illustrations of competitive immunoassay of TC on modified μWEs by
 functionalization with: (A) CMA electroaddressing, (B) CMA electroaddressing followed by
 preconcentration of immunomagnetic nanoparticles cross-linked with Ab-TC antibody, (C)
 Py/Py-COOH/MNPs.

137

2.2.1 Diazonium immobilization onto gold microelectrodes

The cleaned microelectrodes (µWEs) were functionalized by CMA solution (3 mM) 138 previously diazotized in an aqueous solution of HCl (20 mM) and NaNO₂ (20 mM) for 15 139 min at 4 °C. Subsequently, nine-repetitive cyclic voltammograms were applied from 0.3 V to 140 -1.0 V with a scan rate of 80 mV.s⁻¹. This was done sequentially until all desired µWEs were 141 142 modified with CMA. The carboxylic groups formed on the electrode surfaces were activated by a mixture of 0.4 M EDC and 0.1 M NHS, prepared in ethanol for one hour at room 143 temperature. Afterwards, the activated electrode surfaces were incubated in 40 µL of TC at 144 100 μ g.mL⁻¹ for 30 min at 4 °C. Then, the immunosensor was treated with ethanolamine (1% 145 v/v) in PBS buffer (pH 7.4) for 10 min at room temperature. This step is crucial to prevent 146 nonspecific bonding phenomenon at the detection stage of TC (Fig. 1A). 147

148 **2.2.2 Preconcentrating technique**

The core-shell magnetic nanoparticles coated with poly(pyrrole-co-pyrrole carboxylic 149 acid) (Py/Py-COOH/MNPs) were exclusively synthesized by (Tenório-Neto et al., 2016) 150 using a seeded-polymerization technique. The morphology of these nanoparticles was already 151 presented in our previous work (Hassani et al., 2017) using transmission electron microscopy 152 153 (TEM). Here, the magnetic nanoparticles stock solution (100 µL) was washed three times with PBS buffer (pH 7.4) in a magnetic field to separate the nanoparticles from the storage 154 solution. EDC and NHS solutions (each at a concentration of 100 mM) were prepared 155 immediately before use. A volume of 250 µL each of EDC and NHS solutions were added to 156 157 the washed magnetic nanoparticles and incubated with slow rotation at room temperature for 90 min. Then, the reaction mixture was rinsed three times with HCl (1 mM) at 4 °C in a 158 magnetic field. Subsequently, 100 μ L of purified Ab-TC polyclonal antibody (100 μ g.mL⁻¹) 159 was added to the activated nanoparticles and incubated for 2 h at 4 °C. The antibody-coated 160 nanoparticles were washed three times with PBS buffer (pH 7.4). The non-reacted activated 161 carboxylic acid groups were blocked with ethanolamine (1% v/v) in PBS buffer for 10 min. 162 The antibody-coated magnetic nanoparticles were then separated from the mixture, 163 resuspended in 500 µL of PBS buffer and incubated with different concentrations of TC (Fig. 164 1B). Simultaneously, the gold **µWEs** were functionalized by CMA and TC as described in 165 166 paragraph 2.2.1.

167

2.2.3 Functionalization with Py/Py-COOH/MNPs


This functionalization was performed by electrodepositing Py/Py-COOH/MNPs onto 168 microelectrodes by using pulsed chronoamperometric technique (PCA). Briefly, two 169 continuous repetition potentials were applied: the first potential $E_1=-0.2$ V for $t_1=0.1$ s, 170 171 induces a cathodic electrodeposition process, while the second potential $E_2=0.2$ V for $t_2=0.1$ s oxidizes the electrode surface, and in particular, cleans it from eventual extraneous deposited 172 species. The cycling process was iterated 30 times. Then, the gold <u>uWEs</u> were rinsed with 173 174 distilled water. Besides, it is interesting to point out that the amplitude of the charge-discharge process remains practically unchanged during the prolonged cycling process, indicating that 175 the electrode surface exhibits a homogenous Py/Py-COOH/MNPs deposition. The activation 176 of carboxylic groups, present in the Py/Py-COOH/MNPs, was carried out by the mixture of 177 EDC/NHS (0.4 M/0.1 M) prepared as described above. The biofunctionalization and blocking 178 steps were performed by incubating the sensor in TC solution (100 μ g.mL⁻¹) and 179 ethanolamine/PBS (1% v/v), respectively. Finally, the immunosensor was rinsed with PBS 180 and used for TC detection (Fig. 1C). The homogeneity and the roughness of the resulted 181 **WEs** surface were already examined in our preceding study (Hassani et al., 2017). This was 182 conducted by two-, and three-dimensional surface topography using atomic force microscopy 183 (AFM). The interest of using this technique is to ensure that the μ WE surfaces were entirely 184 homogeneous with the presence of spherical shapes indicating the presence of MNPs coated 185 with Py/Py-COOH. 186

187 2.3. Micro-contact printing

The soft-lithographical technique called microcontact printing (µCP) of self-188 assembled monolayers (SAMs) was employed to check the immune TC/Ab-TC reactivity. For 189 this purpose, an elastomeric stamp based on polydimethylsiloxane (PDMS) was fabricated by 190 replica molding (RM) as already described in previous work (Baraket et al., 2013). A mixture 191 192 of pre-polymer PDMS and curing agent (10:1 w/w) was poured onto a silanized silicon-mold which contains micro-pillars on relief of its surface (Fig. 2a). Subsequently, complete 193 degassing of PDMS/silanized silicon-mold was carried out to ensure that all air bubbles which 194 195 may create defects on the PDMS stamp surface have been removed (Fig. 2b). After the polymerization step at 52 °C for 1 hour, the PDMS stamp was peeled-off from the silicon 196

mold bearing the micro holes on its surface (Fig. 2c). The stamp was then inked by immersion 197 for 1 min in heptane solution containing octadecyltrichlorosilane (OTS) (0.1%, v/v) and 198 199 carbon tetrachloride (2%, v/v), then dried with a stream of nitrogen (Fig. 2d). Simultaneously, glass substrates were cleaned with acetone then ethanol, rinsed with distilled water and then 200 dried with nitrogen. The substrates were activated by submersion in piranha solution (3:1 v/v201 202 of H₂SO₄: H₂O₂) for 30 min then thoroughly rinsed with distilled water and dried with nitrogen. Afterwards, the µCP was realized by putting the prepared PDMS stamp in 203 immediate and conformal contact with a freshly activated glass substrate (Fig. 2e). The 204 PDMS stamp was peeled-off from the substrate, leaving then the SAMs of OTS on the glass 205 surface which was incubated afterwards at 100 °C for 45 min to enhance OTS adhesion. 206

The freshly prepared substrates were submerged in an ethanol solution containing 1% of 207 TESUD, rinsed with ethanol, dried with a stream of nitrogen and then heated in a vacuum 208 stove at 100 °C for 60 min to complete the silanization process (Fig. 2f). Subsequently, the 209 functionalized substrate surface (FSS) was incubated in 250 µL of Ab-TC antibody (10 210 μ g.mL⁻¹) already diluted in 4 mM of sodium cyanoborohydride. The use of this reagent was 211 conducted to avoid adversely reducing aldehydes to nonreactive hydroxyls, allowing covalent 212 bonding of Ab-TC with the amine of TESUD (Fig. 2g). After that, the FSS were incubated in 213 TC solution (0.1 $pg.ml^{-1}$) for 1 h to allow the antibody-antigen interaction followed by rinsing 214 in PBS and drying under nitrogen (Fig. 2h). Finally, the rhodamine-labeled polyclonal Ab-215 TC^* (10 µg.ml⁻¹) were dropped on the FSS for 1 h (Fig. 2i). The samples were then rinsed in 216 PBS, dried under nitrogen and observed by fluorescence microscopy. 217



218

Fig. 2. Microcontact printing process to immobilize Ab-TC onto TESUD patterns and sandwich immunoassay with TC and rhodamine-labeled Ab-TC.

221 **2.4. Fluorescence microscopy characterization**

Fluorescence images, as a rapid tool showing the recognition of Ab-TC for its corresponding antigen (TC), were taken using a fluorescence microscope (Zeiss Axioplan 2 Imaging), equipped with $10 \times$ and $40 \times$ lenses and a monochrome camera. Samples were observed by fluorescent light: TC sample was excited with a 550 (±25) nm band-pass filter and fluorescence from the sample was observed with a 605 (±70) nm band-pass filter.


227 **2.5. Electrochemical measurements**

All steps of gold μ WEs modification were characterized by cyclic voltammetry (CV) and electrochemical impedance spectroscopy (EIS) techniques using a multi-channel potentiostat (VMP3 Bio-logic-Science Instrumentation, France). The electrochemical configuration was performed by using integrated μ WEs, RE and CE of the Bio-MEMS. All the measurements were performed in a solution of 5.0 mM [Fe(CN)₆]^{3-/4-} prepared in PBS buffer (pH 7.4) and in a Faraday cage. For the cyclic voltammetric measurements, the potential was scanned from 0.3 V to 0.6 V with a scan rate of 100 mV.s⁻¹. While the electrochemical impedance was measured in the 150 mHz-200 kHz frequency range with a DC potential of -0.3 V/ref and an AC potential of 90 mV.

237 **3. Results and discussion**

238 **3.1 Fluorescence analysis**

Fluorescent imaging is considered as an effective tool for bio-functionalization 239 analysis. It's used to certify that the corresponding immune detection process is taking place 240 based upon the recognition of the antibody for its corresponding antigen. In this study, 241 fluorescent antibodies were immobilized (chemically) onto a glass substrate for the detection 242 of TC antigens. Fig. 3A-B shows the fluorescent pattern of Ab-TC^{*R}. The red fluorescent 243 regions indicate tags in which the antibodies are specifically immobilized, while the extinct 244 areas are those blocked with OTS. In the Fig.3C, we notice an absence of fluorescence due to 245 the non-cross-reactivity of the doxycycline (DX), as interfering molecules, towards Ab-TC. 246 247 This immuno-fluorescent test has reported the strong reactivity of Ab-TC against TC which can, therefore, be applied for immunosensor development. 248



249

Fig. 3. Fluorescent images of $(Ab-TC^{*R})$ chemically immobilized on glass substrates after recognition of TC, (A) Positive homogenous pattern with the blackened regions corresponding to the OTS blocked glass surface. (B) Magnification of the positive pattern showing specific detection between Ab-TC^{*R} and TC. (C) Fluorescent image of the biorecognition assay between doxycycline and Ab-TC^{*R}.

255 **3.2 Electrochemical characterization of modified gold µWEs**

The behaviors of CMA and Py/Py-COOH/MNPs films were checked for each step of the 256 Bio-MEMS functionalization by CV analysis. The process of CMA electro-addressing is 257 shown in Fig. S1A (Supplementary information). The initial cycle shows a large and 258 irreversible cathodic wave with a peak potential at -0.8 V revealing diazotated CMA onto the 259 gold μ WEs surface by diazonium salt reduction. However, the reduction wave has been kept 260 261 unchanged upon the deposition of CMA. This functionalization was checked by variation of the CV before and after gold uWEs modification (Fig. S1B) (Supplementary information). In 262 this figure, a significant decrease in peak-to-peak of the CV cycles was noticed after CMA 263 electro-deposition against the bare gold µWEs. This was referred to the broad passivation 264 region of this microelectrodes. 265

While Fig. S1C (Supplementary information) shows the chrono-amperometric process for 266 Py/Py-COOH/MNPs electrodeposition. During this step it was interesting to specify that the 267 amplitude of the charge-discharge process remains nearly unchanged, indicating that the 268 electrode surface displays homogenous Py/Py-COOH/MNPs layer. CV analysis were also 269 investigated to examine the behavior of the µWEs surfaces after Py/Py-COOH/MNPs 270 electrodeposition. Fig. S1D (Supplementary information) shows the considerable increase in 271 the peak current of cyclic voltammograms after Py/Py-COOH/MNPs deposition compared to 272 bare gold. This was explained by the conductive behavior of Py/Py-COOH coated on MNPs. 273 274 At these steps, the modified gold µWEs were ready for bio-functionalization with the target analyte. 275

3.3 Optimization of conditions for detection

The amount of TC immobilized on the surface of gold µWEs as well as the antibody
 concentration were subject to optimization described in (Supplementary information).

279 **3.4 Competitive detection of TC**

In this study, the quantification of the analyte has been made by the competitive 280 281 approach to enhance the detection signal of TC with a small molecular weight (444 Da) compared to Ab-TC (160 kDa). Furthermore, this approach was advantaged in the present 282 work to make pre-concentration allowing to detect only TC and eliminate other irrelevant 283 molecules. This step was performed by incubating the µWE surfaces of the Bio-MEMS in a 284 mixture of a fixed concentration of Ab-TC previously optimized (10 μ g.mL⁻¹) and decreasing 285 levels of standard solutions of TC from 1000 pg.mL⁻¹ to 0.1 pg.mL⁻¹ during 30 min at 4 °C for 286 each concentration. At least three independents replicated experiments were carried out. 287 Ultimately, the electrode surface was rinsed with PBST (10 mM PBS with 0.05% Tween 20) 288 and analyzed with EIS. Fig. 4A-C show the obtained Nyquist plot semi-circles for the three 289 functionalizations of uWE surfaces. It can be noticed that for the three approaches, the 290 291 Nyquist plot semi-circles increase by decreasing TC concentrations. As awaited, this level decreases produce an increase in the charge transfer resistance owed to the presence of free 292 antibodies reacting with TC immobilized on the uWEs surface. 293

An excellent fitting between the simulated and experimental spectra was obtained for 294 each TC concentration by using the Randles equivalent circuit shown in the inset of Fig. 4D 295 (where R_s is the solution resistance, R_{ct} is the charge-transfer resistance, W is the Warburg 296 impedance and Q is the constant phase element). For each TC concentration, the value of 297 $\Delta R/R_0$ was calculated when R is the charge transfer resistance. The normalized data show 298 three linear equations: $y_a = -0.208 \text{ x} + 0.960$, $y_b = -0.148 \text{ x} + 0.578$ and $y_c = -0.109 \text{ x} + 0.474$ 299 300 with the determination coefficients of 0.962, 0.986 and 0.946, respectively. As can be seen from this figure, the detection sensitivity was highly improved for the Py/Py-COOH/MNPs 301 gold surface modification. For this approach, the configured impedimetric immunosensor 302 have provided the more sensitivity of 0.208 mL.pg^{-1} with a limit of detection (LOD) of 1.2 303 304 pg.mL⁻¹. It is appealing to notice that this LOD value is nearly 83000 times lower than the

- maximum residue limits of TC in honey fixed by European regulation. As listed in Table 1,
- the proposed system provides relatively wide linear ranges and low LODs in comparison with other published studies.




Fig. 4. Nyquist impedance plots in 5 mM of $[Fe(CN)_6]^{3-/4-}$ for competitive detection of TC on the modified µWEs by: (A) Functionalization with CMA, (B) Functionalization with CMA followed by preconcentration step, (C) Functionalization with Py/Py-COOH/MNPs, where (a) Bare gold, (b) After TC immobilization at 100 µg.mL⁻¹, (c-g) After competitive detection of TC at 1000, 100, 10, 1 and 0,1 pg.mL⁻¹, respectively. (D) The sensitivity of the immunosensor by normalization of Nyquist plot data, inset: the equivalent circuit used for EIS fitting.

Table 1. Comparison of different techniques for TC detection.

Methods	Linear range (ng.mL ⁻¹)	LOD (ng.mL ⁻¹)	Real sample	References
High-performance liquid- chromatography with fluorescence detection	15-5000	5	Milk	(Kargin et al., 2016)
Photoelectrochemical aptasensor	0.2-1000.0	0.01	Drug sample	(Han et al., 2018)
Terahertz spectroscopy	0-20×106	0.45	Water and	(Qin et al., 2017)

Enzyme-linked immunosorbent assay and immunochromatographic assay	0.26–2.00	15	Milk and honey	(Chen et al., 2016)
Electrochemical immunoassay	0.05-100	0.006	Honey milk and peanuts	(Que et al., 2013)
Molecularly imprinted polymer mixed with solid- phase extraction	20-600	20	Milk	(Xie et al., 2018)
Electrochemical immunosensor based on the chitosan-magnetic nanoparticles	0.08-1.00	0.03	Milk	(Liu et al., 2016)
Amperometric immunosensor	0.0005 -500	0.86	Milk	(Conzuelo et al., 2013)
This work	0.0001-1	0.0012	Honey	-

milk

318 **3.5 Interferences study**


The selectivity of the proposed system was investigated in the presence of various antibiotics from the same family of cyclines that can have cross-reactivity with Ab-TC antibody. This study was performed by comparing the responses for TC detection with three analogous molecules namely: chlortetracycline (CT), doxycycline (DX), and oxytetracycline (OX) (Fig. S3) (Supplementary information). Consequently, this immunosensor was found to be very selective for TC and can provide reliable results regardless of the presence of the interfering molecules.

326 **3.6 Analysis of spiked honey samples**

The TC determination in real samples was performed by spiking analyte in blank honey matrices which indicate that the calibration curve is suitable to determine the TC content in real unknown samples (Supplementary information). The results of the added spiked samples are shown in Table S1 (Supplementary information). Good recoveries from 80% to 98% were obtained with the RSD ranging between 1% and 6%, which indicate that the calibration curve is suitable to determine the TC content in real unknown samples.

333 3.7 Detection of TC in unknown honey sample

The standard addition method was further employed to determine the TC 334 concentration in unknown honey sample. The competitive detection of TC was performed 335 by a mixture of different standard solutions (3 pg.mL⁻¹, 5 pg.mL⁻¹, 10 pg.mL⁻¹, 15 pg.mL⁻¹) 336 with the presence of Ab-TC (10 µg.mL⁻¹). For all TC concentrations, the normalized charge 337 transfer resistance $\Delta R/R_0$ were measured. The extrapolation of regression line predicts the 338 TC concentration in the diluted honey sample at 2.52 pg.mL⁻¹. Consequently, the 339 concentration of TC in the unknown sample was 25.2 pg.mL⁻¹ according to the dilution 340 341 factor of 10 times (Fig. 5). This level of contamination remains lower than the MRLs of 10 000 pg.mL⁻¹ established by Switzerland, UK, and Belgium. 342



343

Fig. 5. Standard addition plots of different TC spiking in diluted unknown honey sample,
using (a) Ab-TC and (b) Ab-155 antibodies.

346 **4.** Conclusions

In the present work, we have discussed the development process of a novel sensitive 347 348 and highly selective immunosensor based on 2D and 3D network of immobilization with the diazonium salt, and a new structure of MNPs coated with Py-COOH for specific detection of 349 TC. Moreover, the electrochemical analysis performed by these materials allow high 350 reproducibility and rapid response. The high sensitivity of this immunosensor was evaluated 351 for the three functionalization methods and was highly improved for the Pv/Pv-352 COOH/MNPs gold surface modification. For this approach, the configured impedimetric 353 immunosensor have provided a sensitivity of 0.208 mL.pg⁻¹ with a limit of detection of 1.2 354 pg.mL⁻¹. The selectivity of the present immunosensor platform toward TC was higher when 355 compared to analogous molecules, namely DX, CT and OX. We have presented a novel and 356 less costly fabrication process to develop a highly sensing device able to detect low levels of 357 358 TC residues in honey samples without any complicated pretreatment. The miniaturization of 359 this immunosensor will be very beneficial to control the quality of food products on an industrial scale. However, there are still some limitations of this work related to the 360 361 construction time which is a little long. Therefore, our further investigations in this direction are to shorten the development time of this immunosensor, to meet the requirements of a 362 large scientific community which can extend its use for other applications. Future works will 363 also be done, to assess the multi-functionalization of the same device allowing a 364 simultaneous detection of several analytes. 365

366 Acknowledgements

The authors acknowledge the financial support from PHC-Toubkal under the project No.32567YD, and funding from the European Union's Horizon 2020 research and innovation programme entitled HEARTEN under grant agreement No 643694" and microMole grant agreement No 643694 and No 653626, respectively.

371 **References**

Baraket, A., Lee, M., Zine, N., Sigaud, M., Yaakoubi, N., Trivella, M.G., Zabala, M.,
Bausells, J., Jaffrezic-Renault, N., Errachid, A., 2013. Diazonium modified gold
microelectrodes onto polyimide substrates for impedimetric cytokine detection with an
integrated Ag/AgCl reference electrode. Sensors and Actuators B: Chemical 189, 165–
172.

- Bargańska, Ż., Namieśnik, J., Ślebioda, M., 2011. Determination of antibiotic residues in
 honey. TrAC Trends in Analytical Chemistry 30, 1035–1041.
- Bougrini, M., Florea, A., Cristea, C., Sandulescu, R., Vocanson, F., Errachid, A., Bouchikhi,
 B., El Bari, N., Jaffrezic-Renault, N., 2016. Development of a novel sensitive molecularly
 imprinted polymer sensor based on electropolymerization of a microporous-metal-organic
 framework for tetracycline detection in honey. Food Control 59, 424–429.
- Chen, Y., Kong, D., Liu, L., Song, S., Kuang, H., Xu, C., 2016. Development of an ELISA
 and immunochromatographic assay for tetracycline, oxytetracycline, and chlortetracycline
 residues in milk and honey based on the class-specific monoclonal antibody. Food
 Analytical Methods 9, 905–914.
- Chopra, I., Roberts, M., 2001. Tetracycline antibiotics: mode of action, applications,
 molecular biology, and epidemiology of bacterial resistance. Microbiology and molecular
 biology reviews 65, 232–260.
- Conzuelo, F., Campuzano, S., Gamella, M., Pinacho, D.G., Reviejo, A.J., Marco, M.P.,
 Pingarrón, J.M., 2013. Integrated disposable electrochemical immunosensors for the
 simultaneous determination of sulfonamide and tetracycline antibiotics residues in milk.
 Biosensors and Bioelectronics 50, 100–105.
- Deng, B., Xu, Q., Lu, H., Ye, L., Wang, Y., 2012. Pharmacokinetics and residues of
 tetracycline in crucian carp muscle using capillary electrophoresis on-line coupled with
 electrochemiluminescence detection. Food Chemistry 134, 2350–2354.
- Desmarchelier, A., Anizan, S., Minh Tien, M., Savoy, M.-C., Bion, C., 2018. Determination
 of five tetracyclines and their epimers by LC-MS/MS based on a liquid-liquid extraction
 with low temperature partitioning. Food Additives & Contaminants: Part A.
- Economou, V., Gousia, P., 2015. Agriculture and food animals as a source of antimicrobial resistant bacteria. Infection and drug resistance 8, 49.
- Eliopoulos, G.M., Eliopoulos, G.M., Roberts, M.C., 2003. Tetracycline therapy: update.
 Clinical infectious diseases 36, 462–467.
- 404 European Commission, 1999. European Commission (1999) Off J Eur Union L 60:16–52.
- Gaurav, A., Gill, J.P.S., Aulakh, R.S., Bedi, J.S., 2014. ELISA based monitoring and analysis
 of tetracycline residues in cattle milk in various districts of Punjab. Veterinary World 7.
- Han, Q., Wang, R., Xing, B., Chi, H., Wu, D., Wei, Q., 2018. Label-free
 photoelectrochemical aptasensor for tetracycline detection based on cerium doped CdS
 sensitized BiYWO 6. Biosensors and Bioelectronics 106, 7-13
- Hassani, N.E.A.E., Baraket, A., Neto, E.T.T., Lee, M., Salvador, J.-P., Marco, M., Bausells,
 J., Bari, N.E., Bouchikhi, B., Elaissari, A., Errachid, A., Zine, N., 2017. Novel strategy for
 sulfapyridine detection using a fully integrated electrochemical Bio-MEMS: Application
 to honey analysis. Biosensors and Bioelectronics 93, 282–288.
- Joshi, M.D., Anderson, J.L., 2012. Recent advances of ionic liquids in separation science and
 mass spectrometry. Rsc Advances 2, 5470–5484.
- Kargin, I.D., Sokolova, L.S., Pirogov, A.V., Shpigun, O.A., 2016. HPLC determination of
 tetracycline antibiotics in milk with post-column derivatization and fluorescence
 detection. Inorganic Materials 52, 1365–1369.
- Kivrak, I., KIVRAK, Ş., Harmandar, M., 2016. Development of a rapid method for the
 determination of antibiotic residues in honey using UPLC-ESI-MS/MS. Food Science and
 Technology (Campinas) 36, 90–96.
- Lan, L., Yao, Y., Ping, J., Ying, Y., 2017. Recent advances in nanomaterial-based biosensors
 for antibiotics detection. Biosensors and Bioelectronics 91, 504–514.
- Liu, X., Zheng, S., Hu, Y., Li, Z., Luo, F., He, Z., 2016. Electrochemical immunosensor based
 on the chitosan-magnetic nanoparticles for detection of tetracycline. Food Analytical
 Methods 9, 2972–2978.

- Lv, Y.-K., Zhang, J.-Q., He, Y.-D., Zhang, J., Sun, H.-W., 2014. Adsorption-controlled
 preparation of molecularly imprinted hybrid composites for selective extraction of
 tetracycline residues from honey and milk. New Journal of Chemistry 38, 802–808.
- Majdinasab, M., Yaqub, M., Rahim, A., Catanante, G., Hayat, A., Marty, J.L., 2017. An
 Overview on Recent Progress in Electrochemical Biosensors for Antimicrobial Drug
 Residues in Animal-Derived Food. Sensors 17, 1947.
- 433 Mitra, S., 2004. Sample Preparation Techniques in Analytical Chemistry. John Wiley & Sons.
- Nelson, M.L., Dinardo, A., Hochberg, J., Armelagos, G.J., 2010. Brief communication: mass
 spectroscopic characterization of tetracycline in the skeletal remains of an ancient
 population from Sudanese Nubia 350–550 CE. American journal of physical
 anthropology 143, 151–154.
- 438 O'Connor, S., Aga, D.S., 2007. Analysis of tetracycline antibiotics in soil: advances in
 439 extraction, clean-up, and quantification. TrAC Trends in Analytical Chemistry 26, 456–
 440 465.
- Pokrant, E.V., Maddaleno, A.E., Araya, C.E., San Martínb, B.V., Cornejo, J., 2018. In-House
 Validation of HPLC-MS/MS Methods for Detection and Quantification of Tetracyclines
 in Edible Tissues and Feathers of Broiler Chickens. J. Braz. Chem. Soc 29, 659–668.
- 444 Qin, J., Xie, L., Ying, Y., 2017. Rapid analysis of tetracycline hydrochloride solution by
 445 attenuated total reflection terahertz time-domain spectroscopy. Food chemistry 224, 262–
 446 269.
- Qin, J., Xie, L., Ying, Y., 2016. A high-sensitivity terahertz spectroscopy technology for
 tetracycline hydrochloride detection using metamaterials. Food chemistry 211, 300–305.
- Que, X., Chen, X., Fu, L., Lai, W., Zhuang, J., Chen, G., Tang, D., 2013. Platinum-catalyzed
 hydrogen evolution reaction for sensitive electrochemical immunoassay of tetracycline
 residues. Journal of Electroanalytical Chemistry 704, 111–117.
- Shi, X., Meng, Y., Liu, J., Sun, A., Li, D., Yao, C., Lu, Y., Chen, J., 2011. Group-selective
 molecularly imprinted polymer solid-phase extraction for the simultaneous determination
 of six sulfonamides in aquaculture products. Journal of Chromatography B 879, 1071–
 1076.
- Tang, Y., Zhang, J., Liu, J.-H., Gapparov, I., Wang, S., Dong, Y., Su, H., Tan, T., 2017. The
 development of a graphene oxide-based aptasensor used for the detection of tetracycline
 in honey. Analytical Methods 9, 1133–1140.
- Tenório-Neto, E.T., Baraket, A., Kabbaj, D., Zine, N., Errachid, A., Fessi, H., Kunita, M.H.,
 Elaissari, A., 2016. Submicron magnetic core conducting polypyrrole polymer shell:
 Preparation and characterization. Materials Science and Engineering: C 61, 688–694.
- Wang, S., Liu, J., Yong, W., Chen, Q., Zhang, L., Dong, Y., Su, H., Tan, T., 2015. A direct
 competitive assay-based aptasensor for sensitive determination of tetracycline residue in
 honey. Talanta 131, 562–569.
- Wang, S., Yong, W., Liu, J., Zhang, L., Chen, Q., Dong, Y., 2014. Development of an
 indirect competitive assay-based aptasensor for highly sensitive detection of tetracycline
 residue in honey. Biosensors and Bioelectronics 57, 192–198.
- Xie, Y., Hu, Q., Zhao, M., Cheng, Y., Guo, Y., Qian, H., Yao, W., 2018. Simultaneous
 Determination of Erythromycin, Tetracycline, and Chloramphenicol Residue in Raw Milk
 by Molecularly Imprinted Polymer Mixed with Solid-Phase Extraction. Food Analytical
 Methods 11, 374–381.
- Yi, X., Li, L., Peng, Y., Guo, L., 2014. A universal electrochemical sensing system for small
 biomolecules using target-mediated sticky ends-based ligation-rolling circle
 amplification. Biosensors and Bioelectronics 57, 103–109.
- 475