



HAL
open science

Time resolved transient circular dichroism spectroscopy using synchrotron natural polarization

François Auvray, David Denetiere, Alexandre Giuliani, Frédéric Jamme,
Frank Wien, Bastien Nay, Séverine Zirah, François Polack, Claude
Meneglier, Bruno Lagarde, et al.

► To cite this version:

François Auvray, David Denetiere, Alexandre Giuliani, Frédéric Jamme, Frank Wien, et al.. Time resolved transient circular dichroism spectroscopy using synchrotron natural polarization. *Structural Dynamics*, 2019, 6, pp.054307. 10.1063/1.5120346 . hal-02351456

HAL Id: hal-02351456

<https://hal.science/hal-02351456>

Submitted on 6 Nov 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Time resolved transient circular dichroism spectroscopy using synchrotron natural polarisation

François Auvray^{a,b}, David Denetiere^b, Alexandre Giuliani^{b,c}, Frédéric Jamme^b, Frank Wien^b, Bastien Nay^d, Séverine Zirah^e, François Polack^b, Claude Menneglier^b, Bruno Lagarde^e, Jonathan D. Hirst^a,
Matthieu Réfrégiers^{b*}

a School of Chemistry, University of Nottingham, University Park, Nottingham NG7 2RD, United Kingdom

b Synchrotron SOLEIL, L'Orme des Merisiers, Gif-sur-Yvette 91192, France

c UAR 1008 CEPIA, INRA, Nantes 44316, France

d. Laboratoire de Synthèse Organique, CNRS UMR 7652, Ecole Polytechnique, Palaiseau 91128 Cedex, France

e. Unité Molécules de Communication et Adaptation des Microorganismes (MCAM), Muséum national d'Histoire naturelle, CNRS, Paris 75005, France

f. Brain Physiology Lab - CNRS UMR 8118, Université Paris Descartes, 45 Rue des Saints Pères 75270 Paris Cedex 06, France

Abstract

Ultraviolet (UV) synchrotron radiation circular dichroism (SRCD) spectroscopy has made an important contribution to the determination and understanding of the structure of biomolecules. In this paper, we report an innovative approach that we term time-resolved SRCD (tr-SRCD), which overcomes the limitations of current broadband UV SRCD setups. This technique allows accessing ultrafast time scales (down to nanoseconds), previously measurable only by other methods, such as infrared (IR), nuclear magnetic resonance (NMR), fluorescence and absorbance spectroscopies and small angle X-ray scattering (SAXS). The tr-SRCD setup takes advantage of the natural polarisation of the synchrotron radiation emitted by a bending magnet to record broadband UV CD faster than any current SRCD setup, improving the acquisition speed from 10 mHz to 130 Hz and the accessible temporal resolution by several orders of magnitude. We illustrate the new approach by following the isomers concentration changes of an azopeptide after a photoisomerisation. This breakthrough in SRCD spectroscopy opens up a wide range of

* Corresponding author: matthieu.refregiers@synchrotron-soleil.fr

potential applications to the detailed characterisation of biological processes, such as protein folding, protein-ligand binding.

Keywords: time-resolved spectroscopy, circular dichroism, synchrotron, biomolecular dynamics

Introduction

Circular dichroism (CD) is an optical property of molecules having chiral structure(s) and/or spatially oriented arrays of chromophores. It manifests itself as a difference in absorption for left- and right-circularly polarised light. In the ultraviolet (UV) range. This feature has been exploited for decades for the characterisation of organic molecules, materials with supramolecular chirality, and in protein conformation determination, where there are distinctive spectral signatures for each secondary structure type, i.e., α -helices and β -sheets¹. Thus, CD spectroscopy is an important biophysical tool for characterising native and modified proteins. In the biomedical context, protein misfolding can have dramatic consequences on cell physiology, causing serious neurodegenerative diseases, as found in Alzheimer's and Parkinson's². Structural and kinetics studies of protein folding, using time-resolved approaches, are providing crucial insights at the molecular level into the aetiology of these diseases.

There are two main ways to measure CD spectra: the ellipsometric method and the direct absorption difference detection method. The former is based on quantifying the variations of the ellipticity and azimuth orientation of a highly elliptically polarised beam passing through a dichroic sample³. In 2012, Eom *et al.*⁴ innovatively adapted the ellipsometric method to a heterodyne-detection technique, providing both the CD spectrum and the optical rotation dispersion spectrum, by analysis of the phase and the amplitude of the transmitted orthogonal electric field of the incident light polarisation. More recently, Hiramatsu *et al.*⁵ coupled the heterodyne detection technique to a singular value decomposition analysis. This improvement removes linear dichroism and linear birefringence artifacts, allowing accurate time-resolved CD (tr-CD) measurements in the visible range with sub-picosecond temporal resolution. The second way to acquire a CD spectrum is based on absorption difference measurement. The light is alternately circularly right and left polarised using an optical or acousto-optic modulator. Then the detection system records intensity variations and allows CD

determination. This method has been extensively refined, and can provide sub-picosecond resolution for monochromatic measurements⁶. Hache and colleagues used this technique to probe ultra-fast kinetics in biomolecules⁷ and achieved 800 fs temporal resolution⁸. Several groups^{9,10} have developed a setup for picosecond transient broadband CD measurements in the visible range. They used a combination of a Pockels cell and a light continuum generator to obtain a pulsed white light source alternately circularly right and left polarised. Recently Preda and coworkers have introduced a new experimental arrangement based on the combination of time-domain detection with heterodyne amplification¹¹. This approach allows time-resolved optical activity measurements. Both ellipsometric and absorption difference broadband techniques have advantages and disadvantages, but they also often share the same spectral restriction to the visible range, notably due to the lack of stable continuum light sources in the far-UV range.

The synchrotron radiation circular dichroism (SRCD) technique was developed in 1980^{12,13}. Since its first use for protein structure determination¹⁴, it has been used in a wide range of applications¹⁵⁻¹⁷. Indeed, the brilliance of the synchrotron radiation (SR) in the vacuum UV (VUV) range and its stability enables one to measure CD spectra of samples dissolved in buffer down to 160 nm¹⁸ with an acceptable signal to noise ratio for the heterodyne detection.

In this paper, we describe the development of a different approach for SRCD measurements, utilising the natural polarisation of the SR emitted by a bending magnet^{19,20}. The DISCO beamline provides a SR beam composed of two parts (Supplementary Information Fig. 1), which are above and below the electron's orbital plane in the synchrotron storage ring. These two continuum beams are similarly elliptically polarised but with opposite direction (Supplementary Information 2). Schiller and Hormes²¹ have previously demonstrated that the natural elliptical polarisation of the SR can be used for CD spectroscopy. However, they measured the CD signal sequentially, wavelength by wavelength; so the scanning process of

the monochromator limited the temporal resolution of their setup significantly. We have developed a spectrograph that measures simultaneously the intensity of both (oppositely polarised) parts of the SR beam, allowing one to determine a whole UV CD spectrum in just a single measurement. While current SRCD setups only take advantage of the wide spectral range and the brilliance of SR, our setup also uses its natural polarisation and its temporal distribution. This broadband single measurement approach gives access to a combination of a temporal and a spectral range previously inaccessible with the current broadband CD setups, enabling new insights into biomolecular dynamics.

Method

The tr-SRCD setup is annexed to the existing SRCD endstation at the DISCO beamline²². The optical layout is shown Figure 1. The beamline excitation monochromator is used first order for the wavelength calibration of the detection spectrograph (Supplementary Information Fig. 3). With the zero order, we get on the sample a pulsed white beam containing all the wavelengths between 120 nm to 600 nm; this pulsed broadband source is used for the tr-SRCD measurement, limited by the CMOS spectral response. The beam is spatially defined by a spatial filter (labelled S1 in the figure), then centred by the two plane mirrors (M'1, M'2) before being refocused on the sample. In order to avoid damage from beam irradiation, this optical system is designed to allow variation of the illuminated area on the sample. The spherical mirror M'3 and the cylindrical mirror M'4 refocused the beam 2975 mm after the monochromator slits; the calculated beam diameter at the focal plane is about 300 μm (FWHM). By moving the sample cell by 200 mm along the optical axis, we get a controllable probe beam diameter from 300 μm up to 3 mm. Adjustment of the beam diameter to the exposure time helps to minimize the sample irradiation dose. Beyond the sample, the cylindrical mirror M'5 and the spherical mirror M'6 horizontally focused the beam on a

secondary slit (S2). This secondary slit at 3900 mm is used to define the final spectral resolution of the setup. The flat M'7 mirror reflected the beam on a spherical flat field grating with 580 grooves per mm (Horiba Jobin-Yvon) that diffract and focus the incident UV/vis light horizontally onto the 2D detector.




Figure 1: Left: Optical layout of the tr-SRCD setup with distances not at scale. Right: Detection sequence for single pulse measurement combined to a triggered system.

The tr-SRCD setup acquisition frequency is defined by the camera frame rate, which can be set from 0.033 Hz to 100 Hz in full resolution mode and can be increased further by reducing the pixel array size. This range of sampling frequency allows one to study reaction kinetics from tens of milliseconds to minutes through ~~real-time recording~~ and sequential measurements. The image intensifier can be synchronized with the camera output clock and with an external function generator providing trigger signals with a higher frequency. When the intensifier is synchronized with the output clock, the temporal resolution of the measurement corresponds to the duration of the amplification gate or the light pulse length if the gate is sufficiently short (Supplementary Information 4). However, if an external device triggers the image intensifier at a higher frequency than the camera frame rate, the exposure time of the

camera defines the temporal resolution of the setup. The intensifier gate only amplifies the light coming from one SR pulse. The intensifier allows overcoming the limitation brought by the minimal exposure time of the CMOS (50 μ s) and so measures the intensity of one pulse at once (Figure 1 right). We use this protocol for the results presented in Figure 2b. We integrated 500 pulses for each recorded image for this measurement.

The sample was held in a CaF₂ flow cell with a path length of 24 μ m (Aquaspec, Microbiolytics GmbH, Freiburg, D). The sensitivity ranges of 2D detectors, such as CDD or CMOS, are usually limited to 200 nm. We may also use an image intensifier (HS-IRO, LAVISION) to reach lower wavelength range and to increase time resolution. To recover the main part of the visible light provided by the HS-IRO phosphor, a combination of two 50 mm focal length infinity corrected objectives was used. A CMOS detector (Prime 95-B, PHOTOMETRICS) records the phosphor screen image. For the applications that do not need the picosecond temporal resolution, the Prime 95-B was used without intensifier.

Equation 1 defines CD at a given wavelength. The variables A_L and A_R correspond to the absorption for left and right-handed circularly polarised light, respectively. All the spectra associated with the tr-SRCD setup in this study were determined by the following equation.

$$CD = (A_L - A_R) \quad \text{Equation 1}$$

In this case, the absorption difference between left-handed and right-handed circularly polarised light can be associated with the absorption difference between the upper and lower part of the SR (Supplementary Information Note 5). These portions of the image are selected in order to show opposite polarizations. Although such procedure reduces the photon flux at each wavelength, the masks applied remove at most 25% of the images intensity, giving a moderate effect on the photon flux useful for the measurements.

Results

We chose camphorsulfonic acid (CSA), purchased from Sigma-Aldrich, as our calibration sample, as it has two strong peaks in the VUV/deep UV range: the first at 290 nm and the second at 192 nm. It is commonly used as a calibration standard for CD spectrometers^{23,24}. To validate that the recorded signal is CD, the two enantiomers must exhibit opposite signed spectra.

The resulting spectra for D-CSA and L-CSA, measured between 190 nm and 315 nm for 30 mg/mL solution concentration, are shown in Figure 2a. The CMOS detector has been used without the image intensifier. It integrated the light for 500 μ s and the images were acquired at 20 Hz. As expected, we observe two oppositely signed spectra with identical amplitudes. The ratio between the amplitude of the two peaks should be about 2²⁴; we obtained 2.11 \pm 0.11 in the tr-SRCD spectra (curve (d) of Figure 2a). The signal-to-noise ratios at 190 nm for the D-CSA spectra (a), (b), (c) and (d) in Figure 2a are 6.6, 12.7, 39.8 and 44.40, respectively.




Figure 2: **a.** CD spectra of D-CSA (black) and L-CSA (red) acquired with the tr-SRCD setup at 20 Hz with 500 μ s intensifier gate duration. 1, 10, 50 and 100 measurements were integrated to obtain the data shown in curves (a), (b), (c) and (d) respectively. **b.** CD spectra of D-CSA (black) and L-CSA (red) from single pulse measurements acquired with the tr-SRCD setup at 500 Hz with 82 ps intensifier gate duration; 5000 measurements were integrated.

In order to establish the highest theoretical temporal resolution of the setup that is attainable experimentally, we added an image intensifier, thereby obviating the temporal resolution limitation of the CMOS detector. We performed this steady state experiment on the CSA with the maximal temporal resolution as a proof-of-principle. The intensifier was triggered with an external trigger and its amplification was reduced to its shortest value, 100 ns. The measurements were made using the 8-bunch mode of the synchrotron SOLEIL which provides 82 ps pulses at 7.14 MHz. The intensifier amplified only one pulse per gate. So its temporal resolution corresponds to the pulse length, which, in this case, is 82 ps. In Figure 2b, each spectrum comes from ten images; 500 separated pulses were integrated for each image. The spectral range was reduced to between 255 nm and 300 nm, in order to optimize the setup for the detection of the peak at 290 nm. The signal-to-noise ratio at 290 nm for the D-CSA spectrum is 10.33.




Figure 3: Schematic representation of the FK-11-X molecular system for both *cis* and *trans* conformation. Its amino acid sequence is Ac-Glu-Ala-Cys^{AZO}-Ala-Arg-Glu-Ala-Ala-Ala-Arg-Glu-Ala-Ala-Cys^{AZO}-Arg-Gln-NH₂.

We tested the capacity of the setup to follow reactions in real time by studying a photocontrolled and reversible system, the azobenzene crosslinked peptide FK-11-X^{25,26}. It is composed of a 16 amino acid peptide cross-linked to a photoswitchable molecule: azobenzene (Figure 3). This ligand can be isomerised from the *trans* to *cis* and *cis* to *trans* conformation using a 370 nm and a visible light source (460 nm), respectively. Several groups have studied the dynamics of this photo-induced reaction and the isomerisation of the azobenzene appears to occur within one picosecond²⁷⁻³⁰. This conformational change constrains the peptide structure and triggers its folding and unfolding processes. Theoretical^{25,31} and experimental^{32,33} studies agree that, following the azobenzene isomerisation, peptide conformational changes occur on the microsecond scale. The azobenzene cross-linked peptide was synthesized as described previously²⁵ (Supplementary Information Note 6). The aim of our study was to follow the change in concentration of the unfolded (*cis* azobenzene) and folded (*trans* azobenzene, ground

state) FK-11-X peptide. We triggered the isomerisation with two continuous diodes, at 370 nm (5W) and at 460 nm (5W). The concentration of FK-11-X solution was 10 mg/mL in phosphate buffer (70 mM, pH 7). The sample was alternately irradiated with the two diodes. The sample was first irradiated 2.5 seconds with the 370 nm diode to trigger the *trans* to *cis* isomerisation and then 2.5 seconds with the 460 nm diode to trigger the *cis* to *trans* isomerization. The transient CD and absorption spectra of isomers concentration changes were followed, measuring spectra at 130 Hz with a 520 μ s temporal resolution; the measurement has been integrated over 39 cycles of alternate irradiations. The variations in the variations of absorption and CD are shown in Figure 4.




Figure 4: FK-11-X evolution from *trans* to *cis* over time after 460 nm triggering and relaxation after 370 nm triggering. **a.** Evolution of the absorbance difference during the 370 nm irradiation. One spectrum every 70 ms is represented. **b.** 2D representation of the evolution of the global absorbance

over the cycle; 650 spectra are shown **c.** Evolution of the CD difference during the 370 nm irradiation. One spectrum every 70 ms is represented. **d.** Evolution of the CD at 190 nm (x) and 204 nm (+) over the cycle.

The isomers concentration change induces a variation in the absorption of the sample. The evolution of the absorption measured by the detector is shown in Figure 4a and b. The absorption variation (Figure 4.a) shows two peaks, at 198nm and 212nm. It decreases progressively and reaches a plateau where $\Delta A \approx 0$ in the whole range of interest because the intensity reference is measured from these time frames. A clear decrease in intensity occurs from the first milliseconds after the 370 nm irradiation. The irradiation at 460 nm that triggers the *cis* to *trans* isomerisation starts at 2.5 seconds. The measured intensity increases progressively over the following 2.5 seconds to reach the initial value measured at t_0 . This plot shows that the *trans* to *cis* isomerisation is completed for the main part of the molecules after 1 second of irradiation with the 370 nm diode. This plateau does not correspond to an equilibrium between the diode flux and the lifetime of the *cis*-isomer. Indeed, the FK-11-X peptide is relatively stable in its *cis* configuration and it needs tens of minutes to naturally switch from the *cis* to the *trans* configuration. It also shows the complete reversibility of the concentration ratio. The intensity reference I_0 for the determination of the absorbance variation and so for the CD calculation corresponds to the average of the intensity measured between 1 and 2.5 seconds. In this time range, almost all the molecules are in their *cis* configuration. Thus, the absorbance and CD variation measurements shown in the Figure 4 correspond to the variation from the absorbance and CD spectrum of a solution of *cis* FK-11-X peptides. The CD variation spectra (Figure 4c and S5) are close to the signal of the α -helical structure and the increasing amplitude reflects the change of the species concentration. The first spectrum corresponds to the total difference between the CD spectrum of the unfolded (*cis*) FK-11-X peptide and its folded (*trans*) configuration. Then the concentration of the *trans* configuration decreases progressively during the 370 nm irradiation and so the amplitude of the CD difference decreases as well until it reaches approximately 0 before the first second of irradiation. The CD variation at 204 nm and 190 nm (Figure 4d) also confirms the reversibility of the system, the CD difference returns to its initial value at t_0 after the 2.5 seconds of irradiation with the 460 nm diode. The total CD changes from the initial state and final state measured with the tr-SRCD setup (Figure S5) and from steady state measurement (*cis* and *trans* CD spectra) in the literature³² are similar. This comparison shows that being able to calibrate the SR polarisation, we can measure transient CD spectra similar to those measured

with steady state setups. The only difference arising from the elliptical polarisation is the amplitude of the CD, which is linearly correlated to the circularity of the polarisation.

Our study demonstrates that the new tr-SRCD setup developed at the DISCO beamline²² of the SOLEIL synchrotron radiation facility offers new opportunities for the investigation of biomolecular structure and dynamics. The setup is based on the use of the natural properties of the synchrotron radiation provided by the DISCO beamline. It uses both spectral and polarisation characteristics to reduce the time required for the measurement of a transient CD UV spectrum. Our tr-SRCD measurements of the CSA show that an integration time of 500 μs is enough to obtain an acceptable signal-to-noise ratio, i.e., greater than 6. We performed a proof-of-principle high temporal resolution steady state experiment on CSA performed using an image intensifier allowing resolution down to a few tens of picoseconds. The FK11 experiment demonstrates the capacity of the tr-SRCD setup to follow dynamics in real time, with a temporal resolution of 500 μs and acquired at 130 Hz. These values do not correspond to the hardware cap and can be easily improved, by opening the spectrograph slits and integrating more photons in shorter time period and/or coupling the system with the image intensifier. This breakthrough in the SRCD spectroscopy will enable the study of many biological and chemical reactions crucial for our understanding of biomolecular phenomena.

The development of new CaF₂ microlenses windows for sCMOS camera would grant access to direct detection and measurement of spectra on the full DUV range down to 160 nm.

In the near future, the setup could easily be coupled to other reaction triggers, such as a T-jump, lasers and microfluidic stop-flow equipment. Thus, this new approach will provide exciting insights into the dynamics of biomolecules, as well as for molecular and materials systems more broadly. This method will enable one to follow the behaviour of molecules through high quality SRCD spectra on a temporal range from the picoseconds to minutes. Indeed, it covers time scales consistent with the fluctuation and domain motions of proteins

suspected to play a fundamental role in their activity. Ongoing experiments involve the kinetic study of the refolding of model proteins after thermal denaturation and the following of DNA nano-structure formations³⁴.

Supplementary Material See supplementary material for details on 1) the beam, 2) the polarisation characterisation, 3) the wavelength calibration, 4) the detection scheme with an intensifier, 5) the detailed algorithm, 6) the chemical steps of purification.

Acknowledgment Access to synchrotron SOLEIL was supported under project #20151217, #20161257, #20161315. We thank the University of Nottingham and SOLEIL for a PhD studentship for F.A.

Competing Interest None.

Data availability statement Raw data were generated at the synchrotron SOLEIL large scale facility and are available from the corresponding author upon reasonable request.

References

1. Greenfield, N. & Fasman, G. D. “Computed circular dichroism spectra for the evaluation of protein conformation.” *Biochemistry* **8**, 4108–4116 (1969).
2. Hartl, F. U. “Protein misfolding diseases.” *Annu. Rev. Biochem.* **86**, 21-26 (2017).
3. Zhang, C. F., Lewis, J. W., Cerpa, R., Kuntz, I. D. & Kliger, D. S. “Nanosecond circular dichroism spectral measurements: extension to the far-ultraviolet region.” *J. Phys. Chem.* **97**, 5499-5505 (1993).
4. Eom, I., Ahn S. H., Rhee, H. & Cho, M. “Single-shot electronic optical activity

- interferometry: power and phase fluctuation-free measurement.” *Phys. Rev. Lett.* **108**, 103901 (2012).
5. Hiramatsu, K. & Nagata, T. “Broadband and ultrasensitive femtosecond time-resolved circular dichroism spectroscopy.” *J. Chem. Phys.* **143**, 121102 (2015).
 6. Oppermann, M., Bauer, B., Rossi, T., Zinna, F., Helbing, J., Lacour, J., & Chergui, M. “Ultrafast broadband circular dichroism in the deep ultraviolet.” *Optica*, **6(1)**, 56. <http://doi.org/10.1364/OPTICA.6.000056> (2019).
 7. Dartigalongue, T. & Hache, F. “Observation of sub-100 ps conformational changes in photolyzed carbonmonoxy-myoglobin probed by time-resolved circular dichroism.” *Chem. Phys. Lett.* **415**, 313–316 (2005).
 8. Dartigalongue, T., Niezboral, C. & Hache, F. “Subpicosecond UV Spectroscopy of carbonmonoxy-myoglobin: absorption and circular dichroism studies.” *Phys. Chem. Chem. Phys.* **9**, 1-5 (2007).
 9. Mangot, L., Taupier, G., Romeo, M., Boeglin, A., Cregut, O., & Dorkenoo, K. “Broadband transient dichroism spectroscopy in chiral molecules.” *Optics Letters*, **35(3)**, 381–383. (2010).
 10. Trifonov, A., Buchvarov, I., Lohr, A., Würthner, F. & Fiebig, T. “Broadband femtosecond circular dichroism spectrometer with white-light polarization control.” *Rev. Sci. Instrum.* **81**, 043104 (2010).
 11. Preda, F., Perri, A., Réhault, J., Dutta, B., Helbing, J., Cerullo, G., & Polli, D. “Time-domain measurement of optical activity by an ultrastable common-path interferometer.” *Optics Letters*, **43(8)**, 1882. <http://doi.org/10.1364/OL.43.001882> (2018).
 12. Snyder, P. A. & Rowe, E. M. “The first use of synchrotron radiation for vacuum ultraviolet circular dichroism measurements.” *Nucl. Instrum. Methods* **172**, 345-349 (1980).

13. Sutherland, J. C., Desmond, E. J. & Takacs, P. Z. "Versatile spectrometer for experiments using synchrotron radiation at wavelengths greater than 100 nm." *Nucl. Instrum. Methods* **172**, 195-199 (1980).
14. Sutherland, J. C., Emrick A., France, L. L., Monteleone, D. C. & Trunk, J. "Circular dichroism user facility at the national synchrotron light source: estimation of protein secondary structure." *Biotechniques* **13**, 588-590 (1992).
15. Miles, A. J. & Wallace, B. A. "Synchrotron radiation circular dichroism spectroscopy of proteins and applications in structural and functional genomics." *Chem. Soc. Rev.* **35**, 39-51 (2006).
16. Bulheller, B.M., Miles, A.J., Wallace, B.A. & Hirst, J.D. "Charge-transfer transitions in the vacuum-ultraviolet of protein circular dichroism spectra." *J. Phys. Chem. B.* **112**, 1866-1874 (2008).
17. Matsuo K., Yonehara R. & Gekko K. "Secondary-structure analysis of proteins by vacuum-ultraviolet circular dichroism spectroscopy." *J. Biochem.* **135**, 405-411 (2004).
18. Wallace B. A. "Conformational changes by synchrotron radiation circular dichroism spectroscopy." *Nat. Struct. Biol.* **7**, 708-709 (2000).
19. Heinzmann, U., Osterheld, B. & Schafers, F. "Measurements and calculations of the circular polarization and of the absolute intensity of synchrotron radiation in the wavelength range from 40 to 100 nm." *Nucl. Instrum. Methods* **195**, 395-398 (1982).
20. Kang, B. S., Fischer, P., Kim, D. H., Attwood, D., Anderson, E. & Cho, G. "Bending magnet x-ray polarization modulation for magnetic full-field soft x-ray transmission microscopy." *IPAP conference series* **7**, 288-290. (2006).
21. Schiller, J. & Hormes, J. "Using the circularly polarized components of synchrotron radiation for circular dichroism measurements between 400 and 3500Å." *Nucl. Instr. Meth.* **A246**, 772-775 (1986).

22. Giuliani, A., Jamme, F., Rouam, V., Wien, F., Giorgetta, J. L., Lagarde, B., Chubar, O., Bac, S., Yao, I., Rey, S., Herbeaux, C., Marlats, J. L., Zerbib, D., Polack, F. & Refregiers, M. "DISCO: a low-energy multipurpose beamline at synchrotron SOLEIL." *J. Sync. Rad.* **16**, 835-841 (2009).
23. Refregiers, M., Wien, F., Ta, H. P., Premvardhan, L., Bac, S., Jamme, F., Rouam, V., Lagarde, B., Polack, F., Giorgetta, J. L., Ricaud, J. P., Bordessoule, M. & Giuliani, A. "Synchrotron radiation circular dichroism endstation at SOLEIL." *J. Sync. Rad.* **19**, 831-835 (2012).
24. Miles, A. J., Wien, F. & Wallace, B. A. "Redetermination of the extinction coefficient of camphor-10-sulfonic acid, a calibration standard for circular dichroism spectroscopy." *Anal. Biochem.* **335**, 338–339 (2004).
25. Flint D. G., Kumita J. R., Smart O. S. & Woolley G. A. "Using an azobenzene cross-linker to either increase or decrease peptide helix content upon trans-to-cis photoisomerization." *Chem Biol* **9**, 391–397 (2002).
26. Kumita J. R., Smart O. S. & Woolley G. A. "Photo-control of helix content in a short peptide." *Proc. Natl. Acad. Sci. USA* **97**, 3803-3808 (2000).
27. Nenov, A., Borrego-Varillas, R., Oriana, A., Ganzer, L., Segatta, F., Conti, I., Segarra-Marti, J., Omachi, J., Dapor, M., Taioli, S., Manzoni, C., Mukamel, S., Cerullo G. & Garavelli M. "UV-Light-Induced Vibrational Coherences: The key to understand Kasha rule violation in trans-azobenzene." *J. Phys. Chem. Lett.* **9**, 1534–1541 (2018).
28. Nägele, T., Hoche, R., Zinth, W. & Wachtveitl, J. "Femtosecond photoisomerization of cis-azobenzene." *Chem. Phys. Lett.* **272**, 489-495 (1997).
29. Fujino, T. & Tahara, T. "Picosecond time-resolved Raman study of *trans*-azobenzene." *J. Phys. Chem. A* **104**, 4203-4210 (2000).
30. Fujino, T., Arzhantsev, S. Y. & Tahara, T. "Femtosecond time-resolved fluorescence

- study of photoisomerization of trans-azobenzene.” *J. Phys. Chem. A* **105**, 8123-8129 (2001).
31. Xia, S. H., Cui, G., Fang, W. H. & Thiel, W. “How photoisomerization drives peptide folding and unfolding: insights from QM/MM and MM dynamics simulations.” *Angew. Chem. Int. Ed.* **55**, 2067-2072 (2016).
32. Woolley, G. A. “Photocontrolling peptide α helices.” *Acc. Chem. Res.* **38**, 486-493 (2005).
33. Chen, E., Kumita, J. R., Woolley, G. A. & Kliger, D. S. “The kinetics of helix unfolding of an azobenzene cross-linked peptide probed by nanosecond time-resolved optical rotatory dispersion.” *J. Am. Chem. Soc.* **125**, 12443-12449 (2003).
34. Changenet-Barret, P., Hua, Y., Gustavsson, T., & Markovitsi, D. “Electronic Excitations in G-quadruplexes Formed by the Human Telomeric Sequence: A Time-Resolved Fluorescence Study.” *Photochem. Photobiol.*, **91(3)**, 759–765. <http://doi.org/10.1111/php.12379> (2014).