



A Novel Microfluidic Formaldehyde Microanalyser for Continuous Real-Time Monitoring in Indoor Air: Analytical Development and Validation

Stéphane Le Calvé, Claire Trocquet, Pierre Bernhardt

► To cite this version:

Stéphane Le Calvé, Claire Trocquet, Pierre Bernhardt. A Novel Microfluidic Formaldehyde Microanalyser for Continuous Real-Time Monitoring in Indoor Air: Analytical Development and Validation. Proceedings 2017, 1, 740; † Presented at the 5th International Symposium on Sensor Science, 2017, <10.3390/proceedings1080740>. <hal-02414952>

HAL Id: hal-02414952

<https://hal.science/hal-02414952v1>

Submitted on 20 Dec 2019

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

Abstract

A Novel Microfluidic Formaldehyde Microanalyser for Continuous Real-Time Monitoring in Indoor Air: Analytical Development and Validation [†]

Stéphane Le Calvé ^{1,*}, Claire Trocquet ² and Pierre Bernhardt ²

¹ Institute of Chemistry and Processes for Energy, Environment and Health (ICPEES), University of Strasbourg and CNRS (UMR 7515), 25 rue Becquerel, 67087 Strasbourg, France

² In'Air Solutions, 1 rue Blessig, 67084 Strasbourg, France; ctrocquet@inairsolutions.fr (C.T.); pbern@inairsolutions.fr (P.B.)

* Correspondence: slecalve@unistra.fr

[†] Presented at the 5th International Symposium on Sensor Science (ISS 2017), Barcelona, Spain, 27–29 September 2017.

Published: 4 December 2017

Formaldehyde is a major and harmful pollutant of indoor air due to its multiple sources and its carcinogenic effect. This work reports the development of a novel analytical method based on microfluidic technologies for the detection of low airborne Formaldehyde concentrations, representative of those found in indoor air, i.e., 10–100 $\mu\text{g m}^{-3}$. The new analytical technique operates as follows: (1) gas sampling; (2) gaseous Formaldehyde uptake into the aqueous solution using an annular gas/liquid flow at room temperature; (3) derivatization reaction with acetylacetone solution at 65 °C producing 3,5-Diacetyl-1,4-dihydrolutidine (DDL) and (4) fluorimetric DDL detection.

Laboratory experiments were performed to determine the experimental conditions permitting to obtain a stable annular flow, i.e., gas to liquid flow rate ratios greater than 1000. From liquid and gas calibrations, an uptake yield of 100% and a detection limit of 1 $\mu\text{g m}^{-3}$ were determined. Finally, our portable instrument is fully controlled by homemade software and has a response time of 10 min, a temporal resolution of 2 s and an autonomy of 100 h with 100 mL reagent. Finally, this formaldehyde microanalyser was then deployed during several field campaigns and compared with the ISO 16000-3 reference method, i.e., the active sampling on DNPH cartridges.



© 2017 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<http://creativecommons.org/licenses/by/4.0/>).