



HAL
open science

On-chip parallel Fourier transform spectrometer for broadband selective infrared spectral sensing

Alaa Fathy, Yasser M Sabry, Sebastien Nazeer, Tarik Bourouina, Diao Khalil

► **To cite this version:**

Alaa Fathy, Yasser M Sabry, Sebastien Nazeer, Tarik Bourouina, Diao Khalil. On-chip parallel Fourier transform spectrometer for broadband selective infrared spectral sensing. *Microsystems & Nanoengineering*, 2020, 6 (1), 10.1038/s41378-019-0111-0 . hal-02478146

HAL Id: hal-02478146

<https://hal.science/hal-02478146v1>

Submitted on 18 Feb 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

ARTICLE

Open Access

On-chip parallel Fourier transform spectrometer for broadband selective infrared spectral sensing

Alaa Fathy^{1,2,3}, Yasser M. Sabry^{1,2}, Sébastien Nazeer², Tarik Bourouina^{2,3} and Daa A. Khalil^{1,2}

Abstract

Optical spectrometers enable contactless chemical analysis. However, decreasing both their size and cost appears to be a prerequisite to their widespread deployment. Chip-scale implementation of optical spectrometers still requires tackling two main challenges. First, operation over a broad spectral range extending to the infrared is required to enable covering the molecular absorption spectrum of a broad variety of materials. This is addressed in our work with an Micro-Electro Mechanical Systems (MEMS)-based Fourier transform infrared spectrometer with an embedded movable micro-mirror on a silicon chip. Second, fine spectral resolution $\Delta\lambda$ is also required to facilitate screening over several chemicals. A fundamental limit states that $\Delta\lambda$ is inversely proportional to the mirror motion range, which cannot exceed the chip size. To boost the spectral resolution beyond this limit, we propose the concept of parallel (or multi-core) FTIR, where multiple interferometers provide complementary optical paths using the same actuator and within the same chip. The concept scalability is validated with 4 interferometers, leading to approximately 3 times better spectral resolution. After the atmospheric contents of a greenhouse gas are monitored, the methane absorption bands are successfully measured and discriminated using the presented device.

Miniaturization of sensors based on MEMS technologies is now a proven option for performing measurements in low-cost and wide-scale deployments. Their integration in very large volume markets such as smartphones, cars and many other consumer products opens up new prospects for the internet-of-things (IoT). However, this success is still limited to the measurement of physical parameters only. Chemical parameters are so numerous that one can hardly consider having as many chemical sensors as chemical substances, which are quite diverse. To address this difficulty, a paradigm shift regarding implementing concepts of analytical chemistry on a silicon chip has been introduced. Although micro-gas chromatography¹ is among the most interesting options that have been thoroughly investigated, optical spectroscopy seems to be the only way to carry out contactless,

remote chemical analysis. Therefore, microscale optical spectroscopy is attracting increasing interest^{2–11}. Many of the reported solutions are limited to the visible spectral range^{2–4}, while those addressing the near-infrared range have a very limited spectral range^{5,7,8}. In addition, very few attempts extend to the mid-infrared^{9,10}, and those that do still have a very limited bandwidth. In fact, MEMS spectrometers can be based on different configurations¹² such as diffraction gratings¹³, digital micro-mirror devices (DMDs)¹¹, Fourier transform interferometers¹⁴, linear variable filters and tuneable Fabry-Pérot filters¹⁵.

Fourier transform infrared (FTIR) spectrometers appear to be a promising option since their broad spectral range enables covering the absorption molecular spectrum of a broad variety of materials, including gases, liquids and solids. In addition, a high signal-to-noise ratio (SNR) can be achieved due to the detection of all wavelengths simultaneously with a single detector¹⁶. In its benchtop form, it is currently used in medical analysis^{17–21}, food quality control^{22–25} and soil analysis²⁶, to name a few application areas. A vast majority of these applications

Correspondence: Tarik Bourouina (tarik.bourouina@esiee.fr)

¹Faculty of Engineering, Ain-Shams University, 1 Elsarayat St. Abbassia, Cairo, Egypt

²Si-Ware Systems, 3 Khalid Ibn Al-Waleed St., Heliopolis, Cairo, Egypt
Full list of author information is available at the end of the article.

© The Author(s) 2020



Open Access This article is licensed under a Creative Commons Attribution 4.0 International License, which permits use, sharing, adaptation, distribution and reproduction in any medium or format, as long as you give appropriate credit to the original author(s) and the source, provide a link to the Creative Commons license, and indicate if changes were made. The images or other third party material in this article are included in the article's Creative Commons license, unless indicated otherwise in a credit line to the material. If material is not included in the article's Creative Commons license and your intended use is not permitted by statutory regulation or exceeds the permitted use, you will need to obtain permission directly from the copyright holder. To view a copy of this license, visit <http://creativecommons.org/licenses/by/4.0/>.

require both high SNR and fine spectral resolution. For instance, in gas sensing, a resolution between 0.5 and 2 cm^{-1} is needed²⁷. In medical applications, insulin can be analysed in the mid-IR with a resolution of 4 cm^{-1} ²⁸, and the glucose concentration in blood has been determined²⁷. Blood cells can be identified using FTIR with a resolution of $\sim 2\text{ cm}^{-1}$ ²⁹. Food quality control requires a resolution from 4 to 8 cm^{-1} ³⁰, and liquids are commonly measured with a resolution ranging from 4 to 8 cm^{-1} ¹⁹.

Considering the miniaturization format, MEMS-based FTIR spectrometers using different architectures have already been implemented, such as the Michelson interferometer¹⁴, lamellar grating³¹, Mach-Zehnder interferometer^{8,32}, moving wedge interferometer³³, low finesse Fabry-Pérot interferometer³⁴ and cascaded Fabry-Pérot interferometers³⁵. However, all the reported solutions have a limited spectral resolution. On the one hand, the monolithic integration of all the components of the Michelson interferometer led to a resolution of $\sim 33\text{ cm}^{-1}$ ¹⁴. On the other hand, better resolution could be achieved, but at the loss of the advantage of monolithic integration, leading to challenging alignment and high assembly costs; for instance, a resolution of $\sim 25\text{ cm}^{-1}$ in the spectral range from 2 to $5\text{ }\mu\text{m}$ ³⁶ has been obtained with a 3-mm-diameter MEMS mirror that requires further assembly with other discrete components to form the interferometer. A moving mirror with a relatively long physical travel range of $\sim 600\text{ }\mu\text{m}$ was assembled perpendicular to the substrate with other Michelson components and led to a resolution of $\sim 8\text{ cm}^{-1}$ ³⁷. In the case of the MEMS-lamellar grating spectrometer, a resolution of $\sim 15\text{ cm}^{-1}$ within the mid-IR range has been achieved³¹, and gas measurement has been demonstrated with a resolution of $\sim 20\text{ cm}^{-1}$ ³⁸. In all FTIR spectrometers, the spectral resolution $\Delta\lambda$ is inversely proportional to the mirror travel range Δx , i.e., $\Delta\lambda \sim 1/\Delta x$, such that achieving fine spectral resolution requires a large travel range. Therefore, the limitation of the on-chip resolution arises from the restricted travel range of the micro-mirror, limited by the stability and reliability requirements of the MEMS actuators³⁹.

In this work, the parallel (or multi-core) spectrometer concept is introduced, and a silicon chip monolithically integrating multiple interferometers sharing a single MEMS actuator is implemented. This concept is intended to overcome the limitation in the spectral resolution, where the interferometers have complementary optical path differences (OPDs). The moving mirrors of the interferometers are mechanically coupled to the same actuator, while each of the N interferometers in the chip scans a different range of the OPD in the measured interferogram due to the complementary shifts in the mirror position implemented in the design. The interferogram of each interferometer is measured by a


corresponding detector, and then signal processing is applied to produce an overall interferogram as a concatenation of the different recorded interferograms. Therefore, one can increase the OPD N times and thus enhance the wavelength resolution accordingly using the same MEMS actuator.

Results

Concept and modelling

An architecture consisting of four interferometers is depicted in Fig. 1. The input light is injected into the different interferometers by means of an optical fibre bundle, where the interferometers are implemented using the micro-optical bench (MOP) technology⁴⁰. Usually, broadband thermal light sources are larger in size than the input aperture of a single MOP interferometer. For instance, the filament length is a few millimetres with a large divergence angle, while the MEMS device dimensions are hundreds of microns with a limited acceptance angle (low optical throughput). Therefore, one can easily spatially split the source power over the N interferometers without reducing the amount of light received by each of those interferometers. This interestingly means that the SNR per interferometer is not affected by the splitting of the light.

Each interferometer has its own beam splitter, fixed mirror, moving mirror and detector. The four moving mirrors of the four interferometers are coupled to the same actuator, while the mirror positions are shifted with respect to each other by an increment Δ_M such that the resolution can be enhanced up to four times. For instance, if the actuator travel range is $200\text{ }\mu\text{m}$ and the increment Δ_M is $200\text{ }\mu\text{m}$, a full OPD of $\sim 1600\text{ }\mu\text{m}$ is obtained using



the four interferometers. The first interferometer scans the OPD from -800 to -400 μm , the second one scans from -400 to 0 μm , the third one scans from 0 to 400 μm and the last one scans from 400 to 800 μm . The overall interferogram is the concatenation of the four complementary interferograms, and the corresponding spectral resolution is expected to be quadrupled compared to that of a single interferometer.

Concatenation algorithm

Due to the microfabrication non-idealities, such as variations in the over-etching in the silicon across the MEMS chip, the interferometers are not fully matched, and shifts in the OPD (along the horizontal axis of the interferograms) can be encountered with respect to the design. Additionally, the detectors used are not fully matched, leading to mismatch in the power value (along the vertical axis of the interferograms). To overcome these effects, the overlap regions between the OPDs of the different interferograms are introduced for the purpose of calibrating the concatenation positions. The moving mirror of each interferometer is designed to scan the OPD range of $2L + 2\Delta L$, where $2L$ is the actual optical full travel range and $2\Delta L$ represents the overlap region length with respect to the interferogram of the next interferometer. The correction in the OPD depends on introducing an offset Δx in the OPD of one of the two interferograms that gives the best match between the two overlapped parts of the two consecutive interferometers. The matching is calculated using Pearson’s coefficient PC, given by:

$$PC(\Delta x) = \frac{\sum_i [(I_n(x_i + \Delta x) - \overline{I_n(x + \Delta x)}) (I_{n+1}(x_i) - \overline{I_{n+1}(x)})]}{\sqrt{\sum_i (I_n(x_i + \Delta x) - \overline{I_n(x + \Delta x)})^2} \sqrt{\sum_i (I_{n+1}(x_i) - \overline{I_{n+1}(x)})^2}} \tag{1}$$

where I_n and I_{n+1} represent the intensity of the overlapped parts of interferograms of the two consecutive interferometers n and $n + 1$, respectively, x_i is the OPD defined in the overlapped part of the interferograms, and the upper bar represents the mean. The optimum shift Δx corresponds to a maximum value for the PC.

To find the gain correction required for the interferogram power, one of the interferograms is multiplied by a variable gain, which makes the overlapped part of the interferograms nearly equal. The normalized root mean square error NRMSE is used as a merit function, where it is calculated over the overlapped region as

$$NRMSE = \frac{\sqrt{(I_1(x_i) - I_2(x_i))^2}}{\max(I_1(x_i))} \tag{2}$$

where max represents the maximum operator. The best gain corresponds to the minimum NRMSE.

Noise analysis

An analytical formula for the SNR can be derived in the case of additive sources of noise, such as thermal, shot and quantization noises. A noisy interferogram can be expressed by:

$$I(x) = I_o(x)rect\left(\frac{x}{2L}\right) + n_o(x)rect\left(\frac{x}{2L}\right) \tag{3}$$

where $I_o(x)$ and $n_o(x)$ are the desired interferogram signal and noise contribution, respectively. The MEMS average velocity is given by $V_e = L/T_{scan}$, where T_{scan} is the scan time. Thus, the noise power can be expressed by $\sigma^2 = N_o B_f = 2N_o V_e B_v$, where N_o is the noise power spectral density in V^2/Hz , B_f is the electronic bandwidth of the measured spectrum in Hz and B_v is the corresponding wavenumber bandwidth in m^{-1} . Then, the spectrum in the wavenumber domain is given by:

$$S(f) = [B(\nu) + N(\nu)] * 2L sinc(2L\nu) \tag{4}$$

where $B(\nu)$ is the input (desired) spectrum, $N(\nu)$ represents the noise spectrum and $*$ represents the convolution. Then, the average noise energy density is given by¹⁴:

$$E(N(\nu)^2) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} rect\left(\frac{x}{2L}\right) rect\left(\frac{s}{2L}\right) \sigma^2 sinc(2B_v(x-s)) e^{-j2\pi(x-s)\nu} dx ds \tag{5}$$

Using the Fourier transform identities, the above expression can be rewritten as (using $1/2L \ll B_v$)⁴¹:

$$E(N(\nu)^2) = \frac{L\sigma^2}{B_v} rect\left(\frac{\nu}{2B_v}\right) \tag{6}$$

where E represents the expectation. Then, the SNR within a window $\Delta\nu$ around a central wavenumber $\bar{\nu}$ is given by:

$$SNR(\bar{\nu}) = \sqrt{N_{scans} \frac{\int_{\bar{\nu}-\frac{\Delta\nu}{2}}^{\bar{\nu}+\frac{\Delta\nu}{2}} B(\nu) * 2L sinc(2L\nu) d\nu}{\int_{\bar{\nu}-\frac{\Delta\nu}{2}}^{\bar{\nu}+\frac{\Delta\nu}{2}} \sqrt{E(N(\nu)^2)} d\nu}} \tag{7}$$

where N_{scans} is the number of scans. Assuming that the spectrum of the input light is of uniform B_o value, and as $\Delta\nu \rightarrow 0$, the SNR is given by:

$$SNR(\bar{\nu}) = \frac{B_o \sqrt{N_{scans}}}{\sqrt{2N_o L V_e}} = \frac{B_o \sqrt{T_{scan} N_{scans}}}{\sqrt{2N_o L}} \tag{8}$$

In a single-core spectrometer, increasing the resolution N times requires an increase in both the travel range and the actuator velocity by a factor of N while maintaining the same scan time of a single measurement T_{scan} . Thus, the SNR will degrade by a factor of $1/N$. Alternatively, one

can increase the resolution using the same velocity while increasing the time scan of a single measurement T_{scan} by a factor of N . However, this means that the number of scans N_{scans} will be decreased by a factor of $1/N$ for the same total run time ($N_{\text{scan}}T_{\text{scan}}$). Thus, the overall SNR will degrade by a factor of $1/N$ as well for the same total run-time. For the introduced concept of a multi-core spectrometer acting simultaneously with multiple detectors, the velocity of the moving mirrors of the interferometers is not increased, and the full travel range FTR is scanned in the same amount of time as required by the single interferometer. Thus, the SNR will be degraded only by a factor of $1/\sqrt{N}$ while maintaining the same total run time. We call this improvement in the SNR the “parallel spectrometer advantage”.

Experimental device

The main target of the proposed design is to enhance the spectral resolution using four interferometers monolithically integrated on the same chip. This outcome was obtained by replicating the core interferometer, previously reported in ref. ¹⁴, designed on-chip but with shifts in the remaining positions of the moving mirrors. A three-dimensional schematic of the chip with the four interferometers is depicted in Fig. 2a. Since each of the four interferometers has a full travel range $2L$ of $360\ \mu\text{m}$, the total range is $1440\ \mu\text{m}$. Theoretically, this improves the resolution from $33\ \text{cm}^{-1}$ to $8.3\ \text{cm}^{-1}$, according to the full-width at half-maximum (FWHM) criterion¹³. The interferogram of the first interferometer has an OPD range from -860 to $-500\ \mu\text{m}$, the second has an OPD range from -500 to $-140\ \mu\text{m}$, the third has an OPD range from -140 to $220\ \mu\text{m}$, and the fourth has an OPD range from 220 to $580\ \mu\text{m}$. Thus, the centres of their OPDs corresponding to the rest positions of the moving mirrors are $-680\ \mu\text{m}$, $-320\ \mu\text{m}$, $40\ \mu\text{m}$ and $400\ \mu\text{m}$, respectively. Additional overlap regions of the range $2\Delta L$ equal to $40\ \mu\text{m}$ were introduced to apply the concatenation algorithm described previously.

Scanning electron microscope (SEM) images of the fabricated comb and mirror arms are shown in Fig. 2b, c, respectively. In Fig. 2d, a SEM image of the interferometer⁴² is shown, comprising a beam splitter, a fixed mirror and a moving mirror. This design is replicated for the parallel interferometers. Regarding the quality of surfaces, the root mean square (RMS) value of the roughness is $<20\ \text{nm}$ using this technology⁴³, and the measured verticality of the surfaces is better than 0.05° . The mirror reflectivity is greater than 90%. A photo of the fabricated chip assembled on a PCB daughter board is shown in Fig. 2e. The light from the optical source was coupled into the interferometers using a bundle of multimode fibres (MMFs) connected directly to the light source. On the other hand, the optical fibres were cleaved





Fig. 2 Illustration of the fabricated design of the parallel interferometers integrated on the same chip. **a** The chip-scale architecture of the MEMS spectrometer, comprising four interferometers and a comb-drive actuator. The interferometers are denoted by Int. and numbered from 1 to 4. The moving mirrors of all interferometers are attached to the same moving arm. The light path is indicated by the red arrows. **b** A SEM image of the moving mirror arm. **c** A SEM image of an optical interferometer. **d** A SEM image of an optical interferometer. **e** A camera photo of the corresponding fabricated die with the coupling fibres, where the yellow arrows represent the input directions of the light to the interferometers and the red arrows represent the output direction of the light

and inserted into the MEMS chip in micromachined grooves self-aligned with the interferometers. The MEMS actuator is operated at resonance using an actuation voltage with a sinusoidal waveform. It has a maximum displacement of $\sim 250\ \mu\text{m}$. The position of the MEMS actuator was determined by means of a capacitive sensing circuit. The relation between the mirror position and the capacitance was calibrated using a reference light at $1550\ \text{nm}$ ⁴⁴.

The output optical light beams from the interferometers are different, they cannot be summed on the same




detector, and the output of each one should be coupled to its corresponding detector. The output electrical signal from the corresponding detector is amplified by electronic amplifiers. These are followed by an analog-to-digital converter (ADC) to sample the analog signal and convert it to the digital domain, where digital filtering is applied to obtain the corresponding interferogram. After time averaging of each of the four interferograms, the concatenation algorithm is applied, and the final spectrum is obtained. The scan time T_{scan} is ~ 1.5 ms, while N_{scans} is 1000 scans or more.

White light characterization

A white light source was measured by the four interferometers, where each one had its own interferogram versus the respective OPD. The overall interferogram after concatenation is shown in Fig. 3a, where the interferogram comprises four parts of different colours corresponding to the four different interferograms. Applying the Fourier transform to the concatenated interferograms gives the corresponding spectrum of higher resolution, as depicted in Fig. 3b. The intersecting part of the interferograms of interferometers 1 and 2 is shown in Fig. 3c, the intersecting part of the interferograms of interferometers 2 and 3 is shown in Fig. 3d, and the intersecting part of the interferograms of interferometers 3 and 4 is shown in Fig. 3e. The intersection of the interferograms of interferometers 2 and 3 and that of interferometers 3 and 4 are smoother than those of interferometers 1 and 2 because the signal power in these parts is higher and, thus, the noise effect in the transitions is smaller.

Characterization with narrow band filters

To test the resolution enhancement achieved by the parallel interferometers, an optical filter with an FWHM of 2 nm centered at 1.55 μm was placed after the white light lamp to obtain a narrow band spectrum. Additionally, the same filter was measured using a benchtop spectrometer (Bruker Tensor II). The spectrum obtained from a single interferometer, which contains the main burst, is shown in Fig. 4a(i) (depicted by the blue bold line) and compared with a spectrum measured with the benchtop spectrometer for the same corresponding OPD range (depicted by the red dash-dot line). The spectra of the concatenated interferograms of two consecutive interferometers, three consecutive interferometers and four consecutive interferometers are shown in Fig. 4a(i), (ii) and (iv), respectively. The figures show the gradual improvement of the resolution of the spectrum with increasing number of concatenated interferograms. The measured FWHM was improved from 8.4 nm to 2.7 nm (from 35 cm^{-1} to 11 cm^{-1}). Moreover, another filter centered at 2 μm with an FWHM of 1 nm was also measured using the MEMS interferometers and the benchtop spectrometer. The obtained spectra are depicted in Fig. 4b. The measured FWHM was 4.9 nm, obtained using the four interferometers. The corresponding value obtained using a single interferometer was 14.8 nm. At both wavelengths, the improvement in the resolution is ~ 3.1 -fold, which is smaller than the ideal expected fourfold improvement. This result is due to the self-apodization of the interferogram, which affects the achieved resolution¹⁶.



Krypton source spectral analysis

A krypton source is characterized by having distributed emission lines in the infrared region. The emission spectrum of this source was measured using a single interferometer and parallel interferometers. The measured spectra around a wavelength of 1.69 μm are plotted in Fig. 5. For reference, the krypton emission lines taken from the literature⁴⁵ are also shown (dashed black lines). The spectrum obtained by the single interferometer (red curve) does not discriminate the emission lines at 1.678 μm and 1.688 μm due to the limited spectral resolution. Using the parallel interferometers (blue curve), these emission lines are discriminated successfully.

Methane gas spectral analysis

A gas measurement was also conducted to show the ability of the parallel interferometers to discriminate gas absorption bands. A gas cell that contains 10% methane (CH_4) was inserted between the white light source and the spectrometer. The gas cell length is 10 cm. The content of the gas cell was analysed using parallel interferometers and a single interferometer. The obtained absorbance curves are shown in Fig. 6. The absorption bands around 2.32 μm and 2.37 μm are apparently discriminated (blue curve) compared to the spectrum given by the single interferometer (red curve). The absorbance value was also enhanced dramatically, as shown by the wavelength range of 2.3–2.37 μm . For reference, the same gas cell was also measured using the benchtop spectrometer for two different resolutions, namely, 34 cm^{-1} and 8.5 cm^{-1} , and plotted in Fig. 6. These resolutions are the theoretical



resolutions of the single MEMS interferometer and the MEMS parallel interferometers. The comparison shows good agreement with the measurements of the MEMS spectrometer.

Signal-to-noise ratio

To test the SNR of the proposed spectrometer architecture and compare the results with the derived theoretical predictions, the SNR was measured using the 100%-line method¹⁶ for the single interferometer and the parallel interferometers. The method is based on measuring successive spectra and dividing every two consecutive spectra to obtain a group of 100% lines. The measurement setup was in transmission mode. The measurement time was 2 s. The 100% lines of the single interferometer and those of the concatenated interferograms are shown in Fig. 7a, b, respectively. The noise is increased due to the increased OPD. The SNR versus wavelength in the case of the parallel interferometers and the single interferometer (Int. 3) is shown in Fig. 7c. The ratio between the SNR of the parallel interferometers and that of the single interferometer ranges from 0.35 to 0.6 across the wavelength range, with a mean value of ~ 0.5 , as expected from Eq. (8). The peak value of the SNR of the MEMS parallel interferometer MEMS is $\sim 760:1$ at 2.058 μm . This result corresponds to an RMS value of 0.13%, which is lower than the standard required value of 0.2% stability over 2 h.

Discussion

The proposed parallel spectrometer architecture is eventually proven to allow going beyond the fundamental



limit of the conventional FTIR spectrometer, the wavelength resolution $\Delta\lambda$ of which is known to be inversely proportional to the travel range Δx of the movable mirror. This is a significant limitation when considering the chip-scale spectrometers, since Δx is limited by the chip size, which is a few millimetres.

The practical implementation of the concept of parallel FTIR spectrometers was achieved in this work based on standard MEMS technologies on a centimetre-scale chip. The proposed architecture schematically depicted in Figs. 1 and 2 is scalable. It has the advantage of sharing the same actuator for simultaneously acting on the shifting motions of all movable mirrors of the N Michelson interferometers, the individual footprint of which did not exceed a few mm^2 . To validate the predicted performance scalability rules, a MEMS die comprising four parallel interferometers was fabricated and tested. The achieved spectral resolution was 11 cm^{-1} , compared to 35 cm^{-1} for a single interferometer. The corresponding improvement


of $\Delta\lambda$ is 3.1-fold. The proposed structure also did not sacrifice the input power because the thermal infrared light source area is much larger than that of each single MEMS interferometer. Here is the origin of the idea to use multiple interferometers on the same chip to make use of the large source area while simultaneously achieving higher resolution. In this scheme, the SNR was degraded by only 0.5 times while maintaining the same total run time, in agreement with our theoretical modelling, which predicted a scaling as $\text{SNR} \sim 1/\sqrt{N}$. Furthermore, based on its improved resolution, it was shown that the proposed device is capable of recognizing the spectral signature of methane through several of its characteristic lines. The proposed structure used eight optical fibres for light coupling in and out of the interferometers, which adds some complexity to the system. In principle, the fibres can be replaced by waveguides monolithically integrated on the same chip and a micro-optical coupling component^{46,47} to achieve the ultimate goal of compactness and low cost.

Based on these first results of spectral sensing, as well as the experimental validation of the scalability rules established in theoretical modelling, this work opens the way for selective and contactless chemical analysis in-the-field, based on the chip-scale FTIR-MEMS optical spectrometers, especially when one considers the possible wavelength range that can be extended further to the mid-infrared and even far-infrared. Typical target applications include environmental monitoring, including air quality and soil analysis, precision agriculture and medical diagnosis.

Methods

Fabrication

The chip was fabricated using MEMS technology. The chip size was $9.1 \times 12\text{ mm}^2$. A silicon-on-insulator (SOI) wafer with a device layer of $400\text{ }\mu\text{m}$ and oxide thickness of $3\text{ }\mu\text{m}$ was used. The deep reactive ion etching (DRIE) process is used to etch the deep trenches to achieve a high-throughput interferometer. Smooth surfaces were



achieved using post-oxidation. Metallization was further applied for mirrors and pads. More details on the process steps can be found in the Supplementary Data.

Optical measurements

For coupling light into/out of the interferometers, four pairs of MMFs (Thorlabs) were used. The core/cladding diameter was equal to 400/440 μm , and the numerical aperture (NA) was equal to 0.39. Each interferometer has its perspective input and output grooves for placing the fibre inside. Different light sources were used for light injection into the interferometers. For a wide spectrum, a halogen lamp (Avantes Avalight) was used. The source temperature was 3000 K. Multilayer Bragg filters (Omega optics; one with an FWHM of 2 nm at 1.55 μm and the other with an FWHM of 1 nm at 2 μm) were used to narrow the wide light coming from the halogen light source for resolution measurements. The filter FWHM is 2 nm. A krypton source (Ocean optics Kr-1) was used as the multi-wavelength emission source. For the output interferogram detection, extended InGaAs detectors (1.1–2.6 μm) (Hamamatsu G12183) were used for detection. To measure gases using the proposed spectrometer architecture, a standard gas cell (Wavelength references, Inc.) containing 10% methane, 1% acetylene, 0.25% carbon dioxide and 4% carbon monoxide diluted in nitrogen was used. The total pressure inside the cell was 740 Torr. The cell was placed between the light source and the parallel interferometer chip. The gas length was 10 cm. Methane was the most dominant gas absorbed in the spectral range of the spectrometer. Collimation lenses were added at the ends of the gas cell. A benchtop spectrometer (Bruker Tensor II) was used to validate the results of the proposed MEMS spectrometer.

Authors' contributions

A.F. developed the model, participated in designing the chip and conducted the measurements. Y.M.S. participated in designing the chip, optimized the fabrication process, prepared the gas measurement setup and contributed to

the writing and revision of the article. S.N. fabricated the chip. T.B. and D.K. are among the pioneers of the MEMS-FTIR spectrometers. D.K. proposed the idea of parallel architecture, reviewed the design and practical setup, and supervised the work. All authors contributed to the discussion of the results and preparation of the manuscript. T.B. coordinated the writing and made the overall revisions to the article.

Author details

¹Faculty of Engineering, Ain-Shams University, 1 Elsarayat St. Abbassia, Cairo, Egypt. ²Si-Ware Systems, 3 Khalid Ibn Al-Waleed St., Heliopolis, Cairo, Egypt. ³Université Paris-Est, ESYCOM EA 2552, ESIEE Paris, 93162 Noisy-le-Grand, France

Data availability

All published data supporting the findings of this study are available from the corresponding author upon reasonable request.

Conflict of interest

The authors declare that they have no conflict of interest.

Supplementary information accompanies this paper at <https://doi.org/10.1038/s41378-019-0111-0>.

Received: 15 June 2019 Revised: 25 September 2019 Accepted: 1 October 2019

Published online: 10 February 2020

References

1. Qin, Y. & Gianchandani, Y. B. A fully electronic microfabricated gas chromatograph with complementary capacitive detectors for indoor pollutants. *Microsyst. Nanoeng.* **2**, 15049 (2016).
2. Wang, Z. et al. Single-shot on-chip spectral sensors based on photonic crystal slabs. *Nat. Commun.* **10**, 1020 (2019).
3. Faraji-Dana, M. et al. Compact folded metasurface spectrometer. *Nat. Commun.* **9**, 4196 (2018).
4. Calafiore, G. et al. Holographic planar lightwave circuit for on-chip spectroscopy. *Light Sci. Appl.* **3**, e203 (2014).
5. Redding, B., Liew, S. F., Sarma, R. & Cao, H. Compact spectrometer based on a disordered photonic chip. *Nat. Photonics* **7**, 746 (2013).
6. Kita, D. M. et al. High-performance and scalable on-chip digital Fourier transform spectroscopy. *Nat. Commun.* **9**, 4405 (2018).
7. Le Coarer, E. et al. Wavelength-scale stationary-wave integrated Fourier-transform spectrometry. *Nat. Photonics* **1**, 473 (2007).
8. Souza, M. C. M. M., Grieco, A., Frateschi, N. C. & Fainman, Y. Fourier transform spectrometer on silicon with thermo-optic non-linearity and dispersion correction. *Nat. Commun.* **9**, 665 (2018).

9. Muneeb, M. et al. Demonstration of silicon-on-insulator mid-infrared spectrometers operating at 3.8 μm . *Opt. Express* **21**, 11659–11669 (2013).
10. Nedeljkovic, M. et al. Mid-infrared silicon-on-insulator Fourier-transform spectrometer chip. *IEEE Photonics Technol. Lett.* **28**, 528–531 (2016).
11. Wagner, E. P., Smith, B. W., Madden, S., Winefordner, J. D. & Mignardi, M. Construction and evaluation of a visible spectrometer using digital micromirror spatial light modulation. *Appl. Spectrosc.* **49**, 1715–1719 (1995).
12. Crocombe, R. A. Portable spectroscopy. *Appl. Spectrosc.* **72**, 1701–1751 (2018).
13. Pügner, T., Knobbe, J. & Grüger, H. Near-infrared grating spectrometer for mobile phone applications. *Appl. Spectrosc.* **70**, 734–745 (2016).
14. Erfan, M. et al. On-chip micro-electro-mechanical system Fourier transform infrared (MEMS FT-IR) spectrometer-based gas sensing. *Appl. Spectrosc.* <https://doi.org/10.1177/0003702816638295> (2016).
15. Sabry, Y. M. et al. In-plane deeply-etched optical MEMS notch filter with high-speed tunability. *J. Opt.* **17**, 125703 (2015).
16. Griffiths, P. R. & De Haseth, J. A. *Fourier Transform Infrared Spectrometry: Second Edition*. (2006). <https://doi.org/10.1002/047101631X>.
17. Movasaghi, Z., Rehman, S. & Ur Rehman, D. I. Fourier transform infrared (FTIR) spectroscopy of biological tissues. *Appl. Spectrosc. Rev.* **43**, 134–179 (2008).
18. Pijanka, J. et al. Synchrotron-based FTIR spectra of stained single cells. Towards a clinical application in pathology. *Lab. Investig.* **90**, 797 (2010).
19. Yang, H., Yang, S., Kong, J., Dong, A. & Yu, S. Obtaining information about protein secondary structures in aqueous solution using Fourier transform IR spectroscopy. *Nat. Protoc.* **10**, 382 (2015).
20. Byler, D. M. & Susi, H. Examination of the secondary structure of proteins by deconvolved FTIR spectra. *Biopolymers* **25**, 469–487 (1986).
21. Martin, F. L. et al. Distinguishing cell types or populations based on the computational analysis of their infrared spectra. *Nat. Protoc.* **5**, 1748 (2010).
22. de Voort, F. R., Sedman, J., Emo, G. & Ismail, A. A. A rapid FTIR quality control method for fat and moisture determination in butter. *Food Res. Int.* **25**, 193–198 (1992).
23. Shiroma, C. & Rodriguez-Saona, L. Application of NIR and MIR spectroscopy in quality control of potato chips. *J. Food Compos. Anal.* **22**, 596–605 (2009).
24. Anjos, O., Campos, M. G., Ruiz, P. C. & Antunes, P. Application of FTIR-ATR spectroscopy to the quantification of sugar in honey. *Food Chem.* **169**, 218–223 (2015).
25. Das, A. J., Wahi, A., Kothari, I. & Raskar, R. Ultra-portable, wireless smartphone spectrometer for rapid, non-destructive testing of fruit ripeness. *Sci. Rep.* **6**, 32504 (2016).
26. Madejova, J. FTIR techniques in clay mineral studies. *Vib. Spectrosc.* **31**, 1–10 (2003).
27. Simonescu, C. M. Application of FTIR spectroscopy in environmental studies. In *Advanced Aspects of Spectroscopy* (InTech, 2012).
28. Romo-Cárdenas, G. et al. Insulin overlapping in whole blood FTIR spectroscopy in blood glucose measurements. *Results Phys.* **7**, 1221–1222 (2017).
29. Mostaço-Guidolin, L. B. & Bachmann, L. Application of FTIR spectroscopy for identification of blood and leukemia biomarkers: a review over the past 15 years. *Appl. Spectrosc. Rev.* **46**, 388–404 (2011).
30. Sun, D.-W. *Infrared Spectroscopy For Food Quality Analysis And Control* (Academic Press, 2009).
31. Ayerden, N. P. et al. High-speed broadband FTIR system using MEMS. *Appl. Opt.* **53**, 7267–7272 (2014).
32. Omran, H., Medhat, M., Mortada, B., Saadany, B. & Khalil, D. Fully integrated Mach-Zehnder MEMS interferometer with two complementary outputs. *IEEE J. Quantum Electron.* **48**, 244–251 (2012).
33. Al-Saeed, T. A. & Khalil, D. A. Spot size effects in miniaturized moving-optical-wedge interferometer. *Appl. Opt.* **50**, 2671–2678 (2011).
34. Al-Saeed, T. A. & Khalil, D. A. Fourier transform spectrometer based on Fabry–Perot interferometer. *Appl. Opt.* **55**, 5322–5331 (2016).
35. Eltagoury, Y. M., Sabry, Y. M. & Khalil, D. A. Novel Fourier transform infrared spectrometer architecture based on cascaded Fabry-Perot interferometers. In *MOEMS Miniaturized Systems XV*, Vol. 9760, 97600L (SPIE, 2016).
36. Sandner, T., Kenda, A., Drabe, C., Schenk, H. & Scherf, W. Miniaturized FTIR-spectrometer based on optical MEMS translatory actuator. In *MOEMS Miniaturized Systems VI*, Vol. 6466, 646602 (SPIE, 2007).
37. Deutsch, E. R., Reyes, D., Schildkraut, E. R. & Kim, J. High-resolution miniature FTIR spectrometer enabled by a large linear travel MEMS pop-up mirror. In *Next-Generation Spectroscopic Technologies II*, Vol. 7319, 73190J (SPIE, 2009).
38. Scharf, T. et al. Gas detection with a micro FTIR spectrometer in the MIR region. *Procedia Chem.* **1**, 1379–1382 (2009).
39. Legtenberg, R., Groeneveld, A. W. & Elwenspoek, M. C. Comb-drive actuators for large displacements. *J. Micromech. Microeng.* **6**, 320–329 (1996).
40. Sabry, Y. M., Khalil, D. & Bourouina, T. Monolithic silicon-micromachined free-space optical interferometers onchip. *Laser Photon. Rev.* **9**, 1–24 (2015).
41. Kincaid, T. G. & Scudder, H. J. Estimation of periodic signals in noise. *J. Acoust. Soc. Am.* **42**, 1166 (1967).
42. Khalil, D. A., Mortada, B., Nabil, M., Medhat, M. & Saadany, B. A. Compensated MEMS FTIR spectrometer architecture (2013). <https://patents.google.com/patent/US8531675B2/en>.
43. Sabry, Y. M., Khalil, D., Saadany, B. & Bourouina, T. Curved silicon micromirror for linear displacement-to-angle conversion with uniform spot size. *IEEE J. Sel. Top. Quantum Electron.* **21**, 165–173 (2014).
44. Saadany, B. A., Hafez, A. N., Medhat, M. & Haddara, H. Technique to determine mirror position in optical interferometers (2014). <https://patents.google.com/patent/US8873125B2/en>.
45. Humphreys, C. J. & Kostkowski, H. J. Infrared spectra of noble gases (12000 to 19000A). *J. Res. Natl. Bur. Stand. (1934)*. **49**, 73 (1952).
46. Sabry, Y. M. et al. Integrated spectral unit. (2018). <https://patents.google.com/patent/US10060791B2/en>.
47. Sabry, Y. M. et al. Ultra-compact MEMS FTIR spectrometer. In *Next-Generation Spectroscopic Technologies X*, Vol. 10210, 102100H (2017).