



# Interaction of Terahertz Radiation with Bio-Like Objects: Theoretical and Numerical Modelling, Real Objects and Phantom Experiments

O. Smolyanskaya, Q. Cassar, O V Kravtsenyuk, E L Odlyanitskiy, Patrick Mounaix, Jean-Paul Guillet, K. Zaytsev, N V Petrov

## ► To cite this version:

O. Smolyanskaya, Q. Cassar, O V Kravtsenyuk, E L Odlyanitskiy, Patrick Mounaix, et al.. Interaction of Terahertz Radiation with Bio-Like Objects: Theoretical and Numerical Modelling, Real Objects and Phantom Experiments. Bright Far-Infrared Optoelectronic Sources Applied to Field-Matter Interaction Studies, Life Sciences and Environmental Monitoring International Workshop, IRN FIR LAB, Jul 2019, Nizhny Novgorod, Russia. hal-02482611

**HAL Id: hal-02482611**

**<https://hal.science/hal-02482611>**

Submitted on 18 Feb 2020

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# Interaction of Terahertz Radiation with Bio-Like Objects: Theoretical and Numerical Modelling, Real Objects and Phantom Experiments

O. A. Smolyanskaya<sup>1</sup>, M. S. Kulya<sup>1</sup>, Q. Cassar<sup>2</sup>, O. V. Kravtsenyuk<sup>1</sup>, E. L. Odlyanitskiy<sup>1</sup>,  
P. Mounaix<sup>2</sup>, J.-P. Guillet<sup>2</sup>, K. I. Zaytsev<sup>3,4</sup>, N. V. Petrov<sup>1</sup>

<sup>1</sup> ITMO University, Saint-Petersburg, Russia

<sup>2</sup> Bordeaux University, IMS Laboratory, Bordeaux, France

<sup>3</sup> Bauman Moscow State Technical University, Moscow, Russia

<sup>4</sup> Prokhorov General Physics Institute of the Russian Academy of Science, Moscow, Russia

smolyanskaya@corp.ifmo.ru

## I. INTRODUCTION

New biomedical devices require test objects to check their performance and periodic calibration to monitor the system efficiency over the time. A phantom is a bio-like object mimicking the intrinsic real-tissue properties that serves as a test object. A variety of phantoms are commercially available. Phantoms are already used in spectroscopy within various spectral range. However, currently, the THz frequency range suffers from a lack of investigations in this direction. Only the first steps have been performed [1]. In the present work, we investigate interaction of submillimeter radiation with the breast-mimicking skin-mimicking phantoms. Results for breast-mimicking phantom were compared with numerical models based on a double-Debye model of the dielectric permittivity for different content of fat, fibrous and cancerous tissue. The wave front propagation using the angular spectrum representation was used to simulate the theoretical response of breast-mimicking phantoms.

## II. MODEL OF WAVEFRONT PROPAGATION

Electromagnetic fields can be represented in various ways. The angular spectrum representation is a numerical technique describing optical fields in a homogeneous media [2]. Optical fields are described as a superposition of plane waves and evanescent waves. A useful approach to describe optical field diffraction is to conduct the Fourier analysis at a given plane so that the different Fourier components of the field distribution are identified as plane waves propagating away from that plane in different directions. Thus taking less computation time for numerical reconstruction. This technique is preferable for the analysis of light diffraction by histological slides [3]. Angular spectrum representation consists of the following stages: (i) the representation of the field through the angular spectrum of 2D waves (Eq.1); (ii) multiplication by the transfer function that contains a complex refractive index of the object (Eq.2); (iii) the back transition from plane waves to the wave field in the calculated z-plane (Eq.3).

$$C(f_x, f_y, \nu) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} G(\tilde{x}, \tilde{y}, \nu) \exp(-2\pi i(\tilde{x}f_x + \tilde{y}f_y)) d\tilde{x}d\tilde{y}$$

$$g_{x,y}(f_x, f_y, \nu, z) = C(f_x, f_y, \nu) \exp\left(-i \frac{2\pi \nu(z)}{c} \sqrt{1 - \left(\frac{f_x \cdot c}{\nu(\nu)}\right)^2 - \left(\frac{f_y \cdot c}{\nu(\nu)}\right)^2} \cdot z\right)$$

$$G(x, y, \nu, z) = \int_{-\infty}^{\infty} \int_{-\infty}^{\infty} g_{x,y,z}(f_x, f_y, \nu, z) \exp(2\pi i(xf_x + yf_y)) df_x df_y$$

where  $n(\tilde{x}, \tilde{y}, \nu) = n_{re}(\tilde{x}, \tilde{y}, \nu) + i \cdot n_{im}(\tilde{x}, \tilde{y}, \nu)$  is the spatial and frequency distribution of the complex index of refraction.

## III. MATERIALS AND METHODS

### A. Sample preparation

We have used three-component phantoms made of fat, protein and water of different content. Vegetable oil (10–75%), soya (13–75%) and water (0–70%) were homogenized to generate an emulsion. Then, the phantoms were deposited into vacuum packages.

### B. THz-Spectrometer

The commercially available TPS 4000 (Teraview Ltd, UK) spectrometer working in reflection mode was used. The spectral range of the system is from 0.06 to 4.50 THz. The entire system was under an air-dried dome to limit the interaction of water molecules with generated pulses. Samples were mounted on a sapphire substrate during measurement. The sapphire cut was chosen not to exhibit a birefringence that would disrupt results of measurements [4]. Each sample was measured 5 times, independently, *i.e.* the sample was removed and replaced again into system.

### C. Numerical modelling

A double-Debye model of the dielectric permittivity was used to compare optical properties of the three-component phantoms and some biological tissue (adipose, fibrous and cancerous) taken from our previous paper [5].

#### IV. RESULTS AND DISCUSSION

The propagation dynamics of the pulsed THz radiation through a tissue sample was numerically simulated in the temporal and spectral domains. The index of refraction,  $n_{re}$ , absorption coefficient,  $\alpha$ , as well as real,  $\epsilon_{re}$ , and imaginary,  $\epsilon_{im}$ , part of dielectric permittivity was calculated from the experimentally obtained data. We have obtained the time-domain spatial evolution of the THz pulse. Agreement between measured and simulated THz waveforms is confirmed by the comparison was shown. The spectral dependences of the absorption coefficient of the three-component phantoms consisting of fat, protein and water are depicted. It is shown that within the spectral window, the absorption coefficient of phantoms containing water exceeds the absorption coefficient of phantoms free of water. This is due to the fact, that water molecules, which strongly absorb THz radiations, are replaced by low absorbing components (fat and protein). Our results were compared with simulated signals of adipose, fibrous and cancerous tissues taken from our previous paper [5].

#### V. CONCLUSION

To overcome the lack of knowledge on tissue-mimicking phantoms within the submillimeter spectral range, we conducted preliminary studies on skin- and breast tissue-mimicking samples. Simulations performed for skin-mimicking phantoms have match the experimental spectral data. Thus demonstrating advantages of the technique used in this study. The spectral dependences of absorption coefficient extracted from measurements for breast-mimicking phantom are clearly differentiated into two distinct groups. One for phantoms containing water and those being water-free. Such a discrimination is also shown via the double-Debye

model of the dielectric permittivity. Thus, finding of a proper composition of tissue phantom is possible by an appropriate matching of model and experimental data, in particular for prediction of impact on THz radiation probing depth of tissue reversible dehydration using hyperosmotic optical clearing agents [3]. Deeper investigations are however still required to fulfill a set of phantoms that could be used to perform tissue measurement within the THz-range. Such a set would be profitable to design new medical THz-devices without human biopsy.

#### ACKNOWLEDGMENTS

The reported study was funded by RFBR and CNRS according to the research project №18-51-16002, and by the Government of the Russian Federation (Grant 08-08).

#### REFERENCES

- [1] E Liakhov, O Smolyanskaya, A Popov, E Odlyanitskiy, N Balbekin, M Khodzitskiy, Fabrication and characterization of biotissue-mimicking phantoms in the THz frequency range // Journal of Physics: Conference Series. 2016. No. 735. P. 012080.
- [2] Petrov, N. V., Kulya, M. S., Tsyppin, A. N., Bespalov, V. G., & Gorodetskiy, A. Application of terahertz pulse time-domain holography for phase imaging // IEEE Transactions on Terahertz Science and Technology, 2016, V. 6, No. 3, P. 464-472.
- [3] O. A. Smolyanskaya, I. J. Schelkanova, M. S. Kulya, E. L. Odlyanitskiy, I. S. Goryachev, A. N. Tsyppin, Ya V. Grachev, Ya G. Toropova, and V. V. Tuchin // Glycerol dehydration of native and diabetic animal tissues studied by THz-TDS and NMR methods // Biomedical Optics Express, 2018, V. 9, No. 3, P. 1198-1215.
- [4] Q. Cassar, A. Al-Ibadi, Ia. Mavarani, Ph. Hillger, J. Grzyb, G. Macgrogan, Th. Zimmer, U.R. Pfeiffer, J.-P. Guillet, P. Mounaix Pilot study of freshly excised breast tissue response in the 300 – 600 GHz range // Biomed Opt Express. 2018. V. 9 No. 7. P. 2930–2942.
- [5] A. J. Fitzgerald, E. Pickwell-MacPherson, and V. P. Wallace, Use of finite difference time domain simulations and Debye theory for modelling the terahertz reflection response of normal and tumour breast tissue // Plos One 9, 1-9 (2014)