

Diamond detector technology, status and perspectives

H. Kagan, A. Alexopoulos, M. Artuso, F. Bachmair, L. Bäni, M. Bartosik, J. Beacham, H. Beck, V. Bellini, V. Belyaev, et al.

To cite this version:

H. Kagan, A. Alexopoulos, M. Artuso, F. Bachmair, L. Bäni, et al.. Diamond detector technology, status and perspectives. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2019, 924, pp.297-300. $10.1016/j.nima.2018.06.009$. hal-02498265

HAL Id: hal-02498265 <https://hal.science/hal-02498265v1>

Submitted on 19 Mar 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

The University of Manchester Research

Diamond detector technology, status and perspectives

DOI: [10.1016/j.nima.2018.06.009](https://doi.org/10.1016/j.nima.2018.06.009)

Document Version

Accepted author manuscript

[Link to publication record in Manchester Research Explorer](https://www.research.manchester.ac.uk/portal/en/publications/diamond-detector-technology-status-and-perspectives(da6e276c-9de8-4a14-87d7-d08d7b0a68e4).html)

Citation for published version (APA):

Kagan, H., Alexopoulos, A., Artuso, M., Bachmair, F., Bäni, L., Bartosik, M., ... Zavrtanik, M. (2018). [Diamond](https://www.research.manchester.ac.uk/portal/en/publications/diamond-detector-technology-status-and-perspectives(da6e276c-9de8-4a14-87d7-d08d7b0a68e4).html) [detector technology, status and perspectives.](https://www.research.manchester.ac.uk/portal/en/publications/diamond-detector-technology-status-and-perspectives(da6e276c-9de8-4a14-87d7-d08d7b0a68e4).html) Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment. <https://doi.org/10.1016/j.nima.2018.06.009>

Published in:

Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment

Citing this paper

Please note that where the full-text provided on Manchester Research Explorer is the Author Accepted Manuscript or Proof version this may differ from the final Published version. If citing, it is advised that you check and use the publisher's definitive version.

General rights

Copyright and moral rights for the publications made accessible in the Research Explorer are retained by the authors and/or other copyright owners and it is a condition of accessing publications that users recognise and abide by the legal requirements associated with these rights.

Takedown policy

If you believe that this document breaches copyright please refer to the University of Manchester's Takedown Procedures [http://man.ac.uk/04Y6Bo] or contact uml.scholarlycommunications@manchester.ac.uk providing relevant details, so we can investigate your claim.

¹ Diamond detector technology, status and perspectives

⁵⁰ Abstract

 Detectors based on Chemical Vapor Deposition (CVD) diamond have been used extensively ₅₂ and successfully in beam conditions/beam loss monitors as the innermost detectors in the highest radiation areas of Large Hadron Collider (LHC) experiments. The startup of the LHC in 2015 brought a new milestone where the first polycrystalline CVD (pCVD) diamond pixel modules were installed in an LHC experiment and successfully began operation. The RD42 collaboration at CERN is leading the effort to develop polycrystalline CVD diamond as a material for tracking detectors operating in extreme radiation environments. The status of the RD42 project with emphasis on recent beam test results is presented.

⁵⁹ *Keywords:* Chemical Vapor Deposition, pCVD diamond, diamond detectors, 3D diamond ⁶⁰ detectors, radiation tolerant detectors

61 1. Introduction

 62 The RD42 collaboration [1, 2] at CERN is leading the effort to develop radiation tolerant ⁶³ devices based on pCVD diamond as a material for tracking detectors operating in harsh radiation ⁶⁴ environments. Diamond has properties which make it suitable for such detector applications. ⁶⁵ During the last few years the RD42 group has succeeded in producing and measuring a number ⁶⁶ of devices to address specific issues related to use at the HL-LHC [3, 4]. This paper presents 67 the status of the RD42 project with emphasis on recent beam test results. In particular, results ⁶⁸ are presented on the status of the first diamond pixel detector based on pCVD material, on the ⁶⁹ independence of signal size on incident particle rate in pCVD diamond detectors over a range of σ particle fluxes up to 20 MHz/cm² and on the 3D diamond detectors fabricated in pCVD diamond.

⁷¹ 2. Status of the ATLAS Diamond Beam Monitor

⁷² The startup of the LHC in 2015 brought a new milestone for diamond detector development ⁷³ where the first planar diamond pixel modules based on pCVD diamond were installed in an ⁷⁴ LHC experiment, the ATLAS experiment [5], and successfully began operation. The ATLAS ⁷⁵ Diamond Beam Monitor (DBM) [6, 7] was designed to measure the instantaneous luminosity, τ_6 the background rates and the beam spot position. A single DBM module consists of an 18 mm \times 77×21 mm pCVD diamond 500 μ m thick instrumented with a FE-I4 pixel chip [8]. The 26,880 pixels
 78×26 are arranged in 80 columns on 250 μ m pitch and 336 rows on 50 μ m pitch resulting in an active ⁷⁸ are arranged in 80 columns on 250 μ m pitch and 336 rows on 50 μ m pitch resulting in an active area of 16.8 mm \times 20.0 mm. This fine granularity provides high precision particle tracking. The area of 16.8 mm \times 20.0 mm. This fine granularity provides high precision particle tracking. The ⁸⁰ deposited charge from a particle is measured in the FE-I4 by Time-over-Threshold. 81 The ATLAS DBM uses diamonds with a charge collection distance (the average distance an

⁸² electron-hole pair move apart under the influence of the applied electric field) of $200-220 \mu m$ at an applied bias voltage of 500 V. Three telescopes each with 3 diamond DBM modules (plus 1 an applied bias voltage of 500 V. Three telescopes each with 3 diamond DBM modules (plus 1 ⁸⁴ telescope with silicon sensors) mounted as a three layer tracking device were installed inside the ⁸⁵ pixel detector services on each side of the ATLAS interaction point at 90 cm $< |z| < 111$ cm,

Email address: harris.kagan@cern.ch (H. Kagan)

 $3.2 < |\eta| < 3.5$ and at a radial distance from 5 cm to 7 cm from the center of the beam pipe. The modules have an inclination of 10° with respect to the ATLAS solenoid magnetic field direction 87 modules have an inclination of 10° with respect to the ATLAS solenoid magnetic field direction to suppress erratic dark currents [9] in the diamonds. The ATLAS DBM data-acquisition system is shared with the ATLAS IBL [10]. After initial installment, data were collected in the July 2015 run. These data have been analyzed and the first results of the ATLAS DBM tracking capabilities are shown in Fig. 1. A clear separation between background particles from unpaired bunches (open circles) and collision particles from colliding bunches (filled circles) is observed. After two electrical incidents in 2015 with consequent loss of several silicon and diamond modules, the DBM has now been re-commissioned and is again in the operation phase.

Figure 1: Radial distance (left plot) and longitudinal distance (right plot) of the closest approach of the projected particle tracks to the interaction point as recorded by a single DBM telescope with preliminary alignment.

95 3. Rate Studies in pCVD diamond

 In order to study the dependence of signal size on incident particle rate, RD42 performed a 97 series of beam tests in the πM1 beam line of the High Intensity Proton Accelerator (HIPA) at Paul Scherrer Institute (PSI) [11]. This beam line is able to deliver 260 MeV/c π⁺ fluxes from a ⁹⁸ Paul Scherrer Institute (PSI) [11]. This beam line is able to deliver 260 MeV/c π⁺ fluxes from a
₂₈ rate of \approx 5 kHz/cm² to a rate \approx 20 MHz/cm² in bunches spaced 19.8 ps apart 99 rate of ~5 kHz/cm² to a rate ~20 MHz/cm² in bunches spaced 19.8 ns apart.

100 Sensors using pCVD material [12] were tested in a tracking telescope [13] based on $100 \mu m \times 150 \mu m$ silicon pixel sensors read out by the PSI46v2 pixel chip [14]. The diamond signals $101 \times 150 \mu$ m silicon pixel sensors read out by the PSI46v2 pixel chip [14]. The diamond signals

were amplified with custom-built front-end electronics with a peaking time of ~6 ns. returnwere amplified with custom-built front-end electronics with a peaking time of ∼6 ns, return- to-baseline in ∼16 ns and 550*e* noise with 2 pf input capacitance. The amplified signals were recorded with a DRS4 evaluation board [15] operating at 2 GS/s. The entire system was triggered with a scintillator which determined the timing of the beam particles with a precision of ∼0.7 ns. A series of cuts were applied to the data including: removing 60 s of triggers at the beginning of each run, removing triggers from heavily ionizing particles with saturated waveforms (mostly protons), removing calibration triggers, removing triggers in the wrong beam bucket, removing triggers with no tracks in the telescope and removing triggers with large angle tracks in the telescope. After applying this procedure all telescope tracks which project into the diamond fiducial region have a pulse height well separated from the pedestal distribution in the diamond i.e. the diamond is 100% efficient at all rates. The same procedure was applied to all particle flux points and the resulting mean pulse height (in arbitrary units) versus rate is shown in Fig. 2 for both positive and negative bias voltage. The uncertainty on the data points in the plot include both statistical and systematic sources. The systematic uncertainty was determined by assuming 116 any deviations in pulse height for rates below 80 kHz/cm² were due to systematic effects. Thus

¹¹⁷ the spread in the data points at a given rate indicates the reproducibility of the data. Fig. 2 ¹¹⁸ indicates the mean pulse height in pCVD diamond detectors irradiated up to 5×10^{14} n/cm² does

119 not depend strongly on rate up to rates of 20 MHz/cm².

Figure 2: The average pulse height versus rate for an un-irradiated and irradiated pCVD diamond pad detector at positive and negative bias. The beam line parameters were adjusted to set the different particle rates. The data was taken by scanning up and down in rate multiple times. The pulse height units are arbitrary since the un-irradiated and irradiated detectors used different readout electronics. The resulting electronics gain corrections and the relative gain correction for positive versus negative signals in the electronics is still being determined and has not been applied.

¹²⁰ 4. 3D Diamond Pixel Detectors

¹²¹ 3D sensors with electrodes in the bulk of the sensor material were first proposed in 1997 [16] ¹²² in order to reduce the drift distance of the charged carriers to much less than the sensor thickness. ¹²³ In order to achieve this goal a series of alternating + and − electrodes perpendicular to the read ¹²⁴ out face were created in the bulk detector material. This idea is particularly beneficial in detectors ¹²⁵ with a limited mean free path such as trap dominated sensor materials like heavily irradiated ¹²⁶ silicon and pCVD diamond where the observed signal size is related to the mean free path divided ¹²⁷ by the drift distance. Under these circumstances one gains radiation tolerance (larger signals) by ¹²⁸ keeping the drift distance less than the mean free path. With this geometrical structure charge 129 carriers drift inside the bulk parallel to the surface over a typical drift distance of 25-100 μ m
130 instead of perpendicular to the surface over a distance of 250-500 μ m. 130 instead of perpendicular to the surface over a distance of $250-500 \mu m$.

In 2015 RD42 published results of a 3D device fabricated in single-

 In 2015 RD42 published results of a 3D device fabricated in single-crystal CVD diamond [17] 132 showing that the 3D structure works in diamond. In 2016 RD42 fabricated the first 3D device in pCVD diamond [18]. The electrodes in the bulk of the pCVD diamond 3D device were fabricated with lasers as described in [17]. The bias electrodes were placed at the corners and the readout electrodes were placed in the middle of the cells. This pCVD device was shown to collect more than 75% of the deposited charge which translates in more than a factor of two more charge than 137 a planar diamond strip detector fabricated on the same pCVD diamond.

138 In 2017 RD42 successfully constructed the first pCVD diamond 3D pixel detector with 50 μ m
139 \times 50 μ m cells. This pixel device is designed to be read out with the RD53 pixel readout chip [19] $139 \times 50 \mu$ m cells. This pixel device is designed to be read out with the RD53 pixel readout chip [19]

which is not yet available. In order to read this device out with an existing pixel readout chip a which is not yet available. In order to read this device out with an existing pixel readout chip a ¹⁴¹ small number of cells were ganged together to match the pitch of the pixel readout chip. RD42 ¹⁴² is proceeding to make 3D diamond pixel devices compatible with both the CMS pixel readout ¹⁴³ chip (3 × 2 ganging) and the ATLAS pixel readout chip (1 × 5 ganging). The first 50 μ m × 50
¹⁴⁴ μ m nCVD diamond 3D pixel device which was bump-bonded used the CMS pixel readout chip. 144μ m pCVD diamond 3D pixel device which was bump-bonded used the CMS pixel readout chip.
This first diamond 3D pixel device was tested during the Aug 2017 beam test at PSI at a This first diamond 3D pixel device was tested during the Aug 2017 beam test at PSI at a ¹⁴⁶ single voltage and with rates from 7 kHz/cm² to 7 MHz/cm². During the initial lab test it was ¹⁴⁷ discovered that the bump bonding had a small issue on one edge. We decided to take data with the ¹⁴⁸ device rather than try to repair this small bump bonding issue. Fig. 3 (left) shows the preliminary ¹⁴⁹ efficiency as a function of *xy* position for every cell in the device with a 1500*e* pixel threshold. ¹⁵⁰ The red box marks the fiducial region used to measure the hit efficiency. The blue circle indicates ¹⁵¹ the position of the one non-working pixel cell in the central region of the device. Fig. 3 (right) ¹⁵² shows the hit efficiency in the fiducial region with the 1500*e* pixel threshold as a function of time ¹⁵³ during an up-down scan of incident particle rates from 7 kHz/cm² to 7 MHz/cm² and back to 7 $kHz/cm²$. The overall measured efficiency is 99.2% and no change in efficiency as a function of ¹⁵⁵ rate is observed. The corresponding efficiency for a planar silicon CMS pixel detector in this test ¹⁵⁶ was 99.7% with no change in efficiency as a function of rate. The slight loss of efficiency (0.5%) , ¹⁵⁷ assuming it holds through the completion of the analysis, is most likely due to charge loss in the ¹⁵⁸ column electrodes. If this explanation is correct, then this effect can be easily remedied by tilting ¹⁵⁹ the detector at a small angle with respect to the incident beam.

Figure 3: The hit efficiency of the first $50\mu m \times 50 \mu m$ cell pCVD 3D pixel detector with 3×2 ganged cells read out with CMS pixel electronics with a 1500*e* threshold. The left plot shows the efficiency of each ganged cell as a function of *xy* position in the device. The right plot shows the average efficiency in the fiducial region (the red box in the left plot) as a function of time during the run at 67 kHz/cm^2 .

¹⁶⁰ 5. Conclusions

¹⁶¹ The recent progress in the design, fabrication and testing of polycrystalline CVD diamond ¹⁶² detectors was presented. The following milestones have been achieved: successful operation of ¹⁶³ the first pCVD diamond planar pixel detector in the ATLAS experiment at the LHC; demonstration that the average signal pulse height of pCVD diamond detectors irradiated up to 5×10^{14} 164 $n/cm²$ is independent of the particle flux up to ~20 MHz/cm²; successful fabrication and opera-¹⁶⁶ tion of the first pCVD diamond 3D pixel detector with 50 μ m \times 50 μ m pixels read out with CMS
¹⁶⁷ pixel electronics where the efficiency for a MIP was >99% and the average charge collected in ¹⁶⁷ pixel electronics where the efficiency for a MIP was >99% and the average charge collected in
¹⁶⁸ the device was >90% of the deposited charge. In the future RD42 plans to study the pulse height ¹⁶⁸ the device was >90% of the deposited charge. In the future RD42 plans to study the pulse height dependence of CVD diamond sensors with pad and pixel electrodes with radiation doses up to dependence of CVD diamond sensors with pad and pixel electrodes with radiation doses up to

 10^{17} n/cm² and continue the development of 3D diamond detectors with the production of a 50 171μ μ m \times 50 μ m cell pCVD diamond 3D pixel detector compatible with ATLAS readout electronics.

6. Acknowledgments

 The RD42 Collaboration gratefully acknowledges the staff at CERN for test beam time and their help in setting up the beam conditions. We would especially like to thank Henric Wilkins, the test beam coordinator, for his assistance in making our tests a success. We also thank the beam line staff at the PSI High Intensity Proton Accelerator especially Konrad Deiters, Manuel Schwarz and Davide Reggiani for their assistance in carrying out the diamond detector tests. We extend our gratitude to Prof. Lin Li and David Whitehead of the University of Manchester Laser 179 Processing Center for their assistance in the production of 3D diamond devices. The research leading to these results received funding from the European Union's Horizon 2020 research and 181 innovation program under grant agreement No. 654168. This work was also partially supported by the Swiss National Science Foundation grant #20FL20 154216, ETH grant 51 15-1, Royal 183 Society Grant UF120106 and the U.S. Department of Energy through grant DE-SC0010061.

References

- [1] W. Adam, *et al.* [RD42 collaboration], *Development of Diamond Tracking Detectors for High Luminosity Experi- ments at the LHC*, Proposal/RD42 CERN/DRDC 94-21, Status Report/RD42, CERN/LHCC, 95-43, 95-53, 95-58, 97-03, 98-20, 2000-011, 2000-015, 2001-002, 2002-010, 2003-063, 2005-003, 2006-010, 2007-002, 2008-005.
- [2] M. Artuso, *et al.* [RD42 Collaboration], RD42 Status Report: Development of Diamond Tracking Detectors for High Luminosity Experiments at the LHC, CERN-LHCC-2017-006.
- [3] G. Apollinari, *et al.*, *Chapter 1: High-Luminosity Large Hadron Collider HL-LHC* in *High-Luminosity Large Hadron Collider (HL-LHC):Preliminary Design Report*, CERN Yellow Reports: Monographs, CERN-2015-005, CERN Geneva, 2015, https://doi.org/10.5170/CERN-2015-005.1.
- [4] G. Apollinari, *et al.*, *High-Luminosity Large Hadron Collider (HL-LHC): Technical Design Report V.0.1*, CERN Yellow Reports: Monographs, Volume 4/2017 CERN-2017-007-M, CERN Geneva 2017, https://doi.org/10.23731/CYRM-2017-004.
- [5] G. Aad, *et al.* [ATLAS collaboration], *The ATLAS experiment at the CERN Large Hadron Collider*, JINST 3 S08003, 2008.
- [6] H. Kagan, *et al.*, *ATLAS Diamond Beam Monitor TDR*, ATLAS Document DBM 001 (2011).
- 200 [7] M. Červ, et al., *The ATLAS Diamond Beam Monitor*, JINST 9 C02026, 2014.
- [8] M. Barbero, *et al.*, *The FE-I4 pixel readout chip and the IBL module*, PoS(Vertex 2011) 038 (2011).
- [9] A.J. Edwards, *et al.*, *Radiation monitoring with diamond sensors in BABAR*, IEEE Trans. Nucl. Sci. 51 (2004) 1808.
- [10] F. Huegging, *et al.* [ATLAS Collaboration], *The ATLAS Pixel Insertable B-Layer (IBL)*, Nucl. Instrum. Meth. Phys. Res. A 650 (2011) 45.
- [11] PSI High Intensity Proton Accelerator, *High Energy Beam Lines*, http://www.psi.ch/abe/high-energy-beam-lines
- [12] The diamond material for these tests was provided by II-VI Inc., 360 Saxonburg Road, Saxonburg, PA.
- [13] F. Bachmair, *CVD Diamond Sensors In Detectors For High Energy Physics*, Ph.D. Thesis, ETH Zurich (2016), ¨ https://doi.org/10.3929/ethz-a-010748643.
- 210 [14] H.Chr. Kästli, et al., *Design and performance of the CMS pixel detector readout chip*, Nucl. Instrum. Meth. Phys. Res. A 565(2006) 188.
- [15] S. Ritt, *DRS4 Evaluation Board*, http://www.psi.ch/drs/evaluation-board
- [16] S. Parker, C.J. Kenney and J. Segal, *3-D: A proposed new architecture for solid state radiation detectors*, Nucl. Instrum. Meth. Phys. Res. A 395 (1997) 328.
- [17] F. Bachmair, *et al.* [RD42 Collaboration], *A 3D Diamond Detector for Particle Tracking*, Nucl. Instrum. Meth. Phys. Res. A, 786 (2015) 97.
- [18] A. Alexopoulos, *et al.* [RD42 Collaboration], *Diamond detector technology: status and perspectives*, PoS(Vertex 2016) 027 2016.
- [19] M. Garcia-Sciveres, *et al.* [RD53 Collaboration], *The RD53A Integrated Circuit*, CERN-RD53-PUB-17-001 (2017). https://cds.cern.ch/record/2287593?ln=en.