



## Thulium spiked gel for internal standardisation in LA-ICP-MS bio-imaging: quantitative elemental distribution of uranium in kidney tissue

Nagore Grijalba Marijuan, Alexandre Legrand, Yann Gueguen, Valerie Holler,  
Celine Bouvier Capely

### ► To cite this version:

Nagore Grijalba Marijuan, Alexandre Legrand, Yann Gueguen, Valerie Holler, Celine Bouvier Capely. Thulium spiked gel for internal standardisation in LA-ICP-MS bio-imaging: quantitative elemental distribution of uranium in kidney tissue. European Winter Conference on plasma Spectrochemistry, EWCPS 2019, Feb 2019, PAU, France. 2019. hal-02635599

**HAL Id: hal-02635599**

**<https://hal.science/hal-02635599>**

Submitted on 27 May 2020

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Copyright



# Thulium spiked gel for internal standardisation in LA-ICP-MS bio-imaging: quantitative elemental distribution of uranium in kidney tissue

Nagore GRIJALBA, Alexandre LEGRAND, Yann GUEGUEN, Valérie HOLLER, Céline BOUVIER-CAPELY

Institut de Radioprotection et de Sûreté Nucléaire, PSE-SANTE/SESANE/LRSI, 31 Av de la Division Leclerc BP 17, 92262 Fontenay-aux-Roses Cedex, France


The quantitative analysis of trace metals in different organs or cellular structures is a topic of emerging interest for the assessment of toxicological risk. The kidney is recognized as a major site for uranium accumulation able to induce renal toxicity<sup>1,2</sup>. Several studies have shown its heterogeneous distribution within the tissue finding areas (S3 segments in the proximal tubule) of high uranium concentration (100-fold above mean renal concentration)<sup>3-5</sup>. These studies were carried out employing high-energy synchrotron radiation X-ray fluorescence analysis (SR-XRF) whose reduced availability limits its daily use for routine analysis. In this work, mass spectrometry imaging (MSI) using laser ablation inductively coupled plasma mass spectrometry (LA-ICP-MS) has been employed for mapping and quantifying uranium in histological tissue sections of mouse kidney. To the author's knowledge just a single work has been published recently for the semi-quantitative analysis of uranium in mice kidneys due to the lack of an appropriate internal standard<sup>6</sup>. The quantitative monitoring of uranium at tissue level in kidney would facilitate the understanding of its action mechanism in renal toxicity. Therefore, this work presents the development of a correction methodology based on doped gelatine with internal standard as an alternative to current methods<sup>7</sup>. In order to correct matrix effects, lack of tissue homogeneity and instrumental drift, a thulium (Tm) spiked gel was prepared and deposited on the top of glass microscope slides. For quantification purposes, matrix-matched laboratory standards were prepared from a pool of rat kidneys by spiking each level with different concentrations of uranium. The proposed analytical bio-imaging approach was successfully applied for quantification of uranium of rat kidney samples.

## Bio-distribution and quantification of uranium in kidney tissue. How to do it?


### External calibration: matrix-matched standards

The use of matrix-matched standards does not suppress the elemental fractionation but it will happen similarly both in standards and samples.

#### General scheme for the preparation of matrix-matched standards




#### Verification of uranium real concentration: microwave assisted acid digestion + quantification by isotope dilution




### Internal standardization: spiked gelatine

In addition, the external calibration method needs the use of an internal standard (IS) to compensate matrix effects as well as variations in ablated and transported mass and instrumental drifts during analysis. Ideal IS should behave in a similar manner to the analyte during the ablation process and in the ICP. Additionally, it must be in similar concentration and homogeneously distributed within the samples and standard matrices.

#### General scheme for thulium/thallium 100 ppb doped gelatine preparation



#### Optimisation of gelatine %



### Instrumentation and experimental conditions

