



HAL
open science

Ultra-short echo time Magnetic Resonance Elastography

Pilar Sango Solanas, Kevin Tse-Ve-Koon, Eric van Reeth, Cyrielle Caussy, O. Beuf

► **To cite this version:**

Pilar Sango Solanas, Kevin Tse-Ve-Koon, Eric van Reeth, Cyrielle Caussy, O. Beuf. Ultra-short echo time Magnetic Resonance Elastography. ISMRM 2020 28th Annual Meeting & Exhibition, Aug 2020, Virtual Conference, France. hal-02866021

HAL Id: hal-02866021

<https://hal.science/hal-02866021>

Submitted on 12 Jun 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Ultra-short echo time Magnetic Resonance Elastography

Pilar Sango Solanas¹, Kevin Tse Ve Koon¹, Eric Van Reeth¹, Cyrielle Caussy², and Olivier Beuf¹

¹Univ Lyon, INSA-Lyon, Université Claude Bernard Lyon 1, UJM-Saint Etienne, CNRS, Inserm, CREATIS UMR 5220, U1206, F-69616, Lyon, France, Lyon, France, ²Département d'Endocrinologie, Diabète et Nutrition, Centre Hospitalier Lyon Sud, Hospices Civils de Lyon, CarMeN, INSERM U1060, INRA U1397, Lyon, France, Lyon, France

Synopsis

Magnetic Resonance Elastography (MRE) is a valuable technique to quantitatively characterize mechanical properties of tissues based on the properties of shear waves propagation. In this study, a radial acquisition MRE sequence potentially able to quantify viscoelastic parameters of tissues whose T2 values are very short is proposed. To this end, an optimal control-based RF pulse is applied with a constant gradient during the mechanical excitation to simultaneously perform spatially selective excitation and motion encoding. Acquisition is started right after, enabling a very short TE. Results on phantom experiments demonstrated the feasibility of our ultra-short echo time MRE technique.

Introduction

MRE is conventionally performed by the application of oscillating motion encoding gradients (MEG) that are able to encode the spins motion oscillating at a given frequency into the NMR signal phase¹. Usually, several periods of MEG are needed to increase the phase encoding but this results in a long echo time and thus a degraded SNR. A new encoding strategy of the shear wave propagation, by simultaneously applying an Optimal Control (OC) RF pulse and a constant gradient was already demonstrated² when optimized pulses were applied in a conventional RARE MRE sequence, achieving echo times of about 10ms. In order to further reduce the TE, we propose to create a radial acquisition MRE sequence. For this purpose, a UTE sequence was adapted by adding specifically designed OC RF pulses.

Methods

Optimal control theory was applied to MRE to design the optimized RF pulses. The computation is performed with the GRAPE algorithm³. As we know that the macroscopic magnetization \vec{M} is governed by Bloch equations, the OC algorithm provides the RF control parameter to apply in order to manipulate the isochromats from the steady state towards a desired target state.

The RF pulse is simultaneously applied with a constant gradient G to perform slice selectivity⁴ while a shear wave motion at frequency f_e produces static field variations given by $\Delta B_0^{(i,j)}(t) = G(A \sin(-2\pi f_e t + \theta^{(i)}) + z^{(j)})$, where $z^{(j)}$ represents the isochromat (i,j) location in the slice (Figure 1). Finally, different target states are attributed whether the considered isochromat is located in the passband (Δz_{in}) or in the stopband (Δz_{out}). The isochromats phase is directly linked to the phase of the wave motion they are experiencing⁵:

$$\vec{T}^{(i,j)} = \rho(\cos\Phi^i, \sin\Phi^i, 0) \text{ if } j \in \Delta z_{in}$$
$$\vec{T}^{(i,j)} = \rho(0, 0, 1) \text{ if } j \in \Delta z_{out}$$

A UTE readout was experimentally compared to a Turbo Spin Echo using the same OC RF pulse and a conventional MRE using a similar Turbo Spin Echo through phantom MRE experiments. For this purpose, a UTE sequence was modified by adding the optimized pulses dedicated to MRE. Radial acquisitions trajectories starting in the out-center of the k space were used.

All the experiments were carried on an agar phantom doped with NiSO₄ (0.2% NiSO₄, 3.8% agar) with [T₁, T₂]=[230,45]ms. A piezoelectric actuator (CEDRAT Technologies) was used as mechanical transducer for motion generation.

MRI measurements were run on a Bruker 4.7T small animal MRI system. For the OC techniques, the pulse was optimized for an excitation frequency of 400 Hz, a T₂=26ms and a duration of 13.8ms. Three different sequences with unchanged TR and geometry were used: classical RARE MRE, OC RARE and OC UTE (Figure 2). Acquisitions with (400Hz shear wave) and without mechanical excitation were acquired. Two acquisitions with inversion of the polarity were always performed for phase images subtraction to remove static phase errors.

The phase to noise ratio PNR is calculated to evaluate and compare the motion encoding efficacy of the sequences. Since the SNR obtained from magnitude images cannot be directly used to compare radial and cartesian acquisitions, a phase method is used to calculate the noise level of the phase image. We defined the phase encoding Δ_{xy} as the maximal phase variation encoded in the (x,y) pixel along the temporal (phase offsets) direction. A ROI of 3600 pixels was defined to establish the limits of the phantom in the phase image. The PNR was computed as the ratio between the average value of phase encoding $\bar{\Delta}$ obtained when a shear waves is propagating through the phantom and the phase noise σ , calculated as the mean phase encoding in the absence of mechanical excitation:

$$PNR = \frac{\bar{\Delta}}{\sigma} = \frac{\left(\sum_{(x,y) \in ROI} \Delta_{xy} \right)_{MotionON}}{\left(\sum_{(x,y) \in ROI} \Delta_{xy} \right)_{MotionOFF}}$$

Results

Phase images with and without excitation were analyzed. Shear wave phase images are shown in Figure 3 for each acquisition type. The wave propagation is perfectly visualized for all of them.

Figure 4 illustrates phase images of the OC UTE method and the respective temporal phase evolutions with and without excitation.

Figure 5 compares the results of the three presented MRE sequences in terms of Δ , σ and PNR. The OC RARE sequence produces the highest value of phase encoding but the OC UTE scheme achieves the lowest value of phase noise while yielding the best PNR. As a result, despite the difference between phantom's T2-value and the T2 used for the pulse optimization, both OC methods present better PNR values compared to the classical RARE MRE. A gain of 2 in the PNR is reached when comparing the classical RARE MRE (TE=12.5ms) and the OC UTE (TE=2.4ms) methods.

Conclusion

A new motion encoding method consisting in applying OC to MRE with a radial acquisition readout was demonstrated on phantom experiments. The OC pulse, which simultaneously performs frequency selective excitation and motion encoding, allows extremely short echo times yielding the highest PNR values when OC UTE is used. Future work will focus on evaluating the OC pulses robustness to the optimization parameters variation to improve the motion encoding of OC MRE methods. Pre-clinical studies on mice livers will be also conducted to evaluate the feasibility of the OC UTE method on *in vivo* experiments.


Acknowledgements

This work was performed within the scope of LABEX PRIMES (ANR-11-LABX-0063). Experiments were performed on the PILoT facility, part of the France Life Imaging infrastructure (ANR-11-INBS-0006).

References

1. Muthupillai R, Lomas D, Rossman P, Greenleaf J, Manduca A, Ehman R: Magnetic resonance elastography by direct visualization of propagating acoustic strain waves. *Science* 1995; 269:1854–1857.
2. P. Sango Solanas et al., « Constant Gradient Magnetic Resonance Elastography Experiments on Phantom and Bovine Liver. » In ISMRM 27th Annual Meeting & Exhibition, 11-16 May 2019 Montréal, QC, Canada; #06831.
3. Khaneja N, Reiss T, Kehlet C, Schulte-Herbrüggen T, Glaser SJ: Optimal control of coupled spin dynamics: design of NMR pulse sequences by gradient ascent algorithms. *Journal of Magnetic Resonance* 2005; 172:296–305.
4. Vinding MS, Maximov II, Tošner Z, Nielsen NChr: Fast numerical design of spatial-selective rf pulses in MRI using Krotov and quasi-Newton based optimal control methods. *The Journal of Chemical Physics* 2012; 137:054203.
5. Lefebvre PM, Van Reeth E, Ratiney H, et al.: Active control of the spatial MRI phase distribution with optimal control theory. *Journal of Magnetic Resonance* 2017; 281:82–93.


Figures




Experimental setup: the shear wave, of f_e mechanical excitation frequency, propagates (amplitude A) in the phantom along the x axis. The shear wave-dephasing angle is defined as $\theta^i = 2\pi x^i / \lambda$ with x^i representing a specific position.

Sequence type	TE (ms)	Shear wave amplitude (V)
Classical RARE MRE	12.5	2
		0
OC RARE	7.0	2
		0
OC UTE	2.4	2
		0


Acquisition parameters carried out on the phantom. Geometry: matrix size=128x128, FOV=4.5cm, slice thickness=1mm. TR=800ms.



Unwrapped phase images obtained with the different acquisitions types.



a) Phase images and b) phase time evolution profile of a given pixel (in red) obtained with the OC UTE method in the presence (on the top) and the absence (on the bottom) of mechanical excitation. The definition of phase encoding and phase noise are illustrated on the profile figures.



Comparison between the classical and the OC MRE methods for phantom experiments in terms of phase encoding, phase noise and PNR respectively.