



HAL
open science

Reconstruction for Diverging-Wave Imaging Using Deep Convolutional Neural Networks

Jingfeng Lu, Fabien Millioz, Damien Garcia, Sébastien Salles, Wanyu Liu, Denis Friboulet

► **To cite this version:**

Jingfeng Lu, Fabien Millioz, Damien Garcia, Sébastien Salles, Wanyu Liu, et al.. Reconstruction for Diverging-Wave Imaging Using Deep Convolutional Neural Networks. *IEEE Transactions on Ultrasonics, Ferroelectrics and Frequency Control*, 2020, 67 (12), pp.2481-2492. <10.1109/TUFFC.2020.2986166>. <hal-02919384>

HAL Id: hal-02919384

<https://hal.science/hal-02919384v1>

Submitted on 4 Sep 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

Reconstruction for Diverging-Wave Imaging Using Deep Convolutional Neural Networks

Jingfeng Lu, Fabien Millioz, Damien Garcia, Sébastien Salles, Wanyu Liu, and Denis Friboulet

Abstract—In recent years, diverging wave (DW) ultrasound imaging has become a very promising methodology for cardiovascular imaging due to its high temporal resolution. However, if they are limited in number, DW transmits provide lower image quality compared with classical focused schemes. A conventional reconstruction approach consists in summing series of ultrasound signals coherently, at the expense of frame rate, data volume, and computation time. To deal with this limitation, we propose a convolutional neural networks (CNN) architecture, IDNet, for high-quality reconstruction of DW ultrasound images using a small number of transmissions. Given the spatially varying properties of DW images along depth, we adopted the inception model composed of the concatenation of multi-scale convolutional kernels. Incorporating inception modules aims at capturing different image features with multi-scale receptive fields. A mapping between low-quality images and corresponding high-quality compounded reconstruction was learned by training the network using *in vitro* and *in vivo* samples. The performance of the proposed approach was evaluated in terms of contrast ratio (CR), contrast-to-noise ratio (CNR) and lateral resolution (LR), and compared with standard compounding method and conventional CNN methods. The results demonstrate that our method could produce high-quality images using only 3 DWs, yielding an image quality equivalent to that obtained with compounding of 31 DWs and outperforming more conventional CNN architectures in terms of complexity, inference time and image quality.

Index Terms—Diverging wave, image reconstruction, convolutional neural networks, ultrasound imaging.

I. INTRODUCTION

ULTRASOUND imaging has become the modality of choice for cardiovascular imaging because of its non-invasive, cost-efficient, and real time properties. In conventional ultrasound imaging, several sectors of the entire image are reconstructed using sequential narrow beams. The frame rate of conventional schemes mainly depends on the number of transmitted beams required to construct an image. Limitations arise when monitoring highly transient biological phenomena faster than the frame rate delivered by the conventional focused scheme. Tracking mechanical waves, such as remotely induced shear waves [1] and electromechanical waves [2], is a

representative example of such challenge as their propagation speed in myocardium can reach 1 to 10 m/s [3].

To reduce the number of transmissions, multi-line acquisition [4] and multi-line transmit [5] have been introduced. Multi-line acquisition approach uses broader transmit beam and reconstruct multiple image lines (e.g., 2, 4, or 8) for each transmission. In the multi-line transmit method, multiple focused beams are simultaneously transmitted. In order to increase frame rate while preserving the number of scan lines, ultrafast imaging using unfocused transmit beam has been proposed. These approaches use plane waves (PW) [6] or diverging waves (DW) [7], [8] to image a wide field of view. However, in PW or DW imaging, the acoustic energy of unfocused beams is spread onto a wider area, resulting in a deterioration of the quality of reconstructed images if no additional processing is performed.

To alleviate these effects, coherent compounding [8]–[10] of unfocused beams has been proposed. Coherent compounding consists in transmitting multiple consecutive beams at different angles. The backscattered echoes are then coherently summed to improve contrast and resolution. Therefore, a trade-off needs to be made between image quality and frame rate since compounding of more beams produces images of higher quality but decreases the image rate. Therefore, to achieve a high image quality while maintaining the frame rate of original unfocused wave imaging is of high interest.

In recent years, deep learning has achieved state-of-the-art performance in various problems of image processing, including image classification, recognition, and segmentation. The success of deep learning methods lies in its exponentially increasing expressiveness, which can capture modality-specific features [11]. Inspired by the success of deep learning, many researchers have investigated deep learning methods for medical image reconstruction and achieved significant performance [12]–[19]. For instance, Jin et al. [12] proposed to use convolutional neural networks (CNN) to generate high-quality X-ray computed tomography (CT) images from low-quality images reconstructed from sparse views. Lee et al. [19] proposed a deep residual learning network [20] for the reconstruction of magnetic resonance (MR) images from accelerated MR acquisition.

Regarding ultrasound imaging, a number of studies have recently been carried out on image reconstruction based on deep learning [21]–[36]. Among these works, only a few studies were devoted to the specific problem of PW reconstruction [22], [24], [26], [27], [32], [33]. The studies in [26], [27] dealt with PW reconstruction as a beamforming problem (i.e., raw radio frequency (RF) data were processed)

J. Lu is with METISLab, School of Instrumentation Science and Engineering, Harbin Institute of Technology, Harbin, China, and also with University of Lyon, CREATIS, CNRS UMR 5220, Inserm U1044, INSA-Lyon, University of Lyon 1, Villeurbanne, France (email: lujingfeng@hit.edu.cn).

F. Millioz, D. Garcia, S. Salles, and D. Friboulet are with the University of Lyon, CREATIS, CNRS UMR 5220, Inserm U1044, INSA-Lyon, University of Lyon 1, Villeurbanne, France.

W. Liu is with METISLab, School of Instrumentation Science and Engineering, Harbin Institute of Technology, Harbin, China, and also with Sino European School of Technology of Shanghai University, Shanghai, China.

and thus did not consider the possibility of synthetic focusing by compounding multiple PWs. Other studies [22], [24], [32], [33] processed post-beamformed data, possibly taking advantage of several PW transmissions for compounding. Our group previously proposed to reconstruct high-quality images using 3 PW transmissions [21]. A compounding operation was learned using a fully convolutional network. We demonstrated that this approach could produce high-quality images using 3 PWs while preserving the image quality close to that obtained by standard compounding with 31 PWs. In [32], Zhang et al. used a generative adversarial networks (GAN) where the generator was mainly composed of 50 residual blocks, each of them consisting of two convolutional layers. As in [21], the reconstruction was performed from 3 PWs, and the results were found by the authors to be similar to those of [21] in terms of contrast-to-noise-ratio (CNR) and lateral resolution (LR). In [33], Zhou et al. proposed to use a 3-scales CNN with feedback, where the processing of a scale used up to 5 convolutional layers. However, as the proposed CNN did not preserve speckle texture, they resorted to a complex, non-linear, wavelet-based post-processing step to improve the reconstruction. The obtained results were found slightly better than [21] (difference of 1.8 dB in terms of peak signal-to-noise-ratio (PSNR), 0.02 in terms of Structural Similarity Index (SSIM) and 0.08 in terms of Mutual Information (MI)). Khan et al. described a CNN consisting of 27 convolution layers in a contracting path with concatenation [24]. The CNR of the reconstruction obtained with 3 PWs (2.56 dB) was found to be approximately equivalent to that of the standard delay-and-sum (DAS) reconstruction obtained with 11 PWs (2.51 dB). Perdios et al. proposed to use a straightforward U-Net with 5 decomposition levels [22]. The reconstruction was evaluated from simulations and yielded very encouraging results, i.e., a CNR of 17.1 dB and a lateral resolution of 0.38 mm in the far field.

PW and DW have been shown to be suitable for high-frame-rate imaging in echocardiography [37]. They both allow a significant decrease in transmit numbers to examine the region of interest. However, as with focused sequences, echocardiographic PW imaging requires a series of sequential beams that scan the area from edge to edge. In contrast, each DW transmit can cover the entire region of interest. The DW technique has thus two advantages: i) a further gain in image rate; ii) the different regions of the heart are imaged synchronously. As a result, DW-based echocardiography allows simultaneous analysis of intracardiac flow and myocardial motion, as shown by Faurie et al. [38]. Using a DW transmission scheme, it is also possible to cover a wide and deep cardiac view (a four-chamber view) at a very high frame rate (up to 400 cardiac-images/s). Taking advantage of this feature, the motion of the entire myocardium can be deciphered accurately by standard [39] or advanced [40] methods of speckle tracking. The problem of reconstructing images from DW transmissions using deep learning techniques has been addressed only very recently in [35], which was a preliminary version of the present paper, and by Ghani et al. in [36]. The latter study used a six-layer fully convolutional neural network operating patch-wisely from 11 transmissions. As we will show in the

following, the CNN architecture we proposed used only 3 DWs and provided an image quality that competed with the standard compounding of 31 DWs. It is difficult to draw further conclusions from the comparison with [36] to the extent that this study only provided qualitative results (i.e., images).

In this paper, we introduce a novel CNN architecture for high-quality reconstruction for DW imaging using a small number of DW transmissions. To achieve this goal, a unique issue must be resolved. Conventional CNN architectures adopt fixed kernels in one convolutional layer where the same weights are applied over the entire feature map. Such shared-weight architecture contributes to achieving the shift-invariant feature of CNN [41], which is well-adapted to PW images. Nevertheless, due to the geometry of DW, applying deep learning approaches to DW images is specific. Processing RF lines obtained from DW acquisitions implicitly means that the CNN operates in polar coordinates. Maintaining the shift-invariance feature of CNN methods (that was directly applicable in the Cartesian coordinate system associated with PW imaging) requires to adapt the CNN architecture. Inspired by the GoogleNet [42] for image classification problems, we therefore incorporated inception modules to the fully convolutional architecture for the reconstruction of DW images, yielding the so-called IDNet (Inception for DW network). Inception modules employed convolution filters of different sizes for the same input and concatenate all the output for the next layer. As it will be shown in the sequel, inception used in conjunction with maxout activation allowed features from multiple receptive field sizes to be captured, in contrast to conventional CNN architectures where the receptive field size was fixed. Therefore, different image features of different image regions can be learned via multiple convolution filters. This property is demonstrated in more detail in the discussion section.

In summary, the contributions of this work are the following:

- 1) We address the problem of reconstruction for DW imaging using deep learning.
- 2) We introduce a CNN architecture with the inception module to take into account the specific geometry of DW imaging.
- 3) We demonstrate that, using only 3 DWs, our method yielded high-quality images competing with those obtained by compounding with 31 DWs.
- 4) We further show that the proposed method could work at a high frame rate, which made it amenable to real-time reconstruction for DW imaging.

The remainder of this paper is organized as follows: in Section II, the proposed method is described. Section III presents the details of data acquisition and implementation for training. Section IV presents the experiment results that validate the effectiveness of the proposed method. We further discuss the proposed method in Section V and conclude the work in section VI.

II. METHODS

A. Problem formulation

Let X be a tensor that contains the low-quality RF beamformed images of size $m \times w \times h$, where m is the number




Fig. 1. Block diagram of the architecture of the proposed network.

TABLE I
ARCHITECTURE OF PROPOSED NETWORK

block type	feature size	kernel size	padding	number of kernels	activation
	channel \times height \times width	height \times width	height \times width		
inputs	$m \times h \times w$	-	-	-	-
convolution	$64 \times h \times w$	9×3	4×1	256	maxout 4
convolution	$32 \times h \times w$	17×5	8×2	128	maxout 4
convolution	$16 \times h \times w$	33×9	16×4	64	maxout 4
inception	$8 \times h \times w$	41×11	20×5	8	maxout 4
		49×13	24×6	8	maxout 4
		57×15	28×7	8	maxout 4
		65×17	32×8	8	maxout 4
convolution	$1 \times h \times w$	1×1	-	4	maxout 4

of DW transmissions, w is the number of scan lines, and h is the length of each RF signal. Our method aimed at producing one reconstructed RF image of size $w \times h$ using the input X . Standard compounding consists of summing all m DWs to obtain the high-quality image. Considering that there might be useful information which was not exploited by standard compounding, we employed a CNN with trainable parameters Θ to learn the optimal mapping $f(\cdot)$ of $X \rightarrow Y$, where Y was the reference obtained from the standard compounding of n ($n \gg m$) DWs.

B. Network architecture

Fig. 1 is a pictorial description of the proposed IDNet architecture. IDNet was a 2-D convolutional network composed of 5 hidden layers. Two types of basic building modules were employed to build the network. In Fig. 1, each blue block denotes a regular convolutional module followed by a maxout unit activation (white block), and the purple block indicates the inception module. The choices and details related to this architecture are discussed hereunder.

Fully convolution architecture. We excluded the pooling operation from both the regular convolution layers and the inception layer to produce the feature maps with the same dimension. This guaranteed that the spatial information was preserved at the same scale throughout the network, which




Fig. 2. Block diagram of the inception module and the one by one convolution.

was beneficial for maintaining phase in RF signals. The kernel size of each layer was doubled compared to that of its previous layer to achieve the effect of doubling receptive field size, the same as the 2×2 pooling operation.

Inception module. As illustrated in Fig. 2, the inception




Fig. 3. Block diagram of a k pieces maxout unit. A k pieces maxout unit takes the pixel-wise maximum values across the k affine feature maps ($Z_1^{l+1}, Z_2^{l+1}, \dots, Z_k^{l+1}$) produced from the convolution.

module used in our network consisted of four parallel paths. Each path performed convolution of a different kernel size to the same input feature maps. All the paths performed proper padding to ensure their outputs had the same size. The outputs of all paths were stacked along the feature dimension as the final output of the inception module. As the image properties of DW imaging varied along image depth, the multi-size convolution kernels contributed to extract different image features from multiple receptive field sizes. The inception module was only used in the second last layer and followed by the 1×1 convolution layer for two reasons: i) in our network, the number of feature maps in shallow layers were much larger than the deep layers. Incorporating inception modules to shallow layers would drastically increase the number of parameters, since an inception module with four parallel paths would quadruple the parameters of a regular convolution; ii) As deep layers produced high-level features and large receptive sizes, more relevant information could be perceived by employing the inception module with larger kernels in the deeper layers.

One-by-one convolution. The last layer of the network was a 1×1 convolution layer followed by a maxout unit. It served two purposes: i) to reduce the number of output channels and generate the final image; ii) All input pixels of the same position collapsed to one output pixel via the 1×1 convolution in conjunction with the maxout unit, which acted as selecting gates for every element of the inputs. The network was thus trained to select the main elements contributing to form the output element, and thus to learn different image features for different image positions.

Maxout unit. We used maxout units [43] as the activation function for both the regular convolution layers and the inception layer. Maxout units are piecewise-linear convex functions, and a maxout network with more than two maxout units can approximate many popular activation functions and most often outperform them [21], [44]. In a convolution network, a maxout unit takes the pixel-wise maximum values across several affine feature maps to achieve a nonlinear transformation. Fig. 3 is an illustration of the k pieces maxout unit. Given an input X^l to the l th hidden layer of a CNN, the output X^{l+1} after the convolution and the maxout unit is defined as follows:

$$X^{l+1} = \max\{Z_1^{l+1}, \dots, Z_k^{l+1}\} \quad (1)$$

where Z_j^{l+1} ($j \in [1, k]$) is the j th group of feature maps obtained by performing convolution to X^l , and k is the number of maxout pieces.

A more detailed specification of the network architecture, such as the number of feature maps and the size of convolution kernels, is provided in Table I.

III. EXPERIMENT

A. Data Set Acquisition

A Verasonics system research scanner (Vantage 256) equipped with an ATL P4-2 probe (bandwidth: 2-4 MHz, center frequency: 3 MHz, and cycle number: 1.5) was used to acquire the ultrasound data. The samples were acquired by continuously moving the probe on the surface of the imaged objects, at an imaging rate of 50 fps and a packet size of 250 images. Each sample was acquired using 31 DWs with transmission angles ranging between $\pm 30^\circ$, with an incremental step of 2° . For each DW acquisition, the received RF signals were sampled at 12 MHz and beamformed with the conventional DAS method. Each RF beamformed image was of dimension 512×256 , covering a sectorial region of size $11 \text{ cm} \times 90^\circ$. The input images X were composed of a small subset of $m = 3$ DWs ($-30^\circ, 0^\circ$, and 30°), while the reference images Y were the standard compounding of all $n = 31$ DWs. A total of 7000 samples were used in the experiment, and each was made up of low-quality images X and the high-quality corresponding Y . Specifically, 1500 samples were acquired from *in vivo* tissues (femoris muscle, phalanges of fingers, and liver region), and 5500 acquisitions were performed on *in vitro* phantoms.

B. Network Training

5000 samples were randomly selected from the entire data set as the training set, 1000 samples were used as an independent validation set, and the remaining 1000 samples were used as the testing set for evaluation. Learning the reconstruction mapping function $f(\cdot)$ required the estimation of the optimal network parameters Θ by minimizing the loss between the reconstructed images $\hat{Y} = f(X; \Theta)$ and the reference Y . Mean Squared Error (MSE) was used as the loss function:

$$L(\Theta) = \frac{1}{n} \sum_{i=1}^n \|f(X_i; \Theta) - Y_i\|^2, \quad (2)$$

where n is the number of training samples and $i \in [1, n]$ denotes the index of each training sample.

In the training stage, the network weights were initialized with the Xavier initializer [45]. The loss was minimized using mini-batch gradient descent with the Adam optimizer [46], and the batch size was set to 10. The initial learning rate was set to 1×10^{-4} and an early stopping strategy was employed to adjust the learning rate. The learning rate was halved if there had been no decrease in the validation loss for 20 epochs, and 40 epochs without validation loss reduction would end the training. The training was performed using Pytorch [47] library on a NVIDIA Tesla V100 GPU with 32 Gb of memory, resulting in training time of about two days.

C. Evaluation Metrics

To quantitatively show the effectiveness of the proposed network, we used peak-signal-to-noise ratio (PSNR), structure similarity (SSIM), mutual information (MI), contrast ratio (CR), contrast-to-noise ratio (CNR), and lateral resolution (LR) as the evaluation metrics. PSNR and SSIM were used to evaluate the reconstruction quality by comparing it to the reference images (i.e., images obtained through the standard compounding of 31 DWs). The other three quantities (CR, CNR, and LR) were used as specialized ultrasound indices to evaluate the cyst images and the width of the point spread function from point target images.

PSNR is defined as the ratio of the maximum possible power of a signal and the distorting noise which deteriorates the quality of its representation. Given the MSE between the reconstructed image and the reference, the PSNR is calculated as follows:

$$PSNR = 20 \log_{10} \frac{MAX_I}{\sqrt{MSE}}, \quad (3)$$

where MAX_I is the max pixel value of the image.

SSIM measures the similarity between two images, and MI measures mutual dependence between two images. Given the reconstructed image \hat{Y} and the reference Y , the SSIM and MI are calculated as follows:

$$SSIM = \frac{(2\mu_{\hat{Y}}\mu_Y + C_1)(2\sigma_{\hat{Y}Y} + C_2)}{(\mu_{\hat{Y}}^2 + \mu_Y^2 + C_1)(\sigma_{\hat{Y}}^2 + \sigma_Y^2 + C_2)}, \quad (4)$$

$$MI = \sum_{\hat{y}, y} p_{\hat{Y}Y}(\hat{y}, y) \log \frac{p_{\hat{Y}Y}(\hat{y}, y)}{p_{\hat{Y}}(\hat{y})p_Y(y)} \quad (5)$$

where $\mu_{\hat{Y}}$ and μ_Y ($\sigma_{\hat{Y}}^2$ and σ_Y^2) denote the means (variances) of \hat{Y} and Y , $\sigma_{\hat{Y}Y}$ denotes the covariance between \hat{Y} and Y , C_1 and C_2 are two constants that stabilize the division with weak denominator, $p_{\hat{Y}Y}(\hat{y}, y)$ is the joint distribution of \hat{Y} and Y , and $p_{\hat{Y}}(\hat{y})$ and $p_Y(y)$ are the marginal probability distribution of \hat{Y} and Y .

CR and CNR were used to measure the contrast between the object of interest and the surrounding background:

$$CR = -20 \log_{10} \frac{\mu_t}{\mu_b}, \quad (6)$$

$$CNR = 20 \log_{10} \frac{|\mu_t - \mu_b|}{\sqrt{\sigma_t^2 + \sigma_b^2}}, \quad (7)$$

where μ_t and μ_b (σ_t^2 and σ_b^2) denote the means (variances) of the intensity within the target region and the background.

LR was used to assess the width of the point spread function from point target images. The full width at half maximum was used in this work to estimate the LR.

In the testing phase, the PSNR, SSIM, and MI were computed from the full set of testing samples. The CR and CNR were measured on two anechoic regions (in the near field at 40 mm depth and the far field at 120 mm depth) of an image obtained from the Gammex phantom. The LR was measured on 0.1 mm Nylon monofilaments (in the near field at 20 mm and 40 mm depth, the middle field at 60 mm and 80 mm depth, and the far field at 90 mm and 100 mm depth) of an image obtained from the CIRS phantom.

TABLE II
ARCHITECTURES OF THE INCEPTION LAYERS OF IDNET-2, IDNET-4, AND IDNET-8

method	number of kernel types	kernel size height \times width	number of kernels
IDNet-2	2	49 \times 13	16
		65 \times 17	16
IDNet-4	4	41 \times 11	8
		49 \times 13	8
		57 \times 15	8
		65 \times 17	8
		37 \times 11	4
		41 \times 11	4
IDNet-8	8	45 \times 13	4
		49 \times 13	4
		53 \times 15	4
		57 \times 15	4
		61 \times 17	4
		65 \times 17	4

D. Inception and activation layer parameters

In the proposed network, we used the inception module to exploit features with multiple receptive fields of the images. To demonstrate the effect of the inception module of our network, models with different inception layers were trained with the same training data and implementation. Each model employed an inception layer with different convolution kernels. For the sake of clarity, these models were named IDNet-2, IDNet-4, and IDNet-8, whose inception structures are shown in Table II. For a fair comparison, the other components of the network shared the same architecture. Besides, to verify the effectiveness of maxout unit activation, we conducted another experiment which consisted in replacing the maxout unit with the ReLU activation (IDNet-ReLU).

E. Comparison Methods

The proposed method was compared with three methods for the evaluation of the reconstruction quality.

1) Standard compounding method [10]: To assess the improvement of our method over standard compounding method, same input images were used to obtain the compounding reconstruction.

2) CNN of Gasse et al. [21]: IDNet and Gasse's CNN shared the same architecture in the first three convolutional layers, i.e., fully convolutional layers followed by maxout activation without spatial pooling. The difference was that we employed the inception layer composed of four parallel multi-scale convolutions, rather than the fixed convolution. Each path of the inception layer had 8 channels, and stacking of all channels produced 32 output channels, the same as the last layer in Gasse's network.

3) U-Net [22]: U-Net is a typical encoder-decoder structure consisting of a symmetric downsampling and upsampling path. We experimentally observed that the direct use of the




Fig. 4. Example of B-mode images reconstructed from (a) IDNet-2, (b) IDNet-4, (c) IDNet-8, (d) IDNet-ReLU, and standard compounding of 31 DWs (reference).

TABLE III
EVALUATION METRICS OF IDNET AND COMPOUNDING METHOD

model	PSNR [dB]	SSIM	MI	CR [dB]		CNR [dB]		LR [mm]		
				near field	far field	near field	far field	near field	middle field	far field
IDNet-2	31.06 ± 1.49	0.92 ± 0.06	0.81 ± 0.19	19.20	13.71	7.25	3.94	0.95	1.67	2.47
IDNet-4	31.13 ± 1.47	0.93 ± 0.06	0.82 ± 0.20	19.54	14.95	7.63	5.21	0.90	1.64	2.35
IDNet-8	31.07 ± 1.50	0.93 ± 0.06	0.83 ± 0.20	19.05	13.24	7.41	4.83	0.96	1.61	2.44
IDNet-ReLU	30.96 ± 1.50	0.92 ± 0.07	0.76 ± 0.20	17.04	11.86	5.37	2.79	1.01	1.70	2.48
Reference	-	-	-	19.97	15.45	7.71	5.33	0.94	1.63	2.42

architecture proposed in [22] gave poor results. To obtain a better comparison between the U-Net and our network, the convolutional sizes of the U-Net were modified to fit the data used in our experiment. The sequential operation [3 × 3 convolution, ReLU, 3 × 3 convolution, ReLU] used in [22] was replaced by the [3 × 1 convolution, ReLU, 7 × 3 convolution, ReLU] operation, resulting in the same receptive field size as the 9 × 3 convolution of our network.

All compared networks were trained using the dataset and settings described in the previous section.

IV. RESULTS

A. Performance of the Proposed Network

In this section, we compare the results of four IDNet models (IDNet-2, IDNet-4, IDNet-8, and IDNet-ReLU) to obtain the optimal performance and verify the contributions of the components of the proposed network. Fig. 4 displays the images reconstructed from these IDNet models and the reference. Table III shows the quantitative results reached by the four models. From Fig. 4d, IDNet-ReLU yielded the worst

TABLE IV
EVALUATION METRICS OF IDNET AND COMPOUNDING METHOD

model	PSNR [dB]	SSIM	MI
IDNet (3 DWs)	31.13 ± 1.47	0.93 ± 0.06	0.82 ± 0.20
Compounding (3 DWs)	29.24 ± 1.57	0.83 ± 0.15	0.51 ± 0.16

image quality compared with the models that used maxout activation. It can be observed that there was more amplitude loss in the lateral region. The contrast of the cysts appeared to be decreased; in particular, the cyst in the far field was barely visible. Among the models using the maxout unit, IDNet-2 and IDNet-8 obtained a comparable SSIM and MI compared to IDNet-4, while IDNet-4 performed best in PSNR, CR, CNR, and LR. Therefore IDNet-4 was used as the best IDNet model in the next section, where it was compared with the standard compounding method and other CNN architectures.




Fig. 5. B-mode images obtained using IDNet and standard compounding. Top to bottom: *in vivo* tissues from the femoris muscle (a, b, c); *in vitro* tissues from the Gammex phantom (d, e, f); and *in vitro* tissues from the CIRS phantom (g, h, i). Left to right: compounding of 3 DWs; reconstruction of IDNet; and compounding of 31 DWs (reference).




Fig. 6. CR (a) and CNR (b) reached by IDNet and standard compounding. IDNet used a constant number of three DWs, whereas the results of standard compounding are given for an increasing number of DWs. The results are given in the near (blue lines and curves) and far (orange line and curves) fields.

TABLE V
EVALUATION METRICS OF IDNET, GASSE'S CNN [21], AND U-NET [22]

model	PSNR [dB]	SSIM	MI	CR [dB]		CNR [dB]		LR [mm]		
				near field	far field	near field	far field	near field	middle field	far field
Gasse et al.	31.06 ± 1.46	0.93 ± 0.06	0.81 ± 0.20	16.47	12.27	5.95	3.65	0.98	1.69	2.54
U-Net	31.15 ± 1.46	0.94 ± 0.05	0.82 ± 0.23	14.97	12.72	5.29	4.00	1.01	1.67	2.47
IDNet	31.13 ± 1.47	0.93 ± 0.06	0.82 ± 0.20	19.54	14.95	7.63	5.21	0.90	1.64	2.35




Fig. 7. LR reached by IDNet and standard compounding. IDNet used a constant number of three DWs, whereas the results of standard compounding are given for an increasing number of DWs. The results are given in the near (blue line and curve), middle (red line and curve), and far (orange line and curve) fields.

B. Comparison with compounding method

The comparison of the visual quality between the reconstruction of IDNet and the coherent DW compounding method is shown in Fig. 5. It appears that the quality of the reconstruction of IDNet using only three DWs (Fig. 5b, 5e, and 5h) was improved compared to the images obtained from the standard compounding of the same three DWs (Fig. 5a, 5d, and 5g). In particular, the anatomical structures in Fig. 5b was enhanced compared to Fig. 5a. However, it can be observed that the reconstruction of IDNet appeared to have an overall loss in amplitude (the background of the scatterers in Fig. 5e and the lateral region in Fig. 5h), as compared to the reference (Fig. 5f and 5i).

To quantitatively assess the improvement, we report in Table IV the PSNR, SSIM, and MI reached by IDNet and standard compounding of 3 DWs, using the compounding of 31 DWs as the reference. The reconstruction of IDNet showed a gain of 1.89 dB in PSNR, 0.1 in SSIM, and 0.31 in MI, as compared to the compounding of the same three DWs. In

TABLE VI
NUMBER OF PARAMETERS, COMPOUNDING TIME, AND ATTAINABLE FRAME RATE FOR IDNET, GASSE'S CNN, AND U-NET.

model	nb. of parameters [million]	compounding	attainable
		time [ms]	frame rate [fps]
Gasse et al.	1.9	0.48 ± 0.03	2080
U-Net	52.7	2.38 ± 0.06	420
IDNet	1.7	0.75 ± 0.03	1330

Fig. 6 and Fig. 7, we show the CR, CNR, and LR reached by IDNet and standard compounding. The CR and CNR were measured on the anechoic regions in the phantom (as shown in Fig. 5d, 5e and 5f). The LR was measured on the isolated scatterers (as shown in Fig. 5g, 5h and 5i). IDNet used a constant number of three DWs, whereas the results of standard compounding are given for an increasing number of DWs. From Fig. 6, it can be observed that IDNet reached a CR and CNR equivalent to those of the compounding of 21 DWs and 23 DWs, respectively, in both near and far fields. In Fig. 7, the LR evolution of standard compounding was consistent with the experimental observation from Zhang et al. [32]: LR value quickly dropped with two or three DWs and then reached a plateau with more DWs. IDNet reached an LR equivalent to that of the compounding of 3 DWs in the near and far fields, and the compounding of 21 DWs in the middle field.

C. Comparison with other networks

Using the same input (DW images of -30° , 0° , and 30°), the reconstructed images from the IDNet, Gasse's CNN, and U-Net are displayed in Fig. 8. From the figure, it appears that the proposed IDNet produced a better contrast than the other networks, particularly for the cyst in the far field. The quantitative evaluation results of the three models are shown in Table V. The SSIM and MI indices of the three models were rather close, while the PSNR of U-Net was slightly higher than those of IDNet and Gasse's CNN. It can be also observed that the U-Net delivered the worst results in the near field, in terms of CR, CNR, and LR, while producing marginally better values in the middle and far fields, as compared with Gasse's CNN. IDNet produced the best results in terms of CR, CNR, and LR, whatever the depth in the image.

D. Computational complexity and speed

To evaluate the interest of CNN-based methods in terms of attainable frame rates, all the image formation steps (i.e.,




Fig. 8. Example of B-mode images reconstructed from (a) Gasse’s CNN [21], (b) U-Net [22], (c) IDNet, and (d) standard compounding of 31 DWs (reference).

including acquisition, beamforming, and compounding) had to be taken into consideration, and to be compared to the steps involved when using standard compounding. Standard compounding implied using 31 DWs, which yielded an acquisition time of 4.96 ms (travel time for 31 DWs for a tissue depth of 120 mm), a beamforming time of 4.46 ms, and a compounding time of 0.02 ms (Please note that for a fair comparison with CNN-based imaging, the computations were performed on GPU). Therefore, the time bottleneck in this case was the acquisition time. The corresponding attainable frame rate was thus 202 fps. The interest of the CNN-based compounding methods was to rely on the acquisition of only 3 DWs, corresponding to an acquisition time of 0.48 ms and a beamforming time of 0.43 ms, so the time bottleneck was the reconstruction time, as shown in Table VI. Table VI gives for each CNN-based method the number of parameters and the corresponding compounding time (i.e., inference time) and attainable frame rate. Table VI indicates that the U-Net had 52.7 million parameters, which was more than those of IDNet (1.7 million) and Gasse’s CNN (1.9 million). In terms of attainable frame rate, the IDNet reached 1330 fps, which was slower than Gasse’s network (2080 fps) and faster than the U-Net (420 fps).

V. DISCUSSION

A. High-quality reconstruction for DW imaging using CNN

In this study, a methodology for the reconstruction of high-quality DW images with deep learning was proposed. We formulated the reconstruction problem as a mapping problem between the low-quality and high-quality images, which was

solved by training the proposed CNN architecture (IDNet). Although a large number of samples (a total of 7000 samples used for the training, validation, and testing) and long training time (about two days) were required, once the training was completed, the model could be applied at a high frame rate (1330 fps). The experimental evaluation demonstrated that using 3 DW emissions, the proposed method was able to produce images with comparable quality in terms of contrast and resolution as those obtained from standard compounding of 31 DWs (a ten-fold acceleration factor).

B. Inception module to fit sectorial images

CNNs methods have witnessed a gradual increase in the network depth, corresponding to improvements in various challenges. However, from the experimental results of section IV-C, directly employing conventional CNN architectures for PW imaging (Gasse’s CNN using 4 layers or the U-Net using more than 20 layers) yielded a lower performance in terms of contrast and resolution, as shown in Table V. These architectures used convolutions whose shared weights were applied to the entire images or feature maps. This contributed to the shift-invariance property that was applicable in the Cartesian coordinate system associated with PW imaging. Due to the geometry of DW imaging, using the RF channel data obtained from DW acquisitions as the CNN input implied that the CNN operated in polar coordinates. Thus maintaining the shift-invariance feature for sectorial images required a spatially varying convolution kernel in polar coordinates. To this purpose, we employed the inception module used in conjunction with maxout activation, which contributed to the spatially




Fig. 9. Spatial distribution of the inception kernels selected by the last layer of the network. Each pixel indicates which feature map contributed most to generate the output element (i.e., the feature map associated to the maximum coefficient of the 1×1 convolution).

varying property of the IDNet. In Section IV, the performance of the models using different inception modules and their improvement over conventional networks were investigated. With only two convolution scales in the inception layer, the IDNet-2 model was able to produce equivalent or better results, as compared to Gasse’s CNN and U-Net. As the diversity of convolution scales evolved from IDNet-2 to IDNet-4, further improvement of the image quality was obtained. Nevertheless most of the evaluation indices (except the MI and LR in the middle field) of IDNet-8 slightly declined compared with IDNet-4. The performance deterioration of IDNet-8 might be caused by the decreased number of feature maps of each scale in the inception layer. IDNet-4 thus obtained a better compromise between the convolution diversity and feature number of each scale, given a total number of feature maps.

To study the spatially varying property of the proposed method, we analyzed the activation map of IDNet-4. Fig. 9 displays the spatial distribution of the inception kernels selected by the last layer of the network (i.e., the 1×1 convolution / maxout layer). Each pixel in the figure indicates which feature map contributed most to generate the output element (i.e., the feature map associated to the maximum coefficient of the 1×1 convolution). Fig. 10 further summarizes this distribution by showing the contribution of each kernel for each image depth (expressed as a percentage). It can be observed that the elements generated from the larger convolutions (57×15 and 65×17) tended to contribute more in the near field, while the elements generated from smaller convolutions (41×11 and 49×13) tended to contribute more in the far field, which illustrates the depth-dependence of IDNet.

VI. CONCLUSION

In this paper, we presented a CNN architecture for the reconstruction of DW imaging. The proposed method aimed at learning a compounding operator to reconstruct high-quality images using a small number of DWs. We demonstrated that the integration of the inception module followed by the maxout activation allowed to exploit information from DW images




Fig. 10. Contribution of each inception kernel as a function of depth, expressed as a percentage.

more effectively. The experimental results demonstrated the effectiveness of the proposed method, yielding an image quality equivalent to that obtained with standard compounding of 31 DWs, which provided a ten-fold increase in the acceleration factor.

ACKNOWLEDGMENT

This work was performed within the framework of the LABEX PRIMES (ANR-11-LABX-0063) of Universite de Lyon, within the program "Investissements d’Avenir" (ANR-11-IDEX-0007) operated by the French National Research Agency (ANR). Financial support from China Scholarship Council (Grant No. 201806120175) is gratefully acknowledged.

REFERENCES

- [1] M. Couade, M. Pernot, E. Messas, A. Bel, M. Ba, A. Hagege, M. Fink, and M. Tanter, "In vivo quantitative mapping of myocardial stiffening and transmural anisotropy during the cardiac cycle," *IEEE transactions on medical imaging*, vol. 30, no. 2, pp. 295–305, 2010.
- [2] J. Provost, W.-N. Lee, K. Fujikura, and E. E. Konofagou, "Electromechanical wave imaging of normal and ischemic hearts in vivo," *IEEE transactions on medical imaging*, vol. 29, no. 3, pp. 625–635, 2009.
- [3] C. Papadacci, M. Pernot, M. Couade, M. Fink, and M. Tanter, "High-contrast ultrafast imaging of the heart," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 61, no. 2, pp. 288–301, 2014.
- [4] D. P. Shattuck, M. D. Weinschenker, S. W. Smith, and O. T. von Ramm, "Explososcan: A parallel processing technique for high speed ultrasound imaging with linear phased arrays," *The Journal of the Acoustical Society of America*, vol. 75, no. 4, pp. 1273–1282, 1984.
- [5] R. Mallart and M. Fink, "Improved imaging rate through simultaneous transmission of several ultrasound beams," in *New Developments in Ultrasonic Transducers and Transducer Systems*, vol. 1733. International Society for Optics and Photonics, 1992, pp. 120–130.
- [6] L. Sandrin, S. Catheline, M. Tanter, X. Hennequin, and M. Fink, "Time-resolved pulsed elastography with ultrafast ultrasonic imaging," *Ultrasonic imaging*, vol. 21, no. 4, pp. 259–272, 1999.
- [7] H. Hasegawa and H. Kanai, "High-frame-rate echocardiography using diverging transmit beams and parallel receive beamforming," *Journal of medical ultrasonics*, vol. 38, no. 3, pp. 129–140, 2011.
- [8] J. Porée, D. Posada, A. Hodzic, F. Tournoux, G. Cloutier, and D. Garcia, "High-frame-rate echocardiography using coherent compounding with doppler-based motion-compensation," *IEEE transactions on medical imaging*, vol. 35, no. 7, pp. 1647–1657, 2016.

- [9] G. Montaldo, M. Tanter, J. Bercoff, N. Benech, and M. Fink, "Coherent plane-wave compounding for very high frame rate ultrasonography and transient elastography," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 56, no. 3, pp. 489–506, 2009.
- [10] M. Tanter and M. Fink, "Ultrafast imaging in biomedical ultrasound," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 61, no. 1, pp. 102–119, 2014.
- [11] Y. Lecun, Y. Bengio, and G. Hinton, "Deep learning," *Nature*, vol. 521, no. 7553, pp. 436–444, 2015.
- [12] K. H. Jin, M. T. McCann, E. Froustey, and M. Unser, "Deep convolutional neural network for inverse problems in imaging," *IEEE Transactions on Image Processing*, vol. 26, no. 9, pp. 4509–4522, 2017.
- [13] J. M. Wolterink, T. Leiner, M. A. Viergever, and I. Išgum, "Generative adversarial networks for noise reduction in low-dose ct," *IEEE transactions on medical imaging*, vol. 36, no. 12, pp. 2536–2545, 2017.
- [14] Y. Han and J. C. Ye, "Framing u-net via deep convolutional framelets: Application to sparse-view ct," *IEEE transactions on medical imaging*, vol. 37, no. 6, pp. 1418–1429, 2018.
- [15] E. Kang, W. Chang, J. Yoo, and J. C. Ye, "Deep convolutional framelet denosing for low-dose ct via wavelet residual network," *IEEE transactions on medical imaging*, vol. 37, no. 6, pp. 1358–1369, 2018.
- [16] S. Wang, Z. Su, L. Ying, X. Peng, S. Zhu, F. Liang, D. Feng, and D. Liang, "Accelerating magnetic resonance imaging via deep learning," in *2016 IEEE 13th International Symposium on Biomedical Imaging (ISBI)*. IEEE, 2016, pp. 514–517.
- [17] J. Schlemper, J. Caballero, J. V. Hajnal, A. N. Price, and D. Rueckert, "A deep cascade of convolutional neural networks for dynamic mr image reconstruction," *IEEE transactions on Medical Imaging*, vol. 37, no. 2, pp. 491–503, 2017.
- [18] K. Hammernik, T. Klatzer, E. Kobler, M. P. Recht, D. K. Sodickson, T. Pock, and F. Knoll, "Learning a variational network for reconstruction of accelerated mri data," *Magnetic resonance in medicine*, vol. 79, no. 6, pp. 3055–3071, 2018.
- [19] D. Lee, J. Yoo, S. Tak, and J. C. Ye, "Deep residual learning for accelerated mri using magnitude and phase networks," *IEEE Transactions on Biomedical Engineering*, vol. 65, no. 9, pp. 1985–1995, 2018.
- [20] K. He, X. Zhang, S. Ren, and J. Sun, "Deep residual learning for image recognition," in *Proceedings of the IEEE conference on computer vision and pattern recognition*, 2016, pp. 770–778.
- [21] M. Gasse, F. Millioz, E. Roux, D. Garcia, H. Liebgott, and D. Friboulet, "High-quality plane wave compounding using convolutional neural networks," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 64, no. 10, pp. 1637–1639, 2017.
- [22] D. Perdios, M. Vonlanthen, A. Besson, F. Martinez, M. Arditi, and J.-P. Thiran, "Deep convolutional neural network for ultrasound image enhancement," in *2018 IEEE International Ultrasonics Symposium (IUS)*. IEEE, 2018, pp. 1–4.
- [23] D. Hyun, L. L. Brickson, K. T. Looby, and J. J. Dahl, "Beamforming and speckle reduction using neural networks," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 66, no. 5, pp. 898–910, 2019.
- [24] S. Khan, J. Huh, and J. C. Ye, "Deep learning-based universal beamformer for ultrasound imaging," in *Medical Image Computing and Computer Assisted Intervention - MICCAI 2019*. Springer International Publishing, 2019, pp. 619–627.
- [25] A. C. Luchies and B. C. Byram, "Deep neural networks for ultrasound beamforming," *IEEE transactions on medical imaging*, vol. 37, no. 9, pp. 2010–2021, 2018.
- [26] B. Luijten, R. Cohen, F. J. de Bruijn, H. A. Schmeitz, M. Mischi, Y. C. Eldar, and R. J. van Sloun, "Deep learning for fast adaptive beamforming," in *ICASSP 2019-2019 IEEE International Conference on Acoustics, Speech and Signal Processing (ICASSP)*. IEEE, 2019, pp. 1333–1337.
- [27] A. A. Nair, T. D. Tran, A. Reiter, and M. A. L. Bell, "A deep learning based alternative to beamforming ultrasound images," in *2018 IEEE International Conference on Acoustics, Speech and Signal Processing (ICASSP)*. IEEE, 2018, pp. 3359–3363.
- [28] O. Senouf, S. Vedula, G. Zurakhov, A. Bronstein, M. Zibulevsky, O. Michailovich, D. Adam, and D. Blondheim, "High frame-rate cardiac ultrasound imaging with deep learning," in *International Conference on Medical Image Computing and Computer-Assisted Intervention*. Springer, 2018, pp. 126–134.
- [29] S. Vedula, G. Senouf, A. Bronstein, O. Michailovich, and M. Zibulevsky, "Learning beamforming in ultrasound imaging," in *International Conference on Medical Imaging with Deep Learning*, 2019, pp. 493–511.
- [30] S. Vedula, O. Senouf, G. Zurakhov, A. Bronstein, M. Zibulevsky, O. Michailovich, D. Adam, and D. Gaitini, "High quality ultrasonic multi-line transmission through deep learning," in *International Workshop on Machine Learning for Medical Image Reconstruction*. Springer, 2018, pp. 147–155.
- [31] Y. H. Yoon, S. Khan, J. Huh, and J. C. Ye, "Efficient b-mode ultrasound image reconstruction from sub-sampled rf data using deep learning," *IEEE transactions on medical imaging*, vol. 38, no. 2, pp. 325–336, 2018.
- [32] X. Zhang, J. Li, Q. He, H. Zhang, and J. Luo, "High-quality reconstruction of plane-wave imaging using generative adversarial network," in *2018 IEEE International Ultrasonics Symposium (IUS)*. IEEE, 2018, pp. 1–4.
- [33] Z. Zhou, Y. Wang, J. Yu, Y. Guo, W. Guo, and Y. Qi, "High spatial-temporal resolution reconstruction of plane-wave ultrasound images with a multichannel multiscale convolutional neural network," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 65, no. 11, pp. 1983–1996, 2018.
- [34] R. Zhuang and J. Chen, "Deep learning based minimum variance beamforming for ultrasound imaging," in *Smart Ultrasound Imaging and Perinatal, Preterm and Paediatric Image Analysis*, Q. Wang, A. Gomez, J. Hutter, K. McLeod, V. Zimmer, O. Zettinig, R. Licandro, E. Robinson, D. Christiaens, E. A. Turk, and A. Melbourne, Eds. Springer International Publishing, 2019, pp. 83–91.
- [35] J. Lu, F. Millioz, D. Garcia, S. Salles, and D. Friboulet, "Fast diverging wave imaging using deep-learning-based compounding," in *2019 IEEE International Ultrasonics Symposium (IUS)*. IEEE, 2019, pp. 2341–2344.
- [36] M. U. Ghani, F. C. Meral, F. Vignon, and J.-I. Robert, "High frame-rate ultrasound imaging using deep learning beamforming," in *2019 IEEE International Ultrasonics Symposium (IUS)*. IEEE, 2019, pp. 295–298.
- [37] M. Cikes, L. Tong, G. R. Sutherland, and J. D'hooge, "Ultrafast cardiac ultrasound imaging: technical principles, applications, and clinical benefits," *JACC: Cardiovascular Imaging*, vol. 7, no. 8, pp. 812–823, 2014.
- [38] J. Faurie, M. Baudet, K. C. Assi, D. Auger, G. Gilbert, F. Tournoux, and D. Garcia, "Intracardiac vortex dynamics by high-frame-rate doppler vortography-in vivo comparison with vector flow mapping and 4-d flow mri," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 64, no. 2, pp. 424–432, 2016.
- [39] P. Joos, J. Porée, H. Liebgott, D. Vray, M. Baudet, J. Faurie, F. Tournoux, G. Cloutier, B. Nicolas, and D. Garcia, "High-frame-rate speckle-tracking echocardiography," *IEEE transactions on ultrasonics, ferroelectrics, and frequency control*, vol. 65, no. 5, pp. 720–728, 2018.
- [40] J. Porée, M. Baudet, F. Tournoux, G. Cloutier, and D. Garcia, "A dual tissue-doppler optical-flow method for speckle tracking echocardiography at high frame rate," *IEEE transactions on medical imaging*, vol. 37, no. 9, pp. 2022–2032, 2018.
- [41] W. Zhang, K. Itoh, J. Tanida, and Y. Ichioka, "Parallel distributed processing model with local space-invariant interconnections and its optical architecture," *Applied Optics*, vol. 29, no. 32, pp. 4790–4797, 1990.
- [42] C. Szegedy, W. Liu, Y. Jia, P. Sermanet, S. Reed, D. Anguelov, D. Erhan, V. Vanhoucke, and A. Rabinovich, "Going deeper with convolutions," in *The IEEE Conference on Computer Vision and Pattern Recognition (CVPR)*, June 2015, pp. 1–9.
- [43] I. J. Goodfellow, D. Warde-Farley, M. Mirza, A. Courville, and Y. Bengio, "Maxout networks," *Computer Science*, pp. 1319–1327, 2013.
- [44] H. Zhao, F. Liu, and L. Li, "Improving deep convolutional neural networks with mixed maxout units," *PloS one*, vol. 12, no. 7, 2017.
- [45] X. Glorot and Y. Bengio, "Understanding the difficulty of training deep feedforward neural networks," in *Proceedings of the thirteenth international conference on artificial intelligence and statistics*, 2010, pp. 249–256.
- [46] D. P. Kingma and J. Ba, "Adam: A method for stochastic optimization," in *Proceeding of International Conference on Learning Represent*, 2015, pp. 1–41.
- [47] A. Paszke, S. Gross, S. Chintala, G. Chanan, E. Yang, Z. DeVito, Z. Lin, A. Desmaison, L. Antiga, and A. Lerer, "Automatic differentiation in pytorch," in *Advances in Neural Information Processing Systems Workshop*, 2017.