



## **Combining Gradient Ascent Search and Support Vector Machines for Effective Autofocus of a Field Emission – Scanning Electron Microscope.**

Souunkalo Dembele, Olivier Lehmann, Kamal Medjaher, Naresh Marturi, Nadine Piat

### **► To cite this version:**

Souunkalo Dembele, Olivier Lehmann, Kamal Medjaher, Naresh Marturi, Nadine Piat. Combining Gradient Ascent Search and Support Vector Machines for Effective Autofocus of a Field Emission – Scanning Electron Microscope.. Journal of Microscopy, 2016, 264 (1), pp.79 - 87. <hal-02946362>

**HAL Id: hal-02946362**

**<https://hal.science/hal-02946362v1>**

Submitted on 23 Sep 2020

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

# Combining Gradient Ascent Search and Support Vector Machines for Effective Autofocus of a Field Emission – Scanning Electron Microscope

Soukalo Dembélé <sup>a,b</sup>, Olivier Lehmann <sup>b</sup>, Kamal Medjaher <sup>b</sup>, Naresh Marturi <sup>c</sup> and Nadine Piat <sup>b</sup>

<sup>a</sup> Corresponding author: Soukalo Dembélé

E-mail. [Soukalo.dembele@femto-st.fr](mailto:Soukalo.dembele@femto-st.fr)

Tel. +33 381 40 27 91

Fax. +33 381 40 28 02

Postal address.

FEMTO-ST Institute, AS2M department, Univ. Bourgogne Franche-Comté, Univ. de Franche-Comté / CNRS / ENSMM

25 rue Savary, 25000 Besançon, France

<sup>b</sup> FEMTO-ST Institute, AS2M department, Univ. Bourgogne Franche-Comté, Univ. de Franche-Comté / CNRS / ENSMM

25 rue Savary, 25000 Besançon, France

<sup>c</sup> KUKA Robotics

Great Western Street, Wednesbury WS10 7LL, United-Kingdom

## Summary

Autofocus is an important issue in electron microscopy, particularly at high magnification. It consists in searching for sharp image of a specimen, i.e. corresponding to the peak of focus. The paper presents a machine learning solution to this issue. From 7 focus measures, support vector machines (SVM) fitting is used to compute the peak with an initial guess obtained from a gradient ascent search, i.e. search in the direction of higher gradient of focus. The solution is implemented on a Carl Zeiss Auriga FE-SEM with a 3 benchmark specimen and magnification ranging from x300 to x160000. Based on regularized non-linear least squares optimization, the solution overtakes the literature non-regularized search and Fibonacci search methods: accuracy improvement ranges from 1.25 to 8 times, fidelity improvement ranges from 1.6 to 28 times, and speed improvement ranges from 1.5 to 4 times. Moreover the solution is practical by requiring only an off-line easy automatic train with cross-validation of the SVM.

## Key words

Scanning electron microscopy, autofocus, gradient ascent search, machine learning, support vector machines regression, normalized variance.

## I. INTRODUCTION

The scanning electron microscope (SEM) and the transmission electron microscope (TEM) are reference instruments for the microanalysis of materials: they are widely used in material as well as life sciences. The autofocus brings to them a significant ease of use especially at high magnifications, its study started two decades ago and led to significant results. It is seen as a problem of optimization: assuming an unimodal model of the focus with respect to the focal length, it comes to search for the position (focal length) of the maximum of focus (peak). Two types of autofocus can be distinguished.

In the first type, people assume an explicit model and use a fitting method to estimate the peak. Nicolls, de Jager & Sewell (1997) published one of the first work of this type. Assuming a gaussian model, they computed the peak from two linear equations derived from the ratio of 3 focus

measurements (first, intermediate, last). They also showed that the result was better when the intermediate position is close to the peak. Clearly, for practical issue, this method requires an initial estimation of the peak. Moreover, based on approximated relations, it leads to low accurate results. Rudnaya, ter Morsche, Maubach, & Mattheij. (2012) assumed a quadratic model and used linear least squares fitting to compute the peak from at least 3 measurements around its initial estimation which was obtained manually. Unfortunately standard least squares optimization is known to be sensitive to outliers that come with noise. Nishi, Moriyama, Yoshida, Kajimura, Mogaki, Ozawa & Isakozawa. (2013) assumed a quasi-gaussian model and used non-regularized non-linear least squares fitting to estimate the peak from 5 measurements. Initial peak and model parameter guesses were estimated manually. The method is similar to our method, unfortunately it is not robust to outliers, particularly because of the little number of measurements (sparse data): it fails to give accurate value as soon as one measurement is far from the model. However we decided to use this method as a benchmark for comparison with our solution.

In the second type of autofocus, people search for the peak of focus without considering any explicit model. In order to have a reasonable speed, they adopt the coarse-to-fine approach: the peak is searched with lower to higher accuracy. Ong, Phang & Thong (1998a, 1998b) modified hill climbing search by progressively decreasing the sweeping step from far to close to the peak. Batten (2000) implemented a coarse-to-fine approach based on Fibonacci search: coarse and fine searches were performed at x200 and final magnification (x410; x970; x1350; x26000) respectively. Naresh, Tamadazte, Dembélé & Piat. (2013) published a gradient ascent search method that progressively drives the SEM directly to the peak. These three methods are fast enough but they lack accuracy which is particularly sensitive at high magnifications as in the case of our experiments, up to x160000.

This review cannot be ended without considering autofocus in the cases of photon microscope and digital camera. He, Zhou & Hong (2003) modified the standard hill climbing search to a coarse-to-fine search: adaptive big sweeping steps were used to find coarsely the peak where as a constant small step was used to determine it finely. This is somewhat similar to the approach of Ong, Phang & Thong (1998a, 1998b) except that they used square of gradient as focus measure instead of auto-correlation. Wu, Wang & Zhou (2012) used the standard hill climbing search to get the initial estimation of the peak. They assumed an exponential model of the focus left and right of that point (quasi-Laplace model), and used linear least squares regression to find the both models. The peak was then computed accurately by the intersection of the both exponentials. The weaknesses of the method are the lack of robustness to outliers, however it is used as benchmark for comparison with our method. Mir, Xu, Chen & van Beek (2015), unlike above publications, considered the case of multimodal model and implemented a machine learning-based coarse-to-fine search to find all the peaks, typically the foreground and background. This method does not fall with the scope of this study which is the unimodality of the model, indeed our specimens are such they can be imaged at one time by the SEM.

This paper investigates autofocus of the scanning electron microscope, precisely a Zeiss Auriga FE-SEM. The developed solution combines the advantage of coarse-to-fine search, i.e. speed, with those of machine learning fitting search (support vector machines or SVM), i.e. accuracy and fidelity. Gradient ascent search, a native coarse-to-fine search, known for its speed, is chosen to find the initial guess of the focus maximum position. It overtakes the methods presented above (Ong, Phang & Thong (1998a, 1998b), Batten (2000), Wu, Wang & Zhou (2012)) in term of speed, and of starting point that can be far from the peak. Support vector machines (SVM) are chosen for fitting search. Its advantages include accuracy like least squares (Nicolls, de Jager & Sewell (1997), Rudnaya, ter Morsche, Maubach, & Mattheij. (2012), Wu, Wang & Zhou (2012), Nishi, Moriyama, Yoshida, Kajimura, Mogaki, Ozawa & Isakozawa. (2013)) but it overtakes the latter in term of robustness to outliers coming from noise. Indeed, it uses a regularization parameter  $C$  that limits the values of the model parameters and more particularly the slack parameter  $\epsilon$  that tunes the acceptable variations of the data (Bishop (2006)). It is easier to implement than the method of Mir, Xu, Chen & van Beek

(2015) since it is just a regression and only requires an off-line automatic training to determine its parameters. Finally, the solution is an efficient autofocus method that works well at low and high magnifications.

## II. PROBLEM STATEMENT

The setup consists of a Zeiss Auriga FE-SEM (Oberkochen, Germany) along with its computer (SEM computer) and a remote computer (Figure 1). The SEM features Schottky field emission Gemini electron column and two SE detectors (Everhart-Thornley in the chamber and Inlens in the column). The remote computer runs C++ client applications while the SEM computer runs C# server applications. The autofocus is implemented as a client application and is based on OpenCV (Bradski, G. (2000)) and particularly the Machine Learning Library which includes the Support Vector Machines implementation of Chang & Lin (2001).

The focusing in this SEM consists of the direct control of the focal length  $F$ . The literature (Batten (2000), Rudnaya, Mattheij & Maubach (2010), Marturi, Tamadazte, Dembélé & Piat (2013) demonstrates the superiority of variance over others focus measures (gradient, auto-correlation, ...) in terms of speed, accuracy and fidelity. Then we chose normalized variance as focus measure (Figure 2). Let  $I(u, v)$  be the image intensity at the pixel  $(u, v)$ , the focus measure  $S$  may be written:

$$S = \frac{1}{WH\mu} \sum_w \sum_H (I(u, v) - \mu)^2$$

(eq. 1)

with  $W$ ,  $H$  and  $\mu$  the width, height and mean intensity of the image  $I$ , respectively.

Assuming the specimen is installed inside the SEM at an unknown position with the settings (brightness, contrast, astigmatism, scan speed) defined, the problem is how to drive the instrument to the peak of focus. The expected properties are accuracy, robustness to outliers coming with noise and flexibility.

Three benchmark specimens were used to validate the solution: a gold-coated gripper over 20  $\mu\text{m}$  polymere balls on aluminium substrate, tin-on-carbon (5-30  $\mu\text{m}$  particles) and gold-on-carbon (5-150 nm particles) test specimens (Figure 3).

## III. DEVELOPED SOLUTION

The block diagram of the solution is depicted in Figure 4. The main stages include coarse-to-fine search for the initial guess of peak position, by means of the gradient ascent search and ultra-fine search, for the final peak position, by means of SVM fitting. Both the stages are performed at predefined lower region of interest, magnification and scan speed, and at the end, the original settings are restored.

The gradient ascent search method formerly described and applied to a thermo-SEM working at low magnifications by Marturi, Tamadazte, Dembélé & Piat (2013) is revisited. It consists in moving the SEM to the peak of focus by steps relative to the focus: the step is great far from the peak and decreases progressively when approaching the peak.

Let  $S_k$  and  $\Delta_k$  be the focus and the step at the iteration  $k$ .  $\Delta_k$  is written as:

$$\Delta_k = \Gamma_k \Lambda_k$$

(eq. 2)

with  $\Gamma_k$  the direction of motion,

$$\Gamma_k = \frac{\nabla S_k}{\|\nabla S_k\|}$$

and  $\Lambda_k$  the decreasing factor,

$$\Lambda_k = \begin{cases} \alpha \left(\frac{S_0}{S_k}\right)^2 & \text{if } \left(\frac{S_0}{S_k}\right) < 1 \\ \alpha_{else} & \text{else} \end{cases}$$

The parameter  $\alpha$  enables to tune the speed of focusing: the highest is this parameter the fastest is the autofocusing but the highest will be the level of instability. The value of  $\alpha$  is empirically chosen as the 3/2 of the depth of field.

From the current focal length  $F_k$ , the next focal length  $F_{k+1}$  is defined by:

$$F_{k+1} = F_k + \Delta_k$$

(eq. 3)

The control is stopped when the peak of variance is reached, which is detected by the zero-crossing of its derivative with respect to focal length (Figure 5). That value defines the initial guess of the peak position, i.e. it is used for the SVM fitting.

SVM is a convex optimization method to estimate a linear model, Smola & Scholkopf (2004), Bishop (2006):

$$y(x) = \omega^T \Phi(x) + b$$

(eq. 4)

with  $y$  the output or labelled data,  $x$  the input or training data,  $\omega$  the vector of model parameters,  $b$  the bias parameter and  $\Phi$  the vector of kernel functions.

A regularized error function is minimized:

$$C \sum_{k=1}^N E_{\epsilon}(y(x_k) - t_k) + \frac{1}{2} \|\omega\|^2$$

(eq. 5)

with  $C$  the regularization parameter,  $t_k$  the target value of  $y(x_k)$ ,  $\epsilon$  the slack parameter,  $N$  the number of samples and  $E_{\epsilon}$  the  $\epsilon$ -insensitive error function:

$$E_{\epsilon}(y(x_k) - t_k) = \begin{cases} 0 & \text{if } |y(x_k) - t_k| < \epsilon \\ |y(x_k) - t_k| & \text{otherwise} \end{cases}$$

(eq. 6)

The problem is solved from Lagrangian multipliers  $a_k$  and  $\hat{a}_k$ :

$$\omega = (\dots; a_k - \hat{a}_k; \dots); b = \frac{1}{N} \sum_k t_k - \epsilon - \omega^T \Phi(x_k)$$

(eq. 7)

The regularization parameter  $C$  avoids over-fitting, i.e. avoid the values of the model parameters to reach large values. The slack parameter  $\epsilon$  determines the admitted variations in the values of the input data with respect to the model. The both parameters,  $C$  and  $\epsilon$ , explain the robustness of SVM, Bishop (2006) over non regularized least squares.

SVM is known to fit sparse training data i.e. the estimation of new inputs only depends on the kernel to be evaluated at few training points. We use the  $\epsilon$ -Support Vector Regression ( $\epsilon$ -SVR) with a RBF (Radial Basis Function) kernel:

$$\Phi_i(x_k, z_k) = e^{-\gamma \|x_k - z_k\|^2}$$

(eq. 8)

The 3 parameters,  $C$ ,  $\epsilon$  and  $\gamma$ , have to be estimated accurately, for that a gradient ascent search is first performed to get peak position, several images (301) are acquired around that position from which training (position) and labels (focus) data are generated. Following the references Smola & Scholkopf (2004), Chang & Lin (2001), an automatic train including a cross-validation is performed to get the parameter values.

With the values of  $C$ ,  $\gamma$  and  $\epsilon$  the SVM is trained from the data of 7 images to compute the model from which the peak position (best focal length) is derived.

In addition to our solution, four benchmark solutions were considered. The first was an expert of the microscope who daily used it. The second method was the Zeiss solution that comes with their application development kit. Its main stages comprises coarse autofocus by searching from 0 to 20 mm (full focal length range) to find the initial guess of peak, and fine autofocus around that position by Fibonacci searches over increasing magnifications. The third solution was that of Nishi, Moriyama, Yoshida, Kajimura, Mogaki, Ozawa & Isakozawa. (2013). They manually estimated the initial guess of the peak, and assuming a quasi-gaussian model ( $a \exp\left(-\frac{|x-b|^{1.3}}{d}\right) + c$ ) they used non-regularized non-linear least squares search to find the peak. The last benchmark was the solution of Wu, Wang & Zhou (2012). They used hill climbing search for initial guess. Assuming an exponential model at the left of the peak ( $a_l e^{b_l x}$ ), and at the right of the peak ( $a_r e^{b_r x}$ ) they used non-regularized non-linear least squares search to find both models and to compute the peak by their intersection. We replaced manual and hill climbing of the last two methods, respectively, by our gradient ascent search and only implemented the fitting stages.

The results of the four methods are compared with our solution, called FEMTO in the tables.

## IV. RESULTS

The first experiment was performed with the gold-coated gripper over 20  $\mu\text{m}$  polymere balls on an aluminium substrate, the final objective was the handling of the balls. The following stable settings were used: secondary electron detector, 3kV voltage, 60  $\mu\text{m}$  aperture, 49.8% brightness (i.e. the ratio of image intensity mean with respect to 255) and 19.6% contrast (i.e. the ratio of the higher image intensity with respect to the lower intensity).

In order to evaluate the accuracy of the method, the autofocus was performed at x300 and the magnification was switched to x900, x1200, x1500 and x2100, respectively, and at each magnification the scan speed took the value 155.5 ns/pixel (noisy images), 480.5 ns/pixel and 1780 ns/pixel (sharp images), respectively. In every case the focus was computed. The results are summarized in Table 1. Except low magnification (x300), our method gives high focus images than any other method. Assuming the accuracy is defined by the value of the focus, these results show that our method overtakes all the other methods. It is slightly better than Zeiss method, 2 times, 1.5 time, 1.25 time better than Wu method, Nishi method, and Expert, respectively.

To evaluate the fidelity of the focal length obtained, we changed the starting point of the autofocus to 10 mm, 9mm, 8 mm, 7 mm and 6 mm, respectively. The results are summarized in Table 2 where it can be seen that our method has a standard deviation in the measurement

of the focal length of 0.025 mm vs. 0.021 mm for Nishi method. Unfortunately the latter fails in some cases, e.g. 8 mm. The method is 1.44 time better than the Wu's method. For a starting point of 8 mm, the speed of the autofocus were 11 s, 33 s, 40 s (+ eventually 10 s for extra fine autofocus) for our method, Expert and Zeiss, respectively: our method is the fastest.

The second experiment was performed with the tin-on-carbon test specimen. The stable settings were: secondary electron detector, 3kV voltage, 60  $\mu\text{m}$  aperture, 50.4% brightness and 22.3% contrast.

For accuracy evaluation the autofocus was performed at x300 and the magnification was switched to x3000, x9000, x30000, x60000, x90000, x120000 and x160000, respectively, and the scan speed was switched to 155.5 ns/pixel (noisy images), 480.5 ns/pixel and 1780 ns/pixel (sharp images), respectively. The results are summarized in Table 3. Our method is slightly less better than Zeiss method, but 3 times and 8 times better than Wu method and Nishi method, respectively.

For fidelity evaluation, the starting point was changed to 8 mm, 7mm, 6 mm and 4 mm, respectively. The results are summarized in Table 4. Our method has the same standard deviation of Zeiss method, i.e. 0.014, that is 6 times better than Wu and Nishi methods.

For a starting point of 10 mm, the speed of the autofocus were 13 s, 30 s, 40 s (+ eventually 10 s for extra fine autofocus) for our method, Expert and Zeiss, respectively: our method overtakes all the other methods.

The third experiment was performed with the gold-on-carbon test specimen. The same procedure was used as the previous experiment.

Table 5 summarized the accuracy of the methods. Our method overtakes all the other method: it is 6 times, 6.5 times and 1.6 time better than Wu's method, Nishi's method and Zeiss's method respectively.

Table 6 summarized the fidelity of the methods. Our method overtakes all the other methods: it is 28 times, 14 times, 1.6 time better than Wu's method, Nishi's method and Zeiss's method respectively.

For a starting point of 7 mm our method overtakes the other methods with a speed of 10 s, 15 s, 40 s (+ eventually 10 s for extra fine autofocus) for our method, Expert and Zeiss, respectively.

Above results can be easily explained. If the model is close to the trained points, the 3 methods (our method, Wu's method and Nishi' method) find the peak with high accuracy (Figures 6 and 7). This is normal because the 3 methods are based on non-linear least squares optimization. In the other cases, i.e. presence of outliers Nishi's method gives inaccurate peak (Figure 8) or fails to find the peak (Figure 9). Wu's method also gives inaccurate peak (Figure 10), but does not failed.

Finally, our method out-performs all the other methods with respect to accuracy and fidelity of measurements.

## VI. CONCLUSION

The paper has investigated the problem of autofocus in scanning electron microscopy. A solution is developed that combines gradient ascent search, a native coarse-to-fine search approach, to find the peak of focus, with the machine learning SVM fitting, a regularized and non-linear least squares optimization method, of 7 focus data to compute the peak. It has been applied to a Carl Zeiss Auriga FE-SEM with 3 specimen, a gold-coated gripper over 20



$\mu\text{m}$  polymere balls on an aluminium substrate, a tin-on-carbon test specimen with 5-30  $\mu\text{m}$  particles and a gold-on-carbon test specimen with 5-150 nm particles.

The results have shown an improvement of accuracy with respect to literature non-regularized optimization methods and Fibonacci search method ranging from 1.25 to 8 times. The improvement of fidelity ranged from 1.6 to 28 times, that of speed from 1.5 to 4 times.

Finally the work led to a practical and efficient autofocus method for electron microscopes and probably for other imaging systems. The main drawback of this solution is the gradient ascent search, which gives the initial guess of peak, it sometimes gets stuck at local maxima. An improvement would come with the use of a robust method like Newton's method.

Autofocus is interesting for standard SEM use: analysis of specimen from two-dimension images. It becomes essential for the real-time depth estimation during robotic handling of specimen, Fatikow, Wich, Hülsen, Sievers & Jähnisch (2007), and more particularly for the reconstruction of three-dimension images by means of the structure-from-motion approach: rotation of the specimen under the electron column, acquisition of focused images, processing of images and reconstruction of the three-dimension mode, Kratochvil, Dong, Zhang & Nelson (2010). At high magnification the focus can get lost during image acquisition and it is required to perform an autofocus.

## **ACKNOWLEDGMENTS**

This work has been supported by the Equipex ROBOTEX project (contract "ANR-10-EQPX-44-01), the Labex ACTION project (contract "ANR-11-LABX-0001-01) and the NANOROBUST project (contract "ANR-11-NANO-006").

We thanks Julien Derivet for his participation and the team ROBOTEX for their support.

## **REFERENCES**

Batten, C. F. (2000) Autofocusing and astigmatism correction in the scanning electron microscope. Master's thesis, University of Cambridge, Cambridge.

Bishop, C. M. (2006) Pattern Recognition and Machine Learning. Springer, New York, USA.

Bradski, G. (2000) The OpenCV Library. Doctor Dobbs Journal, 25 (11), 120-126.

Chang, C-C & Lin, C-J (2001) Libsvm: a library for Support Vector Machines. ACM Transactions on Intelligent Systems and Technology, 2 (27) : 1-27.

Fatikow, S., Wich, T., Hülsen, H., Sievers, T. & Jähnisch, M. (2007) Microrobot system for automatic nanohandling inside a scanning electron microscope. IEEE/ASME Trans, 12 : 244-252.

He, J., Zhou, R. & Hong, Z. (2003) Modified fast climbing search auto-focus algorithm with adaptive step size searching technique for digital camera. IEEE Transactions on Consumer Electronics, 49 (2) : 257-262.

Kratochvil, B.E., Dong, L.X., Zhang, L. & Nelson, B.J. (2010) Image-based 3D reconstruction using helical nanobelts for localized rotations. Journal of Microscopy, 237:122-135.



Marturi, N., Tamadazte, B., Dembélé, S. & Piat, N. (2013) Visual servoing-based approach for efficient autofocus in scanning electron microscope. In 2013 IEEE/RSJ International Conference on Intelligent Robots and Systems, Tokyo, pp. 2677-2682.

Mir, H., Xu, P., Chen, R., van Beek, P. (2015) An autofocus heuristic for digital cameras based on supervised machine learning. *Journal of Heuristics*, 21 (5) : 599-622.

Nicolls, F.C., de Jager, G. & Sewell, B.T. (1997) Use of a general imaging model to achieve predictive autofocus in the scanning electron microscope. *Ultramicroscopy*, 69(1) : 25-37.

Nishi, R., Moriyama, Y., Yoshida, K., Kajimura, N., Mogaki, H., Ozawa, M. & Isakozawa, S. (2013) An autofocus method using quasi-gaussian fitting of image sharpness in ultra-high-voltage electron microscopy. *Microscopy*, 62 (5) : 515-519.

Ong, K.H., Phang, J.C.H. & Thong, J.T.L. (1998) A robust focusing and astigmatism correction method for the scanning electron microscope - Part II: Autocorrelation-Based Coarse Focusing Method. *Scanning*, 20 : 324-334.

Ong, K.H., Phang, J.C.H. & Thong, J.T.L. (1998) A robust focusing and astigmatism correction method for the scanning electron microscope - Part III: An improved technique. *Scanning*, 20 : 357-368.

Rudnaya, M.E., Mattheij, R.M.M. & Maubach, J.M.L. (2010) Evaluating sharpness functions for automated scanning electron microscopy. *Journal of Microscopy*, 240(1) : 38-49.

Rudnaya, M.E., ter Morsche, H.G., Maubach, J.M.L. & Mattheij, R.M.M. (2012) A derivative-based fast autofocus method in electron microscopy. *Journal of Mathematical Imaging and Vision*, 44 : 38-51.

Smola, A.J. & Scholkopf, B. (2004) A tutorial on support vector regression. *Statistics and Computing*, 14 : 199-222.

Wu, Z.M., Wang, D.H. & Zhou, F. (2012) Bilateral prediction and intersection calculation autofocus method for automated microscopy. *Journal of Microscopy*, 2048(3) : 271-280.

## TABLES

Table 1: Comparison of method accuracy with the gripper over 20 µm polymere balls on aluminium substrate (x means the data were not available).

Magnification	Scan Speed (ns/pixel)	Focus				
		FEMTO	WU	NISHI	ZEISS	EXPERT
300	155,5	2,697	2,394	2,559	<b>3,040</b>	2,519
	480,5	2,912	2,326	2,710	<b>3,271</b>	2,636
	1780	3,169	2,292	2,911	<b>3,536</b>	2,878
900	155,5	<b>4,928</b>	2,935	3,519	4,615	3,768
	480,5	<b>6,203</b>	2,974	4,344	5,537	4,901
	1780	<b>6,430</b>	2,977	4,258	5,652	4,936
1200	155,5	<b>7,175</b>	3,767	5,002	6,507	5,468
	480,5	<b>9,130</b>	4,025	6,397	7,801	7,299
	1780	<b>9,447</b>	4,081	6,077	7,976	7,213
1500	155,5	8,868	4,450	6,369	<b>9,074</b>	6,927
	480,5	<b>10,877</b>	4,800	7,569	10,305	8,644
	1780	<b>10,756</b>	4,898	7,008	10,264	8,252
2100	155,5	<b>11,632</b>	x	x	11,521	9,152
	480,5	<b>13,909</b>	x	x	13,553	11,453
	1780	<b>15,565</b>	x	x	15,285	12,706
Average		<b>8,247</b>	3,493	4,894	7,862	6,583

Table 2: Comparison of method fidelity with the gripper over 20 µm polymere balls on aluminium substrate (x means the data were not available).

Starting point	Peak position /Focal length (mm)				
(mm)	FEMTO	WU	NISHI	ZEISS	EXPERT
10	<b>11,059</b>	11,034	<b>11,080</b>	x	x
9	<b>11,017</b>	11,105	<b>11,037</b>	x	x
8	<b>11,079</b>	11,089	<b>failed</b>	x	x
7	<b>11,068</b>	11,127	<b>11,078</b>	11,057	x
6	<b>11,075</b>	11,115	<b>11,078</b>	x	11,058
Average	<b>11,059</b>	11,094	<b>11,068</b>	x	x
Std Deviation	<b>0,025</b>	0,036	<b>0,021</b>	x	x

Table 3: Comparison of method accuracy with the tin-on-carbon specimen.

Magnification	Scan Speed (ns/pixel)	Focus			
		FEMTO	WU	NISHI	ZEISS
3000	155,5	12,231	8,376	3,585	12,332
	480,5	12,845	7,82	2,806	13,102
	1780	12,864	7,442	2,398	13,195
9000	155,5	16,548	10,47	2,552	17,128
	480,5	16,598	9,698	1,735	17,492
	1780	16,392	9,307	1,324	17,388
30000	155,5	9,582	1,968	1,418	9,934
	480,5	9,124	1,158	0,638	9,505
	1780	8,734	0,753	0,245	9,127
60000	155,5	11,319	1,69	1,406	11,546
	480,5	10,781	0,89	0,627	11,063
	1780	10,405	0,481	0,234	10,723
90000	155,5	9,143	1,584	1,406	9,218
	480,5	8,474	0,787	0,624	8,6
	1780	8,097	0,379	0,229	8,216
120000	155,5	6,589	1,516	1,399	7,741
	480,5	5,814	0,717	0,621	7,155
	1780	5,402	0,311	0,227	6,786
160000	155,5	4,455	1,471	1,395	6,69
	480,5	3,524	0,675	0,619	6,039
	1780	3,095	0,27	0,226	5,659
Average		9,620	3,227	1,224	10,411

Table 4: Comparison of method fidelity with the tin-on-carbon specimen balls (x means the data were not available).

Starting point (mm)	Peak position /Focal length (mm)				
	FEMTO	WU	NISHI	ZEISS	EXPERT
8	9,278	9,274	9,278	9,261	
7	9,288	9,291	failed	9,244	9,291
6	9,255	9,462	8,037	9,261	x
4	9,273	9,332	failed	9,278	x
Average	9,274	9,340	8,657	9,261	x
Std Deviation	0,014	0,085	0,877	0,014	x

Table 5: Comparison of method accuracy with the gold-on-carbon specimen.

Magnification	Scan Speed (ns/pixel)	Focus			
		FEMTO	WU	NISHI	ZEISS
3000	155,5	<b>3,899</b>	1,863	1,806	3,348
	480,5	<b>5,402</b>	0,925	0,868	4,164
	1780	<b>6,437</b>	0,440	0,378	4,683
9000	155,5	<b>7,032</b>	1,852	1,828	5,203
	480,5	<b>10,030</b>	0,920	0,884	6,643
	1780	<b>12,155</b>	0,429	0,389	7,619
30000	155,5	<b>6,681</b>	1,928	1,887	4,788
	480,5	<b>7,622</b>	0,991	0,959	4,776
	1780	<b>7,676</b>	0,500	0,470	4,519
60000	155,5	<b>6,474</b>	1,811	1,782	4,075
	480,5	<b>6,476</b>	0,882	0,858	3,502
	1780	<b>6,095</b>	0,398	0,378	3,061
90000	155,5	<b>6,168</b>	1,814	1,721	3,662
	480,5	<b>5,755</b>	0,889	0,803	2,935
	1780	<b>5,305</b>	0,409	0,327	2,490
120000	155,5	<b>6,227</b>	1,840	1,691	3,746
	480,5	<b>5,604</b>	0,918	0,781	2,962
	1780	<b>5,166</b>	0,435	0,305	2,535
160000	155,5	<b>6,141</b>	1,866	1,692	3,949
	480,5	<b>5,451</b>	0,946	0,781	3,160
	1780	<b>4,984</b>	0,465	0,305	2,722
Average		<b>6,513</b>	1,072	0,995	4,026

Table 6: Comparison of method fidelity with the gold-on-carbon specimen balls (x means the data were not available).

Starting point (mm)	Peak position /Focal length (mm)				
	FEMTO	WU	NISHI	ZEISS	EXPERT
7	<b>8,946</b>	8,993	9,035	8,945	8,953
6	<b>8,943</b>	8,858	8,948	x	x
5	<b>8,951</b>	9,057	8,947	8,955	x
4	<b>8,947</b>	9,053	8,924	8,954	x
Average	<b>8,947</b>	8,990	8,964	8,951	x
Std Deviation	<b>0,0033</b>	0,0929	0,0489	0,0055	x

## FIGURES



Figure 1: The setup with the Zeiss Auriga FE-SEM and the associated PC.











Figure 2: Illustration of the normalized variance as focus with the tin-on-carbon specimen at x300 : the image with the maximum of focus is shown completely, the others are shown partially (cropping of an region-of-interest).






Figure 3: Experimental benchmark specimens.






Figure 4: Block diagram of the developed solution.




Figure 5: Gradient ascent search illustration with the gold-on-carbon specimen at x300.




Figure 6: Our method vs. Nishi's method when the trained points were closed to the model, both methods located the peak at 9.278 mm.




Figure 7: Our method vs. Wu's method when the trained points were closed to the model, our method and Wu's method located the peak at 9.278 mm and 9.274, respectively, i.e. an error of 0.4%.




Figure 8: Our method vs. Nishi's method when one trained point was far from the model, i.e. an outlier, our method and Nishi's method located the peak at 11.059 mm and 11.080 mm, respectively, i.e. an error of 2% of Nishi's method to locate the peak.




Figure 9: Our method vs. Nishi's method when two trained points were far from the model, our method located the peak at 11.079 and Nishi's method failed.




Figure 10: Our method vs. Wu's method when the trained points were not symmetric with respect to the peak guess, our method and Wu's method located the peak at 9.255 mm and 9.462 mm, respectively, i.e. an error of 2%.