



HAL
open science

LIBS-Based Imaging: Recent Advances and Future Directions

Vincent Motto-Ros, V Gardette, L Sancey, M Leprince, D Genty, Stéphane Roux, B
Busser, F Pelascini

► **To cite this version:**

Vincent Motto-Ros, V Gardette, L Sancey, M Leprince, D Genty, et al.. LIBS-Based Imaging: Recent Advances and Future Directions. *Journal of Spectroscopy*, 2020, 35 (2), pp.34-40. <hal-02998713>

HAL Id: hal-02998713

<https://hal.science/hal-02998713v1>

Submitted on 10 Nov 2020

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization

LIBS-Based Imaging: Recent Advances and Future Directions

V. Motto-Ros, V. Gardette, L. Sancey, M. Leprince, D. Genty, S. Roux, B. Busser, and F. Pelascini

Laser-induced breakdown spectroscopy (LIBS)-based imaging is becoming a promising technique in the panel of spatially-resolved elemental approaches. This method has outstanding advantages, such as all optical instrumentation, fast acquisition speed, operation at ambient atmosphere, and detection limits at the ppm scale for most elements. LIBS-based imaging has an extensive range of applications, including biology, medicine, industry, and geology. In this paper, we aim to report recent advances in LIBS imaging, focusing on instrumentation, performance, and applicability. Two examples are shown; first, a speleothem with application to paleoclimate studies, and second, a biological organ with implications for toxicology of new drugs based on nanoparticles. Finally, some future directions are discussed.

In the past few years, the application of laser-induced breakdown spectroscopy (LIBS) to microscopic elemental imaging has known strong developments, both in terms of instrumentation and applications (1–4). In LIBS imaging, a series of laser-induced plasma is generated at different positions on the sample surface with a predefined sequence covering the region of interest. Such plasma sources allow specific optical responses resulting from the electronic relaxation of atoms and ions excited by the high plasma temperature to be elicited from the elements constituting the sample. The light emitted by the plasma is collected and analyzed using an optical spectrometer. The elemental “signal” (atomic and ionic emissions) is then extracted from the recorded spectra, and elemental maps can be obtained in a pixel-by-pixel manner (Figure 1) (5,6).

The main advantage of LIBS certainly lies in its simple implementation. This is because a single laser pulse simultaneously samples the material by laser ablation, atomizes, and then excites the vaporized mass by heating the plasma plume. Such apparent simplicity endows this technique with a series of advantages, including an all-optical design compatible with conventional optical microscopy, operation in ambient atmosphere, and fast operating speed up to kHz. In addition to its table-top instrumentation, this technique has multielement capabilities, no restrictions in the detection of light

elements, detection limits in the range of ppm for most elements, and microscopic-scale resolution (2). It also has quantitative capabilities through the use of standards with known concentrations or an external calibration using different strategies (6–9). All of these advantages make LIBS imaging highly promising, with the potential to become a reference technique in the area of spatially-resolved elemental techniques with good complement to gold standard methods, such as laser ablation inductively coupled plasma mass spectrometry (LA-ICP-MS), synchrotron radiation microanalysis, or electron probe microanalysis (EPMA) (10–13).

Recently, the LIBS community demonstrated the interest of this technique for various ranges of applications, covering geological, industrial, and biomedical domains. For example, the capability of LIBS imaging to characterize complex samples was explored with the analysis of a multiphase hydrothermal ore sample (14). More than 30 elements were identified, with the concentration ranges extending from major to trace elements. For application to paleoclimate, large-scale speleothems and coral samples ($> 10 \text{ cm}^2$) were analyzed, with the aim of revealing the laminar structures related to vegetation or climatic variations (5). Several studies applied to industrial materials have also been conducted on crystals (7,9) or heterogeneous catalysts (8,15). The possibility of obtaining elemental images of biological tissues was

also demonstrated in different biomedical situations (16–18). In this paper, we illustrate recent advances in LIBS imaging in terms of instrumentation and applicability. Two different examples are shown to illustrate the specificities and the current performance of this technique.

Experimental

The micro-LIBS system developed in our group is based on a custom microscope (Figure 2). The opto-mechanical structure has been designed with the use of cage and tube systems commercialized by Thorlabs. The device includes a laser injection line, monitoring cameras (sample surface, plasma), and illuminators (white light, laser pointer), as well as up to three collection lines connected to spectrometers with optical fibers (19). The device is also equipped with a Raman system operating at 735 nm, and allowing punctual analysis to be performed. The laser used is a nanosecond Nd:YAG laser working at the fundamental wavelength (1064 nm)




FIGURE 1: LIBS-based imaging principle. (a) Schematic view of the micro-LIBS setup showing the 15x microscope objective used to focus the laser pulse, the motorized platform supporting the sample and an optical detection system connected to the optical spectrometer via an optical fiber. (b) Example of single-shot emission spectra recorded in four different regions of the sample (a rat kidney sampled one hour after gold nanoparticle administration) with the characteristic emission lines of iron (Fe), phosphorous (P), and gold (Au). (c) Principle of data extraction using LasMap Software. (d) Example of relative-abundance images of Au (yellow), Fe (red), and P (blue).

with a frequency rate of 100 Hz (Centurion, Quantel). The typical pulse energy is

in the range of 1 mJ for most applications. The laser energy is stabilized throughout

Avantes

ppm, but strongly deteriorates for organic and halogen elements. It is important to emphasize that such relative LoDs are obtained in a single pulse configuration and from an ablated mass in the range of 0.1 nanogram (typical ablated mass from a single laser pulse). Expressed in absolute detection limits (for example, the smallest mass quantity that can be detected), values typically lower than the femto-gram can be achieved.

Paleoclimate Application

Thanks to the 100 Hz scanning speed, and because there is no constraint on the size of the sample (such as, for example, no need for an ablation cell), large surfaces can be easily analyzed. Such aspects may be valuable in the case of geological samples, particularly for the characterization of speleothems. Stalagmites, stalactites, or columns precipitate from seepage waters in caves, and usually consist of calcium carbonate, with calcite being the most common mineral phase. Climate proxies, such as Mg, Sr, or Ba, can be retrieved from these materials, and may indicate changes in climate or environmental conditions that occurred during the growth periods and provide information on past temperature, precipitation, and vegetation changes over the last hundreds of thousands of years (21). For recording paleoclimate proxies, it is interesting to analyze several centimeters, or even more, of a sample while keeping a resolution on the scale of the annual speleothem growth (typical rates of 0.01 to 2 mm per year) (22).

An example is shown in Figure 4 for a speleothem extracted from a Gallo-Roman copper ore mine situated close to the "Grotte Bleue" in Aude, France (courtesy of D. Genty and P. Cabrol of DREAL, in Toulouse, France). At the collection site, there is a strong abundance of copper, which gives a particular blue color to this mineral. This site is also known to be rich in iron, lead and arsenic. Amphoras discovered on the site suggested Roman exploitation of the cave from the 1st century BC, but it is suspected that the site was also occupied in the Bronze Age (23). In the example shown in Figure 4, a mapping sequence of 2300 x 1600 pixels was performed with a




FIGURE 4: Example of elemental images obtained from a speleothem section recovered from the Gallo-Roman copper ore mine located close to the "Grotte Bleue" in Aude, France. Images shown are (from left to right): an optical image, and highlighted composition images for magnesium (Mg), lead (Pb), and copper (Cu).

step size of 24 μm , which corresponds to a surface of analysis of more than 20 cm^2 . The optical image of the sample and elemental images of Mg and Pb are shown. These results illustrate that the resolution capabilities of the LIBS imaging method and annual growth layers can be observed. It is worth mentioning that other elements, such as Cu, Ag, Sr, Ba, Al, and Si, were also detected. The interpretation of such results is currently being explored by specialists.

Imaging Biological Specimens

LIBS-imaging can also be applied on soft biological tissues. Our first work was conducted on a mouse kidney to image the biodistribution of nanoparticles before administration to animals (6,24,25). These works were followed by several articles reporting biological studies conducted on murine animals, where LIBS was used in complement with more conventional techniques such as transmission elec-



Orbis Micro-XRF Analysis System

- Non-destructive X-ray elemental analysis from Na to Bk
- Excitation geometry perpendicular to sample for accurate sample targeting
- High-intensity analysis for micro to millimeter areas
- Increased sensitivity with filters as standard
- Fastest analytical speed with state of the art detectors

For more information, please visit:
<https://www.edax.com/micro-xrf>

AMETEK
MATERIALS ANALYSIS DIVISION

[edax.com](https://www.edax.com)

EDAX
Smart Insight




FIGURE 5: Multielemental image of a rat kidney after intravenous injection of gold (Au) nanoparticles. Merged image composed of the combination of three false color elemental images for Au (yellow), P (blue) and Fe (red). These animal experiments were approved by the local ethics committee under agreement A21231016EA.

tronic microscopy (TEM) or fluorescence microscopy (17,18,26). More recently, some studies were also conducted with human specimens of medical interest (16,27). In health infrastructures, the possibility of coupling LIBS imaging with conventional histopathology could be highly valuable for the completion of some medical diagnoses. Because LIBS imaging is compatible with the light microscopy used in pathology laboratories, it may provide important support for uncovering the elemental composition of tissues, particularly in the case of numerous pathologies relating to exposure to metals (such as respiratory diseases, dermatological diseases, and cancers, for example).

As an example, Figure 5 shows an image obtained for a rat kidney section. The organ was sampled 1 h after intravenous administration of gold nanoparticles. This experiment was performed with a resolution of 10 μm , and represents 2 million individual pixels. Three elements are represented with a false color scale: Au in yellow, P in blue (the P distribution reflects the histological structure of the tissue), and Fe in red (in this case, the threshold allows for visualizing Fe from hemoglobin, and by extent, blood vessels). As observed in several previous work (6,26), small NPs are mainly eliminated by the renal route with partial retention in the cortex region of the organ (peripheral region), indicating a normal elimination process. More generally, the possibility of imaging nanoparticle

distribution at the entire organ scale and without any labeling is highly relevant for studies involving the development of new drugs based on metals.

Status and Future Directions

LIBS-based imaging is a mature technique for various classes of materials. These materials consist of a homogeneous matrix in which it is desired to observe the possible heterogeneity of minor or trace elements (both examples shown above (Figures 4 and 5)). The advantage of such a simple material is that matrix effects can be avoided, greatly facilitating the calibration of raw intensities. In addition, since we are interested in minor or trace elements, line saturation phenomena (such as self-absorption and detector saturation) are nonexistent, and the intensity of the measured lines is linearly dependent on the concentration.

However, difficulties arise when considering complex materials with multiple matrices. These difficulties are inherent to all elemental imaging methods. Regardless of the analytical technique, the interaction between the primary beam and the material will always depend on the sample properties. The recorded signal will then be dependent on the material (nature, hardness, composition, and so on), so it is difficult to generate a quantitative analysis on such complex material, regardless of the technique used. Currently, the classical way to calibrate the intensities is to use reference samples for each phase

of the material, knowing that in practice these standards do not always exist. An interesting possibility for the future is to implement a calibration-free approach. Such a method does not require external references because it only relies on the physical analysis of the spectra. By considering a given plasma model and measuring its parameters (such as temperature or electronic density, for example), it is possible to extract the concentrations of the various detected elements, regardless the type of the matrix. Such a method could be highly promising for the development of the technique, and several groups are currently working on its implementation.

An additional difficulty related to emission spectroscopy comes from the possible complexity of the signal. Some elements, such as transition metals, have a large number of emission lines from the ultraviolet (UV) to near infrared (NIR) range. This is the case in particular for iron or titanium. If one or more of these elements are a matrix constituent, the observed spectra will be spectrally dense, and the desired lines (generally from minor or trace elements) will likely be masked. In this case, the performance in terms of sensitivity will deteriorate.

Managing the processing of data obtained on complex samples also represents a challenge and limits the fields of possible exploration of the technique. It must be kept in mind that the data to be processed are very large (up to several tens of gigabytes), and the software used cannot be supervised. It is indeed impractical to interpret spectra one by one for a megapixel analysis. A promising possibility that is currently being explored is the use of statistical analysis methods (such as chemometrics). In general, these methods allow the dimensionality of a series of data to be reduced by looking for correlations among the different variables, thus making the information less redundant and easier to handle. We recently published an article detailing the implementation of principal component analysis (28), and several groups in the LIBS community are currently exploring other methods, such as the possibility to use machine learning approaches.

Emission

Conclusion

LIBS imaging is currently experiencing strong development with an increase in the number of published papers during the three last years. Due to the feasibility studies conducted in laboratories, this approach is becoming increasingly mature, and is currently a good candidate to become a reference approach for spatially-resolved elemental characterization in the near future. LIBS has unique specificities, such as the possibility of reaching micrometric scale resolution with limits of detection down to ppm, operation at ambient conditions with table top and all optical instrumentation. Regarding the acquisition rate, recent work has demonstrated the possibility of access to kHz acquisition rates (29), and higher operating speeds should be achieved very soon.

The future will tell if this LIBS-based imaging can be generalized and used outside academic laboratories. What is certain is that, at some point, its development will have to be supported by large industrial groups. In the meantime, researchers still have some work to explore. In particular, the analysis of complex materials (those composed of several matrices) appears challenging, both for quantification and data processing. Further studies dedicated to these issues need to be conducted, and the use of calibration-free LIBS and advanced chemometric tools represent interesting opportunities.

Acknowledgments

This work was partially supported by the French region Rhône Alpes Auvergne (Optolyse, CPER2016) and the French ANR (projects MediLIBS ANR-17-CE18-0028-04 and Imazinc ANR-16-CE29-0024). In addition, we gratefully acknowledge Florian Trichard from Ablatom S.A. S, Nicolo Omenetoo from the University of Florida, Ludovic Duponchel from the University of Lille, and Christophe Dujardin from ILM for fruitful discussions.

References

1. M.P. Mateo and G. Nicolas, *App. Spectrosc. Rev.* **48**, 357–383 (2013).
2. L. Jolivet, M. Leprince, S. Moncayo, L. Sorbier, C.P. Lienemann, and V. Motto-Ros, *Spectrochim. Acta Part B* **151**, 41–53 (2019).
3. B. Busser, S. Moncayo, J.-L. Coll, L. Sancey, and V. Motto-Ros, *Coord. Chem. Rev.* **358**, 70–79 (2018).
4. R. Gaudiuso et al., *Spectrochim. Acta Part B* **152**, 123–148 (2019).
5. J.O. Cáceres, F. Pelascini, V. Motto-Ros, S. Moncayo, F. Trichard, G. Panczer, A. Marín-Roldán, J.A. Cruz, I. Coronado, and J. Martín-Chivelet, *Sci. Rep.* **7**, 5080 (2017).
6. L. Sancey, V. Motto-Ros, B. Busser, S. Kotb, J.M. Benoit, A. Piednoir, F. Lux, O. Tillement, G. Panczer, and J. Yu, *Sci. Rep.* **4**, 6065 (2014).
7. G. Alombert-Goget, F. Trichard, H. Li, C. Pezzani, M. Silvestre, N.s Barthalay, V. Motto-Ros, and K. Lebbou, *Optical Materials*, ISSN 0925-3467 (2016). **(NEED MORE INFO)**
8. F. Trichard, L. Sorbier, S. Moncayo, Y. Blouët, C.-P. Lienemann, and V. Motto-Ros, *Spectrochim. Act. B* **133**, 45–51 (2017).
9. P. Veber et al., *CrystEngComm* **21**, 3844–3853 (2019).
10. K. Janssens, W. de Nolf, G. van der Snickt, L. Vincze, B. Vekemans, R. Terzano, and F.E. Brenker, *Trends Anal. Chem.* **29**, 464–478 (2010).
11. K.L. Moore, E. Lombi, F.-J. Zhao, and C.R.M. Grovenor, *Anal. Bioanal. Chem.* **402**, 3263–3273 (2012).
12. D. Pozebon, G.L. Scheffler, V.L. Dressler, and M.A.G. Nunes, *J. Anal. At. Spectrom.* **29**, 2204–2228 (2014).
13. M. Reich, R. Large, and P.A. Deditius, *Ore Geol. Rev.* **81**, 1215–1217 (2017).
14. C. Fabre et al., *J. Anal. At. Spectrom.* **33**, 1345–1353 (2018).
15. F. Trichard, F. Gaulier, J. Barbier, D. Espinat, B. Guichard, C.-P. Lienemann, L. Sorbier, P. Levitz, and V. Motto-Ros, *J. Catal.* **363**, 183–190 (2018).
16. B. Busser et al., *Mod. Pathol.* **152**, 1–7 (2017).
17. S. Kunjachan et al., *Nano Lett.* **15**, 7488 (2015).
18. X. Le Guével et al., *Nanoscale* **10**, 18657–18664 (2018).
19. V. Motto-Ros, E. Negre, P. Pelascini, G. Panczer, and J. Yu, *Spectrochim. Act. B* **92**, 60–69 (2014).
20. V. Motto-Ros, S. Moncayo, F. Trichard, and F. Pelascini, *Spectrochim. Act. B* **155**, 127–133 (2019).
21. J. Ian Fairchild et al., *Earth-Sci. Rev.* **75**, 105–153 (2016).
22. D. Genty, A. Baker, and B. Vokal, *Chem. Geol.* **176**, 191–212 (2001).
23. G. Munteanu, E. Kammenthaler, J. Manteau, and C. Rico, *ArcheoSciences* **40**, 163–180 (2016).
24. V. Motto-Ros, L. Sancey, Q. L. Ma, F. Lux, X. S. Bai, X. C. Wang, J. Yu, G. Panczer, and O. Tillement, *App. Phys. Lett.* **101**, 223702 (2012).
25. Y. Gimenez et al., *Nature Sci. Rep.* **6**, 29936 (2016).
26. L. Sancey et al., *ACS Nano* **9**, 2477 (2015).
27. S. Moncayo, F. Trichard, B. Busser, M. Sabatier-Vincent, F. Pelascini, N. Pinel, I. Templier, J. Charles, L. Sancey, and V. Motto-Ros, *Spectrochim. Act. B* **133**, 40–44 (2017).
28. S. Moncayo, L. Duponchel, N. Mousavipak, G. Panczer, F. Trichard, B. Bousquet, F. Pelascini, and V. Motto-Ros, *J. Anal. Atom. Spectrom.* **33**, 210–220 (2018).
29. K. Rifai, F. Doucet, L. Özcan, and F. Vidal, *Spectrochim. Act. B* **150**, 43–48 (2018).

V. Motto-Ros, V. Gardette and **M. Leprince** are with the Institut Lumière Matière at the Université de Lyon, in Villeurbanne, France. **Lep-rince, L. Sancey** and **B. Busser** are with the Institute for Advanced Biosciences at the Univ. Grenoble Alpes, in Grenoble, France. **B. Busser** is also with Grenoble University Hospital, in Grenoble, France. **D. Gentry** is with the Université de Bordeaux, in Pessac, France. **S. Roux** is with the Université de Franche-Comté, in Besançon, France. **F. Pelascini** is with Cetim Grand Est, in Illkirch-Graffenstaden, France. Direct correspondence to: vincent.motto-ros@univ-lyon1.fr