

Development of a label-free electrochemical aptasensor based on diazonium electrodeposition: Application to cadmium detection in water

Selma Rabai, Messaoud Benounis, Gaëlle Catanante, Abdoullatif Baraket, Abdelhamid Errachid, Nicole Jaffrezic Renault, Jean Marty, Amina Rhouati

▶ To cite this version:

Selma Rabai, Messaoud Benounis, Gaëlle Catanante, Abdoullatif Baraket, Abdelhamid Errachid, et al.. Development of a label-free electrochemical aptasensor based on diazonium electrodeposition: Application to cadmium detection in water. Analytical Biochemistry, 2021, 612, pp.113956. 10.1016/j.ab.2020.113956. hal-03140071

HAL Id: hal-03140071 https://hal.science/hal-03140071

Submitted on 17 Oct 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

Version of Record: https://www.sciencedirect.com/science/article/pii/S0003269720304887 Manuscript_51924897d41a19c3a1420b4f92253bd1

1	Development of a label-free electrochemical aptasensor based on diazonium
2	electrodeposition: Application to cadmium detection in water
3	Selma Rabai ¹ , Messaoud Benounis ¹ , Gaëlle Catanante ² , Abdoullatif Baraket ³ , Abdelhamid
4	Errachid ³ , Nicole Jaffrezic- Renault ³ , Jean-Louis Marty ² and Amina Rhouati ^{2, 4*}
5	¹ Laboratory of Sensors, Instrumentations and Process (LCIP), University of Khenchela,
6	Khenchela, Algeria.
7	² Laboratory of Biosensors, Analysis and Environment (BAE), University of Perpignan Via
8	Domitia, Perpignan, France.
9 10	³ Laboratory of Analytical Sciences, University of Claude Bernard 1, Doua 69100 Villeurbanne, Lyon, France.
11	⁴ Higher National School of Biotechnology, Constantine, Algeria.
12	*a.rhouati@ensbiotech.edu.dz
13	Abstract:
14	In this study we have developed a new aptasensor for cadmium (Cd^{2+}) detection in water. Gold
15	electrode surface has been chemically modified by electrochemical reduction of diazonium salt (CMA) with
16	carboxylic acid outward from the surface. This was used for amino-modified cadmium aptamer immobilization
17	through carbodiimide reaction. Chemical surface modification was characterized by cyclic voltammetry (CV)
18	and electrochemical impedance spectroscopy (EIS). This latter was also used for Cd ²⁺ detection. The aptasensor
19	has exhibited a good linear relationship between the logarithm of the Cd^{2+} concentration and the impedance
20	changes in the range from 10^{-3} to 10^{-9} M with a correlation R ² of 0.9954. A high sensitivity was obtained with a
21	low limit of detection (LOD) of 2.75*10 ⁻¹⁰ M. Moreover, the developed aptasensor showed a high selectivity
22	towards Cd^{2+} when compared to other interferences such as Hg^{2+} , Pb^{2+} and Zn^{2+} . The developed aptasensor
23	presents a simple and sensitive approach for Cd^{2+} detection in aqueous solutions with application for trace Cd^{2+}
24	detection in spring water samples.

25 Keywords: Label-free detection, aptasensor, diazonium electrografting, gold electrode, cadmium detection.

26 Introduction

27 Recently, pollution caused by heavy metals is becoming more and more severe. Heavy metals are 28 naturally occurring elements characterized by high atomic weight and density, which is at least five times greater 29 than that of water [1]. The distribution of heavy metals; like Arsenic, Cadmium, Chromium, Lead, and Mercury 30 in the environment is a result of their multiple applications, and raising concerns about their potential effects on 31 human health and environment in most regions around the world, because of their high degree of toxicity [2]. 32 Cadmium (Cd²⁺) is a toxic heavy metal whose presence in the environment is mainly issued from human 33 activities [3]. It is produced in the manufacture of batteries and pigments which can permeate, soil and water 34 supply. Thereby, exposing the general population to low Cd^{2+} doses can affect a variety of cellular mechanisms 35 proliferation, apoptosis, differentiation, cell signalling and gene expression. For these reasons, such as 36 Cadmium has been classified as carcinogenic to humans by the World Health Organization (WHO) even at low 37 doses, the maximum limit is 0.003 mg.L⁻¹. Indeed, it has been reported that exposure to this ion plays a particular 38 role in prostate cancer[4, 5].[4]. Numerous analytical methods traditionally used for detection of Cadmium (II) 39 ions have been widely reviewed in the literature; like atomic absorption/emission spectroscopy (AAS/AES) [6, 40 7], ultraviolet-visible spectrometry [8], Square-wave anodic stripping voltammetry [9] and atomic fluorescence 41 spectrometry (AFS) [10]. Despite their performance, most of these methods are often tedious, time consuming 42 and require great amounts of solvents, reagents and sophisticated apparatus. Therefore, inexpensive, sensitive 43 and selective techniques are very desirable for cadmium monitoring.

Biosensors are considered as very promising tools for sensitive, rapid and low cost monitoring of wide range of analytes in different fields of applications. A wide range of biomolecules or chemical species have been used in the literature as sensitive layer for chemical or biosensors manufacturing[11]. Aptamers are considered as bio-inspired receptors which have gained great interest since their discovery in 1990 [12].

Aptamers are RNA or DNA molecules with specific 3D structures, selected by *in vitro* selection process called Systematic Evolution of Ligands by Exponential Enrichment (SELEX) [13]. Due to their unique conformations, aptamers can bind a wide range of targets such as small molecules, proteins, and even whole cells. Their advantageous characteristics, such as high stability, simple chemical synthesis, possibility of modification, made them strong competitors of antibodies in molecular analysis [14-16]. They constitute thus 53 good candidates as effective probes for chemical sensors. Several aptasensors have been developed in the 54 literature for Cadmium ions detection. Hamid et al[17] have developed an electrochemical cadmium sensor using a DNA aptamer, where the sensing mechanism was based on target binding-induced conformational 55 56 change. Here, magnetic beads were used as immobilization support for the aptamer. Despite the biosensor 57 selectivity toward Cd²⁺, the detection range [250-1000 nM] of this electrochemical sensor was still rather 58 narrow. In another report, Guo et al. have developed a new colorimetric method for sensing Cd²⁺ by using 59 unmodified AuNPs [18]. Here, the linear detection range [5-40 μ M] of this method was narrow and could not be used for the detection of Cd²⁺ in various types of samples. 60

61 The critical point in aptasensor fabrication technology remains in the immobilization of active 62 biomolecules onto the transducer surface. Several immobilization techniques can be adopted according to the 63 biomolecules architecture and the desired application. For instance, direct attachment of bioreceptors to gold 64 electrodes has been reported, by using 3' or 5' thiol-labelled aptamers self-assembly through gold-sulfur bond 65 interaction by forming ordered single-carrier membrane [19, 20]. However, this technique suffers from non-66 specific adsorption (random adsorption) and cannot be applied on other materials. Attachment of aptamers onto 67 sensor surfaces based on hybridization with partially complementary oligonucleotides has been also adopted 68 [20]. Nevertheless, annealing and hybridization steps are required, enduring thus the experimental conditions. 69 Covalent attachment to chemically-modified sensor surfaces is a promising method based on the interaction of a 70 labelling group and a chemically functionalized surface resulting in a layer of ordered film of aptamers. 71 Chemical covalent attachment may increase the specificity and selectivity of the aptasensing strategy [20].

72 In this study we have chosen an amino-modified aptamer for Cd^{2+} detection that we immobilized on 73 screed printed gold electrode through diazonium salt chemistry. To the best of our knowledge, this is the first 74 report describing an electrochemical aptasensor exploring diazonium-coupling reaction mechanism for Cd²⁺ 75 determination. To achieve that, Cd²⁺aptamer was immobilized onto gold electrodes through 4-carboxymethyl 76 aryl diazonium (CMA) molecules which are electrochemically reduced on the electrode surface. Unlike 77 conventional grafting methods, diazonium salt-based electrochemical deposition involves simple reagents 78 without isolation and purification steps [21-24]. Gold working electrodes (WE) have been characterized with 79 cyclic voltammetry (CV) and electrochemical impedance spectroscopy (EIS) before and after WE chemical 80 surface modification and after aptamer immobilization. EIS was then used to monitor the extent of binding 81 between Cadmium and its specific capture aptamer.

82 Experimental

83 Chemicals and Reagents

84 Cadmium (II) nitrate [Cd (NO₃)₂],4-aminophenylacetic acid (4-carboxymethylaniline CMA), Sodium nitrite 85 (NaNO₂), isopropyl alcohol (IPA) 96%, N-hydroxysuccinimide (NHS) and N-(3-dimethylaminopropyl)-N'-86 ethyl-carbodiimide hydrochloride (EDC) were obtained from Sigma-Aldrich, France. Phosphate buffer 87 $(NaH_2PO_4.H2O; Na_2HPO_4.2H_2O)$ and KCl solution containing ferricyanide K₃ [Fe(CN)₆] and ferrocyanide K₄ [Fe(CN) 6]) were used as supporting electrolytes. Cadmium aptamer was synthesized and purchased by 88 Microsynth (Switzerland). The specific sequence of the aptamer was as follows: (5' amino-modified):5'-ACC 89 90 GAC CGT GCT GGA CTC TGG ACT GTT GTG GTA TTA TTT TTG GTT GTG CAG TAT GAG CGA GCG 91 TTG CG-3' [25].

92 Electrochemical analysis

93 The electrochemical experiments were carried out with a potentiostat (Biologic EC-Lab SP-300). The 94 electrochemical measurements were performed in PBS with pH 7.4 at room temperature in a conventional 95 electrochemical glass cell containing three-electrode system [26].

96 Cyclic Voltammetry (CV) measurements were used before and after chemical gold surface modifications. CV 97 analysis were performed in 5 mM of $[Fe(CN)_6]^{3-/4-}$ solution within a potential range of (- 0.2V to + 0.6V) and a 98 scan rate of 80 mV.s⁻¹.

Electrochemical impedance spectroscopy (EIS) was used to study the barrier properties of the modified electrode
[27, 28]. The EIS measurements were conducted at a potential of 128 mV on the Gold WEs over a frequency
range between 100 mHz and 200 kHz. The data of EIS were fitted using Randles equivalent circuit model. The
modelling of the obtained EIS data was made by the EC-Lab software using the Randomize + Simplex method.
Here, randomize was stopped on 100 000 iterations and the fit stopped on 5000 iterations.

104 Real sample analysis

In order to validate and evaluate the accuracy and applicability of the our aptasensor, analyses were carried out on a real river water which is polluted by public work wastes and containing nitrate and cadmium [29, 30]. This river is crossed by a volcanic water source from Hamman Essalihine valley (khenchela, Algeria) which is rich in minerals, such as bicarbonates, sulphates, calcium and magnesium. We have made sampling from the river 109 which contains cadmium and other potential interferences. Real samples were tested by EIS combined with 110 standard addition method [31]. This method consists of adding known amounts of a standard solution to a sample of interest with unknown concentrations. Then, the analytical signals of standard solutions are recorded 111 112 in response to each addition [31]. In brief, a constant volume (2.5mL) of water real sample was added to each 113 flask of 25mL. Then in the first flask, a 22.5mL of PBS were added to the 2.5mL of real samples in order to 114 reach a final volume of 25 mL and this flask was called Level 1. The standard solution containing the analyte 115 was then added in increasing volumes: 0.25, 2.5 and 5 mL, implying the addition of 10, 100 and 200 nM of Cd²⁺ 116 to the subsequent flasks 2, 3 and 4. Finally theses flasks were made up to the volume of 25 mL with PBS giving 117 thus Level 2, Level 3 and Level 4, respectively.

118 Results and discussion

119 Electrochemical aptasensor preparation

120 Fig.1 illustrates the detection principle of Cd^{2+} by using the electrochemical aptasensor. Cd^{2+} -aptamer 121 immobilization was achieved through the covalent linkage between amino groups of the aptamer and the 122 abundant reactive carboxylic groups on the gold surface. The electrode surface was first modified by deposition 123 of a diazonium salt film, and activation of the terminal carboxylic groups with EDC/NHS. Then, the aptamer was linked to the gold surface via its activated carboxylic groups. In the presence of Cd²⁺, a variation in the 124 125 aptamer conformation takes place from a random coil structure to an aptamer-Cd²⁺ complex [32, 33]. This 126 aptamer-target interaction hampers the electron transfer, enhancing thus the surface resistance which is 127 proportional to cadmium concentration in the sample.



Fig. 1 Schematic illustration of gold WE chemical surface modification by CMA followed by aptamer and immobilization for Cadmium detection[32, 33]using integrated RE and CE.

131 Characterization of the functionalized gold surface

132 Cyclic voltammetry studies

133 For the Cyclic Voltammetry studies, Fig.2A shows the different cyclic voltammograms of the construction steps 134 of this aptasensor. The oxidation-reduction peaks of the redox couple are visible for the bare Gold WE. Fig.2A 135 (a) showing a pair of quasi-reversible peaks characteristic of the oxidation-reduction with cathodic and anodic 136 current peak ratio of approximately one and peak-to-peak separation of 0,1V.Then, a significant decrease in the 137 current Fig.2A (b) has been observed after the electrochemical grafting of diazonium salt Fig.2B. The 138 electrochemical reduction of the diazonium salt was generated in situ by a reaction of CMA with NaNO₂ in a 139 solution of HCl, with a scanning speed of 80 mV.s⁻¹, and within a potential range of (-0.2 to -1V). In the first 140 cycle, the current variation showed a reduction wave localized at -0.6V. This current corresponds to the 141 reduction of the diazonium salt in solution which leads to the formation of a radical amine. Afterwards, we 142 noticed from the second scan cycle that the reduction current decreases remarkably. This can be attributed to the decrease of the electron transfer rate that was created by the CMA blocking layer. This was confirmed by a 143 144 significant decrease in the current Fig.2A (b). After aptamer immobilization, CV peaks at the electrode have 145 significantly increased compared to CMA modified gold WE electrode Fig.2A(c). This implies that most of the 146 surface of the electrode was covered with aptamer and demonstrates successful aptamer immobilization. Then, 147 the modified gold WE was incubated in a PBS solution containing 1nM cadmium Cd²⁺during 30mn followed by 148 CV measurement Fig.2A (d). Here we have again a decrease of CV peaks which means that the Cd²⁺have been successfully adsorbed onto the gold WE surface. The Cd²⁺ adsorption prevents or decreases the transfer of 149 150 electrons between gold WE surface and the redox couples in solution due to the opening of the rod portion of aptamer and formation of the complex Cd²⁺- aptamer 151



Fig. 2 (A) Cyclic voltammograms of (a) bare gold WE (b) after CMA deposition(c) after aptamer immobilization
 and (d) after the detection for 10⁻⁹M of Cd²⁺ solution.(B) Cyclic voltammograms for electrodeposition of CMA
 for repetitive potential cycles from -0.2V to -1V with a scan rate of 0.8V.s⁻¹

157 Electrochemical impedance spectroscopy studies

158 Impedance spectra were employed to assess the electrochemical behaviour of the aptasensor fabrication and 159 realized in ferric potassium ferrocyanide solution. The Nyquist plot semi-circles allow to assess the transfer 160 charge resistance R_{ct} generated by any modification on the gold WEs surface. The more the surface is passivated, 161 the higher is the resistance R_{ct} value Fig.3. Here, the first Nyquist plot semi-circle Fig.3 (a) corresponds to the 162 bare gold which presents negligible resistance Ret. Then, the formation of an organic layer of CMA onto the gold 163 WE surface has increased the R_{ct} to 25 290 Ω Fig.3 (b). This was due to the CMA layer which decreases the electron transfer. Subsequently, the surface carboxylic groups were activated using carbodiimide chemistry to 164 165 allow the Cd²⁺-aptamer immobilization through the covalent linkage between amino groups of aptamer and the 166 abundant reactive carboxylic groups of CMA. Aptamers immobilization has induced a strong increase in the 167 semi-circular diameter, with an R_{ct} of 28 096 Ω Fig. 3(c). Finally, the developed aptasensor was tested for Cadmium detection by incubating the aptasensor in Cd²⁺ solution at 10⁻⁹Mfor 30 min. Here we observed an 168 169 increase of R_{ct} up to 42 537 Ω Fig.3(d) due to the conformational changes induced by aptamer switching from a 170 random coil structure to an aptamer-target complex[32, 33].



171Re(Z)/KΩ172Fig. 3: Nyquist diagrams of the bare gold electrode (\blacksquare), the electrochemically grafted CMA (\bullet), after deposition173of the aptamer (\blacktriangle) and after incubation of Aptamer/CMA/AU in 10⁻⁹M Cd²⁺ (\blacktriangledown) in the presence of 5 mM of174[Fe(CN)₆]^{3-/4-}solution in a frequency range of 100 mHz - 200 kHz. The potential polarization used for these175measurements was optimized at 128 mV.

176 Optimization of aptamer concentration

177 The aptamer concentration determines its final density onto the gold WE surface [34]. The optimization of aptamer concentration is a crucial step which affects the Cd²⁺ions detection and the analytical performance of the 178 179 aptasensor. Therefore, three different concentrations of aptamer (0.25, 0.5 and 0.75 µM) were tested within the range of 10⁻¹⁰ to 10⁻³M of cadmium by using EIS. Fig. 4 shows the variation of R_{ct} as function of Cd²⁺ 180 181 concentration for the three aptamer concentrations. As it can be observed, when using the aptamer concentration 182 of 0.25 μ M, the Rct did not change significantly, to the weak complexation of the aptamer with its target. This 183 aptamer concentration generated a low signal with a slope of 1.744 k Ω (R²= 0.927). The aptamer concentration 184 of 0.5μ M resulted a narrow range of linearity with a slope of $3.891 \text{ k}\Omega$ (R²=0.911). Finally, by using the aptamer 185 concentration of 0.75 μ M, the signal was improved and the analytical response of the aptasensor was linear within the range of 10^{-9} to 10^{-4} M with a LOD of $2.75*10^{-10}$ M. In addition, the sensitivity increased to 4.683 k Ω 186

187 when compared to that obtained with the lower concentrations.



Log C Fig. 4 Optimization of Cd²⁺-aptamer concentration by EIS; (\blacktriangle) 0.25 µM, (\bullet) 0.5 µM and (\blacksquare) 0.75µM of aptamer.

191 Analytical performance of the aptasensor

Under the optimal conditions, the aptasensor was used for cadmium detection in PBS buffer using EIS. To achieve that, the modified aptamer gold WE surface was incubated with different concentrations of cadmium for 30 minutes. Experimental results showed a significant increase in the diameter of Nyquist plot semi-circles by increasing Cd²⁺ concentration **Fig. 5A**. The first Nyquist plot semi-circle corresponds to the aptamer-modified gold WE before the addition of cadmium. After incubation of aptasensor with increasing concentrations of the target, the Nyquist plot semi-circles continue to increase indicating the interaction between cadmium and its specific aptamer.



199



Fig.5: (A) Nyquist plots of Cd²⁺-aptamer-Gold WE surface incubated with increasing Cd²⁺ concentrations ranging from 10⁻³ to 10⁻¹⁰ M, (B) sensitivity of the aptasensor by normalization of the data towards the varying Cadmium concentrations. Inset: equivalent circuit used for the EIS Nyquist plot semi-circles fitting.

204 The transfer charge resistance R_{ct} was extracted from Nyquist plot semi-circles by using the equivalent 205 circuit in Figure 5B Inset [21]. Where Rs represents the solution resistance, Rct; the charge-transfer resistance 206 and Q1; the constant phase element (an equivalent model of double-layer capacitance). The real part of 207 impedance (Re(Z)) increased as function of cadmium concentration. The linear regression equation was 208 normalized to $\Delta R_{ct} = R_{ct}$ (aptamer-cadmium) – Rct (aptamer) as a function logarithm of Cadmium concentration. Rct (aptamer-cadmium) is the value of the electron transfer resistance after Cd²⁺ binding to the immobilized 209 210 aptamer on the modified electrode surface. Rct (aptamer) represents the value of the electron transfer resistance 211 after the aptamer immobilization on the CMA-modified electrode surface. The resistance of the aptasensor 212 exhibited a linear correlation to Cadmium concentration through the range of (10⁻¹⁰-10⁻⁴M), with a good 213 correlation coefficient of ($R^2 = 0.9954$) and a sensitivity of 4.683k Ω per cadmium concentration decade. The 214 limit of detection (LOD) of the aptasensor is the concentration of cadmium corresponding to the 3 times "s/m" 215 ratio, where "s" is the standard deviation of the blank impedance signal (three replicates) and "m" is the slope of the related calibration curve [35]. Here, the calculated LOD was 2.75×10^{-10} M.The detection limit obtained in this 216 217 work was lower the detection limits of Cd²⁺aptamer strategies and other methods (Table 1).

Table 1: Comparison of other sensors with different detection methods for the determination of Cd^{2+} .

Detection methods	Linear range (nM)	Limit of detection (LOD)	Refs.
		(nM)	
Fluorescent aptasensor	0 to 1000	40	[36]
Fluorescent aptasensor	7.19 to 200	2.15	[37]
Colorimetric aptasensor	10 to 400	4.6	[25]

Electrochemical aptasensor	1to1000	0.337	[38]
Electrochemical aptasensor	250 to 1000	92	[17]
Electrochemical aptasensor	$0.1 \text{to} 10^6$	0.27	This work

236 237

238

The reproducibility of the developed aptasensor was tested with interassay precision. The interassay precision was evaluated by incubating three electrodes, independently prepared under the same experimental conditions with a same cadmium concentration of 10⁻⁶ M. A relative standard deviation from of 4.49% was calculated, indicating acceptable aptasensor reproducibility.

225 For the specificity study, the **Fig.6** depicts the impedance changes after incubation with Cd^{2+} and other 226 heavy metals such as Hg²⁺, Pb²⁺ and Zn²⁺during 30 minutes. Three concentrations have been tested for each 227 analyte; 10-8, 10-7 and 10-6M. At this stage, the interactivity of the aptasensor and cadmium detection was 228 performed only in PBS buffer. As it can be observed in (Fig.6); the impedance responses show clearly that the 229 aptasensor was highly selective to cadmium as compared to the other tested heavy metals. No significant signals have been recorded, even in the presence of Pb²⁺ and Zn²⁺. However, a slight cross-reactivity was noted with 230 Hg²⁺,it may be attributed to the special T-Hg-T mismatched base pair between Hg²⁺ and Cd²⁺- aptamer[39-231 232 41]. These results show that the selected aptamer immobilized on the electrode surface maintains its affinity and 233 specificity in reaction with the cadmium, giving significant Δ Rct increase. This may be considered as a 234 validation of the proposed aptasensor selectivity, because it allows the discrimination between the tested heavy 235 metals into the entire selected range.





240 The developed aptasensor was applied for the detection of Cadmium in water samples of Hammam Essalihine 241 valley using standard addition method as described previously. Four flasks containing the real water samples 242 were prepared as mentioned in experimental, and then analysed by EIS measurements Fig.7 depicts the first 243 Nyquist plot corresponding to the aptamer-modified gold surface before cadmium addition, while the second one 244 is attributed to the aptasensor after being incubated in the first flask (level 1) for 30 min. A shift in the semi-245 circle from the first one was observed, it was explained by the increase of resistance provided from aptamer 246 reaction with cadmium. The second flask, corresponding to level 2, showed the same behaviour as the first flask. 247 The analysis of the other flasks was performed in the same parameters. An increase in the semi-circular diameter 248 was noticed and was absolutely explained by the increase of cadmium amount in the flasks. In order to determine the concentration of Cd^{2+} in the real sample, slope (m) and y-intercept (b) of the resulting calibration curve ($R^2 =$ 249 250 0.987) were used, and the unknown Cd²⁺ concentration of real sample was estimated to be $4.29.10^{-8}$ M.

Table 2: Determination of Cd2+ by proposed aptasensors and atomic absorption spectroscopy (AAS) in water
 samples and recovery study of the aptasensor for determining cadmium

Ion in sample	AAS(C)	The aptasensor(Cs)	Recovery (%)	
Cd2+	4.76*10 ⁻⁸ M	4.29*10-8M	90.12	

253 Recovery% = Cs/C×100%[42]

Aiming to validate our aptasensor, the tested sample was also analyzed by atomic adsorption spectroscopy (AAS) which is a reference method usually used for Cadmium detection[6, 7]. The obtained results are presented in (Table2). They are in a good agreement with those obtained by the standard addition method.



258





Fig. 7 (A) Nyquist impedance plots obtained from the performance application on a real sample. (**n**) PBS; (**n**) Level 1 (corresponding to an addition of 0 M); (**•**) Level 2 (addition of 10^{-8} M); (**•**) Level 3 (addition of 10^{-7} M) and (**▲**) Level 4 (addition of $2*10^{-7}$ M). (B) Sensitivity curve used to calculate the concentration of analyte in the unknown sample by the standard addition method. Error bars represent the relative standard deviation on the measurements performed in triplicate.

266 Conclusion

267 In summary, we have developed an electrochemical aptasensor for Cadmium detection. The aptamer was 268 covalently immobilized onto gold WE through CMA molecules. The studied aptasensor showed a linear 269 relationship between the impedance changes and the logarithm of Cd²⁺concentration in a broad range from 10⁻³ 270 to 10⁻¹⁰ M, with a LOD about 2.75*10⁻¹⁰M. Indeed, the present aptasensor displays excellent sensitivity and 271 selectivity, in addition to a good reproducibility. Moreover, the aptasensor was highly selective to cadmium in the presence of other interferences, such as Hg^{2+} , Pb^{2+} and Zn^{2+} . In order to prove the method applicability, the 272 273 constructed aptamer sensor has been employed to detect cadmium in real spring water samples with high 274 accuracy. These promising results show that the developed method could be an excellent alternative for 275 traditional techniques because of its accuracy, reliability, simplicity and low cost. This miniaturized sensor could 276 be used wide wearable devices in rivers or directly in spring water sources. In addition, the working electrode 277 could be functionalized by other aptamers to allow the concomitant determination of several heavy metals.

278 References

- [1] S.J. Hawkes, What is a" heavy metal"?, Journal of Chemical Education, 74 (1997) 1374.
- 280 [2] P.B. Tchounwou, C.G. Yedjou, A.K. Patlolla, D.J. Sutton, Heavy metal toxicity and the environment,
- 281 Molecular, clinical and environmental toxicology, Springer2012, pp. 133-164.

- [3] A. Turner, Cadmium pigments in consumer products and their health risks, Science of The TotalEnvironment, (2018).
- 284 [4] D.G. Bostwick, H.B. Burke, D. Djakiew, S. Euling, S.m. Ho, J. Landolph, H. Morrison, B. Sonawane,
- T. Shifflett, D.J. Waters, Human prostate cancer risk factors, Cancer: Interdisciplinary International
 Journal of the American Cancer Society, 101 (2004) 2371-2490.
- 287 [5] F. Edition, Guidelines for drinking-water quality, WHO chronicle, 38 (2011) 104-108.
- [6] N.A. Kasa, D.S. Chormey, Ç. Büyükpınar, F. Turak, T.B. Budak, S. Bakırdere, Determination of
 cadmium at ultratrace levels by dispersive liquid-liquid microextraction and batch type hydride
 generation atomic absorption spectrometry, Microchemical Journal, 133 (2017) 144-148.
- [7] K.S. Rao, T. Balaji, T.P. Rao, Y. Babu, G. Naidu, Determination of iron, cobalt, nickel, manganese,
 zinc, copper, cadmium and lead in human hair by inductively coupled plasma-atomic emission
 spectrometry, Spectrochimica Acta Part B: Atomic Spectroscopy, 57 (2002) 1333-1338.
- [8] D. Xu, W. Fan, H. Lv, Y. Liang, Y. Shan, G. Li, Z. Yang, L. Yu, Simultaneous determination of traces amounts of cadmium, zinc, and cobalt based on UV–Vis spectrometry combined with wavelength selection and partial least squares regression, Spectrochimica Acta Part A: Molecular and Biomolecular Spectroscopy, 123 (2014) 430-435.
- [9] L.C.S. Figueiredo-Filho, B.C. Janegitz, O. Fatibelilo-Filho, L.H. Marcolino-Junior, C.E. Banks,
 Inexpensive and disposable copper mini-sensor modified with bismuth for lead and cadmium
 determination using square-wave anodic stripping voltammetry, Analytical Methods, 5 (2013) 202 207.
- 302 [10] Y.-k. Lü, H.-W. Sun, C.-G. Yuan, X.-P. Yan, Simultaneous determination of trace cadmium and
 303 arsenic in biological samples by hydride generation-double channel atomic fluorescence
 304 spectrometry, Analytical chemistry, 74 (2002) 1525-1529.
- 305 [11] H. Kaur, R. Kumar, J.N. Babu, S. Mittal, Advances in arsenic biosensor development–a 306 comprehensive review, Biosensors and Bioelectronics, 63 (2015) 533-545.
- 307 [12] C. Tuerk, L. Gold, Systematic evolution of ligands by exponential enrichment: RNA ligands to
 308 bacteriophage T4 DNA polymerase, science, 249 (1990) 505-510.
- [13] X. Sun, F. Li, G. Shen, J. Huang, X. Wang, Aptasensor based on the synergistic contributions of
 chitosan-gold nanoparticles, graphene-gold nanoparticles and multi-walled carbon nanotubes cobalt phthalocyanine nanocomposites for kanamycin detection, Analyst, 139 (2014) 299-308.
- [14] W. Tan, M.J. Donovan, J. Jiang, Aptamers from cell-based selection for bioanalytical applications,
 Chemical reviews, 113 (2013) 2842-2862.
- [15] K. Sefah, J. Phillips, C. Wu, Cell-Specific Aptamers for Disease Profiling and Cell Sorting,
 Aptamers Selected by Cell-SELEX for Theranostics, Springer2015, pp. 197-213.
- [16] K.-M. Song, S. Lee, C. Ban, Aptamers and their biological applications, Sensors, 12 (2012) 612-631.
- [17] H.R.L.Z. Zhad, Y.M.R. Torres, R.Y. Lai, A reagentless and reusable electrochemical aptamer-based
 sensor for rapid detection of Cd (II), Journal of Electroanalytical Chemistry, 803 (2017) 89-94.
- 320 [18] Y. Guo, Y. Zhang, H. Shao, Z. Wang, X. Wang, X. Jiang, Label-free colorimetric detection of 321 cadmium ions in rice samples using gold nanoparticles, Analytical chemistry, 86 (2014) 8530-8534.
- 322 [19] I.M. Taylor, Z. Du, E.T. Bigelow, J.R. Eles, A.R. Horner, K.A. Catt, S.G. Weber, B.G. Jamieson, X.T.
- Cui, Correction: Aptamer-functionalized neural recording electrodes for the direct measurement of cocaine in vivo, Journal of Materials Chemistry B, 5 (2017) 8417-8417.
- 325 [20] Z. Ling, W. Ming-Hua, W. Jian-Ping, Y. Zhun-Zhong, Application of biosensor surface
 326 immobilization methods for aptamer, Chinese Journal of Analytical Chemistry, 39 (2011) 432-438.
- 327 [21] A. Baraket, M. Lee, N. Zine, N. Yaakoubi, M.G. Trivella, M. Zabala, J. Bausells, N. Jaffrezic-Renault,
 328 A. Errachid, Cytokine detection using diazonium modified gold microelectrodes onto polyimide
 329 substrates with integrated Ag/AgCl reference electrode, 2012.
- 330 [22] L. Barhoumi, A. Baraket, F.G. Bellagambi, G.S. Karanasiou, M.B. Ali, D.I. Fotiadis, J. Bausells, N.
- 331 Zine, M. Sigaud, A. Errachid, A novel chronoamperometric immunosensor for rapid detection of TNF-
- 332 α in human saliva, Sensors and Actuators B: Chemical, 266 (2018) 477-484.

- 333 [23] A. Longo, A. Baraket, M. Vatteroni, N. Zine, J. Baussells, F. Di Francesco, G.S. Karanasiou, D.I.
 334 Fotiadis, A. Menciassi, A. Errachid, Highly sensitive Electrochemical BioMEMS for TNF-α detection in
 335 humansaliva: Heart Failure, Procedia Engineering, 168 (2016) 97-100.
- [24] S. Baranton, D. Bélanger, Electrochemical derivatization of carbon surface by reduction of in situ
 generated diazonium cations, The Journal of Physical Chemistry B, 109 (2005) 24401-24410.
- 338 [25] Y. Wu, S. Zhan, L. Wang, P. Zhou, Selection of a DNA aptamer for cadmium detection based on 339 cationic polymer mediated aggregation of gold nanoparticles, Analyst, 139 (2014) 1550-1561.
- 340 [26] D.M. Jenkins, B. Chami, M. Kreuzer, G. Presting, A.M. Alvarez, B.Y. Liaw, Hybridization probe for
- 341 femtomolar quantification of selected nucleic acid sequences on a disposable electrode, Analytical
- 342 chemistry, 78 (2006) 2314-2318.
- 343 [27] A.B. Kharitonov, L. Alfonta, E. Katz, I. Willner, Probing of bioaffinity interactions at interfaces
 344 using impedance spectroscopy and chronopotentiometry, Journal of Electroanalytical Chemistry, 487
 345 (2000) 133-141.
- 346 [28] A. Hayat, L. Barthelmebs, A. Sassolas, J.-L. Marty, An electrochemical immunosensor based on 347 covalent immobilization of okadaic acid onto screen printed carbon electrode via diazotization-348 coupling reaction, Talanta, 85 (2011) 513-518.
- 349 [29] H. Araar, M. Benounis, A. Direm, A. Touati, S. Atailia, H. Barhoumi, N. Jaffrezic-Renault, A new
- thin film modified glassy carbon electrode based on melaminium chloride pentachlorocuprate (II) for
 selective determination of nitrate in water, Monatshefte für Chemie-Chemical Monthly, 150 (2019)
 1737-1744.
- [30] K. Hanane, B. Messaoud, B. Houcine, T. Moncef, Highly sensitive modified glassy carbon sensor
 based on TDAN for nitrate detection in real water, Monatshefte für Chemie-Chemical Monthly, 151
 (2020) 153-158.
- 356 [31] F.G. Bellagambi, A. Baraket, A. Longo, M. Vatteroni, N. Zine, J. Bausells, R. Fuoco, F. Di Francesco,
 357 P. Salvo, G.S. Karanasiou, Electrochemical biosensor platform for TNF-α cytokines detection in both
- artificial and human saliva: Heart failure, Sensors and Actuators B: Chemical, 251 (2017) 1026-1033.
- [32] Y. Luan, A. Lu, J. Chen, H. Fu, L. Xu, A label-free aptamer-based fluorescent assay for cadmium
 detection, Applied Sciences, 6 (2016) 432.
- [33] Z. Bagheryan, J.-B. Raoof, M. Golabi, A.P. Turner, V. Beni, Diazonium-based impedimetric
 aptasensor for the rapid label-free detection of Salmonella typhimurium in food sample, Biosensors
 and Bioelectronics, 80 (2016) 566-573.
- [34] R.J. White, N. Phares, A.A. Lubin, Y. Xiao, K.W. Plaxco, Optimization of electrochemical aptamer based sensors via optimization of probe packing density and surface chemistry, Langmuir, 24 (2008)
 10513-10518.
- 367 [35] C.M. Riley, T.W. Rosanske, Development and validation of analytical methods, Elsevier1996.
- [36] H. Wang, H. Cheng, J. Wang, L. Xu, H. Chen, R. Pei, Selection and characterization of DNA
 aptamers for the development of light-up biosensor to detect Cd (II), Talanta, 154 (2016) 498-503.
- [37] Y.-F. Zhu, Y.-S. Wang, B. Zhou, J.-H. Yu, L.-L. Peng, Y.-Q. Huang, X.-J. Li, S.-H. Chen, X. Tang, X.-F.
 Wang, A multifunctional fluorescent aptamer probe for highly sensitive and selective detection of
- 372 cadmium (II), Analytical and bioanalytical chemistry, 409 (2017) 4951-4958.
- [38] X. Wang, W. Gao, W. Yan, P. Li, H. Zou, Z. Wei, W. Guan, Y. Ma, S. Wu, Y. Yu, A novel aptasensor
 based on graphene/graphite carbon nitride nanocomposites for cadmium detection with high
 selectivity and sensitivity, ACS Applied Nano Materials, 1 (2018) 2341-2346.
- [39] C.-K. Chiang, C.-C. Huang, C.-W. Liu, H.-T. Chang, Oligonucleotide-based fluorescence probe for
 sensitive and selective detection of mercury (II) in aqueous solution, Analytical chemistry, 80 (2008)
 3716-3721.
- [40] C. Gao, Q. Wang, F. Gao, F. Gao, A high-performance aptasensor for mercury (II) based on the formation of a unique ternary structure of aptamer–Hg 2+–neutral red, Chemical Communications,
- 381 50 (2014) 9397-9400.
- 382 [41] C.-X. Tang, Y. Zhao, X.-W. He, X.-B. Yin, A "turn-on" electrochemiluminescent biosensor for 383 detecting Hg2+ at femtomole level based on the intercalation of Ru (phen) 32+ into ds-DNA,
- 384 Chemical Communications, 46 (2010) 9022-9024.

- 385 [42] B. Dindar, E. Karakuş, F. Abasıyanık, New urea biosensor based on urease enzyme obtained from
- Helycobacter pylori, Applied biochemistry and biotechnology, 165 (2011) 1308-1321.

_ _ _

