



HAL
open science

Passive RF shimming for 7T head coil with hybridized meta-atom: In vivo investigation

Marc Dubois, Lisa Leroi, Zo Raolison, Redha Abdeddaim, Tryfon Antonakakis, Julien de Rosny, Alexandre Vignaud, Pierre Sabouroux, Elodie Georget, Benoit Larrat, et al.

► To cite this version:

Marc Dubois, Lisa Leroi, Zo Raolison, Redha Abdeddaim, Tryfon Antonakakis, et al.. Passive RF shimming for 7T head coil with hybridized meta-atom: In vivo investigation. International Society for Magnetic Resonance in Medicine 28th Annual Meeting, Aug 2020, Sydney, Australia. hal-03164015

HAL Id: hal-03164015

<https://hal.science/hal-03164015v1>

Submitted on 9 Mar 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Passive RF shimming for 7T head coil with hybridized meta-atom: In vivo investigation

Marc Dubois^{1,2}, Lisa Leroi³, Zo Raolison², Redha Abdeddaim¹, Tryfon Antonakakis⁴, Julien de Rosny⁵, Alexandre Vignaud³, Pierre Sabouroux¹, Elodie Georget³, Benoit Larrat³, Gérard Tayeb¹, Nicolas Bonod¹, Alexis Amadon³, Franck Mauconduit³, Cyril Poupon³, Denis Le Bihan³, Stefan Enoch¹

¹ Aix Marseille Univ, CNRS, Centrale Marseille, Institut Fresnel, Marseille, France

² Aix Marseille Univ, CNRS, CRMBM, Marseille, France

³ CEA, DRF, JOLIOT, NeuroSpin, UNIRS, Université Paris-Saclay, Gif-sur-Yvette, France

⁴ Multiwave Imaging, Geneva, Switzerland

⁵ ESPCI Paris, PSL Research University, CNRS, Institut Langevin, Paris, France

* Authors contributed equally

Background

Ultra-high field scanners benefits in terms of signal-to-Noise ratio (SNR) and Contrast-to-Noise ratio (CNR) are undermined by the inhomogeneous propagation of the radiofrequency (RF) field in the human body at 7T. Parallel transmission technique using multi-channels RF coils is being explored to solve this issue ^{1,2}. Our aim is to tackle this hardware problem *via* passive RF shimming, *i.e.* inserting high dielectric constant materials ^{3,4,5} or metamaterial structures ⁶ around the subject to tailor the B_1^+ amplitude. We demonstrated *in vitro* that the hybridization of four parallel metallic wires arranged on a square unit cell provides the ability to control RF field inside a 7T head birdcage equipped with a receive array (1Tx/32Rx - Nova Medical, Wilmington, MA, USA) ⁷. Here, we show the first *in vivo* testing of the hybridized meta-atom (HMA) on two healthy volunteers with safe sequence protocols designed for RF prototype assessment at 7T ⁸.

Methods

Four 1-mm-diameter brass wires are assembled at the four corners of a 1-cm-side square to form the HMA structure. The latter is then placed between the birdcage used for transmission and the receive array. Figure 1 presents a schematics of the experimental setup. MRI acquisitions were performed using a SC72 gradient dedicated 1Tx/32Rx proton head coil (Nova Medical, Wilmington, MA, USA) in a 7T Magnetom MRI scanner (Siemens Healthineers, Erlangen, Germany). Table 1 presents the parameters of the *in vivo* protocol used. It includes a localizer (Proton density weighted), a B_0 mapping and a flip angle (FA) mapping (XFL sequence ⁹). These parameters were enforced such that even if all the input power was focused and absorbed within 10g of tissues, the SAR would not exceed the regulatory limits (global and local SAR). The localizer and FA maps were acquired on three orthogonal slices sequentially. *In vivo* acquisitions have been done under local IRB rules on two healthy volunteers.

Results

Results on the first volunteer are presented in Figure 2. It shows the FA map obtained for the reference scan without HMA in the three orientations (V_{ref} 250V) and the same maps acquired when a 42-cm-long HMA is inserted (V_{ref} 320V). The bottom row presents the relative gain between these two scans in percentage. Within the coronal slice, the right hemisphere of the brain indicates a 28% average gain in FA whereas the left one shows a reduction of 10%. The axial slice located at the eyes level exhibits an increase of +17% on the HMA side and a reduction of 27% on the opposite side. Meanwhile the loss in the cerebellum region is 34% on average. Figure

3 displays the results obtained on the second volunteer. The same methodology is applied, a reference scan was done first (Vref 250V), followed by two scans: one with a 40-cm-long HMA (Vref 275V) and another one with 44-cm-long HMA (Vref 325V) in the same location. Relative FA gain in percentage is presented in comparison with the reference FA map in each case. The relative gain in the coronal slice with 40-cm-long HMA is + 9% (similar for 44cm) on the HMA side and -2% (-9% for 44cm) on the opposite side. The loss in the cerebellum area is around 10% in each case. The effect in the axial slice is more localized than for the previous volunteer and the 40-cm-long HMA shows benefits on a larger brain volume.

Discussions

The relative FA gain observed in Figure 2 is consistent with the observation made on specific anthropomorphic phantom ⁷. While the opposite side of the HMA shows some significant loss, the improvement on the HMA side is significant considering the poor FA distribution in the reference map, as usually encountered at 7T due to the volunteer's head diameter. The second volunteer, displayed in Figure 3 presents a smaller head diameter which yields a more homogeneous FA distribution from the coil alone. In this case, we tested two different HMA lengths. First, we observe that the 40-cm-long HMA only increased the reference voltage by 10% which is positive regarding the transmit efficiency of the coil (FA with respect to the input power). Moreover, the size reduction (44cm to 40cm) does not affect strongly the performances of the HMA in terms of relative FA gain. The results indicate that a reduced length HMA should be a better solution but also show that the relative gain of the HMA depends on the head diameter as the load of the birdcage coil changes.

Conclusions

The HMA solution previously studied *in vitro* was successfully tested on two volunteers. This technique features the ability to enhance locally the FA distribution together with the possibility to fill one of the gap usually observed in the brain temporal lobes. Future work will target the homogenization of FA in the whole brain with possibly the use of multiple elements of various dimensions. Finally, let us stress the benefits of the HMA solution which provides a cost effective, long-lasting, and non-toxic solution without being in direct contact with the subject's body. This last point is crucial for improving the patient's comfort during examination.

Acknowledgements

This work has received funding from the European Union Horizon 2020 Research and Innovation program under Grant Agreement No. 736937. It is also supported by the Leducq Foundation (large equipment ERPT program, NEUROVASC7T project).

References

- [1] Katscher U. *et al*, NMR Biomed 2006; 19: 393–400
- [2] Grissom W. *et al*, Magn. Reson. Med. 2006; 56: 620–629
- [3] A. G. Webb, Concepts Magn. Reson. Part A 2011; 38A, 148
- [4] T. P. A. O'Reilly, A. G. Webb, and W. M. Brink, J. Magn. Reson. 2016; 270, 108-114
- [5] A. L. Neves, *et al*, Magn. Reson. Med. 2018; 79, 1753
- [6] M. Dubois, *et al*, Phys. Rev. X 2018; 8, 031083
- [7] M. Dubois, *et al*, Proc. Intl. Soc. Mag. Reson. Med. 2019; 27, 1491
- [8] A. Vignaud, *et al*, J. Phys.: Conf. Ser. 2018; 1092 012159




Figure 1 – Schematics of the experimental setup. The HMA (four wires) is represented in brown and positioned between the receive helmet and the transmit birdcage as in the experiments. Birdcage shield is not shown.

	Localizer	B ₀ map	FA map
Sequence type	2D TFL	2D GRE	2D XFL
RF pulse type	Sinc	Sinc	Sinc
Duration (μs)	1000	1000	1000
FA (°)	3	5	3
FR saturation	-	-	Rect
Sat duration	-	-	1000
Sat FA (°)	-	-	60
REF voltage	470	470	470
TR (ms)	20000	400	20000
N _y	128	64	64
Multislice mode	Sequential	Sequential	Sequential
Resolution (mm ³)	2x2x4	4x4x4	4x4x4
Slice number	3	5	3
TA (min)	2	2	2

Table 1 – Parameters of the “restricted SAR mode” protocol sequences. REF voltage gives the maximum input allowed by the coil. It is used for the worst case scenario SAR estimation. After a calibration procedure, approximately half of this value was used during the measurements.




Figure 2 – Three orientation slices presenting (top row) the reference FA maps (in degrees) obtained with the coil alone (central row) the FA maps obtained when a 42-cm-long HMA is set. Orange boxes denote the position of the HMA (not to scale), (bottom row) the relative FA gain in percentage after the addition of the HMA in the coil. The reference voltage is increased by 28% for the HMA scan. Black dashed lines denotes the ROI locations.




Figure 3 – Three orientation slices presenting (top row) the reference FA maps (in degrees) obtained with the coil alone (second row) the FA maps obtained when a 40-cm-long HMA is set and (third row) with a 44-cm-long HMA in the same location. Orange boxes denote the position of the HMA (not to scale). The last two rows present the relative FA gain in percentage after the addition of the HMA (40cm and 44cm). The reference voltage was increased by 10% (40cm) and 30% (44cm). Black dashed lines denotes the ROI locations.