

Mitral valve repair based on physical characterization of coaptation forces

Daniel Grinberg, Alexis Bruhat, Pierre-Jean Cottinet, Minh Quyen Le, David

Adams, Anthony Costa

► To cite this version:

Daniel Grinberg, Alexis Bruhat, Pierre-Jean Cottinet, Minh Quyen Le, David Adams, et al.. Mitral valve repair based on physical characterization of coaptation forces. Journal of Thoracic and Cardiovascular Surgery, 2020, 159 (3), pp.e181-e183. 10.1016/j.jtcvs.2019.07.097. hal-03211608

HAL Id: hal-03211608 https://hal.science/hal-03211608

Submitted on 7 Mar 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

Version of Record: https://www.sciencedirect.com/science/article/pii/S0022522319316514 Manuscript_53396cbe79f098e49bb2fc2abb43e748

1	Mitral valve repair based on physical characterization of
2	coaptation forces
3	Daniel GRINBERG, M.D., M.S. ^{123*}
4	Alexis BRUHAT, M.S. ²
5	Pierre-Jean COTTINET, Ph.D. ³
6	Minh Quyen LE, Ph.D. ³
7	David H. ADAMS, M.D. ¹
8	Anthony COSTA, Ph.D. ²
9	-
10	¹ Department of Cardiovascular Surgery – Mount Sinai Hospital – New York, NY, USA
11	² Department of Neurosurgery, Mount Sinai Hospital – New York, NY, USA
12	³ University of Lyon, INSA-Lyon, LGEF, EA682, Villeurbanne, F-69621 France
13	-
14	* Corresponding Author: Daniel GRINBERG, M.D., M.S.
15	Department of cardiovascular surgery at Mount Sinai Hospital
16	Sinai BioDesign lab at Icahn School of Medicine at Mount Sinai
17	1428 Madison Ave, BMC-24
18	New York, NY 10029, USA
19	Phone : +1 646 683 8599
20	daniel.grinberg.pro@gmail.com

21	Conflict of interest statement
22	- Daniel Grinberg
23	 No conflict of interest
24 25	• Received funding from: Federation Française de Cardiologie, Philippe Foundation Foundation Servier Mitral Foundation INSAVALOR
26	- Alexis BRUHAT
27	• No conflict of interest
28	• No Funding
29	- Pierre-Jean COTTINET
30	\circ No conflict of interest
31	o No Funding
32	- Minh-Ouven LE
33	• No conflict of interest
34	• No Funding
35	- David H. ADAMS
36	• No conflict of interest
37	• Patents: The Icahn School of Medicine at Mount Sinai receives royalty
38	payments from Edwards Lifesciences and Medtronic for intellectual property
39	related to development of valve repair rings.
40	• Potentially influence: National Co-PI of the Medtronic APOLLO FDA Pivotal
41	Trial, the NeoChord ReChord FDA Pivotal Trial, The Medtronic CoreValve
42	US Pivotal Trial and the Abbott TRILUMINATE Pivotal Trial.
43	- Anthony COSTA
44	• No conflict of interest
45	• No Funding
46	-
47	And source of funding: Federation Française de Cardiologie, Philippe Foundation, Foundation
48	Servier, Mitral Foundation, INSAVALOR
49	-
50	meeting presentation: mitral conclave
51	-
52	World count: 734
53	-
54	Abbreviations :
55	MVr - mitral valve repair
56	PMR - primary mitral regurgitation
57	MR- Mitral Regurgitation
58	FFT - Fast Fourier Transform

59 **<u>CENTRAL MESSAGE</u>**

- 60 Technological breakthrough allows surgeons to obtain a precise coaptation characterization of
- 61 the mitral valve (measurement of surface and mapping of coaptation forces), in order to assist
- 62 surgeons during mitral valve repair.

CENTRAL PICTURE (GRAPHICAL ABSTRACT)





67 **TEXT:**

Open heart mitral valve repair (MVr) is the gold standard treatment for severe primary mitral 68 regurgitation (PMR). Several approaches have been described to intraoperatively mimic 69 ventricular systole like the "saline test" and the "ink test"¹. Such methods provide only a 70 71 morphologic and approximate assessment. Thus, MVr success today still primarily depends 72 on the expertise of the individual surgeon. Consequently, major heterogeneities are observed 73 globally on the rate of MVr repair, surgical techniques, and outcomes. In Western countries, 74 only two in three PMR patients benefit from MVr, while the others are treated with an implanted prosthesis that is associated with worse outcomes². Furthermore, depending on the 75 series, up to 25 percent of patients develop recurrent Mitral Regurgitation (MR) after a MVr³. 76 77 Among these patients a large majority have no MR at the discharge, highlighting the insufficient power of the echocardiography to predict mid-term repair failure⁴. New tools are 78 79 required to better understand the coaptation behavior and to assist surgeons in order to 80 increase the repair rate and long-term repair success.

81

82 The restoration of a large surface of coaptation is one of the key principles of mitral 83 reconstruction and a high coaptation height is correlated with lower postoperative residual MR. In a previous work, we developed an innovative technique to measure the tension 84 85 applied on chordae during transapical beating-heart implantations of neochordae⁵. Our results 86 revealed a decrease in subvalvular apparatus stress when strong and homogeneous coaptation 87 was restored. Up to now, coaptation forces have been poorly characterized due to challenges 88 in materials and device design (e.g., compactness, flexibility, sensing accuracy), as well as its deployment in complex clinical environments (e.g., beating heart, coaptation zone. 89

In the present work, we have developed a device enabling precise physical and quantitative
characterization of the coaptation (e.g., mapping coaptation forces [perpendicular forces, one
leaflet against the other] and measurement of coaptation surface).

94

95 The tool is composed of three principal parts:

96 - A piezoresistive pressure sensor including 28 strain gages made from resistive ink

97 printed on a flexible polyimide substrate. Its design (total area of 5.8 cm^2 and 150 mm

98 thick comprised of 28 individual 2mm² sensing elements) provides easy insertion

between the two leaflets without changing the natural 3-dimensional coaptation shape,

100 due to the low thickness and stiffness of the sensing element. During the measurement

101 period the sensor sits in the mitral valve orifice and

A support device (holder) enables facile maintenance of the sensor in the mitral valve
 and removal after measurement.

- An acquisition system leads to record and analyze data in real-time. The time domain

105 of 28 signals is then post-treated based a Fast Fourier Transform (FFT). Frequency

analysis of all data is performed in different conditions of mitral valves (i.e. normal orprolapse).

108

The device is specifically designed for application to MVr procedures. It is inserted inside of the heart through the left atriotomy (repair access) during the procedure and positioned between the two leaflets. The device can be kept in the heart when the atriotomy is closed making continuous force measurements in a beating-heart state possible, and removed before the closure of chest without risk of air embolism.

114

The prototype has been iteratively tested on three porcine *ex vivo* total heart models activated by an apical pump (e.g. BioSimulator, Figure 1, video). Initial results achieved precise surface mapping of mitral coaptation forces (Figure 2). The coaptation signal was first recorded in a

normal valve and then in an induced prolapse-state valve after creation of a posterior central prolapse (section of marginal chordae). We demonstrate that the forces recorded were notably higher in the normal valve compared to the prolapsed one under similar hemodynamic conditions (Figure 3). Our tests also demonstrated the limitations of piezoresistive technology (noise due to bending, loss of calibration), and that our chosen BioSimulator is not an ideal testing model for this measurement since the pump ejection generated a functional MR due to annular dilatation.

125

126 A provisional patent based on our novel method has been filed. Further development of the 127 device is in progress as we move our invention toward clinical trials: selection of optimal 128 sensing technologies, optimization of force sensitivity, substrate flexibility and 129 biocompatibility. In the near future, this technique could be useful to assist surgeons during 130 MVr while providing real-time objective measurements in support of the conventional 131 intraoperative morphological assessment. Coaptation forces characterization could also be 132 useful to improve our understanding of physical changes occurring during and after MVr.



- Figure 1- Testing on *ex vivo* beating heart model. The surgeon is holding the device, that was introduced through the posterior wall of the left atrium into the mitral valve. This *ex vivo* model is
- activated by an apical pump.



138

139 Figure 2 - Design of the sensors used (left), and graphical render of mitral coaptation forces

140 mapping (right). Red colors correspond to high forces. Every colored rectangle corresponds to a sensing

141 areal of 1x3mm. This mapping provides a 2-dimensional representation of pressure repartition.



- Figure 3- The mapping of forces was obtained in a normal valve (left) and in a prolapsed valve (right).
- 145 A simple visual observation recalls higher coaptation forces in a normal valve than in a 146 pathological one.
- 147

142

148

149

- 150 Caption for Video : render of mitral coaptation forces mapping in "BioSimulator" model. Red colors
- 151 correspond to high forces. Every colored rectangle corresponds to a sensing area.

153 **BIBLIOGRAPHY**

Anyanwu AC, Adams DH. The intraoperative "ink test": A novel assessment tool in
 mitral valve repair. *J Thorac Cardiovasc Surg.* 2007;133(6):1635-1636.
 doi:10.1016/j.jtcvs.2007.01.035

Chikwe J, Toyoda N, Anyanwu AC, et al. Relation of Mitral Valve Surgery Volume
 to Repair Rate, Durability, and Survival. *J Am Coll Cardiol.* 2017;69(19):2397-2406.
 doi:10.1016/j.jacc.2017.02.026

160 3. Flameng W, Herijgers P, Bogaerts K. Recurrence of mitral valve regurgitation after
161 mitral valve repair in degenerative valve disease. *Circulation*. 2003;107(12):1609-1613.

162 doi:10.1161/01.CIR.0000058703.26715.9D

4. Gasser S, Stumm M Von, Sinning C, Schaefer U, Reichenspurner H, Girdauskas E.
Can We Predict Failure of Mitral Valve Repair ? 2019:1-12.

165 5. Grinberg D, Cottinet P-J, Thivolet S, et al. Measuring chordae tension during
166 transapical neochordae implantation: Toward understanding objective consequences of mitral
167 valve repair. *J Thorac Cardiovasc Surg*. October 2018. doi:10.1016/j.jtcvs.2018.10.029

168

169 **KEYWORDS:**

170 Mitral valve repare, Coaptation surface, beating heat, sensor MAPPING device, force171 measurments

172

173 ACKNOWLEDGMENT:

We would like to thank to INSAVALOR for their BQI funding that allows to perform the exvivo test and sensors development, as well as Federation Francaise de Cardiologie, Philippe
Foundation, Foundation Servier, Mitral Foundation.

