



**HAL**  
open science

## Estimation of cementless femoral stem stability using an impact hammer

Anne-Sophie Poudrel, Hugues Albine Lomami, Giuseppe Rosi, Arnaud Dubory, Charles-Henri Flouzat-Lachaniette, Guillaume Haiat

► **To cite this version:**

Anne-Sophie Poudrel, Hugues Albine Lomami, Giuseppe Rosi, Arnaud Dubory, Charles-Henri Flouzat-Lachaniette, et al.. Estimation of cementless femoral stem stability using an impact hammer. Forum Acusticum, Dec 2020, Lyon, France. pp.1803-1807, 10.48465/fa.2020.1130 . hal-03235369

**HAL Id: hal-03235369**

**<https://hal.science/hal-03235369v1>**

Submitted on 27 May 2021

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# ESTIMATION OF CEMENTLESS FEMORAL STEM STABILITY USING AN IMPACT HAMMER

A.-S. Poudrel<sup>1</sup>, H. Albin Lomami<sup>1</sup>, G. Rosi<sup>2</sup>, A. Dubory<sup>3</sup>, C.-H. Flouzat-Lachaniette<sup>3</sup> and G. Haiat<sup>1</sup>

<sup>1</sup> CNRS, Laboratoire Modélisation et Simulation Multi Echelle, MSME UMR CNRS 8208, Equipe Biomécanique, Université Paris-Est Créteil Val de Marne (UPEC), Faculté des Sciences et Technologie, 61, Avenue du Général de Gaulle, Créteil Cedex 94010, France

<sup>2</sup> Université Paris-Est, Laboratoire Modélisation et Simulation Multi Echelle, MSME UMR 8208 CNRS, Créteil Cedex 94010, France

<sup>3</sup> Hopital Henri Mondor, 50, av. du Mal. de Lattre de Tassigny, 94000 Creteil, France  
anne-sophie.poudrel@u-pec.fr

## ABSTRACT

The success of cementless hip arthroplasty depends on the primary stability of the femoral stem (FS). It remains difficult to assess the optimal number of impacts to guarantee the FS stability while avoiding bone fracture. The aim of this study is to validate a method using a hammer instrumented with a force sensor to monitor the insertion of the FS in bovine femoral samples. Cementless FS of different sizes were impacted into five bovine femur samples, leading to 99 configurations. Three methods were used to quantify the insertion endpoint of the FS: the impact hammer, video motion tracking and the surgeon proprioception. For each configuration, the number of impacts performed by the surgeon until he felt a correct insertion was noted  $N_{surg}$ . The insertion depth  $E$  was measured through video motion tracking, and the impact number  $N_{vid}$  corresponding to the end of the insertion was estimated. Two indicators, noted  $I$  and  $D$ , were determined from the analysis of the time variation of the force, and the impact number  $N_d$  corresponding to a threshold reached in  $D$  variation was estimated. The pullout force of the FS was significantly correlated with  $I$  ( $R^2=0.81$ ). The values of  $N_{surg}$ ,  $N_{vid}$  and  $N_d$  were similar for all configurations. The results validate the use of the impact hammer to assess the primary stability of the FS and the moment when the surgeon should stop the impaction procedure for an optimal insertion, which could lead to the development of a decision support system.

## 1. INTRODUCTION

Cementless hip arthroplasty aims at replacing defective hip joints. The femoral stem (FS) primary stability is an important factor of surgical success. Nevertheless, it remains difficult for the surgeon to assess the optimal number of impacts to guarantee the FS stability while avoiding bone fracture.

A few quantitative methods aiming at assessing the insertion of the FS into the host bone have been reported in the literature. One of them consists in using vibrational analysis to follow the FS insertion [1–3]. Such approach has been validated in vitro and in vivo [4] but it remains difficult to implement because of the device is in contact

with bone tissue. Another diagnostic method based on acoustic analysis has been developed to monitor the FS insertion by studying the resulting sound produced by the hammer blows on the ancillary during the implant insertion process (5,6). This method has the advantage to be non-contact based but it is influenced by the environmental noise and the operational variability. Further, a study was performed to evaluate modal features as potential fixation predictors to assess cementless implant fixation [7], [8]. An *in vitro* study has been made on artificial bone but more work is needed to validate the method in a situation closer the operative room [9].

Another diagnostic method based on the analysis of the variation of the force applied between an instrumented hammer and the ancillary during impacts was developed by our group [10–12]. This method was first validated *in vitro* [13], [14] and on anatomical subjects for the acetabular cup implant (ACI) primary stability estimation [15]. More recently, the same technique using an instrumented hammer was applied to follow the FS insertion in bone mimicking phantoms [16] and with anatomical subjects [17].

The aim of the present study is to validate this method to FS insertion in bovine femoral samples. The signal processing technique aims at quantifying the insertion endpoint of the FS using three methods: i) the instrumented hammer, ii) video motion tracking and iii) the surgeon proprioception. To do so, the instrumented hammer was used to insert different sizes of FS in five bovine femurs. A system using optical markers was used to follow the FS insertion with video motion tracking (VMT) and the results were compared to the FS insertion endpoint estimated by the surgeon using his proprioception.

## 2. MATERIALS AND METHODS

### 2.1 Implants and specimens

Cementless femoral stems made of titanium alloy (TiAl6Al4V) and coated by hydroxyapatite (CERAFIT R-MIS, Ceraver, Roissy, France) and their corresponding rasps were used. Seven sizes were tested, from size 7 to 13. Five bovine femurs were collected and were prepared similarly to [16]. A home-made ancillary was screwed

into the FS so as to obtain a rigid bilateral fixation between the two parts.

## 2.2 Experimental set-up and insertion protocol

The femoral stem was inserted in the femur with a 1.3 kg hammer equipped with a dynamic piezoelectric force sensor (208C05, PCB Piezoelectronics, Depew, New York, USA) screwed in the center of the impacting face as in [16], [17]. The variation of the force  $s(t)$  applied to the ancillary was recorded by the force sensor. Video tracking was carried out to follow the insertion of the FS into the femoral shaft during the impaction procedure. The relative displacement of the FS compared to the bone was calculated with the software Tracker (Cabrillo College, Aptos, CA, USA). A schematic representation of the experimental set-up is shown in Fig. 1.




Figure 1 : Schematic representation of the experimental set-up.

Two series of experiments were performed by a trained orthopedic surgeon. In the first series (similar to what was done previously in [16]) all FS were inserted in a cavity that was reamed with the size of the rasp corresponding to the FS, while in the second series which aimed at obtaining relatively unstable configurations, the size of the FS was smaller than the rasp size used to ream the cavity. Four bovine samples were considered for the first series, while only one bovine sample was used for the second series. All FS were inserted in a cavity created in each bovine bone with the smallest available rasp (size  $n=7$ ). The impaction procedure was performed: the FS - ancillary system was impacted with the instrumented hammer until the surgeon considered that the FS was stable. Then, 12 additional impacts were carried out in order to assess whether the implant could be further inserted into the femoral shaft, and to determine the average value of the indicator  $I$  obtained for stable conditions of the FS. The aforementioned procedure was repeated four times with the same implant as far as i) the sample was not fractured and ii) the surgeons felt that it was possible to obtain a good FS stability. Then, the surgeon checked if the femur was fractured. If so, the tests were stopped for the sample. If not, a bigger size of rasp and the corresponding FS was used. Eventually, the implant was pulled out axially and the pullout force  $F$  was measured.

## 2.3 Signal processing and video motion tracking

A dedicated signal processing method was developed to analyze the force signal  $s(t)$  as illustrated in Fig. 2. Two specific indicators were employed. First, a specific indicator noted  $D$  and given in ms based on the time difference between the first and second local maxima of the force signal was calculated for each impact as follow :

$$D = f_2(s) - f_1(s) \quad (1)$$

The second indicator, noted  $I$  and given in kN, corresponds to the impact momentum, and was defined similarly to our previous works [16]:

$$I = \frac{1}{t_2 - t_1} \int_{f_1(s)+t_1}^{f_2(s)+t_2} s(t) dt \quad (2)$$




Figure 2 Illustration of a signal  $s(t)$  corresponding to the variation of the force

The third indicator, noted  $E$  and given in cm based on VMT was defined by the relative motion of the FS compared to the femoral bone in relation to the initial position following:

$$E(i) = d(0) - d(i) \quad (3)$$

## 2.4 Post processing and data analysis

Three different analyses were performed for each insertion procedure in order to estimate the required number of impacts to reach the insertion endpoint, corresponding to the end of the migration phase of the FS into the femur. The first method consists in following a parameter  $N_d$  defined as the number of the first impact for which the indicator  $D$  respects :

$$D(i) < D_{th} \quad (4)$$

where  $D_{th}$  is a threshold defined empirically to 0.47 ms.

The second method consists in defining a parameter  $N_{vid}$  defined as the lowest impact number for which the indicator  $E$  respects:

$$E(N_{vid}) \geq E_m - \delta \times E_{sd} \quad (5)$$

where  $E_m$  and  $E_{sd}$  are respectively the average and the standard deviation of the values of  $E$  for the last ten

impacts and  $\delta$  is a parameter empirically chosen equal to 7.

The third method is based on the surgeon proprioception as in clinic: when the surgeon felt that the insertion endpoint of the FS was reached, the parameter  $N_{surg}$  was noted. In order to determine the accuracy of the three methods presented in 2.4, it was necessary to quantify the difference obtained between the different parameters values  $N_d$ ,  $N_{vid}$ , and  $N_{surg}$ . To do so, we defined three parameters as follow:

$$M_d = N_d - N_{surg} \quad (6)$$

$$M_{vid} = N_{vid} - N_{surg} \quad (7)$$

$$M_c = N_{vid} - N_d \quad (8)$$

### 3. RESULTS

A total of 99 configurations were tested. Tab. 1 shows an example of the number of impaction procedures considered for each bovine bone sample for the first series of experiments were stable configurations were considered.

Implant size	Sample n°1	Sample n°2	Sample n°3	Sample n°4	Total
7	4	4	4	4	16
8	3	4	3	4	14
9	4	4	4	4	16
10	4	4	X	4	12
11	1	X	X	4	5
12	X	X	X	4	4
13	X	X	X	4	4
<b>Total</b>	16	16	11	28	71

Table 1 Number of configurations set up considered in the first series of experiments

Figure 3 shows the variation of the two parameters  $D$  and  $E$  as a function of the impact number  $i$  obtained for the configuration sample#4, implant size 9 and test #3. The horizontal back dashed line represents a threshold penetration for indicator  $E$  and the horizontal gray dashed line represents the threshold  $D_{th} = 0.47$  ms. The vertical dashed lines represent the different parameters values,  $N_d$ ,  $N_{vid}$ , and  $N_{surg}$  corresponding to three values of insertion endpoint obtained for each insertion procedure using the instrumented hammer, the video motion tracking and the surgeon proprioception, respectively.




Figure 3 Variation of  $D$  (gray line) and  $E$  (black line) as a function of the impact number for sample#4, implant size 9 and test #3

Figure 4 represents the distribution of the values of  $M_d$ ,  $M_{vid}$  and  $M_c$ , corresponding to the difference between the different methods developed for the estimation of the FS insertion. The average and standard deviation values of  $M_d$ ,  $M_{vid}$ , and  $M_c$  obtained for the 99




Figure 3 Distribution of the values obtained for  $M_d$ ,  $M_{vid}$  and  $M_c$

show that the three indicators are in the same range of variation.

Finally, a significant correlation is obtained between  $I_m$  and  $F$  and the determination coefficient  $R^2$  is equal to 0.81. The error bars correspond to the standard deviation  $I_{sd}$  obtained for each configuration.

### 4. DISCUSSION

The decrease of  $D$  as a function of the number of impacts can be explained to the increase of the bone-implant contact ratio during the insertion of the FS, resulting in an increase of the stiffness of the bone-implant system that may in turn explain the increase of its resonance frequency. Note that other authors also found a similar evolution of the resonance frequency during the insertion, with a comparable approach, using force signal tracking.

Figure 4 shows a good agreement between the three methods developed to assess the number of impacts required to reach the FS insertion endpoint. Again, these results are consistent with the previous in vitro results [16]. However, there remain some discrepancies between the methods. The average value of  $M_c$ , equal to  $-0.19$ , shows a good agreement between the results obtained with VMT ( $N_{vid}$ ) and the instrumented hammer ( $N_d$ ). However, the average values of  $M_{vid}$  ( $-1.17$ ) and  $M_c$  ( $-0.98$ ) indicate that the surgeon evaluation of the insertion endpoint is assessed in average one impact after the results obtained with VMT and the instrumented hammer. Several factors can explain these differences. First, the surgeon proprioception is mainly qualitative and depends on the surgeon's tactile feelings, hearing and vision. Second, errors may occur using VMT due to i) changes of angular position, ii) possible macroscopic 3D movements and plastic bone deformation, iii) errors based of the image processing, such as the precision of markers tracking, which can be observed in the fluctuation of  $E$  indicator even after reaching stability (Fig. 3).

This study presents several limitations. First, the biomechanical properties of the pelvic bone and of

bovine femoral bone are different. Second, the multiple use of the bovine specimen is not representative of clinical situations and could lead to changes of bone properties. Third, our study only focused on one type of FS (CERAFIT R-MIS) and ancillary and further studies should be performed to evaluate the method for implants manufactured by others.

## 5. CONCLUSION

This study is the first *ex vivo* validation of the use of an instrumented hammer to assess the insertion endpoint of the FS as well as its primary stability. The results are in agreement with a previous *in vitro* study. The time variation analysis of the force applied by the hammer on the ancillary during the impacts allows the assessment of the number of impacts required for obtaining a satisfactory press-fit condition into the host femoral bone. Further work will focus on the adaptation of this method to clinical conditions, in order to develop a medical device that could help the surgeons estimating the FS stability, with minimal changes of the surgical protocol. The reader is referred to [18] for further details on the present study.

## 6. REFERENCES

- [1] L. C. Pastrav, J. Devos, G. Van der Perre, et S. V. N. Jaecques, « A finite element analysis of the vibrational behaviour of the intra-operatively manufactured prosthesis-femur system », *Med Eng Phys*, vol. 31, n° 4, p. 489-494, mai 2009, doi: 10.1016/j.medengphy.2008.11.017.
- [2] M. Lannocca, E. Varini, A. Cappello, L. Cristofolini, et E. Bialoblocka, « Intra-operative evaluation of cementless hip implant stability: A prototype device based on vibration analysis », *Medical Engineering & Physics*, vol. 29, n° 8, p. 886-894, oct. 2007, doi: 10.1016/j.medengphy.2006.09.011.
- [3] L. C. Pastrav, S. V. N. Jaecques, M. Mulier, et G. Van Der Perre, « Detection of the insertion end point of cementless hip prostheses using the comparison between successive frequency response functions », *J Appl Biomater Biomech*, vol. 6, n° 1, p. 23-29, avr. 2008.
- [4] L. C. Pastrav, S. V. N. Jaecques, I. Jonkers, G. V. der Perre, et M. Mulier, « In vivo evaluation of a vibration analysis technique for the per-operative monitoring of the fixation of hip prostheses », *J Orthop Surg Res*, vol. 4, p. 10, avr. 2009, doi: 10.1186/1749-799X-4-10.
- [5] Q. Goossens *et al.*, « Acoustic Analysis to Monitor Implant Seating and Early Detect Fractures in Cementless Tha: An in Vivo Study », *J. Orthop. Res.*, août 2020, doi: 10.1002/jor.24837.
- [6] Q. Goossens *et al.*, « Development of an acoustic measurement protocol to monitor acetabular implant fixation in cementless total hip Arthroplasty: A preliminary study », *Medical Engineering & Physics*, vol. 49, p. 28-38, nov. 2017, doi: 10.1016/j.medengphy.2017.07.006.
- [7] P. Henys, L. Capek, J. Fencel, et E. Prochazka, « Evaluation of acetabular cup initial fixation by using resonance frequency principle », *Proc Inst Mech Eng H*, vol. 229, n° 1, p. 3-8, janv. 2015, doi: 10.1177/0954411914561485.
- [8] P. Henyš *et al.*, « Modal frequency and shape curvature as a measure of implant fixation: A computer study on the acetabular cup », *Medical Engineering & Physics*, vol. 60, p. 30-38, oct. 2018, doi: 10.1016/j.medengphy.2018.07.003.
- [9] P. Henyš et L. Čapek, « Impact Force, Polar Gap and Modal Parameters Predict Acetabular Cup Fixation: A Study on a Composite Bone », *Ann Biomed Eng*, vol. 46, n° 4, p. 590-604, avr. 2018, doi: 10.1007/s10439-018-1980-3.
- [10] V. Mathieu *et al.*, « Variation of the impact duration during the *in vitro* insertion of acetabular cup implants », *Medical Engineering & Physics*, vol. 35, n° 11, p. 1558-1563, nov. 2013, doi: 10.1016/j.medengphy.2013.04.005.
- [11] A. Michel *et al.*, « Finite element model of the impaction of a press-fitted acetabular cup », *Med Biol Eng Comput*, vol. 55, n° 5, p. 781-791, mai 2017, doi: 10.1007/s11517-016-1545-2.
- [12] R. Bosc *et al.*, « Influence of soft tissue in the assessment of the primary fixation of acetabular cup implants using impact analyses », *Clinical Biomechanics*, vol. 55, p. 7-13, juin 2018, doi: 10.1016/j.clinbiomech.2018.03.013.
- [13] A. Michel, R. Bosc, R. Vayron, et G. Haiat, « In Vitro Evaluation of the Acetabular Cup Primary Stability by Impact Analysis », *Journal of Biomechanical Engineering*, vol. 137, n° 3, p. 031011, mars 2015, doi: 10.1115/1.4029505.
- [14] A. Michel, R. Bosc, F. Sailhan, R. Vayron, et G. Haiat, « Ex vivo estimation of cementless acetabular cup stability using an impact hammer », *Medical Engineering & Physics*, vol. 38, n° 2, p. 80-86, févr. 2016, doi: 10.1016/j.medengphy.2015.10.006.
- [15] A. Michel, R. Bosc, J.-P. Meningaud, P. Hernigou, et G. Haiat, « Assessing the Acetabular Cup Implant Primary Stability by Impact Analyses: A Cadaveric Study », *PLoS ONE*, vol. 11, n° 11, p. e0166778, nov. 2016, doi: 10.1371/journal.pone.0166778.
- [16] A. Tijou *et al.*, « Monitoring cementless femoral stem insertion by impact analyses: An *in vitro* study », *Journal of the Mechanical Behavior of Biomedical Materials*, vol. 88, p. 102-108, déc. 2018, doi: 10.1016/j.jmbbm.2018.08.009.
- [17] A. Dubory, G. Rosi, A. Tijou, H. A. Lomami, C.-H. Flouzat-Lachaniette, et G. Haiat, « A cadaveric validation of a method based on impact analysis to monitor the femoral stem insertion », *J Mech Behav Biomed Mater*, vol. 103, p. 103535, nov. 2019, doi: 10.1016/j.jmbbm.2019.103535.
- [18] H. Albini Lomami *et al.*, « Ex vivo estimation of cementless femoral stem stability using an instrumented hammer », *Clinical Biomechanics*, vol. 76, p. 105006, juin 2020, doi: 10.1016/j.clinbiomech.2020.105006.

