

MicroMator: Open and Flexible Software for Reactive Microscopy

Zachary Fox, Steven Fletcher, Achille Fraisse, Chetan Aditya, Sebastián Sosa-Carrillo, Sébastien Gilles, François Bertaux, Jakob Ruess, Gregory Batt

▶ To cite this version:

Zachary Fox, Steven Fletcher, Achille Fraisse, Chetan Aditya, Sebastián Sosa-Carrillo, et al.. Micro-Mator: Open and Flexible Software for Reactive Microscopy. 2021. hal-03261134v1

HAL Id: hal-03261134 https://hal.science/hal-03261134v1

Preprint submitted on 29 Jun 2021 (v1), last revised 15 Jun 2022 (v2)

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

1	
2	MicroMator: Open and Flexible Software for Reactive Microscopy
3	
4	Zachary R Fox ^{1,2,3,*} , Steven Fletcher ^{1,2,*} , Achille Fraisse ^{2,1,§} , Chetan Aditya ^{1,2,§} , Sebastián Sosa-
5	Carrillo ^{1,2} , Sébastien Gilles ⁴ , François Bertaux ^{2,1} , Jakob Ruess ^{1,2} , Gregory Batt ^{1,2}
6	¹ Inria Paris, 2 rue Simone Iff, 75012 Paris, France
7	² Institut Pasteur, 28 rue du Docteur Roux, 75015 Paris, France
8	³ Center for Nonlinear Studies, Los Alamos National Laboratory, Los Alamos, NM, USA
9	⁴ Inria Saclay – Ile-de-France, 1 rue H. d'Estienne d'Orves, 91120 Palaiseau, France
10	*These authors contributed equally to this work
11	[§] These authors contributed equally to this work
12	
10	

13 Abstract

Microscopy image analysis has recently made enormous progress both in terms of accuracy and speed thanks to machine learning methods. This greatly facilitates the online adaptation of microscopy experimental plans using real-time information of the observed systems and their environments. Here we report MicroMator, an open and flexible software for defining and driving reactive microscopy experiments, and present applications to single-cell control and single-cell recombination.

19

20 Introduction

21 Software for microscopy automation are essential to support reproducible high-throughput 22 microscopy experiments¹. Samples can now be routinely imaged using complex spatial and temporal 23 patterns. Yet, in the overwhelming majority of cases, executions of experiments are still cast in stone 24 at the beginning, with little to no possibility for human or computer-driven interventions during the 25 experiments. This is all the more surprising given that image analysis has recently made a giant leap 26 in terms of accuracy and rapidity thanks to deep learning methods, thus opening the way for 27 implementing elaborate protocols. Software empowering microscopy with real-time adaptation 28 capabilities is needed to exploit the full potential of automated microscopes.

Several dedicated microscopy software solutions have been developed for applications requiring real-29 30 time analysis. This is notably the case for the efficient scanning of large and complex microscopy samples (eq, Refs^{2–6}). For other important applications, such as real-time control of cellular processes 31 (eg, Refs⁷⁻¹³), results are generally obtained using ad hoc software solutions. Very few generic tools 32 have been developed so far to facilitate the realization of complex, reactive microscopy experiments. 33 One notable exception is Pycro-Manager¹⁴. This powerful framework is built on top of µManager, a 34 35 widely-used software^{15,16} controlling a large range of microscopy hardware. In Pycro-Manager, reactive protocols are built from the ground up. Whereas this gives maximal flexibility, this also 36 37 increases the difficulty to rapidly design experiments, especially for non-expert users. Moreover, no in-depth case studies demonstrating its practical applicability – and showing possible limitations – have 38 39 been reported so far. One can also mention Cheetah, a simple to use Python library to support the 40 development of real-time cybergenetic control platforms that combines microscopy imaging and 41 microfluidics control¹⁷. In its current state, the possibilities to programmatically control the 42 microscope appear limited.

43 MicroMator Software

44 In this paper, we present MicroMator, a simple software solution supporting reactive microscopy 45 experiments, and demonstrate its potential via two challenging case studies. In MicroMator, events play a fundamental role (Fig 1A). They consist of Triggers and Effects. They can be defined by the user 46 in a flexible manner. Examples of triggers include "at the 10th frame", "if more than 100 cells are in 47 the field of view", "if the fluorescence of the 3rd newborn cell exceeds 100 a.u." and "if message 48 lupdate position=10 frame=last is received from Discord". Examples of effects include changing a 49 50 microscope configuration, sending light in the field of view with a given pattern, actuating a 51 microfluidic pump, or starting an optimization routine. Microscopy experiments are defined by a main 52 image acquisition loop, that serves as a backbone for the experiment, and by event creation functions (Fig 1B). Naturally, the main acquisition loop itself can be modified by event effects in the course of 53 54 the experiment. MicroMator is written in Python 3, is open-source and has a modular design. For 55 controlling hardware, MicroMator primarily uses the powerful Python API of µManager pymmcore, 56 but can also use other dedicated Python or web-based APIs provided by vendors, as done for our 57 CellAsic ONIX microfluidic platform. Various types of analysis can be performed using dedicated 58 software modules, such as on-line image analysis or real-time control and optimization. 59 Communication modules can be used to interface MicroMator with digital distribution platforms such 60 as Discord to track experiment progress and potential issues. Lastly, MicroMator leverages Python's 61 multiprocessing module to perform computations concurrently and possesses an extensive and 62 customizable logging system, gathering logs of all modules in a unique file and fostering reproducible 63 research (see SI Text).

64 We also provide SegMator, a software which uses U-Net for bright-field yeast cell segmentation 65 (provided by DeLTA¹⁸), and tracking using TrackPy¹⁹. U-Net is a convolutional neural network with a 66 structure that excels at image segmentation²⁰. U-Net can analyze dense fields of cells in a few seconds 67 and with good accuracy (see SI Text and Fig S1 and S2 and Movie S1).

68 To showcase the full potential of reactive experiments performed with MicroMator, we designed 69 experiments in which cellular processes are controlled in real-time. Single-cell stimulations are 70 computed on line based on the cell state and/or position, demonstrating that reactive loops can be implemented at the level of individual cells. These experiments are inspired by previously-published 71 studies^{10,11,13,21} and show how these could be repeated and further extended using generic software. 72 73 We also provide a tutorial application in which cells with fluorescent proteins are imaged with 74 increasing durations such that the measured intensity reaches a given threshold (Supplementary Text 75 1). This could typically be used to guarantee a good signal to noise ratio irrespectively of the initial 76 fluorescence of the cells.



Figure 1. MicroMator overview. A. Modular software architecture. MicroMator consists of an extensible set of
modules that control various hardware and software aspects of microscopy experiments and of a central core
that handles user-defined events. It is written in the high-level programming language Python. B. Event-based
reactive microscopy workflow. Imaging can be followed by online analysis of the samples. This typically involves
segmentation, tracking, quantification of cell properties, and possibly advanced additional computations. Effects
may then be triggered based on the result of the analysis. These may include the physical actuation of the
hardware or the initiation of communications or of additional computations.

85 Model predictive control of gene expression at the single-cell level in yeast

86 For our first application, we use the EL222 optogenetic system and the mScarletl fluorescent reporter 87 to engineer light-responsive yeast cells (Fig 2A). Using real-time imaging, segmentation and cell 88 tracking, different cells can be stimulated differently in the field of view using a digital micromirror device (DMD). Our goal is to implement different model predictive control (MPC) strategies for 89 controlling the expression levels of a protein in a cell population. The cellular response of our 90 91 engineered cells was characterized for different light stimulation profiles in the same experiment (Fig 92 2B). Only the most central part of the cell receives significant light stimulation. This erosion of the 93 stimulation region helps improving the precision of single-cell light stimulations in dense cell regions 94 because of illumination bleed-through of DMD systems (see SI Text and Fig S4). We then developed 95 and calibrated an "average cell" (deterministic) and a "single cell" (stochastic) model of light-driven 96 gene expression (SI Text and Fig S5, and S6).

97 In open loop control, the average cell model is used to precompute a temporal pattern of light 98 stimulation so that cells follow a target behavior. This light pattern is then applied to all cells in the 99 field of view (Fig 2C and Movie S2). In closed loop population-based control, the average cell model 100 and the average of the measured fluorescence of cells are used by classical state estimators and model 101 predictive controllers to compute in real-time the appropriate light stimulation to drive the mean 102 fluorescence to its target (Fig 2D and Movie S3). Finally, in closed loop single-cell control, a stochastic 103 model of gene expression and single-cell fluorescence measurements are used by advanced state 104 estimators and controllers to compute in real-time the appropriate light stimulations to drive the 105 fluorescence of each and every cell in the field of view to its target (Fig 2E and Movie S4). This control problem is quite challenging and needs to be solved for hundreds of cells in parallel. Advanced 106 107 methods for numerical simulation and state estimation were essential (see SI Text and Fig S7).

- 108 Defining control performance as the time averaged deviation to target, we found that the single-cell
- 109 control method leads to a modest reduction of error of the population averaged fluorescence but to
- a drastic improvement of the average error of the single-cell fluorescence (Fig 2F).



111

112 Figure 2. Control gene expression at the single cell level in yeast. A. The red fluorescent protein mScarletl is 113 placed under the control of the light-responsive transcription factor EL222. B. To efficiently characterize cell 114 responses to light stimulations, cells in the field of view are partitioned in 3 groups, each group being stimulated 115 with a different temporal profile. Bright-field images are segmented and cells are tracked. Then, based on their 116 groups, cells are stimulated during the appropriate time with eroded masks. The temporal evolution of the 117 mean mScarletl fluorescence of the cells in the three groups is shown with envelopes indicating one standard 118 deviation. C. Open-loop control experiment in which a model of the response of the cell population is used to 119 precompute a light stimulation profile that drives the cell population to the target behavior. The application of 120 the light profile leads to significant deviations from the target of the individual cell trajectories. **D.** Closed-loop 121 control experiment in which the same model is used jointly with real-time observations of the population state 122 to decide which light profile to apply to all cells, using a receding horizon strategy. E. A stochastic model of 123 individual cell response is used jointly with single-cell observations to decide which light profile to apply to each 124 cell. **F.** The different strategies have similar performances to drive the mean fluorescence to its target, but the 125 single-cell feedback strategy is significantly better to drive individual cells to their target profiles.

126 Patterns of recombined yeast cells

For our second application, we constructed a light-driven artificial recombination system in yeast and
 employed different light stimulation strategies to obtain various structures of recombined cells.

We again used the EL222 optogenetic induction system but this time to drive the expression of the Cre recombinase. The Cre recombinase induces the expression of a fluorescent reporter, mCerulean, via an amplification step using the ATAF1 transcription factor (Fig 3A & 3B). This strain has been designed as in Ref ²².

133 Firstly, we applied a ring-like recombination signal. More specifically, every cell that was in the 134 designated zone at any moment throughout the experiment has been targeted for recombination 135 (Movie S5). As a result, we did obtain a ring-like pattern of recombined cells (Fig 3C). Experimental 136 and biological limitations can be revealed by the analysis of the tails of the distributions of the 137 recombination readout (i.e., mCerulean fluorescence) within the cell populations (Fig 3C). For 138 example, we found that some cells have been erroneously targeted for recombination because of 139 tracking issues, and that only a few cells have not shown the recombined phenotype at the end of the 140 experiment despite having being effectively targeted for recombination (Fig 3C and S9).

141 Secondly, we tried to create islets of recombined cells. To this end, we dynamically searched for cells 142 that were far from any previously-targeted cell, and targeted these cells for recombination. To 143 maximize the chances that the chosen cells do recombine, we tracked each chosen cell and targeted 144 it repeatedly with light stimulations (Movie S4). Our strategy was effective in creating isolated micro-145 colonies of recombined cells (Fig 3D). Analysis of the lineage trees of targeted cells and non-targeted 146 cells confirmed that recombined cells have a slow growth phenotype. Previous works demonstrating optogenetically-driven recombination use static masks for light targeting²¹. Obtaining single-cell 147 148 resolution as demonstrated in Fig 3D necessitates real-time image analysis and the use of reactive 149 software.





151 Figure 3. Patterns of recombined yeast cells. A. Upon light exposure, the Cre recombinase is expressed and 152 triggers recombination, leading to the expression of ATAF1 and then of mCerulean-Far1M. Stars indicate nuclear 153 localization of the protein. B. Targeted cells are stimulated for 1 second every 6 minutes until the end of the 154 experiment. Fluorescence levels emitted by targeted cells can be recorded. At the end of the experiment, all 155 cells are imaged and a recombined or non-recombined phenotype is attributed. C. A ring-like region in the field 156 of view is selected at the beginning of the experiment and all cells entering the designated region at some time 157 point are targeted for recombination. The distributions of the fluorescence levels of the targeted and non-158 targeted cells can be computed at the end of the experiment. The vast majority of cells present the expected 159 phenotype and outliers can be further analyzed. D. Cells are dynamically selected such that no target cells are 160 close to each other. Cell lineages of targeted and non-targeted cells can be manually reconstructed and statistics 161 can be extracted.

162 Discussion

163 We presented MicroMator together with two challenging applications demonstrating how this 164 software can help using automated microscopy platforms to their full potential. Each application goes beyond the state of the art. We showed the first demonstration of control of protein expression at 165 166 the single cell level in dense field of cells. This requires to jointly solve two challenges, namely 167 obtaining sufficiently precise single-cell stimulations with DMDs and segmenting and tracking cells

168 with sufficient accuracy over extended durations. We also provide the first demonstration of cell 169 recombination targeted at the single-cell level, enabling single-cell resolution patterns. In comparison 170 with Pycro-Manager, MicroMator uses the Micro-Manager GUI to create a main acquisition backbone 171 for the experiment and reactive events are then used to enhance or even dynamically modify this 172 initial plan. Events are created by default as separated threads and an extended logging system 173 gathers messages from all modules that might be running in parallel. This structure provides 174 robustness to real-time issues and facilitates error identification, two critical aspects for developing 175 long and complex experiments.

176 Yet, we foresee that reactiveness in microscopy will primarily be used to enhance and automate 177 classical experiments. Examples of simple use cases abound: triggering autofocus only when needed, 178 dynamically adjusting the imaging condition to the signal strength, identifying novel regions of 179 interest, or following the course of experiments via easily accessible online services (e.g. warning 180 messages sent on Discord), to provide but a few examples. Thanks to its modular nature and to its use 181 of a simple but powerful event system, MicroMator capacities can be conveniently expanded to drive 182 novel hardware or perform a wide range of analyses. MicroMator is a relatively simple software 183 extension that significantly empowers laboratory equipment that is present in most quantitative 184 biology laboratory worldwide.

185 Methods

186 Software and data availability

187 MicroMator is an open-source software. It contains a core part and an extensible list of modules. The 188 MicroMator core manages the user-specified events and also the metadata and logging system. The current list 189 of modules includes a Microscope Controller module, an Image Analysis module, a Model Predictive Control 190 module, and a Discord Bot module. The Microscope Controller module is an interface with the Python wrapper 191 for MicroManager pymmcore. The Image Analysis module uses deep learning methods to segment yeast cells 192 from bright-field images. It also uses an efficient algorithm for cell tracking. This module is also available as a 193 standalone tool called SegMator. The Model Predictive Control module implements state estimation and model 194 predictive control routines for deterministic and stochastic systems, at either the population or single-cell level. 195 The Discord Bot module uses a web app running on the microscope's computer and connected to the Discord 196 communication system.

197 MicroMator, SegMator, event definitions for representative experiments (Fig 2E and S4), and data analysis code 198 for the experiments (Fig 2E, 3D, and S4), as well as a tutorial example (Supplementary Text 1), can be found 199 online: <u>https://gitlab.inria.fr/InBio/Public/micromator</u>. Raw and processed data for Fig 2C-E, 3C-D, and S4 are 200 freely available on the zenodo repository: <u>https://doi.org/10.5281/zenodo.4616659</u> (45GB).

- 201 Supplementary movies
- Movie S1: real-time_segmentation_and_tracking_with_SegMator.mov. Time-lapse movie showing the real-time segmentation and tracking quality obtained with SegMator. Left: bright-field image. Right: bright-field image overlaid with segmentation mask in cyan.
- **Movie S2**: optogenetic_control_of_gene_expression-Open_loop.mov. Time-lapse movie showing the response of the cells (mScarletl fluorescence) in an open-loop control experiment. Corresponds to Fig. 2C.
- Movie S3: optogenetic_control_of_gene_expression-Population_closed_loop.mov. Time-lapse movie showing the response of the cells (mScarletI fluorescence) in a population closed-loop control experiment. Corresponds to Fig. 2D.
- Movie S4: optogenetic_control_of_gene_expression-Single_cell_closed_loop.mov. Time-lapse movie
 showing the response of the cells (mScarletI fluorescence) in a single-cell closed-loop control experiment.
 Corresponds to Fig. 2E.
- Movie S5: single_cell_recombination-Islets.mov. Time-lapse movie showing the light signal sent to cells in order to create small islets of recombined cells. Corresponds to Fig 3C.

Movie S6: single_cell_recombination-Ring.mov. (Left) Time-lapse movie showing the light signal sent to cells
 in order to recombine all cells that have been at one moment in a ring-like pattern. (Right) Image showing
 the recombined state of the cells at the end of the experiment. Corresponds to Fig 3D.

218 Genetic constructions and yeast strains

219 All plasmids and strains were constructed using the Yeast Tool Kit, a modular cloning framework for yeast 220 synthetic biology²³, the common laboratory strain BY4741 (Euroscarf), and the EL222 optogenetic system²⁴. The 221 light responsive strain (IB44) harbors a constitutively expressed EL222 light-responsive transcription factor (NLS-222 VP16AD-EL222) and an EL222-responsive promoter (5xBS-CYC180pr) driving the expression of the mScarletI 223 protein. The IB44 strain genotype is MATa his3Δ1 leu2Δ0::5xBS-CYC180pr-mScarletI-Leu2 met15Δ0 ura3Δ::NLS-224 VP16AD-EL222-URA3. The recombining strain (IB237) harbors a constitutively expressed EL222 light-responsive 225 transcription factor (NLS-VP16AD-EL222) floxed between two LoxP sites that upon recombination expresses the 226 ATAF1 transcription factor. This factor expresses (pATAF1 4x) in turn the mCerulean fluorescent protein fused 227 to a constitutively active Far1 protein (FAR1M_mCerulean). Lastly, the strain also harbors the Cre recombinase 228 placed under the control of an EL222-responsive promoter (5BS-Gal1pr). The IB237 strain genotype is MATa 229 his3Δ1::pATAF1 4x-FAR1M mCerulean-tDIT1-HIS3 leu2Δ::5BS-Gal1pr-CRE-tENO1-LEU2 met15Δ0 ura3Δ:: 230 pTDH3-LoxP-NLS-VP16AD-EL222-tENO1-LoxP-ATAF1-tTDH1-URA3. Lastly, we also used the IB84 strain as a 231 constitutive 3-color strain to characterize DMD precision. The genotype of this strain is MATa his3Δ1 232 leu2Δ0::pTDH3-mCerulean-tTDH1-pTDH3-NeonGreen-tTDH1-pTDH3-mScarlet-tTDH1-LEU2 met15Δ0 ura3Δ:: 233 NLS-VP16AD-EL222-URA3.

234 Culture preparation

Cells were grown at 30°C in synthetic medium (SD) consisting of 2% glucose, low fluorescence yeast nitrogen
 base (Formedium CYN6510), and complete supplement mixture of amino acids and nucleotides (Formedium
 DCS0019). For each experiment, cells were grown overnight in SC media at 30°C, then diluted 50 times and
 grown for 4 to 5 hours before being loaded in microfluidic plates.

239 Microscopy setup, microfluidics and imaging

240 Images were taken under a Leica DMi8 inverted microscope (Leica Microsystems) with a ×63 oil-immersion 241 objective (HC PL APO), an LTM200 V3 scanning stage, and an sCMOS camera Zyla 4.2 (ANDOR). Bright-field 242 images were acquired using a 12V LED light source from Leica Microsystems. Fluorescence images were acquired 243 using a pE-4000 light source from CoolLED and the following filter cubes: EX:436/20nm DM:455nm 244 EM:480/40nm (CFP), EX:500/20nm DM:515nm EM:535/30nm (YFP), and EX:546/10nm DM:560nm 245 EM:585/40nm (RHOD) from Leica Microsystems. Light stimulation was performed using the pE-4000 light source 246 and the CFP filter. Spatially-resolved illuminations were obtained thanks to a digital mirror device (DMD) 247 reflecting the light of a pE-4000 light source. We used a MOSAIC3 DMD from ANDOR. The device is used both 248 for targeted fluorescence imaging and for optogenetic stimulations. A CellASIC ONIX2 system (Merck) was used 249 together with the Y04C CellASIC microfluidic plates to grow yeast cells in monolayers. Media flow was 250 maintained by a 7.5 kPa pressure gradient. The media was the same as for pre-culture. The temperature was 251 maintained at 30 °C by an opaque environmental box and a temperature controller 2000-2, both from PECON. 252 The microscope was operated using MicroMator.

253 Model predictive control of gene expression

254 To compare single-cell and population control strategies, we developed stochastic and deterministic models of 255 gene expression. Both have been calibrated with respect to the dataset presented in Fig 2B and Fig S5. For 256 population control, we used the deterministic model, assumed Gaussian measurement noise and used a Kalman 257 filter for state estimation. Each model assumes a deterministic delay between the time the light signal is applied 258 and the time protein production is effective. For MPC, fluorescence measurements were taken every 6 minutes 259 and we considered receding time horizons of 24 minutes. The controller explores the set of light stimulation 260 profiles in which a 1000ms light stimulation is either applied or not for each measurement time interval, and 261 selects the profile minimizing mean square deviations. For tracking purposes, brightfield measurements were 262 taken every 3 minutes. For single-cell control, we used the stochastic model and simulated the cell behavior 263 using a finite state projection approximation. For each and every cell, state estimation is performed using a

264 Bayesian approach which conditions the probability distribution for each cell on the most recent measurement,

- and light stimulation profiles are selected using the approach outlined above and the expected absolute deviation as selection criterion. More information is provided in SI Text. Box plots of Figure 2F indicate the lower
- 266 deviation as selection criterion. More information is provided in SI Text. Box plots of Figure 2F indicate the lower 267 guartile, the median, and the upper guartile of the data, with the whiskers corresponding to 1.5 interguartile
- 267 quartile, the median, and the upper quartile of the data, with the whiskers corres 268 ranges.

269 References

- 270 1. Eisenstein, M. Smart solutions for automated imaging. Nat Methods 17, 1075–1079 (2020).
- 271 2. Conrad, C. *et al.* Micropilot: automation of fluorescence microscopy–based imaging for systems biology.
- 272 Nat Methods 8, 246–249 (2011).
- 273 3. Carro, A., Perez-Martinez, M., Soriano, J., Pisano, D. G. & Megias, D. iMSRC: converting a standard
- automated microscope into an intelligent screening platform. *Sci Rep* **5**, 10502 (2015).
- 275 4. Pinkard, H., Stuurman, N., Corbin, K., Vale, R. & Krummel, M. F. Micro-Magellan: open-source, sample-
- adaptive, acquisition software for optical microscopy. *Nat Methods* **13**, 807–809 (2016).
- 277 5. Li, T. *et al*. MAARS: a novel high-content acquisition software for the analysis of mitotic defects in fission
- 278 yeast. *MBoC* 28, 1601–1611 (2017).
- 279 6. Politi, A. Z. et al. Quantitative mapping of fluorescently tagged cellular proteins using FCS-calibrated four-
- 280 dimensional imaging. *Nat Protoc* **13**, 1445–1464 (2018).
- 281 7. Toettcher, J. E., Gong, D., Lim, W. A. & Weiner, O. D. Light-based feedback for controlling intracellular
- signaling dynamics. Nat Methods 8, 837–839 (2011).
- 283 8. Uhlendorf, J. et al. Long-term model predictive control of gene expression at the population and single-cell
- 284 levels. PNAS 109, 14271–14276 (2012).
- 285 9. Lugagne, J.-B. *et al.* Balancing a genetic toggle switch by real-time feedback control and periodic forcing.
- 286 Nat Commun 8, 1671 (2017).
- 287 10. Chait, R., Ruess, J., Bergmiller, T., Tkačik, G. & Guet, C. C. Shaping bacterial population behavior through
- 288 computer-interfaced control of individual cells. *Nat Commun* **8**, 1535 (2017).
- 289 11. Rullan, M., Benzinger, D., Schmidt, G. W., Milias-Argeitis, A. & Khammash, M. An Optogenetic Platform for
- 290 Real-Time, Single-Cell Interrogation of Stochastic Transcriptional Regulation. *Molecular Cell* **70**, 745-756.e6
- 291 (2018).
- 292 12. Perrino, G., Wilson, C., Santorelli, M. & di Bernardo, D. Quantitative Characterization of α-Synuclein
- Aggregation in Living Cells through Automated Microfluidics Feedback Control. *Cell Reports* 27, 916-927.e5
- 294 (2019).

- 295 13. Perkins, M. L., Benzinger, D., Arcak, M. & Khammash, M. Cell-in-the-loop pattern formation with
- 296 optogenetically emulated cell-to-cell signaling. Nat Commun 11, 1355 (2020).
- 297 14.Pinkard, H. et al. Pycro-Manager: open-source software for customized and reproducible microscope
- 298 control. Nat Methods (2021) doi:10.1038/s41592-021-01087-6.
- 299 15. Edelstein, A., Amodaj, N., Hoover, K., Vale, R. & Stuurman, N. Computer Control of Microscopes Using
- 300 μManager. Current Protocols in Molecular Biology 92, 14.20.1-14.20.17 (2010).
- 301 16.Edelstein, A. D. et al. Advanced methods of microscope control using μManager software. J Biol Methods 1,
- 302 10 (2014).
- 303 17. Pedone, E. et al. Cheetah: A Computational Toolkit for Cybergenetic Control. ACS Synth. Biol.
- 304 acssynbio.0c00463 (2021) doi:10.1021/acssynbio.0c00463.
- 305 18.Lugagne, J.-B., Lin, H. & Dunlop, M. J. DeLTA: Automated cell segmentation, tracking, and lineage
- 306 reconstruction using deep learning. *PLoS Comput Biol* **16**, e1007673 (2020).
- 307 19.Allan, D. et al. soft-matter/trackpy: Trackpy v0.4.2. (Zenodo, 2019). doi:10.5281/ZENODO.3492186.
- 308 20. Falk, T. *et al.* U-Net: deep learning for cell counting, detection, and morphometry. *Nat Methods* **16**, 67–70
- 309 (2019).
- 310 21. Polstein, L. R., Juhas, M., Hanna, G., Bursac, N. & Gersbach, C. A. An Engineered Optogenetic Switch for
- 311 Spatiotemporal Control of Gene Expression, Cell Differentiation, and Tissue Morphogenesis. ACS Synth. Biol.
- **6**, 2003–2013 (2017).
- 313 22.Aditya, C., Bertaux, F., Batt, G. & Ruess, J. A light tunable differentiation system for the creation and control
- 314 of consortia in yeast. *bioRxiv* 2021.06.09.447744 (2021).
- 315 23.Lee, M. E., DeLoache, W. C., Cervantes, B. & Dueber, J. E. A Highly Characterized Yeast Toolkit for Modular,
- 316 Multipart Assembly. ACS Synth Biol 12 (2015).
- 317 24.Benzinger, D. & Khammash, M. Pulsatile inputs achieve tunable attenuation of gene expression variability
- 318 and graded multi-gene regulation. *Nat Commun* 9, 3521 (2018).

319 Acknowledgments

320 This work was supported by ANR grants CyberCircuits (ANR-18-CE91-0002), MEMIP (ANR-16-CE33-0018), and

- 321 Cogex (ANR-16-CE12-0025), by the H2020 Fet-Open COSY-BIO grant (grant agreement no. 766840) and by the
- 322 Inria IPL grant COSY. We acknowledge the support of the U.S. Department of Energy through the LANL/LDRD
- 323 Program and the Center for Non Linear Studies for this work.

324 Author contributions

- 325 S.F. developed the MicroMator software and performed experiments. Z.F. wrote the image analysis and real-
- 326 time control modules, and analyzed the data. A.F. performed experiments and helped develop the MicroMator
- 327 software and analyze the data. C.A. developed strains and helped perform experiments. S.S.-C. developed
- 328 strains. S.G. helped with software development. F.B. helped with software development and platform
- 329 integration. J.R. helped with controller development. Z.F., F.B., J.R., and G.B. supervised the project. Z.F. and
- 330 G.B. wrote the manuscript with input from all authors.

331 Declaration of interests

332 The authors declare no competing financial interests.