



HAL
open science

Biomechanical effects of using a passive back support exoskeleton during prone-positioning maneuver: A pilot study

Waldez Gomes, Pauline Maurice, N Settembre, J Theurel, L Wioland, J J Atain-Kouadio, L Claudon, B Chenuel, H Hani, A Kimmoun, et al.

► To cite this version:

Waldez Gomes, Pauline Maurice, N Settembre, J Theurel, L Wioland, et al.. Biomechanical effects of using a passive back support exoskeleton during prone-positioning maneuver: A pilot study. SB 2021 - 46th Congress of the Society of Biomechanics, Oct 2021, Saint-Etienne, France. hal-03285178

HAL Id: hal-03285178

<https://hal.science/hal-03285178>

Submitted on 13 Jul 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Biomechanical effects of using a passive back support exoskeleton during prone-positioning maneuver: A pilot study

W. Gomes^a, P. Maurice^{a*}, N. Settembre^b, J. Theurel^c, L. Wioland^c, J.J. Atain-Kouadio^c, L. Claudon^c, B. Chenuel^d, H. Hani^e, A. Kimmoun^f, B. Levy^f, J.M. Sellal^g, J. Paysant^h, S. Malikov^b and S. Ivaldi^a

^a University of Lorraine, CNRS, Inria, Loria, F-54000 Nancy, France; ^bCHRU-Nancy, Inserm U1116, University of Lorraine, F-54000 Nancy, France; ^cWorking Life Dept., INRS, F-54500 Vandoeuvre-les-Nancy, France; ^dCHRU-Nancy, University Centre of Sports Medicine and Adapted Physical Activity, Pulmonary Function and Exercise Testing Dept., EA DevAH, Dept. of Medical Physiology, University of Lorraine, F-54000 Nancy, France; ^eVirtual Hospital of Lorraine, CUESim, University of Lorraine, F-54000 Nancy, France; ^fMedical Intensive Care Unit Brabois, CHRU-Nancy, INSERM U1116, University of Lorraine, F-54000 Nancy, France; ^gCHRU-Nancy, Dept. of Cardiology, F-54000 Nancy, France; ^hCHRU-Nancy, Dept. of Rehabilitation Medicine, EA DevAH, University of Lorraine, F-54000 Nancy, France.

Keywords: Exoskeleton; Ergonomics; Workload Assessment; Digital Human Simulation ; Healthcare Sector.

1. Introduction

Musculoskeletal disorders associated with patients care and manipulation are frequent among healthcare workers, due to force exertion and sustained awkward postures (EU Commission 2011, Davis 2015). During the COVID-19 pandemic, intensive care units (ICUs) staffs had to perform prone-positioning (PP) of a large number of intubated patients suffering from Acute Respiratory Distress Syndrome (ARDS) many times a day. PP maneuver consists in repositioning a patient from a supine position (on the back) to a prone position (on the front side), which improves oxygenation and ventilatory mechanisms in ARDS patients (Guérin *et al.* 2013). PP is a delicate procedure (risk of extubation) which requires the medical staff to remain bent forward for several minutes (Fig. 1). This posture result in back loading and in the long-term potential injuries in the low-back. Similar postures are common in industrial tasks, and recently occupational exoskeletons have started to be commercialized to help reduce physical load on workers (DeLooze *et al.* 2016). Exoskeletons therefore appear as a promising solution to support healthcare workers, though few studies focused on this sector (e.g. Miura *et al.* 2020; Iishi *et al.* 2015). But while studies overall agree that back-support exoskeletons, and specifically passive devices, have the potential to reduce lumbar muscle activity in tasks involving trunk bending, the exact efficacy and usability of such devices appear largely device- and task-dependent (Theurel *et al.* 2019). A preliminary study of our team (Settembre *et al.* 2020) identified the Laevo passive exoskeleton for lumbar support (Laevo, the Netherlands) as the exoskeleton that best matched the PP maneuver and ICU constraints, among 4 exoskeletons that were available at the time of the test during the COVID-19 first wave. The present study presents a pilot investigation of the biomechanical effects of using the Laevo during PP maneuver.

2. Methods

2.1 Experimental protocol



Figure 1: Medical staff equipped with the Laevo exoskeleton performing a PP maneuver on a manikin.

Two healthy volunteers with PP experience performed PP maneuvers both with and without wearing the Laevo, on a patient simulator (manikin) at the Hospital Simulation Center of the University of Lorraine (Fig. 1). Each maneuver lasted about 4 min. In a first experiment, the participants were equipped with the Xsens MVN Link inertial motion capture system (capture rate 240 Hz) to record whole-body kinematics. In a subsequent experiment, the participants were equipped with a Delsys Trigno EKG sensor to measure heart rate, and 12 Delsys Trigno sEMG sensors (sampling rate 4370 sa/sec) placed according to the Seniam recommendations bilaterally on the erector spinae longissimus (ESL), erector spinae iliocostalis (ESI), trapezius ascendens (TA), and biceps femoris long head (BF), and laterally on the right rectus abdominis at T10 level (RA), rectus femoris (RF), gluteus maximus (GM) and tibialis anterior (TAL).

2.2 Data analysis

In order to analyze the kinematics and dynamics of the maneuver, we replayed the recorded motions in a physics-engine based simulation using a 43 degrees of freedom digital human model (DHM). The recorded motions were retargeted on the DHM directly in the Cartesian space by using a quadratic programming controller (QP) typically used for humanoid robots (Maurice *et al.* 2019). The QP control directly estimates the joint torques associated with the motion. This method has the advantage to blend the traditional inverse kinematics and inverse dynamics steps into one faster step, since the computation runs in

close to real-time. Since we did not have a direct measurement of the support torque provided by the Laevo, we used the empirical calibration curve published by (Koopman *et al.* 2019) to estimate the lumbar support torque depending on the trunk flexion angle. We used the resulting L5/S1 flexion/extension estimated joint torque as a complement to the physiological measurements obtained with the sEMG.

3. Results and discussion

Kinematic results: Using the Laevo did not significantly modify the kinematics of the motion during PP. Specifically, the profile of the lumbar flexion angle remained overall similar (Fig. 2). This result was confirmed by the subjective feedback of the participants who reported that they did not feel the exoskeleton was affecting the way they performed the task.

Dynamic results: The median value of the estimated L5/S1 flexion/extension torque decreased by about 12% when using the Laevo (Fig. 2). This result agrees with the reduced perceived lumbar effort reported by participants, and is aligned with the 15% reduction observed by (Koopman *et al.* 2019) in a similar static forward bending task. Yet those results should be considered cautiously, since only the assistive torque of the exoskeleton was modeled, and not the full effect of the load transfer.

Physiological results: We did not observe any major change in heart rate when using the Laevo. This might be due to the short duration of the experimental task (4 min). The activity of the back extensor muscles (except ESI) and hip extensor muscles overall decreased when using the Laevo (Table 1). The other recorded muscles remained mostly unaffected by the exoskeleton. Those results are consistent with the perceived reduced effort reported by the participants.

4. Conclusions

Our pilot study, though limited to a small number of participants, suggests that using the Laevo could help reduce the musculoskeletal load on the low back during PP maneuver, without causing any significant negative side-effect nor modifying the practice. Following this study, 2 Laevos have been successfully deployed in the ICU of CHRU Nancy and have been used there for the past 6 months. Recording quantitative measurements during this field-testing is not possible. However subjective feedback from users is collected through questionnaires and interviews, and will be analyzed in a future study.

References

EU Commission. 2011. Occupational Health and Safety Risks in the Healthcare Sector. Publications Office of the European Union.

*Corresponding author. Email: pauline.maurice@loria.fr




Figure 2: Time-series (top) and distribution across time (bottom) of the L5/S1 flexion angle (left) and torque (right) during PP maneuver with and without using the Laevo.

	TA	ESL	ESI	GM	RF	BF	RA	TAL
P1 H	-35	-12	-1	-30	0	-16	-1	-2
P2 H	+2	+5	0	+5	NA	-20	0	+6
P1 S	-15	-9	-1	-12	-6	-53	0	-10
P2 S	-5	-11	-2	+8	NA	-34	0	-20

Table 1: Relative difference (%) in RMS value of the EMG signal with the Laevo vs. without it, for both participants and both PP positions (patient's head (H) and side (S)).

- Davis KG, Kotowski SE. 2015. Prevalence of musculoskeletal disorders for nurses in hospitals, long-term care facilities, and home health care: A comprehensive review. *Hum. Factors* 57:754–792.
- Guérin C, et al. 2013. Prone positioning in severe acute respiratory distress syndrome. *N Engl J Med.* 368:2159-2168.
- De Looze M, et al. 2016. Exoskeletons for industrial application and their potential effects on physical work load. *Ergonomics*, 59(5):671–681.
- Koopman AS, et al. 2019. Effects of a passive exoskeleton on the mechanical loading of the low back in static holding tasks. *J Biomech* 83:97-103.
- Miura K, et al. 2020. Successful Use of the hybrid assistive limb for care support to reduce lumbar load in a simulated patient transfer. *Asian Spine J.*
- Iishi C, et al. 2015. Development of a new type of lightweight power assist suit for transfer work. *Proc Asia-Pacific Conf Comp Aided Syst Eng* 208-213.
- Settembre N, et al. 2020. The use of exoskeletons to help with prone positioning in the intensive care unit during COVID-19. *Ann Phys Rehab Medicine.*
- Maurice P, et al. 2019. Assessing and improving human movements using sensitivity analysis and digital human simulation. *Int J Comp Integ Manuf* 32(6):546–558.
- Theurel J, et al. 2019. Occupational exoskeletons: Overview of their benefits and limitations in preventing work-related musculoskeletal disorders. *Trans Occup Ergo Hum Factors*, 7:3-4, 264-280.