

#### Probing the interactions between bubbles (gas / liquid) and (bio)interfaces at the molecular scale using FluidFM technology

Irem Demir, Ines Lüchtefeld, Claude Lemen, Etienne Dague, Pascal Guiraud, Tomaso Zambelli, Cécile Formosa-Dague

#### ▶ To cite this version:

Irem Demir, Ines Lüchtefeld, Claude Lemen, Etienne Dague, Pascal Guiraud, et al.. Probing the interactions between bubbles (gas / liquid) and (bio)interfaces at the molecular scale using FluidFM technology. Journées plénières du GDR Bioingénierie des Interfaces (B2I), Sep 2021, Toulouse, France. hal-03429997

#### HAL Id: hal-03429997 https://hal.science/hal-03429997v1

Submitted on 16 Nov 2021

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Bio & Chemical Engineering

4<sup>ÈMES</sup> JOURNÉES **PLÉNIÈRES** GDR B21

# Probing the interactions between bubbles (gas / liquid) and (bio) interfaces at the molecular scale using FluidFM technology Irem DEMIR-YILMAZ<sup>1,2,\*</sup>, Ines Lüchtefeld<sup>3</sup>, Claude Lemen<sup>1</sup>, Etienne Dague<sup>2,4</sup>, Pascal

## Guiraud<sup>1,4</sup>, Tomaso Zambelli<sup>3</sup>, and Cécile Formosa-Dague<sup>1,4\*</sup>

<sup>1</sup> TBI, Université de Toulouse, INSA, INRAE, CNRS, Toulouse, France

<sup>2</sup> LAAS, Université de Toulouse, CNRS, Toulouse, France

<sup>3</sup> Laboratory of Biosensors and Bioelectronics, Institute for Biomedical Engineering, ETH Zurich, Zurich, Switzerland

<sup>4</sup> Fédération de Recherche Fermat, CNRS, Toulouse, France

demir@insa-toulouse.fr

### ABSTRACT

Understanding the molecular mechanisms underlying bubble-(bio)surfaces interactions is currently a challenge that if overcame, would allow to understand and control the various processes in which they are involved. Atomic force microscopy is a valuable tool to measure such interactions, but it is limited by the large size and instability of bubbles. To overcome these challenges, we here develop a new method to probe more accurately the interactions between bubbles and (bio)-interfaces by using the fluidic force microscopy technology (FluidFM) that combines AFM with microfluidics. In this study, we use FluidFM in an original manner, to produce microsized bubbles. We characterized the bubbles produced using this method, and their interactions with hydrophobic surfaces were probed, showing that bubbles behave like hydrophobic surfaces. Thus they can be used to measure the hydrophobic properties of microorganisms' surfaces.

Finally we developed a strategy to functionalize their surface, thereby modulating their interactions with microorganisms' surfaces. This new method provides a valuable tool to understand bubble-(bio)surfaces interactions but also to engineer them.

## INTRODUCTION

Our aim is to understand the biological and physicochemical nature of thebubbles and (bio)-surfaces




New tools to measure the interactions with bubbles

**Develop strategies to** functionalize bubbles


We used FluidFM in an original manner to produce bubble

#### FluidFM Technology to produce bubbles and measure interactions




#### **Technological challenges:**

- Produce stable microbubbles
- Functionalize bubbles with surfactants
- Measure interactions with cells at the molecular level




### **METHODS**

#### **Production of stable microbubbles**




### **RESULTS**<sup>[3]</sup> **Probing the hydrophobicity of surfaces**



P. aeruginosa


(bacteria)




C. albicans

## **Probing the interaction with living cells**

100 80



#### Adhesion (nN)





## **Bubble behave like hydrophobic surface**

Bubble probe method can be used with reliability to measure the hydrophobicity of complex surfaces

> **Different type of** microorganisms show different hydrophobicity.

> > Interactions of cells with

**bubbles are controlled by** 


**bubble functionalization** 

REFERENCES

[1] Poulain et al., Physical Review Letter, 2018

[3] Demir et al., Journal of Colloid and Interface Science, 2021

[2] Meister, et al. Nano Letters, 2009



C. vulgar<mark>is</mark>

roalaae)

### **CONCLUSIONS**

Here we provide the possibility to (i) modify easily the surface of the bubbles produced using FluidFM, and (ii) to show in what way the modification of the bubble surface influences the nature and strength of the interaction with cells. In future projects, this strategy could also be used to specifically separate cell populations from each other; for example to separate bacterial cells from human blood cells in the case of sepsis, but many other applications can now be envisioned.

• **TBI** - Toulouse Biotechnology Institute, *Bio & Chemical Engineering* www.toulouse-biotechnology-institute.fr



colistin