

Towards absorption spectroscopy by means of mid-infrared supercontinuum generation in a step index tellurite fiber

Arnaud Lemière, Anthony Maldonado, Frédéric Désévédavy, Bertrand Kibler, Pierre Mathey, Grégory Gadret, Jean-Charles Jules, Nguyen Phuoc Trung Hoa, Takenobu Suzuki, Yasutake Ohishi, et al.

▶ To cite this version:

Arnaud Lemière, Anthony Maldonado, Frédéric Désévédavy, Bertrand Kibler, Pierre Mathey, et al.. Towards absorption spectroscopy by means of mid-infrared supercontinuum generation in a step index tellurite fiber. Laser Physics, 2021, 31 (2), pp.025702. 10.1088/1555-6611/abd9af. hal-03436982

HAL Id: hal-03436982

https://hal.science/hal-03436982

Submitted on 25 Nov 2021

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Towards absorption spectroscopy by means of mid-infrared supercontinuum generation in a step index tellurite fiber

Arnaud Lemière¹, Anthony Maldonado¹, Frédéric Désévédavy¹, Bertrand Kibler¹, Pierre Mathey¹, Grégory Gadret¹, Jean-Charles Jules¹, Nguyen Phuoc Trung Hoa², Takenobu Suzuki², Yasutake Ohishi² and Frédéric Smektala¹

surson e to your trace.jp (1x.1x.), takenoon e to your trace.jp (1.5.), omsin e to your trace.

E-mail: frederic.smektala@u-bourgogne.fr

Received xxxxxx Accepted for publication xxxxxx Published xxxxxx

Abstract

We report on mid-infrared gas spectroscopy measurements by means of the supercontinuum absorption spectroscopy method. We developed a 80-MHz-repetition-rate mid-infrared supercontinuum light source with an average power of 40 mW generated through nonlinear propagation of 200-fs pulses in a 10-cm-long segment of step-index tellurite fiber. A compact multipass cell is used for analyzing absorption of CO_2 and CH_4 gases and their mixture in the 2-3.5 μ m spectral range. Experimental results are compared to the HITRAN database.

Keywords: supercontinuum absorption spectroscopy; mid-infrared light sources; tellurite glasses; optical fibers

1. Introduction

Mid- and Long-Wave Infrared (MIR and LWIR) are two spectral regions (2-8 μ m for MIR and 8-17 μ m for LWIR) of great interest for sensing applications because of many chemical species spectral signature located between 1 and 18 μ m [1–3]. Especially, carbon dioxyde (CO₂) and methane (CH₄) are two greenhouse pollutants absorbing in both MIR and LWIR regions. CO₂ absorbs light at 2, 2.7, 4.26, 4.84 μ m in the MIR region. In the LWIR region, one large absorption band is located between 13.2 and 17 μ m with a maximum peak at 15 μ m. CH₄ presents narrow absorption bands centered at 1.3, 1.65, 2.36, 2.58 and 6.5 μ m as well as two large absorption bands from 3.11 to 4.12 μ m (maximum peak at 3.27 μ m) and from 7.15 to 8.5 μ m (maximum peak at 7.66 μ m) [4]. Table 1 summarizes the different absorption bands of these two chemical species and gives the corresponding line intensities expressed in cm⁻¹/(molec.cm⁻²). This unit represents the integral of the absorption band relative to the volume density of molecules at a given pressure.

¹ Laboratoire Interdisciplinaire Carnot de Bourgogne, UMR 6303 CNRS-UBFC, 9 av. Alain Savary, BP 47870, 21078 Dijon, France;

² Research Center for Advanced Photon Technology, Toyota Technological Institute 2-12-1, Hisataka, Tempaku, Nagoya 468-8511, Japan; sd15501@toyota-ti.ac.jp (K.N.); takenobu@toyota-ti.ac.jp (T.S.); ohishi@toyota-ti.ac.jp (Y.O.)

Table 1. Absorption bands of CO_2 and CH_4 between 1 and 17 μm with the corresponding line intensities.

Wavelength (µm)	Line intensities (cm ⁻¹ /(molec.cm ⁻²)) x10 ⁻²²				
CO ₂					
2.0	13.3				
2.68	580				
2,76	390				
4.26	34951.0				
4.84	2.06				
13.2-17	2967.2				
CH ₄					
1.3	1.6				
1.65	15.6				
2.36	52.6				
2.58	5.0				
3.11-4.12	2083.05 (@ 3.27 μm)				
6.5	8.4				
7.15-8.5	976.0 (@ 7.5 μm)				

From Table 1, we notice that line intensities are generally increasing as long as we increase the wavelength from near-infrared to MIR. There is a factor 2600 by comparing the line intensities centered at 2.0 μ m and at 4.26 μ m for CO₂. For CH₄, there is a factor 1300 between the line intensities located at 1.3 μ m and the ones around 3.11-4.12 μ m. These values confirms the interest that lies gas absorption spectroscopy in the 3-5 μ m atmospheric window (see figure 1).




Figure 1. Atmospheric spectrum (black) [5] compared to line intensities of methane and carbon dioxide [4]

Supercontinuum absorption spectroscopy (SAS) is a powerful technique for multi-species analysis of a gas mixture or liquids. This method has already been extensively studied in various configurations (open paths or local sensing) [6, 7]. Nowadays, it mainly consists in the developpement of suitable fiber-based supercontinuum (SC) light sources according to the detection spectral range, to benefit from the higher power of SC pulsed sources than black body radiators. Note that the addition of a strongly-dispersive propagation stage of SC pulses offers the possibility of time-domain based broadband spectroscopy [8, 9]. The recent developpment of fiber-based MIR light sources has then largely extended the spectral range of applications of SAS where many gases exhibit stronger absorption lines [10]. It is also worth mentionning that one can combine photoacoustic detection with SAS, thus providing higher sensitivity than conventional Fourier transform infrared (FTIR) spectrometers [11, 12].

We previously reported on SAS measurements of CH₄ performed between 1 and 2.5 μ m with a SC generated in a 75% TeO₂ - 15% ZnO - 5% Na₂O - 5% ZnF₂ (% molar) microstructured tellurite fiber [13]. Because of the better understanding of aging issues of tellurite glasses [14] in particular in the 2.5-3.5 μ m range (OH absorption), the development of all-solid tellurite fibers appears now essential. Besides, it has been demonstrated that the simple step-index design allows to overcome aging issue, and it offers large capabilities in controlling the fiber dispersion properties and subsequent SC generation [15, 16, 17], just by slightly varying the core size and refractive index difference between core and cladding.

In this work we demonstrate the application of such tellurite fiber-based supercontinuum sources for multi-species analysis of a gas mixture by means of the SAS method in the $2-3.5~\mu m$ spectral range. To this regard, we make use of a compact multipass cell (70cm-long optical path length, 8 reflexions on gold mirror) and a convential FTIR spectrometer. First, we describe the tellurite glass preparation, fiber fabrication, and the resulting fiber properties as well. Next, we investigate supercontinuum absorption spectroscopy for detection of CO_2 and CH_4 gases. We analyze the response to nominal gas concentrations and then apply our SAS setup for sensing gas mixtures.

2. Glass preparation and fiber properties

The step-index tellurite fiber is developed by means of conventionnal built-in-casting and rod-in-tube methods as described in Ref. [18]. Two compatible glasses are selected to develop a small-core step-index tellurite fiber (with core diameter below 5 μ m for flexible dispersion engineering): TZNF and TZNG for core and clad glasses, respectively. The composition of TZNF core glass is 80TeO_2 -5ZnO- $10\text{Na}_2\text{O}$ -5ZnF₂ (molar fraction) and the composition of TZNG clad glass is 60TeO_2 -5ZnO- $20\text{Na}_2\text{O}$ -15GeO₂ (molar fraction). These two glasses have a good thermal and optical compatibility (see Table 2). Their optical index difference is about 0.13 at 633 nm (0.12 at 1550 nm) and the difference between both glass transition temperatures is only 8 °C. Furthermore, the thermal expansion coefficients are very similar with a difference of 3.7×10^{-6} °C⁻¹, thus allowing an excellent compatibility for co-drawing.


Table 2. Principal properties of the tellurite glasses used for fabrication of our step-index fiber.

	n (633 nm)	n (1064 nm)	n (1550 nm)	Tg (°C)	α (x10 ⁻⁶ °C ⁻¹)
TZNF	2.037	/	1.989	280	16.7
TZNG	1.905	1.872	1.867	272	13.0

The synthesis of the glass preform is realized under pure O₂ atmosphere by means of a furnace connected to a dry air glove box. This avoids water contaminations and reduces the amount of OH residual groups in the glass matrix. To improve the purification of our core glass, we substituted 5% of zinc oxide by zinc fluoride [18–21]. OH groups will react with ZnF₂ to produce HF (and ZnO), evacuated by the constant furnace O₂ gas flow (2 L/min). This gas flow is used to guarantee the oxidation conditions during synthesis. The two glasses are synthesized from high purity precursors provided by Fox chemicals, 5N for TeO₂, GeO₂, ZnO and 6N for Na₂CO₃. Alfa Aesar 4.5N ZnF₂ is also used. Both melted glasses are quenched into a brass preheated mold at 240 °C after a synthesis at 900 °C during around 1h30. The resulting preform is then annealed for 8 hours at 270 °C. We obtain a 45mm-long and 16mm-large cylindrical preform [18].

The large-core step-index fiber preform (core-clad ratio $\approx 50\%$) from the build-in-casting stage is used to manufacture capillaries and measure fiber losses (see figure 2a). The fiber under test has an outer diameter of 150 µm and a core diameter of 70 µm. The infrared attenuation spectrum of the large-core step-index fiber was obtained by using the cutback method on several meters-long fiber samples with a FTIR spectrometer (NICOLET 6700) working in the 1–5 µm range. For technical purposes, an external halogen lamp emitting from 0.1 to 5 µm was used as the light source. The attenuation of the fiber exhibits low losses (below 1 dB/m between 1.5 and 2.8 µm). Its attenuation increases until 8 dB/m at 4 µm. We observe that the OH residual groups between 3 and 4 µm are very low. Considering the absorption coefficient of the OH-group in silica ($\epsilon_{OH} = 49.1 \times 10^3 \text{ cm}^3 \text{mol}^{-1} \text{cm}^{-1}$) [20,22] glass we estimate the amount of OH residual groups at a level below 0.4 ppm. Another clad-glass preform is synthesized and then drilled at the convenient diameter to insert inside a capillary from the previous large-core preform. An additional drawing process is then required to fabricate a small-core step-index fiber. The resulting fiber is chosen with an external diameter of 150 µm, which corresponds to a core diameter of 3.5 µm. This particular value of core diameter gives rise to low and flat anomalous dispersion in the 2 -2.5 µm spectral range (as shown in figure 2b) as well as strong optical confinement [15], which is fully relevant for SC generation with sub-nanojoule femtosecond pump pulses located beyond 2 µm

[16]. It was shown that coherent MIR supercontinuum generation can be obtained for various dispersion regimes and nonlinear propagation over 10 cm, provided that a moderate injected input power is used. More particularly, the optimized dispersion profile (figure 2b) for SC generation is the one resulting from the 3.5 µm core with two close zero-dispersion wavelengths (and pumping into the anomalous dispersion), while SC coherence is preserved for input pulse energy about a few hundreds of pJ (i.e., in the present case below 3-kW peak power for 200-fs pulses). Such a configuration is then chosen in the following SAS experiments for enhanced pulse-to-pulse stability over the whole SC spectrum.






Figure 2. (a) Attenuation of the large core step-index fiber. (b) Wavelength-dependent curves of dispersion parameter D of our step-index fiber design for the following core diameters \emptyset =3.0, 3.25, 3.5, 3.75, 4.0 μ m (Inset: corresponding image of the small-core fiber cross section).

3. Experimental setup of supercontinuum absorption spectroscopy

The experimental set-up of SAS is depicted in figure 3. A Ti:Sapphire laser pumps an optical parametric oscillator (OPO) that delivers 200-fs idler pulses at a repetition rate of 80 MHz and with an average power about 200 mW in the 2.2-2.4 μ m spectral window. A half-wave plate is inserted between the laser and the tellurite fiber to excite the fiber under test along a defined polarization axis. A silica microscope objective allows an efficient beam coupling (2.5-kW injected peak power) into the fundamental mode of our step-index tellurite fiber. A 10-cm-long segment of tellurite fiber is then carefully cleaved with a scalpel blade and the output SC signal is collimated with a chalcogenide lens. The light beam is re-focalized in the multipass cell with a 10-cm focal length CaF2 lens. The multipass cell provides an optical path length of 70 cm for 8 reflexions on gold mirrors. The output signal of the cell is collimated with a silver off-axis parabolic mirror and then focalized in an InF3 multimode fiber (with high transmission between 0.35-5.5 μ m) connected to a compact FTIR spectrometer (Arcoptix FT-MIR Rocket 2-6 μ m). The resolution of the spectrometer is here 10 cm⁻¹. The multipass cell is connected to a vacuum pump and a pressure gauge. It is also connected to particles filter to protect the inner part of the cell from dust. Then two gas bottles are connected, which allow to fill the cell with a neutral gas (N₂) and the analyte gas (CO₂ or CH₄). The neutral gas is used to purge the cell. We can also adjust some binary mixtures composed of CO₂ and CH₄ by connecting the two corresponding bottles.




Figure 3. Experimental setup for supercontinuum generation and SAS experiment.

The SC measured at the output of the step-index tellurite fiber is shown in figure 4. A SC spectrum covering at least the 2-3.5 μ m window is obtained. At wavelengths beyond 3.5 μ m, the FTIR sensitivity did not allow us to have a suitable signal-to-noise ratio for next SAS measurements. The input pump spectrum is also indicated in figure 4. We superposed the CH₄ and CO₂ line intensities to clearly locate their spectral positions in comparison with the measured SC. One can easily note that the CO₂ absorption band located at 2.7 μ m and three CH₄ absorption bands located at 2.36, 2.58 and 3.27 μ m (see Table 1) can be analysed in SAS experiments. To this regard, we carried out the following procedure of measurements. First, a reference SC spectrum is recorded when the cell is filled with neutral gas (N₂ at 1 bar). Next, the cell is purged of N₂ with a flow of analyte gas (CH₄ or CO₂) for a few seconds and this gas flow is stopped by closing the last valve. Then, another SC spectrum is measured for several pressure values (e.g., 990, 600, 200, 100 and 50 mbar) by gradually evacuating the cell by means of the vacuum pump.




Figure 4. Recorded SC spectrum at the output of the 10-cm-long tellurite fiber (blue curve). Input pump spectrum (red curve). Gray and black lines indicate the CH_4 and CO_2 absorption bands, respectively.

4. CO₂ detection

Multiple SC spectra were recorded at the output of the gas cell for various CO_2 pressures. Figure 5a presents a typical SC measurement for 600 mbar of CO_2 (red curve) compared to the SC recorded when the cell is filled with N_2 (blue curve). We clearly observe the CO_2 absorptions located in the 2.68-2.84 μm range on the SC spectrum after the gas cell, in fairly good

agreement with the HITRAN database. The measured absorbance *A* through the cell is then plotted for each gas pressure in figure 5b according to the following equation:

$$A = log\left(\frac{I_0}{I}\right) \tag{1}$$

where I_0 is the reference SC spectrum (cell filled with N_2) and I is the intensity of the measured spectrum through the cell filled with CO_2 . We observe the increase of CO_2 absorbance up to 600 mbar in the cell. Between 50 and 600 mbar, we measure the absorbance maximum increasing from 0.38 to 1.45. For higher pressure, a saturation effect is visible over the full band of interest due to the increasing of the absorption cross section directly related to the gas pressure. There is no saturation for the band located at 2 μ m because the line intensities at such wavelength are significantly weaker. Then, the evolution of the absorbance is only noticed beyond 600 mbar (i.e., thus requiring a ten times increase of the pressure when comparing to the 2.7- μ m absorption band). As reported in Table 1, absorption line intensities of CO_2 show three main absorbance bands between 2 and 3 μ m. The first one (and the weakest one) is at 2 μ m with a value of 1.3×10^{-21} cm⁻¹/(molec.cm⁻²). The second and the third ones are located at 2.68 μ m and 2.76 μ m and have respectively a value of 5.8×10^{-20} and 3.9×10^{-20} cm⁻¹/(molec.cm⁻²). This intensity difference induces a pressure range for which an absorption band is useful. For example, the two absorption bands located at 2.68 and 2.76 μ m are useful for low pressures as it can be seen between 50 and 600 mbar (see figure 5b). In the case of CO_2 , it would be useful to get access to the absorption band located at 4.26 μ m because its high value of absorption intensity (3.5×10⁻¹⁸ cm⁻¹/(molec.cm⁻²)) would allow us to detect a lower amount of gas and thus increase sensitivity.




Figure 5. (a) SAS spectra measured at the output of the cell with (red) and without (blue) CO_2 gas. (b) Evolution of the measured absorbance of CO_2 as a function of pressure.

The line intensities given by HITRAN database for the absorbance bands at 2 μ m and 2.68-2.84 μ m are shown in figure 6a (grey line), as well as the comparison of the measured absorbance (red) with the predicted absorbance (blue) for the pressure of 600 mbar. In the case of the absorptions located between 2.6 and 2.9 μ m, we observe a good correlation. By measuring the area of this absorption band, we can plot a calibration curve (see figure 6b) to determine the response of our SAS system and retrieve the gas content in next detection measurements. This curve shows the increasing measured area of the absorption band with pressure until reaching a saturation above 300 mbar, with a rather good agreement between theoretical and experimental absorption for all the different pressures. Note that our measurements are not located in the linear regime of our SAS system (here well below 50 mbar). This would require another gas cell configuration for higher sensitivity. For instance, air contains 400 ppm of carbon dioxide which would correspond to 0.4 mbar of CO_2 in our cell at atmospheric pressure. As a result, the present sensitivity of our simple SAS system is not high enough for such a low content of CO_2 due to low signal-to-noise ratio in our spectral measurements when gas pressure is below 10 mbar. A longer optical path in the gas cell would be required for detecting CO_2 in air at these wavelengths.




Figure 6. (a) Measured absorbance of CO_2 (red) compared to theoretical absorbance calculated from HITRAN database (blue). The grey curve corresponds to the line intensities of CO_2 given by HITRAN database. (b) Experimental calibration (red points) of CO_2 detection for the 2.6-2.9 μ m waveband by our SAS system and comparison with HITRAN predictions (blue dashed line). Magenta points correspond to sensing of gas mixtures.

5. CH₄ detection

The same procedure has been applied for CH_4 detection. Figure 7a shows a typical example of SAS measurement, namely two SC spectra recorded when the multipass cell is filled with N_2 or with 300 mbar of CH_4 . Here we observe three absorption bands centered at 2.3 μ m, 2.58 μ m and 3.27 μ m, respectively (see also Table 1). Corresponding measured absorbance as a function of gas pressure is depicted in figure 7b. We observe again a continuous increase of absorbance with pressure until reaching a saturation beyond 600 mbar, in particular for the absorption bands centered on 2.3 and 3.27 μ m. Those differences on the useful range of pressures is related with a strong difference between CH_4 line intensities in accordance with HITRAN database. Indeed, the CH_4 absorption line intensities at 2.3 μ m are about 6.3×10^{-21} cm⁻¹/(molec.cm⁻²) instead of 1.3×10^{-21} cm⁻¹/(molec.cm⁻²) at 2.6 μ m. Beyond 3 μ m, the absorption intensity has a value around 2.1×10^{-19} cm⁻¹/(molec.cm⁻²). Thus, CH_4 SAS measurement beyond 3 μ m leads to higher sensitivity and allow detection of lower gas content.




Figure 7. (a) SAS spectra measured at the output of the cell with (red) and without (blue) CH_4 gas. (b) Evolution of the measured absorbance of CH_4 .

The line intensities given by HITRAN database for the absorbance bands are shown in figure 8a (grey line), as well as the comparison of the measured absorbance (red) with the predicted absorbance (blue) for the pressure of 200 mbar. A calibration curve is plotted by measuring the area of the absorption band beyond 3 µm as a function of gas pressure in order to qualify the response of our SAS system for CH₄ (see figure 8b). This waveband is chosen since it significantly differs from the CO₂ absorption band studied above, thus facilitating the investigation of gas mixtures in the following. The analysis of area of the long-wavelength absorption band indicates a saturation effect that appears around 100 mbar (see figure 8b). The prediction from HITRAN data shows that this saturation effect occurs even at lower pressures (before 50 mbar of CH₄). As in the case of CO₂ measurements, our cell configuration doesn't correspond to a linear regime.




Figure 8. (a) Measured absorbance of CH₄ (red) compared to theoretical absorbance given by HITRAN database (blue). The grey curve corresponds to the line intensities of CH4 given by HITRAN database. (b) Calibration curves of CH₄ detection by our SAS system for the 3.0-3.34 μ m waveband by our SAS system and compared with HITRAN predictions (blue dashed line). Magenta points correspond to sensing of gas mixtures.

6. Detection of gas mixtures

As a proof of principle, we demonstrate the capability of our simple SAS system to detect and differentiate gases in a binary mixture. Two gas mixtures composed of CH₄ and CO₂ are analysed at a pressure of 1 bar in our multipass cell. Figure 9 shows the absorbance measurements for 66% of CO₂ with 33% of CH₄ (red curve) and 33% of CO₂ with 66% of CH₄ (blue curve). All the absorption signatures detected earlier are observed and their discrimination is possible due to negligible spectral overlap. The absorption band between 2.65-2.85 µm corresponds to CO₂ while the absorption bands between 2,15-2,45 µm, around 2,58 and beyond 3 µm correspond to CH₄. As expected, the CO₂-rich atmosphere has an absorbance at 2.65-2.85 µm higher than in the CH₄-rich one. We note that the 2-µm CO₂-related absorption band is absent in the CH₄-rich atmosphere. On the other hand, this absorption band is easily monitored for the CO₂-rich atmosphere. The same phenomenon is noticed for the 2.58 µm CH₄-related absorption band, which can be properly monitored mainly in CH₄-rich atmosphere. Finally, for a pressure above 330 mbar for each gas respectively, we have access to all absorption bands in the wavelength range of interest. Then we check the reliability of our SAS system by indicating the corresponding absorption areas of the two gas mixtures on each previous calibration curves of CH₄ and CO₂ (see magenta points on figures 6b and 8b). We observe that our measurements are consistent and that this developed SAS system can simultaneously detect CO₂ and CH₄ from their absorption bands between 2 and 3.5 µm and quantify the gas mixture. From these proof of principle experiments, a way to increase significantly the sensitivity of the set-up would be to increase the optical path of the multipass cell.




Figure 9: Absorbance curves of two gas mixtures for total pressure of 1-bar: 66% of CO2 and 33% of CH4 (red curve) and 33% of CO2 and 66% of CH4 (blue curve)

7. Conclusion

In conclusion, we have investigated mid-infrared gas sensing by means of a simple configuration of supercontinuum absorption spectroscopy based on a high-repetition-rate mid-infrared supercontinuum light generated through nonlinear propagation of 200-fs pulses in a 10-cm-long segment of step-index tellurite fiber. In particular, a compact multipass cell is used for analyzing absorption of CO₂ and CH₄ gases and their mixture in the 2-3.5 µm spectral range. It is worth mentionning that methane has a polluting power 32 times higher than CO₂ [23] and it deserves to be easily measured as well as CO₂. In the case of CO₂, there is about 400 ppm of CO₂ in the air, which would correspond to a CO₂ pressure in the cell of our experimental set-up of 0.4 mbar. In the case of methane its atmospheric concentration is about 1800 ppb corresponding to a cell CH₄ pressure of 1.8 μbar). These values are below the sensitivity threshold of our set-up. However, this work is a proof-of-principle experiment which illustrates the feasibility of multiple detection of greenhouse gases using SAS method. It can be much improved in particular by incresasing the optical path of the sensing cell and by using a more intense and larger supercontinuum up to 5.3 µm which is reachable with this type of fibers [17]. In particular, in this case, the wavelength of 4.26 µm may be reached allowing to measure low concentrations of CO₂ and CH₄. In addition, we can also mention that we use free space coupling at the input and output of our tellurite fiber. An optimization towards an all-fiber spliced device would also improve the set-up as well as the use of new pulsed fs fibered lasers entering the market as a pump source. In addition, by increasing the signal to noise ratio of our set-up, it will be possible to work at low pressures in the linear regime of the detection. Finally, this proof of principle experiment needs now to mature to develop its potential as a new multi-gas infrared optical sensor.

Acknowledgements

We acknowledge the financial support of the French Investissements d'Avenir program (ISITE-BFC ANR-15-IDEX-0003, SCUVIRA project), the Council of the French Région Bourgogne Franche-Comté, the European program FEDER, through the grant 'Jeune Chercheur Entrepreneur' scheme, and the French CNRS and the Japanese JSPS through the PRC CNRS-JSPS Program.

References

- [1] Ferreira M F et al 2017 J. Opt. **19** 083001
- [2] Zhao Z et al 2017 Laser Photonics Rev. 11 1700005
- [3] Petersen C R et al 2018 Opt. Lett. 43 999–1002

- [4] https://hitran.org/
- [5] Hughes R et al 2014 J Quant Spectrosc Radiat Transf 148 18–21
- [6] Brow D M et al 2008 Opt. Express 16 8457-8471
- [7] Nakaema W M et al 2011 Sensors 11 1620-1640
- [8] Hult J et al 2007 Opt. Express 15 11385-11395
- [9] Dupont S et al 2014 Las. Phys. Lett. 11 075601
- [10] Amiot C et al 2017 Appl. Phys. Lett. 111 061103
- [11] Mikkonen T et al 2018 Opt Lett 43 5094–5097
- [12] Dasa M K et al 2020 Photoacoustics 18 100163
- [13] Strutynski C et al 2016 Laser Phys Lett 13 075402
- [14] Strutynski C et al 2017 Opt. Fiber Technol. 38 154–159
- [15] Strutynski C et al 2017 Opt. Lett. 42 247-250
- [16] Froidevaux P et al 2018 Appl. Sci. 8 1875
- [17] Kedenburg S et al 2017 JOSA B **34** 601–607
- [18] Strutynski C et al 2016 JOSA B 33 D12–D18
- [19] O'Donnell M et al 2003 J. Non-Cryst. Solids **331** 48–57
- [20] Massera J et al 2011 J. Am. Ceram. Soc. 94 130–136
- [21] Savelii I et al 2013 Opt. Mater. 35 1595–1599
- [22] Humbach O et al 1996 J. Non-Cryst. Solids 203 19–26
- [23] Etminan M et al 2016 Geophys. Res. Lett. 43 12,614-12,623