



**HAL**  
open science

## Intraoperative electrophysiology during deep brain surgeries in disorders of consciousness

Jerome Coste, Bénédicte Pontier, Anna Sontheimer, Jean-Jacques Lemaire,  
Dbs-Post-Coma Study Group

### ► To cite this version:

Jerome Coste, Bénédicte Pontier, Anna Sontheimer, Jean-Jacques Lemaire, Dbs-Post-Coma Study Group. Intraoperative electrophysiology during deep brain surgeries in disorders of consciousness. Regis, Jean (Marseille) Gonçalves-Ferreira António (Lisbon). 24th congress, European Society for Stereotactic and Functional Neurosurgery, ESSFN 2021, Sep 2021, Marseille, France. Karger, Stereotactic and Functional Neurosurgery, 99 (Suppl. 1), 2021, 24th Congress, ESSFN 2021, Marseille, September 2021: Abstracts. hal-03468989

**HAL Id: hal-03468989**

**<https://hal.science/hal-03468989v1>**

Submitted on 7 Dec 2021

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

# #23424 Intraoperative electrophysiology during deep brain surgeries in disorders of consciousness

Jérôme Coste<sup>1,2</sup>, Bénédicte Pontier<sup>1,2</sup>, Anna Sontheimer<sup>1,2</sup>, Jean-Jacques Lemaire<sup>1,2</sup>, DBS-Post-Coma Study group<sup>3</sup>

<sup>1</sup>Service de Neurochirurgie, CHU de Clermont-Ferrand, France

<sup>2</sup>Institut Pascal, UMR6602 (UCA, CNRS), France

<sup>3</sup>CHU de Clermont-Ferrand, France; Centre Jean Perrin, France; Centres régionaux de Soins de Suite et de Réadaptation, France

## Introduction

The efficacy on deep brain stimulation in disorders of consciousness (DC) is inconclusive. We investigated bilateral 30-Hz low-frequency stimulation designed to overdrive neuronal activity by dual pallido-thalamic targeting to promote conscious behavior in five patients. Five adult patients (mean age 38 years old) were included: one unresponsive-wakefulness-syndrome male (traumatic brain injury); and four patients in a minimally conscious state, one male (traumatic brain injury) and three females (two hemorrhagic strokes and one traumatic brain injury). Pallido-thalamic electrode targeting were performed under sevoflurane general anesthesia with recording of extracellular multi-unit neuronal activity and local field potential (LFP). These data were compared with a motor disease group: five GPi DBS implantations (three males) for Parkinson's disease (n=4) or dystonia (59 years old) and three VIM DBS surgeries (two males) for Parkinson's disease (n=1) or essential tremor (40 years old).

## Method

The electrode implantation was realized after electrophysiological mapping using an intraoperative electrophysiological system for functional neurosurgery (MicroGuide Pro, Alpha Omega Eng., Israel). For each brain hemisphere, the mapping was performed using two exploration electrodes (Alpha Omega Eng.) steered by rigid guide tubes: one on the planned track and a second one on the parallel track located posteriorly at 2mm.

Electrophysiological approach

systematic recordings of the spontaneous neuronal activity every 1 mm during 30 seconds along the distal 8.2 mm (min: 5; max: 13 mm) in the pallidum and the thalamus previously labelled and manually outlined on stereotactic MRI.

## Results


**A.** Spikes were encountered for 50 out of 130 recordings in the pallidum (38.46%) and for 43 out of 185 (23.24%) in the thalamus of patients with disorders of consciousness. Concerning motor disease (MD) patients implanted under general anesthesia, spikes were found in 52 out of 150 recordings in the pallidum (34.66%) and for 54 out of 82 (65.85%) in the thalamus.

**B.** Mean firing rate of neurons in the pallidum was 9.84 +/-8.71 Hz and 11.95 +/-10.31 Hz during surgery for patients with DC and MD respectively. Concerning Thalamus, mean firing rate was 3.78 +/-4.87 Hz versus 2.25 +/-2.15 Hz (DC vs MD).

**C.** Percentage of raw power spectral density of LFP was similar in GPi for patients with DC and MD. Delta (47.8%), Theta (19.5%) and Alpha (17.0%) bands were preponderant for DC, while Delta (37.2%), Alpha (29.1%) and Beta (17.0%) range were powerful for MD patients. Pallidal neuronal activity is similar for patients under general anesthesia suffering from DC or MD. Differences were encountered for Thalamus: higher 7-30Hz frequency power was encountered for patients with MD and higher low (<7Hz) and high (>30Hz) band power was observed for DC.

## Discussion

Thalamic neuronal activity differs between MD and DC patients. Few records have been obtained and published for patients with impaired consciousness. These results should be considered with caution given the small number of patients in the study.



Jérôme Coste, PhD

jcoste@chu-clermontferrand.fr

CHU de Clermont Ferrand  
Hôpital Gabriel Montpied  
Service de Neurochirurgie, 4ème étage, extension 1  
58 rue Montalembert  
F-63003 CLERMONT-FERRAND Cedex 1, FRANCE

## Reference

Lemaire JJ, Sontheimer A, Pereira B, Coste J, Rosenberg S, Sarret C, Coll G, Gabrillargues J, Jean B, Gillart T, Coste A, Roche B, Kelly A, Pontier B, Feschet F. (2018) Deep brain stimulation in five patients with severe disorders of consciousness *Annals of Clinical and Translational Neurology* **5(11)**, 1372-1384

## Funding



This clinical trial (NCT01718249) was funded by the Fondation de l'Avenir (ET1-615 2011; "Deroche Legacy" Grant BO4-001), the University Hospital of Clermont-Ferrand, the Centre Jean Perrin and Medtronic France. DBS-Post-coma study group (3) and clinical research protocol are detailed here:

<https://onlinelibrary.wiley.com/action/downloadSupplement?doi=10.1002%2Facn3.648&file=acn3648-sup-0001-Supinfo.docx>