

Simultaneous quantification of dabrafenib, hydroxy-dabrafenib and trametinib in human plasma by liquid chromatography-tandem mass spectrometry

David Balakirouchenane, Nihel Khoudour, Sarah Guégan, Nora Kramkimel, Nathalie Franck, Thomas Rodier, François Goldwasser, Nicolas Dupin, Selim Aractingi, Michel Vidal, et al.

▶ To cite this version:

David Balakirouchenane, Nihel Khoudour, Sarah Guégan, Nora Kramkimel, Nathalie Franck, et al.. Simultaneous quantification of dabrafenib, hydroxy-dabrafenib and trametinib in human plasma by liquid chromatography-tandem mass spectrometry. Journal of Pharmaceutical and Biomedical Analysis, 2021, 193, pp.113718 -. 10.1016/j.jpba.2020.113718 . hal-03493871

HAL Id: hal-03493871 https://hal.science/hal-03493871v1

Submitted on 7 Nov 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

Version of Record: https://www.sciencedirect.com/science/article/pii/S0731708520316046 Manuscript_43572a991f59d0d00db3ae72411b31f6

1 Simultaneous quantification of dabrafenib, hydroxy-

2 dabrafenib and trametinib in human plasma by

3 liquid chromatography-tandem mass spectrometry

4

5	David Balakirouchenane ^{1,2*} , Nihel Khoudour ¹ , Sarah Guégan ^{3,4} , Nora Kramkimel ³ , Nathalie
6	Franck ³ , Thomas Rodier ^{1,2} , François Goldwasser ⁵ , Nicolas Dupin ^{3,4} , Selim Aractingi ^{3,4} ,
7	Michel Vidal ^{1,2} and Benoit Blanchet ^{1,2}
8	
9	¹ Department of Pharmacokinetics and Pharmacochemistry, Cochin Hospital, AP-HP,
10	CARPEM, 75014 Paris, France;
11	² UMR8038 CNRS, U1268 INSERM, Faculty of Pharmacy, University of Paris, PRES
12	Sorbonne Paris Cité, CARPEM, 75006 Paris, France;

- ³ Department of Dermatology, Cochin Hospital AP-HP, 75014 Paris, France;
- ⁴ Cochin Institute, INSERM U1016, University of Paris, 75014 Paris, France;
- ⁵ Department of Medical Oncology, Cochin Hospital AP-HP, 75014 Paris, France;

16

17 * Correspondence: david.balakirouchenane@aphp.fr

18 Abstract

A new liquid chromatography-tandem mass spectrometry (LC-MS/MS) method for the 19 20 simultaneous quantification of dabrafenib (DAB), its main metabolite hydroxy-dabrafenib (OHD) and trametinib (TRA) in human plasma has been developed and validated. After 21 22 addition of internal standard (dabrafenib-d9), extraction was achieved after protein precipitation with acetonitrile containing 1% (v/v) formic acid. Chromatographic separation 23 was performed on an Accucore[®] C18 (2.1×50 mm; 2.6μ m) column using a gradient elution 24 25 of water acidified with 0.1% (v/v) formic acid (A) and acetonitrile containing 0.1% (v/v) 26 formic acid (B) at a flow rate of 500 µL/min. The calibration ranged from 10 to 2000 ng/mL 27 for DAB and OHD and from 5 to 50 ng/mL for TRA. This method was validated with 28 satisfactory results including good precision (intra- and inter-assay coefficient of variation 29 from 2.0% to 14.9%) and good accuracy (inter- and intra-day bias between -1.2 % and 10.9 30 %), as well as long term stability in unprocessed plasma at -20°C. This newly proposed 31 method is useful for clinical research purposes as well as therapeutic drug monitoring for 32 patients with a Rapidly Accelerated Fibrosarcoma kinase B (BRAF)-mutated cancer.

33

34 Keywords: Therapeutic drug monitoring, Dabrafenib, Hydroxy-dabrafenib, Trametinib, Melanoma, LC-MS/MS

35 <u>1. Introduction</u>

Increased understanding of molecular pathways leads to the development of effective 36 targeted therapies for the treatment of solid tumors. Thus, a number of frequent driver 37 mutations (ie, genetic errors that control the disease) including Rapidly Accelerated 38 39 Fibrosarcoma kinase B (BRAF) have been identified in cancers such as melanoma and to a 40 lesser extent non-small cell lung cancer (NSCLC) and anaplatic thyroid cancer (ATC). 41 Targeted therapies in the form of signal transduction inhibitors, notably BRAF and MEK 42 inhibitors, have proven their efficacy in clinical trials [1]. Dabrafenib (DAB) and trametinib 43 (TRA) target the RAS-RAF-MEK-ERK transduction signal that regulates cellular growth and survival. DAB is a selective and potent inhibitor of serine/threonine kinase BRAF^{V600E/K} 44 45 mutants, whereas TRA is a MEK inhibitor [2]. These two drugs have been approved by the 46 US Food and Drug Administration (FDA) and the European Medicines Agency (EMA), alone 47 or in combination, for unresectable or metastatic melanoma, advanced NSCLC and ATC 48 [3,4].

49 DAB and TRA exhibit moderate to large interindividual variability in their 50 pharmacokinetics. Pharmacokinetic (PK) data from clinical trials highlight that predictive 51 factors such as age, sex or concomitant treatment can explain only in part this variability 52 which remains largely unexplained [2,5,6]. The magnitude of this PK variability could be 53 greater in unselected patients from "real life" who are more fragile than those included in 54 clinical trials for different reasons such as comorbidities, polymedication and age. Several 55 clinical studies reported higher plasma concentrations in patients experiencing dose-limiting 56 toxicities or specific toxicities such as pyrexia and fatigue for DAB [7]. The relationship 57 between plasma drug exposure and efficacy for DAB and TRA is currently controversial 58 [8,9], and requires further investigation. Moreover, DAB undergoes oxidative metabolism via cytochrome P450 (CYP) 3A4 and CYP2C8 to form hydroxy-dabrafenib (OHD), an active 59

metabolite, with a twofold higher potency than an inhibitor of mutant BRAF. The mean
OHD-to-DAB exposure ratio in plasma is 0.8 [2]; therefore, the pharmacodynamic
contribution of OHD in clinical outcomes deserves to be investigated.

63 Several DAB and TRA liquid chromatography-tandem mass spectrometry (LC-64 MS/MS) multi-analyte methods in human plasma have been described [10-16]. One of these 65 methods allows assaying simultaneously DAB and six metabolites (including OHD) in 66 plasma [17]. Because metabolite standards were not available, it is only semi-quantitative, 67 thus limiting its use to investigate pharmacokinetic/pharmacodynamic (PK/PD) relationship 68 for OHD. As far as we know, no quantitative method for OHD assay in plasma has been 69 reported yet.

In our laboratory, we recently conducted a PK/PD study in BRAF-mutated cancer patients treated concomitantly with DAB and TRA [6]. For this purpose, OHD was custom synthetized; then a LC-MS/MS method for OHD assay in plasma was developed to investigate exposure–response relationship for toxicity and efficacy in BRAF-mutated patients. The present study presents the development and the validation of a LC-MS/MS method for the simultaneous quantification of DAB, OHD and TRA in plasma from BRAFmutated cancer patients.

77 <u>2. Materials & methods</u>

78 <u>2.1. Standards and reagents</u>

DAB (519.6 g/mol, purity ≥98%) and TRA (615.4 g/mol, purity ≥98%) were
purchased from Toronto Research Chemical (Toronto, Canada). AlsaChim (Illkirch, France)
provided the isotopically labelled internal standard dabrafenib-d9 (528.6 g/mol, purity >98%).
OHD (535.6 g/mol, purity >99%) was synthesized by Shanghai Medicilon INC (Shanghai,
China). The chemical structures of these analytes are shown in Fig. 1.

Water was supplied by J.T.Baker (Gliwice, Poland); acetonitrile, methanol, dimethylsulfoxide
(DMSO) and formic acid (98%) were purchased from Sigma-Aldrich (St. Louis, MO, USA).
All these reagents were LC-MS grade. UTAK Laboratories (Valencia, Spain) was the
provider of drug-free heparinized human plasma.

88 <u>2.2. Preparation of calibration standards and internal quality control samples</u>

89 Stock solutions of DAB, OHD and TRA were prepared in DMSO at a concentration of 90 1 mg/mL. Two independent stock solutions were stored separately at -20°C, one for the 91 preparation of calibration standards (CS) and the other for internal quality controls (IQC). 92 Spiking solutions for calibration and control samples were prepared from two sets of working 93 solutions prepared in a mixture of methanol-DMSO (70:30, v/v) at 40 μ g/mL and 1 μ g/mL for 94 DAB and OHD, at 20 µg/mL and 1 µg/mL for TRA. A stock solution of internal standard (IS) 95 (1 mg/mL) was prepared in DMSO; then diluted in methanol-DMSO (70:30, v/v) to obtain a 96 working solution at 10 µg/mL. All these stock solutions were stored at -20°C in darkness.

97 The day of the analysis, CS and IQC were prepared by diluting spiking solutions with blank
98 plasma. CS consist of six levels: 10, 20, 50, 100, 500 and 2000 ng/mL for DAB and OHD; 5,
99 7.5, 10, 20, 30 and 50 ng/mL for TRA. IQC samples were prepared for low, medium and high
100 concentrations: 25, 200, 1500 ng/mL for DAB and OHD, and 8, 15 and 40 ng/mL for TRA.

101 <u>2.3. Sample preparation</u>

102 A 200 μ L aliquot of plasma (CS, IQC and patient sample) was mixed with 20 μ L of 103 working solution of IS (10 μ g/mL), then vortexed for 5 sec. Six hundreds microliters of 104 acetonitrile with 1% (v/v) formic acid were added for protein precipitation. Samples were 105 vortexed again before centrifugation at 13.000g for 10 min at room temperature (RT). The 106 supernatants were transferred into glass tubes to evaporate to dryness at 45°C under a nitrogen 107 stream. The dried extract was recovered with 200 μ L of water-methanol (70:30, v/v) with 108 0.1% (v/v) formic acid, then vortexed before the transfer into a 1.5 mL Eppendorf tube for a second centrifugation (13.000g for 10 min, RT). The final extracts were transferred into a
polypropylene vials; then 10 µL were injected into the LC-MS/MS system.

111 <u>2.4. Instrumentation</u>

112 The samples were analyzed using an Accela AS[®] auto-sampler with a quaternary 113 pump and degasser for chromatographic separation (ThermoFisher Scientific, Les Ulis, 114 France). A TSQ QuantumUltra[®] mass spectrometer equipped with an Ion Max electrospray 115 ionization (ESI) interface (ThermoFisher Scientific) was used for mass detection and 116 quantification. Data acquisition and processing were performed using Xcalibur[®] (version 117 2.0.7, ThermoFisher Scientific).

118 Chromatographic separation of three analytes (DAB, OHD and TRA) and IS was achieved on 119 Accucore[®] C18 (2.1×50 mm; 2.6μ m) analytical column (ThermoFisher Scientific, Les Ulis, 120 France) associated with a guard column packed with the same bonded phase. Column 121 temperature was maintained at 30°C and the auto-sampler at 4°C.

122 2.5. LC-MS/MS conditions

123 The mobile phase consisted of water (solution A) and acetonitrile (solution B), both 124 with 0.1% (v/v) formic acid. A solution of methanol-water (1:1, v/v) was used to flush the 125 auto-sampler after each injection. The mobile phase was delivered at a flow rate of 500 126 µL/min throughout the 13.2 min-run. The analytes were eluted following this multistep 127 gradient: starting at 95% A and 5% B, the proportion of B is linearly increased to 61.7% up to 10.0 min. From 10.0 min to 10.1 min the proportion of B is increased to 90% and the 128 129 conditions were maintained until 12.1 min. Finally, starting conditions (95% A and 5% B) 130 were maintained from 12.2 min to the end of run at 13.2 min.

131 The chromatography system was coupled with a TSQ Quantum triple quadrupole mass 132 spectrometer (ThermoFisher Scientific, Les Ulis, France). The ESI positive spray voltage was 133 3500 V; the ion transfer tube and vaporizer temperatures were 320°C and 300°C, respectively.

134 The sheath and auxiliary gas were nitrogen at a flow-rate set of 45 and 30 (arbitrary units) 135 respectively; the Q2 (second quadrupole) collision gas (argon) pressure was 1 mTorr. Q1 and 136 Q3, respectively the first and third quadrupole, mass resolution was 0.7 amu full width at half 137 maximum (FWHM). Scan width and scan time were 0.01 amu and 15 msec, respectively. 138 Two optimal m/z transitions were chosen for each analyte: one for quantification and the 139 second one for confirmation in order to have a specific method. Analyses were performed 140 using the selected reaction monitoring (SRM) detection mode. The transitions and general 141 parameters are presented in Table 1.

142 <u>2.6. Method validation</u>

143 The method validation was based on international recommendations [18,19].

144 <u>2.6.1. Limit of quantification and linearity</u>

This method was developed for therapeutic drug monitoring and clinical research. The
CS values for DAB and TRA were determined from observed concentrations during clinical
trials [3,4]. Given the mean OHD-to-DAB exposure ratio in plasma is 0.8 [2], CS values for
OHD were similar to those of DAB.

Calibration curves for each compound were plotted using the ratio of the analyte's peak area over the IS's peak area versus analyte concentration. The back-calculated concentration for each standard was to be less than 20% of the lower limit of quantification (LLOQ) nominal concentration and less than 15% of the other calibration values.

153

2.6.2. Selectivity and carryover

The selectivity of this analytical method was assessed by analyzing plasma samples (n = 10) from polymedicated cancer patients (including BRAF mutated cancer patients) to rule out any interference from others drugs with DAB, OHD, TRA and dabrafenib-d9. The potential interfering signal was to be less than 20 % of the LLOQ for each analyte and less than 5 % for IS. The carry-over was studied by processing blank plasma (n = 6) after the CS at the upper limit of quantification (ULOQ). Carry-over should not be greater than 20% of the
LLOQ for each analyte, and 5% of the peak area for internal standard.

161 <u>2.6.3. Accuracy and precision</u>

Accuracy was calculated as the percentage of deviation (% bias) between the measured and the nominal concentration. Precision was evaluated as the coefficient of variation (% CV) of the measured concentrations. Intra-day accuracy and precision were assessed by six replicates of LLOQ and each IQC level analyzed in one run the same day. Inter-assay accuracy and precision were evaluated by three replicates of LLOQ and each IQC level on six different days. The analytical assay is considered valid if bias and precision are within \pm 15% (\pm 20% for LLOQ and the low IQC).

169

2.6.4. Matrix effect, extraction recovery and process efficiency

Three sets of each IQC level (n = 6) with IS were prepared to study matrix effect (ME), extraction recovery (ER) and process efficiency (PE): A) IQC and IS solution were prepared in the recovery phase (water-methanol (70:30, v/v) with 0.1% (v/v) formic acid), B) Blank plasma from 6 different pools were extracted and then spiked with the IQC and IS solution, C) IQC were prepared in 6 different batch of drug-free plasma (UTAK) and processed accordingly to the method described in 2.3.

ME, ER and PE were calculated as the ratio of each compounds peak areas (%) of B/A, C/B
and C/A, respectively.

178

2.6.5. Dilution integrity

Dilution integrity was investigated by charging a plasma sample with the three analytes at a concentration 1.5 times higher than their respective ULOQ, and then by diluting it 2- and 5-fold with drug-free plasma. Dilutions were repeated 6 times. To meet the validation criterion, accuracy and precision were to be within \pm 15% of the expected value for each dilution. 184

2.6.6. Stability

The stability of DAB, OHD and TRA in plasma was investigated in different conditions for each level of IQC (n = 3). Three freeze-thaw cycles were performed with thawing at RT and freezing at -20°C in one day, plasma samples were assayed after storage for 24 hours at RT, 48 hours at 4°C, 30 days and 6 months at -20°C. The stability in the autosampler was studied by re-analyzing three levels of IQC (n = 3 for each level) after 48 hours at 4°C. The results were accepted if the mean percentage change from nominal concentration was less than 85%.

192 <u>2.7. Clinical applications</u>

193 The present LC-MS/MS method was used for PK/PD study conducted in mutated 194 BRAF cancer patients treated with DAB and TRA [6]. This study was approved by the local ethics committee in oncology (N° CLEC 211218ACBB1). Patients gave their informed 195 196 consent before blood collection. In the present study, all patients were concomitantly treated 197 with the recommended daily doses of DAB (150 mg twice a day) and TRA (2 mg once a day). 198 Twenty-five blood samples were collected at steady state in heparin lithium-containing tubes 199 at different time-after-dose over administration intervals. After centrifugation (1850g, 4°C, 10 200 min), until plasma was collected and transferred into polypropylene test tubes before storage 201 at -20°C up to subsequent analysis.

Reanalysis of 12 incurred samples (including high and low concentrations) was performed to
ensure the reliability of the present method. The two analyses were considered equivalent if at
least 67% of the results percentage difference [(repeat-original)/mean*100] was within 20%
[19].

- 206 **<u>3. Results and discussion</u>**
- 207 <u>3.1. Development</u>

Finding the best calibration range was challenging since DAB plasma concentrations are 50 to 100 times higher than TRA concentrations. We based LLOQ and ULOQ on previously published methods and on our experience of TDM [10,12]. Since these molecules are highly lipophilic [3,4], stock solutions were made out of DMSO and the four working solutions in methanol-DMSO (70:30, v/v). Dabrafenib-d9 was chosen as sole internal standard for the three analytes in order to have a simple method that achieved sufficient sensibility.

The extraction process and choice of an Accucore[®] C18 column was based on a previously developed method by our laboratory[20] which resulted in an adequate chromatographic separation and satisfactory results. The retention times (RT) and MS/MS settings are reported in Table 1.

219 <u>3.2. Chromatograms</u>

Fig. 2 displays typical chromatographic profiles of blank plasma and plasma samples at the LLOQ and ULOQ for the three analytes and the IS.

222 <u>3.3. Limit of quantification and linearity</u>

Calibration curves ranged from 10 to 2000 ng/mL for DAB and OHD and from 5 to 50 ng/mL for TRA. They were best described by a 1/x weighted linear regression of the ratio of each molecule over IS peak's area versus the concentrations of the respective molecule in each standard sample. Out of 6 calibrations curves, each CS of the three analytes had a CV lower than 8%. Moreover, regression coefficients (r²) of calibration curves for DAB, OHD and TRA were higher than 0.99 showing a good linearity of the method for each analyte.

229 3.4. Selectivity and carry-over

No interference was detected in 10 plasma samples from polymedicated cancerpatients for whom we knew their medications. Overall, these plasma samples contained 60

different drugs frequently administered in cancer patients. Endogenous signals were lowerthan 1% of the signal at the LLOQ of each analyte and IS.

Regarding carry-over, the signal of blank plasma samples processed after injection of ULOQ samples was less than 5% of the corresponding peak area of the LLOQ sample and the response of the IS was less than 1% of the normal response.

237 <u>3.5. Accuracy and precision</u>

The results of accuracy and precision are presented in Table 2. Mean intra- and interday accuracies were below 9.3% and 10.9% for each analyte, respectively. Regarding intraand inter-day precisions, CVs were less than 13.5% and 14.9% respectively. Overall, the results for precision and accuracy met the validation criterion, which proves both the reproducibility and reliability of the analytical method.

243 <u>3.6. Matrix effect, extraction recovery and process efficiency</u>

The mean ME, ER and PE ranged from 88.0% to 112.2%, from 61.4% to 87.1% and from 59.1% to 89.9%, respectively (Table 3). The higher polarity of OHD could contribute to its greater ER and PE compared with DAB.

247 <u>3.7. Dilution integrity</u>

Accuracy and precision for the 2-fold dilution ranged from -2.1% to 0.2% and from 3.1% to 4.3%, respectively; for the 5-fold dilution, from -0.8% to 2.2% and from 3.3% to 13.0%, respectively.

251 <u>3.8. Stability</u>

Following three freeze-thaw cycles, results met the validation criterion for all analytes (supplementary Table 1). Short-term stability was confirmed at RT and at 4°C, which will allow sufficient time to transport samples from clinical department to the laboratory in daily practice. Long-term stability studies showed that DAB, OHD and TRA were stable at -20°C for up to 6 months. Finally, all analytes were stable up to 48 hours in the auto-sampler at 4°C.

This new LC-MS/MS method was successfully applied to 25 plasma samples from 258 259 BRAF-mutated cancer patients including metastatic melanoma (n = 20), NSCLC (n = 2) and 260 ATC (n = 3). The median concentrations [range] were 332 ng/mL [36-2273], 335 ng/mL [29-261 1416] and 12.6 ng/mL [5.5-21.6] for DAB, OHD and TRA respectively. As previously 262 reported, trough and peak concentrations for DAB and TRA were in the expected range for 263 patients treated with the recommended daily dose. In the same way, median OHD-to-DAB 264 ratio was 0.8 [0.5-2.8], which is in accordance with the expected value [2]. The results of the 265 incurred sample reanalysis showed the lack of any significant difference (<20%) between the 266 initial analysis and the second one for all the analytes (Fig. 3), which confirmed the satisfying 267 reproducibility observed with IQC. Taken together, these results prove the suitability of our 268 method to monitor DAB, OHD and TRA in plasma from BRAF-mutated cancer patients. 269 Vikingsson et al. [17] reported a method to assay DAB and six metabolites but only through 270 semi-quantitative calculation using the calibration curve for DAB. To our best knowledge, the 271 present method is the first to simultaneously quantify DAB and its main active metabolite 272 OHD. We recently reported that plasma ratio $AUC_{OHD}/AUC_{DAB} \ge 1$ is independently 273 associated with shorter overall survival (Hazard Ratio: 10.61 (2.34-48.15), p < 0.05), which 274 suggests a clinical benefit to monitor OHD in daily clinical practice [6].

Finally, this method offers the advantage of simultaneously monitoring TRA in patients treated in association with DAB and TRA. This may contribute to more widespread plasma monitoring of these drugs in BRAF-mutated cancer patients in daily clinical practice.

278 **<u>4. Conclusion</u>**

We have developed and validated a LC-MS/MS method to simultaneously quantify DAB, OHD and TRA in human plasma. Patients treated with DAB and TRA exhibit a large interindividual variability in clinical outcomes in terms of toxicity and efficacy, which has led to a growing interest in PK/PD investigations. This method is suitable for studying OHD
 pharmacological activity especially in daily clinical practice.

284 CRediT authorship contribution statement

285 David Balakirouchenane: Data curation, Formal analysis, Methodology, 286 Visualization, Writing - original draft. Nihel Khoudour: Formal analysis, Methodology, 287 Supervision, Writing - review & editing. Sarah Guégan: Data curation, Investigation, 288 Writing - review & editing. Nora Kramkimel: Data curation, Investigation, Writing - review 289 & editing. Nathalie Franck: Data curation, Investigation, Writing - review & editing. 290 Thomas Rodier: Data curation, Writing - review & editing. François Goldwasser: Funding 291 acquisition, Writing - review & editing. Nicolas Dupin: Data curation, Investigation, Writing 292 - review & editing. Selim Aractingi: Data curation, Investigation, Writing - review & editing. 293 Michel Vidal: Conceptualization, Methodology, Resources, Supervision, Writing - review & 294 editing. Benoit Blanchet: Conceptualization, Methodology, Supervision, Writing - review & 295 editing.

296 Declaration of Competing Interest

Benoit Blanchet has served on advisory boards and received honoraria from AstraZeneca, Boehringer Ingelheim, Bristol-Myers Squibb, Leo Pharma, and Pfizer. Francois Goldwasser has received a travel, accommodation, and research grant from Bristol-Myers Squibb. Nicolas Dupin reports personal fees, travel, accommodations, and expenses from Roche. Selim Aractingi has been a speaker for Pierre Fabre, Novartis, Janssen, and Abbvie and a member of boards of meetings for BMS and Sanofi.

303 Other authors have declared no conflict of interest.

304 **<u>References</u>**

305	[1] C. Robert, JJ. Grob, D. Stroyakovskiy, B. Karaszewska, A. Hauschild, E. Levchenko, V.
306	Chiarion Sileni, J. Schachter, C. Garbe, I. Bondarenko, et al, Five-Year Outcomes with
307	Dabrafenib plus Trametinib in Metastatic Melanoma. N Engl J Med. 381 (2019) 626-636.

- 308 [2] A. Puszkiel, G. Noé, A. Bellesoeur, N. Kramkimel, NM. Paludetto, A. Thomas309 Schoemann, M. Vidal, F. Goldwasser, E. Chatelut, B. Blanchet, Clinical
 310 Pharmacokinetics and Pharmacodynamics of Dabrafenib, Clin Pharmacokinet. 58 (2019)
 311 451-467.
- 312 [3] European Medicines Agency, Committee for Medicinal Products for Human Use (CHMP)
 313 assessment report: Tafinlar (dabrafenib), 2013,
- 314 https://www.ema.europa.eu/en/documents/product-information/tafinlar-epar-product-
- 315 information_en.pdf (accessed February 20, 2020).
- 316 [4] European Medicines Agency, Committee for Medicinal Products for Human Use (CHMP)
 317 assessment report: Mekinist (trametinib), 2014,
- 318 https://www.ema.europa.eu/en/documents/product-information/mekinist-epar-product-
- 319 information_en.pdf (accessed February 20, 2020).
- [5] D. Ouellet, N. Kassir, J. Chiu, MS. Mouksassi, C. Leonowens, D. Cox, DJ. DeMarini, O.
 Gardner, W. Crist, K. Patel, Population pharmacokinetics and exposure-response of
 trametinib, a MEK inhibitor, in patients with BRAF V600 mutation-positive melanoma,
 Cancer Chemother Pharmacol. 77 (2016) 807-17.
- 324 [6] D. Balakirouchenane, S. Guégan, C. Csajka, A. Jouinot, V. Heidelberger, A. Puszkiel, O. 325 Zehou. N. Khoudour. P. Courlet, N. Kramkimel. et al, Population 326 Pharmacokinetics/Pharmacodynamics of Dabrafenib Plus Trametinib in Patients with 327 BRAF-Mutated Metastatic Melanoma, Cancers. 12 (2020) 931.

- [7] KT. Flaherty, C. Robert, P. Hersey, P. Nathan, C. Garbe, M. Milhem, LV. Demidov, JC.
 Hassel, P. Rutkowski, P. Mohr, et al, Improved survival with MEK inhibition in BRAFmutated melanoma, N Engl J Med. 367 (2012) 107-14.
- [8] M. Rousset, C. Dutriaux, P. Bosco-Lévy, S. Prey, A. Pham-Ledard, L. Dousset, E.
 Gérard, S. Bouchet, M. Canal-Raffin, K. Titier, et al, Trough dabrafenib plasma
 concentrations can predict occurrence of adverse events requiring dose reduction in
 metastatic melanoma. Clin Chim Acta. 472 (2017) 26-29.
- 335 [9] HY. Kim, JK. Duong, M. Gonzalez, GV. Long, AM. Menzies, H. Rizos, SY. Lim, J. Lee,
- AV. Boddy, Pharmacokinetic and cytokine profiles of melanoma patients with dabrafenib
- and trametinib-induced pyrexia. Cancer Chemother Pharmacol. 83 (2019) 693-704.
- [10] M. Rousset, K. Titier, S. Bouchet, C. Dutriaux, A. Pham-Ledard, S. Prey, M. CanalRaffin, M. Molimard, An UPLC-MS/MS method for the quantification of BRAF
 inhibitors (vemurafenib, dabrafenib) and MEK inhibitors (cobimetinib, trametinib,
 binimetinib) in human plasma. Application to treated melanoma patients. Clin Chim Acta.
 470 (2017) 8-13.
- [11] M. Herbrink, N. de Vries, H. Rosing, ADR. Huitema, B. Nuijen, JHM. Schellens, JH.
 Beijnen, Development and validation of a liquid chromatography-tandem mass
 spectrometry analytical method for the therapeutic drug monitoring of eight novel
 anticancer drugs. Biomed Chromatogr. 32 (2018) e4147.
- 347 [12] CM. Nijenhuis, H. Haverkate, H. Rosing, JH. Schellens, JH. Beijnen, Simultaneous
 348 quantification of dabrafenib and trametinib in human plasma using high-performance
 349 liquid chromatography-tandem mass spectrometry. J Pharm Biomed Anal. 125 (2016)
 350 270-279.

- [13] H. Huynh, C. Pressiat, H. Sauvageon, I. Madelaine, P. Maslanka, C. Lebbé, C.
 Thieblemont, L. Goldwirt, S. Mourah, Development and Validation of a Simultaneous
 Quantification Method of 14 Tyrosine Kinase Inhibitors in Human Plasma Using LCMS/MS. Ther Drug Monit. 39 (2017) 43-54.
- [14] E. Cardoso, T. Mercier, AD. Wagner, K. Homicsko, O. Michielin, K. Ellefsen-Lavoie, L.
 Cagnon, M. Diezi, T. Buclin, N. Widmer, et al, Quantification of the next-generation oral
 anti-tumor drugs dabrafenib, trametinib, vemurafenib, cobimetinib, pazopanib,
 regorafenib and two metabolites in human plasma by liquid chromatography-tandem mass
 spectrometry. J Chromatogr B Analyt Technol Biomed Life Sci. 1083 (2018) 124-136.
- 360 [15] C. Merienne, M. Rousset, D. Ducint, N. Castaing, K. Titier, M. Molimard, S. Bouchet,
 361 High Throughput Routine Determination of 17 Tyrosine Kinase Inhibitors by LC362 MS/MS. J Pharm Biomed Anal. 150 (2018) 112-120.
- 363 [16] SD. Krens, E. van der Meulen, FGA. Jansma, DM. Burge, NP. van Erp, Quantification of
 364 cobimetinib, cabozantinib, dabrafenib, niraparib, olaparib, vemurafenib, regorafenib and
 365 its metabolite regorafenib M2 in human plasma by UPLC-MS/MS. Biomed Chromatogr.
 366 34 (2020) e4758.
- 367 [17] S. Vikingsson, JO. Dahlberg, J. Hansson, V. Höiom, H. Gréen, Simple and cost-effective
 368 liquid chromatography-mass spectrometry method to measure dabrafenib quantitatively
 369 and six metabolites semi-quantitatively in human plasma. Anal Bioanal Chem. 409 (2017)
 370 3749-3756.
- 371 [18] EMA, Guideline on bioanalytical method validation,
- 372 http://www.ema.europa.eu/docs/en_GB/document_library/Scientific_guideline/2011/08/
- 373 WC500109686.pdf (accessed April 28, 2020).

- [19] FDA, Homepage, Guidance for Industry: Bioanalytical Method Validation,
- 375 http://www.fda.gov/downloads/Drugs/GuidanceComplianceRegulatoryInformation/
- 376 Guidances/UCM070107.pdf (accessed April 28, 2020).
- [20] R. Reis, L. Labat, M. Allard, P. Boudou-Rouquette, J. Chapron, A. Bellesoeur, A.
 Thomas-Schoemann, J. Arrondeau, F. Giraud, J. Alexandre, et al, Liquid
 chromatography-tandem mass spectrometric assay for therapeutic drug monitoring of the
 EGFR inhibitors afatinib, erlotinib and osimertinib, the ALK inhibitor crizotinib and the
 VEGFR inhibitor nintedanib in human plasma from non-small cell lung cancer patients. J
 Pharm Biomed Anal. 158 (2018) 174-183.







396 Table 1

397	MS/MS general settings for dabrafenib (DAB), hydroxy-dabrafenib (OHD), trametinib (TRA)
398	and the internal standard (IS).

	Analyte	Expected RT (min)	ESI	Precursor ion (m/z)	Product ion (m/z)		CE (eV)	TL (V)
	DAB	7.9	+	520.1	Quantification Confirmation	306.9 276.9	29 57	154
	OHD	5.9	+	536.1	Quantification Confirmation	322.9 292.9	30 44	121
	TRA	8.2	+	616.1	Quantification Confirmation	254.0 491.0	37 30	124
200	IS	7.8	+	529.2	Quantification Confirmation	316.0 279.9	30 66	142
399 400	CE, Collision Ener	rgy; eV, electron volt; I	S , internal st	andard; ESI, Electrospra	ay Ionization; m/z, masse/charge;	min, minute; RT, R	etention Time; TL, Tu	be Lens; V, volt.
401 402								
403 404								
405 406								
400								
408 409								
410								
411 412								
413 414								
415								
416 417								
418								
419 420								
421								
422								
423 424								
425								
426								
427								
428								
430								
431								
432 433								
434								
435								



Fig. 2. Chromatograms of A) blank plasma, B) the lower limit of quantification and C) the
upper limit of quantification containing dabrafenib, hydroxy-dabrafenib, trametinib and
internal standard, from top to bottom respectively.

Table 2

Intra- and inter- assay accuracy and precision of dabrafenib (DAB), hydroxy-dabrafenib (OHD) and trametinib (TRA) at lower limit of quantification, low, medium and high

concentrations

400	concentratio	ns.				
	Nominal concentration	Level	Intra-day accuracy $(\% \text{ bias, } n = 6)$	Intra-day precision $(\% \text{ CV}, n = 6)$	Inter-day accuracy $(\% \text{ bias, } n = 18)$	Inter-day precision $(\% \text{ CV}, n = 18)$
	(ng/mL)					
	DAB					
	10	LLOQ	8.1	4.3	9.9	4.0
	25	IQC1	1.3	2.0	4.2	6.9
	200	IQC2	6.9	2.3	1.6	8.6
	1500	IQC3	0.9	3.1	10.9	4.8
	OHD					
	10	LLOQ	5.2	6.7	6.2	7.8
	25	IQC1	0.9	4.1	1.2	6.6
	200	IQC2	2.9	3.3	-1.2	4.7
	1500	IQC3	5.9	4.8	10.1	2.8
	TRA	-				
	5	LLOQ	6.6	12.3	0.6	11.0
	8	IQC1	1.4	12.5	2.5	14.9
	15	IQC2	3.9	10.5	2.5	10.4
161	40	IQC3	9.3	13.5	0.8	8.7
482	CV, Coefficient of variat quantification	ion; IQC, interna	al quality control; IQC1, IQC2 and IQ	C3 correspond to the IQC concentratio	ns at low, medium and high level, res	pectively; LLOQ, lower limit of
463	•					
464						
465						
466						
407						
408 469						
470						
471						
472						
473						
474						
475						
476						
477						
4/8						
4/9						
400						
482						
483						
484						
485						
100						



516 Table 3

Recovery and matrix effect for dabrafenib (DAB), hydroxy-dabrafenib (OHD), trametinib (TRA) and internal standard (IS) in human heparinized plasma samples (n = 6 for each 517

518

analyte and n = 18 for the IS). 519

		Matrice effect (%)		Extraction recovery (%)		Process efficiency (%)	
Nominal concentration (ng/mL)	Level	Mean	CV	Mean	CV	Mean	CV
DAB							
25	IQC1	94.6	5.7	64.2	7.6	60.7	9.0
200	IQC2	96.5	2.2	65.2	9.5	62.9	9.2
1500	IQC3	88.0	6.5	67.4	10.4	59.1	10.0
OHD							
25	IQC1	104.5	6.1	85.3	5.0	86.7	10.6
200	IQC2	106.9	3.6	86.3	4.1	89.9	4.5
1500	IQC3	93.7	4.5	87.1	4.3	81.5	2.9
TRA							
8	IQC1	112.2	13.8	61.4	6.1	68.6	12.9
15	IQC2	105.1	13.8	67.6	10.7	67.6	14.7
40	IQC3	99.8	11.8	66.3	11.8	63.2	5.4
IS		93.9	5.3	70.1	9.0	65.7	9.7

520 CV. coefficient of variation. IS. internal standard.