



**HAL**  
open science

# Implantable Blood Pressure Sensors with Analogic Thermal Drift Compensation

Serigne Modou Die Mbacke, Mohammed El Gibari, Benjamin Lauzier,  
Chantal Gautier, Hongwu Li

► **To cite this version:**

Serigne Modou Die Mbacke, Mohammed El Gibari, Benjamin Lauzier, Chantal Gautier, Hongwu Li. Implantable Blood Pressure Sensors with Analogic Thermal Drift Compensation. 8th International Symposium on Sensor Science, May 2021, Dresde (Fully online event), Germany. pp.34, 10.3390/I3S2021Dresden-10126 . hal-03512283

**HAL Id: hal-03512283**

**<https://hal.science/hal-03512283v1>**

Submitted on 2 Jun 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Abstract

# Implantable Blood Pressure Sensors with Analogic Thermal Drift Compensation <sup>†</sup>

Serigne Modou Die Mbacke <sup>1,\*</sup>, Mohammed El Gibari <sup>1,\*</sup>, Benjamin Lauzier <sup>2,\*</sup> , Chantal Gautier <sup>2</sup>  
and Hongwu Li <sup>1,\*</sup>

<sup>1</sup> Faculté des Sciences et Techniques, Université de Nantes, CNRS, IETR UMR 6164,  
2 Chemin de la Houssinière, 44322 Nantes, France; serigne.mbacke@univ-nantes.fr

<sup>2</sup> L'Institut Du Thorax, Université de Nantes, INSERM, CNRS, INSERM UMR 1087/CNRS UMR 6291,  
8 Quai Moncoussu, 44007 Nantes, France; gautier.chant@orange.fr

\* Correspondence: mohammed.el-gibari@univ-nantes.fr (M.E.G.); benjamin.lauzier@univ-nantes.fr (B.L.);  
hongwu.li@univ-nantes.fr (H.L.)

<sup>†</sup> Presented at the 8th International Symposium on Sensor Science, 17–28 May 2021; Available online:  
<https://i3s2021dresden.sciforum.net/>.

**Abstract:** Implantable pressure sensors represent an important part of the research activity in laboratories. Unfortunately, their use is limited by cost, autonomy and temperature-related drifts. The cost of use depends on several parameters, particularly their low battery life and the need for miniaturization to be able to implant the animals and monitor them over a time that is long enough to be physiologically relevant. This paper studied the possibility of reducing the thermal drift of implantable sensors. To quantify and compensate for the thermal drift, we developed the equivalent model of the piezoresistive probe by using the Cadence software. Our model takes into account the temperature (34–39 °C) as well as the pressure (0–300 mmHg). We were thus able to identify the source of the drift and thanks to our model, we were able to compensate for it thanks to the compensation circuits added to the conditioning circuits of the sensor. The maximum relative drift of the sensor is (0.1 mV/°C)/3.6 mV (2.7%), a drift of the conditioning circuit is (0.98 mV/°C)/916 mV (0.1%) and the whole is (13.4 mV/°C)/420 mV (32%). The compensated sensor shows a relative maximum drift of (0.371 mV/°C)/405 mV (0.09%). The output voltage remains stable over the measurement temperature range.

**Keywords:** thermal drift; Cadence simulation; blood pressure sensors; analogic compensation; piezoresistive



**Citation:** Mbacke, S.M.D.; El Gibari, M.; Lauzier, B.; Gautier, C.; Li, H. Implantable Blood Pressure Sensors with Analogic Thermal Drift Compensation. *Eng. Proc.* **2021**, *6*, 34. <https://doi.org/10.3390/I3S2021Dresden-10126>

Academic Editors: Gianarelio Cuniberti and Larysa Baraban

Published: 17 May 2021

**Publisher's Note:** MDPI stays neutral with regard to jurisdictional claims in published maps and institutional affiliations.



**Copyright:** © 2021 by the authors. Licensee MDPI, Basel, Switzerland. This article is an open access article distributed under the terms and conditions of the Creative Commons Attribution (CC BY) license (<https://creativecommons.org/licenses/by/4.0/>).

**Supplementary Materials:** The following are available online at <https://www.mdpi.com/article/10.3390/I3S2021Dresden-10126/s1>.

**Institutional Review Board Statement:** Not applicable.

**Informed Consent Statement:** Not applicable.

**Data Availability Statement:** The data presented in this study are available on request from the corresponding author.