

## DIFFUSE REFLECTANCE SPECTROSCOPY WITH POLARIZATION GATING

## <u>Anabela Da Silva,\*</u> Susmita Shridar, Hind Oulhaj, Callum M. Macdonald, Ugo Tricoli, Vadim Markel <sup>1</sup>Aix Marseille Univ, CNRS, Centrale Marseille, Institut Fresnel UMR 7249, 13013 Marseille, France \*anabela.dasilva@fresnel.fr

Keywords: polarization gating imaging, Monte Carlo simulations

Spectrometry involves the analysis of chromophore content by observing the wavelength dependent attenuation of light which has propagated through a sample (Beer-Lambert law). Knowledge of the pathlength through which light has travelled is needed to treat this problem. In the presence of scattering, as in biological tissues, the optical pathlength depends non-trivially on the wavelength. As a result, the relationship between the measured attenuation, and the medium absorption, becomes non-linear. Polarization filtering to linearize this problem in the transmission geometry [1-2] allows simple models, such as the modified Beer-Lambert law to still be applied. We are interested in solving this problem in the backscattering geometry, using wide field illumination, and wide field detection (Fig.1).




Fig. 1 : Left : Polarization gating Diffuse Reflectance setup ; Right : Examination of a burn scar, examples of measurements obtained with different ellipticities of polarization.

If a constant pathlength can be achieved through polarization modulation, then consistency of the probing volume will also likely improve, reducing errors incurred by mesoscopic inhomogeneities in the medium under investigation. In previous work, we have shown that modulation of the ellipticity of polarization, from linear to circular, allowed to sense the tissues at increasing depth [3-5]. Here, with the development of an optimized Monte Carlo simulation software [6-7], we investigate the use of polarization filtering with improved combinations of linear, elliptical, or circular, in order to improve this selectivity of the effective optical pathlength.

[1] I. M. Stockford, B. Lu, J. A Crowe, S. P. Morgan, Applied Spectroscopy, 61, 1379 (2007).

[2] J. M. Schmitt, A. H. Gandjbakhche, R. F. Bonner, Applied Optics, 31, 6535 (1992).

[3] A. Da Silva, C. Deumié, and I. Vanzetta, "Elliptically polarized light for depth resolved optical imaging", Biomedical Optics Express 3(11), 2907-2915, 2012.

[4] S. Rehn, A. Planat-Chrétien, M. Berger, J.-M. Dinten, C. Deumié, and A. da Silva," Depth Probing of Diffuse Tissues Controlled with Elliptically Polarized Light," J. Biomedical Optics 18(1), 016007, 2013.

[5] S. Sridhar, and A. Da Silva, Journal of Biomedical Optics, 21, 071107 (2016).

[6] C. M. Macdonald, U. Tricoli, A. Da Silva, And V. A. Markel, "Numerical investigation of polarization filtering for direct optical imaging within scattering media", JOSAA 34(8), (2017).

[7] U. Tricoli, C. M. Macdonald, A. Da Silva, And V. A. Markel, "The optimized diffusion approximation," JOSAA 35(2), 2018