

Characterization of stilbene and EJ-276 scintillators coupled with a large area SiPM array for a fast neutron dose rate detector

F. Ferrulli, N. Dinar, L. Gallego Manzano, M. Labalme, M. Silari

► To cite this version:

F. Ferrulli, N. Dinar, L. Gallego Manzano, M. Labalme, M. Silari. Characterization of stilbene and EJ-276 scintillators coupled with a large area SiPM array for a fast neutron dose rate detector. Nuclear Instruments and Methods in Physics Research Section A: Accelerators, Spectrometers, Detectors and Associated Equipment, 2021, 1010, pp.165566. 10.1016/j.nima.2021.165566 . hal-03552889

HAL Id: hal-03552889 https://hal.science/hal-03552889v1

Submitted on 2 Aug 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution - NonCommercial 4.0 International License

Characterisation of stilbene and EJ-276 scintillators coupled with a large area SiPM array for a fast neutron dose rate detector.

F. Ferrulli^{1,2,*}, N. Dinar², L. Gallego Manzano^{2,3}, M. Lablme¹, M. Silari²

¹Université de CAEN, CAEN, France

²CERN, 1211 Geneve 23, Switzerland

³Now at the Institute of Radiation Physics, Lausanne University Hospital and Lausanne University, Switzerland

1 Abstract—A 1-inch stilbene and EJ-276 (plastic) organic scintillators were characterised in view of their use in a portable survey meter for fast neutron detection and dose rate measurements 2 in radiation protection. They were coupled with a large area Silicon Photomultiplier (SiPM) array 3 and characterised in terms of neutron/ γ -ray (n/ γ) discrimination capability, count rate linearity, 4 5 count rate saturation effects, temperature stability and neutron detection efficiency. A Pile-Up Rejection (PUR) algorithm was developed and tested to increase their count rate linearity with 6 increasing dose rate. Results show that the stilbene exhibits better performance compared to the 7 8 plastic in terms of: Pulse Shape Discrimination (PSD) with a Figure Of Merit (FOM) of 1.5 vs. 1.1, energy resolution (15% vs 20% at 662 keV) and fast neutron efficiency (12% vs. 9% with an AmBe 9 source). For both scintillators, the dose rate response is linear up to 1 mSv/h with a 137 Cs γ -ray 10 source. With a neutron source, the response is linear, and the FOM is almost constant for dose rates 11 up to 1.5 mSv/h. The PUR reduces the false neutron events by one order of magnitude when the 12 detectors are irradiated with a y-ray source up to 10 mSv/h, and it correctly works when irradiated 13 with an AmBe source at 1.5 mSv/h together with a 137 Cs source up to 60 μ Sv/h. The proposed PUR 14 15 algorithm is promising but needs to be tested with a higher intensity mixed n/γ field. The temperature stability of both detectors was also studied in the temperature range from -10°C to 16 17 +40°C. The light yield increases by around 25% for both detectors when reducing the temperature. 18 The variation of the neutron count rate is around 10% in the same temperature range. This work 19 provides a complete overview of the performance of these detectors and a coherent comparison between the two types of scintillators with the same experimental setup. 20

21 Keywords— Stilbene, EJ-276, SiPM, Fast neutrons, Dose rate.

22 **1. Introduction**

23 Organic scintillators are widely used in particle detection. They are often preferred to inorganic crystals since their fabrication is easier, they have short decay times and lower cost. Moreover, the 24 capability to perform particle discrimination [1] makes them suitable for charged particle detection. 25 As neutron detectors, organic scintillators have been investigated for several applications such as 26 fission neutron detection, fast neutron imaging [2, 3], neutron spectrometry (with unfolding) [4] and 27 as an alternative to ³He counters in non-proliferation application, homeland security and fusion 28 research [5, 6]. In this paper, we consider the stilbene and EJ-276 (plastic) organic scintillators for a 29 30 neutron probe of the B-RAD [7, 8], a radiation survey meter developed by CERN and the 31 Polytechnic of Milan and commercialized by ELSE NUCLEAR [9].

32 The B-RAD is specifically designed for operation in strong magnetic fields. Its detection system 33 consists of a scintillator (currently a LaBr₃ crystal for γ dose rate and γ -spectrometry) coupled with 34 a Silicon Photomultiplier (SiPM) array, instead of a traditional Photomultiplier Tube (PMT), and 35 specifically designed electronics. The purpose of this paper is to study two potential candidates for 36 an additional probe for fast neutron measurements.

Here, we analyse and compare the main features of the stilbene and EJ-276 coupled with a large area SiPM array. The scintillators and the experimental setup are described in section 2. Section 3 addresses the energy calibration and energy resolution, the neutron/ γ -ray (n/ γ) discrimination via the Pulse Shape Discrimination (PSD) technique, the count rate linearity, the neutron detection efficiency, a Pile-Up Rejection (PUR) algorithm developed for application in high-intensity and mixed radiation fields and the temperature dependence. The results are discussed and compared with literature data when available. Finally, the conclusions are summarized in section 6.

44 **2. Detectors and experimental setup**

The stilbene and the EJ-276 are organic scintillators characterised by a similar chemical
composition, low density (close to 1 g cm⁻³) and short scintillation decay times (less than 500 ns).
Table 1 summarizes and compares the main properties of the two scintillators. The quenching factor
for recoil protons is also reported in view of their application as fast neutron detectors.

49

50

Table 1. Comparison of the main properties of the stilbene and EJ-276 scintillators.

	Stilbene	EJ-276	
Manufacturer	Inrad Optic	Scionix	
Type of scintillator	Organic crystal	Organic plastic	
Shape and dimensions	1-inch right cylinder	1-inch right cylinder	
Chemical composition	$C_{14}H_{12}[10]$	4.546×10 ²² H atoms cm ⁻³ 4.906×10 ²² C atoms cm ⁻³ [11]	
Density	1.15 g cm^{-3} [10]	1.096 g cm ⁻³ [11]	
Housing	Yes	No (wrapped in Teflon tape)	
Peak scintillation wavelength	390 nm [10]	425 nm	
Scintillation light yield	14,000 ph MeV ⁻¹ [10]	8,600 ph MeV ⁻¹ [11]	
Decay constants [ns]	4, 35, 332 [12]	13, 35, 270 (γ) 13, 59, 460 (n) [11]	
Recoil proton quenching factor ⁽¹⁾ (proton recoil energy)	0.1 – 0.17 [13] (300 keV – 3 MeV)	0.08 – 0.3 ⁽²⁾ [14] (300 keV – 3 MeV)	

51

52 The scintillators were coupled with an 8×8 SiPM Array J-30035-64P from SensL [15, 16] through optical grease and placed inside a light-tight box. A bias voltage of +28.5 V, corresponding to 4 V 53 54 of overvoltage, was supplied to the SiPM array. The Photon Detection Efficiency (PDE) for the 55 selected overvoltage is around 40% for the stilbene and 43% for the EJ-276 [16]. The summed 56 anode signal was acquired with a suitably designed board [17, 18] and sent to a CAEN DT5720 digitizer (12 bit, 12 Ms/s). The digitizer threshold was set at 100 keVee for the stilbene and 250 57 keVee for the EJ-276. The measurements described in section 3.3 were carried out sending the 58 59 output signal to a Teledyne LeCroy Waverunner 8104 Oscilloscope (1 GHz, 20 Gs/s) with a 50 Ω

⁽¹⁾ The quenching factor is defined as E_{vis}/E_r where E_{vis} is the energy in electron equivalent measured with the scintillator after γ -ray calibration and E_r the energy of the recoil proton emitted by the neutron reaction.

⁽²⁾ Data have been extrapolated from the light yield data relative to 477 keVee.

60 termination (instead of the digitizer) because it allows a higher sampling frequency and more 61 flexibility of parameter setting. Around 30,000 signals were saved for each measurement and 62 analysed offline with a Python code [19].

63 **3. Characterisation of the stilbene and the EJ-276 scintillators**

64 **3.1 Energy calibration and energy resolution**

65 Organic scintillators such as the stilbene and the EJ-276 (see Table 1) are made of low Z materials and have low density. Therefore, their interaction with γ -rays mainly occurs through Compton 66 scattering [1]. Because of the lack of the photoelectric peak, the Compton edge has to be assumed 67 as reference point for the energy calibration. However, one of the major obstacles is to accurately 68 set its position [20-23]. In this work, the energy calibration was performed according to the 69 following method [23]. The spectra of three γ -ray sources (¹³⁷Cs, ⁶⁰Co and ²²Na) were acquired. The 70 theoretical spectra of ¹³⁷Cs, ⁶⁰Co and ²²Na in the stilbene and EJ-276 scintillators were calculated 71 using the FLUKA Monte Carlo code version 4-0.1 [24–26]. A Gaussian Energy Broadening (GEB) 72 was applied to the simulated spectra to take into account the finite energy resolution of the detectors 73 74 assuming a σ proportional to \sqrt{E} , where E is the γ -ray energy [22]. In analogy with [22], the energy 75 resolution is defined as $R_E = FWHM/E$, where FWHM is the Full Width at Half Maximum. The 76 broadened spectrum was normalized to the maximum of the experimental Compton distribution and 77 its x-axis (in energy scale) was linearly scaled in order to match the experimental spectrum (in ADC 78 channels). Both the scaling factor and the R_E values were varied to find the broadened spectrum, 79 which best fitted the experimental one (through least square minimization) in the Compton edge 80 area.

Figure 1 shows the three spectra (experimental, simulated and broadened) of ¹³⁷Cs and ⁶⁰Co for the 81 stilbene in the top, and the plots of the residuals between the experimental and broadened spectra in 82 83 the bottom, in the selected energy range. Figure 2 shows the same results for the EJ-276. The position of the Compton edge of ¹³⁷Cs (447 keV) for the stilbene and EJ-276 was found at 92% and 84 90% of the position of the maximum, respectively (at 74% and 73% of the peak height). For the EJ-85 276, the two Compton edges of ⁶⁰Co were not well distinguishable in the experimental spectrum. 86 87 For this reason, a single fit was performed and the position of the mean energy of the two Compton 88 edges was considered as the calibration point (i.e. 1041 keV) [27]. Figure 3 shows the calibration 89 points and the linear fits for both scintillators. Table 2 reports the linear fitting equations and the 90 corresponding R^2 . The different slopes of the energy calibration equations for the two scintillators 91 are in agreement with the different light yields (see table 1). The small intercept is also expected 92 since it is due to the non-linearity of the light yield at low energies (below 100 keVee) [28].

93 The calculation of the energy calibration and the energy resolution in organic scintillators is not 94 unique as explained in [27]. In this study, the energy resolution for the stilbene and the EJ-276 95 scintillators is 15% and 20% respectively. From the plots of the residuals, the Compton edges are 96 well fitted in the selected energy ranges, and the comparison between the two detectors is reliable. 97 The fitting region was restricted to the Compton edge range since the photon backscattering and the source geometry, which affect the spectrum shape below the Compton edge, were not considered. A 98 99 few percent variation on the energy resolution is acceptable for the scope of this study since the two 100 detectors were studied for use in a dose rate survey meter.



Figure 1. Comparison of the normalized spectrum of a ¹³⁷Cs source (on the left) and of a ⁶⁰Co source (on the right) measured with the stilbene scintillator (blue line), the FLUKA simulated spectrum (black line) and the result of the GEB of the simulated spectrum which best fits the experimental data (red line). On the bottom, the plots of the residuals between the experimental and broadened spectra.

102



Figure 2. Comparison of the normalized spectrum of a ¹³⁷Cs source (on the left) and of a ⁶⁰Co source (on the right) measured with the EJ-276 scintillator (blue line), the FLUKA simulated spectrum (black line) and the result of the GEB of the simulated spectrum which best fits the experimental data (red line). On the bottom, the plots of the residuals between the experimental and broadened spectra.



116 Figure 3. Compton edge positions of ¹³⁷Cs, ⁶⁰Co and ²²Na sources for the stilbene and EJ-276

117 scintillators. The red lines show the linear fits to the data points.

118

119 Table 2. Energy calibration equation for the stilbene and EJ-276, where E is the energy in keV and 120 ch. the corresponding channel. The R^2 of the linear regression is also reported.

	Equation	\mathbb{R}^2
Stilbene	$E (keV) = (0.166 \pm 0.005) \times ch. + (0.952 \pm 27.426)$	0.998
EJ-276	$E (keV) = (0.238 \pm 0.002) \times ch + (42.440 \pm 5.507)$	0.999

121

122 **3.2 Pulse Shape Discrimination (PSD)**

123 The PSD capability is one of the most interesting parameters for a neutron detector. In this work, 124 the PSD was defined according to the charge integration method and calculated for each signal as 125 follows:

126 127

$$PSD = \frac{Q_{long} - Q_{short}}{Q_{long}} \tag{1}$$

128

129 where Q_{short} and Q_{long} are the integrals of the signal (i.e. the charge) calculated over two different 130 integration times (t_{short} and t_{long} respectively). A Figure of Merit (FOM) is defined to quantify the 131 goodness of the PSD according to Eq. 2:

132

$$FOM = \frac{x_n - x_{\gamma}}{FWHM_n + FWHM_{\gamma}}$$
(2)

134

where x_n and x_{γ} are the mean values of the gaussian fitting equations of the neutron and γ -ray distributions respectively, and *FWHM_n* and *FWHM_γ* the corresponding FWHM.

In the present work, the PSD analysis was performed with the DPP-PSD software provided with the CAEN digitizer. The FOM was 1.5 for the stilbene and 1.1 for the EJ-276 scintillator. In the latter case, a higher energy threshold of 350 keVee had to be set to reach a sufficient discrimination

140 (FOM > 1.15), while it was not necessary for the stilbene. The time windows were selected so as to 141 optimize the FOM and set at 300 ns and 3 μ s for both scintillators. The long-time gates selected in 142 this work, if compared to the decay times of both scintillators (see table 1), are due to the high 143 capacitance of the 64 pixels array, which dominates the shape of the output pulse.

Table 3 compares the FOM calculated in this work with the one derived in other studies. The FOM found in this work is within the range of values found in the literature. The deviation from the best FOM can be explained by the different energy threshold and/or by the use of a different

147 photodetector.

148Table 3. Comparison between the FOM calculated in this work and literature data. The size of the149scintillator, the photodetector and the source are specified. The energy cut-off or the energy range

150 selected for the PSD calculation are given in brackets.

	This work	Steinberger et al. [29]	Taggart et al. [30]	Grodzika-Kobylka et al. [31]
Dimensions	Ø1 inch \times 1 inch cylinder	6 mm × 6 mm × 50 mm	6 mm × 6 mm × 6 mm	Ø1 inch \times 1 inch cylinder
Photodetector	64 pixels J-Series SiPM array	SensL C-Series SiPM	J-series SiPM (single pixel)	R6233-100 Hamamatsu PMT
Source	AmBe	²⁵² Cf	AmBe	PuBe
FOM Stilbene	1.5 (100 keVee)	1.17 (100 – 200 keVee)	_	1.74 (100 keVee)
FOM EJ-276	1.1 (250 keVee)		2.39 (500 – 600 keVee)	1.09 (100 keVee)

151

3.3 Count rate linearity

The count rate as a function of the dose rate of both detectors was measured. As expected, the count 153 rate linearly increases with the dose rate up to a limit where the count rate starts to saturate because 154 155 of pile-up events. The pile-up also affects the PSD analysis when the detectors are exposed to a mixed n/γ radiation field. This effect was also studied. It is worth mentioning that the results 156 presented in this work are mainly affected by the electronic read-out system (i.e. the 64 pixels array) 157 and by the selected PSD parameters. However, they can be used for comparison with similar 158 159 detectors, especially because of the lack in the literature of data measured with similar photodetectors (large SiPM array). 160

161 The scintillators were exposed to different photon dose rates from a 137 Cs source at the Calibration 162 Laboratory of the Radiation Protection group at CERN [32], from 1 µSv/h to 15 mSv/h. For both 163 detectors, count rate saturation effects start at around 1 mSv/h, which corresponds to 4×10⁴ and 164 3×10⁴ counts per seconds (cps) in the stilbene and EJ-276, respectively. Figure 4 shows the count 165 rate as a function of the dose rate and the linear fitting equations calculated outside the count rate 166 saturation region (for a better visualization the plot in figure 4 starts from 300 µSv/h).

167 The measurements were repeated with an AmBe neutron source at dose rates between 5 μ Sv/h and 168 1.5 mSv/h, which is the maximum available dose rate at the Calibration Laboratory. The neutron 169 count rate was derived from the PSD analysis. Figure 5 shows the results: the left y-axis is the 170 measured count rate, while the right y-axis represents the count rate corresponding to neutron 171 events only calculated after the PSD. As can be seen, the response is linear over the entire range. 172 The maximum count rate is 1.5×10^3 cps and 1.2×10^3 cps for the stilbene and the plastic 173 respectively, quite lower than the count rate saturation value found with the photon source.



176 Figure 4. Count rate versus dose rate measured with the stilbene (blue triangles) and the EJ-267

- 177 (black dots) scintillators when exposed to photons from a ¹³⁷Cs source before the count rate 178 saturation region. The red line represents the linear fits calculated outside the count rate saturation
- region to the data points. The error bars are smaller than the data points.
- 1/9 region to the data points. The error bars are smaller than the data points



180

Figure 5. On the left y-axis, count rate versus dose rate measured with the stilbene (blue triangles) and the EJ-267 (black dots) scintillator when exposed to an AmBe neutron source. On the right yaxis the neutron count rate calculated after the PSD analysis for the same measurements. The red

184 lines represent the linear fits to the data points. The error bars are smaller than the data points.

The variation of the PSD with increasing pile-up was studied by means of the 2D PSD histogram 185 plot. Each detector was first irradiated with the AmBe source alone at the maximum available dose 186 rate, i.e around 1.5 mSv/h. Afterward, the detectors were exposed simultaneously to the ¹³⁷Cs and 187 AmBe sources. The neutron source activity and the source to detector distance were kept constant to 188 189 maintain the neutron dose rate at 1.5 mSv/h, whereas the photon dose rate was progressively 190 increased by increasing the source activity. Figures 6, 7 and 8 compare the 2D histogram of the PSD vs. energy measured with the stilbene at around 1×10^3 cps, 5×10^3 cps and 40×10^3 cps, 191 respectively. At 5×10^3 cps (figure 7) many pile-up events are detected in the energy range of the 192

- ¹³⁷Cs source, which become dominant at 40×10^3 cps. The same behaviour is observed with the EJ-276 starting from 3.5×10^3 cps. Figures 7 and 8 clearly show that the signals affected by pile-up are spread along the entire PSD range leading to a bad discrimination and to a completely wrong estimation of the dose rate if the energy threshold is not modified.
- 197 The PSD was also monitored by calculating the FOM while keeping constant the energy threshold 198 of each detector. Above 3×10^3 cps the FOM drastically drops compared to the value calculated 199 below 1.5×10^3 cps, where the variation between the maximum and the minimum value is about 200 5%. A correct discrimination can be restored, for example by increasing the energy threshold, as 201 can be seen in figure 7, where the discrimination is still feasible if the energy threshold is set above 202 1.5 MeV.
- As stated before, the count rate range in our detectors is mainly limited by the high capacitance of the employed photodetector. For this reason, it can be extended simply by using a PMT in place of the large area SiPM array. This option was not considered since the B-RAD is designed to work in high magnetic field regions and a PMT cannot be employed. A PUR algorithm was instead tested as discussed in the following section.
- 208
- 209



- 210
- 211

Figure 6. 2D histogram plot of the PSD for the stilbene when irradiated with an AmBe neutron source at 1.5 mSv/h (around 1,000 cps).



215 Figure 7. 2D histogram plot of the PSD for the stilbene when irradiated with an AmBe neutron

source at 1.5 mSv/h and a 137 Cs photon source at 60 μ Sv/h (around 5,000 cps).

217



218

Figure 8. 2D histogram plot of the PSD for the stilbene when irradiated with an AmBe neutron source at 1.5 mSv/h and a 137 Cs photon source at 650 μ Sv/h (around 40,000 cps).

221 **3.4 Pile-up rejection (PUR)**

A PUR algorithm was developed to improve the PSD analysis and increase the count rate linearity of the two detectors. The PUR algorithm was defined similarly to the fractional double-pulse technique described in [5, 6]: a pulse is classified as double if consecutive samples (points of the digitalized pulse) exceed a certain fraction of the pulse height in the tail region. The algorithm was tested for a single combination of the parameters (time windows and threshold) since the results were sufficiently good and proved the possibility to apply the proposed PUR. However, for its application in the B-RAD, these parameters must be optimised.

The time windows were selected by a qualitative analysis, i.e. by observing the neutron and photon standard pulses (see figure 9). For both scintillators, the tail region was selected starting from 1 μ s (indicated by the dashed black vertical line in figure 9) and the total recorded length was set at 4 μ s. The threshold (represented by the horizontal line in figure 9) was selected as the fraction of the γ ray pulse amplitude at 1 µs with respect to the pulse height. The number of samples was chosen to exceed the time interval between the dashed black and dashed-dotted red vertical lines in figure 9. The latter was selected as the point where the neutron pulse crosses the threshold. Following this approach, the threshold was set at 19% and 21% of the pulse height, for the stilbene and the EJ-276, respectively. The sampling frequency was set to 250 MHz for both scintillators.



238

Figure 9. The neutron and γ -ray standard pulses in stilbene. The horizontal line represents the amplitude threshold selected for the PUR, the vertical lines the minimum time intervals that the signal amplitude must exceed, in the tail region, to be discarded by the PUR.

At first, the PUR was applied after irradiating the detectors with a 137 Cs source in the 1 μ Sv/h -242 243 15 mSv/h range. Figures 10 and 11 show the 2D histogram plots acquired with the stilbene before and after applying the PUR at 1 mSv/h. One should note that γ -ray events are expected in the 244 bottom region of the plot, delimited in figures 10 and 11 by the continuous red line. Figure 10 245 246 shows that because of pile-up γ -ray events are erroneously classified as neutron events, while figure 247 11 demonstrates that the PUR correctly removes them. In analogy to [5] the erroneous classification 248 ratio (ER) was calculated as the ratio between the signals detected in the neutron region over the 249 total detected pulses, before and after the PUR. The results are plotted in figures 12 and 13 for the 250 stilbene and the EJ-276, respectively. At low dose rate, the stilbene is characterised by an ER one 251 order of magnitude lower than the EJ-276, while for both detectors the ER decreases by a factor of 10 after the PUR, also at high dose rates. The two curves of the ER after the PUR seem to saturate 252 253 above 1 mSv/h. This is probably due to the spreading of the pile-up events along the entire PSD 254 range, i.e., also in the γ -ray region (below the continuous red line in figures 10 and 11)

A corrected ER factor (ER_{cor}) was defined as the ratio between the signals detected in the neutron region over the signals detected in the region delimited by the γ -ray distribution, confined in figures 10 and 11 by the continuous and dashed red lines. The ER_{cor} is also plotted in figures 12 and 13 (red crosses). The ER_{cor} is linear up to 10 mSv/h for both detectors. Above this value it starts to saturate for the EJ-276 probably because pile-up events are also detected inside the γ -ray region, which cannot be discriminated by the PUR. However, the PUR is still able to well discard false events in the neutron region.



Figure 10. The Q_{long} vs Q_{short} 2D histogram plot of the stilbene irradiated with a ¹³⁷Cs source at 1 mSv/h. The two red lines delimit the y-ray region.



Figure 11. The Q_{long} vs Q_{short} 2D histogram plot of the stilbene irradiated with a ¹³⁷Cs source at 1 mSv/h after applying the PUR algorithm. The two red lines delimit the γ -ray region.



Figure 12. The ER and the ER_{cor} for the stilbene irradiated with a 137 Cs at different dose rate values, before and after applying the PUR algorithm.

273



274

Figure 13. The ER and the ER_{cor} for the EJ-276 irradiated with a ¹³⁷Cs at different dose rate values, before and after applying the PUR algorithm.

The PUR was also tested in a mixed n- γ field from an AmBe source, at around 1.5 mSv/h, and from 277 a 137 Cs source, at 5 μ Sv/h and 60 μ Sv/h. Figures 14 and 15 show the PSD plot before and after the 278 PUR for the stilbene for the two cases. The green lines were calculated from the PSD plot (Qlong vs 279 Q_{short}) acquired with the AmBe source alone as the lines which enclose the region of the neutron 280 signals. As can be seen, the PSD is dominated by the ¹³⁷Cs signals, especially with the 60 μ Sv/h 281 ¹³⁷Cs source, however the AmBe source is still detected below the ¹³⁷Cs distribution. The tails of 282 the neutron and γ -ray distributions can be identified. After applying the PUR, the pile up signals are 283 correctly rejected. In fact, more than 95% of the signals above the neutron region are always 284 discarded. With the 5 μ Sv/h ¹³⁷Cs source, less than 5% of the signals inside the neutron region are 285 erroneously rejected, whereas with 60 µSv/h source the rejection rate is around 20% for the stilbene 286 and 50% for the EJ-276. Nonetheless, neutrons from the AmBe source are still detectable in spite of 287 the low n/γ ratio. 288



Figure 14. The Q_{long} vs Q_{short} 2D histogram plot for the stilbene irradiated with an AmBe source (at 21) 1.5 mSv/h) and a ¹³⁷Cs source (at 5 μ Sv/h on the left and at 60 μ Sv/h on the right), before applying

292 *the PUR algorithm. The green lines delimit the detected neutron signals.*

293

289



294

Figure 15. The Q_{long} vs Q_{short} 2D histogram plot for the stilbene irradiated with an AmBe source (at 1.5 mSv/h) and a ¹³⁷Cs source (at 5 μ Sv/h on the left and at 60 μ Sv/h on the right), after applying the PUR algorithm. The green lines delimit the detected neutron signals.

298 Due to the lack of data measured with similar photodetectors (large SiPM arrays), the study of the 299 linearity and the pile-up events was performed using as a reference the work by Bourne et al. [5]. Bourne et al. coupled a \emptyset 5.08 cm \times 5.08 cm cylindrical stilbene scintillator with a PMT. They 300 demonstrated that their detection system is still able to detect neutrons when exposed to a radiation 301 field of 1000 γ -rays per neutron impinging on the detector (with a maximum count rate of 10⁵ cps), 302 by implementing a more sophisticate PUR technique. The lower limit set in this work (at around 3-4 303 $\times 10^3$ cps) with respect to Bourne et al. is mainly due to the longer time profile of the signal 304 305 produced by the SiPM array as compared to the PMT (the time window in [5] was 300 ns while in 306 this work it was set to $4 \mu s$). Below this limit, the proposed PUR algorithm is promising for 307 application in neutron dose rate meters. Due to the dead time of our acquisition system, mainly

induced by the oscilloscope, at higher count rates the number of detected neutrons (among 30,000signals) was too low to evaluate the efficiency of the PUR.

310 **3.5 Neutron detection efficiency**

311 The neutron detection efficiency was calculated as the average of the intrinsic neutron efficiencies

312 (ϵ) (as defined in [33]) measured in the linear range between 5 μ Sv/h and 1.5 mSv/h (see figure 4). 313 The obtained values are 12% for the stilbene and 9% for the plastic, which correspond to 2.0 counts

314 per nSv and 1.1 counts per nSv, respectively.

315 It should be noted that the calculated efficiencies are hugely affected by the threshold because the 316 pulse height spectrum in organic scintillators has a continuum shape and hence, all the counts below 317 the threshold are not counted. If we apply the same energy threshold to both scintillators, e.g. 318 250 keVee, which is the minimum threshold for the EJ-276, the calculated efficiency of the two 319 detectors is comparable: 1.1 counts per nSv for the EJ-276 and 1.3 counts per nSv for the stilbene.

320 3.7 Temperature dependence

Both detectors, inside the light-tight box, were placed in a climate chamber in which the temperature was varied from -10° C to $+40^{\circ}$ C at a rate of 10° C h⁻¹. Measurements were performed in 10° C steps with a ¹³⁷Cs and an AmBe source independently. A thermocouple was placed close to the SiPM because of the different thermal inertia of the detector inside the box and the climate chamber.

The temperature dependence of the detectors depends on both the variation of the scintillators' light yield and the breakdown voltage. The combined effect was evaluated by measuring the variation of the Compton edge position (in the ¹³⁷Cs spectrum), calculated as described in section 3.1.

The contribution of the SiPM was calculated as in [34], considering that the breakdown voltage increases with temperature at a rate of 21.5 mV $^{\circ}C^{-1}$ and the gain linearly decreases accordingly [16]. The contribution of the SiPM to the variation of the Compton edge position was assumed proportional to the gain variation. Finally, the contribution of each scintillator was obtained from the comparison of the measured and calculated variation of the Compton edge position. Please note that the former is ascribed to the SiPM + scintillator system, while the latter is attributed to the SiPM only.

Figure 16 compares the variation of the measured and calculated Compton edge position for both detectors. The x-axis is the temperature measured by the probe close to the SiPM. For the stilbene, below ~20°C the measured Compton edge position linearly decreases with temperature with a slope of -18.7 \pm 0.4 channel °C⁻¹ versus a variation of -12.7 channel °C⁻¹ due to the SiPM only, i.e. the light yield of the stilbene decreases inducing a faster variation of the Compton edge position. Above 20°C the stilbene light output increases and partially compensates the SiPM effect.

Similarly, for the EJ-276, the Compton edge position decreases below 20°C (see figure 16). 342 Nonetheless, the difference between the measured and calculated curves is smaller: -343 10.5 ± 0.6 channel°C⁻¹ versus -7.6 channel °C⁻¹ respectively. Above 20°C, the temperature 344 dependence is mainly attributed to the SiPM. This is partially in agreement with [35], where the 345 variation of the Compton edge position measured by a PMT with a plastic scintillator is almost flat 346 347 in the temperature range - 30° C - 20° C. However, in [35] the exact composition of the plastic 348 scintillator is not specified. The total percent variation of the Compton edge position is around 25% for both detectors. 349

350 The variation of the PSD capability was also investigated with the AmBe source by observing both 351 the variation of the PSD distribution, and of the FOM. Figure 17 shows the PSD distribution at 352 different temperatures for the stilbene and the EJ-276, and figure 18 compares the variation of their 353 FOM. In figure 17, the left and right Gaussian distributions (on both plots) are associated to γ -rays 354 and neutrons respectively.

In both detectors, up to 20°C, the shape of the PSD distribution remains unchanged but, as 355 expected, the number of counts decreases with temperature because of the increasing breakdown 356 voltage. For the stilbene, at 30°C and 40°C, the neutron distribution shifts to the right and the right 357 tail of the γ -ray distribution increases. The intensity of the fast decay appears to decrease with 358 359 temperature in favour of the slow component (according to equation 1). This effect is generally 360 possible since the temperature can affect different scintillation mechanisms in different ways. For the EJ-276, the PSD distribution above 20°C shows an opposite trend as compared to the stilbene. 361 The peaks of the two Gaussians slightly move to the left and in addition, some signals are detected 362 in the region between the two distributions. At higher temperatures, the lower light yield of the EJ-363 276 combined with the higher breakdown voltage of the SiPM has a large impact on the PSD 364 analysis. A greater number of signals is in fact detected just above the digitizer threshold and it is 365 well known that for low amplitude signals the discrimination is not clear. To overcome this limit 366 and for a better comprehension, a higher bias voltage can be set in the EJ-276 detector. For both 367 scintillators the temperature increase induces a broadening of the neutron and γ -ray distributions in 368 369 the PSD plot resulting in the decrease of the FOM (see figure 18).

370 To our knowledge, the variation with temperature of the light yield in organic scintillators has not been quantified yet. Baker et al. [36] reported a weak decreasing trend of the stilbene light yield 371 with increasing temperature between $-90 - 70^{\circ}$ C, partially in agreement with this work. In [36] the 372 373 stilbene was irradiated both with a γ -ray source and a neutron source. A study was performed with an anthracene crystal (comparable to the stilbene) in [37] only for a 6°C temperature range (22 – 374 375 28° C) and only for γ -ray detection. In this study, a slightly linear decreases of the light output was 376 measured of less than 10%. In [35], the temperature dependence of the stilbene is deeply 377 investigated between -30°C and 60°C and the pulse shapes at different temperature are also 378 reported. The pulse shape seems to be constant in the abovementioned range. However, the 379 discussion is lacking and a more accurate study is necessary (e.g. fitting of the average signals). On 380 the other hand, the light yield of plastic scintillators was considered temperature independent [38] 381 until recent studies have demonstrated some dependence [39,40]. In [40], a small or almost null 382 temperature dependence was measured for some plastic scintillators according to the substrate 383 material.

Figure 19 shows the variation of the neutron counts with temperature (15 minutes of acquisition time). The variation of the neutron counts with temperature is around 10% for both detectors, which is still acceptable for applications in radiation protection. The measured neutron counts with the stilbene has a minimum at 10°C, while the EJ-276 is almost flat up to 20°C and shows a decreasing trend above that temperature. This variability makes the temperature correction not an easy task. Further investigation such as measurements with a PMT and a pulse shape analysis should be performed.



392 Figure 16. Variation with temperature of the Compton edge position of a ¹³⁷Cs source for the

393 stilbene and the EJ-276. The x-axis is the temperature measured by a thermocouple close to the 394 SiPM array while the temperature chamber was varied between $-10^{\circ}C$ and $+40^{\circ}C$. The error bars

554 Sil in array while the temperature chamber was varied between -10 C and 140 C. The error bar

395 *are smaller than the data points.*



396

397 Figure 17. Variation of the PSD distribution measured with an AmBe source with the stilbene (left)

and the EJ-276 (right) while the temperature varied between $-10^{\circ}C$ and $+40^{\circ}C$.



400 Figure 18. Variation with temperature of the FOM measured with an AmBe source for the stilbene

401 and the EJ-276. The x-axis is the temperature measured by a thermocouple close to the SiPM array, 402 while the temperature is the shareheavening distance 10% cm d $\pm 40\%$ c

402 while the temperature in the chamber varied between $-10^{\circ}C$ and $+40^{\circ}C$.



403

404 Figure 19. Variation with temperature of the measured neutron counts (15 minutes of acquisition
405 time) with an AmBe source for the stilbene and the EJ-276. The x-axis is the temperature measured
406 by a thermocouple close to the SiPM array while the temperature in the chamber varied between 407 10°C and +40°C.

408 **6.** Conclusions

409 In this paper, we evaluate and compare the performance of a stilbene and an EJ-276 plastic 410 scintillators coupled with an 8 × 8 SiPM array J-30035-64P from SensL, in view of their potential 411 application in the B-RAD portable survey meter for fast neutron detection. The reported results show that the energy resolution is 15% for the stilbene and 20% for the plastic with a ¹³⁷Cs source. 412 Moreover, the stilbene shows a better PSD capability compared to the EJ-276: a FOM of 1.5 is 413 414 achieved with the minimum threshold of 100 keVee, while a threshold of 350 keVee had to be set to reach a sufficient discrimination with the plastic (FOM = 1.15). The calculated neutron detection 415 efficiency is 12% for the stilbene and 9% for the plastic. The response of both detectors starts to 416 deviate from linearity at 1 mSv/h photon dose rate, whereas the PSD is stable up to 1.5 mSv/h 417

418 neutron dose rate. When exposed to a mixed $n-\gamma$ radiation field, the FOM of both scintillators drastically drops above 3 - 4×10^3 cps. Below this limit, a Pile-Up Rejection algorithm can reduce 419 the detection of false neutron events by one order of magnitude without affecting the detection 420 421 efficiency. The two scintillators show similar performance in terms of temperature variation in the 422 range -10°C to +40°C. The variation of the measured Compton edge position is around 25%, while the variation of the neutron count rate is around 10%. A deeper investigation of the PSD 423 424 distribution as a function of the temperature shows a different behaviour of the two scintillators 425 above 20°C. which has to be confirmed by further measurements.

426 **References**

- 427 [1] G.F.Knoll, Radiation Detection and Measurements, 4th ed., 2010.
- 428 [2] U. Bravar, P.J. Bruillard, E.O. Flückiger, J.R. Macri, M.L. McConnell, M.R. Moser, J.M. 429 Ryan, R.S. Woolf, Design and testing of a position-sensitive plastic scintillator detector for 430 neutron IEEE Trans. Nucl. Sci. 53 3894-3903. fast imaging, (2006)https://doi.org/10.1109/TNS.2006.886046. 431
- 432 [3] E. Brubaker, J. Steele, Neutron imaging using the anisotropic response of crystalline organic
 433 scintillators, IEEE Nucl. Sci. Symp. Conf. Rec. (2010) 1647–1652.
 434 https://doi.org/10.1109/NSSMIC.2010.5874055.
- Y.A. Kaschuck, B. Esposito, L.A. Trykov, V.P. Semenov, Fast neutron spectrometry with
 organic scintillators applied to magnetic fusion experiments, Nucl. Instruments Methods
 Phys. Res. Sect. A Accel. Spectrometers, Detect. Assoc. Equip. 476 (2002) 511–515.
 https://doi.org/10.1016/S0168-9002(01)01499-1.
- M.M. Bourne, S.D. Clarke, N. Adamowicz, S.A. Pozzi, N. Zaitseva, L. Carman, Neutron detection in a high-gamma field using solution-grown stilbene, Nucl. Instruments Methods Phys. Res. Sect. A Accel. Spectrometers, Detect. Assoc. Equip. 806 (2016) 348–355.
 https://doi.org/10.1016/j.nima.2015.10.025.
- M.M. Bourne, S.D. Clarke, M. Paff, A. DiFulvio, M. Norsworthy, S.A. Pozzi, Digital pile-up rejection for plutonium experiments with solution-grown stilbene, Nucl. Instruments Methods Phys. Res. Sect. A Accel. Spectrometers, Detect. Assoc. Equip. 842 (2017) 1–6. https://doi.org/10.1016/j.nima.2016.10.023.
- 447 [7] A. Fazzi, M. Silari, Portable radiation detection device for operation in intense magnetic field.
 448 CERN/Polytechnic of Milan, joint patent. Patent Grant number 9977134 (13 July 2017).
- [8] D. Celeste, A. Curioni, A. Fazzi, M. Silari, V. Varoli, B-RAD: A radiation survey meter for
 operation in intense magnetic fields, J. Instrum. 14 (2019). https://doi.org/10.1088/17480221/14/05/T05007.
- 452 [9] Online: Products for nuclear safety, radiation detection and measurement ELSE Nuclear,
 453 http://www.elsenuclear.com/it/ (accessed May 8, 2021).
- 454 [10] Online: ScintinelTM Stilbene. https://www.inradoptics.com/scintinel-stilbene (accessed May 8, 2021).
- 456 [11] Online: https://eljentechnology.com/products/plastic-scintillators/ej-276 (accessed May 7, 2021).

- [12] N. Zaitseva, A. Glenn, L. Carman, H. Paul Martinez, R. Hatarik, H. Klapper, S. Payne,
 Scintillation properties of solution-grown trans-stilbene single crystals, Nucl. Instruments
 Methods Phys. Res. Sect. A Accel. Spectrometers, Detect. Assoc. Equip. 789 (2015) 8–15.
 https://doi.org/10.1016/j.nima.2015.03.090.
- 462 [13] Y. Shimizu, M. Minowa, H. Sekiya, Y. Inoue, Directional scintillation detector for the
 463 detection of the wind of WIMPs, Nucl. Instruments Methods Phys. Res. Sect. A Accel.
 464 Spectrometers, Detect. Assoc. Equip. 496 (2003) 347–352. https://doi.org/10.1016/S0168465 9002(02)01661-3.
- 466 [14] T.A. Laplace, B.L. Goldblum, J.E. Bevins, D.L. Bleuel, E. Bourret, J.A. Brown, E.J.
 467 Callaghan, J.S. Carlson, P.L. Feng, G. Gabella, K.P. Harrig, J.J. Manfredi, C. Moore, F.
 468 Moretti, M. Shinner, A. Sweet, Z.W. Sweger, Comparative scintillation performance of EJ469 309, EJ-276, and a novel organic glass, J. Instrum. 15 (2020). https://doi.org/10.1088/1748470 0221/15/11/P11020.
- 471 [15] Online: Products ON Semiconductor. https://www.onsemi.com/products/sensors/silicon 472 photomultipliers-sipm/j-series-sipm-array (accessed May 8, 2021).
- 473 [16] Online: DS-MicroJseries.pdf (sensl.com) (accessed May 7, 2021).
- 474 [17] D. Nesrine, Development of neutron detectors for use in radiation protection, University of475 Paris-Saclay, 2019.
- [18] N. Dinar, D. Celeste, M. Silari, V. Varoli, A. Fazzi, Pulse shape discrimination of CLYC
 scintillator coupled with a large SiPM array, Nucl. Instruments Methods Phys. Res. Sect. A
 Accel. Spectrometers, Detect. Assoc. Equip. 935 (2019) 35–39.
 https://doi.org/10.1016/j.nima.2019.04.099.
- 480 [19] G. Van Rossum, F.L. Drake, Python 3 Reference Manual, CreateSpace, Scotts Valley, CA,
 481 2009.
- 482 [20] H.H. Knox, T.G. Miller, A technique for determining bias settings for organic scintillators,
 483 Nucl. Instruments Methods. 101 (1972) 519–525. https://doi.org/10.1016/0029484 554X(72)90040-7.
- 485 [21] C.H. Lee, J. Son, T.H. Kim, Y.K. Kim, Characteristics of plastic scintillators fabricated by a
 486 polymerization reaction, Nucl. Eng. Technol. 49 (2017) 592–597.
 487 https://doi.org/10.1016/j.net.2016.10.001.
- 488 [22] S. Ashrafi, M. Ghahremani Gol, Energy calibration of thin plastic scintillators using Compton
 489 scattered γrays, Nucl. Instruments Methods Phys. Res. Sect. A Accel. Spectrometers, Detect.
 490 Assoc. Equip. 642 (2011) 70–74. https://doi.org/10.1016/j.nima.2011.04.003.
- 491 [23] G. Dietze, H. Klein, Gamma-calibration of NE 213 scintillation counters, Nucl. Instruments
 492 Methods. 193 (1982) 549–556. https://doi.org/10.1016/0029-554X(82)90249-X.
- 493 [24] Online: https://fluka.cern.
- 494 [25] T.T. Böhlen, F. Cerutti, M.P.W. Chin, A. Fassò, A. Ferrari, P.G. Ortega, A. Mairani, P.R. 495 Sala, G. Smirnov, V. Vlachoudis, The FLUKA Code: Developments and challenges for high 496 and medical applications, 211-214. energy Nucl. Data Sheets. 120 (2014)497 https://doi.org/10.1016/j.nds.2014.07.049.

- 498 [26] G. Battistoni, T. Boehlen, F. Cerutti, P.W. Chin, L.S. Esposito, A. Fassò, A. Ferrari, A.
 499 Lechner, A. Empl, A. Mairani, A. Mereghetti, P.G. Ortega, J. Ranft, S. Roesler, P.R. Sala, V.
 500 Vlachoudis, G. Smirnov, Overview of the FLUKA code, Ann. Nucl. Energy. 82 (2015) 10–
 501 18. https://doi.org/10.1016/j.anucene.2014.11.007.
- 502 [27] E. V. Pagano, M.B. Chatterjee, E. De Filippo, P. Russotto, L. Auditore, G. Cardella, E.
 503 Geraci, B. Gnoffo, C. Guazzoni, G. Lanzalone, S. De Luca, C. Maiolino, N.S. Martorana, A.
 504 Pagano, M. Papa, T. Parsani, S. Pirrone, G. Politi, F. Porto, L. Quattrocchi, F. Rizzo, A.
 505 Trifirò, M. Trimarchi, Pulse shape discrimination of plastic scintillator EJ 299-33 with
 506 radioactive sources, Nucl. Instruments Methods Phys. Res. Sect. A Accel. Spectrometers,
 507 Detect. Assoc. Equip. 889 (2018) 83–88. https://doi.org/10.1016/j.nima.2018.02.010.
- 508 [28] G. Dietze, Energy calibration of NE-213 scintillation counters by x-rays, IEEE Trans. Nucl.
 509 Sci. 26 (1979) 398–402. https://doi.org/10.1109/TNS.1979.4329665.
- [29] W. Steinberger, M.L. Ruch, S.A. Pozzi, Performance of stilbene bars coupled to silicon
 photomultipliers using Different Reflectors, 2017 IEEE Nucl. Sci. Symp. Med. Imaging Conf.
 NSS/MIC 2017 Conf. Proc. (2018) 12–14. https://doi.org/10.1109/NSSMIC.2017.8532739.
- 513 [30] M.P. Taggart, P.J. Sellin, Comparison of the pulse shape discrimination performance of
 514 plastic scintillators coupled to a SiPM, Nucl. Instruments Methods Phys. Res. Sect. A Accel.
 515 Spectrometers, Detect. Assoc. Equip. 908 (2018) 148–154.
 516 https://doi.org/10.1016/j.nima.2018.08.054.
- 517 [31] M. Grodzicka-Kobylka, T. Szczesniak, M. Moszyński, K. Brylew, L. Swiderski, J.J. Valiente518 Dobón, P. Schotanus, K. Grodzicki, H. Trzaskowska, Fast neutron and gamma ray pulse
 519 shape discrimination in EJ-276 and EJ-276G plastic scintillators, J. Instrum. 15 (2020).
 520 https://doi.org/10.1088/1748-0221/15/03/P03030.
- 521 [32] F. Pozzi, R.G. Alia, M. Brugger, P. Carbonez, S. Danzeca, B. Gkotse, M.R. Jaekel, F. Ravotti,
 522 M. Silari, M. Tali, CERN irradiation facilities, Radiat. Prot. Dosimetry. 180 (2018) 120–124.
 523 https://doi.org/10.1093/RPD/NCX187.
- 524 [33] D.S. McGregor, J. Kenneth Shultis, Reporting detection efficiency for semiconductor neutron
 525 detectors: A need for a standard, Nucl. Instruments Methods Phys. Res. Sect. A Accel.
 526 Spectrometers, Detect. Assoc. Equip. 632 (2011) 167–174.
 527 https://doi.org/10.1016/j.nima.2010.12.084.
- [34] N. Dinar, D. Celeste, P. Puzo, M. Silari, Characterization of CLYC scintillator coupled with
 photomultipliers and a large SiPM array, Radiat. Prot. Dosimetry. 180 (2018) 399–402.
 https://doi.org/10.1093/RPD/NCX203.
- [35] E. Mullin, K. Mesick, S. Nowicki, D. Coupland, N. Zaitseva, K. Ianakiev, Performance
 characterization of organic pulse shape discrimination scintillators at Los Alamos National
 Laboratory (2018) 30517.
- J.H. Baker, N.Z. Galunov, O.A. Tarasenko, Variation of scintillation light yield of organic
 crystalline solids for different temperatures, IEEE Trans. Nucl. Sci. 55 (2008) 2736–2738.
 https://doi.org/10.1109/TNS.2008.2002146.

- 537 [37] P. Schuster, E. Brubaker, Investigating the anisotropic scintillation response in anthracene
 538 through neutron, gamma-ray, and muon measurements, IEEE Trans. Nucl. Sci. 63 (2016)
 539 1942–1954. https://doi.org/10.1109/TNS.2016.2542589.
- 540 [38] A.S. Beddart, T.R. Mackie, F.H. Attix, Water-equivalent plastic scintillation detectors for 541 high-energy beam dosimetry: I. Physical characteristics and theoretical considerations, 1992.
- 542 [39] L. Wootton, S. Beddar, Temperature dependence of BCF plastic scintillation detectors, Phys.
 543 Med. Biol. 58 (2013) 2955–2967. https://doi.org/10.1088/0031-9155/58/9/2955.
- [40] L. Peralta, Temperature dependence of plastic scintillators, Nucl. Instruments Methods Phys.
 Res. Sect. A Accel. Spectrometers, Detect. Assoc. Equip. 883 (2018) 20–23. https://doi.org/10.1016/j.nima.2017.11.041.