



# **Microfluidic Chip for the Electrochemical Detection of MicroRNAs: Methylene Blue Increasing the Specificity of the Biosensor**

Claire Poujouly, Jérémy Le Gall, Martina Freisa, Djamila Kechkeche, David Bouville, Jihed Khemir, Pedro Gonzalez-Losada, Jean Gamby

## **► To cite this version:**

Claire Poujouly, Jérémy Le Gall, Martina Freisa, Djamila Kechkeche, David Bouville, et al.. Microfluidic Chip for the Electrochemical Detection of MicroRNAs: Methylene Blue Increasing the Specificity of the Biosensor. *Frontiers in Chemistry*, 2022, 10, pp.868909. <10.3389/fchem.2022.868909>. <hal-03623585>

**HAL Id: hal-03623585**

**<https://hal.science/hal-03623585v1>**

Submitted on 29 Mar 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



HAL Authorization



# Microfluidic Chip for the Electrochemical Detection of MicroRNAs: Methylene Blue Increasing the Specificity of the Biosensor

Claire Poujouly, Jérémy Le Gall, Martina Freisa, Djamila Kechkeche, David Bouville, Jihed Khemir, Pedro Gonzalez-Losada and Jean Gamby\*

Université Paris-Saclay, CNRS, Centre de Nanosciences et de Nanotechnologies, Palaiseau, France

## OPEN ACCESS

### Edited by:

Ismael Díez-Pérez,  
King's College London,  
United Kingdom

### Reviewed by:

Hong Zhou,  
Qingdao University of Science and  
Technology, China  
Seung Woo Lee,  
Georgia Institute of Technology,  
United States  
Emily Maria Kerr,  
Deakin University, Australia

### \*Correspondence:

Jean Gamby  
jean.gamby@c2n.upsaclay.fr

### Specialty section:

This article was submitted to  
Electrochemistry,  
a section of the journal  
Frontiers in Chemistry

**Received:** 03 February 2022

**Accepted:** 07 March 2022

**Published:** 29 March 2022

### Citation:

Poujouly C, Le Gall J, Freisa M,  
Kechkeche D, Bouville D, Khemir J,  
Gonzalez-Losada P and Gamby J  
(2022) Microfluidic Chip for the  
Electrochemical Detection of  
MicroRNAs: Methylene Blue  
Increasing the Specificity of  
the Biosensor.  
Front. Chem. 10:868909.  
doi: 10.3389/fchem.2022.868909

MicroRNAs (miRNAs) are biomarkers involved in biological processes that are released by cells and found in biological fluids such as blood. The development of nucleic acid-based biosensors has significantly increased in the past 10 years because the detection of such nucleic acids can easily be applied in the field of early diagnosis. These biosensors need to be sensitive, specific, and fast in order to be effective. This work introduces a newly-built electrochemical biosensor that enables a fast detection in 30 min and, as a result of its integration in microfluidics, presents a limit of detection as low as 1 aM. The literature concerning the specificity of electrochemical biosensors includes several studies that report one base-mismatch, with the base-mismatch located in the middle of the strand. We report an electrochemical nucleic acid biosensor integrated into a microfluidic chip, allowing for a one-base-mismatch specificity independently from the location of the mismatch in the strand. This specificity was improved using a solution of methylene blue, making it possible to discriminate a partial hybridization from a complete and complementary hybridization.

**Keywords:** microfluidics, miRNA electrochemical detection, methylene blue (MB), microelectrodes, surface functionalization

## INTRODUCTION

MicroRNAs are non-coding RNAs of 21 to 25 bases acting as regulators of protein translation. Since many diseases are caused by the misregulated activity of proteins, researchers actively studied microRNAs as biomarkers for early diagnosis of different types of cancer (Calin and Croce, 2006), as well as heart diseases (Ji et al., 2009; Adachi et al., 2010), and muscle damage (Siracusa et al., 2016; Siracusa et al., 2018). Standard methods for detecting microRNAs include northern blotting, microarrays, and real-time PCR (RT-PCR) (Cissell and Deo, 2009). However, these methods are time consuming and their limits of detection can be improved with electrochemical biosensors. Electrochemistry is an attractive technique in terms of sensitivity and ease of use (Drummond et al., 2003; Rosario and Mutharasan, 2014; Minaei et al., 2015). Electrochemical biosensors are based on converting a biological binding event to an electronic signal (Gooding, 2002; Grieshaber et al., 2008). In order to simplify the biosensor and reduce its number of detection steps, the direct detection of nucleic acids is generally favored. This direct detection is mostly based on the detection of the hybridization of two complementary strands of nucleic acids (Cissell and Deo, 2009). This hybridization induces a change in the electronic signal measured on the

electrodes of the electrolytic cell. The change of signal can be generated by either direct charge transfer (Ariksoysal et al., 2005; Abbaspour and Noori, 2008), or by the phenomenon of long-range electron transfer (LET) (Kelley et al., 1999), playing an important role in the specificity of the biosensor (Boon et al., 2000). LET is based on the hypothesis that DNA-mediated electron transfer from an electrode to a redox mediator intercalated into the DNA duplex allows electrocatalysis of a redox tracer reduction in solution on DNA layers. For instance, Barton et al. (Boon et al., 2003a) described the electrocatalytic reduction of ferricyanide mediated by methylene blue (MB) using linear sweep voltammetry in a 3-electrode setup including a rotating gold electrode. The DNA duplexes (15-basis DNA with thiol linker) self-assembled into a dense monolayer in an upright position, blocking the electrochemical reduction. The authors reported that micromolar concentrations of the redox-active DNA intercalator are sufficient to enhance catalytic currents. In brief, the chemical reaction between MB and  $\text{FeCN}_6^{3-}$  happens at the film/solution interface. The system needs a few seconds to reach a steady-state current density for  $\text{FeCN}_6^{3-}$  reduction, which is a function of MB bulk concentration in the solution. For MB concentrations higher than 15  $\mu\text{M}$ , they assumed that the reaction is supposed to be purely controlled by diffusion convection and its rate is proportional to the MB concentration in solution. To conclude, DNA films seem to be saturated with MB above 2  $\mu\text{M}$  where binding of MB seems to be reversible with one intercalation site per 15-basis duplex. The specificity of the detection is one of the three most important key factors, together with the limit of detection and the rapidity, for nucleic acid biosensors to perform a reliable diagnosis (Horny et al., 2016; Nassi et al., 2016). Moreover, numerous electrode materials and electrochemical techniques can be used for sensitive measurements [cyclic voltammetry (Shoae et al., 2018), differential pulse voltammetry (Butterworth et al., 2019), square wave voltammetry (Tsaloglou et al., 2018), and electrochemical impedance spectroscopy (Quan Li et al., 2017)].

The integration of biosensors into microfluidics presents many advantages, such as reducing the number of samples needed as well as decreasing the time of experiments in parallel (Squires and Quake, 2005). In addition, microfluidic electrochemical nucleic acid biosensors enable a low detection threshold due to the diffusion layer's reduction under forced convection, bringing more nucleic acids to the sensor surface (Ferguson et al., 2009).

In order to resolve the above issues, we developed a microfluidic chip for the electrochemical detection of nucleic acids based on the LET using the system's catalytic nature to improve the sensor's specificity. We report an experimental study about enhancing the biosensor specificity by using methylene blue in a redox solution. We studied the specificity by evaluating the hybridization rate for target sequences with single-base mismatches in the middle or at the extremity of the nucleic acid strands. The intercalation of MB in double-stranded nucleic acids then enabled us to discriminate a partial hybridization from a complementary hybridization.

**TABLE 1 |** Nucleic acids sequences.

Name	Sequence
P	5'—Thiol modifier C6—CAAACACCATTGTCACTGC—3'
T	5'—GCAGTGTGACAATGGTGTGG—3'
TRNA	5'—GCAGUGUGACAAUGGUGUUUG—3'
T—1 M middle	5'—GCAGTGTGACCATGGTGTGG—3'
T—1 M beginning	5'—GCAGTGTGACAATGGTGTGG—3'
T—1 M end	5'—ACAGTGTGACAATGGTGTGG—3'
Nc T	5'—TTTGGTCCCTTCAACCAGCTG—3'


## EXPERIMENTAL

### Chemical Products and Nucleic Acids

Sodium chloride, potassium hexacyanoferrate (II) trihydrate, potassium hexacyanoferrate (III), and methylene blue (MB) were purchased from Sigma Aldrich. All aqueous solutions were prepared using deionized water with a resistivity of 18 M $\Omega$ . All DNA and RNA sequences were purchased from Eurogentec with reverse phase HPLC purification in a dried format. A 21-base thiol-labeled DNA probe (P) mimics the complementary sequence of the micro-ribonucleic acid 122 (miR 122). A DNA target (T) and an RNA target ( $T_{\text{RNA}}$ ) mimic the miR 122 sequence; these two targets are complementary to the DNA probe. A DNA target with one mismatch in the middle of the strand (T—1 M middle). A DNA target with one mismatch at the beginning of the strand (T—1 M beginning). A DNA target with one mismatch at the end of the strand (T—1 M end). A non-complementary DNA target (Nc T) mimics the miR 133-3p, involves muscle damage and serves as a negative test. **Table 1** presents these sequences. The nucleic acids were first diluted in deionized water to obtain a  $10^{-4}$  M concentration, and then diluted in a 0.5 M NaCl solution to get the wanted concentration.

### Microfluidic Chip Fabrication

The microfluidic device is formed by a glass substrate containing a pair of gold microelectrodes, and on top, a polydimethylsiloxane (PDMS) cover containing the microfluidic channel (300  $\mu\text{m}$  wide, 60  $\mu\text{m}$  high, 2 cm long). The pair of gold microelectrodes is composed of a working microelectrode (300  $\mu\text{m}$  wide, 30  $\mu\text{m}$  long) and a counter electrode (300  $\mu\text{m}$  wide, 2 mm long) (**Figure 1A**). The microelectrodes were patterned onto a 4-inch glass wafer using a lift-off process. In order to realize the lift-off process, an initial photolithography step was realized using a negative photoresist (AZ nLOF 2020, MicroChemicals). Then we sputtered a 20 nm TiW layer on the whole wafer to improve the adhesion of the metal to the glass substrate. We completed the process by sputtering another 200 nm layer of Au. The process was completed by removing the photoresist in an acetone ultrasound bath for 15 mins, immediately followed by an isopropanol rinse. Once the electrodes were patterned, the wafer was diced into several chips. The microchannel structures were fabricated by molding PDMS on a SU-8 master mold already set on a silicon 4-inch wafer used as a substrate. For the mold fabrication the first layer of SU-8 2002



**FIGURE 1 | (A)** Biosensor geometry, designed using KLayout software, with the detection area dimensions—working electrode (WE) (width 300  $\mu\text{m}$ , length 30  $\mu\text{m}$ ), counter electrode (CE) (width 300  $\mu\text{m}$ , length 2 mm). **(B)** Photography of the final microfluidic device composed of two pairs of gold electrodes (WE, CE) for repeatability of measurements.

(MicroChemicals) of 2  $\mu\text{m}$  was spin-coated on the wafer and baked to serve as an adhesion layer. Then a 60  $\mu\text{m}$  layer of SU-8 2050 (MicroChemicals) was spin-coated, and microchannels were patterned by photolithography. Finally, we prepared a mixture of PDMS (RTV 615, Neyco), mixing 10 parts of silicone elastomer and one part of the curing agent. This mixture was then poured on the mold, placed under a vacuum for 2 h, and cured in an oven at 60°C for at least 4 h. After the PDMS was polymerized, it was peeled off from the mold and cut out. The inlet and outlet were punched using a biopsy needle (0.5 mm, Elveflow). At the end of the process, the diced electrode chips and the PDMS microchannels were bonded using oxygen plasma (**Figure 1B**). Once the electrode glass substrate and the PDMS microchannels are bonded, the device must be used within a week. If it is not used in this time period the effect of the oxygen plasma treatment on the PDMS microchannels disappears and the channels become hydrophobic. Considering the durability of the electrochemical sensor itself, once the first measurement is executed on the chip, all experiments must be carried out during the day. After that, the gold electrode surfaces start oxidizing.

### Experimental Protocol of Probe Immobilization and Target Hybridization

A syringe pump (neMESYS, Cetoni GmbH, Germany) and syringes of 1 to 5 ml are used to load the solutions into the microfluidic channel. The syringes are connected to the microfluidic channel via a Tygon tubing (1/16" OD  $\times$  0.51 mm ID, Elveflow) and stainless steel couplers (23 G 0.025" OD  $\times$  0.013" ID, Elveflow) inserted into the inlet and outlet previously punched. The DNA probe immobilization was performed by loading the DNA probe solution, diluted to  $10^{-7}$  M, into the microfluidic channel for 2 h. The thiol-modified DNA probe sequence spontaneously bound to the gold microelectrodes during the immobilization, forming the self-assembled monolayer (SAM) of DNA probes. Between the probe immobilization and the target recognition step, a 0.5 M NaCl solution was loaded into the microchannel for 30 min to test the stability of the SAM. The target hybridization was realized by loading into the microfluidic chip the nucleic acid target solution, containing the sequence of interest, for 30 min at a concentration

varying from  $10^{-18}$  M to  $10^{-6}$  M. Electrochemical measurements were performed before and after this target hybridization step in order to determine the hybridization rate for the probe by comparing the measured currents.


### Electrochemical Measurements

An electrochemical workstation (Biologic SP-300, France) was used to record the electrochemical signals of the two-electrode system integrated into the microfluidic chip. At each step of electrochemical measurements (bare gold, ssDNA immobilization, dsDNA hybridization), the measured current can be attributed to the WE current response even if both the WE and the CE are functionalized with the probe sequence (Horny et al., 2016). EC-Lab software was used for the acquisition, processing, and display of all the electrochemical measurements. The cyclic voltammetry measurements were recorded from  $-0.2$  to  $0.2$  V with a scan rate of  $10 \text{ mV.s}^{-1}$  in a 3 mM equimolar  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  redox couple added to a 0.5 M NaCl solution. The chronoamperometry measurements were recorded at  $-0.2$  V for 150 s in a 3 mM equimolar  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  in a 0.5 M NaCl solution adding 5  $\mu\text{M}$  of methylene blue solution. A micromolar concentration of MB is necessary to be electrochemically detected. Both electrochemical measurements were carried out at a flow rate of 0.5  $\mu\text{L/s}$ . The flow rate was programmed using the syringe pump connected to the input of the microfluidic chip.

### RESULTS AND DISCUSSION

The electrochemical system used consists of a two-electrode setup with a counter electrode (CE) surface area about 60-fold higher than the working electrode (WE) surface. Consequently, the CE current density variation is lower in comparison to the WE current density, allowing the CE to be considered as a pseudo-reference electrode. Therefore, the highest current density circulating in the electrolytic cell can be attributed to the WE current response.

All the experiments presented in this paper resulted from the functionalization of the WE with the complementary DNA probe



**FIGURE 2 |** Cyclic voltammetry response of a microband WE in a solution of 3 mM  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  and 0.5 M NaCl, with a  $10 \text{ mV.s}^{-1}$  scan rate and a  $0.5 \mu\text{L.s}^{-1}$  flow rate, showing the current absolute value decreasing after hybridization.

for miR 122 capture, diluted at  $10^{-7}$  M. The experimental section above describes the immobilization protocol leading to the DNA probe's self-assembled monolayer (SAM). The specificity of the biosensor was tested with different sequences of nucleic acid targets in an equimolar solution of  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  in NaCl with or without MB intercalation.

### Study of the Specificity in $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$

In a solution of  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  without MB, the hybridization of the nucleic acid target to the probe was deduced from the decreased current (in absolute value) compared to the measured current after single-stranded DNA-SAM formation (Figure 2).

The current decrease, measured by cyclic voltammetry, was due to the double-stranded nucleic acid limiting the electron transfer between the gold microelectrode and the redox tracer. To quantify the hybridization rate, we subtracted the current level of double-stranded nucleic acid ( $I_{\text{dsDNA}}$ ) from the current level of single-stranded DNA ( $I_{\text{ssDNA}}$ ). Since the gold microelectrodes' surface reproducibility depends on external conditions during the microfabrication process in the clean room, the current level of different microelectrodes might present certain differences. In order to compare all electrochemical measurements, a normalization procedure was systematically applied to the data by plotting the difference between  $I_{\text{ssDNA}} - I_{\text{dsDNA}}$  divided by the ssDNA current level (Eq. 1).

$$I_{\text{hyb}} = \frac{|I_{\text{ssDNA}} - I_{\text{dsDNA}}|}{|I_{\text{ssDNA}}|} \quad (1)$$


$I_{\text{ssDNA}}$  is the current level after immobilization of the DNA probe as SAM on the WE, and  $I_{\text{dsDNA}}$  is the current level after loading the nucleic acid target, both selected at  $-0.2$  V on the voltammograms.

To study the specificity of the electrochemical microfluidic chip, we tested different target sequences. The five sequences we focused on were a complementary sequence, a non-complementary sequence, and a sequence with a 1-base mismatch at different locations in the strand (at the beginning, in the middle, or at the end). The target concentration varied from  $10^{-18}$  to  $10^{-6}$  M. The hybridization detection was analyzed according to Eq. 1, and plotted as a calibration curve. The limit of detection (LoD)  $3\sigma$  is equal to 0.115 was calculated using the definitions established by Armbruster and Pry (2008). In Figure 3A, the probe and its complementary target (duplex named miR-122) serve as a reference for all hybridization tests since they can be considered as the perfect hybridization match. The hybridization detection experiment between the probe and the non-complementary target (named Nc T) served as a control test since no hybridization is expected (Figure 3A). As shown in Figure 3B, the sensor is highly specific regarding the target sequences with a 1-base mismatch when the mismatch is in the middle of the strand (experiment named T-1 M middle). The calibration curve is below the LoD  $3\sigma$ , meaning there was no detection from the sensor for this sequence. In contrast when the single-base mismatch is located at the beginning of the strand (experiment named T-1 M beginning, Figure 3C) or at the end of the strand (experiment named T-1 M end, Figure 3D), the discrimination between these experiments and the complementary one is impossible. This inability to discriminate can be explained by a partial hybridization of the two strands, even if the sequences are not perfectly complementary. In these cases, the double-stranded nucleic acid limits the electron transfer between the electrode and the electrolyte as efficiently as a complementary hybridization. Consequently, in a ferri/ferrocyanide solution without a DNA intercalator, the sensor does not enable discrimination of a partial hybridization (T-1 M beginning or T-1 M end) from a complementary hybridization with no mismatch.

### Study of the Specificity in $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$ and Methylene Blue

In order to improve the specificity of the sensor by distinguishing the hybridization detection of a single-base mismatch (independently of its location in the strand) from a complementary hybridization, we repeated the previous experiments in a solution of  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  in NaCl with MB. Methylene blue is well-known for its intercalation in double-stranded nucleic acids (Tuite and Norden, 1994; Kelley et al., 1997; Boon et al., 2003a; Boon et al., 2003b; Nafisi et al., 2007; Furst et al., 2015; Kékedy-Nagy and Ferapontova, 2018; Furst et al., 2019). In this electrocatalytic process, electrons pass from the electrode surface to the intercalated  $\text{MB}^+$ . Leucomethylene blue ( $\text{LB}^+$ ), the reduced form of  $\text{MB}^+$ , reduces the ferricyanide solution, allowing the catalytic cycle to continue (Boon et al., 2000). Due






to the catalytic nature of the system, the overall electrocatalytic response increases with the integration times. Therefore, the hybridization of the nucleic acid target to the probe was measured by chronoamperometry. In the case of double-stranded nucleic acids, more  $\text{MB}^+$  molecules are electrochemically reduced, which increases the concentration of active catalyst, thus increasing the absolute value of the current level measured after hybridization (Figure 4).

For these reasons, the normalization procedure explained above for Eq. 1 is modified to quantify the hybridization, by


subtracting  $I_{\text{ssDNA}}$  from  $I_{\text{dsDNA}}$  and normalizing by  $I_{\text{dsDNA}}$  (Eq. 2) as follows:

$$I_{\text{hyb}} = \frac{|I_{\text{dsDNA}} - I_{\text{ssDNA}}|}{|I_{\text{dsDNA}}|} \quad (2)$$

In Figure 5, the probe and its complementary target (duplex named miR 122) serve as a reference for all hybridization tests. As presented on Figure 5, when using 5  $\mu\text{M}$  MB in a 3 mM  $[\text{Fe}_{\text{III}}(\text{CN})_6]^{3-}/[\text{Fe}_{\text{II}}(\text{CN})_6]^{4-}$  solution for the detection of a single-mismatch located at the end of the strand, the calibration curve




**FIGURE 4** | Chronoamperometry at  $-0.2$  V in  $3$  mM  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  and  $0.5$  M NaCl with  $5$   $\mu\text{M}$  MB and a  $0.5$   $\mu\text{L}\cdot\text{s}^{-1}$  flow rate, showing the absolute value of the current increasing with the DNA target concentration.



**FIGURE 5** | Calibration curves in  $3$  mM  $[\text{Fe}(\text{III}) (\text{CN})_6]^{3-}/[\text{Fe}(\text{II}) (\text{CN})_6]^{4-}$  and  $0.5$  M NaCl with  $5$   $\mu\text{M}$  MB for the detection of DNA target at concentrations varying from  $10^{-18}$  M to  $10^{-6}$  M. For a DNA target complementary to the DNA probe sequence (red dots—called miR 122) compared to the calibration curves of a sequence with a one-base mismatch located (A) at the end (purple dots), or (B) at the beginning (pink dots) of the strand.

obtained is independent from the increasing targets concentration. Indeed, since the mismatch is far from the electrode surface, and even if the strands are partially

hybridized, the electrons cannot flow from the electrode to the electrolyte. In the opposite case, for a mismatch at the beginning of the strand, electrons can pass from the electrode to the



electrolyte more easily than for a mismatch at the end of the strand, resulting in an increase in the hybridization current. However, this current is still lower than that for the complementary strand, and can be distinguished from the complementary target by a concentration of the complementary target higher than  $10^{-12}$  M.

The next step will be to test the specificity of the sensor in a biological blood sample. In a biological sample, the sequence of interest is a micro-ribonucleic acid, while until now, the sequences studied were deoxyribonucleic acids. In this context, an intermediate step towards real blood samples was realized: the detection of DNA/DNA duplexes versus DNA/RNA duplexes was compared. **Figure 6** compares the calibration curves for the detection of a DNA target sequence and an RNA target sequence, both complementary to the same DNA probe sequence. The current levels measured being roughly the same, it underlines that there is no difference between the detection of a DNA target or an RNA target. As a result, the sensor's specificity results could be suitable for a blood sample analysis, for which a pre-treatment with commercially available miRNA extraction kits should be necessary.

## REFERENCES

- Abbaspour, A., and Noori, A. (2008). Electrochemical Studies on the Oxidation of Guanine and Adenine at Cyclodextrin Modified Electrodes. *Analyst* 133 (12), 1664–1672. doi:10.1039/b806920d
- Adachi, T., Nakanishi, M., Otsuka, Y., Nishimura, K., Hirokawa, G., Goto, Y., et al. (2010). Plasma MicroRNA 499 as a Biomarker of Acute Myocardial Infarction. *Clin. Chem.* 56 (7), 1183–1185. doi:10.1373/clinchem.2010.144121

## CONCLUSION

In conclusion, we developed a nucleic acid biosensor composed of an electrochemical cell integrated into a microfluidic chip, enabling the detection of a microRNA sequence in 30 min. Its specificity was improved by adding methylene blue (MB), a nucleic acid intercalator electrochemically active. Without MB, the electrochemical detection of a partial hybridization is indiscernible from the detection of a complete and complementary hybridization. By using MB at a micromolar concentration, the possibility of discriminating between these two types of hybridization was demonstrated, and the specificity of the biosensor improved, independent of the location of the mismatch in the sequence. The value of 1 a.m. is the LoD for two complementary strands. However, when the mismatch is located at the beginning of the strand, and for concentrations below 1 p.m., the calibration curve measured does not enable discrimination between this kind of mismatch and the perfect match, leading to the biosensor's sensitivity being limited to 1 p.m. Moreover, this protocol is promising for the electrochemical detection of nucleic acids in biological samples.

## DATA AVAILABILITY STATEMENT

The original contributions presented in the study are included in the article/Supplementary Material; further inquiries can be directed to the corresponding author.

## AUTHOR CONTRIBUTIONS

CP, JLG, JK, PL, and JG designed the experiments; CP, JLG, DB, and PL fabricated the biosensors in a clean room; CP and MF performed the electrochemical experiments; CP, JLG, MF, and JG analyzed the results; JG supervised the project; CP wrote the paper with the feedback of JG. All authors contributed to the article and approved the submitted version.

## ACKNOWLEDGMENTS

CP and MF thank the doctoral school “Electrical, Optical, Bio-Physics and Engineering” (ED575) and Paris-Saclay University for PhD grants. The authors would like to thank the DIMELEC project ANR-19-CE09-0016 for funding and the RENATECH clean room facilities at C2N, Palaiseau, France.

- Ariksoyal, D. O., Karadeniz, H., Erdem, A., Sengonul, A., Sayiner, A. A., and Ozsoz, M. (2005). Label-free Electrochemical Hybridization Genosensor for the Detection of Hepatitis B Virus Genotype on the Development of Lamivudine Resistance. *Anal. Chem.* 77 (15), 4908–4917. doi:10.1021/ac050022+
- Armbruster, D. A., and Pry, T. (2008). Limit of Blank, Limit of Detection and Limit of Quantitation. *Clin. Biochem. Rev.* 29 (Suppl. 1), S49–S52.
- Boon, E. M., Barton, J. K., Bhagat, V., Nersissian, M., Wang, W., and Hill, M. G. (2003). Reduction of Ferricyanide by Methylene Blue at a DNA-Modified Rotating-Disk Electrode. *Langmuir* 19 (22), 9255–9259. doi:10.1021/la030266u



- Boon, E. M., Ceres, D. M., Drummond, T. G., Hill, M. G., and Barton, J. K. (2000). Mutation Detection by Electrocatalysis at DNA-Modified Electrodes. *Nat. Biotechnol.* 18 (10 Suppl. L), 1096–1100. doi:10.1038/80301
- Boon, E. M., Jackson, N. M., Wightman, M. D., Kelley, S. O., Hill, M. G., and Barton, J. K. (2003). Intercalative Stacking: A Critical Feature of DNA Charge-Transport Electrochemistry. *J. Phys. Chem. B* 107 (42), 11805–11812. doi:10.1021/jp030753i
- Butterworth, A., Blues, E., Williamson, P., Cardona, M., Gray, L., and Corrigan, D. K. (2019). SAM Composition and Electrode Roughness Affect Performance of a DNA Biosensor for Antibiotic Resistance. *Biosensors* 9 (22), 22. doi:10.3390/bios9010022
- Calin, G. A., and Croce, C. M. (2006). MicroRNA Signatures in Human Cancers. *Nat. Rev. Cancer* 6 (11), 857–866. doi:10.1038/nrc1997
- Cissell, K. A., and Deo, S. K. (2009). Trends in microRNA Detection. *Anal. Bioanal. Chem.* 394 (4), 1109–1116. doi:10.1007/s00216-009-2744-6
- Drummond, T. G., Hill, M. G., and Barton, J. K. (2003). Electrochemical DNA Sensors. *Nat. Biotechnol.* 21 (10), 1192–1199. doi:10.1038/nbt873
- Ferguson, B. S., Buchsbaum, S. F., Swensen, J. S., Hsieh, K., Lou, X., and Soh, H. T. (2009). Integrated Microfluidic Electrochemical DNA Sensor. *Anal. Chem.* 81 (15), 6503–6508. doi:10.1021/ac900923e
- Furst, A. L., Hill, M. G., and Barton, J. K. (2015). A Multiplexed, Two-Electrode Platform for Biosensing Based on DNA-Mediated Charge Transport. *Langmuir* 31, 6554–6562. doi:10.1021/acs.langmuir.5b00829
- Furst, A. L., Muren, N. B., and Hill, M. G. (2019). Electrochemistry toward Multimarker and Functional Assays from Crude Cell Lysates: Controlling Spacing and Signal Amplification in DNA-CT – Based Bioelectrochemical Devices. *Curr. Opin. Electrochem.* 14, 104–112. doi:10.1016/j.coelec.2018.12.008
- Gooding, J. J. (2002). Electrochemical DNA Hybridization Biosensors. *Electroanalysis* 14 (17), 1149–1156. doi:10.1002/1521-4109(200209)14:17<1149::aid-elant1149>3.0.co;2-8
- Grieshaber, D., MacKenzie, R., Vörös, J., and Reimhult, E. (2008). Electrochemical Biosensors – Sensor Principles and Architectures. *Sensors* 8, 1400–1458. doi:10.3390/s8031400
- Horny, M.-C., Lazerges, M., Siaugue, J.-M., Pallandre, A., Rose, D., Bedioui, F., et al. (2016). Electrochemical DNA Biosensors Based on Long-Range Electron Transfer: Investigating the Efficiency of a Fluidic Channel Microelectrode Compared to an Ultramicroelectrode in a Two-Electrode Setup. *Lab. Chip* 16 (22), 4373–4381. doi:10.1039/c6lc00869k
- Ji, X., Takahashi, R., Hiura, Y., Hirokawa, G., Fukushima, Y., and Iwai, N. (2009). Plasma miR-208 as a Biomarker of Myocardial Injury. *Clin. Chem.* 55 (11), 1944–1949. doi:10.1373/clinchem.2009.125310
- Kékedy-Nagy, L., and Ferapontova, E. E. (2018). Sequence-Specific Electron Transfer Mediated by DNA Duplexes Attached to Gold through the Alkanethiol Linker. *J. Phys. Chem. B* 122 (44), 10077–10085. doi:10.1021/acs.jpcc.8b07505
- Kelley, S. O., Barton, J. K., Jackson, N. M., and Hill, M. G. (1997). Electrochemistry of Methylene Blue Bound to a DNA-Modified Electrode. *Bioconjug. Chem.* 8 (1), 31–37. doi:10.1021/bc960070o
- Kelley, S. O., Jackson, N. M., Hill, M. G., and Barton, J. K. (1999). Long-range Electron Transfer through DNA Films. *Angew. Chem. Int. Ed.* 38 (7), 941–945. doi:10.1002/(sici)1521-3773(19990401)38:7<941::aid-anie941>3.0.co;2-7
- Minaei, M. E., Saadati, M., Najafi, M., and Honari, H. (2015). DNA Electrochemical Nanobiosensors for the Detection of Biological Agents. *J. Appl. Biotechnol. Rep.* 2 (1), 175–185.
- Nafisi, S., Saboury, A. A., Keramat, N., Neault, J.-F., and Tajmir-Riahi, H.-A. (2007). Stability and Structural Features of DNA Intercalation with Ethidium Bromide, Acridine orange and Methylene Blue. *J. Mol. Struct.* 827, 35–43. doi:10.1016/j.molstruc.2006.05.004
- Nassi, A., Guillon, F.-X., Amar, A., Hainque, B., Amriche, S., Maugé, D., et al. (2016). Electrochemical DNA-Biosensors Based on Long-Range Electron Transfer: Optimization of the Amperometric Detection in the Femtomolar Range Using Two-Electrode Setup and Ultramicroelectrode. *Electrochimica Acta* 209, 269–277. doi:10.1016/j.electacta.2016.04.144
- Quan Li, P., Piper, A., Schmueser, I., Mount, A. R., and Corrigan, D. K. (2017). Impedimetric Measurement of DNA-DNA Hybridisation Using Microelectrodes with Different Radii for Detection of Methicillin Resistant *Staphylococcus aureus* (MRSA). *Analyst* 142, 1946–1952. doi:10.1039/c7an00436b
- Rosario, R., and Mutharasan, R. (2014). Nucleic Acid Electrochemical and Electromechanical Biosensors: a Review of Techniques and Developments. *Rev. Anal. Chem.* 33 (4), 213–230. doi:10.1515/revac-2014-0017
- Shoaei, N., Forouzandeh, M., and Omidfar, K. (2018). Voltammetric Determination of the *Escherichia coli* DNA Using a Screen-Printed Carbon Electrode Modified with Polyaniline and Gold Nanoparticles. *Microchim. Acta* 185 (4), 1–9. doi:10.1007/s00604-018-2749-y
- Siracusa, J., Koulmann, N., Bourdon, S., Goriot, M.-E., and Banzet, S. (2016). Circulating miRNAs as Biomarkers of Acute Muscle Damage in Rats. *Am. J. Pathol.* 186 (5), 1313–1327. doi:10.1016/j.ajpath.2016.01.007
- Siracusa, J., Koulmann, N., Sourdille, A., Chapus, C., Verret, C., Bourdon, S., et al. (2018). Phenotype-specific Response of Circulating miRNAs Provides New Biomarkers of Slow or Fast Muscle Damage. *Front. Physiol.* 9, 684. doi:10.3389/fphys.2018.00684
- Squires, T. M., and Quake, S. R. (2005). Microfluidics: Fluid Physics at the Nanoliter Scale. *Rev. Mod. Phys.* 77 (3), 977–1026. doi:10.1103/revmodphys.77.977
- Tsaloglou, M.-N., Nemiroski, A., Camci-Unal, G., Christodouleas, D. C., Murray, L. P., Connelly, J. T., et al. (2018). Handheld Isothermal Amplification and Electrochemical Detection of DNA in Resource-Limited Settings. *Anal. Biochem.* 543, 116–121. doi:10.1016/j.ab.2017.11.025
- Tuite, E., and Norden, B. (1994). Sequence-specific Interactions of Methylene Blue with Polynucleotides and DNA: a Spectroscopic Study. *J. Am. Chem. Soc.* 116, 7548–7556. doi:10.1021/ja00096a011

**Conflict of Interest:** The authors declare that the research was conducted in the absence of any commercial or financial relationships that could be construed as a potential conflict of interest.

**Publisher's Note:** All claims expressed in this article are solely those of the authors and do not necessarily represent those of their affiliated organizations, or those of the publisher, the editors, and the reviewers. Any product that may be evaluated in this article, or claim that may be made by its manufacturer, is not guaranteed or endorsed by the publisher.

Copyright © 2022 Poujouly, Le Gall, Freisa, Kechkeche, Bouville, Khemir, Gonzalez-Losada and Gamby. This is an open-access article distributed under the terms of the Creative Commons Attribution License (CC BY). The use, distribution or reproduction in other forums is permitted, provided the original author(s) and the copyright owner(s) are credited and that the original publication in this journal is cited, in accordance with accepted academic practice. No use, distribution or reproduction is permitted which does not comply with these terms.