

# A New Sitting-type Lower-Limb Rehabilitation Robot based on a Spatial Parallel Kinematic Machine

J Mohanta, S Mohan, Philippe Wenger, Christine Chevallereau

# ▶ To cite this version:

J Mohanta, S Mohan, Philippe Wenger, Christine Chevallereau. A New Sitting-type Lower-Limb Rehabilitation Robot based on a Spatial Parallel Kinematic Machine. Springer. Lecture Notes in Mechanical Engineering, 2021, 978-981-15-4476-7. hal-03639735

# HAL Id: hal-03639735 https://hal.science/hal-03639735

Submitted on 13 Apr 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers. L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

## A New Sitting-type Lower-Limb Rehabilitation Robot based on a Spatial Parallel Kinematic Machine

J.K Mohanta<sup>1</sup>,S. Mohan<sup>2</sup>, P. Wenger<sup>3</sup>, C.Chevallereau<sup>3</sup> <sup>1</sup>Department of Electrical Engineering, IIT Kanpur, India Email: jkmjayant@gmail.com <sup>2</sup>Department of Mechanical Engineering, IIT Pallakad, India Email: santhakumar@iitpkd.ac.in <sup>3</sup>Laboratoire des Sciences du Numérique de Nantes (LS2N), CNRS, France Email: {Philippe.Wenger, Christine.Chevallereau}@ls2n.fr

#### Abstract

This paper proposes a lying/sittingtype lower-limb rehabilitation robot based on a three degrees of freedom spatial parallel kinematic manipulator namely "The Orthoglide" along with an actuated rotational degree of freedom at the endeffector for the purpose of lower-limb rehabilitation treatments. This rehabilitation robot is an end-effector/foot-plate based mechanism which controls the ankle-joint movements to assist physiotherapists in performing therapeutic treatments along with a passive orthosis (supporting system). The conceptual and detail design of the rehabilitation device is presented along with its kinematics. Further, the functional validation of the proposed robot in terms of the kinematic motion capabilities to provide physiological motions like knee flexion-extension, hip adduction-abduction and hip flexion-extension are also investigated. The required safety features for the proposed rehabilitation robot in terms of mechanical constraints and non-actuated joints are also discussed.

**Keywords:** Lower limb rehabilitation robot; Orthoglide; foot-plate based; parallel manipulator; passive orthosis;

#### **1** Introduction

Lower limb treatments using mechanized systems has improved the speed of recovery of patients from the bedridden state [1]. The lower limb impairments are caused by motor nerves dysfunction, which are generally caused by spinal cord injury (SCI), stroke and accidental disability. Continuous physical motion of the impaired limb, known as motor rehabilitation, required for this process [2]. These are done in numerous ways for example with a body weight support (BWS), using supportive orthosis, using foot-pate motion and limb motion in sitting/lying posture. The sitting/lying treatments are appropriate for the bedridden patients. Example of sitting/lying type systems are the ANKUR-LL2 [8], an improved Lambda [6], the Supine [4], Physiotherabot [3], the Lambda [5] and the ANKUR-LL1 [7].

Lower limb mechanisms with high mechanical advantage joints, outfits the system with a self-locking capability, which helps in case of emergency stop [9]. Structures of the mechanisms consist of bulky actuators and system appears dynamically heavy. To avoid such difficulty a passive light-weight orthosis, with no actuators, is used along with joint feedback sensors [6]. Here the orthosis is indirectly controlled.

Rehabilitation devices are generally having motion in sagittal plane[1]. Although many therapies need off-planar movements but complexity of the arrangement and

safety of the patient is major concern. In this direction a device namely Hipbot [10] is designed to give off-planar motions to the lower limbs, when the patient is in lying posture. Still Hipbot is having limitation that it can give adduction-abduction away from the body not towards the body because of physical limitations. To address this issue a new manipulator configuration is proposed. The major difficulty in off-planar movement is movement of the actuator assembly as they are mounted in single plane. So, here we have used a parallel manipulator based mechanism with passive orthosis which can render four degrees of freedom to the lower limb.

The rest of the paper is organized as follows:Section 2 discusses the architecture of the proposed rehabilitation robot and its geometrical aspects. Section3 describes thesystemkinematics, Section 4 demonstrates the physiological motion validation of the proposed manipulator and section 5 concludes the paper with safety aspects and scope of future works.

#### 2 Proposed Manipulator's Architecture

The proposed rehabilitation robot shown in Figure 1 consists of a driving mechanism and a passive orthosis. The driving mechanism is made using the concept of a parallel manipulator namely "The Orthoglide" [11]. Instead of an overhead structure a modified version with actuators on ground is taken for usage in rehabilitation purpose. Here three linear joints are placed orthogonally and three links with each leg having configuration PUU is connecting to the end-effector. The orthosis consists of three links corresponding to the thigh, crus and foot of the limb with a vertical hip joint. These three links are joined in series at the hip, knee and ankle positions, with rotary joint. The solid model of the proposed orthosis is shown in Figure 2. The passive orthosis can be seen in Figure 1. Forward kinematics for the passive orthosis is as follows:

$$x = \left(L_{thigh}\cos\phi_1 + L_{crus}\cos\left(\phi_1 - \phi_2\right)\right)\cos\phi_3$$
$$y = \left(L_{thigh}\cos\phi_1 + L_{crus}\cos\left(\phi_1 - \phi_2\right)\right)\sin\phi_3 \tag{1}$$
$$z = L_{thigh}\sin\phi_1 + L_{crus}\sin\left(\phi_1 - \phi_2\right)$$





Figure 1: Proposed Lower limb rehabilitation robot

Here  $\phi_1$ ,  $\phi_2$  and  $\phi_3$  are the hip flexion-extension, knee flexion-extension and hip adduction-abduction angles, which are shown in Figure 2.  $L_{thigh}$  and  $L_{crus}$  are the length of the thigh support and crus support. Where (x, y, z) are the coordinate location of the ankle joint. The required workspace based on the physiological motion specifications are plotted and shown in Figure 3. Here the reachable workspace is shown by considering the hip flexion-extension joint as origin. The possible joint movements on the sagittal plane for the human limbs are as:  $-30^\circ \le \phi_1 \le 60^\circ$ ,  $0^\circ \le \phi_2 \le 105^\circ$  and  $-15^\circ \le \phi_3 \le 15^\circ$ . In the Figure 3,  $\theta_S = \phi_1 - \phi_2$ , shows orientation of the crus link at the ankle jointHere, the ankle rotation is controlled by an independent actuator which is placed at the ankle joint. the angle requirement for the foot link at the end effector is.  $\theta_s + \phi_{ankle}$ .



Figure 4 shows the junction end-effector arrangement for the system, where the orthosis ankle joint and Orthoglide's end-effector is connected. The orthosis ankle joint is mounted on a body. This body has a passive vertical rotary joint at the bottom

Paper no. Asian MMS-2018-128



of the Orthoglide's end-effector's cube. This enable the orthosis not to hinder the hip adduction-abduction motion.

Figure 4: End-Effector orientation arrangement

## **3 Manipulator Kinematics**

For the Orthoglide kinematic solution can be determined by the following relations

$$(x - x_1)^2 + y^2 + z^2 = l^2$$
  

$$x^2 + (y - y_2)^2 + z^2 = l^2$$
  

$$x^2 + y^2 + (z - z_3)^2 = l^2$$
(2)

Here,  $x_1 = r_1 - d_1$ ,  $y_2 = r_2 + d_2$  and  $z_3 = r_3 + d_3$ . Where  $r_1$ ,  $r_2$  and  $r_3$  are the joint space variable. Hence our joint space vector consist of  $[r_1 r_2 r_3 \theta_{ankle}]^T$  and task space consists of  $[\phi_1 \phi_2 \phi_3 \phi_{ankle}]$ 

Inverse kinematics for the Orthoglide part can be written as follows:

$$r_{1} = \sqrt{l^{2} - y_{o}^{2} - z_{o}^{2}} + x_{o} + d_{1}$$

$$r_{2} = \sqrt{l^{2} - x_{o}^{2} - z_{o}^{2}} - y_{o} - d_{2}$$

$$r_{3} = \sqrt{l^{2} - x_{o}^{2} - y_{o}^{2}} + z_{o} - d_{3}$$
(3)

Where the  $x_o$ ,  $y_o$  and  $z_o$  are the Orthoglide end-effector point.



Figure 5: Kinematic parameters of Orthoglide

Kinematics of the orthosis system is shown in eq. (1) relation between the orthosis coordinates to the to the end-effector coordinates of the Orthoglide can be written as  $x = x - p - y \sin \phi$ 

$$y_o = y - p_y - y_t \cos \phi_3$$

$$z_o = z + p_z - z_t$$
(4)

By substituting eq. (4) and eq.(1) in eq. (3) we can obtain complete inverse kinematics relation for the assembly. The parameters  $y_t$  and  $z_t$  are shown in Figure 4. The point (x, y, z) is with respect to (w.r.t.) frame {2}, while  $(x_o, y_o, z_o)$  is w.r.t. frame {1}. To bring all the points to one coordinates system linear transformation is applied to the frames the distances  $\begin{bmatrix} p_x & p_y & p_z \end{bmatrix}$  are the linear transformation for the coordinates. In motion assembly, the parameters  $\begin{bmatrix} r_1 & r_2 & r_3 \end{bmatrix} = f(\phi_1, \phi_2, \phi_3)$  while the ankle joint is having independent actuator whose actuation is represented by  $\phi_{ankle}$  which is a function of hip and knee adduction-abduction angle,  $\theta_{ankle} = f(\phi_1, \phi_2, \phi_{ankle})$ .



Figure 6: Passive orthosis

## **4** Physiological Motions

Physiological motions are decided by the therapist according to rehabilitation status of the patient. Here in this section basic physiological motions are shown to validate the kinematics of the proposed system. Each individual joint are given motion according to the angular limits used in therapeutic treatments [12].

### 4.1 Hip flexion-extension

Hip flexion-extension motion refers to motion of hip joint about the sagittal plane. To visualize this motion the values input is shown in Table 1.

Table 1: Task-space for validation of hip flexion-extension motion

| Joint | $\phi_1$   | $\phi_2$ | $\phi_3$ | $\phi_{ankle}$ |
|-------|--|----------|----------|----------------|
| Value | $\left(\frac{60(168.75 - (t - 15)^2)}{225}\right)^o$ | $10^{o}$ | $0^o$    | $0^{o}$        |

To validate the motion possibility each joint is simulated for a 30 second motion trajectory. Here in this case the hip flexion-extension,  $-15^{o} < \phi_{1} < 45^{o}$  is simulated the geometric parameters are taken as:  $L_{thigh} = 0.42$  m,  $L_{crus} = 0.4$  m, l = 1.4 m,  $d_{1} = 0.28$ m,  $d_{2} = 0.4$  m,  $d_{3} = 1.6$  m,  $p_{x} = 1.2$  m,  $p_{y} = 0.14$  m,  $p_{z} = 0.917$  m,  $y_{t} = 0.05$  m and  $z_{t} = 0.1$  m.



Figure 7: Hip flexion-extension joint space movement

The joint-space movement shown in Figure 7. shows that the motion is possible and mechanism is not encountering any singularities while performing the task. Parallel singularity can be avoided by regulation the parameter l, higher values are more stable, which is the length of the connecting link from the actuator to end endeffector with universal joints. Furthermore, optimization can be done as this parameter. The stroke length of the linear actuator is taken as 1.2 m. Hence, during the motion system is not overruling the limits.

#### 4.2 Hip adduction-abduction

The trajectory taken for hip adduction-abduction is shown in Table 2. Here we have given motion to  $\phi_3$  as,  $-15^{\circ} < \phi_3 < 15^{\circ}$  using as sinusoidal function with a time period of 30 seconds. The trajectory would be a circle in transverse plane as the patient is in lying posture.

Table 2: Task-space for validation of hip adduction-abduction motion

| Joint | $\phi_{\mathrm{l}}$ | $\phi_2$ | $\phi_3$        | $\phi_{ankle}$ |
|-------|---------------------|----------|-----------------|----------------|
| Value | $30^{o}$            | $20^{o}$ | $15\sin(12t)^o$ | $0^o$          |



This joint-space in Figure 8 clearly shows that mechanism is able to providebidirectional adduction and abduction which was not possible in case of Hipbot. Only one condition is to be verified that the lower limb under treatment is should not hit the other limb which is not being treated.

#### 4.3 Knee flexion-extension

For knee flexion-extension the  $\phi_2$  angle is to be varied, this angle cannot be less than zero as it is constrained by human body constraints, also a safety feature in the knee joint is also required which stops the orthosis to go less than zero physically, for this simulation the task-space is shown in Table 3 and the joint-space output is shown in

Figure 9. Here we have taken  $0 < \phi_2 < 40^\circ$  using a polynomial equation.

| Joint | $\phi_{l}$ | $\phi_2$  | $\phi_3$ | $\phi_{ankle}$ |
|-------|------------|---|----------|----------------|
| Value | $10^{o}$   | $\left(\frac{40(225 - (t - 15)^2)}{225}\right)^o$ | $0^{o}$  | $0^o$          |

Table 3: Task-space for validation of knee flexion-extension motion

All the joint must be protected by physical restriction and numerical restriction through integrated software. Other than this two emergency switches must be installed. One near patient's hand and another with the physiotherapist. These treatments must be provided in presence of a therapist. This devices are supportive devices which increases the efficiency of the therapist to give long duration treatments without making the therapist tired.



#### **5** Conclusions

Lower limb rehabilitation techniques are designed in order to provide continuous therapy for a longer period of time at a stretch. For such purpose the mechanisms should perform continuous operation without facing any kind of singularity and joint space limitations. Also, it should be safe for the patient and the therapist. The proposed system is designed to provide and addition dimension to the therapists that is hip adduction-abduction which is not present with proper features in other devices. The design and kinematics is explained in order to realize the system. The basic physiological motions are kinematically simulated to verify the manipulator behavior towards the joint input. This architecture has potential to provide wide range of rehabilitation treatments. In future we would like to find dynamic based control schemes and real-time validation on human subject.

### References

- 1. I. Diaz, J.J. Gil and E. S'anchez, "Lower-limb robotic rehabilitation: Literature review and challenges", Journal of Robotics, 2011, Article ID 759764, 2011
- M.J.Johnson, H.Schmidt, "Robot assisted neurological rehabilitation at home: motivational aspects and concepts for tele-rehabilitation", Public Health Forum, 17(4) 2009,8e1–8e4
- S.Mohan, J.K. Mohanta,S. Kurtenbach,J. Paris,B. Corves,M.Huesing, "Design, development and control of a 2PRP-2PPR planar parallel manipulator for lower limb rehabilitation therapies", Mechanism and Machine Theory, 112, 2017, 272–294
- J.K. Mohanta, C. Saxena,G.Gupta, M.Santhakumar, "Kinematic analysis of a passive sitting/lying type lower limb rehabilitation robot, in: Proc. of the 2nd international Conference on Machines and Mechanisms, Kanpur, India, 2015, 1–12
- 5. F.M. Lim, R. Foong, H.Yu, "A supine gait training device for stroke rehabilitation", Journal of Medical Devices 8 (2), 2014
- 6. E.Akdoğan, M.A. Adli, "The design and control of a therapeutic exercise robot for lower limb rehabilitation: Physiotherabot", Mechatronics, 21(3), 2011, 509-522

- M.Bouri,B.Le Gall,R.Clavel, "A new concept of parallel robot for rehabilitation and fitness: The Lambda" In IEEE International Conference on Robotics and Biomimetics (ROBIO), 2009, 2503-2508
- J.K. Mohanta, S. Mohan, P. Deepasundar, R.Kiruba-Shankar, "Development and control of a new sitting-type lower limb rehabilitation robot", Computers and Electrical Engineering, 67, 2018, 330-347
- W. Wang, Z. Hou,L.Tong, F.Zhang,Y. Chen,M.H. Tan, "A novel leg orthosis for lower limb rehabilitation robots of the sitting/lying type", Mechanism and Machine Theory, 74, 2014,337-353
- C.H. Guzmán-Valdivia, A.Blanco-Ortega, M.A. Oliver-Salazar, F.A.Gómez-Becerra, J.L. Carrera-Escobedo, "HipBot – The design, development and control of a therapeutic robot for hip rehabilitation", Mechatronics, 30, 2015, 55-64
- 11. D. Chablat, P. Wenger. "Architecture optimization of a 3-DOF translationalparallel mechanism for machiningapplications, the Orthoglide",IEEE Transactions on robotics and automation, 19(3), 2003, 403-410
- B.W. Stansfield,S.J.Hillman, M.E.Hazlewood,J.E.Robb, "Regression analysis of gait parameters with speed in normal children walking at self-selected speeds", Gait & Posture, 23, 2006, 288–294