



HAL
open science

Validation of a novel system to assess end-expiratory lung volume and alveolar recruitment in an ARDS model

Laurent Bitker, Nadja Cristinne Carvalho, Sascha Reidt, Christoph Schranz, Dominik Novotni, Maciej Orkisz, Eduardo Davila Serrano, Jean-Pierre Revelly, Jean-Christophe Richard

► To cite this version:

Laurent Bitker, Nadja Cristinne Carvalho, Sascha Reidt, Christoph Schranz, Dominik Novotni, et al.. Validation of a novel system to assess end-expiratory lung volume and alveolar recruitment in an ARDS model. *Intensive Care Medicine Experimental*, 2021, 9 (1), pp.46. 10.1186/s40635-021-00410-x . hal-03688174

HAL Id: hal-03688174

<https://hal.science/hal-03688174>

Submitted on 20 Mar 2024

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

RESEARCH ARTICLES

Open Access



Validation of a novel system to assess end-expiratory lung volume and alveolar recruitment in an ARDS model

Laurent Bitker^{1,2}, Nadja Cristinne Carvalho³, Sascha Reidt³, Christoph Schranz³, Dominik Novotni³, Maciej Orkisz², Eduardo Davila Serrano², Jean-Pierre Revelly³ and Jean-Christophe Richard^{1,2*} 

*Correspondence:

j-christophe.richard@chu-lyon.fr

¹ Service de Médecine Intensive - Réanimation, Hôpital De La Croix Rousse, Hospices Civils de Lyon, 103 Grande Rue de la Croix Rousse, 69004 Lyon, France Full list of author information is available at the end of the article

Abstract

Background: Personalizing mechanical ventilation requires the development of reliable bedside monitoring techniques. The multiple-breaths nitrogen washin–washout (MBNW) technique is currently available to measure end-expiratory lung volume (EELV_{MBNW}), but the precision of the technique may be poor, with percentage errors ranging from 28 to 57%. The primary aim of the study was to evaluate the reliability of a novel MBNW bedside system using fast mainstream sensors to assess EELV in an experimental acute respiratory distress syndrome (ARDS) model, using computed tomography (CT) as the gold standard. The secondary aims of the study were: (1) to evaluate trending ability of the novel system to assess EELV; (2) to evaluate the reliability of estimated alveolar recruitment induced by positive end-expiratory pressure (PEEP) changes computed from EELV_{MBNW}, using CT as the gold standard.

Results: Seven pigs were studied in 6 experimental conditions: at baseline, after experimental ARDS and during a decremental PEEP trial at PEEP 16, 12, 6 and 2 cmH₂O. EELV was computed at each PEEP step by both the MBNW technique (EELV_{MBNW}) and CT (EELV_{CT}). Repeatability was assessed by performing replicate measurements. Alveolar recruitment between two consecutive PEEP levels after lung injury was measured with CT (Vrec_{CT}), and computed from EELV measurements (Vrec_{MBNW}) as ΔEELV minus the product of ΔPEEP by static compliance.

EELV_{MBNW} and EELV_{CT} were significantly correlated ($R^2 = 0.97$). An acceptable non-constant bias between methods was identified, slightly decreasing toward more negative values as EELV increased. The conversion equation between EELV_{MBNW} and EELV_{CT} was: $EELV_{MBNW} = 0.92 \times EELV_{CT} + 36$. The 95% prediction interval of the bias amounted to ± 86 mL and the percentage error between both methods amounted to 13.7%.

The median least significant change between repeated measurements amounted to 8% [CI_{95%}: 4–10%]. EELV_{MBNW} adequately tracked EELV_{CT} changes over time (concordance rate amounting to 100% [CI_{95%}: 87%–100%] and angular bias amounting to $-2^\circ \pm 10^\circ$).

Vrec_{MBNW} and Vrec_{CT} were significantly correlated ($R^2 = 0.92$). A non-constant bias between methods was identified, slightly increasing toward more positive values as Vrec increased.

Conclusions: We report a new bedside MBNW technique that reliably assesses EELV in an experimental ARDS model with high precision and excellent trending ability.

Keywords: Acute respiratory distress syndrome, Positive end-expiratory pressure, Lung injury, End-expiratory lung volume, Functional residual capacity, Computed tomography, Alveolar recruitment, Nitrogen washin–washout

Background

Despite improvement in therapeutic management, acute respiratory distress syndrome (ARDS) mortality remains high in observational studies [1]. Personalizing mechanical ventilation may improve ARDS prognosis [2], but requires the development of reliable bedside monitoring techniques. End-expiratory lung volume (EELV) monitoring is appealing as it provides information on the size of the baby lung and lung strain, and may help individualize respiratory settings during mechanical ventilation of ARDS patients, by estimating alveolar recruitment related to positive end-expiratory pressure (PEEP) changes [3]. However, this clinical application has not been validated so far, using computed tomography (CT) as the reference technique.

The multiple-breaths nitrogen washin–washout (MBNW) technique is currently available on some ICU ventilators, and has been shown reliable to assess EELV in both mechanically ventilated patients at PEEP 5 cmH₂O [4] and in experimental studies at higher PEEP [5, 6], using CT as the reference technique to measure EELV. However, the precision of the technique may be poor, with percentage errors ranging from 28 to 57%, and is accompanied with a systematic bias at high PEEP levels identified in one study [6]. Such bias at high PEEP levels was merit of attention of a previous study, where the presence of overinflated lung region and increased alveolar dead space at higher PEEP levels partially explained the substantial difference between EELV measured by CT and MBNW [7]. A new non-invasive MBNW monitor was designed, using fast mainstream sensors enabling continuous and synchronous measurements of proximal flow and O₂/CO₂ concentrations. We hypothesized that this new system might be more reliable to assess EELV.

The aim of the study was primarily to evaluate the reliability of this new non-invasive MBNW monitor to measure EELV at bedside (bias, precision, percentage error, and repeatability) in an experimental model of ARDS using CT as the gold standard. Secondly, the trending ability of the new MBNW monitor and the reliability of estimated alveolar recruitment induced by PEEP changes computed from EELV_{MBNW} measurements were evaluated, using CT as the gold standard.

Methods

The study was approved by an Institutional Review Board for the care of animal subjects (Comité d'éthique en experimentation animale n°042, APAFIS#21542–2019071909353561), and carried out on 7 female Large White pigs weighing 29 [29, 30] kg, in accordance with Good Laboratory Practice Regulation [8].

Animal preparation

Pigs were premedicated with an intramuscular injection of xylazine 2% (0.7 mg kg⁻¹), ketamine (17 mg kg⁻¹), and droperidol (0.17 mg kg⁻¹). Surgical preparation was

performed in the supine position under continuous general anesthesia by propofol (7 mg kg h^{-1}) and fentanyl ($10 \text{ } \mu\text{g kg h}^{-1}$). Body temperature was maintained constant at $38 \pm 1 \text{ } ^\circ\text{C}$ using a heating pad.

A 6.0 tracheal tube was introduced via median surgical tracheotomy, secured hermetically after cuff inflation, and connected to a HAMILTON-C6 ventilator (Hamilton Medical AG, Bonaduz, SWITZERLAND). Mechanical ventilation was initially performed through a heated humidifier (HAMILTON-H900, Hamilton Medical AG, Bonaduz, SWITZERLAND) in volume-controlled mode, constant inspiratory flow, tidal volume (V_T) 10 mL kg^{-1} , inspired fraction of oxygen (FiO_2) 50%, PEEP 5 cm H_2O and respiratory rate adjusted to achieve normocapnia. Muscle relaxation was obtained with cisatracurium (30 mg h^{-1}).

The left jugular vein was surgically cannulated with a 3 lm 8.5 Fr catheter (model Arrow # CV-12853, Teleflex, Wayne, PA, USA) for drug administration. A 7-french pulmonary artery catheter (model #131F7, Edwards, Irvine, CA, USA) was advanced through the right external jugular vein to assess central venous pressure, pulmonary artery pressure and cardiac output. The left carotid artery was surgically cannulated with an 8.5 Fr catheter for continuous monitoring of arterial blood pressure.

Experimental protocol (Fig. 1)

Experimental ARDS was achieved by saline lavage under FiO_2 90%, V_T 10 mL kg^{-1} and PEEP 5 cm H_2O . Repeated intra-tracheal instillations of 1000 mL 0.9% sodium chloride warmed at $37 \text{ } ^\circ\text{C}$ was performed, and lung fluid was removed by both gravity and endotracheal suctioning. The procedure aimed to achieve a $\text{PaO}_2/\text{FiO}_2$ ratio $< 100 \text{ mm Hg}$ 15 min after the last lung lavage. Respiratory rate (RR) could be increased up to 30 breaths per min to maintain pH above 7.20. A recruitment maneuver was then executed by applying a continuous airway pressure of 40 cm H_2O during 20 s, V_T was set to 6 mL kg^{-1} , FiO_2 was set to 80%, and a decremental PEEP trial (16, 12, 6, and 2 cm H_2O) was performed. PEEP could possibly be increased by 1 cm H_2O increments at the last two PEEP steps if the peripheral oxygen saturation fell below 88% to maintain the animal alive.

The following measurements were successively performed at each protocol step, i.e., at baseline, 15 min after the achievement of experimental ARDS criteria, and after a 10-min stabilization period at each PEEP step (Fig. 1): arterial blood gas, hemodynamic measurements, respiratory mechanics assessment, first assessment of EELV by MBNW, assessment of EELV by CT, and ultimately replicate assessment of EELV by MBNW.


Measurements

Arterial blood gas

Arterial blood gases were analyzed with an ABL80 FLEX CO-OX blood gas analyzer (Radiometer, Copenhagen, DENMARK) for oxygen and carbon dioxide partial pressures, pH and lactate.

Hemodynamic measurements

Pressure transducers (Abbott, San Clemente, CA, USA) were positioned at the mid-chest level and connected to a Philips monitor (M1205A Philips Healthcare, Best,



NETHERLAND). Cardiac output was measured in triplicate with the thermodilution technique, using 10 ml saline injected at room temperature.

Respiratory measurements

Airway pressure and flow were measured by the HAMILTON proximal flow sensor of the mechanical ventilator. Total PEEP and plateau pressure were measured at the end of 3 s end-expiratory and end-inspiratory pauses, respectively. Peripheral capillary oxygen saturation was measured continuously on the tip of the animal tongue.

Multiple-breaths nitrogen washin–washout determination of end-expiratory lung volume (EELV_{MBNW})

EELV_{MBNW} was determined using a MBNW maneuver, consisting of a change in *FiO₂* by 10% and reaching a new steady-state, before resetting the *FiO₂* to the original value. A novel mainstream *CO₂/O₂* gas sensor was placed between the proximal flow sensor and the tracheal tube, measuring *O₂/CO₂* partial pressures at a frequency of 100 Hz. Gas flow, airway pressure and *O₂/CO₂* partial pressures were acquired by the ventilator and relayed via a serial connection to a computer running a custom MATLAB program (MathWorks, Natick, MA, USA). The signals were processed online and automatically calculating the breath-by-breath volumetric gas exchange [9]. The cumulative volume

of nitrogen washed-out by the increase in FiO_2 was determined by integrating gas flow, O_2 and CO_2 partial pressure signals and subtracting the nitrogen baseline determined during the steady-states. The cumulative volume of nitrogen retained during the wash-in phase was computed accordingly. $\text{EELV}_{\text{MBNW}}$ was then computed as:

$$\text{EELV}_{\text{MBNW}} = \frac{\text{Cumulative } \Delta \text{N}_2 \text{ volume}}{\Delta \text{N}_2 \text{ fraction}}.$$

$\text{EELV}_{\text{MBNW}}$ measured during wash-in and wash-out were averaged. To allow meaningful comparisons with EELV_{CT} , instrumental dead-space, i.e., the volume from the O_2/CO_2 sensor to the tip of the tracheal tube was measured and subtracted from $\text{EELV}_{\text{MBNW}}$. The stability of carbon dioxide production (VCO_2) and oxygen consumption (VO_2) before and after the MBNW maneuver was verified. Repeated $\text{EELV}_{\text{MBNW}}$ measurements (i.e., two consecutive nitrogen washin–washout) at each PEEP level were carried out in five pigs of the study to assess repeatability of the technique.

An estimation of the recruited volume between two consecutive PEEP levels after lung injury was performed with the MBNW technique ($\text{Vrec}_{\text{MBNW}}$) as follows [3]:

$$\text{Vrec}_{\text{MBNW}} = (\text{EELV}_{\text{PEEP}_{n+1}} - \text{EELV}_{\text{PEEP}_n}) - \text{Compliance}_{\text{LowerPEEP}} \bullet (\text{PEEP}_{n+1} - \text{PEEP}_n),$$

with n = rank of measurement and EELV assessed with MBNW technique.

CT measurements

EELV_{CT} was measured on a Biograph mCT (Siemens, Munich, GERMANY) previously calibrated with the manufacturer phantom, using the following settings: voltage 120 kV, pitch 1.2, current time-product 80 mAs, field-of-view diameter 500 mm, slice thickness 1 mm, Kernel B31f (smooth), matrix size 512 by 512, and pixel size 0.8 mm by 0.8 mm.

Lung scanning was performed from apex to base during both end-expiratory and end-inspiratory pauses, with clamped tracheal tubes. The lungs were semi-automatically segmented, followed by visual inspection by one author (JCR) with a previously validated in-house software [10], excluding pleural effusions, hilar and mediastinal structures. Segmented lung volumes were analyzed using MATLAB (MathWorks, Natick, MA, USA).

Gas and tissue volume in each lung voxel and end-expiratory lung volume (EELV_{CT}) were computed from the CT number (HU) according to the following formulae [11]:

Voxel gas volume = $\frac{\text{CTnumber}}{-1000} \bullet \text{voxelvolume}$ for lung voxels with $-1000 \leq \text{CT number} \leq 0$.

Voxel gas volume = 0 for lung voxels with CT number > 0 .

Voxel gas volume = voxel volume for lung voxels with CT number < -1000 .

Voxel tissue volume = voxel volume – voxel gas volume.

$\text{EELV}_{\text{CT}} = \sum_{i=1}^k \text{voxelgasvolume}(i)$ with k = total number of voxels in segmented lung volume at end-expiration.

Lung parenchyma was then classified into four compartments, according to CT number [11]: non-inflated (density between +100 and –100 Hounsfield units (HU)), poorly inflated (density between –101 and –500 HU), normally inflated (density between –501 and –900 HU), and overinflated tissue (density between –901 and –1000 HU). The volume of each compartment was measured at end-expiration and end-inspiration.

CT-derived recruited volume ($V_{rec_{CT}}$ in mL) between two consecutive PEEP levels after lung injury was estimated on end-expiratory CT scans as follows [12]:

$$V_{rec_{CT}} = \frac{V_{tis_{PEEP_n}} - V_{tis_{PEEP_{n+1}}}}{1 - F_{gas_{PEEP_n}}} \bullet F_{gas_{PEEP_n}},$$

with n = rank of measurement, $V_{tis_{PEEP_n}}$ = tissue volume of the non-aerated compartment at PEEP_{*n*} at end-expiration, $V_{tis_{PEEP_{n+1}}}$ = tissue volume of the non-aerated compartment at PEEP_{*n+1*} at end-expiration, $F_{gas_{PEEP_n}}$ = gas fraction of the aerated compartment at PEEP_{*n*} at end-expiration.

Statistical analysis

Statistical analyses were performed using R for Windows v4.0.2 [13] with the following packages: boot [14, 15], lme4 [16], lmerTest [17], MuMIn [18], MethComp [19], and multcomp [20]. A p-value below 0.05 was chosen for statistical significance. Values were expressed as median [1st quartile–3rd quartile] unless otherwise stated. 95% confidence intervals (CI_{95%}) of proportions were computed with the Wilson score method.

Ventilatory, respiratory, hemodynamic and CT variables were compared across experimental conditions with a linear mixed model, using experimental condition as a factor with fixed effect, and pigs as a factor with random effect to consider multiple measurements. Multiple comparisons were performed with Dunnett’s test using baseline measurements as a reference.

Repeatability of the MBNW technique was assessed as follows. The coefficient of variation (CV) of EELV_{MBNW} repeated measurements was computed as the standard deviation divided by the mean of the 2 replicates. The precision of replicates for EELV_{MBNW} measurement was computed as 2 × CV. The least significant change (LSC) between repeated measurements was computed as CV × 1.96 × √2 [21]. The bias corrected and accelerated bootstrap method with 10,000 replicates was used to compute CI_{95%} for precision and LSC [22].

EELV_{MBNW} and EELV_{CT} were compared using a linear mixed-effect model, and Bland and Altman representation, using the first EELV_{MBNW} replicate [23]. Bias between methods was computed as EELV_{MBNW} minus EELV_{CT}. Since bias was non-constant and the experimental design involved repeated measurements, limits of agreement and conversion equation between EELV_{MBNW} and EELV_{CT} were computed from the posterior medians of a Monte Carlo chain simulation [24]. Percentage error was computed as 2 × SD_{Bias}/mean EELV_{CT} [25], with SD_{Bias} = standard deviation of the bias. Similar analyses were performed with Vrec_{MBNW} and Vrec_{CT}.

To explore the impact of potential confounding variables on the bias between methods, a linear mixed-effects model was built using ventilatory, respiratory, hemodynamic and CT variables as fixed effects, pigs as factor with random effect [26], and bias between methods as the dependent variable. Variables with *p* values < 0.1 in univariate analyses were considered for inclusion in a multivariable model. Model simplification was performed using a backward stepwise algorithm.

Ability of the MBNW technique to track directional changes in EELV between experimental conditions was assessed using four-quadrant and polar plots. Concordance rate was defined as the percentage of data points falling into one of the two quadrants of agreement on the four-quadrants plot (i.e., quadrants in which both $EELV_{MBNW}$ and $EELV_{CT}$ have the same directional changes) [27]. Since the four-quadrant plot does not quantify the distance between each data point and the line of identity, a polar plot analysis was performed [27]. Angular bias (difference in calibration between the reference and test method) was computed as the mean angle between all data points and polar axis [27], and compared to zero using Mann–Whitney U test. Radial limits of agreement (defined as the radial sector containing 95% of the data points, after conversion of negative deflections to positive ones) were computed similarly to the limits of agreement in Bland and Altman analysis [27].

Results

Ten measurements were missing due to various reasons, including premature death of one animal as reported in Additional file 1 and ending up in 32 available measurements. Variables repeatedly studied over time are reported in Table 1. After experimental lung injury, PaO_2/FiO_2 , pH and EELV decreased significantly while $PaCO_2$, plateau pressure and mean pulmonary artery pressure increased significantly. $EELV_{CT}$ values spanned from 254 to 1200 mL, while $EELV_{MBNW}$ ranged from 298 to 1157 mL.

Repeatability of the MBNW technique

27 replicate $EELV_{MBNW}$ measurements were obtained. The median precision of repeated $EELV_{MBNW}$ measurements was 6% [$CI_{95\%}$: 3–7%]. The median LSC between repeated $EELV_{MBNW}$ measurements was 8% [$CI_{95\%}$: 4–10%].

Comparison of $EELV_{MBNW}$ and $EELV_{CT}$

$EELV_{MBNW}$ and $EELV_{CT}$ were significantly correlated (marginal $R^2 = 0.97$, $p < 0.001$, Fig. 2A), and the regression equation between $EELV_{MBNW}$ and $EELV_{CT}$, had an intercept of 36 mL ($p < 0.001$) and a slope of 0.92. A non-constant bias between methods was identified on Bland and Altman plot (Fig. 2B), with a slight albeit significant trend of more negative values as EELV increased with PEEP. The conversion equation between $EELV_{MBNW}$ and $EELV_{CT}$ was computed as $EELV_{MBNW} = 0.92 \times EELV_{CT} + 36$. The 95% prediction interval of the bias between methods amounted to ± 86 mL, and the percentage error between both methods was 13.7%.

Multivariate analysis identified the volume of the overinflated compartment at end-expiration as the only variable independently associated with the bias between methods (Table 2), with a linear coefficient amounting to -8.7 ± 1.8 (i.e., for each mL of overinflated volume increase, the $EELV_{MBNW}$ is 9 ml lower than the reference EELV value measured with CT).

Table 1 Variables repeatedly assessed over time

Variables	Baseline	Post-injury	PEEP#1	PEEP#2	PEEP#3	PEEP#4	<i>p</i> value
PEEP (cm H ₂ O)	5 [5–5]	5 [5–5]	16 [16–16] [†]	12 [12–12] [†]	6 [6–8] [†]	3 [2–5] [†]	< 0.001
VT (ml kg ⁻¹)	10.0 [10.0–10.0]	10.0 [10.0–10.0]	5.9 [5.9–6.1] [†]	5.9 [5.9–6.1] [†]	5.9 [5.9–6.1] [†]	5.9 [5.9–6.0] [†]	< 0.001
PEEP _{tot} (cm H ₂ O)	5 [5–5]	6 [5–6]	17 [16–17] [†]	13 [12–13] [†]	7 [6–8] [†]	4 [3–6]	< 0.001
PEEP _{int} (cm H ₂ O)	0 [0–0]	0 [0–1]	1 [0–1] [†]	1 [0–1] [†]	0 [0–1]	1 [1–1] [†]	< 0.01
P _{plat} (cm H ₂ O)	16 [16–18]	29 [26–30] [†]	28 [26–29] [†]	22 [22–23] [†]	22 [19–22] [†]	24 [21–25] [†]	< 0.001
RR (min ⁻¹)	15 [15–16]	19 [15–24]	25 [25–28] [†]	25 [25–28] [†]	28 [25–28] [†]	28 [27–29] [†]	< 0.001
HR (min ⁻¹)	97 [86–118]	113 [108–123]	105 [100–107]	102 [97–133]	128 [109–130]	128 [117–142]	0.11
MAP (mm Hg)	84 [80–92]	92 [87–97]	95 [89–95]	86 [85–102]	104 [101–109] [†]	106 [104–108] [†]	< 0.05
MPAP (mm Hg)	24 [23–27]	32 [31–37] [†]	35 [33–39] [†]	34 [31–38] [†]	40 [37–47] [†]	41 [34–46] [†]	< 0.01
PAOP (mm Hg)	11 [11–11]	13 [12–14] [†]	14 [14–14] [†]	13 [13–13] [†]	11 [11–12]	10 [10–10]	< 0.001
CVP (mm Hg)	10 [10–12]	10 [9–14]	13 [12–13] [†]	12 [11–12]	10 [10–10]	9 [8–9]	< 0.001
CO (L.min ⁻¹)	3.5 [3.4–4.2]	4.4 [3.8–4.5]	2.9 [2.5–3.1] [†]	3.2 [2.8–3.3]	4.1 [3.6–4.4]	4.6 [4.3–4.8]	< 0.001
pH	7.45 [7.40–7.48]	7.28 [7.25–7.38] [†]	7.32 [7.29–7.34] [†]	7.34 [7.32–7.36] [†]	7.33 [7.28–7.36] [†]	7.25 [7.25–7.31] [†]	< 0.001
PaO ₂ /FiO ₂ (mm Hg)	435 [378–466]	92 [80–96] [†]	341 [335–342]	370 [342–398]	131 [85–237] [†]	83 [66–96] [†]	< 0.001
PaCO ₂ (mm Hg)	42 [37–47]	59 [44–61] [†]	53 [51–56] [†]	52 [52–53] [†]	60 [53–61] [†]	68 [58–74] [†]	< 0.001
Lactate (mmol/L ⁻¹)	1.2 [1.0–1.4]	0.9 [0.8–1.1]	0.7 [0.6–1.0]	0.7 [0.5–0.9]	0.6 [0.5–0.7]	0.6 [0.5–0.8]	0.15
EELV _{MBNW} (mL)	610 [579–656]	392 [357–455] [†]	987 [901–1059] [†]	850 [831–872] [†]	466 [427–553]	363 [348–428] [†]	< 0.001
EELV _{CT} (mL)	604 [533–615]	353 [324–407] [†]	1019 [988–1122] [†]	846 [824–888] [†]	470 [463–472] [†]	323 [295–420] [†]	< 0.001

CO = cardiac output; CVP central venous pressure; EELV_{CT} end-expiratory lung volume assessed with computed tomography; EELV_{MBNW} end-expiratory lung volume assessed with the multiple breaths N₂ washin–washout technique; HR heart rate; MAP mean arterial pressure; MPAP mean pulmonary artery pressure; PaCO₂ arterial carbon dioxide partial pressure; PaO₂ arterial oxygen partial pressure; PAOP pulmonary arterial occlusion pressure; PEEP positive end-expiratory pressure; PEEP_{int} intrinsic PEEP (i.e., PEEP_{tot}–PEEP); PEEP_{tot} total PEEP; P_{plat} plateau pressure; RR respiratory rate; VT tidal volume

Values are median [1st quartile–3rd quartile]. [†]*p* < 0.05 vs baseline

Ability of the MBNW technique to track directional changes in EELV

EELV_{MBNW} adequately tracked EELV_{CT} changes over time (Fig. 3), with a concordance rate amounting to 100% [CI_{95%}: 87%–100%]. Results of the polar plot analysis are reported in Fig. 4. The angular bias amounted to $-2^\circ \pm 10^\circ$ and was not statistically different from 0 (*p* = 0.21). Radial limits of agreement amounted to $\pm 26^\circ$.

Comparison of recruited volume assessed by the MBNW technique and CT

Vrec_{MBNW} and Vrec_{CT} were significantly correlated (marginal *R*² = 0.92, *p* < 0.001, Fig. 5A), and the regression equation between EELV_{MBNW} and EELV_{CT}, had an





Table 2 Uni- and multivariate analysis of variables associated with the bias between methods

Variables	Univariate coefficient	Univariate <i>p</i> value	Multivariate coefficient	Multivariate <i>p</i> value
PEEP (per 1 cm H ₂ O increase)	- 7.2 ± 1.5	< 0.0001	-	-
VT (per 10 ml increase)	3.1 ± 1.4	< 0.05	-	-
PEEP _{TOT} (per 1 cm H ₂ O increase)	- 7.0 ± 1.4	< 0.0001	-	-
P _{PLAT} (per 1 cm H ₂ O increase)	- 2.1 ± 1.9	< 0.05	-	-
Δ <i>P</i> (per 1 cm H ₂ O increase)	4.5 ± 1.7	< 0.01	-	-
RR (per 1 cycle.min ⁻¹ increase)	- 3.5 ± 1.5	< 0.05	-	-
CO (per 1 L.min ⁻¹ increase)	36.7 ± 9.8	< 0.05	-	-
Overinf. cpt at expiration (per 1 mL increase)	- 8.5 ± 1.7	< 0.0001	- 8.7 ± 1.8	< 0.0001
Overinf. cpt at inspiration (per 1 mL increase)	- 3.8 ± 1.1	< 0.01	-	-
Non-inf. cpt at expiration (per 1 mL increase)	0.12 ± 0.05	< 0.05	-	-

Bias between methods was computed as $EELV_{MBNW}$ minus $EELV_{CT}$. PEEP and Δ*P* were not included in the multivariate model for collinearity with PEEP_{TOT}

CO cardiac output; Δ*P* driving pressure $EELV_{CT}$ end-expiratory lung volume assessed with computed tomography; $EELV_{MBNW}$ end-expiratory lung volume assessed with the multiple breaths N₂ washin–washout technique; non-inf. cpt non-inflated compartment on CT; overinf. cpt. overinflated compartment on CT; PEEP positive end-expiratory pressure; PEEP_{TOT} total positive end-expiratory pressure; P_{PLAT} plateau pressure; RR respiratory rate; VT tidal volume

intercept of 0 mL (*p* < 0.001) and a slope of 1.43. A non-constant bias between methods was identified on Bland and Altman plot, slightly increasing toward more positive values as Vrec increased (Fig. 5B). The 95% prediction interval of the bias between methods amounted to ± 147 mL.




Discussion

The main findings of the study are the following: (1) the new MBNW monitoring tool is reliable to assess EELV at bedside, with statistically significant trending ability; (2) the precision of the technique is good, with a percentage error against CT amounting to 13.7%, and a LSC between repeated measurements $\leq 10\%$; (3) estimation of alveolar recruitment from $EELV_{MBNW}$ and measured with CT are highly correlated.

Reliability and trending ability of $EELV_{MBNW}$

As previously reported, EELV assessed by the MBNW technique exhibited a systematic bias, as compared to $EELV_{CT}$ [6], with a slight underestimation at high values of EELV. Although considered as a “gold standard” for lung volume assessment, it is important to highlight that CT scans accurately measures the “anatomical lung volume”, in contrast with MBNW technique, or other gas-dilution technique, where the lung volume measured contributes to ventilation, i.e., “ventilable lung volume” [4]. Thus, in the presence of overinflated lung areas (at high PEEP levels) the “anatomical lung volume” is expected to be different from the “ventilable lung volume” [7, 28]. The current study supports previous results, since the bias observed between the tested methods was independently related to the amount of overinflation at end-expiration, it may be hypothesized that overinflated lung areas did not participate to nitrogen mixing during washin–washout, which may explain EELV underestimation at high PEEP. From the clinical standpoint, the bias observed between techniques in the current study is likely to be of minor relevance if we consider the potential clinical application of the proposed novel technology and the drawbacks of the CT technique.

Nevertheless, the upper limit of the $CI_{95\%}$ of the LSC between repeated measurements amounted to 10%, suggesting a good ability to detect relatively small changes in EELV. This was confirmed by a 100% concordance rate and an angular bias not significantly different from 0, suggesting a good trending ability of the new technique. The percentage error of this new technique was the lowest of previously published experimental and



clinical studies (ranging from 28 to 57% [4–6]). This is presumably related to the use of the fast mainstream sensors, which ensured that the measurements of proximal gas concentrations and gas flow are precisely aligned in time [29]. Signal alignment may be more challenging when a side stream sensor is used, as in a previous study using the same ARDS animal model [6]. One might argue that despite considerable improvement on $EELV_{MBNW}$ accuracy provided by this novel technology, lower PEEP levels were used in the current study. However, the experimental protocol presented here is closer to the clinical setting in terms of applied PEEP levels, improving its extrinsic validity.

Reliability of estimated alveolar recruitment by the MBNW technique

To our knowledge, the present study is the first to provide concomitant measurements of $Vrec_{MBNW}$ with CT as the reference technique. As proposed by Dellamonica et al. [3], $Vrec_{MBNW}$ is computed from the difference between change in EELV induced by PEEP, and




Fig. 5 Correlation between $Vrec_{MBNW}$ and $Vrec_{CT}$ (A) and corresponding Bland and Altman plot (B). Each symbol represents a concomitant measurement of alveolar recruitment assessed with the multi-breath nitrogen washin–washout technique ($Vrec_{MBNW}$) and computed tomography ($Vrec_{CT}$). Negative recruited volumes denote alveolar derecruitment. **A:** Solid line is regression line. Dashed line is line of identity. **B:** continuous line and the 2 broken lines are the mean bias and 95% prediction interval limits of the bias between methods, respectively. 95% p.i. = 95% prediction interval of the bias between methods

minimally predicted increase in lung volume between PEEP levels (i.e., the product of respiratory system compliance assessed at lower PEEP and PEEP level change). When the change in EELV is larger than this minimum predicted volume gain, the difference is considered as an estimate of alveolar recruitment. However, this method assumes that compliance is linear over the change in pressure induced by PEEP; an assumption we are unable to confirm with our data. Furthermore, this method combines measurement errors from several devices (ventilator pneumotachograph, ventilator pressure sensors, MBNW technique), questioning the ability of this technique to detect small recruited volumes. Nevertheless, a fair correlation was identified between $Vrec_{MBNW}$ and alveolar recruitment computed from the pressure–volume curves in ARDS patients [3], but mathematical coupling between both techniques ($Vrec_{MBNW}$ and the pressure volume curves are computed using the ventilator pneumotachograph) could question the validity of this study. Of note, 4 subjects with the highest PEEP levels (16 cm H_2O) were excluded of Dellamonica et al. study as their $EELV_{MBNW}$ was underestimated [3], a finding in line with our results. Nevertheless, we observed in the present study a strong correlation with CT measurements, although $Vrec_{MBNW}$ was overestimated at high recruited volume values (a consequence of the underestimation of $EELV_{MBNW}$). Whether this technique could reliably detect low and high recruiters by PEEP remains to be determined.

Clinical perspective

As compared to other MBNW technologies [4–7], the current monitoring tool exhibited substantially lower percentage error (28% to 57% vs 13.7%) and hence higher precision, and narrower radial limits of agreements ($\pm 51^\circ$ [6] vs. $\pm 26^\circ$), i.e., higher trending ability. Furthermore, the current monitoring tool is virtually unbiased, provided the conversion equation presented above is used to account for the non-constant bias [30]. Although we did not perform a direct comparison, we hypothesize that an improved accuracy could result from fast mainstream gas concentration measurements able to better

accommodate the heterogeneities in the distribution of the ventilation than a device based only on end-tidal gas composition. The availability of a technology providing accurate bedside assessment and trending ability of end-expiratory lung volume allows foreseeing the regular assessment of a recruitment maneuver or prone positioning. Likewise direct assessment of the effects on EELV of a PEEP change should make PEEP optimization part of systematic daily practice. Furthermore, measurement of EELV allows computation of lung strain (the lung deformation related to its original status) and lung stress (using a presumed value for specific elastance), both being important parameters to evaluate the risk of ventilator-induced lung injury. This should help implement a personalized lung protective mechanical ventilation.

Study limits

Some limitations of the present study should, however, be acknowledged. First, the experimental ARDS model used is highly recruitable by PEEP, and external validity in poorly recruitable ARDS may be questionable. Second, the reliability of the technique was not tested at VT lower than 6 ml.kg⁻¹, at FiO₂ higher than 80% and higher respiratory rates (*i.e.*, situations in which the MBNW technique may be less reliable [6]). Furthermore, the present study was performed in deeply sedated animals under neuromuscular blocking agents and controlled ventilation, and the reliability of the present MBNW technique during mechanical ventilation with assisted ventilation modes remains unknown. Third, lung segmentation on CT excluded the trachea and the main bronchus (*i.e.*, structures involved in EELV computed by the MNBW technique), which could explain a slight underestimation of EELV_{CT}, although this would be a systematic error as these structures are cartilaginous and are not expected to increase their volumes at increasing PEEP levels.

Conclusions

We report a novel technique that reliably assesses EELV at bedside, and might add valuable information to further personalization of mechanical ventilation. This novel system designed to assess EELV shows high measurement precision and with excellent trending ability, compared to EELV estimated using CT, as the reference technique. Finally, estimation of PEEP-induced alveolar recruitment using this novel system appears reliable in the specific setting of a highly recruitable experimental ARDS model.

Abbreviations

ARDS: Acute respiratory distress syndrome; CI_{95%}: 95% Confidence interval; CT: Computed tomography; CV: Coefficient of variation; EELV: End-expiratory lung volume; FiO₂: Inspired fraction of oxygen; LSC: Least significant change; MBNW: Multiple-breaths nitrogen washin-washout; PEEP: Positive end-expiratory pressure; RR: Respiratory rate; VCO₂: Carbon dioxide production; VO₂: Oxygen consumption; Vrec_{CT}: Recruited alveolar volume assessed with CT; Vrec_{MBNW}: Recruited alveolar volume assessed with the MBNW technique; VT: Tidal volume.

Supplementary Information

The online version contains supplementary material available at <https://doi.org/10.1186/s40635-021-00410-x>.

Additional file 1. Available and missing measurements.

Acknowledgements

None.

Authors' contributions

LB made substantial contributions to study design, to data acquisition, study analysis, and interpretation of data, and drafted the manuscript, and approved the version to be published. NCC made substantial contributions to study design, to data acquisition, and interpretation of data, and revised the manuscript for important intellectual content, and approved the version to be published. SR made substantial contributions to study design, to data acquisition, and interpretation of data, and revised the manuscript for important intellectual content, AND approved the version to be published. CS made substantial contributions to study design, to data acquisition, and interpretation of data, and revised the manuscript for important intellectual content, and approved the version to be published. DN made substantial contributions to study design, to data acquisition, and interpretation of data, and revised the manuscript for important intellectual content, and approved the version to be published. MO made substantial contributions to study analysis, and interpretation of data, and revised the manuscript for important intellectual content, and approved the version to be published. EDS made substantial contributions to study analysis, and interpretation of data, and revised the manuscript for important intellectual content, and approved the version to be published. JPR made substantial contributions to study design, to data acquisition, and interpretation of data, and revised the manuscript for important intellectual content, and approved the version to be published. JCR made substantial contributions to study design, to data acquisition, study analysis, and interpretation of data, and drafted the manuscript, and approved the version to be published and agreed to be accountable for all aspects of the work in ensuring that questions related to the accuracy or integrity of any part of the work are appropriately investigated and resolved. All authors read and approved the final manuscript.

Funding

The study was funded by Hamilton Medical AG. The funding body was involved in the design of the study, during data acquisition, and reviewed the manuscript, but was not involved in data analysis. LB, MO, EDS, and JCR are not employees of Hamilton Medical AG, and did not receive reimbursements, fees, or salary by the company.

Availability of data and material

The datasets used and/or analyzed during the current study are available from the corresponding author on reasonable request.

Declarations

Ethics approval and consent to participate

The study was approved by an Institutional Review Board for the care of animal subjects (Comité d'éthique en expérimentation animale n°042, APAFIS#21542–2019071909353561),

Consent for publication

Not applicable.

Competing interests

LB declares that he has no competing interests. MO declares that he has no competing interests. EDS declares that he has no competing interests. JCR declares that he has no competing interests. Drs. Carvalho, Reidt, Schranz, Revely and Novotni disclosed they are full-time employees at the Research & New Technology Department at Hamilton Medical AG, Switzerland.

Author details

¹Service de Médecine Intensive - Réanimation, Hôpital De La Croix Rousse, Hospices Civils de Lyon, 103 Grande Rue de la Croix Rousse, 69004 Lyon, France. ²Université de Lyon, Université Claude Bernard Lyon 1, INSA-Lyon, UJM-Saint Etienne, CNRS, Inserm, CREATIS UMR 5220, U1206, Villeurbanne, France. ³Research and New Technology Department, Hamilton Medical AG, Bonaduz, Switzerland.

Received: 10 May 2021 Accepted: 10 August 2021

Published online: 10 September 2021

References

1. Bellani G, Laffey JG, Pham T et al (2016) Epidemiology, patterns of care, and mortality for patients with acute respiratory distress syndrome in intensive care units in 50 countries. *JAMA* 315:788–800. <https://doi.org/10.1001/jama.2016.0291>
2. Bos LDJ, Artigas A, Constantin J-M et al (2021) Precision medicine in acute respiratory distress syndrome: workshop report and recommendations for future research. *Eur Respir Rev*. <https://doi.org/10.1183/16000617.0317-2020>
3. Dellamonica J, Lerolle N, Sargentini C et al (2011) PEEP-induced changes in lung volume in acute respiratory distress syndrome. Two methods to estimate alveolar recruitment. *Intensive Care Med* 37:1595–1604. <https://doi.org/10.1007/s00134-011-2333-y>
4. Chiumello D, Cressoni M, Chierichetti M et al (2008) Nitrogen washout/washin, helium dilution and computed tomography in the assessment of end expiratory lung volume. *Crit Care* 12:R150. <https://doi.org/10.1186/cc7139>
5. Graf J, Santos A, Dries D et al (2010) Agreement between functional residual capacity estimated via automated gas dilution versus via computed tomography in a pleural effusion model. *Respir Care* 55:1464–1468
6. Richard JC, Pouzot C, Morales Pinzon A et al (2014) Reliability of the nitrogen washin-washout technique to assess end-expiratory lung volume at variable PEEP and tidal volumes. *Intensive Care Med* Exp 2:10. <https://doi.org/10.1186/2197-425X-2-10>

7. Tang R, Huang Y, Chen Q et al (2012) The effect of alveolar dead space on the measurement of end-expiratory lung volume by modified nitrogen wash-out/wash-in in lavage-induced lung injury. *Respir Care* 57:2074–2081. <https://doi.org/10.4187/respcare.01800>
8. Food and Drugs, Chapter 1, Subchapter A, Part 58—Good Laboratory Practice for Nonclinical Laboratory Studies. US Government Publishing Office; Electronic Code of Federal Regulations (current updates). <http://www.ecfr.gov>
9. Olegard C, Sondergaard S, Houltz E et al (2005) Estimation of functional residual capacity at the bedside using standard monitoring equipment: a modified nitrogen washout/washin technique requiring a small change of the inspired oxygen fraction. *Anesth Analg* 101:206–212. <https://doi.org/10.1213/01.ANE.0000165823.90368.55>
10. Dávila Serrano E, Dhelft F, Bitker L, et al (2020) Software for CT-image Analysis to Assist the Choice of Mechanical-Ventilation Settings in Acute Respiratory Distress Syndrome. In: International Conference on Computer Vision and Graphics. ICCVG 2020. Lecture Notes in Computer Science. Springer, Warsaw, Poland, pp 48–58
11. Gattinoni L, Pesenti A, Bombino M et al (1988) Relationships between lung computed tomographic density, gas exchange, and PEEP in acute respiratory failure. *Anesthesiology* 69:824–832. <https://doi.org/10.1097/0000542-198812000-00005>
12. da Paula LFS, C, Wellman TJ, Winkler T, et al (2016) Regional Tidal Lung Strain in Mechanically Ventilated Normal Lungs. *J Appl Physiol* 121:1335–1347. <https://doi.org/10.1152/jappphysiol.00861.2015>
13. R Core Team (2020) R: A language and environment for statistical computing. R Foundation for statistical Computing, Vienna, Austria
14. Davison AC, Hinkley DV (1997) *Bootstrap Methods and Their Applications*. Cambridge University Press, Cambridge
15. Canty A, Ripley B (2020) *Boot: Bootstrap R (S-Plus) Functions*
16. Bates D, Maechler M, Bolker B, Walker S (2015) Fitting linear mixed-effects models using lme4. *J Stat Softw* 67:1–48. <https://doi.org/10.18637/jss.v067.i01>
17. Kuznetsova A, Brockhoff PB, Christensen RHB (2017) lmerTest package: tests in linear mixed effects models. *J Stat Softw* 82:1–26. <https://doi.org/10.18637/jss.v082.i13>
18. Barton K (2020) MuMin: Multi-Model Inference
19. Carstensen B, Gurrin L, Ekstrom CT, Figurski M (2020) MethComp: Functions for analysis of agreement in method comparison studies
20. Hothorn T, Bretz F, Westfall P (2008) Simultaneous inference in general parametric models. *Biom J* 50:346–363. <https://doi.org/10.1002/bimj.200810425>
21. Cecconi M, Rhodes A, Poloniecki J et al (2009) Bench-to-bedside review: The importance of the precision of the reference technique in method comparison studies - with specific reference to the measurement of cardiac output. *Crit Care* 13:201. <https://doi.org/10.1186/cc7129>
22. Efron B (1987) Better bootstrap confidence intervals. *J Am Stat Assoc* 82:171–185. <https://doi.org/10.2307/2289144>
23. Bland JM, Altman DG (2007) Agreement between methods of measurement with multiple observations per individual. *J Biopharm Stat* 17:571–582. <https://doi.org/10.1080/10543400701329422>
24. Carstensen B (2011) *Comparing Clinical Measurement Methods: A Practical Guide (Statistics in Practice)*, 2nd edn. Wiley, Chichester
25. Critchley LA, Critchley JA (1999) A meta-analysis of studies using bias and precision statistics to compare cardiac output measurement techniques. *J Clin Monit Comput* 15:85–91. <https://doi.org/10.1023/a:1009982611386>
26. Carstensen B, Simpson J, Gurrin LC (2008) Statistical models for assessing agreement in method comparison studies with replicate measurements. *Int J Biostat* 4:1–26. <https://doi.org/10.2202/1557-4679.1107>
27. Critchley LA, Yang XX, Lee A (2011) Assessment of trending ability of cardiac output monitors by polar plot methodology. *J Cardiothorac Vasc Anesth* 25:536–546. <https://doi.org/10.1053/jjvca.2011.01.003>
28. Tran MC, Crockett DC, Cronin JN et al (2021) Bedside monitoring of lung volume available for gas exchange. *ICMx* 9:3. <https://doi.org/10.1186/s40635-020-00364-6>
29. Yang J, Chen B, Burk K et al (2016) A mainstream monitoring system for respiratory CO₂ concentration and gasflow. *J Clin Monit Comput* 30:467–473. <https://doi.org/10.1007/s10877-015-9739-y>
30. Bland JM, Altman DG (1986) Statistical methods for assessing agreement between two methods of clinical measurement. *Lancet* 327:307–310

Publisher's Note

Springer Nature remains neutral with regard to jurisdictional claims in published maps and institutional affiliations.