



HAL
open science

Review on Needle Insertion Haptic Simulation

Benjamin Delbos, Rémi Chalard, Richard Moreau, Minh Tu Pham, Arnaud Lelevé

► **To cite this version:**

Benjamin Delbos, Rémi Chalard, Richard Moreau, Minh Tu Pham, Arnaud Lelevé. Review on Needle Insertion Haptic Simulation. *Current Robotics Reports*, 2022, Medical and Surgical Robotics, 10.1007/s43154-022-00093-6 . hal-03780282

HAL Id: hal-03780282

<https://hal.science/hal-03780282>

Submitted on 19 Sep 2022

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Review on Needle Insertion Haptic Simulation

Benjamin Delbos¹, Rémi Chalard^{1†}, Richard Moreau^{1†}, Minh Tu Pham^{1†} and Arnaud Lelevé^{1*†}

^{1*}Univ Lyon, INSA Lyon, Université Claude Bernard Lyon 1, Ecole Centrale de Lyon, CNRS, Ampère, UMR5005, 69621 Villeurbanne, France.

*Corresponding author(s). E-mail(s): arnaud.leleve@insa-lyon.fr;
Contributing authors: benjamin.delbos@insa-lyon.fr;
remi.chalard@insa-lyon.fr; richard.moreau@insa-lyon.fr;
minh-tu.pham@insa-lyon.fr;

†These authors contributed equally to this work.

Abstract

a) Purpose of review: This short review updates an exhaustive one written by Correa et al in 2019 about haptic training simulation on needle insertion in the medical field.
b) Recent findings: Latest works refine well-known models and enhance setups and methods to facilitate generically getting experimental data.
c) Summary: We provide a complementary focus on device specifications and recent models to render this specific haptic feedback on Computer-Based Simulators. Assessment approaches and the issues encountered when introducing such simulators into curricula are also discussed. FEM-based approaches still do not permit real-time computation but hybrid approaches as proposed by [1] may become a good compromise. Nonetheless, psychophysical studies should be performed to determine the haptic fidelity of the various approaches found in the literature, and embed them efficiently in medical curricula. This would permit to delay the necessary final hands-on training on patients that raises ethical issues.

Keywords: Haptic Training, Computer Simulation, Gesture Training, Needle Insertion

Introduction

Many medical procedures (blood sampling, biopsy, puncture, catheterization ... in anesthesia, brachytherapy, neurosurgery, ...) require needle insertion but this common and important gesture differs a lot according to the goal, the concerned areas of the body, and the visibility in the area. By nature, the part of the needle already inside the body is not directly visible, which makes this gesture performed almost blindly. Practitioners then require another source of information to determine if the tip of the needle has reached the target location, knowing that the needle may deflect from its initial trajectory (notably for beveled ones), may cross various layers of anatomic tissues with different mechanical behaviors (skin, fat, tendons, nerves, ...) requiring mastered insertion forces and penetration velocities from the practitioner.

One important source is the haptic¹ feedback: the needle-patient body interaction forces felt by the practitioners in their hand(s) while inserting the needle. In this way, they feel whether the needle penetrates or slips around a blood vessel wall, or enters in contact with a bone, for instance. But in some cases, this force feedback is not sufficient, such as in epidural anesthesia or intraarticular injections. Complementary information must be provided to the practitioners to help them in their gestures. For instance, in the case of epidural anesthesia, a syringe filled with a neutral solution is mounted on the needle; the way it empties through the needle provides to the practitioner crucial information about the reach of the epidural area. In other cases (brachytherapy, intraarticular injection, some epidural anesthesia, ...) real-time medical imaging provides this complementary information. In all these cases, the practitioners must learn how to manipulate and coordinate these tools taking into account these sensations and complementary information, during a long apprenticeship. Some of them require much practice before being efficient. For instance, 90 epidural insertions are necessary to obtain an 80% success rate [2].

However, this training is, in general, performed first on manikins and next on real patients that may suffer from unsuccessful first attempts. This widespread ethical issue in the medical discipline has encouraged the use of more realistic simulators that could permit safely and efficiently acquire the technical skills and delay the necessary final training on patients [3]. Computer-Based Simulation (CBS) has been the first response to this general requirement. However, they lack the force feedback rendering dimension. Haptic training simulators (HTS) add this feedback with the help of haptic devices, raising the fidelity of CBS [4]. For instance, a recent review of haptic training for laparoscopy is proposed in [5]. Such simulators can render the aforementioned haptic feedback to help practitioners train themselves on these sensations as many times as necessary without any risk for patients.

This short review deals with hands-on training for needle insertion with haptic training simulators. It updates a more exhaustive one written by Correa et al in 2019 [6] with a complementary focus on devices and models used to

¹The word “haptic” comes from the Greek word “haptomai” (ἅπτομαι) which means “touch”, gathering kinaesthetic (force) and tactile senses.

render this haptic feedback, and assessment approaches recently developed to provide trainees an objective evaluation of their gestures and information on how to improve them. Therefore, the following section introduces the haptic devices that could be used for such a purpose, while Section 2 details the various models that permit the control of the aforementioned haptic devices to render realistic force feedback. Section 3 deals with gesture assessment.

1 Haptic Devices Used For Needle Insertion Simulators

Virtual reality simulators are most of the time insufficient because the haptic part is missing [7]. The haptic part can be passively reproduced by basic mannequins. Nevertheless, if they can be sufficient for tactile feedback, they are, in general, not realistic enough in terms of force feedback. Since the democratization of additive manufacturing, some researchers have been developing multi-material components to provide haptic feedback. For example, the combination of Thermoplastic Polyurethane (TPU) and Acrylonitrile Butadiene Styrene (ABS) can be used to make a phantom allowing needle insertion training [8]. Models based on gelatin can also be used [9]. However, these solutions require manufacturing new products to reproduce different behaviors. To solve this issue, the authors of [10] propose using a specific cartridge to reproduce the penetration of the needle into different layers.

However, using passive materials to provide feedback has one main issue: they will wear out over time and can be damaged by piercing. One solution is to use active haptic interfaces. These interfaces can offer configurable simulators without any damage to materials. They also permit embedding sensors to record data for gesture assessment purposes.

1.1 Commercial Haptic Devices

Usual haptic interfaces are based on electric actuators such as DC motors embedded in robots with serial (such as Touch by 3D System or Virtuoso 6D by Haption) or parallel (such as Falcon by Novint Technologies Inc. or Omega 6 by Force Dimensions) architectures. This not exhaustive list gathers the products usually found in the literature. Figure 1 includes photos of these interfaces. Their main difference lies in their numbers of Degrees of Freedom (DoF) and of Degrees of Force Feedback (DoFF), the maximum force they can produce, their workspace, and their cost. Table 1 gathers these characteristics.

1.2 Specifications For Needle Insertion

Needle insertion simulators require at least one DoFF to feel the axial tissue resistance during the penetration. It can be useful in terms of fidelity with real cases and pedagogical requirements also to enable the orientation of the needle around the insertion hole, which then requires 5 DoF (the rotation around the needle axis has usually no interest). It could be also interesting to reproduce

**Fig. 1:** Examples of electric haptic interface

Device	DoF	DoFF	Max Force [N]	Stiffness [N.mm]	Workspace [mm]	Cost [k€]
Touch	6	3	3	1 to 2.31	160 x 120 x 70	2
Virtuose 6D	6	6	35	3	1330 x 575 x 1020	90
Falcon	3	3	9	??	100 x 100 x 100	0.2
Omega 6	6	6	14.5	14.5	ϕ160 x 110	25

Table 1: Characteristics of most encountered electric haptic interfaces

the lateral tissue forces while the operator changes the orientation of the needle during the penetration. It is therefore recommended to reproduce 5 DoFF to get a realistic simulator.

1.3 Recent Samples Using Commercial Haptic Devices

To increase trainees' immersion into the simulation, it is advised to provide a mock needle on the interface. With the development of additive manufacturing, it is quite common to insert it on the haptic interface terminal tool. For instance, in [11], where authors used a Novint Falcon for their epidural simulator, they have developed a custom end effector to substitute the Novint Falcon one (See Fig. 2). It allows increasing the realism of the simulator and thus improve skills transfers to real-life situations.

In [12], authors used two Touch X interfaces (improved version of the Touch interface by 3D Systems) to reproduce forces during ophthalmic surgical procedures. They also developed a specific end effector to allow practitioners to use similar tools as in real procedures. To improve immersion, a virtual world is added to the simulator.

In [13], authors simulated a Central Venous Catheterization (CVC) with a virtual ultrasound probe featuring a 3D tracker and a Touch interface to simulate the CVC needle. The trainee handles the mock US probe with one hand and the needle with the other hand. The real-time position and orientation of the probe permit providing synchronized fake US images integrating the virtual needle when visible. Li *et al* enhanced this setup by simulating the fake probe with a second Touch interface to render probe-patient interaction forces (see Fig. 3) [14]. The originality of this study mainly concerns the methods




Fig. 2: Epidural simulator based on a Novint Falcon [11]

to compute in real-time the force to be reproduced by the haptic interfaces, taking into account the respiration of the virtual patient.



Fig. 3: Simulator for renal biopsy based on Touch interfaces [14]

1.4 Custom Haptic Interfaces

The aforementioned commercially available interfaces can be used for many medical applications. However, some procedures require to develop dedicated haptic interfaces to be more realistic.

In [15], the authors developed a custom interface based on electric actuators and a hexapod design (see Fig. 4). This original structure, which allows obtaining 6 DoFF, lies on the unbound tool handled by the trainees. They can thus move their tool without any constraints when they are outside the virtual body. They can also change their tool and once they touch the virtual object, the tip of the hexapod is linked to the tooltip.





Fig. 4: Simulator using a custom haptic device based on an electric actuated hexapod [15]

In epidural procedures, practitioners handle a needle mounted on a syringe. Epidural anesthesia is a blind procedure as practitioners cannot see through the human body and do not use any ultrasound probes on daily use. To bypass this issue, practitioners connect a syringe filled with a neutral solution and push on the piston while introducing the needle. The piston resistance provides them haptic information about the localization of the tip of the needle. This resistance quickly decreases as soon as the tip reaches the area of interest. In [16], the authors used a pneumatic cylinder coupled with an artificial needle mounted on a commercial haptic interface (Virtuose6D) (see Fig. 5). Using this simulator, trainees are provided with force feedback not only from the needle (through the electric haptic device) but also from the syringe (through the pneumatic cylinder).



(a) Global view showing both electric and pneumatic haptic devices
 (b) Pneumatic cylinder used to mimic a syringe

Fig. 5: PeriSIM: training haptic simulator for epidural procedures

1.5 Conclusion On Haptic Devices And Needle Insertion Simulators

Training simulators are becoming more and more popular to learn and assess medical gestures [17]. To be efficient in terms of hands-on training, configurable, repetitive, and providing objective assessment feedback, needle insertion simulators should feature an active haptic interface. These haptic interfaces allow producing forces to allow trainees to become familiar with real procedures. However, these forces need to be realistic. For that purpose, it is necessary to compute them using biomechanical models. Next section deals with this aspect.

2 Needle Insertion Simulation Models

Extensive work has focused on force modeling for needle insertion into soft tissues. Some comprehensive reviews have been written either considering haptic simulation applications [6, 18], or robot-assisted procedures [19]. In this review, we focus on works related to haptic simulations, either already in use or ought to be integrated into haptic training simulators in the coming years.

According to Azar *et al*, models used to render needle insertion forces in simulations belong to two categories: Deformation and Fracture Mechanics based models. Deformation-based models come from the observation of forces due to the penetration of the needle into the tissues without considering underlying physics. In the second category, the needle insertion is modeled as a crack that propagates with the help of an energetic approach [20].

2.1 Deformation Based Models

Forces exerted on the shaft of the needle during insertion into soft tissues are commonly considered as the sum of "cutting, sliding, stick-slip, tissue deformation, and displacement and peeling" [21].

In the early 2000s, Simone and Okamura [22] proposed a method to measure these different forces and to gather them into 3 components corresponding to a) cutting forces, applied on the tip of the needle, b) friction forces, applied along the shaft of the needle and c) stiffness forces, due to elasticity of the tissue before puncture, when the needle pushes against the organ causing visco-elastic deformations. They distinguished three phases with different force patterns, corresponding to the I/ prepuncture, II/ penetration, and IV/ extraction motions (in phase III, the needle is immobile). These models were obtained through *ex vivo* experiments with computation of values a posteriori, but later, Barbé *et al* performed online estimation on *in vivo* specimen [23]. Figure 6 describes these phases and the total axial force pattern due to the penetration of the needle. These models have still been used in recent works such as [7, 11, 13, 13, 14, 24–34]. Recently, works focused on the ability to render multiple layers of tissues, using piece-wise exponential models [13] or nested boxes [25] rendering each one its stiffness and cutting forces according to [35]. To




Fig. 6: Needle insertion into soft tissues, divided into four phases: I/ Pre-puncture II/ Puncture III/ Relaxation (stop) IV/ Extraction, [23]. The total axial force as a function of the penetration is provided on the plot inspired from [36].

benefit from the use of active haptic feedback, these works also proposed to render different patient morphologies by adapting the size of the tissue layers [13] or box depths [25]. Extrapolation of these models to other patient types was validated with experts.

To improve rendering of cutting forces, Daniel *et al.* proposed the "tracking wall" algorithm as an enhancement of traditional proxy-based algorithms [24]. It allows rendering constant cutting force without chatter and capture small rejections forces occurring when the needle stops inside tissues as stated by [36]. In recent works, lateral forces or clamping forces applied on the sidewalls of the needle were modeled using proportional approaches, mimicking the elasticity of the tissue [25, 26]. In some cases, elasticity was pondered by the depth of insertion, meaning that diverging from the insertion path would become harder as the needle penetrates the tissue. In these approaches, deflection of the needle was not considered.

However, all these models consider forces as a function of penetration depth only, without considering the velocity that affects the needle insertion into viscoelastic tissues [37, 38]. Therefore, Wu *et al.* proposed a non-linear model rendering the force feedback as a function of the needle's instantaneous position and velocity during the insertion phase into soft tissue. This model uses a piece-wise split into two areas separated by the moment when the tissue stops deforming while the needle proceeds. In the first area, an exponential-fitting of the penetration depth, and in the second one, the force is proportional to it. In both cases, velocity impacts the magnitude proportionally to the penetration depth. This model takes also the needle diameter as a parameter which

influence is demonstrated (doubling the needle diameter, nearly doubles the force magnitude). Unfortunately, the model parameter values for pork and beef livers were not provided. Sadeghnejad *et al.* proposed, on the same approach a 3-phase piece-wise model (tissue loading deformation, fracture (point separating pre-puncture and puncture phases) and cutting) taking into account the tissue fracture event and cutting forces as a function of penetration and velocity [28]. Note that they do not model the fracture physics, only consider the change of behavior at this point. Nonetheless, these both studies do not consider the retraction/withdrawal phase.

Note that, instead of stiffness-like functions, Castro *et al.* use non-holonomic constraints to render forces due to contacts. This approach will be integrated into surgical simulators in future works [39]. Table 2 sums up the recent approaches to render forces using deformation-based models.

Author	Paper	Year	Contact	Friction	Cutting	Clamping
Daniel	[24]	2020	Second-order polynomial [22]	Lugre model [40]	Tracking wall algorithm	Virtual fixture plasticity
Senac	[25]	2019	Second-order polynomial [22]	Not modeled	Constant with low-pass filter	Linear and angular virtual spring
DiVece	[26]	2021	Constant damping effect	Constant static and dynamic friction	Not modeled	Virtual spring
Wu	[27]	2019	Exponential and polynomial	Damping	Constant force (proportional to diameter)	Not modeled
Pepley	[13]	2018	Piece-wise exponential [41]			Not modeled
Sadeghnejad	[28]	2019	Kelvin-Voigt modified non linear model	Not modeled	Proportional displacement according to constant velocity	Not modeled
Pepley	[42]	2016	Piece-wise exponential [41]			Small bounding forces
Correa	[30]	2017	Hooke's law	Not modeled		Lateral forces prop to penetration
Li	[14]	2019	Fourier series			Not modeled
Barnouin	[31]	2020	2nd order polynomial [22]	Stick-slip [22]	Constant force	Stiffness K proportional to divergence and insertion depth.
Moo-Young	[11]	2021	Hooke's law	Not modeled		PID controller
Esterer	[32]	2020	Polynomial peaks	Linear up to a constant		Not modeled
El-Monajjed	[33]	2021	Piece-wise polynomial and exponential functions			Not modeled

Table 2: Overview of functions used to render forces due to needle insertion and its components (friction, cutting, contact and clamping forces) in deformation-based models from recent years

Despite numerous works and advancements on deformation-based Finite Element Modelling (FEM) of needle insertion, [6], FEM based simulations are still found to be too computationally demanding for haptic simulations. Recent research focuses on improving efficiency of FEM based simulations by avoiding remeshing during cutting [43]. To provide realistic force feedback behavior and overcome computing limitations of FEM based approaches (see next subsection) at the same time, Wittek *et al* use Meshless Total Lagrangian Explicit Dynamics to simulate tissue deformations. This model has the advantage of

requiring only patient-specific geometry and two parameters (being easy to identify from intra-operative images) to render patient-specific simulations [1]. As the model computes needle-tissue interaction forces, it may be used to compute real-time force feedback in the future but this still requires some validation.

2.2 Fracture-based Models

Two main approaches consider the modeling of the tissue fracture performed by the insertion of the needle: Energy-Based Modeling (EBM) and Finite-Element Modeling (FEM). This modeling should provide more precision around the fracture point as, according to [38] for porcine skin, 61% of the total insertion force comes from the fracture, 21% from the friction, and 18% from the tissue deformation. Note that these models take into account the needle velocity. However, as far as we know, no study compared the relative force rendering quality of this family of approaches in an end-user physiological study.

In EBM, the cutting is considered as the consequence of exchanges of energy between the needle and the tissue, causing crack propagation. Early works demonstrated the relevance of this approach [20, 37] but only *Barnett et al.* exploited this approach to predict insertion forces. Yet, this has not been embedded in a training simulator.

Extensive research on Finite Element Method (FEM) based needle insertion modeling has been done and synthesized in [6]. Recent progress is reported in [44] where the modeling of the tissue fracture enables realistic FEM simulations of needle trajectories and estimation of the interaction forces matching the experimental data for deep insertion cases. In [45], multilayer tissues are considered [45], but this kind of approach is reserved for design validation or preplanning purposes as such simulations last several hours: they are still computationally complex to provide real-time realistic force feedback in haptic simulators. To enable this kind of modeling for real-time simulation, Bui *et al.* proposed, in [46], a method to minimize the complexity of mesh generation and an algorithm named CutFEM solving equations with constraints twice faster as classical FEM on a liver model (around 450 ms per iteration). They validated it on an electrode implantation simulation in Deep Brain Stimulation. This computation velocity enables 2D or 3D rendering in simulations but is still a little slow to render directly force feedback in a haptic application where 1 kHz is the minimum sampling period to reproduce hard contacts.

Needle deflection during insertion is a function of the diameter and shape of the needle [45] and insertion forces [38]. Despite early works to investigate the force-needle deflection relationship [47], *Correa et al.* confirm that the "*deformation of [...] tissues or organs received more attention than the deflection of needles*" [6]. In recent FEM-based methods, needles were either considered rigid [44] or deformable, using for instance the Euler-Bernoulli beam theory in [46].

2.3 Constitutive Models Representing Specific Tissues And Parameters Identification

Forces occurring during needle insertion are tied to specific properties of both needles and crossed tissues. If general assumptions are made on force distributions during needle insertion (see Fig. 6), intrinsic parameters depend on the mechanical properties of the simulated tissues. Therefore, parameters of aforementioned models need to be determined to realistically render needle insertion in specific tissues. Table 3 sums up the various recent studies that permitted to determine these parameters for various kinds of needle insertion applications.

In deformation based models, a method to obtain parameters is fitting functions to experimental data obtained during needle insertions [27, 28, 32, 33] or traction/compression experiments [45]. Usually, these experiments are performed on cadaveric (*ex vivo*) or animal tissues (*in vivo* or *ex vivo*). However, *in vivo* data are difficult to obtain and *ex vivo* tissues cannot reliably replicate life-like conditions as in *in vivo* tissues [48, 49]. Moreover, the accuracy of mechanical properties from cadaveric samples would be hindered by freezing and thawing [32]. Therefore, it is complicated to get accurate parameters only using these methods. Several authors thus preferred to rely on literature or the experience of expert surgeons to evaluate the accuracy of force feedback through iterative try-and-test experiments. [24–26] Tissue phantoms can also be used to test behaviors of needle insertion in various conditions and verify models in early stages but were not used to identify parameters. [1, 50]

Efforts are still made to propose measurement schemes to acquire reliable force-displacement data to help in the design of needle insertion haptic simulators. Measuring devices usually rely on 6 DoF force sensors and means to track positions during insertions. The measurement apparatus can either be handheld by an expert surgeon [13, 14, 49] or be robotically inserted through tissues [14, 51]. To improve the reliability of insertion tests into *ex vivo* tissues, Li et al. developed a test platform that would record forces during needle insertion while simulating movements due to respiration [14].

Authors	Papers	Anatomy parts	Identification
Daniel	[24]	Shoulder	Literature and expert feedback
Senac	[25]	Lumbar region	Expert feedback
DiVece	[26]	Abdominal region	Cohort of 10 expert urologists feedback
Wu	[27]	liver	Measurements: insertion tests on bovine and porcine samples
Pepley	[13]	Neck tissues	Cadaver needle insertion experiments and expert feedback
Sadeghnejad	[28]	Simus and skull basis	Sheep sample curette insertion tests
Pepley	[42]	Neck tissues	Expert feedback and method developed in [41]
Correa	[30]	Alveolar nerve block	Novice and Expert Feedback
Sainsbury	[7]		Validation through novice and expert users
Li	[14]	Kidney	Custom test-bench, <i>in vivo</i> measurements, evaluation with novice and experts
Esterer	[32]	Epidural needle insertion	Measurements on cadaveric and phantom tissues
El-Monajjed	[33]	Intervertebral disc	Measurements on cadaveric tissues
Mohammadi (FEM)	[44]	Skin	Experimental data obtained from literature

Table 3: Overview of parameters identification methods in recent years

2.4 Conclusion On Needle Insertion Simulation Models

Various approaches have been proposed to render forces during needle insertion into tissues. They vary in computation complexity and their ability to accurately render forces and deformations. The majority of recent works still rely on displacement-force models that heavily depend on experimental data and/or expert feedback to obtain realistic force feedback. These approaches are still being improved to grasp all the complexity of needle insertion into soft tissues and motivate the prototyping of measurement apparatus.

3 Gesture Assessment

To integrate haptic simulators into a medical training routine, it is essential to define two validation criteria for its use. Indeed, the simulator must allow the evaluation of the performance to qualify the gesture and quantify the progress of a trainee [52]. On the other hand, it is necessary to study the impact of the use of new technology in the learning process compared to the classical training method [53].

3.1 Gesture Evaluation

Whatever the practice and the application, it is necessary to objectively evaluate the mastery of the gesture and the progress of the trainee during his/her curriculum. There are many evaluation methods specific to each type of skill. In the medical field, many evaluation methods have been developed to study surgical procedures. We can classify them into two categories [54]:

- Subjective tests based on responses to targeted questionnaires [55, 56];
- Objective tests from the study of skills based on measurable metrics [57, 58].

The first method to evaluate and improve the learning of a gesture is the use of subjective tests. These are usually carried out in the form of questionnaires. In the medical field, there are three well-known questionnaires:

- The NASA TLX [59];
- The ASQ - IBM [60];
- Bibliographic Collection and Usability Scale System [61].

The main issue with the subjective tests is that they do not allow for the quantification of learning because they are based on the learner's feelings. However, they allow us to measure the degree of confidence of a learner concerning the action he/she is performing. The other problem with these tests is that they focus exclusively on a specific application (e.g. NASA TLX for robot-assisted surgery). As far as we know, for the moment, there are no subjective tests that can be used for needle insertion. This gesture focuses, most of the time, exclusively on the haptic perception of the learner. This sense, although described and studied mechanically in the literature, is only rarely present in psychomotor studies.

Concerning the objective tests, there are several methods to analyze the gesture allowing to evaluate its mastery such as the OSATS method [62–64] (for classical surgery) or the GOALS method [65, 66] (for minimally invasive surgery and in particular laparoscopy). Other metrics for surgical skill evaluation have been developed based on the orientation of surgical instruments [58]. These methods make it possible to score a specific gesture of a practitioner. This rating is based on the analysis of predefined metrics and can be compared with a reference score from an expert procedure [67]. It is thus possible to evaluate the performance of a learner and to use the results of these algorithms for training purposes to evaluate the mastery of each user [68].

Although each application has its properties, standard and commonly used metrics for the evaluation of medical procedures have been isolated. These metrics are thus highlighted in various review studies [69, 70]. The main metrics used to qualify a gesture are:

- the TCT [71, 72] (Task Completion Time) which corresponds to the time necessary to complete a task;
- the deviation from an optimal path [70, 73];
- the regularity of movements [69, 74];
- the economy of movement [70];
- the length of movement [69, 75];
- the trajectory curvature and affine velocity [74, 76].

Most of the metrics presented above are applied to gestures performed in a three-dimensional environment and are particularly relevant to surgical gestures. These gestures require dexterity and navigation in space, hence the relevance of metrics such as affine velocity, curvature, regularity of movement, etc. Gesture evaluation methods provide good results for classical surgical gestures [77]. However, they have two main problems. The first is that they are difficult to generalize because they use predefined metrics and require the notation of reference gestures. In general, specific metrics are found in many medical applications. Indeed, it is important to note that the metrics are likely to change according to each application. This is for example the case for epidural anesthesia [78] or ventricular puncture [79] where the main metric of gesture mastery focuses on the interpretation of haptic sensations called "overshoot" [80–82]. Another example is prostate or uterine biopsy, where the main metric is the reached position [83, 84]. It is important to note that new automatic metric extraction methods are increasingly developed to address this issue [85].

The other recurring problem is the creation and use of a database to classify the different gestures and extract the metrics needed for evaluation. One of the possible ways to improve the creation of a usable database is to introduce gesture simulators. These simulators aim to reproduce clinical contexts to allow the learner to train for a specific task. They can be purely virtual [86], augmented reality [87, 88] or haptic [78, 89]. The main advantage of using simulators comes from their ability to extract data from learners' gestures. This data is derived from the many sensors that the simulator may contain and

can then be used to evaluate the gesture and provide accurate feedback to the learner to improve their learning. However, the use of simulators is still very little implemented in practice because they require a modification of the clinical learning routine and their impact on the learner's curriculum is not yet detailed in the literature.

3.2 Impact of Robotics in Clinical Routine

There are several works in the field of Human-Machine Interfaces (HMI) that have focused on the impact of robotic interfaces in surgery. For now, most of its work focuses on soft skills [90] (e.g., workflow, communication, situational awareness, teamwork [53, 91–93]) and it uses mostly subjective metrics. In addition, a 2015 neuroscience study by Heuer et al [94] on robotic assistance for surgery, showed that it was difficult to show that a surgeon keeps the same level of practice before and after the use of robotic systems in their clinical routine [94]. He demonstrated the risk of dependence of the learner on the specific robotic system in the context of teleoperation. However, a more recent study based on objective metrics also concluded that the use of robotic systems in the learning process achieved the desired level of competence faster than with the conventional learning routine [95].

In general, the addition of a haptic simulator in the learning process of a gesture has beneficial aspects such as the creation and use of databases allowing objective or subjective feedback to the learner or the saving of learning time [95]. It should be noted, however, that such devices are still not widely used in practice because they often require too great a change in the learning routine.

Conclusions

In this paper, we provided an updated short review about haptic training simulation on needle insertion. As an extensive overview was provided by Correa et al. in 2019 [6], we focused here on three important and complementary aspects of such simulators. We summed up specifications concerning the haptic devices for this hands-on training application and exposed current commercial devices and recent custom realizations. We reviewed recent force modeling approaches, and detailed the issues of assessment with simulators and their integration into learning routine. We could determine that recent results refine well-known models and enhance experimental setups and methods to facilitate generically getting experimental data. FEM-based approaches still do not permit real-time computation but hybrid approaches as proposed by [1] may become a good compromise. Despite these progress, the variety of case studies make it still difficult comparing the various aforementioned modeling approaches as no comparative work has yet been performed, as far as we know. As exposed in [3], training simulators should be evaluated in terms of rendered haptic fidelity. Each complete chain (from experimental data gathering to haptic rendering evaluated using psychophysical studies) should be evaluated to determine the most efficient ones in each family of medical cases. Classifying these simulators

in low/medium/high haptic fidelity families would help embed them efficiently into medical curricula.

Compliance with Ethical Standards

Conflict of Interest

No author has any conflicts of interest.

Human and Animal Rights and Informed Consent

This article does not contain any studies with human or animal subjects performed by any of the authors.

References

- [1] Wittek, A. *et al.* Mathematical modeling and computer simulation of needle insertion into soft tissue. *PloS one* **15** (12), e0242704 (2020). <https://doi.org/10.1371/journal.pone.0242704> .
- [2] Vaughan, N., Dubey, N., Venketesh, M. Y., Wee, K. & Isaacs, R. A review of epidural simulators : where are we today ? *Medical Engineering and Physics* **35** (9), 1235–1250 (2013) .
- [3] Favier, V., Subsol, G., Duraes, M., Captier, G. & Gallet, P. Haptic fidelity: The game changer in surgical simulators for the next decade? *Frontiers in Oncology* **11**, 3110 (2021). <https://doi.org/10.3389/fonc.2021.713343> .
- [4] Lelevé, A., McDaniel, T. & Rossa, C. Haptic training simulation. *Frontiers in Virtual Reality* **1**, 3 (2020). <https://doi.org/10.3389/frvir.2020.00003> .
- [5] Overtoom, E. M., Horeman, T., Jansen, F.-W., Dankelman, J. & Schreuder, H. W. R. Haptic feedback, force feedback, and force-sensing in simulation training for laparoscopy: A systematic overview. *Journal of Surgical Education* **76** (1), 242–261 (2019). <https://doi.org/10.1016/j.jsurg.2018.06.008> .
- [6] ●● Corrêa, C. G., Nunes, F. L., Ranzini, E., Nakamura, R. & Tori, R. Haptic interaction for needle insertion training in medical applications: The state-of-the-art. *Medical Engineering and Physics* **63**, 6–25 (2019). <https://doi.org/10.1016/j.medengphy.2018.11.002>,
This paper is an important recent and detailed review about haptic training simulators .

- [7] • Sainsbury, B. *et al.* Evaluation of a virtual reality percutaneous nephrolithotomy (pcnl) surgical simulator. *Frontiers in Robotics and AI* **6**, 145 (2020). <https://doi.org/10.3389/frobt.2019.00145> .
- [8] Yin, J. *et al.* 3d printed multi-material medical phantoms for needle-tissue interaction modelling of heterogeneous structures. *Journal of Bionic Engineering* **18** (2), 346–360 (2021). <https://doi.org/10.1007/s42235-021-0031-1> .
- [9] Ng, S. Y., Kuo, Y.-L. & Lin, C.-L. Low-cost and easily fabricated ultrasound-guided breast phantom for breast biopsy training. *Applied Sciences* **11** (16) (2021). <https://doi.org/10.3390/app11167728> .
- [10] DF, P. *et al.* Low-cost haptic simulation using material fracture. *IEEE Trans Haptics* **12** (4), 563–570 (2019). <https://doi.org/10.1109/TOH.2019.2914441> .
- [11] Moo-Young, J., Weber, T. M., Kapralos, B., Quevedo, A. & Alam, F. Development of unity simulator for epidural insertion training for replacing current lumbar puncture simulators. *Cureus* **13** (2), e13409–e13409 (2021). URL <https://pubmed.ncbi.nlm.nih.gov/33758704> .
- [12] Heimann, F. *et al.* A custom virtual reality training solution for ophthalmologic surgical clinical trials. *Advances in Simulation* **6** (1), 12 (2021). <https://doi.org/10.1186/s41077-021-00167-z> .
- [13] Pepley, D. F. *et al.* Integrating cadaver needle forces into a haptic robotic simulator. *Journal of medical devices* **12** (1) (2018). <https://doi.org/10.1115/1.4038562> .
- [14] Li, F. *et al.* Real-time needle force modeling for vr-based renal biopsy training with respiratory motion using direct clinical data. *Applied Bionics and Biomechanics* **2019**, 9756842 (2019). <https://doi.org/10.1155/2019/9756842> .
- [15] Aygün, M. M., Ögüt, Y. C., Baysal, H. & Taşcıoğlu, Y. Visuo-haptic mixed reality simulation using unbound handheld tools. *Applied Sciences* **10** (15) (2020). <https://doi.org/10.3390/app10155344> .
- [16] Sénac, T. *et al.* Simulating a syringe behavior using a pneumatic cylinder haptic interface. *Control Engineering Practice* **90**, 231–240 (2019). <https://doi.org/https://doi.org/10.1016/j.conengprac.2019.07.005> .
- [17] Pozner, C. N. & Eyre, A. in *Simulation in graduate medical education* (eds Strother, C., Okuda, Y., Wong, N. & McLaughlin, S.) *Comprehensive Healthcare Simulation: Emergency Medicine* 173–180 (Springer International Publishing, Cham, 2021). URL <https://doi.org/10.1007/>

978-3-030-57367-6_16.

- [18] Ravali, G. & Manivannan, M. Haptic feedback in needle insertion modeling and simulation. *IEEE Reviews in Biomedical Engineering* **10**, 63–77 (2017). <https://doi.org/10.1109/RBME.2017.2706966> .
- [19] Yang, C., Xie, Y., Liu, S. & Sun, D. Force modeling, identification, and feedback control of robot-assisted needle insertion: A survey of the literature. *Sensors* **18** (2) (2018). <https://doi.org/10.3390/s18020561> .
- [20] Azar, T. & Hayward, V. Bello, F. & Edwards, P. J. E. (eds) *Estimation of the fracture toughness of soft tissue from needle insertion*. (eds Bello, F. & Edwards, P. J. E.) *Lecture Notes in Computer Science, Conference on Biomedical Simulation (ISBMS 2008)*, 166–175 (Springer Berlin Heidelberg, Berlin, Heidelberg, 2008).
- [21] Brett, P. N., Parker, T., Harrison, A. J., Thomas, T. A. & Carr, A. Simulation of resistance forces acting on surgical needles. *Proceedings of the Institution of Mechanical Engineers, Part H: Journal of Engineering in Medicine* **211** (4), 335–347 (1997). <https://doi.org/10.1243/0954411971534467> .
- [22] Okamura, A. M., Simone, C. & O’leary, M. D. Force modeling for needle insertion into soft tissue. *IEEE transactions on biomedical engineering* **51** (10), 1707–1716 (2004). <https://doi.org/10.1109/TBME.2004.831542> .
- [23] Barbé, L., Bayle, B., de Mathelin, M. & Gangi, A. Needle insertions modeling: Identifiability and limitations. *Biomedical signal processing and control* **2** (3), 191–198 (2007). <https://doi.org/10.1016/j.bspc.2007.06.003> .
- [24] Alamilla Daniel, M. d. I. A., Moreau, R. & Tanneguy, R. EMBC (ed.) *Development of haptic simulator for practicing the intraarticular needle injection under echography*. (ed.EMBC) *Proc. of the 42nd Annual International Conference of the IEEE Engineering in Medicine Biology Society (EMBC 2020)*, 4713–4716 (2020).
- [25] Senac, T. *et al.* IEEE (ed.) *Designing an accurate and customizable epidural anesthesia haptic simulator*. (ed.IEEE) *Proc. of the International Conference on Robotics and Automation (ICRA 2019)*, 8353–8359 (IEEE, Montreal, Canada, 2019).
- [26] Di Vece, C., Luciano, C. & De Momi, E. Psychomotor skills development for veress needle placement using a virtual reality and haptics-based simulator. *International Journal of Computer Assisted Radiology and Surgery* **16** (4), 639–647 (2021). <https://doi.org/10.1007/s11548-021-02341-0> .

- [27] Wu, H., Chen, C., Zhou, Y., Wang, J. & Xie, Y. IEEE/ASME (ed.) *VR-based haptic simulation for dynamic needle insertion*. (ed.IEEE/ASME) *Proc. of the International Conference on Advanced Intelligent Mechatronics (AIM 2019)*, 924–929 (2019).
- [28] Sadeghnejad, S., Farahmand, F., Vossoughi, G., Moradi, H. & Hosseini, S. M. S. Phenomenological tissue fracture modeling for an endoscopic sinus and skull base surgery training system based on experimental data. *Medical engineering & physics* **68**, 85–93 (2019). <https://doi.org/10.1016/j.medengphy.2019.02.004> .
- [29] Mostafa, A. E. *et al.* Designing neurosimvr: A stereoscopic virtual reality spine surgery simulator (2017). URL <https://prism.ucalgary.ca/handle/1880/52230>.
- [30] Correa, C. G., Machado, M. A. d. A. M., Ranzini, E., Tori, R. & Nunes, F. d. L. S. Virtual reality simulator for dental anesthesia training in the inferior alveolar nerve block. *J. Appl. Oral Sci.* **25** (4) (2017). <https://doi.org/10.1590/1678-7757-2016-0386> .
- [31] Barnouin, C., Zara, F. & Jaillet, F. SciTePress (ed.) *A real-time ultrasound rendering with model-based tissue deformation for needle insertion*. (ed.SciTePress) *Proc. of the 15th International Conference on Computer Graphics Theory and Applications (GRAPP 2020)* (Valletta, Malta, 2020).
- [32] Esterer, B. *et al.* Characterization of tissue properties in epidural needle insertion on human specimen and synthetic materials. *Journal of the Mechanical Behavior of Biomedical Materials* **110**, 103946 (2020). <https://doi.org/10.1016/j.jmbbm.2020.103946> .
- [33] El-Monajjed, K. & Driscoll, M. Analysis of surgical forces required to gain access using a probe for minimally invasive spine surgery via cadaveric-based experiments towards use in training simulators. *IEEE Transactions on Biomedical Engineering* **68** (1), 330–339 (2021). <https://doi.org/10.1109/TBME.2020.2996980> .
- [34] Ricca, A., Chellali, A. & Otmane, S. Comparing touch-based and head-tracking navigation techniques in a virtual reality biopsy simulator. *Virtual Reality* **25** (1), 191–208 (2021). <https://doi.org/10.1007/s10055-020-00445-7> .
- [35] Kuchenbecker, K. J., Fiene, J. & Niemeyer, G. Improving contact realism through event-based haptic feedback. *IEEE transactions on visualization and computer graphics* **12** (2), 219–230 (2006) .

- [36] Maurin, B. *et al.* in *In vivo study of forces during needle insertions* (ed.Scientific, W.) *Perspective in Image-Guided Surgery* 415–422 (World Scientific, 2004).
- [37] Mahvash, M. & Dupont, P. E. Mechanics of dynamic needle insertion into a biological material. *IEEE Transactions on Biomedical Engineering* **57** (4), 934–943 (2009). <https://doi.org/10.1109/TBME.2009.2036856> .
- [38] Barnett, A. C., Lee, Y.-S. & Moore, J. Z. Fracture mechanics model of needle cutting tissue. *Journal of Manufacturing Science and Engineering* **138** (1) (2016). <https://doi.org/10.1115/1.4030374> .
- [39] Castro-Díaz, J. D., Sánchez-Sánchez, P., Gutiérrez-Giles, A., Arteaga-Pérez, M. A. & Pliego-Jiménez, J. Experimental results for haptic interaction with virtual holonomic and nonholonomic constraints. *IEEE Access* **8**, 120959–120973 (2020). <https://doi.org/10.1109/ACCESS.2020.3006715> .
- [40] Asadian, A., Patel, R. V. & Kermani, M. R. *A distributed model for needle-tissue friction in percutaneous interventions*, 1896–1901 (2011).
- [41] Gordon, A., Kim, I., Barnett, A. C. & Moore, J. Z. ASME (ed.) *Needle insertion force model for haptic simulation*. (ed.ASME) *Proc. of the International Manufacturing Science and Engineering Conference*, Vol. 56833, V002T03A003 (American Society of Mechanical Engineers, 2015).
- [42] Pepley, D. *et al.* A virtual reality haptic robotic simulator for central venous catheterization training. *Journal of medical devices* **10** (3) (2016). <https://doi.org/10.1115/1.4033867> .
- [43] Perrusi, P. H. S., Baksic, P. & Courtecuisse, H. *Interactive finite element model of needle insertion and laceration* (2021).
- [44] Mohammadi, H., Ebrahimian, A. & Maftoon, N. Fracture behaviour of human skin in deep needle insertion can be captured using validated cohesive zone finite-element method. *Computers in Biology and Medicine* **139**, 104982 (2021). <https://doi.org/10.1016/j.combiomed.2021.104982> .
- [45] Jushiddi, M. G. *et al.* A computational multilayer model to simulate hollow needle insertion into biological porcine liver tissue. *Acta Biomaterialia* **136**, 389–401 (2021). <https://doi.org/10.1016/j.actbio.2021.09.057> .
- [46] Bui, H. P., Tomar, S. & Bordas, S. P. Corotational cut finite element method for real-time surgical simulation: Application to needle insertion simulation. *Computer methods in applied mechanics and engineering* **345**, 183–211 (2019). <https://doi.org/10.1016/j.cma.2018.10.023> .

- [47] Kataoka, H., Washio, T., Audette, M. & Mizuhara, K. Heidelberg, S.-V. B. (ed.) *A model for relations between needle deflection, force, and thickness on needle penetration.* (ed.Heidelberg, S.-V. B.) *Proc. of the International Conference on Medical Image Computing and Computer-Assisted Intervention (MICCAI 2001)*, 966–974 (Springer, 2001).
- [48] Mirza, S. & Athreya, S. Review of simulation training in interventional radiology. *Academic Radiology* **25** (4), 529–539 (2018). URL <https://www.sciencedirect.com/science/article/pii/S107663321730435X>. <https://doi.org/https://doi.org/10.1016/j.acra.2017.10.009> .
- [49] Schimmoeller, T., Neumann, E. E., Nagle, T. F. & Erdemir, A. Reference tool kinematics-kinetics and tissue surface strain data during fundamental surgical acts. *Scientific Data* **7** (1), 21 (2020). <https://doi.org/10.1038/s41597-020-0359-0> .
- [50] Jushiddi, M. G. *et al.* Simulation of biopsy bevel-tipped needle insertion into soft-gel. *Computers in Biology and Medicine* **111**, 103337 (2019). <https://doi.org/10.1016/j.compbiomed.2019.103337> .
- [51] Brown, D., Gonzalez-Vargas, J. M., Han, D., Miller, S. & Moore, J. ASME (ed.) *Incremental needle insertion system for force and position sensing.* (ed.ASME) *Frontiers in Biomedical Devices Proc. of the 2020 Design of Medical Devices Conference*, Vol. 2020 Design of Medical Devices Conference (Minneapolis, Minnesota, USA, 2020). V001T06A002.
- [52] • Marvel, J. A., Bagchi, S., Zimmerman, M. & Antonishek, B. Towards effective interface designs for collaborative hri in manufacturing: Metrics and measures. *J. Hum.-Robot Interact.* **9** (4) (2020). <https://doi.org/10.1145/3385009>,
This work highlights the necessity of evaluating gestures during robotic interaction and the metrics which can be used .
- [53] • Avellino, I. *et al.* ACM (ed.) *Impacts of telemanipulation in robotic assisted surgery.* (ed.ACM) *Proc. of the CHI Conference on Human Factors in Computing Systems Proceedings (CHI 2019)*, CHI Conference on Human Factors in Computing Systems Proceedings (CHI 2019) (Glasgow, United Kingdom, 2019).
This study focuses on the impact of the robotics on the clinical routine of the surgeons.
- [54] • Close, M. *et al.* Subjective vs computerized assessment of surgeon skill level during mastoidectomy. *Otolaryngology–Head and Neck Surgery* 1255–1257 (2020). <https://doi.org/10.1177/0194599820933882>,
Findings from this study discuss the evaluation of the medical gesture and explain the two different ways to evaluate it .

- [55] Charles, R. & Nixon, J. Measuring mental workload using physiological measures: A systematic review. *Applied ergonomics* **74**, 221–232 (2019) .
- [56] Adrian Cornelius, M. *et al.* Physiological parameter response to variation of mental workload. *Human Factors* **60** (1), 31–56 (2018). <https://doi.org/10.1177/0018720817733101>, PMID: 28965433 .
- [57] Seung-kook, J. *et al.* Evaluation of robotic minimally invasive surgical skills using motion studies.. In *Proceedings of the Workshop on Performance Metrics for Intelligent Systems (PerMIS '12)* 198–205 (2012). <https://doi.org/10.1145/2393091.2393129> .
- [58] Sharon, Y., Jarc, A. M., Lendvay, T. S. & Nisky, I. Rate of orientation change as a new metric for robot-assisted and open surgical skill evaluation. *IEEE Transactions on Medical Robotics and Bionics* **3** (2), 414–425 (2021). <https://doi.org/10.1109/tmrb.2021.3073209> .
- [59] Hart, S. G. & Staveland, L. E. in *Development of nasa-tlx (task load index): Results of empirical and theoretical research* (eds Hancock, P. A. & Meshkati, N.) *Human Mental Workload*, Vol. 52 of *Advances in Psychology* 139–183 (North-Holland, 1988). URL <https://www.sciencedirect.com/science/article/pii/S0166411508623869>.
- [60] Lewis, J. R. IBM computer usability satisfaction questionnaires: Psychometric evaluation and instructions for use. *International Journal of Human–Computer Interaction* **7** (1), 57–78 (1995). <https://doi.org/10.1080/10447319509526110> .
- [61] Klug, B. An overview of the system usability scale in library website and system usability testing. *Weave: Journal of Library User Experience* **1** (2017). <https://doi.org/10.3998/weave.12535642.0001.602> .
- [62] Martin, J. A. *et al.* Objective structured assessment of technical skill (osats) for surgical residents. *Br J Surg.* **84** (2), 273–8 (1997). URL <http://www.csats.com/osats>. <https://doi.org/10.1046/j.1365-2168.1997.02502.x> .
- [63] Niitsu, H. *et al.* Using the objective structured assessment of technical skills (osats) global rating scale to evaluate the skills of surgical trainees in the operating room. *Surgery today* **43** (2012). <https://doi.org/10.1007/s00595-012-0313-7> .
- [64] Asif, H. *et al.* Objective structured assessment of technical skill (osats) in the surgical skills and technology elective program (sstep): Comparison of peer and expert raters. *The American Journal of Surgery, Elsevier* (2021). <https://doi.org/10.1016/j.amjsurg.2021.03.064> .

- [65] Higuchi, M. *et al.* Development and validation of a porcine organ model for training in essential laparoscopic surgical skills. *International Journal of Urology* **27** (10), 929–938 (2020). <https://doi.org/10.1111/iju.14315> .
- [66] Vassiliou, M. C. *et al.* A global assessment tool for evaluation of intra-operative laparoscopic skills. *The American Journal of Surgery* **190** (1), 107–113 (2005). <https://doi.org/10.1016/j.amjsurg.2005.04.004> .
- [67] Zia, A. & Essa, I. Automated surgical skill assessment in rmis training. *International Journal of Computer Assisted Radiology and Surgery* **13**, 731–739 (2018). <https://doi.org/10.1007/s11548-018-1735-5> .
- [68] Zia, A. *et al.* Automated video-based assessment of surgical skills for training and evaluation in medical schools. *International Journal of Computer Assisted Radiology and Surgery* **11** (9), 1623–1636 (2016). <https://doi.org/10.1007/s11548-016-1468-2> .
- [69] Cotin, S. *et al.* “metrics for laparoscopic skills trainers: The weakest link!”. *Medical Image Computing and Computer-Assisted Intervention — MICCAI 2002* **2488**, 35–43 (2002) .
- [70] Feng, C., Rozenblit, J. & Hamilton, A. IEEE (ed.) *Fuzzy logic-based performance assessment in the virtual, assistive surgical trainer (vast)*. (ed.IEEE) *Proc. of the 15th IEEE Intl. Conference and Workshops on the Engineering of Computer Based Systems (ECBS '08)*, 203–209 (Belfast, Ireland, 2008).
- [71] Mishra, S., Ganpule, A., Kurien, A., Muthu, V. & Desai, M. Task completion time: Objective tool for assessment of technical skills in laparoscopic simulator for urology trainees. *Indian Journal of Urology* **24** (1), 35–38 (2008). <https://doi.org/10.4103/0970-1591.38601> .
- [72] Fuerst, D., Hollensteiner, M. & Schrempf, A. ACM (ed.) *A novel augmented reality simulator for minimally invasive spine surgery*. (ed.ACM) *Proc. of the 2014 Summer Simulation Multiconference, SummerSim '14* (Society for Computer Simulation International, San Diego, CA, USA, 2014).
- [73] Fortmeier, D., Mastmeyer, A., Schröder, J. & Handels, H. A virtual reality system for ptdc simulation using direct visuo-haptic rendering of partially segmented image data. *IEEE Journal of Biomedical and Health Informatics* **20**, 355–366 (2016) .
- [74] Fard, M. J. *et al.* Machine learning approach for skill evaluation in robotic-assisted surgery. *CoRR* **abs/1611.05136** (2016). URL <http://arxiv.org/abs/1611.05136> . 1611.05136 .

- [75] Chalasani, V. *et al.* Development and validation of a virtual reality transrectal ultrasound guided prostatic biopsy simulator. *Canadian Urological Association journal = Journal de l'Association des urologues du Canada* **5**, 19–26 (2011). <https://doi.org/10.5489/cuaj.09159> .
- [76] Cifuentes Quintero, J. A. *Development of a new technique for objective assessment of gestures in mini-invasive surgery*. Theses, INSA de Lyon ; Universidad nacional de Colombia (2015). URL <https://tel.archives-ouvertes.fr/tel-01368173>.
- [77] Millan, M. *L'apprentissage profond pour l'évaluation et le retour d'information lors de l'apprentissage de gestes*. Theses, Sorbonne Université (2020). URL <https://tel.archives-ouvertes.fr/tel-03191291>.
- [78] Senac, T. *et al.* Skill assessment of an epidural anesthesia using the PeriSIM simulator. *IEEE Transactions on Medical Robotics and Bionics* **3** (1), 106–114 (2021). URL <https://hal.archives-ouvertes.fr/hal-03092054>. <https://doi.org/10.1109/TMRB.2020.3048247> .
- [79] Brenke, C., Fürst, J., Katsigiannis, S. & Carolus, A. High accuracy of external ventricular drainage placement using anatomical landmarks. *Neurochirurgie* **66** (6), 435–441 (2020). <https://doi.org/10.1016/j.neuchi.2020.09.009> .
- [80] Manoharan, V., van Gerwen, D., van den Dobbelsteen, J. & Dankelman, J. Design and validation of an epidural needle insertion simulator with haptic feedback for training resident anaesthesiologists. *2012 IEEE Haptics Symposium (HAPTICS)* 341–348 (2012) .
- [81] van Adrichem, L. Avoiding overshoot. *DSPE Mikroniek* **3**, 36–40 (2009) .
- [82] Cometa, A. When millimeters count, epidural loss of resistance techniques differ : A simulator study. *UFHealth, University of Florida* (2018) .
- [83] Chalard, R., Fazel, A. & Vitrani, M.-A. Real time estimator to perform targeted biopsies with a free-wrist robot despite large deformations of the insertion orifice. *Frontiers in Robotics and AI* (2021). URL <https://hal.archives-ouvertes.fr/hal-03385568> .
- [84] Goksel, O., Sapchuk, K. & Salcudean, S. E. Haptic simulator for prostate brachytherapy with simulated needle and probe interaction. *IEEE Transactions on Haptics* **4** (3), 188–198 (2011). <https://doi.org/10.1109/TOH.2011.34> .
- [85] Guerin, S., Huaultmé, A., Lavoue, V., Jannin, P. & Timoh, K. N. Review of automated performance metrics to assess surgical technical skills in robot-assisted laparoscopy. *Surgical Endoscopy* (2021). <https://doi.org/>

[10.1007/s00464-021-08792-5](https://doi.org/10.1007/s00464-021-08792-5) .

- [86] Alaker, M., Wynn, G. R. & Arulampalam, T. Virtual reality training in laparoscopic surgery: A systematic review and meta-analysis. *International Journal of Surgery* **29**, 85–94 (2016). <https://doi.org/10.1016/j.ijssu.2016.03.034> .
- [87] Christensen, N. H. *et al.* ACM (ed.) *Depth cues in augmented reality for training of robot-assisted minimally invasive surgery.* (ed.ACM) *Proc. of the 21st International Academic Mindtrek Conference*, Academic-Mindtrek'17, 120–126 (Association for Computing Machinery, New York, NY, USA, 2017). URL <https://doi.org/10.1145/3131085.3131123>.
- [88] Huang, Y.-H. *et al.* Catar: A novel stereoscopic augmented reality cataract surgery training system with dexterous instruments tracking technology. *Proceedings of the 2018 CHI Conference on Human Factors in Computing Systems* (2018) .
- [89] Galvan, A., Da Costa, A. K., Shields, J., Kho, K. & Fey, A. M. *Haptic simulator for trocar insertion training*, 397–402 (2021).
- [90] Yule, S., Flin, R., Maran, N. & Paterson-Brown, S. Non-technical skills for surgeons in the operating room: A review of the literature. *Surgery* **139** (2), 140–149 (2006). <https://doi.org/10.1016/j.surg.2005.06.017> .
- [91] Cheatle, A., Pelikan, H., Jung, M. & Jackson, S. Sensing (co)operations: Articulation and compensation in the robotic operating room. *Proc. ACM Hum.-Comput. Interact.* **3** (CSCW) (2019). URL [10.1145/3359327](https://doi.org/10.1145/3359327). <https://doi.org/10.1145/3359327> .
- [92] Pelikan, H., Cheatle, A., Jung, M. & Jackson, S. Operating at a distance - how a teleoperated surgical robot reconfigures teamwork in the operating room. *Proceedings of the ACM on Human-Computer Interaction* **2**, 1–28 (2018). <https://doi.org/10.1145/3274407> .
- [93] Randell, R., Honey, S., Hindmarsh, J. & et al. A realist process evaluation of robot-assisted surgery: integration into routine practice and impacts on communication, collaboration and decision-making. *Southampton (UK): Health Services and Delivery Research (NIHR) Journals Library* (2017). <https://doi.org/10.3310/hsdr05200> .
- [94] Heuer, H. & Lüttgen, J. Robot assistance of motor learning: A neuro-cognitive perspective. *Neuroscience & Biobehavioral Reviews* **56**, 222–240 (2015). <https://doi.org/10.1016/j.neubiorev.2015.07.005> .
- [95] Ferrier-Barbut, E., Gauthier, P., Luengo, V., Canlorbe, G. & Vitrani, M.-A. Measuring the quality of learning in a human-robot collaboration:

a study of laparoscopic surgery. *ACM Transactions on Human-Robot Interaction* (2021). URL <https://hal.archives-ouvertes.fr/hal-03355055> .