



**HAL**  
open science

# Mapping the relative shear modulus within biological tissues from internal displacement fields measured in quasi-static ultrasound elastography

Elisabeth Brusseau, Lorena Petrusca, Elie Bretin, Pierre Millien, Laurent Seppecher

## ► To cite this version:

Elisabeth Brusseau, Lorena Petrusca, Elie Bretin, Pierre Millien, Laurent Seppecher. Mapping the relative shear modulus within biological tissues from internal displacement fields measured in quasi-static ultrasound elastography. 10th International Conference 'Inverse Problems: Modeling and Simulation' (IPMS-2022), The Eurasian Association on Inverse Problems (EAIP), May 2022, Mellieha, Malta. hal-03793238

**HAL Id: hal-03793238**

**<https://hal.science/hal-03793238>**

Submitted on 13 Nov 2022

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.



Distributed under a Creative Commons Attribution 4.0 International License

## MAPPING THE RELATIVE SHEAR MODULUS WITHIN BIOLOGICAL TISSUES FROM INTERNAL DISPLACEMENT FIELDS MEASURED IN QUASI-STATIC ULTRASOUND ELASTOGRAPHY

E. Brusseau<sup>a</sup>, L. Petrusca<sup>a</sup>, E. Bretin<sup>b</sup>, P. Millien<sup>c</sup>, L. Seppecher<sup>d</sup>

<sup>a</sup>CREATIS, Univ Lyon, France, [elisabeth.brusseau@creatis.insa-lyon.fr](mailto:elisabeth.brusseau@creatis.insa-lyon.fr); [lorena.petrusca@creatis.insa-lyon.fr](mailto:lorena.petrusca@creatis.insa-lyon.fr)

<sup>b</sup>Institut Camille Jordan, INSA de Lyon, Univ Lyon, France, [elie.bretin@insa-lyon.fr](mailto:elie.bretin@insa-lyon.fr)

<sup>c</sup>Institut Langevin, PSL Research University, Paris, France, [pierre.millien@espci.fr](mailto:pierre.millien@espci.fr)

<sup>d</sup>Institut Camille Jordan, Ecole Centrale de Lyon, Univ Lyon, France, [laurent.seppecher@ec-lyon.fr](mailto:laurent.seppecher@ec-lyon.fr)

### Abstract

This presentation deals with relative shear modulus mapping within linear elastic and isotropic media, from the knowledge of displacement fields estimated in quasi-static ultrasound elastography (QSUE). QSUE mainly provides images of the internal displacements and strains that biological tissues experience under compression. Typically, the operator compresses the biological tissues with the probe while radiofrequency (RF) ultrasound images are acquired. With this technique, no boundary information is therefore available. Moreover, internal displacement fields are estimated from the acquired RF images, leading to 2D displacements, whereas the inverse problem is inherently a 3D one. In our recent work, mathematical developments for analyzing this inverse elasticity problem were introduced, as well as some initial results [1,2]. Here, to overcome the 2D/3D data problem, a plane stress-based approach is considered. The proposed method is based on the variational formulation of the equilibrium equations. This approach involves the use of test functions, chosen in this work to vanish at the domain boundary, because of the lack of information. Using a suitable finite element discretization technique leads to a finite dimensional homogeneous system of linear equations and the problem of reconstructing a relative shear modulus map is finally reduced to a null space identification problem of a large sparse matrix. With this method, only the displacement fields within the analyzed media need to be known beforehand. They are estimated using a method we previously developed for strain imaging [3].

Relative shear modulus mapping was assessed with plane stress and 3D numerical simulations, as well as data from phantom experiments (CIRS models 049 and 059). These different media consist of spherical inclusions embedded in homogeneous background materials. In all cases, the reconstructed maps allow a clear identification of the regions differing in stiffness, with inclusion-to-background modulus ratios (IBRs) that unambiguously reveal whether inclusions are stiffer or softer than the surrounding material. More specifically, for the plane stress simulations, resulting IBRs are very close to the theoretical values, whereas those obtained with the 3D simulations are, as expected, less accurate since the plane stress conditions are no longer satisfied. For the experimental data, a comparison with a clinical ultrasound scanner (Aixplorer, SuperSonic Imagine, France) was also conducted. Overall, for the phantom cases, the IBRs obtained with our approach are closer to the values given by the phantom manufacturer than the IBRs from the clinical system. Finally, results obtained with *in vivo* data of breast lesions will also be presented and discussed.

### References

1. H. Ammari et al., A direct linear inversion for discontinuous elastic parameters recovery from internal displacement information only, *Numerische Mathematik*, 147, 189-226 (2021).
2. E. Brusseau et al., Reconstructing the shear modulus contrast of linear elastic and isotropic media in quasi-static ultrasound elastography, *IEEE EMBS*, 3171-74 (2021).
3. E. Brusseau et al., In vivo response to compression of 35 breast lesions observed with a two-dimensional locally regularized strain estimation method, *UMB*, 40, 300-312 (2014).