



Towards an HD-sEMG mask to measure orofacial muscle activity during speech production

Maëva Garnier, Christophe Karam, Ladislas Nalborczyk, Alberto Botter,
Sofiane Boudaoud

► To cite this version:

Maëva Garnier, Christophe Karam, Ladislas Nalborczyk, Alberto Botter, Sofiane Boudaoud. Towards an HD-sEMG mask to measure orofacial muscle activity during speech production. SMC 2022 - 8th International Conference on Speech Motor Control (SMC2022), Aug 2022, Groningen, Netherlands. . hal-03920151

HAL Id: hal-03920151

<https://hal.science/hal-03920151>

Submitted on 3 Jan 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

Towards an HD-sEMG mask to measure orofacial muscle activity during speech production



Maëva GARNIER¹, Christophe KARAM¹, Ladislas NALBORCZYK^{1,2}, Alberto BOTTER³, Sofiane BOUDAOUED⁴



¹ Univ. Grenoble Alpes, CNRS, Grenoble INP, GIPSA-lab, 38000 Grenoble, France

² Aix Marseille Univ, CNRS, LPC, Marseille, France / Aix Marseille Univ, CNRS, LNC, Marseille, France

³ LISIN, Laboratory for Engineering of the Neuromuscular System, Department of Electronics and Telecommunications - Politecnico di Torino, Turin, Italy

⁴ Alliance Sorbonne Université, Université de Technologie de Compiègne, Laboratoire de Biomécanique et de Bioingénierie CNRS-UMR 7338 BMBI, 60200, Compiègne, France



Context


Orofacial movements

- Facial expressions
- Speech
- Eating (swallowing, suction, mastication)
- Music (whistle, brass instruments)

Applications

- Speech pathology
- Modeling
- Automatic recognition
- Animation of avatars

Acquisition systems




→ In particular: characterization of effort and coordination of speech articulatory movements in typical and disordered speech

Objectives

Interest and feasibility of an HD-sEMG facial mask ?

- Compatibility with natural movements
- « Sticking power » of the electrodes
- Information redundancy
- Discrimination of basic orofacial movements
- Prediction of articulatory effort (as measured through inter-lip force)





1- Feasibility

- Layout of the arrays that respects the "naso-labial fold" on the cheek and the "mental crease" on the chin
- Only intense rounding and spreading of the lips restrained by the non-elasticity of the electrode arrays




- 2.3 "bad channels" per electrode array (over 32)
- Not increased by some movements or over time
- Significantly higher for chin array (4.2 ± 2.0), compared to the others (1.9 ± 0.9).

2- Information redundancy




- 21.1% of electrode pairs highly correlated signals ($|R| > 0.75$)
- 55.5% of the electrode pairs poorly correlated ($|R| < 0.5$)
- Not too redundant information.
- Most of the high-correlated pairs of electrodes are adjacent ones in a same array
- High correlation between upper cheek array and lower lip array ($R=0.53$), and lower cheek array ($R=0.70$)
- Muscle synergy ?


3- Discrimination of basic movements




Projection in the two first PCA dimensions



Hierarchical Agglomerative Clustering




→ Unsupervised classification of 5 movements (T3, T4, T5, T6, T7)




→ Only 3 movement (T3, T5 and T7)

Material & Method




- 1 participant
- 6 arrays of 8×4 electrodes, with Inter Electrode Distance : 5mm




- Microphone
- Force sensor between the lips
- 7 orofacial movements
- Speech (/ama/)

4- Prediction of articulatory efforts



Force $\sim \sum \text{Dim}_i$

(Adjusted $R^2 = 0.61$)



Force $\sim \sum \text{RMS}_i$

(Adjusted $R^2 = 0.37$)

Conclusions

Feasibility

Proof of concept. Elastic support appears necessary to respect skin deformation and enable unconstrained gestures

Interest, compared to multiple pairs of electrodes

High inter-electrode correlation, but enables a better discrimination of movements and better prediction of articulatory efforts.

A greater IED would remain acceptable, especially on the cheek