

## Acquisition and Physico-Chemical Data Analysis of Oxygenated Compounds From Biomass Using Microfluidics

Rosa Moreno Jimenez, Claire Marliere, Benoit Creton, Olivier Nguyen, Lionel Teule-gay, Samuel Marre

### ▶ To cite this version:

Rosa Moreno Jimenez, Claire Marliere, Benoit Creton, Olivier Nguyen, Lionel Teule-gay, et al.. Acquisition and Physico-Chemical Data Analysis of Oxygenated Compounds From Biomass Using Microfluidics. 8th World Congress on Mechanical, Chemical, and Material Engineering (MCM'22), Jul 2022, Prague, Czech Republic. pp.HTFF 181, 10.11159/htff22.181. hal-03945868

HAL Id: hal-03945868

https://hal.science/hal-03945868

Submitted on 18 Jan 2023

**HAL** is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

 $Proceedings\ of\ the\ 8^{th}\ World\ Congress\ on\ Mechanical,\ Chemical,\ and\ Material\ Engineering\ (MCM'22)$ 

Prague, Czech Republic – July 31, 2022 - August 02, 2022

Paper No. HTFF 181 DOI: 10.11159/htff22.181

# Acquisition and Physico-Chemical Data Analysis of Oxygenated Compounds from Biomass Using Microfluidics.

Rosa Moreno Jimenez<sup>1,2\*</sup>, Claire Marliere<sup>1</sup>, Benoit Creton<sup>1</sup>, Olivier Nguyen<sup>2</sup>, Lionel Teule-Gay<sup>2</sup>, Samuel Marre<sup>2</sup>

<sup>1</sup>IFP Energies nouvelles, 1 et 4 avenue de Bois-Préau, 92852 Rueil-Malmaison, France.

<sup>2</sup>CNRS, Univ. Bordeaux, ICMCB, UMR 5026, F-33600 Pessac, France
rosa.moreno@ifp.fr

Global warming-related climate change demands prompt actions to reduce greenhouse gas (GHG) emissions, particularly carbon dioxide. To reduce GHGs, biomass-based biofuels containing oxygenated compounds represent a promising alternative of energy source. To convert biomass into energy, a variety of conversion processes performed at high pressure and high temperature conditions are required, and the design of such processes need as support, thermophysical property data, particularly thermal conductivity. The conventional methods to measure thermal conductivity are often time consuming and/or requires important quantities of products. Microfluidics has been proven as an appropriate support to overcome these issues thanks to its low reagent consumption, fast screening, low operating time, improvement of heat and mass transfers etc. It allows the automated manipulation, performing high throughput experimentation. In addition, over the last 10 years, a new field of investigation called "high pressure and high temperature (HP-HT) microfluidics" [1] has gained increasing interest, in particular for the determination of the thermo-physical properties of fluids systems[2] [3]. Currently, available methods for measuring thermal conductivity in microfluidics are not adapted to HP-HT conditions. Also, thermal conductivity data of oxygenated compounds are scarce in literature or not available in extreme conditions. Therefore, the use of alternative methods such as models, combined with microfluidics, are essential to complement experimental data. Machine learning (ML) provides powerful predictive tools with the ability to learn from available data. The aim of this thesis is to develop a microfluidic device capable of measuring thermal conductivity of oxygenated compounds, operating at HP-HT (up to 100 bars and 100 °C), complementing it with modeling as a view to "high throughput" production of experimental data. In this study, two types of sensor devices are implemented within a microfluidic device, allowing the measurement of thermal conductivity using different methodologies. In parallel, a data compilation resulting in a database has been established for the development of a ML based predictive model to estimate thermal conductivity of oxygenated compounds.

### The reference section at the end of the paper should be edited based on the following:

#### References

- [1] Marre, S.; Adamo, A.; Basak, S.; Aymonier, C.; Jensen, K. F. Design and Packaging of Microreactors for High Pressure and High Temperature Applications. Ind. Eng. Chem. Res. [Online] 2010, 49 (22), 11310–11320.
- [2] B. Pinho, S. Girardon, F. Bazer-Bachi, G. Bergeot, S. Marre, C. Aymonier, Microfluidic Investigation of Multicomponent Systems at high pressure and high temperature, Lab Chip, 2014, 14, 3843-3849
- [3] T. Gavoille, N. Pannacci, G. Bergeot, C. Marlière, S. Marre, Microfluidics approaches for accessing thermophysical properties of fluids systems, Reaction Chemistry and Engineering, 2019, 4, 1721-.1739