

Application of new aptasensor modified with nanocomposite for selective estradiol valerate determination in pharmaceutical and real biological samples

Amina Touati, Mohamed Braiek, Messaoud Benounis, Nicole Jaffrezic-Renault

To cite this version:

Amina Touati, Mohamed Braiek, Messaoud Benounis, Nicole Jaffrezic-Renault. Application of new aptasensor modified with nanocomposite for selective estradiol valerate determination in pharmaceutical and real biological samples. Chemical Monthly = Monatshefte für Chemie, 2021, 152 (6), pp.577-585. 10.1007/s00706-021-02773-z hal-03997032

HAL Id: hal-03997032 <https://hal.science/hal-03997032v1>

Submitted on 20 Feb 2023

HAL is a multi-disciplinary open access archive for the deposit and dissemination of scientific research documents, whether they are published or not. The documents may come from teaching and research institutions in France or abroad, or from public or private research centers.

L'archive ouverte pluridisciplinaire **HAL**, est destinée au dépôt et à la diffusion de documents scientifiques de niveau recherche, publiés ou non, émanant des établissements d'enseignement et de recherche français ou étrangers, des laboratoires publics ou privés.

 $\stackrel{\bullet}{=}$

MCCM_Template_Vers 4 May 2014

- aptasensor to estradiol valerate with a very low limit of detection of 0.56 fM.
- This was very promising for rapid analysis of estradiol valerate detection in
- pharmaceutical and biological samples.
-
- *Graphical abstract*

Introduction

 Oestradiol (E2) is a major female sex hormone. According to Ettawfik Medical Analysis Laboratory, normal levels of E2 for menstruating women 4 range from 18 to 147 pg/cm³. For postmenopausal, E2 levels is lower than 5 .10 pg/cm³. Lower levels of E2 in the body may suggest late puberty, hypopituitarism, hypogonadism, Turner syndrome, and menopause. Whereas E2 levels that are higher than normal may suggest tumors in the ovaries, early puberty, and scarring of the liver. Estradiol valerate (EV, Fig. 1) is essentially the ester of E2 characterized by the presence of a substitution radical at carbon 17 such as the radical. Following absorption in the body, the ester was cleaved, and the natural estrogen (E2) resulted [1]. EV is very significant in clinical medicine and could be used to treat menopause syndrome and prostate cancer. This hormone can be utilized together with a progestogen for the inhibition of ovulation [2]. As a result, many clinical studies have reported the use of different detection techniques such as HPLC [3, 4] and LC-MS/MS [5], CG/MS [6, 7] for the quantification of EV in diverse samples. Thus, given the high cost, time requirements, and expensive instruments. The development of rapid, simple, selective, less expensive, sensitive, and specific analytical methods for the detection of EV is of great importance.

 Several biosensors have been developed for electrochemical detection of natural E2 [8-12] and, to our knowledge; few research articles have been published for synthetic E2 detection [13-15]. Mofidi et al. [13] have developed an electrochemical sensor for EV in blood samples; a carbon paste electrode was modified with terbium carbonate nanoparticles for EV determination. The method showed a limit detection of 0.1 ng/cm³ with a 7 linear range of 13×10^2 - 10×10^3 ng/cm³. In this study, EV was extracted from the blood sample with a recovery percentage of 67%. Various electrochemical methods are based on aptamer-based biosensors for the detection of E2 in real samples [16-20].

 Recently, many aptasensor based on nanomaterial electrochemical electrode-modifying has received a huge amount of attention due to their inherent advantages such as high specific surface area and excellent electrical performance of nanomaterials such as simplicity, high sensitivity, low cost, and high stability [21-24]. Hlavata et al. [25] have developed an electrochemical biosensor based on nanomaterials, by combining the benefits of chitosan and carboxylate single-walled carbon nanotubes for evaluation of damage to DNA induced by UV-C radiation. They were studied the role of these nanomaterials (with and without carbon nanotube) for the performance of biosensors and they deduced that the utilization of carbon nanotubes increasing the electro-conducting properties of the

MCCM Template Vers 4 May 2014

 biosensor. In another report, Nameghi et al. [26] have proposed a simple electrochemical aptasensor for E2 in a blood sample. In this sensor platform, a split DNA aptamer was applied as recognizing agents with E2. The aptasensor was indicated a wide linear range with detection limits of 0.7 pM and 0.5 pM for E2 in milk samples and tap water, respectively.

 In this study, we developed a highly sensitive aptasensor based on PANI-MWCNT@Chs-APT composite for the electrochemical detection of estradiol valerate (EV) in real samples. The elaboration of the aptasensor involves a three-step modification process. First, we carried out the electrochemical polymerization of aniline (polyaniline (PANI)) on the gold electrode (AuE). Secondly, the MWCNT@Chs was prepared and dropped to the surface of electropolymerized gold to form MWCNT@Chs/PANI/AuE. This new membrane has been developed by Ying Ou et al.; the MWCNTs were functionalized by the polyelectrolyte Chs, via a noncovalent surface deposition method [27]. Finally, the free carboxyl groups on the MWCNT@Chs/PANI/AuE surface were conjugated 17 with the NH_2 -Aptamer to afford an aptasensor APT/MWCNT@Chs/PANI/AuE interface. Electrochemical impedance spectroscopy (EIS) was used for gold electrodes (AuE) characterization and it combined with a standard added method to evaluate the performance of

MCCM_Template_Vers 4 May 2014

18 shows the variation of R_{ct}^{f} as function of the EV concentrations for the

- aptamer concentrations. The calibration curves (Fig. 3) show a comparison
- between the relative slopes of three aptasensors; as it can be observed, when

17 (R_{ct}^f) as a function of EV concentrations is plotted in Fig. 4(B). The R_{ct}^f was 18 calculated according to the equation:

19
$$
R_{\rm ct}^{\rm f} = \frac{R_{\rm ct}^{\rm c} - R_{\rm ct}^{\rm c=0}}{R_{\rm ct}^{\rm c=0}}
$$
 (1)

1 where R_{ct}^{C} and $R_{\text{ct}}^{\text{C=0}}$ are the values of the charge transfer resistance before and after the incubation in EV concentrations. Figure 4(A) showed that the Nyquist plot of the interface aptasesor at 4 different concentrations of EV. The R_{CT} values were found to rise (from 12 to 51.5 kΩ) with increasing concentration of EV from 4×10^{-16} to 4×10^{-9} mg/cm³. This was due to the increased complexation by adsorption of EV onto the aptasensor interface; it was directly related to the good sensitivity of the modified electrode to EV. It could be seen in the calibration plot (Fig. 4(B)), a linear dependence 10 between the R_{ct}^{f} values and the logarithmic values of EV concentration. The calibration curve equation was obtained by Origin 9.1 version via plotting and analyzing: $R_{ct}^{f}/\Omega = 0.46$ Log ([EV] /mg cm⁻³) + 7.41 (2)

 The aptasensor had a linear response in the range of EV concentrations 15 between 4×10^{-16} to 4×10^{-9} mg/cm³ with a good correlation coefficient of 16 about 0.996 and a high sensitivity of 0.46 Ω per EV concentration decade. While the limit of detection was calculated to be 0.56 fM from the calibration curve by three times the standard deviation of the average value of the blank without adding the EV standard solution (n=5), divided by the sensitivity. This LOD was lower than those reported in the references [23, 31, 32, 34] and it was in good agreement with those attained by other sensors [29, 30,

36] (Table 1). This lower limit detection was attributed by Ke et al. [30];

1
$$
R_{ct}^f/\Omega = 6.473 \text{ Log ([EV] /mg cm}^3) + 79.21
$$
 (3)

 The concentration of EV in the unknown serum sample obtained by the 3 standard addition method was 43.15 pg/cm^3 .

 The concentration of this plasma sample was also tested using the immuno-electrochimiluminescene "ECLIA" (Cobas, Roche). The result 6 obtained was 41.25 pg/cm³. Both results were nearly the same with a recovery = 95 %. However, this result was in line with the estradiol levels determined in the blood samples of healthy women during menstruation 9 range from (18 to 147 pg/cm³). These results indicated that the developed aptasensor could be applied in real pharmaceutical applications.

-
- $12 \,$ \leq Figure 6 $>$
- $\langle \text{Table 1} \rangle$
-

Conclusion

 In this work, a new sensitive and selective aptasensor was developed and successfully applied to determine EV in pharmaceutical and real female blood samples. We showed that the APT/MWCNT@Chs/PANI/AuE aptasensor exhibited excellent performances with high sensitivity and 20 selectivity. A good linear relationship was also obtained between the R_{CT} values and the logarithmic values of EV concentration with a lower LOD of

 about 0.56 fM both in pharmaceutical and women's blood. Therefore, this work could provide a potential tool for the rapid detection of EV in real samples.

Experimental

 (Progynova®) was employed from Delpharm Lille S.A.S (France). Progesterone (P4) was obtained from Laboratoires BESINS International (France). Methylprednisolone (DCI) was produced from Valdepharm (France). 17β-Estradiol (E2), multi-walled carbon nanotube (MWNTs), chitosan (Chs, *75%* deacetylated), glutaraldehyde (GA) (grade II, 25% 11 aqueous solution), potassium ferricyanide $(K_3[Fe(CN)_6])$, potassium 12 ferrocyanide $(K_4[Fe(CN)_6])$, ammonia, phosphate buffer solution (pH 7.4), nitric acid, hydrogen peroxide solution, acetone, sulfuric acid, hydrogen peroxide, and ethanol were obtained from Sigma-Aldrich (France). Aniline and acetic acid were obtained from VWR International S.A.S (France). *N*- Hydroxysuccinimide (NHS) and *N*-(3-dimethylaminopropyl)-*N′*- ethylcarbodiimide (EDC) were received from Shanghai Yuanye BioTechnology Co., Ltd. (China). Gold substrates were provided by the French RENATECH network (LAAS, CNRS Toulouse). The single-strand DNA Aptaanti-oestradiol was chosen according to the prior reported literature [37]; then, was obtained from Eurogentec, the sequence of the 76-

128 mV and an amplitude of 10 mV. The obtained data EIS were fitted by

MCCM Template Vers 4 May 2014

 Randles equivalent circuit model. The validation of this model is obtained by comparing the experimental data to the theoretical EIS model using EC- Lab software with the Randomize + Simplex method. This device was then used to detect EV, both in pharmaceutical and blood women, these samples 5 were diluted in PBS at pH 7.4 within the range 4.0×10^{-16} to 4.0×10^{-9} mg/cm³. The impedance response was recorded by successive incubations for 30 min of interface aptasensor in real samples containing different concentrations of EV.

Elaboration of the aptasensor (APT/MWCNTs@Chs/PANI/AuE)

Electropolymerization of aniline on gold electrode

 Before modification, the gold electrodes were cleaned with acetone and sonicated in ultrapure water. Then, dipped into a piranha solution (3:1 14 mixture of concentrated H_2SO_4 with H_2O_2 for 4 min and washed with ethanol and ultrapure water. Next, the electrodes were electrochemically 16 cleaned in 0.5 M H_2SO_4 in the potential scanning between -0.5 V and 1.5 V for 5 min, followed by washing them with ultrapure water and drying under nitrogen. After that, the electrochemical polymerization of AuE electrode had been performed in 0.5 M sulfuric acid solution containing 0.1 M of aniline by potential cycling within the potential scan limits of -0.2 to 1.4 V

1 remove the adsorbed and uncross-linked Chs [27]. Then, 5 mm³ of Chs@MWCNTs dispersion was dropped by drop added to the surface of the electropolymerized gold electrode and dried at room temperature for 5 h.

Immobilization of aptamer

 Before the aptamer immobilization, NHS/EDC technique was used to 7 activate the –COOH groups of the aptamer. Briefly, 50 mm³ of aptamer (4.04) 8 mM) were mixed with 50 mm³ of the PBS solution containing (0.1 M 9 NHS/0.4 M EDC). 6 mm³ of the activated aptamer was then dropped on the MWCNT@Chs/PANI/AuE electrode and maintained for 24 h at room temperature for immobilization. Afterward, the remained active carboxylic acid groups were deactivated by incubating them in the ethanolamine solution (1% in PBS) at room temperature for 20 min. This stage is very important to prevent nonspecific bonding at the detection step [46]. The whole construction procedure of the modified aptasensor was displayed in Fig. 8 and the resulting electrode is called APT/MWCNT@Chs/PANI/AuE electrode.

-
- 19 \langle Figure 8 >
-

- 9. Lahcen AA, Baleg AA, Baker P, Iwuoha E, Amine A (2017) Sens Actuators B Chem 241:698
- 10. Li J, Jiang J, Zhao D, Xu Z, Liu M, Deng P, Liu X, Yang C, Qian D, Xie
- H (2018) J Alloys Compd 769:566
- 11. Wang Y, Luo J, Liu J, Li X, Kong Z, Jin H, Cai X (2018) Biosens Bioelectron 107:47
- 12. Wang A, Ding Y, Li L, Duan D, Mei Q, Zhuang Q, Cui S, He X (2019)
- Talanta 192:478
- 13. Mofidi Z, Norouzi P, Larijani B, Seidi S, Ganjali MR, Morshedi M
- (2018) J Electroanal Chem 813:83
- 14. Triviño JJ, Gómez M, Valenzuela J, Vera A, Arancibia V (2019) Sens Actuators B Chem 297:126728
- 15. Bergamin B, Pupin RR, Wong A, Sotomayor MD (2019) J Braz Chem Soc 30:2344
- 16. Liu M, Ke H, Sun C, Wang G, Wang Y, Zhao G (2019) Talanta 194:266
- 17. Raj M, Goyal RN (2019) Sens Actuators B Chem 284:759
- 18. Zhang G, Li T, Zhang J, Chen A (2018) Sens Actuators B Chem 273:1648
- 19. Antoniazzi C, de Lima CA, Marangoni R, Spinelli A, de Castro EG
- (2018) J Solid State Electrochem 22:1373

137:146

Figure Captions

 Fig. 6 (A) Nyquist impedance plots (Zr vs Zi) obtained from the standard addition method performed on a serum blood sample at 5 mM of PBS pH 7.4 solution. (B) Calibration curve used to determine the concentration of 5 EV in the unknown Serum sample by the standard addition method $(n=3)$, 6 $R^2 = 0.996$ **Fig. 7** Cyclic voltammograms of electrochemical polymerization of aniline on the gold electrode from the aqueous solution of 0.5 M of sulfuric acid 10 containing 0.1 M of aniline, as obtained at a potential scan rate of 100 mV s⁻ $11⁻¹$ within potential scan limits of -0.2 to 1.4 V **Fig. 8** Construction of the aptasensor APT/MWCNTs@Chs/PANI/AuE interface

1 **Table 1** Performance comparison of APT/MWCNT@Chs/PANI/AuE for

2 the detection of estradiol with other sensors

- 4 APT–ERGO/GCE: electrochemically reduced graphene oxide
5 DPV: differential pulse voltammetry
6 CuO/CPE : copper(II) oxide-modified carbon paste electrode
7 GN: graphite nanoparticle
8 SPR: surface plasmon resonance
9 DNI
- 5 DPV: differential pulse voltammetry
- 6 CuO/CPE : copper(II) oxide-modified carbon paste electrode
- GN: graphite nanoparticle
- 8 SPR: surface plasmon resonance
- 9 pNIPAm: poly(*N*-isopropylacrylamide) microgel-based etalons
10 SERS: surface-enhanced raman spectroscopy
- 10 SERS: surface-enhanced raman spectroscopy
11 Ru@ethyl-SiO₂: Ru(bpy)₃²⁺ Doped vinyl-SiO₂
- 11 Ru@ethyl-SiO₂: Ru(bpy)₃²⁺ Doped vinyl-SiO₂ [nanoparticles](https://www.sciencedirect.com/topics/chemistry/nanoparticle) (Ru@ethyl-SiO₂) and E2 imprinted polypyrrole (MIP)
- 12 [polypyrrole](https://www.sciencedirect.com/topics/chemistry/polypyrrole) (MIP)
13 ECL: electrochemi
- 13 ECL: electrochemiluminescence
14 GOx/APT/AuNPs/CuS: glucose
- 14 GOx/APT/AuNPs/CuS: glucose oxidase / aptamer / gold nanoparticles / copper sulfide
-
- 15 nanosheets
16 Au@Ag CS
- 16 Au@Ag CS NPs: gold-silver core-shell nanoparticles
17 Anti-E2/ITO/ErG/AuNP: anti estradiol antibody/indiu Anti-E2/ ITO/ErG/AuNP: anti estradiol antibody/indium tin oxide/graphene/gold nanoparticle.
- 18

- 2
- 3

4 *Figure 8*

